

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090821\
 Data File : BP006887.D
 Acq On : 08 Sep 2021 17:32
 Operator : CG/JU
 Sample : SST04018
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040618

Quant Time: Sep 09 02:00:23 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP090821.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 09 01:57:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	268978	20.000	ng/u1	0.00
20) Naphthalene-d8	10.740	136	1088140	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.563	164	692049	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.316	188	1462709	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.398	240	1487062	20.000	ng/u1	0.00
88) Perylene-d12	23.833	264	1574402	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	107695	13.542	ng/uL	0.00
4) Pyridine-d5	3.781	84	739546	32.350	ng/u1	0.00
7) Phenol-d5	7.099	99	814565	30.124	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.269	67	471570	28.377	ng/u1	0.00
11) 2-Chlorophenol-d4	7.469	132	675340	33.534	ng/u1	0.00
15) 4-Methylphenol-d8	8.640	113	648165	30.004	ng/u1	0.00
21) Nitrobenzene-d5	9.099	128	290384	31.438	ng/u1	0.00
24) 2-Nitrophenol-d4	9.822	143	292382	30.251	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.357	165	663197	40.613	ng/u1	0.00
31) 4-Chloroaniline-d4	10.875	131	889269	33.505	ng/u1	0.00
46) Dimethylphthalate-d6	13.975	166	1889285	35.598	ng/u1	0.00
49) Acenaphthylene-d8	14.257	160	2404291	36.933	ng/u1	0.00
54) 4-Nitrophenol-d4	14.751	143	335040	30.222	ng/u1	0.00
60) Fluorene-d10	15.557	176	1633254	37.838	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.669	200	270260	27.953	ng/u1	0.00
73) Anthracene-d10	17.416	188	2526337	36.504	ng/u1	0.00
81) Pyrene-d10	19.645	212	2895710	37.081	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.680	264	3216508	37.886	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.399	88	119810	14.828	ng/uL	89
5) Pyridine	3.799	79	730394	30.863	ng/u1	88
6) Benzaldehyde	7.075	77	528657	39.763	ng/u1	90
8) Phenol	7.122	94	847921	30.704	ng/u1	90
10) Bis(2-Chloroethyl)ether	7.363	93	677236	31.817	ng/u1	96
12) 2-Chlorophenol	7.499	128	690827	34.019	ng/u1	92
13) 2-Methylphenol	8.375	108	647453	31.865	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.475	45	789098	31.728	ng/u1	97
16) Acetophenone	8.763	105	1000553	29.422	ng/u1	90
17) N-Nitroso-di-n-propyla...	8.752	70	476016	26.023	ng/u1#	96
18) 4-Methylphenol	8.704	108	700757	31.723	ng/u1	97
19) Hexachloroethane	9.022	117	272998	30.893	ng/u1#	80
22) Nitrobenzene	9.140	77	686780	28.400	ng/u1#	88
23) Isophorone	9.669	82	1371054	30.404	ng/u1#	94
25) 2-Nitrophenol	9.851	139	334386	32.802	ng/u1#	84
26) 2,4-Dimethylphenol	9.910	107	748520	34.180	ng/u1	93
27) Bis(2-Chloroethoxy)met...	10.151	93	880519	33.685	ng/u1	98
29) 2,4-Dichlorophenol	10.387	162	646980	40.839	ng/u1#	94
30) Naphthalene	10.793	128	2159885	36.574	ng/u1	100
32) 4-Chloroaniline	10.899	127	908966	34.901	ng/u1	100
33) Hexachlorobutadiene	11.087	225	427756	45.796	ng/u1	97
34) Caprolactam	11.669	113	89057	14.041	ng/u1	94
35) 4-Chloro-3-methylphenol	12.016	107	673463	33.232	ng/u1	89
36) 2-Methylnaphthalene	12.398	142	1457451	35.915	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090821\
 Data File : BP006887.D
 Acq On : 08 Sep 2021 17:32
 Operator : CG/JU
 Sample : SST04018
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040618

Quant Time: Sep 09 02:00:23 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP090821.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 09 01:57:53 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.616	142	1493107	36.972	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.763	216	815335	43.717	ng/ul#	95
40) Hexachlorocyclopentadiene	12.751	237	524922	42.901	ng/ul	99
41) 2,4,6-Trichlorophenol	12.998	196	504409	39.382	ng/ul	99
42) 2,4,5-Trichlorophenol	13.069	196	551871	40.470	ng/ul	98
43) 1,1'-Biphenyl	13.404	154	1974397	36.699	ng/ul#	98
44) 2-Chloronaphthalene	13.445	162	1522384	37.298	ng/ul	92
45) 2-Nitroaniline	13.645	65	351038	23.850	ng/ul#	75
47) Dimethylphthalate	14.022	163	1883250	35.875	ng/ul	99
48) 2,6-Dinitrotoluene	14.140	165	356480	35.184	ng/ul#	75
50) Acenaphthylene	14.287	152	2470936	38.828	ng/ul	99
51) 3-Nitroaniline	14.463	138	369372	31.834	ng/ul	86
52) Acenaphthene	14.628	153	1603723	35.343	ng/ul	98
53) 2,4-Dinitrophenol	14.669	184	174335	26.603	ng/ul#	68
55) 4-Nitrophenol	14.763	109	246769	24.823	ng/ul#	85
56) Dibenzofuran	14.963	168	2314965	38.156	ng/ul#	86
57) 2,4-Dinitrotoluene	14.922	165	504289	33.832	ng/ul#	76
58) 2,3,4,6-Tetrachlorophenol	15.187	232	451689	39.902	ng/ul#	77
59) Diethylphthalate	15.387	149	1839724	32.971	ng/ul#	94
61) Fluorene	15.610	166	1822851	36.033	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.610	204	930157	40.615	ng/ul#	80
63) 4-Nitroaniline	15.628	138	375688	34.577	ng/ul#	83
66) 4,6-Dinitro-2-methylph...	15.686	198	278292	28.794	ng/ul#	74
67) N-Nitrosodiphenylamine	15.822	169	1602213	35.661	ng/ul	99
68) 4-Bromophenyl-phenylether	16.504	248	577345	40.021	ng/ul#	83
69) Hexachlorobenzene	16.616	284	663894	40.722	ng/ul#	87
70) Atrazine	16.775	200	613413	38.486	ng/ul	94
71) Pentachlorophenol	16.957	266	382100	35.781	ng/ul	98
72) Phenanthrene	17.357	178	2980879	36.074	ng/ul	99
74) Anthracene	17.451	178	3016046	35.943	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.369	216	847074	42.588	ng/ul	99
76) Pentachlorobenzene	14.887	250	817184	41.699	ng/ul	99
77) Carbazole	17.716	167	2736841	34.920	ng/ul	97
78) Di-n-butylphthalate	18.269	149	3150871	31.094	ng/ul#	98
80) Fluoranthene	19.322	202	3600018	36.678	ng/ul	94
82) Pyrene	19.675	202	3657490	36.078	ng/ul#	93
83) Butylbenzylphthalate	20.545	149	1394335	27.820	ng/ul#	88
84) 3,3'-Dichlorobenzidine	21.316	252	1321411	38.941	ng/ul#	98
85) Benzo(a)anthracene	21.380	228	3605933	37.228	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.310	149	2046612	26.500	ng/ul#	93
87) Chrysene	21.433	228	3523689	37.732	ng/ul	100
89) Di-n-octyl phthalate	22.245	149	3556557	26.133	ng/ul	100
90) Benzo(b)fluoranthene	23.086	252	3950254	38.624	ng/ul	97
91) Benzo(k)fluoranthene	23.139	252	3510331	35.451	ng/ul	98
93) Benzo(a)pyrene	23.727	252	3738064	40.346	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	26.398	276	4365608	36.037	ng/ul	99
95) Dibenzo(a,h)anthracene	26.415	278	3763872	36.989	ng/ul	97
96) Benzo(g,h,i)perylene	27.180	276	3702839	36.814	ng/ul	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090821\
 Data File : BP006887.D
 Acq On : 08 Sep 2021 17:32
 Operator : CG/JU
 Sample : SSTD040618
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD040618

Quant Time: Sep 09 02:00:23 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP090821.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 09 01:57:53 2021
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 14.828 ng/uL
RT: 3.399 min Scan# 53
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion: 88 Resp: 119810
Ion Ratio Lower Upper
88 100
43 25.5 20.1 30.1
58 71.9 48.4 72.6



#5
Pyridine
Concen: 30.863 ng/uL
RT: 3.799 min Scan# 121
Delta R.T. -0.006 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion: 79 Resp: 730394
Ion Ratio Lower Upper
79 100
52 72.5 49.2 73.8
51 32.4 22.7 34.1









#6
Benzaldehyde
Concen: 39.763 ng/ul
RT: 7.075 min Scan# 678
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion: 77 Resp: 528657
Ion Ratio Lower Upper
77 100
105 101.5 74.4 111.6
106 97.8 70.3 105.5



#8
Phenol
Concen: 30.704 ng/ul
RT: 7.122 min Scan# 686
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion: 94 Resp: 847921
Ion Ratio Lower Upper
94 100
65 25.5 24.4 36.6
66 33.0 31.8 47.6









#10

Bis(2-Chloroethyl)ether
 Concen: 31.817 ng/ul
 RT: 7.363 min Scan# 727
 Delta R.T. 0.000 min
 Lab File: BP006887.D
 Acq: 08 Sep 2021 17:32

Instrument :
 BNA_P
 ClientSampleId :
 SSTD040618




Tgt Ion: 93 Resp: 677236
 Ion Ratio Lower Upper
 93 100
 63 70.3 59.7 89.5
 95 32.9 25.3 37.9




#12

2-Chlorophenol
 Concen: 34.019 ng/ul
 RT: 7.499 min Scan# 750
 Delta R.T. 0.000 min
 Lab File: BP006887.D
 Acq: 08 Sep 2021 17:32

Tgt Ion: 128 Resp: 690827
 Ion Ratio Lower Upper
 128 100
 64 43.0 40.6 60.8
 130 31.6 26.4 39.6









#13
2-Methylphenol
Concen: 31.865 ng/ul
RT: 8.375 min Scan# 899
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:108 Resp: 647453
Ion Ratio Lower Upper
108 100
107 88.2 72.2 108.2



#14
2,2'-oxybis(1-Chloropropane)
Concen: 31.728 ng/ul
RT: 8.475 min Scan# 916
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion: 45 Resp: 789098
Ion Ratio Lower Upper
45 100
77 16.2 13.8 20.8
79 12.6 10.9 16.3









#16
Acetophenone
Concen: 29.422 ng/ul
RT: 8.763 min Scan# 965
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:105 Resp: 1000553
Ion Ratio Lower Upper
105 100
77 72.3 67.0 100.4
51 23.2 19.0 28.6



#17
N-Nitroso-di-n-propylamine
Concen: 26.023 ng/ul
RT: 8.752 min Scan# 963
Delta R.T. 0.006 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion: 70 Resp: 476016
Ion Ratio Lower Upper
70 100
42 49.8 40.4 60.6
101 10.5 7.2 10.8
130 23.5 14.9 22.3#









#18
4-Methylphenol
Concen: 31.723 ng/ul
RT: 8.704 min Scan# 955
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:108 Resp: 700757
Ion Ratio Lower Upper
108 100
107 114.1 88.6 133.0



#19
Hexachloroethane
Concen: 30.893 ng/ul
RT: 9.022 min Scan# 1009
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:117 Resp: 272998
Ion Ratio Lower Upper
117 100
201 106.0 68.2 102.2#
199 65.6 43.5 65.3#









#22
Nitrobenzene
Concen: 28.400 ng/ul
RT: 9.140 min Scan# 1029
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion: 77 Resp: 686780
Ion Ratio Lower Upper
77 100
123 48.1 31.1 46.7#
65 12.8 10.6 16.0



#23
Isophorone
Concen: 30.404 ng/ul
RT: 9.669 min Scan# 1119
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion: 82 Resp: 1371054
Ion Ratio Lower Upper
82 100
95 6.9 5.8 8.8
138 18.9 12.5 18.7#









#25
2-Nitrophenol
Concen: 32.802 ng/ul
RT: 9.851 min Scan# 1150
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:139 Resp: 334386
Ion Ratio Lower Upper
139 100
65 46.2 48.4 72.6#
109 29.1 28.2 42.2



#26
2,4-Dimethylphenol
Concen: 34.180 ng/ul
RT: 9.910 min Scan# 1160
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:107 Resp: 748520
Ion Ratio Lower Upper
107 100
121 51.3 38.1 57.1
122 86.7 64.2 96.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 33.685 ng/ul
 RT: 10.151 min Scan# 1201
 Delta R.T. 0.000 min
 Lab File: BP006887.D
 Acq: 08 Sep 2021 17:32




Instrument :
 BNA_P
 ClientSampleId :
 SSTD040618


Tgt Ion:	93	Resp:	880519
Ion	Ratio	Lower	Upper
93	100		
95	32.4	25.6	38.4
123	11.8	8.2	12.4



#29
 2,4-Dichlorophenol
 Concen: 40.839 ng/ul
 RT: 10.387 min Scan# 1241
 Delta R.T. 0.006 min
 Lab File: BP006887.D
 Acq: 08 Sep 2021 17:32

Tgt Ion:	162	Resp:	646980
Ion	Ratio	Lower	Upper
162	100		
164	65.7	52.0	78.0
98	33.7	34.3	51.5#









#30
Naphthalene
Concen: 36.574 ng/ul
RT: 10.793 min Scan# 1310
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:128 Resp: 2159885
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 13.1 10.6 15.8



#32
4-Chloroaniline
Concen: 34.901 ng/ul
RT: 10.899 min Scan# 1328
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:127 Resp: 908966
Ion Ratio Lower Upper
127 100
129 32.2 25.7 38.5









#33
Hexachlorobutadiene
Concen: 45.796 ng/ul
RT: 11.087 min Scan# 1360
Delta R.T. 0.006 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:225 Resp: 427756
Ion Ratio Lower Upper
225 100
223 63.1 48.2 72.2
227 64.3 49.6 74.4



#34
Caprolactam
Concen: 14.041 ng/ul
RT: 11.669 min Scan# 1459
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:113 Resp: 89057
Ion Ratio Lower Upper
113 100
55 150.6 125.8 188.6
56 115.1 98.2 147.2









#35
4-Chloro-3-methylphenol
Concen: 33.232 ng/ul
RT: 12.016 min Scan# 1518
Delta R.T. 0.006 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:107 Resp: 673463
Ion Ratio Lower Upper
107 100
144 26.7 18.8 28.2
142 83.2 58.4 87.6



#36
2-Methylnaphthalene
Concen: 35.915 ng/ul
RT: 12.398 min Scan# 1583
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:142 Resp: 1457451
Ion Ratio Lower Upper
142 100
141 89.5 72.3 108.5









#37
1-Methylnaphthalene
Concen: 36.972 ng/ul
RT: 12.616 min Scan# 1620
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:142 Resp: 1493107
Ion Ratio Lower Upper
142 100
141 91.5 74.9 112.3
116 3.6 3.3 4.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 43.717 ng/ul
RT: 12.763 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:216 Resp: 815335
Ion Ratio Lower Upper
216 100
214 79.1 64.3 96.5
179 20.5 19.3 28.9
108 16.6 19.8 29.6#









#40
Hexachlorocyclopentadiene
Concen: 42.901 ng/ul
RT: 12.751 min Scan# 1643
Delta R.T. 0.006 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:237 Resp: 524922
Ion Ratio Lower Upper
237 100
235 62.5 50.4 75.6
272 13.0 9.7 14.5



#41
2,4,6-Trichlorophenol
Concen: 39.382 ng/ul
RT: 12.998 min Scan# 1685
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:196 Resp: 504409
Ion Ratio Lower Upper
196 100
198 94.5 76.2 114.4
200 31.2 24.6 36.8









#42
2,4,5-Trichlorophenol
Concen: 40.470 ng/ul
RT: 13.069 min Scan# 1697
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:196	Resp: 551871
Ion Ratio	Lower Upper
196	100
198	96.7 79.3 118.9
200	31.3 25.8 38.8



#43
1,1'-Biphenyl
Concen: 36.699 ng/ul
RT: 13.404 min Scan# 1754
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:154	Resp: 1974397
Ion Ratio	Lower Upper
154	100
153	40.4 32.4 48.6
76	12.5 12.9 19.3#









#44
2-Chloronaphthalene
Concen: 37.298 ng/ul
RT: 13.445 min Scan# 1761
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:162 Resp: 1522384
Ion Ratio Lower Upper
162 100
164 33.3 25.6 38.4
127 37.1 36.2 54.2



#45
2-Nitroaniline
Concen: 23.850 ng/ul
RT: 13.645 min Scan# 1795
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion: 65 Resp: 351038
Ion Ratio Lower Upper
65 100
92 75.7 53.7 80.5
138 123.0 71.4 107.2#









#47
Dimethylphthalate
Concen: 35.875 ng/ul
RT: 14.022 min Scan# 1859
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:163 Resp: 1883250
Ion Ratio Lower Upper
163 100
194 4.1 3.0 4.4
164 10.1 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 35.184 ng/ul
RT: 14.140 min Scan# 1879
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:165 Resp: 356480
Ion Ratio Lower Upper
165 100
89 51.7 62.2 93.2#
121 21.0 22.1 33.1#









#50
Acenaphthylene
Concen: 38.828 ng/ul
RT: 14.287 min Scan# 1904
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:152 Resp: 2470936
Ion Ratio Lower Upper
152 100
151 20.1 15.9 23.9
153 13.5 10.4 15.6



#51
3-Nitroaniline
Concen: 31.834 ng/ul
RT: 14.463 min Scan# 1934
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:138 Resp: 369372
Ion Ratio Lower Upper
138 100
108 13.7 9.7 14.5
92 115.4 106.7 160.1









#52
Acenaphthene
Concen: 35.343 ng/ul
RT: 14.628 min Scan# 1962
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:153 Resp: 1603723
Ion Ratio Lower Upper
153 100
152 46.9 38.6 57.8
154 88.0 69.0 103.6



#53
2,4-Dinitrophenol
Concen: 26.603 ng/ul
RT: 14.669 min Scan# 1969
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:184 Resp: 174335
Ion Ratio Lower Upper
184 100
63 52.6 73.4 110.2#
154 56.4 59.4 89.0#









#55
4-Nitrophenol
Concen: 24.823 ng/ul
RT: 14.763 min Scan# 1985
Delta R.T. 0.006 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:109 Resp: 246769
Ion Ratio Lower Upper
109 100
139 146.9 90.3 135.5#
65 128.4 102.2 153.2



#56
Dibenzofuran
Concen: 38.156 ng/ul
RT: 14.963 min Scan# 2019
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:168 Resp: 2314965
Ion Ratio Lower Upper
168 100
139 35.0 35.3 52.9#









#57
2,4-Dinitrotoluene
Concen: 33.832 ng/ul
RT: 14.922 min Scan# 2012
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:165 Resp: 504289
Ion Ratio Lower Upper
165 100
63 38.4 45.2 67.8#
182 2.9 1.9 2.9#



#58
2,3,4,6-Tetrachlorophenol
Concen: 39.902 ng/ul
RT: 15.187 min Scan# 2057
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:232 Resp: 451689
Ion Ratio Lower Upper
232 100
131 41.4 51.8 77.6#
130 2.2 3.1 4.7#
166 29.8 31.0 46.4#









#59
Diethylphthalate
Concen: 32.971 ng/ul
RT: 15.387 min Scan# 2091
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:149 Resp: 1839724
Ion Ratio Lower Upper
149 100
177 23.3 15.5 23.3#
150 12.8 9.9 14.9



#61
Fluorene
Concen: 36.033 ng/ul
RT: 15.610 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:166 Resp: 1822851
Ion Ratio Lower Upper
166 100
165 97.5 79.6 119.4
167 13.5 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 40.615 ng/ul
RT: 15.610 min Scan# 2129
Delta R.T. 0.006 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:204 Resp: 930157
Ion Ratio Lower Upper
204 100
206 33.1 26.3 39.5
141 53.1 62.0 93.0#



#63
4-Nitroaniline
Concen: 34.577 ng/ul
RT: 15.628 min Scan# 2132
Delta R.T. 0.006 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:138 Resp: 375688
Ion Ratio Lower Upper
138 100
92 51.0 47.1 70.7
108 72.7 73.8 110.8#









#66
4,6-Dinitro-2-methylphenol
Concen: 28.794 ng/ul
RT: 15.686 min Scan# 2142
Delta R.T. 0.006 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:198 Resp: 278292
Ion Ratio Lower Upper
198 100
182 3.9 4.5 6.7#
77 21.6 31.7 47.5#



#67
N-Nitrosodiphenylamine
Concen: 35.661 ng/ul
RT: 15.822 min Scan# 2165
Delta R.T. 0.006 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:169 Resp: 1602213
Ion Ratio Lower Upper
169 100
168 66.6 53.9 80.9
167 36.4 30.0 45.0









#68
4-Bromophenyl-phenylether
Concen: 40.021 ng/ul
RT: 16.504 min Scan# 2281
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:248 Resp: 577345
Ion Ratio Lower Upper
248 100
250 98.6 78.4 117.6
141 66.9 80.6 121.0#



#69
Hexachlorobenzene
Concen: 40.722 ng/ul
RT: 16.616 min Scan# 2300
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:284 Resp: 663894
Ion Ratio Lower Upper
284 100
142 34.3 37.8 56.6#
249 31.9 27.1 40.7









#70
Atrazine
Concen: 38.486 ng/ul
RT: 16.775 min Scan# 2327
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:200 Resp: 613413
Ion Ratio Lower Upper
200 100
173 25.9 22.2 33.4
215 49.4 35.4 53.0



#71
Pentachlorophenol
Concen: 35.781 ng/ul
RT: 16.957 min Scan# 2358
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:266 Resp: 382100
Ion Ratio Lower Upper
266 100
264 64.0 51.0 76.4
268 64.2 49.4 74.2









#72
Phenanthrene
Concen: 36.074 ng/ul
RT: 17.357 min Scan# 2426
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:178 Resp: 2980879
Ion Ratio Lower Upper
178 100
179 15.9 12.0 18.0
176 19.3 15.6 23.4



#74
Anthracene
Concen: 35.943 ng/ul
RT: 17.451 min Scan# 2442
Delta R.T. 0.006 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:178 Resp: 3016046
Ion Ratio Lower Upper
178 100
179 15.6 12.1 18.1
176 18.7 14.9 22.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 42.588 ng/uL
RT: 13.369 min Scan# 1748
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:216 Resp: 847074
Ion Ratio Lower Upper
216 100
214 79.0 62.9 94.3
218 48.5 37.8 56.8



#76
Pentachlorobenzene
Concen: 41.699 ng/uL
RT: 14.887 min Scan# 2006
Delta R.T. 0.006 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:250 Resp: 817184
Ion Ratio Lower Upper
250 100
248 63.4 51.0 76.6
252 64.4 51.0 76.4









#77
 Carbazole
 Concen: 34.920 ng/ul
 RT: 17.716 min Scan# 2487
 Delta R.T. 0.000 min
 Lab File: BP006887.D
 Acq: 08 Sep 2021 17:32

Instrument :
 BNA_P
 ClientSampleId :
 SSTD040618

Tgt Ion:167 Resp: 2736841




Ion	Ratio	Lower	Upper
167	100		
166	22.0	17.6	26.4
139	12.2	11.9	17.9




#78
 Di-n-butylphthalate
 Concen: 31.094 ng/ul
 RT: 18.269 min Scan# 2581
 Delta R.T. 0.000 min
 Lab File: BP006887.D
 Acq: 08 Sep 2021 17:32

Tgt Ion:149 Resp: 3150871

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.3	10.9
104	4.7	4.8	7.2#









#80
Fluoranthene
Concen: 36.678 ng/ul
RT: 19.322 min Scan# 2760
Delta R.T. 0.006 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:202 Resp: 3600018
Ion Ratio Lower Upper
202 100
101 11.0 11.0 16.4
100 8.5 8.3 12.5



#82
Pyrene
Concen: 36.078 ng/ul
RT: 19.675 min Scan# 2820
Delta R.T. 0.006 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:202 Resp: 3657490
Ion Ratio Lower Upper
202 100
101 12.7 12.6 19.0
100 10.4 10.6 15.8#









#83
Butylbenzylphthalate
Concen: 27.820 ng/ul
RT: 20.545 min Scan# 2968
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:149 Resp: 1394335
Ion Ratio Lower Upper
149 100
91 68.0 63.0 94.6
206 22.4 14.1 21.1#



#84
3,3'-Dichlorobenzidine
Concen: 38.941 ng/ul
RT: 21.316 min Scan# 3099
Delta R.T. 0.006 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:252 Resp: 1321411
Ion Ratio Lower Upper
252 100
254 63.9 51.6 77.4
126 9.8 10.7 16.1#









#85
Benzo(a)anthracene
Concen: 37.228 ng/ul
RT: 21.380 min Scan# 3110
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:228 Resp: 3605933
Ion Ratio Lower Upper
228 100
229 19.9 15.8 23.8
226 26.9 21.6 32.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 26.500 ng/ul
RT: 21.310 min Scan# 3098
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:149 Resp: 2046612
Ion Ratio Lower Upper
149 100
167 28.9 20.2 30.2
279 4.8 2.7 4.1#









#87
Chrysene
Concen: 37.732 ng/ul
RT: 21.433 min Scan# 3119
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:228 Resp: 3523689
Ion Ratio Lower Upper
228 100
226 30.7 24.5 36.7
229 20.1 15.9 23.9



#89
Di-n-octyl phthalate
Concen: 26.133 ng/ul
RT: 22.245 min Scan# 3257
Delta R.T. -0.006 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:149 Resp: 3556557









#90
Benzo(b)fluoranthene
Concen: 38.624 ng/ul
RT: 23.086 min Scan# 3400
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:252 Resp: 3950254
Ion Ratio Lower Upper
252 100
253 22.4 17.7 26.5
125 10.2 10.1 15.1



#91
Benzo(k)fluoranthene
Concen: 35.451 ng/ul
RT: 23.139 min Scan# 3409
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:252 Resp: 3510331
Ion Ratio Lower Upper
252 100
253 22.1 17.2 25.8
125 10.4 9.4 14.0









#93
Benzo(a)pyrene
Concen: 40.346 ng/ul
RT: 23.727 min Scan# 3509
Delta R.T. 0.000 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32




Instrument :
BNA_P
ClientSampleId :
SSTD040618


Tgt Ion:252 Resp: 3738064
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 11.0 11.3 16.9#



#94
Indeno(1,2,3-cd)pyrene
Concen: 36.037 ng/ul
RT: 26.398 min Scan# 3963
Delta R.T. 0.012 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:276 Resp: 4365608
Ion Ratio Lower Upper
276 100
138 25.9 21.8 32.8
277 24.9 19.9 29.9









#95
Dibenzo(a,h)anthracene
Concen: 36.989 ng/ul
RT: 26.415 min Scan# 3966
Delta R.T. 0.012 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Instrument :
BNA_P
ClientSampleId :
SSTD040618

Tgt Ion:278 Resp: 3763872
Ion Ratio Lower Upper
278 100
139 16.4 15.4 23.2
279 23.7 19.0 28.4



#96
Benzo(g,h,i)perylene
Concen: 36.814 ng/ul
RT: 27.180 min Scan# 4096
Delta R.T. 0.012 min
Lab File: BP006887.D
Acq: 08 Sep 2021 17:32

Tgt Ion:276 Resp: 3702839
Ion Ratio Lower Upper
276 100
138 22.2 20.7 31.1
277 23.9 19.0 28.4

