




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090822\
 Data File : BP011652.D
 Acq On : 08 Sep 2022 10:32
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 09 06:02:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP090222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 03 04:07:09 2022
 Response via : Initial Calibration



TIC: BP011652.D\data.ms

(70) Pentachlorophenol (C)




17.004min (-0.006) 204262.90 ng

response 552559

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.10	63.13
264.00	63.00	63.01
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090822\
 Data File : BP011652.D
 Acq On : 08 Sep 2022 10:32
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 09 06:02:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP090222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 03 04:07:09 2022
 Response via : Initial Calibration



TIC: BP011652.D\data.ms

(77) Benzidine

19.545min (-0.000) 756843.26 ng

response 2907739

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.40	14.03
183.00	11.30	11.50
0.00	0.00	0.00