

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091520\
 Data File : BP003288.D
 Acq On : 15 Sep 2020 14:10
 Operator : CG/JU
 Sample : SSTDICC10
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010

Quant Time: Sep 15 16:27:50 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 15 16:17:32 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.622	152	74969	20.00	ng	0.00
21) Naphthalene-d8	10.404	136	302169	20.00	ng	0.00
39) Acenaphthene-d10	14.269	164	201333	20.00	ng	0.00
64) Phenanthrene-d10	17.028	188	456100	20.00	ng	0.00
76) Chrysene-d12	21.127	240	537623	20.00	ng	0.00
86) Perylene-d12	23.351	264	586463	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.240	112	99032	23.36	ng	0.00
7) Phenol-d6	6.822	99	131542	24.75	ng	0.00
23) Nitrobenzene-d5	8.775	82	113901	27.17	ng	0.00
42) 2,4,6-Tribromophenol	15.775	330	66979	24.60	ng	0.00
45) 2-Fluorobiphenyl	12.887	172	315237	22.93	ng	0.00
79) Terphenyl-d14	19.604	244	588494	24.09	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.164	88	20238	12.48	ng	96
3) Pyridine	3.552	79	52279	12.25	ng	92
4) n-Nitrosodimethylamine	3.470	42	15964	14.33	ng	85
6) Aniline	6.958	93	73789	12.75	ng	98
8) 2-Chlorophenol	7.199	128	55127	11.66	ng	95
9) Benzaldehyde	6.769	77	38286	15.45	ng	90
10) Phenol	6.846	94	62164	12.62	ng	90
11) bis(2-Chloroethyl)ether	7.052	93	48997	12.57	ng	99
12) 1,3-Dichlorobenzene	7.516	146	66586	11.58	ng	# 96
13) 1,4-Dichlorobenzene	7.658	146	66151	11.40	ng	96
14) 1,2-Dichlorobenzene	7.975	146	64958	11.81	ng	96
15) Benzyl Alcohol	7.869	79	41865	13.98	ng	91
16) 2,2'-oxybis(1-Chloropr...	8.163	45	36386	11.01	ng	83
17) 2-Methylphenol	8.087	107	44947	12.81	ng	95
18) Hexachloroethane	8.699	117	24582	12.95	ng	97
19) n-Nitroso-di-n-propyla...	8.428	70	35790	14.90	ng	97
20) 3+4-Methylphenols	8.422	107	59809	12.65	ng	# 88
22) Acetophenone	8.440	105	82413	12.25	ng	# 95
24) Nitrobenzene	8.816	77	56453	13.71	ng	89
25) Isophorone	9.340	82	103749	14.03	ng	93
26) 2-Nitrophenol	9.528	139	29350	11.96	ng	# 85
27) 2,4-Dimethylphenol	9.610	122	43623	11.83	ng	90
28) bis(2-Chloroethoxy)met...	9.828	93	70305	12.76	ng	98
29) 2,4-Dichlorophenol	10.087	162	52865	11.55	ng	96
30) 1,2,4-Trichlorobenzene	10.275	180	65091	11.95	ng	99
31) Naphthalene	10.451	128	177710	11.45	ng	99
32) Benzoic acid	9.752	122	12867	5.37	ng	94
33) 4-Chloroaniline	10.575	127	74132	11.43	ng	# 92
34) Hexachlorobutadiene	10.751	225	44577	13.63	ng	99
35) Caprolactam	11.328	113	17480	12.71	ng	95
36) 4-Chloro-3-methylphenol	11.728	107	59125	13.84	ng	89
37) 2-Methylnaphthalene	12.081	142	129732	11.68	ng	97
38) 1-Methylnaphthalene	12.298	142	123912	11.74	ng	98
40) 1,2,4,5-Tetrachloroben...	12.457	216	72098	12.08	ng	100
41) Hexachlorocyclopentadiene	12.440	237	6722	2.38	ng	97
43) 2,4,6-Trichlorophenol	12.710	196	45064	11.48	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091520\
 Data File : BP003288.D
 Acq On : 15 Sep 2020 14:10
 Operator : CG/JU
 Sample : SSTDICC10
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010


Quant Time: Sep 15 16:27:50 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 15 16:17:32 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.798	196	47707	11.49	ng	97
46) 1,1'-Biphenyl	13.098	154	169041	11.19	ng	98
47) 2-Chloronaphthalene	13.140	162	139559	11.29	ng	97
48) 2-Nitroaniline	13.351	65	33525	14.48	ng #	79
49) Acenaphthylene	13.992	152	214250	11.46	ng	99
50) Dimethylphthalate	13.734	163	182101	11.86	ng	100
51) 2,6-Dinitrotoluene	13.851	165	37171	11.65	ng #	85
52) Acenaphthene	14.334	154	131093	11.42	ng	99
53) 3-Nitroaniline	14.187	138	39818	11.01	ng #	82
54) 2,4-Dinitrophenol	14.428	184	8340	5.85	ng	99
55) Dibenzofuran	14.675	168	210325	11.66	ng	97
56) 4-Nitrophenol	14.551	139	17573	6.76	ng	83
57) 2,4-Dinitrotoluene	14.651	165	51901	12.10	ng #	83
58) Fluorene	15.328	166	170662	12.28	ng	98
59) 2,3,4,6-Tetrachlorophenol	14.916	232	43503	11.59	ng	97
60) Diethylphthalate	15.104	149	187158	12.42	ng	98
61) 4-Chlorophenyl-phenyle...	15.322	204	91968	12.82	ng	96
62) 4-Nitroaniline	15.357	138	41923	11.28	ng	83
63) Azobenzene	15.616	77	142897	14.29	ng	93
65) 4,6-Dinitro-2-methylph...	15.434	198	19097	8.76	ng	99
66) n-Nitrosodiphenylamine	15.539	169	153252	11.49	ng	99
67) 4-Bromophenyl-phenylether	16.222	248	60806	12.08	ng	97
68) Hexachlorobenzene	16.339	284	73012	12.18	ng	98
69) Atrazine	16.498	200	59172	13.16	ng	98
70) Pentachlorophenol	16.704	266	16728	5.84	ng	99
71) Phenanthrene	17.069	178	282812	11.44	ng	99
72) Anthracene	17.163	178	274956	11.62	ng	99
73) Carbazole	17.439	167	258497	11.39	ng	99
74) Di-n-butylphthalate	18.004	149	337935	12.66	ng	99
75) Fluoranthene	19.051	202	351687	12.31	ng	99
77) Benzidine	19.239	184	173448	14.12	ng	98
78) Pyrene	19.404	202	366089	10.49	ng	99
80) Butylbenzylphthalate	20.286	149	165714	11.53	ng	95
81) Benzo(a)anthracene	21.110	228	383836	11.40	ng	99
82) 3,3'-Dichlorobenzidine	21.051	252	148956	12.02	ng	98
83) Chrysene	21.163	228	372754	11.56	ng	99
84) Bis(2-ethylhexyl)phtha...	21.051	149	244178	11.77	ng	98
85) Di-n-octyl phthalate	21.916	149	433604	12.19	ng	99
87) Indeno(1,2,3-cd)pyrene	25.615	276	491926	11.25	ng	98
88) Benzo(b)fluoranthene	22.680	252	408865	11.22	ng	99
89) Benzo(k)fluoranthene	22.721	252	391593	11.24	ng	99
90) Benzo(a)pyrene	23.251	252	383059	11.33	ng	98
91) Dibenzo(a,h)anthracene	25.627	278	405838	11.30	ng	98
92) Benzo(g,h,i)perylene	26.304	276	395198	10.96	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 15 16:27:50 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 15 16:17:32 2020
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 7.622 min Scan# 805
 Delta R.T. 0.000 min
 Lab File: BP003288.D
 Acq: 15 Sep 2020 14:10

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010

Tgt Ion:152 Resp: 74969




Ion	Ratio	Lower	Upper
152	100		
150	153.9	123.8	185.6
115	61.2	43.8	65.6




#2
 1,4-Dioxane
 Concen: 12.48 ng
 RT: 3.164 min Scan# 47
 Delta R.T. -0.000 min
 Lab File: BP003288.D
 Acq: 15 Sep 2020 14:10

Tgt Ion: 88 Resp: 20238

Ion	Ratio	Lower	Upper
88	100		
58	52.6	44.2	66.4
43	20.7	15.2	22.8









#3
Pyridine
Concen: 12.25 ng
RT: 3.552 min Scan# 113
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Instrument :
BNA_P
ClientSampleId :
SSTDIC010

Tgt Ion: 79 Resp: 52279




Ion	Ratio	Lower	Upper
79	100		
52	44.7	41.3	61.9
51	25.0	21.2	31.8




#4
n-Nitrosodimethylamine
Concen: 14.33 ng
RT: 3.470 min Scan# 99
Delta R.T. 0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion: 42 Resp: 15964

Ion	Ratio	Lower	Upper
42	100		
74	163.8	148.6	223.0
44	6.9	5.3	7.9









#5
2-Fluorophenol
Concen: 23.36 ng
RT: 5.240 min Scan# 400
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:112 Resp: 99032
Ion Ratio Lower Upper
112 100
64 46.3 32.8 49.2
63 26.0 17.0 25.4#



#6
Aniline
Concen: 12.75 ng
RT: 6.958 min Scan# 692
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion: 93 Resp: 73789
Ion Ratio Lower Upper
93 100
66 32.5 25.4 38.0
65 16.7 12.5 18.7









#7
Phenol-d6
Concen: 24.75 ng
RT: 6.822 min Scan# 669
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Instrument :
BNA_P
ClientSampleId :
SSTDIC010

Tgt Ion: 99 Resp: 131542




Ion	Ratio	Lower	Upper
99	100		
42	12.6	8.6	13.0
71	34.9	23.4	35.2




#8
2-Chlorophenol
Concen: 11.66 ng
RT: 7.199 min Scan# 733
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion: 128 Resp: 55127

Ion	Ratio	Lower	Upper
128	100		
130	32.6	13.0	53.0
64	36.2	10.4	50.4









#9
Benzaldehyde
Concen: 15.45 ng
RT: 6.769 min Scan# 660
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Instrument :
BNA_P
ClientSampleId :
SSTDIC010

Tgt Ion: 77 Resp: 38286




Ion	Ratio	Lower	Upper
77	100		
105	108.3	97.1	137.1
106	101.7	93.7	133.7




#10
Phenol
Concen: 12.62 ng
RT: 6.846 min Scan# 673
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion: 94 Resp: 62164

Ion	Ratio	Lower	Upper
94	100		
65	25.3	2.3	42.3
66	37.7	10.9	50.9









#11
bis(2-Chloroethyl)ether
Concen: 12.57 ng
RT: 7.052 min Scan# 708
Delta R.T. -0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleID :
SSTDICC010


Tgt Ion:	93	Resp:	48997
Ion	Ratio	Lower	Upper
93	100		
63	58.9	39.0	79.0
95	32.2	12.8	52.8



#12
1,3-Dichlorobenzene
Concen: 11.58 ng
RT: 7.516 min Scan# 787
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:	146	Resp:	66586
Ion	Ratio	Lower	Upper
146	100		
148	63.9	52.5	78.7
75	27.7	18.1	27.1#









#13
1,4-Dichlorobenzene
Concen: 11.40 ng
RT: 7.658 min Scan# 811
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:146 Resp: 66151
Ion Ratio Lower Upper
146 100
148 63.3 51.3 76.9
111 40.9 28.9 43.3



#14
1,2-Dichlorobenzene
Concen: 11.81 ng
RT: 7.975 min Scan# 865
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:146 Resp: 64958
Ion Ratio Lower Upper
146 100
148 65.5 51.6 77.4
111 44.3 31.0 46.6









#15
Benzyl Alcohol
Concen: 13.98 ng
RT: 7.869 min Scan# 847
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Instrument :
BNA_P
ClientSampleId :
SSTDIC010

Tgt Ion: 79 Resp: 41865




Ion	Ratio	Lower	Upper
79	100		
108	78.6	73.5	110.3
77	63.5	50.3	75.5




#16
2,2'-oxybis(1-Chloropropane)
Concen: 11.01 ng
RT: 8.163 min Scan# 897
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion: 45 Resp: 36386

Ion	Ratio	Lower	Upper
45	100		
77	28.5	2.3	42.3
79	26.2	0.0	36.6









#17
2-Methylphenol
Concen: 12.81 ng
RT: 8.087 min Scan# 884
Delta R.T. 0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	107	Resp:	44947
Ion	Ratio	Lower	Upper
107	100		
108	108.7	89.6	134.4
77	43.8	31.8	47.6
79	43.4	31.1	46.7



#18
Hexachloroethane
Concen: 12.95 ng
RT: 8.699 min Scan# 988
Delta R.T. 0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:	117	Resp:	24582
Ion	Ratio	Lower	Upper
117	100		
119	93.1	78.8	118.2
201	113.8	92.3	138.5









#19
n-Nitroso-di-n-propylamine
Concen: 14.90 ng
RT: 8.428 min Scan# 942
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	70	Resp:	35790
Ion	Ratio	Lower	Upper
70	100		
42	40.3	32.8	49.2
101	12.6	9.6	14.4
130	27.1	24.3	36.5



#20
3+4-Methylphenols
Concen: 12.65 ng
RT: 8.422 min Scan# 941
Delta R.T. 0.012 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:	107	Resp:	59809
Ion	Ratio	Lower	Upper
107	100		
108	87.7	69.0	109.0
77	54.0	8.5	48.5#
79	29.3	4.8	44.8









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.404 min Scan# 1278
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Instrument :
BNA_P
ClientSampleId :
SSTDIC010

Tgt Ion:136 Resp: 302169




Ion	Ratio	Lower	Upper
136	100		
137	11.2	8.6	13.0
54	4.7	3.6	5.4
68	4.7	3.3	4.9




#22
Acetophenone
Concen: 12.25 ng
RT: 8.440 min Scan# 944
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:105 Resp: 82413

Ion	Ratio	Lower	Upper
105	100		
71	4.9	2.6	3.8
51	19.6	13.5	20.3
120	22.7	19.6	29.4









#23
Nitrobenzene-d5
Concen: 27.17 ng
RT: 8.775 min Scan# 1001
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Instrument :
BNA_P
ClientSampleId :
SSTDIC010

Tgt Ion: 82 Resp: 113901




Ion	Ratio	Lower	Upper
82	100		
128	48.2	45.6	68.4
54	36.3	30.1	45.1




#24
Nitrobenzene
Concen: 13.71 ng
RT: 8.816 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion: 77 Resp: 56453

Ion	Ratio	Lower	Upper
77	100		
123	51.0	48.6	72.8
65	13.8	10.0	15.0









#25
Isophorone
Concen: 14.03 ng
RT: 9.340 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion: 82 Resp: 103749
Ion Ratio Lower Upper
82 100
95 8.0 6.4 9.6
138 19.1 18.7 28.1



#26
2-Nitrophenol
Concen: 11.96 ng
RT: 9.528 min Scan# 1129
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion: 139 Resp: 29350
Ion Ratio Lower Upper
139 100
109 34.1 22.1 33.1#
65 42.5 26.5 39.7#









#27
2,4-Dimethylphenol
Concen: 11.83 ng
RT: 9.610 min Scan# 1143
Delta R.T. 0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Instrument :
BNA_P
ClientSampleId :
SSTDIC010

Tgt Ion:122 Resp: 43623




Ion	Ratio	Lower	Upper
122	100		
107	123.1	87.3	130.9
121	61.7	47.3	70.9




#28
bis(2-Chloroethoxy)methane
Concen: 12.76 ng
RT: 9.828 min Scan# 1180
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion: 93 Resp: 70305

Ion	Ratio	Lower	Upper
93	100		
95	32.6	26.7	40.1
123	11.3	10.5	15.7









#29
2,4-Dichlorophenol
Concen: 11.55 ng
RT: 10.087 min Scan# 1224
Delta R.T. 0.018 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampled :
SSTDIC010


Tgt Ion:	162	Resp:	52865
Ion	Ratio	Lower	Upper
162	100		
164	63.1	45.8	85.8
98	33.3	10.2	50.2



#30
1,2,4-Trichlorobenzene
Concen: 11.95 ng
RT: 10.275 min Scan# 1256
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:	180	Resp:	65091
Ion	Ratio	Lower	Upper
180	100		
182	94.1	76.2	114.2
145	30.0	23.8	35.8









#31
Naphthalene
Concen: 11.45 ng
RT: 10.451 min Scan# 1286
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:128	Resp:	177710
Ion	Ratio	Lower Upper
128	100	
129	11.3	8.8 13.2
127	12.9	10.6 15.8



#32
Benzoic acid
Concen: 5.37 ng
RT: 9.752 min Scan# 1167
Delta R.T. -0.023 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:122	Resp:	12867
Ion	Ratio	Lower Upper
122	100	
105	113.3	94.1 134.1
77	82.0	50.1 90.1









#33
4-Chloroaniline
Concen: 11.43 ng
RT: 10.575 min Scan# 1307
Delta R.T. 0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	127	Resp:	74132
Ion	Ratio	Lower	Upper
127	100		
129	34.3	25.7	38.5
65	26.6	17.3	25.9#
92	21.1	12.9	19.3#



#34
Hexachlorobutadiene
Concen: 13.63 ng
RT: 10.751 min Scan# 1337
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:	225	Resp:	44577
Ion	Ratio	Lower	Upper
225	100		
223	61.3	50.1	75.1
227	63.8	51.4	77.2









#35
Caprolactam
Concen: 12.71 ng
RT: 11.328 min Scan# 1435
Delta R.T. -0.012 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	113	Resp:	17480
Ion	Ratio	Lower	Upper
113	100		
55	102.5	77.3	117.3
56	79.8	55.9	95.9



#36
4-Chloro-3-methylphenol
Concen: 13.84 ng
RT: 11.728 min Scan# 1503
Delta R.T. 0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:	107	Resp:	59125
Ion	Ratio	Lower	Upper
107	100		
144	23.8	22.7	34.1
142	77.1	70.7	106.1









#37
2-Methylnaphthalene
Concen: 11.68 ng
RT: 12.081 min Scan# 1563
Delta R.T. 0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:142 Resp: 129732
Ion Ratio Lower Upper
142 100
141 89.7 71.4 107.2
115 36.9 25.8 38.6



#38
1-Methylnaphthalene
Concen: 11.74 ng
RT: 12.298 min Scan# 1600
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:142 Resp: 123912
Ion Ratio Lower Upper
142 100
141 91.1 74.4 111.6
116 4.2 2.9 4.3









#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.269 min Scan# 1935
Delta R.T. -0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:164 Resp: 201333
Ion Ratio Lower Upper
164 100
162 102.4 79.4 119.2
160 44.7 36.1 54.1



#40
1,2,4,5-Tetrachlorobenzene
Concen: 12.08 ng
RT: 12.457 min Scan# 1627
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:216 Resp: 72098
Ion Ratio Lower Upper
216 100
214 78.8 63.0 94.6
179 20.7 16.9 25.3
108 16.2 12.6 19.0









#41
Hexachlorocyclopentadiene
Concen: 2.38 ng
RT: 12.440 min Scan# 1624
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	237	Resp:	6722
Ion	Ratio	Lower	Upper
237	100		
235	61.3	43.2	83.2
272	11.7	0.0	33.3



#42
2,4,6-Tribromophenol
Concen: 24.60 ng
RT: 15.775 min Scan# 2191
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:	330	Resp:	66979
Ion	Ratio	Lower	Upper
330	100		
332	97.4	77.1	115.7
141	31.9	23.4	35.2









#43
2,4,6-Trichlorophenol
Concen: 11.48 ng
RT: 12.710 min Scan# 1670
Delta R.T. 0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampled :
SSTDICC010


Tgt Ion:196 Resp: 45064
Ion Ratio Lower Upper
196 100
198 99.0 77.0 115.6
200 30.3 25.1 37.7



#44
2,4,5-Trichlorophenol
Concen: 11.49 ng
RT: 12.798 min Scan# 1685
Delta R.T. 0.012 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:196 Resp: 47707
Ion Ratio Lower Upper
196 100
198 94.7 77.3 115.9
97 43.1 31.0 46.4
132 25.3 20.9 31.3









#45
2-Fluorobiphenyl
Concen: 22.93 ng
RT: 12.887 min Scan# 1700
Delta R.T. -0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	172	Resp:	315237
Ion Ratio	Lower	Upper	
172	100		
171	35.0	28.5	42.7
170	23.4	19.2	28.8



#46
1,1'-Biphenyl
Concen: 11.19 ng
RT: 13.098 min Scan# 1736
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:	154	Resp:	169041
Ion Ratio	Lower	Upper	
154	100		
153	41.0	20.4	60.4
76	10.7	0.0	29.3









#47
2-Chloronaphthalene
Concen: 11.29 ng
RT: 13.140 min Scan# 1743
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	162	Resp:	139559
Ion	Ratio	Lower	Upper
162	100		
127	40.5	29.9	44.9
164	32.5	26.1	39.1



#48
2-Nitroaniline
Concen: 14.48 ng
RT: 13.351 min Scan# 1779
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:	65	Resp:	33525
Ion	Ratio	Lower	Upper
65	100		
92	81.0	70.6	106.0
138	128.7	133.6	200.4









#49
Acenaphthylene
Concen: 11.46 ng
RT: 13.992 min Scan# 1888
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	152	Resp:	214250
Ion Ratio	Lower	Upper	
152	100		
151	20.6	15.9	23.9
153	12.8	10.5	15.7



#50
Dimethylphthalate
Concen: 11.86 ng
RT: 13.734 min Scan# 1844
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:	163	Resp:	182101
Ion Ratio	Lower	Upper	
163	100		
194	4.2	3.3	4.9
164	10.1	8.2	12.2









#51
2,6-Dinitrotoluene
Concen: 11.65 ng
RT: 13.851 min Scan# 1864
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:165 Resp: 37171
Ion Ratio Lower Upper
165 100
63 44.0 27.8 41.6#
89 52.6 34.6 52.0#



#52
Acenaphthene
Concen: 11.42 ng
RT: 14.334 min Scan# 1946
Delta R.T. -0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:154 Resp: 131093
Ion Ratio Lower Upper
154 100
153 113.0 91.0 136.6
152 55.6 43.0 64.6









#53
3-Nitroaniline
Concen: 11.01 ng
RT: 14.187 min Scan# 1921
Delta R.T. 0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:138 Resp: 39818
Ion Ratio Lower Upper
138 100
108 11.6 7.5 11.3#
92 112.3 75.0 112.6



#54
2,4-Dinitrophenol
Concen: 5.85 ng
RT: 14.428 min Scan# 1962
Delta R.T. 0.024 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:184 Resp: 8340
Ion Ratio Lower Upper
184 100
63 39.8 32.3 48.5
154 61.1 49.1 73.7









#55
Dibenzofuran
Concen: 11.66 ng
RT: 14.675 min Scan# 2004
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:168	Resp: 210325
Ion Ratio	Lower Upper
168 100	
139 38.8	29.1 43.7
169 13.4	10.4 15.6



#56
4-Nitrophenol
Concen: 6.76 ng
RT: 14.551 min Scan# 1983
Delta R.T. 0.024 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:139	Resp: 17573
Ion Ratio	Lower Upper
139 100	
109 75.8	44.6 84.6
65 76.8	41.6 81.6









#57
2,4-Dinitrotoluene
Concen: 12.10 ng
RT: 14.651 min Scan# 2000
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:165	Resp:	51901
Ion	Ratio	Lower Upper
165	100	
63	35.6	20.6 31.0#
89	65.6	42.6 63.8#
182	3.3	2.4 3.6



#58
Fluorene
Concen: 12.28 ng
RT: 15.328 min Scan# 2115
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:166	Resp:	170662
Ion	Ratio	Lower Upper
166	100	
165	99.1	77.8 116.8
167	13.3	11.0 16.4









#59
2,3,4,6-Tetrachlorophenol
Concen: 11.59 ng
RT: 14.916 min Scan# 2045
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	232	Resp:	43503
Ion	Ratio	Lower	Upper
232	100		
131	41.1	34.1	51.1
130	2.2	1.8	2.8
166	27.8	24.6	36.8



#60
Diethylphthalate
Concen: 12.42 ng
RT: 15.104 min Scan# 2077
Delta R.T. -0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:	149	Resp:	187158
Ion	Ratio	Lower	Upper
149	100		
177	21.4	18.0	27.0
150	13.0	10.0	15.0









#61
4-Chlorophenyl-phenylether
Concen: 12.82 ng
RT: 15.322 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	204	Resp:	91968
Ion	Ratio	Lower	Upper
204	100		
206	33.8	25.9	38.9
141	62.1	46.5	69.7



#62
4-Nitroaniline
Concen: 11.28 ng
RT: 15.357 min Scan# 2120
Delta R.T. 0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:	138	Resp:	41923
Ion	Ratio	Lower	Upper
138	100		
92	47.2	20.7	60.7
108	84.4	47.8	87.8









#63
Azobenzene
Concen: 14.29 ng
RT: 15.616 min Scan# 2164
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Instrument :
BNA_P
ClientSampled :
SSTDIC010

Tgt Ion: 77 Resp: 142897




Ion	Ratio	Lower	Upper
77	100		
182	30.4	16.9	56.9
105	23.0	5.6	45.6
51	23.4	2.9	42.9




#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.028 min Scan# 2404
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion: 188 Resp: 456100

Ion	Ratio	Lower	Upper
188	100		
94	7.9	6.2	9.2
80	8.7	6.2	9.2









#65
4,6-Dinitro-2-methylphenol
Concen: 8.76 ng
RT: 15.434 min Scan# 2133
Delta R.T. 0.012 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:198	Resp:	19097
Ion Ratio	Lower	Upper
198	100	
51	26.1	6.4 46.4
105	38.4	18.9 58.9



#66
n-Nitrosodiphenylamine
Concen: 11.49 ng
RT: 15.539 min Scan# 2151
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:169	Resp:	153252
Ion Ratio	Lower	Upper
169	100	
168	66.1	53.8 80.6
167	36.7	28.9 43.3









#67
4-Bromophenyl-phenylether
Concen: 12.08 ng
RT: 16.222 min Scan# 2267
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:	248	Resp:	60806
Ion	Ratio	Lower	Upper
248	100		
250	94.2	76.7	115.1
141	70.9	53.0	79.6



#68
Hexachlorobenzene
Concen: 12.18 ng
RT: 16.339 min Scan# 2287
Delta R.T. -0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:	284	Resp:	73012
Ion	Ratio	Lower	Upper
284	100		
142	31.6	24.2	36.4
249	30.6	23.8	35.8









#69
Atrazine
Concen: 13.16 ng
RT: 16.498 min Scan# 2314
Delta R.T. -0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	200	Resp:	59172
Ion Ratio	Lower	Upper	
200	100		
173	26.8	5.9	45.9
215	46.7	28.4	68.4



#70
Pentachlorophenol
Concen: 5.84 ng
RT: 16.704 min Scan# 2349
Delta R.T. 0.012 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:	266	Resp:	16728
Ion Ratio	Lower	Upper	
266	100		
268	64.7	51.1	76.7
264	64.5	50.6	76.0









#71
Phenanthrene
Concen: 11.44 ng
RT: 17.069 min Scan# 2411
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:178 Resp: 282812
Ion Ratio Lower Upper
178 100
176 19.3 15.3 22.9
179 15.1 12.5 18.7



#72
Anthracene
Concen: 11.62 ng
RT: 17.163 min Scan# 2427
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:178 Resp: 274956
Ion Ratio Lower Upper
178 100
176 19.5 15.0 22.4
179 15.2 12.5 18.7









#73
 Carbazole
 Concen: 11.39 ng
 RT: 17.439 min Scan# 2474
 Delta R.T. 0.006 min
 Lab File: BP003288.D
 Acq: 15 Sep 2020 14:10




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010


Tgt Ion:	167	Resp:	258497
Ion Ratio	Lower	Upper	
167	100		
166	21.6	17.6	26.4
139	13.0	10.2	15.4



#74
 Di-n-butylphthalate
 Concen: 12.66 ng
 RT: 18.004 min Scan# 2570
 Delta R.T. 0.000 min
 Lab File: BP003288.D
 Acq: 15 Sep 2020 14:10

Tgt Ion:	149	Resp:	337935
Ion Ratio	Lower	Upper	
149	100		
150	9.0	7.3	10.9
104	5.0	3.4	5.0









#75
Fluoranthene
Concen: 12.31 ng
RT: 19.051 min Scan# 2748
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	202	Resp:	351687
Ion	Ratio	Lower	Upper
202	100		
101	9.6	0.0	29.4
203	17.3	0.0	37.8



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.127 min Scan# 3101
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:	240	Resp:	537623
Ion	Ratio	Lower	Upper
240	100		
120	8.3	7.8	11.6
236	26.2	20.8	31.2









#77
Benzidine
Concen: 14.12 ng
RT: 19.239 min Scan# 2780
Delta R.T. 0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:184 Resp: 173448
Ion Ratio Lower Upper
184 100
185 13.4 11.5 17.3
183 11.5 9.6 14.4



#78
Pyrene
Concen: 10.49 ng
RT: 19.404 min Scan# 2808
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:202 Resp: 366089
Ion Ratio Lower Upper
202 100
200 20.9 17.1 25.7
203 17.7 14.2 21.2









#79
 Terphenyl-d14
 Concen: 24.09 ng
 RT: 19.604 min Scan# 2842
 Delta R.T. 0.000 min
 Lab File: BP003288.D
 Acq: 15 Sep 2020 14:10




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010


Tgt Ion:	244	Resp:	588494
Ion	Ratio	Lower	Upper
244	100		
212	7.2	5.8	8.6
122	8.8	7.8	11.8



#80
 Butylbenzylphthalate
 Concen: 11.53 ng
 RT: 20.286 min Scan# 2958
 Delta R.T. 0.006 min
 Lab File: BP003288.D
 Acq: 15 Sep 2020 14:10

Tgt Ion:	149	Resp:	165714
Ion	Ratio	Lower	Upper
149	100		
91	61.2	45.4	68.0
206	22.5	17.1	25.7









#81
Benzo(a)anthracene
Concen: 11.40 ng
RT: 21.110 min Scan# 3098
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampled :
SSTDIC010


Tgt Ion:	228	Resp:	383836
Ion Ratio	Lower	Upper	
228	100		
226	27.1	22.1	33.1
229	19.8	15.9	23.9



#82
3,3'-Dichlorobenzidine
Concen: 12.02 ng
RT: 21.051 min Scan# 3088
Delta R.T. 0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:	252	Resp:	148956
Ion Ratio	Lower	Upper	
252	100		
254	64.9	51.0	76.4
126	9.0	7.9	11.9









#83
Chrysene
Concen: 11.56 ng
RT: 21.163 min Scan# 3107
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt	Ion:228	Resp:	372754
Ion	Ratio	Lower	Upper
228	100		
226	29.7	24.4	36.6
229	19.5	15.8	23.6



#84
Bis(2-ethylhexyl)phthalate
Concen: 11.77 ng
RT: 21.051 min Scan# 3088
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:149	Resp: 244178
Ion Ratio	Lower Upper
149 100	
167 27.2	22.6 34.0
279 5.2	4.2 6.2









#85
Di-n-octyl phthalate
Concen: 12.19 ng
RT: 21.916 min Scan# 3235
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Instrument :
BNA_P
ClientSampleId :
SSTDIC010

Tgt Ion:149 Resp: 433604




Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	5.2	3.7	5.5




#86
Perylene-d12
Concen: 20.00 ng
RT: 23.351 min Scan# 3479
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:264 Resp: 586463

Ion	Ratio	Lower	Upper
264	100		
260	24.1	18.8	28.2
265	21.2	17.4	26.2









#87
Indeno(1,2,3-cd)pyrene
Concen: 11.25 ng
RT: 25.615 min Scan# 3864
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Instrument :
BNA_P
ClientSampleId :
SSTDIC010

Tgt Ion: 276 Resp: 491926




Ion	Ratio	Lower	Upper
276	100		
138	21.0	18.7	28.1
277	25.4	20.4	30.6




#88
Benzo(b)fluoranthene
Concen: 11.22 ng
RT: 22.680 min Scan# 3365
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion: 252 Resp: 408865

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.6	26.4
125	8.4	7.3	10.9





#89

Benzo(k)fluoranthene

Concen: 11.24 ng

RT: 22.721 min Scan# 3372





Delta R.T. -0.006 min

Lab File: BP003288.D

Acq: 15 Sep 2020 14:10

Tgt Ion:252 Resp: 391593

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.8	26.8
125	8.6	7.4	11.2



#90

Benzo(a)pyrene

Concen: 11.33 ng

RT: 23.251 min Scan# 3462




Delta R.T. 0.000 min


Lab File: BP003288.D

Acq: 15 Sep 2020 14:10

Tgt Ion:252 Resp: 383059

Ion	Ratio	Lower	Upper
252	100		
253	21.1	17.4	26.0
125	9.3	8.2	12.4









#91
Dibenzo(a,h)anthracene
Concen: 11.30 ng
RT: 25.627 min Scan# 3866
Delta R.T. 0.000 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Instrument :
BNA_P
ClientSampleId :
SSTDIC010

Tgt Ion:	278	Resp:	405838
Ion	Ratio	Lower	Upper
278	100		
139	13.3	12.4	18.6
279	23.7	19.0	28.6



#92
Benzo(g,h,i)perylene
Concen: 10.96 ng
RT: 26.304 min Scan# 3981
Delta R.T. -0.006 min
Lab File: BP003288.D
Acq: 15 Sep 2020 14:10

Tgt Ion:	276	Resp:	395198
Ion	Ratio	Lower	Upper
276	100		
277	23.1	19.0	28.6
138	17.5	16.6	24.8

