

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091520\
 Data File : BP003289.D
 Acq On : 15 Sep 2020 14:44
 Operator : CG/JU
 Sample : SSTDICC20
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC20

Quant Time: Sep 15 16:28:28 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 15 16:17:32 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.622	152	80768	20.00	ng	0.00
21) Naphthalene-d8	10.404	136	313018	20.00	ng	0.00
39) Acenaphthene-d10	14.269	164	209000	20.00	ng	0.00
64) Phenanthrene-d10	17.028	188	471090	20.00	ng	0.00
76) Chrysene-d12	21.127	240	538657	20.00	ng	0.00
86) Perylene-d12	23.351	264	594560	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.240	112	186561	40.84	ng	0.00
7) Phenol-d6	6.822	99	251196	43.86	ng	0.00
23) Nitrobenzene-d5	8.775	82	214423	49.38	ng	0.00
42) 2,4,6-Tribromophenol	15.769	330	128629	45.51	ng	0.00
45) 2-Fluorobiphenyl	12.887	172	580532	40.68	ng	0.00
79) Terphenyl-d14	19.604	244	1068141	43.63	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.164	88	36888	21.12	ng	98
3) Pyridine	3.552	79	100656	21.90	ng	95
4) n-Nitrosodimethylamine	3.464	42	29825	24.85	ng	86
6) Aniline	6.952	93	143405	23.00	ng	96
8) 2-Chlorophenol	7.199	128	104575	20.53	ng	94
9) Benzaldehyde	6.769	77	71165	26.65	ng	91
10) Phenol	6.846	94	118503	22.33	ng	89
11) bis(2-Chloroethyl)ether	7.057	93	93112	22.17	ng	98
12) 1,3-Dichlorobenzene	7.516	146	124555	20.11	ng	# 96
13) 1,4-Dichlorobenzene	7.657	146	126661	20.26	ng	97
14) 1,2-Dichlorobenzene	7.975	146	121170	20.44	ng	96
15) Benzyl Alcohol	7.869	79	81764	25.34	ng	90
16) 2,2'-oxybis(1-Chloropr...	8.157	45	68773	19.32	ng	88
17) 2-Methylphenol	8.087	107	85322	22.57	ng	96
18) Hexachloroethane	8.699	117	46323	22.65	ng	96
19) n-Nitroso-di-n-propyla...	8.428	70	69180	26.74	ng	98
20) 3+4-Methylphenols	8.416	107	115816	22.74	ng	94
22) Acetophenone	8.440	105	156385	22.43	ng	# 95
24) Nitrobenzene	8.816	77	106651	25.01	ng	88
25) Isophorone	9.340	82	197321	25.76	ng	95
26) 2-Nitrophenol	9.528	139	57859	22.76	ng	# 86
27) 2,4-Dimethylphenol	9.604	122	82264	21.53	ng	88
28) bis(2-Chloroethoxy)met...	9.828	93	132385	23.20	ng	98
29) 2,4-Dichlorophenol	10.075	162	103179	21.76	ng	94
30) 1,2,4-Trichlorobenzene	10.275	180	122601	21.72	ng	99
31) Naphthalene	10.451	128	334326	20.79	ng	99
32) Benzoic acid	9.746	122	35460	14.29	ng	88
33) 4-Chloroaniline	10.575	127	142819	21.25	ng	96
34) Hexachlorobutadiene	10.751	225	84080	24.82	ng	98
35) Caprolactam	11.328	113	33824	23.74	ng	88
36) 4-Chloro-3-methylphenol	11.722	107	111792	25.26	ng	88
37) 2-Methylnaphthalene	12.075	142	242794	21.10	ng	97
38) 1-Methylnaphthalene	12.298	142	230292	21.06	ng	100
40) 1,2,4,5-Tetrachloroben...	12.457	216	137414	22.18	ng	100
41) Hexachlorocyclopentadiene	12.440	237	22553	7.70	ng	95
43) 2,4,6-Trichlorophenol	12.704	196	87006	21.36	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091520\
 Data File : BP003289.D
 Acq On : 15 Sep 2020 14:44
 Operator : CG/JU
 Sample : SSTDICC20
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Quant Time: Sep 15 16:28:28 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 15 16:17:32 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.792	196	91460	21.23	ng	97
46) 1,1'-Biphenyl	13.098	154	312920	19.96	ng	98
47) 2-Chloronaphthalene	13.140	162	260678	20.31	ng	97
48) 2-Nitroaniline	13.345	65	66166	27.54	ng	# 75
49) Acenaphthylene	13.987	152	401300	20.68	ng	99
50) Dimethylphthalate	13.734	163	337988	21.21	ng	99
51) 2,6-Dinitrotoluene	13.851	165	71402	21.56	ng	# 86
52) Acenaphthene	14.334	154	244768	20.53	ng	100
53) 3-Nitroaniline	14.181	138	78372	20.88	ng	# 82
54) 2,4-Dinitrophenol	14.410	184	21569	14.58	ng	# 93
55) Dibenzofuran	14.675	168	390664	20.87	ng	97
56) 4-Nitrophenol	14.534	139	42631	15.79	ng	85
57) 2,4-Dinitrotoluene	14.651	165	100843	22.64	ng	# 84
58) Fluorene	15.328	166	309793	21.48	ng	99
59) 2,3,4,6-Tetrachlorophenol	14.916	232	84902	21.78	ng	97
60) Diethylphthalate	15.104	149	346987	22.19	ng	99
61) 4-Chlorophenyl-phenyle...	15.322	204	168365	22.61	ng	95
62) 4-Nitroaniline	15.351	138	83908	21.75	ng	# 76
63) Azobenzene	15.616	77	267945	25.82	ng	94
65) 4,6-Dinitro-2-methylph...	15.422	198	45074	20.01	ng	95
66) n-Nitrosodiphenylamine	15.539	169	281551	20.44	ng	99
67) 4-Bromophenyl-phenylether	16.222	248	115374	22.19	ng	99
68) Hexachlorobenzene	16.345	284	136677	22.08	ng	97
69) Atrazine	16.498	200	110410	23.77	ng	98
70) Pentachlorophenol	16.698	266	41736	14.10	ng	96
71) Phenanthrene	17.069	178	523443	20.51	ng	99
72) Anthracene	17.163	178	516782	21.15	ng	99
73) Carbazole	17.439	167	490130	20.90	ng	100
74) Di-n-butylphthalate	18.004	149	629084	22.82	ng	99
75) Fluoranthene	19.051	202	658316	22.31	ng	99
77) Benzidine	19.233	184	335912	27.29	ng	99
78) Pyrene	19.404	202	680757	19.47	ng	99
80) Butylbenzylphthalate	20.280	149	312067	21.68	ng	90
81) Benzo(a)anthracene	21.110	228	703004	20.84	ng	99
82) 3,3'-Dichlorobenzidine	21.045	252	280136	22.57	ng	98
83) Chrysene	21.163	228	691654	21.41	ng	99
84) Bis(2-ethylhexyl)phtha...	21.051	149	453676	21.82	ng	100
85) Di-n-octyl phthalate	21.916	149	809070	22.70	ng	99
87) Indeno(1,2,3-cd)pyrene	25.609	276	908572	20.49	ng	98
88) Benzo(b)fluoranthene	22.680	252	768110	20.79	ng	99
89) Benzo(k)fluoranthene	22.727	252	721746	20.44	ng	98
90) Benzo(a)pyrene	23.251	252	708816	20.68	ng	99
91) Dibenzo(a,h)anthracene	25.621	278	752588	20.66	ng	98
92) Benzo(g,h,i)perylene	26.304	276	737984	20.19	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 15 16:28:28 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 15 16:17:32 2020
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.622 min Scan# 805
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:152 Resp: 80768
Ion Ratio Lower Upper
152 100
150 154.4 123.8 185.6
115 60.6 43.8 65.6



#2
1,4-Dioxane
Concen: 21.12 ng
RT: 3.164 min Scan# 47
Delta R.T. -0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion: 88 Resp: 36888
Ion Ratio Lower Upper
88 100
58 54.4 44.2 66.4
43 20.3 15.2 22.8









#3
Pyridine
Concen: 21.90 ng
RT: 3.552 min Scan# 113
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion: 79 Resp: 100656
Ion Ratio Lower Upper
79 100
52 46.9 41.3 61.9
51 26.1 21.2 31.8



#4
n-Nitrosodimethylamine
Concen: 24.85 ng
RT: 3.464 min Scan# 98
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion: 42 Resp: 29825
Ion Ratio Lower Upper
42 100
74 165.1 148.6 223.0
44 7.1 5.3 7.9









#5
2-Fluorophenol
Concen: 40.84 ng
RT: 5.240 min Scan# 400
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:112 Resp: 186561
Ion Ratio Lower Upper
112 100
64 45.6 32.8 49.2
63 24.4 17.0 25.4



#6
Aniline
Concen: 23.00 ng
RT: 6.952 min Scan# 691
Delta R.T. -0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion: 93 Resp: 143405
Ion Ratio Lower Upper
93 100
66 33.6 25.4 38.0
65 17.3 12.5 18.7









#7
Phenol-d6
Concen: 43.86 ng
RT: 6.822 min Scan# 669
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	99	Resp:	251196
Ion	Ratio	Lower	Upper
99	100		
42	11.5	8.6	13.0
71	34.5	23.4	35.2



#8
2-Chlorophenol
Concen: 20.53 ng
RT: 7.199 min Scan# 733
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:	128	Resp:	104575
Ion	Ratio	Lower	Upper
128	100		
130	32.2	13.0	53.0
64	36.2	10.4	50.4









#9
Benzaldehyde
Concen: 26.65 ng
RT: 6.769 min Scan# 660
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion: 77 Resp: 71165




Ion	Ratio	Lower	Upper
77	100		
105	108.4	97.1	137.1
106	101.8	93.7	133.7




#10
Phenol
Concen: 22.33 ng
RT: 6.846 min Scan# 673
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion: 94 Resp: 118503

Ion	Ratio	Lower	Upper
94	100		
65	25.8	2.3	42.3
66	38.4	10.9	50.9









#11
bis(2-Chloroethyl)ether
Concen: 22.17 ng
RT: 7.057 min Scan# 709
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion: 93 Resp: 93112
Ion Ratio Lower Upper
93 100
63 56.8 39.0 79.0
95 32.0 12.8 52.8



#12
1,3-Dichlorobenzene
Concen: 20.11 ng
RT: 7.516 min Scan# 787
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion: 146 Resp: 124555
Ion Ratio Lower Upper
146 100
148 64.9 52.5 78.7
75 28.2 18.1 27.1#









#13
1,4-Dichlorobenzene
Concen: 20.26 ng
RT: 7.657 min Scan# 811
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:146 Resp: 126661
Ion Ratio Lower Upper
146 100
148 64.1 51.3 76.9
111 41.5 28.9 43.3



#14
1,2-Dichlorobenzene
Concen: 20.44 ng
RT: 7.975 min Scan# 865
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:146 Resp: 121170
Ion Ratio Lower Upper
146 100
148 65.7 51.6 77.4
111 43.7 31.0 46.6









#15
Benzyl Alcohol
Concen: 25.34 ng
RT: 7.869 min Scan# 847
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion: 79 Resp: 81764
Ion Ratio Lower Upper
79 100
108 77.8 73.5 110.3
77 60.5 50.3 75.5



#16
2,2'-oxybis(1-Chloropropane)
Concen: 19.32 ng
RT: 8.157 min Scan# 896
Delta R.T. -0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion: 45 Resp: 68773
Ion Ratio Lower Upper
45 100
77 28.1 2.3 42.3
79 21.5 0.0 36.6









#17
2-Methylphenol
Concen: 22.57 ng
RT: 8.087 min Scan# 884
Delta R.T. 0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	107	Resp:	85322
Ion	Ratio	Lower	Upper
107	100		
108	112.7	89.6	134.4
77	44.1	31.8	47.6
79	44.9	31.1	46.7



#18
Hexachloroethane
Concen: 22.65 ng
RT: 8.699 min Scan# 988
Delta R.T. 0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:	117	Resp:	46323
Ion	Ratio	Lower	Upper
117	100		
119	94.4	78.8	118.2
201	119.1	92.3	138.5









#19
n-Nitroso-di-n-propylamine
Concen: 26.74 ng
RT: 8.428 min Scan# 942
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	70	Resp:	69180
Ion	Ratio	Lower	Upper
70	100		
42	41.0	32.8	49.2
101	12.0	9.6	14.4
130	27.2	24.3	36.5



#20
3+4-Methylphenols
Concen: 22.74 ng
RT: 8.416 min Scan# 940
Delta R.T. 0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:	107	Resp:	115816
Ion	Ratio	Lower	Upper
107	100		
108	90.7	69.0	109.0
77	37.1	8.5	48.5
79	28.9	4.8	44.8









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.404 min Scan# 1278
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:136 Resp: 313018




Ion	Ratio	Lower	Upper
136	100		
137	11.3	8.6	13.0
54	5.1	3.6	5.4
68	4.7	3.3	4.9




#22
Acetophenone
Concen: 22.43 ng
RT: 8.440 min Scan# 944
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:105 Resp: 156385

Ion	Ratio	Lower	Upper
105	100		
71	4.5	2.6	3.8
51	19.2	13.5	20.3
120	22.1	19.6	29.4









#23
Nitrobenzene-d5
Concen: 49.38 ng
RT: 8.775 min Scan# 1001
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion: 82 Resp: 214423




Ion	Ratio	Lower	Upper
82	100		
128	47.5	45.6	68.4
54	37.5	30.1	45.1




#24
Nitrobenzene
Concen: 25.01 ng
RT: 8.816 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion: 77 Resp: 106651

Ion	Ratio	Lower	Upper
77	100		
123	49.6	48.6	72.8
65	12.2	10.0	15.0









#25
Isophorone
Concen: 25.76 ng
RT: 9.340 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion: 82 Resp: 197321
Ion Ratio Lower Upper
82 100
95 8.4 6.4 9.6
138 20.0 18.7 28.1



#26
2-Nitrophenol
Concen: 22.76 ng
RT: 9.528 min Scan# 1129
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion: 139 Resp: 57859
Ion Ratio Lower Upper
139 100
109 33.7 22.1 33.1#
65 42.4 26.5 39.7#









#27
2,4-Dimethylphenol
Concen: 21.53 ng
RT: 9.604 min Scan# 1142
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:122 Resp: 82264




Ion	Ratio	Lower	Upper
122	100		
107	126.6	87.3	130.9
121	60.8	47.3	70.9




#28
bis(2-Chloroethoxy)methane
Concen: 23.20 ng
RT: 9.828 min Scan# 1180
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion: 93 Resp: 132385

Ion	Ratio	Lower	Upper
93	100		
95	32.3	26.7	40.1
123	12.3	10.5	15.7









#29
2,4-Dichlorophenol
Concen: 21.76 ng
RT: 10.075 min Scan# 1222
Delta R.T. 0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	162	Resp:	103179
Ion	Ratio	Lower	Upper
162	100		
164	62.3	45.8	85.8
98	34.5	10.2	50.2



#30
1,2,4-Trichlorobenzene
Concen: 21.72 ng
RT: 10.275 min Scan# 1256
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:	180	Resp:	122601
Ion	Ratio	Lower	Upper
180	100		
182	95.8	76.2	114.2
145	30.7	23.8	35.8









#31
Naphthalene
Concen: 20.79 ng
RT: 10.451 min Scan# 1286
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:128	Resp:	334326
Ion Ratio	Lower	Upper
128	100	
129	10.6	8.8 13.2
127	13.6	10.6 15.8



#32
Benzoic acid
Concen: 14.29 ng
RT: 9.746 min Scan# 1166
Delta R.T. -0.029 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:122	Resp:	35460
Ion Ratio	Lower	Upper
122	100	
105	121.8	94.1 134.1
77	85.5	50.1 90.1









#33
4-Chloroaniline
Concen: 21.25 ng
RT: 10.575 min Scan# 1307
Delta R.T. 0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:127 Resp: 142819




Ion	Ratio	Lower	Upper
127	100		
129	31.8	25.7	38.5
65	24.8	17.3	25.9
92	19.1	12.9	19.3




#34
Hexachlorobutadiene
Concen: 24.82 ng
RT: 10.751 min Scan# 1337
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:225 Resp: 84080

Ion	Ratio	Lower	Upper
225	100		
223	64.2	50.1	75.1
227	65.2	51.4	77.2



Abundance Scan 1441 (11.363 min): BP003089.D (-1429) (-)





#35
Caprolactam
Concen: 23.74 ng
RT: 11.328 min Scan# 1435
Delta R.T. -0.012 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44


Tgt Ion:	113	Resp:	33824
Ion	Ratio	Lower	Upper
113	100		
55	111.1	77.3	117.3
56	84.6	55.9	95.9

Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Abundance Scan 1435 (11.328 min): BP003289.D\data.ms



Abundance Scan 1435 (11.328 min): BP003289.D\data.ms (-)





Abundance Scan 1506 (11.745 min): BP003089.D (-1495) (-)




#36
4-Chloro-3-methylphenol
Concen: 25.26 ng
RT: 11.722 min Scan# 1502
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44


Tgt Ion:	107	Resp:	111792
Ion	Ratio	Lower	Upper
107	100		
144	25.0	22.7	34.1
142	75.9	70.7	106.1

Abundance Scan 1502 (11.722 min): BP003289.D\data.ms



Abundance Scan 1502 (11.722 min): BP003289.D\data.ms (-)









#37
2-Methylnaphthalene
Concen: 21.10 ng
RT: 12.075 min Scan# 1562
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:142 Resp: 242794




Ion	Ratio	Lower	Upper
142	100		
141	89.0	71.4	107.2
115	38.2	25.8	38.6




#38
1-Methylnaphthalene
Concen: 21.06 ng
RT: 12.298 min Scan# 1600
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:142 Resp: 230292

Ion	Ratio	Lower	Upper
142	100		
141	92.9	74.4	111.6
116	4.2	2.9	4.3









#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.269 min Scan# 1935
Delta R.T. -0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDICC020


Tgt Ion:164 Resp: 209000
Ion Ratio Lower Upper
164 100
162 102.2 79.4 119.2
160 44.6 36.1 54.1



#40
1,2,4,5-Tetrachlorobenzene
Concen: 22.18 ng
RT: 12.457 min Scan# 1627
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:216 Resp: 137414
Ion Ratio Lower Upper
216 100
214 78.5 63.0 94.6
179 21.0 16.9 25.3
108 15.7 12.6 19.0









#41
Hexachlorocyclopentadiene
Concen: 7.70 ng
RT: 12.440 min Scan# 1624
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	237	Resp:	22553
Ion	Ratio	Lower	Upper
237	100		
235	58.6	43.2	83.2
272	13.1	0.0	33.3



#42
2,4,6-Tribromophenol
Concen: 45.51 ng
RT: 15.769 min Scan# 2190
Delta R.T. -0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:	330	Resp:	128629
Ion	Ratio	Lower	Upper
330	100		
332	98.0	77.1	115.7
141	31.1	23.4	35.2









#43
2,4,6-Trichlorophenol
Concen: 21.36 ng
RT: 12.704 min Scan# 1669
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:196 Resp: 87006
Ion Ratio Lower Upper
196 100
198 97.1 77.0 115.6
200 30.1 25.1 37.7



#44
2,4,5-Trichlorophenol
Concen: 21.23 ng
RT: 12.792 min Scan# 1684
Delta R.T. 0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:196 Resp: 91460
Ion Ratio Lower Upper
196 100
198 96.0 77.3 115.9
97 45.0 31.0 46.4
132 26.0 20.9 31.3









#45
2-Fluorobiphenyl
Concen: 40.68 ng
RT: 12.887 min Scan# 1700
Delta R.T. -0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	172	Resp:	580532
Ion	Ratio	Lower	Upper
172	100		
171	35.2	28.5	42.7
170	22.9	19.2	28.8



#46
1,1'-Biphenyl
Concen: 19.96 ng
RT: 13.098 min Scan# 1736
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:	154	Resp:	312920
Ion	Ratio	Lower	Upper
154	100		
153	41.1	20.4	60.4
76	11.2	0.0	29.3









#47
2-Chloronaphthalene
Concen: 20.31 ng
RT: 13.140 min Scan# 1743
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:162 Resp: 260678




Ion	Ratio	Lower	Upper
162	100		
127	40.6	29.9	44.9
164	33.1	26.1	39.1




#48
2-Nitroaniline
Concen: 27.54 ng
RT: 13.345 min Scan# 1778
Delta R.T. -0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion: 65 Resp: 66166

Ion	Ratio	Lower	Upper
65	100		
92	77.4	70.6	106.0
138	124.5	133.6	200.4









#49
Acenaphthylene
Concen: 20.68 ng
RT: 13.987 min Scan# 1887
Delta R.T. -0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:152	Resp: 401300
Ion Ratio	Lower Upper
152 100	
151 20.2	15.9 23.9
153 13.7	10.5 15.7



#50
Dimethylphthalate
Concen: 21.21 ng
RT: 13.734 min Scan# 1844
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:163	Resp: 337988
Ion Ratio	Lower Upper
163 100	
194 4.2	3.3 4.9
164 10.0	8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 21.56 ng
RT: 13.851 min Scan# 1864
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:165 Resp: 71402
Ion Ratio Lower Upper
165 100
63 43.4 27.8 41.6#
89 51.3 34.6 52.0



#52
Acenaphthene
Concen: 20.53 ng
RT: 14.334 min Scan# 1946
Delta R.T. -0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:154 Resp: 244768
Ion Ratio Lower Upper
154 100
153 113.9 91.0 136.6
152 53.2 43.0 64.6









#53
3-Nitroaniline
Concen: 20.88 ng
RT: 14.181 min Scan# 1920
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:138 Resp: 78372
Ion Ratio Lower Upper
138 100
108 11.6 7.5 11.3#
92 111.9 75.0 112.6



#54
2,4-Dinitrophenol
Concen: 14.58 ng
RT: 14.410 min Scan# 1959
Delta R.T. 0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:184 Resp: 21569
Ion Ratio Lower Upper
184 100
63 49.1 32.3 48.5#
154 63.2 49.1 73.7









#55
Dibenzenofuran
Concen: 20.87 ng
RT: 14.675 min Scan# 2004
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:168 Resp: 390664
Ion Ratio Lower Upper
168 100
139 38.8 29.1 43.7
169 13.4 10.4 15.6



#56
4-Nitrophenol
Concen: 15.79 ng
RT: 14.534 min Scan# 1980
Delta R.T. 0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:139 Resp: 42631
Ion Ratio Lower Upper
139 100
109 76.7 44.6 84.6
65 72.5 41.6 81.6









#57
2,4-Dinitrotoluene
Concen: 22.64 ng
RT: 14.651 min Scan# 2000
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:165 Resp: 100843




Ion	Ratio	Lower	Upper
165	100		
63	34.0	20.6	31.0#
89	64.7	42.6	63.8#
182	2.9	2.4	3.6




#58
Fluorene
Concen: 21.48 ng
RT: 15.328 min Scan# 2115
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:166 Resp: 309793

Ion	Ratio	Lower	Upper
166	100		
165	98.4	77.8	116.8
167	13.2	11.0	16.4




Abundance Scan 2049 (14.939 min): BP003089.D (-2043) (-)




m/z-->

Abundance Scan 2045 (14.916 min): BP003289.D\data.ms




m/z-->

Abundance Scan 2045 (14.916 min): BP003289.D\data.ms (-)




m/z-->

Abundance Scan 2084 (15.145 min): BP003089.D (-2076) (-)




m/z-->

Abundance Scan 2077 (15.104 min): BP003289.D\data.ms



m/z-->

Abundance Scan 2077 (15.104 min): BP003289.D\data.ms (-)




m/z-->

#59

2,3,4,6-Tetrachlorophenol
Concen: 21.78 ng
RT: 14.916 min Scan# 2045
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44


Tgt Ion:	232	Resp:	84902
Ion	Ratio	Lower	Upper
232	100		
131	40.7	34.1	51.1
130	2.3	1.8	2.8
166	29.0	24.6	36.8




#60

Diethylphthalate
Concen: 22.19 ng
RT: 15.104 min Scan# 2077
Delta R.T. -0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:	149	Resp:	346987
Ion	Ratio	Lower	Upper
149	100		
177	22.8	18.0	27.0
150	12.8	10.0	15.0









#61
4-Chlorophenyl-phenylether
Concen: 22.61 ng
RT: 15.322 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:204 Resp: 168365
Ion Ratio Lower Upper
204 100
206 33.8 25.9 38.9
141 62.9 46.5 69.7



#62
4-Nitroaniline
Concen: 21.75 ng
RT: 15.351 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:138 Resp: 83908
Ion Ratio Lower Upper
138 100
92 52.8 20.7 60.7
108 89.7 47.8 87.8#









#63
Azobenzene
Concen: 25.82 ng
RT: 15.616 min Scan# 2164
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion: 77 Resp: 267945




Ion	Ratio	Lower	Upper
77	100		
182	30.9	16.9	56.9
105	23.1	5.6	45.6
51	22.9	2.9	42.9




#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.028 min Scan# 2404
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion: 188 Resp: 471090

Ion	Ratio	Lower	Upper
188	100		
94	7.5	6.2	9.2
80	8.2	6.2	9.2









#65
4,6-Dinitro-2-methylphenol
Concen: 20.01 ng
RT: 15.422 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:198	Resp:	45074
Ion Ratio	Lower	Upper
198	100	
51	28.7	6.4 46.4
105	41.8	18.9 58.9



#66
n-Nitrosodiphenylamine
Concen: 20.44 ng
RT: 15.539 min Scan# 2151
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:169	Resp:	281551
Ion Ratio	Lower	Upper
169	100	
168	66.5	53.8 80.6
167	36.6	28.9 43.3









#67
4-Bromophenyl-phenylether
Concen: 22.19 ng
RT: 16.222 min Scan# 2267
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	248	Resp:	115374
Ion	Ratio	Lower	Upper
248	100		
250	96.6	76.7	115.1
141	67.1	53.0	79.6



#68
Hexachlorobenzene
Concen: 22.08 ng
RT: 16.345 min Scan# 2288
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:	284	Resp:	136677
Ion	Ratio	Lower	Upper
284	100		
142	29.0	24.2	36.4
249	31.6	23.8	35.8









#69
Atrazine
Concen: 23.77 ng
RT: 16.498 min Scan# 2314
Delta R.T. -0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:200 Resp: 110410
Ion Ratio Lower Upper
200 100
173 24.3 5.9 45.9
215 47.8 28.4 68.4



#70
Pentachlorophenol
Concen: 14.10 ng
RT: 16.698 min Scan# 2348
Delta R.T. 0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:266 Resp: 41736
Ion Ratio Lower Upper
266 100
268 59.1 51.1 76.7
264 62.3 50.6 76.0









#71
Phenanthrene
Concen: 20.51 ng
RT: 17.069 min Scan# 2411
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:178 Resp: 523443
Ion Ratio Lower Upper
178 100
176 19.4 15.3 22.9
179 15.4 12.5 18.7



#72
Anthracene
Concen: 21.15 ng
RT: 17.163 min Scan# 2427
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:178 Resp: 516782
Ion Ratio Lower Upper
178 100
176 18.9 15.0 22.4
179 15.3 12.5 18.7









#73
 Carbazole
 Concen: 20.90 ng
 RT: 17.439 min Scan# 2474
 Delta R.T. 0.006 min
 Lab File: BP003289.D
 Acq: 15 Sep 2020 14:44




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion:167	Resp: 490130
Ion Ratio	Lower Upper
167 100	
166 21.8	17.6 26.4
139 13.0	10.2 15.4



#74
 Di-n-butylphthalate
 Concen: 22.82 ng
 RT: 18.004 min Scan# 2570
 Delta R.T. 0.000 min
 Lab File: BP003289.D
 Acq: 15 Sep 2020 14:44

Tgt Ion:149	Resp: 629084
Ion Ratio	Lower Upper
149 100	
150 9.4	7.3 10.9
104 4.9	3.4 5.0









#75
Fluoranthene
Concen: 22.31 ng
RT: 19.051 min Scan# 2748
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:202	Resp: 658316
Ion Ratio	Lower Upper
202 100	
101 9.6	0.0 29.4
203 17.5	0.0 37.8



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.127 min Scan# 3101
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:240	Resp: 538657
Ion Ratio	Lower Upper
240 100	
120 9.0	7.8 11.6
236 25.6	20.8 31.2









#77
Benzidine
Concen: 27.29 ng
RT: 19.233 min Scan# 2779
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:184 Resp: 335912
Ion Ratio Lower Upper
184 100
185 13.7 11.5 17.3
183 12.3 9.6 14.4




#78
Pyrene
Concen: 19.47 ng
RT: 19.404 min Scan# 2808
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:202 Resp: 680757
Ion Ratio Lower Upper
202 100
200 21.3 17.1 25.7
203 17.3 14.2 21.2



Abundance Scan 2848 (19.639 min): BP003089.D (-2835) (-)



#79

Terphenyl-d14

Concen: 43.63 ng

RT: 19.604 min Scan# 2842

Delta R.T. 0.000 min


Lab File: BP003289.D

Acq: 15 Sep 2020 14:44


Tgt Ion:244 Resp: 1068141

Ion	Ratio	Lower	Upper
244	100		
212	7.0	5.8	8.6
122	8.8	7.8	11.8


Abundance Scan 2842 (19.604 min): BP003289.D\data.ms




Abundance



Abundance Scan 2842 (19.604 min): BP003289.D\data.ms (-)



Abundance Scan 2963 (20.316 min): BP003089.D (-2954) (-)



#80

Butylbenzylphthalate

Concen: 21.68 ng

RT: 20.280 min Scan# 2957

Delta R.T. 0.000 min


Lab File: BP003289.D

Acq: 15 Sep 2020 14:44


Tgt Ion:149 Resp: 312067

Ion	Ratio	Lower	Upper
149	100		
91	66.8	45.4	68.0
206	21.9	17.1	25.7


Abundance Scan 2957 (20.280 min): BP003289.D\data.ms




Abundance



Abundance Scan 2957 (20.280 min): BP003289.D\data.ms (-)









#81
Benzo(a)anthracene
Concen: 20.84 ng
RT: 21.110 min Scan# 3098
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:228 Resp: 703004




Ion	Ratio	Lower	Upper
228	100		
226	27.3	22.1	33.1
229	19.6	15.9	23.9




#82
3,3'-Dichlorobenzidine
Concen: 22.57 ng
RT: 21.045 min Scan# 3087
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:252 Resp: 280136

Ion	Ratio	Lower	Upper
252	100		
254	65.0	51.0	76.4
126	9.4	7.9	11.9









#83
Chrysene
Concen: 21.41 ng
RT: 21.163 min Scan# 3107
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	228	Resp:	691654
Ion Ratio	Lower	Upper	
228	100		
226	29.9	24.4	36.6
229	19.6	15.8	23.6



#84
Bis(2-ethylhexyl)phthalate
Concen: 21.82 ng
RT: 21.051 min Scan# 3088
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:	149	Resp:	453676
Ion Ratio	Lower	Upper	
149	100		
167	28.1	22.6	34.0
279	5.4	4.2	6.2









#85
Di-n-octyl phthalate
Concen: 22.70 ng
RT: 21.916 min Scan# 3235
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:149 Resp: 809070




Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	5.1	3.7	5.5




#86
Perylene-d12
Concen: 20.00 ng
RT: 23.351 min Scan# 3479
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:264 Resp: 594560

Ion	Ratio	Lower	Upper
264	100		
260	23.9	18.8	28.2
265	22.0	17.4	26.2









#87
Indeno(1,2,3-cd)pyrene
Concen: 20.49 ng
RT: 25.609 min Scan# 3863
Delta R.T. -0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion: 276 Resp: 908572




Ion	Ratio	Lower	Upper
276	100		
138	21.1	18.7	28.1
277	25.6	20.4	30.6




#88
Benzo(b)fluoranthene
Concen: 20.79 ng
RT: 22.680 min Scan# 3365
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion: 252 Resp: 768110

Ion	Ratio	Lower	Upper
252	100		
253	21.5	17.6	26.4
125	8.3	7.3	10.9









#89
Benzo(k)fluoranthene
Concen: 20.44 ng
RT: 22.727 min Scan# 3373
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:252 Resp: 721746




Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.8	26.8
125	8.0	7.4	11.2




#90
Benzo(a)pyrene
Concen: 20.68 ng
RT: 23.251 min Scan# 3462
Delta R.T. 0.000 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:252 Resp: 708816

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.4	26.0
125	9.3	8.2	12.4









#91
Dibenzo(a,h)anthracene
Concen: 20.66 ng
RT: 25.621 min Scan# 3865
Delta R.T. -0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	278	Resp:	752588
Ion	Ratio	Lower	Upper
278	100		
139	14.0	12.4	18.6
279	23.2	19.0	28.6



#92
Benzo(g,h,i)perylene
Concen: 20.19 ng
RT: 26.304 min Scan# 3981
Delta R.T. -0.006 min
Lab File: BP003289.D
Acq: 15 Sep 2020 14:44

Tgt Ion:	276	Resp:	737984
Ion	Ratio	Lower	Upper
276	100		
277	23.5	19.0	28.6
138	17.9	16.6	24.8

