

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091521\
 Data File : BP006981.D
 Acq On : 15 Sep 2021 18:30
 Operator : CG/JU
 Sample : PB138901BS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS901

Manual Integrations
 APPROVED

mohammad
 9/16/2021 11:44:04 AM

Quant Time: Sep 15 18:56:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP090821.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 15 17:01:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.922	152	406181	20.000	ng/ul	0.00
20) Naphthalene-d8	10.722	136	1700191	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.551	164	1036258	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.298	188	1977735	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.374	240	1519303	20.000	ng/ul	0.00
88) Perylene-d12	23.798	264	1218768	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	71366	6.992	ng/uL	0.00
4) Pyridine-d5	3.769	84	958728	36.045	ng/ul	0.00
7) Phenol-d5	7.081	99	1177656	39.448	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.251	67	686321	40.030	ng/ul	0.00
11) 2-Chlorophenol-d4	7.451	132	934912	38.792	ng/ul	0.00
15) 4-Methylphenol-d8	8.628	113	910194	38.825	ng/ul	0.00
21) Nitrobenzene-d5	9.081	128	434896	42.230	ng/ul	0.00
24) 2-Nitrophenol-d4	9.804	143	440473	43.445	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.339	165	903927	37.811	ng/ul	0.00
31) 4-Chloroaniline-d4	10.857	131	1176327	35.581	ng/ul	0.00
46) Dimethylphthalate-d6	13.963	166	2431112	36.222	ng/ul	0.00
49) Acenaphthylene-d8	14.239	160	3182966	37.445	ng/ul	0.00
54) 4-Nitrophenol-d4	14.739	143	433434	37.173	ng/ul	0.00
60) Fluorene-d10	15.539	176	2123197	35.844	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.657	200	355554	40.639	ng/ul	0.00
73) Anthracene-d10	17.398	188	3075668	37.564	ng/ul	0.00
81) Pyrene-d10	19.627	212	3195472	44.910	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.645	264	2180841	37.091	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.387	88	152660	13.441	ng/uL#	87
5) Pyridine	3.787	79	974247	36.514	ng/ul	87
6) Benzaldehyde	7.057	77	757130m	39.801	ng/ul	
8) Phenol	7.110	94	1240970	40.213	ng/ul	90
10) Bis(2-Chloroethyl)ether	7.345	93	981438	39.724	ng/ul	99
12) 2-Chlorophenol	7.487	128	980726	39.665	ng/ul	93
13) 2-Methylphenol	8.363	108	927904	39.594	ng/ul	97
14) 2,2'-oxybis(1-Chloropr...	8.451	45	1203905	42.187	ng/ul	96
16) Acetophenone	8.745	105	1433704	39.878	ng/ul	91
17) N-Nitroso-di-n-propyla...	8.739	70	681276	40.670	ng/ul#	96
18) 4-Methylphenol	8.692	108	1001536	39.846	ng/ul	98
19) Hexachloroethane	9.004	117	387726	38.509	ng/ul#	82
22) Nitrobenzene	9.122	77	1033804	41.484	ng/ul	91
23) Isophorone	9.657	82	1905681	38.503	ng/ul#	94
25) 2-Nitrophenol	9.834	139	506437	44.690	ng/ul	87
26) 2,4-Dimethylphenol	9.892	107	1026356	37.384	ng/ul	93
27) Bis(2-Chloroethoxy)met...	10.134	93	1273073	38.857	ng/ul	98
29) 2,4-Dichlorophenol	10.369	162	919495	38.837	ng/ul	95
30) Naphthalene	10.775	128	3011592	36.517	ng/ul	100
32) 4-Chloroaniline	10.881	127	1205925	35.813	ng/ul	99
33) Hexachlorobutadiene	11.063	225	568542	34.810	ng/ul	97
34) Caprolactam	11.663	113	235085m	33.313	ng/ul	
35) 4-Chloro-3-methylphenol	11.998	107	927061	38.606	ng/ul	89
36) 2-Methylnaphthalene	12.380	142	2140757	38.716	ng/ul	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091521\
 Data File : BP006981.D
 Acq On : 15 Sep 2021 18:30
 Operator : CG/JU
 Sample : PB138901BS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS901

Manual Integrations
 APPROVED

mohammad
 9/16/2021 11:44:04 AM

Quant Time: Sep 15 18:56:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP090821.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 15 17:01:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.598	142	2039047	36.150	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.745	216	1110717	37.432	ng/ul#	94
40) Hexachlorocyclopentadiene	12.728	237	621573	32.457	ng/ul	99
41) 2,4,6-Trichlorophenol	12.980	196	699329	40.718	ng/ul	100
42) 2,4,5-Trichlorophenol	13.051	196	756328	39.503	ng/ul	96
43) 1,1'-Biphenyl	13.386	154	2696223	37.369	ng/ul#	97
44) 2-Chloronaphthalene	13.427	162	2061829	37.065	ng/ul	92
45) 2-Nitroaniline	13.627	65	513209	45.392	ng/ul#	79
47) Dimethylphthalate	14.010	163	2463115	36.388	ng/ul	99
48) 2,6-Dinitrotoluene	14.127	165	481019	42.550	ng/ul#	75
50) Acenaphthylene	14.269	152	3045271	34.528	ng/ul	99
51) 3-Nitroaniline	14.451	138	497602	38.398	ng/ul	84
52) Acenaphthene	14.610	153	2179451	37.319	ng/ul	97
53) 2,4-Dinitrophenol	14.651	184	248804	39.178	ng/ul#	74
55) 4-Nitrophenol	14.751	109	319999	37.337	ng/ul#	84
56) Dibenzofuran	14.945	168	3044298	36.123	ng/ul	87
57) 2,4-Dinitrotoluene	14.910	165	671349	41.556	ng/ul#	76
58) 2,3,4,6-Tetrachlorophenol	15.169	232	593996	38.503	ng/ul#	81
59) Diethylphthalate	15.369	149	2379398	36.102	ng/ul	95
61) Fluorene	15.592	166	2449946	36.474	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.592	204	1235043	36.067	ng/ul#	79
63) 4-Nitroaniline	15.616	138	509431	39.375	ng/ul#	82
66) 4,6-Dinitro-2-methylph...	15.669	198	375629	42.169	ng/ul#	78
67) N-Nitrosodiphenylamine	15.804	169	2069289	39.727	ng/ul	99
68) 4-Bromophenyl-phenylether	16.486	248	722891	38.662	ng/ul#	83
69) Hexachlorobenzene	16.598	284	796909	36.756	ng/ul#	88
70) Atrazine	16.757	200	689798	34.969	ng/ul	95
71) Pentachlorophenol	16.945	266	468361	37.714	ng/ul	98
72) Phenanthrene	17.339	178	3670641	37.258	ng/ul	99
74) Anthracene	17.433	178	3690037	37.682	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.351	216	1134513	40.928	ng/uL	99
76) Pentachlorobenzene	14.869	250	999553	37.252	ng/uL	99
77) Carbazole	17.698	167	3205115	36.827	ng/ul	97
78) Di-n-butylphthalate	18.251	149	3623902	36.739	ng/ul#	98
80) Fluoranthene	19.304	202	3941032	44.533	ng/ul#	93
82) Pyrene	19.657	202	3993646	43.996	ng/ul#	93
83) Butylbenzylphthalate	20.527	149	1327801	42.009	ng/ul#	90
84) 3,3'-Dichlorobenzidine	21.292	252	1091127	34.611	ng/ul	98
85) Benzo(a)anthracene	21.362	228	3344492	37.809	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.286	149	1904775	40.424	ng/ul	95
87) Chrysene	21.415	228	3245394	37.075	ng/ul	100
89) Di-n-octyl phthalate	22.221	149	2875817	45.656	ng/ul	100
90) Benzo(b)fluoranthene	23.056	252	2885574	40.463	ng/ul	98
91) Benzo(k)fluoranthene	23.109	252	2922722	43.784	ng/ul	97
93) Benzo(a)pyrene	23.692	252	2565136	37.385	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	26.339	276	2953135	37.101	ng/ul	97
95) Dibenzo(a,h)anthracene	26.350	278	2526569	36.687	ng/ul	97
96) Benzo(g,h,i)perylene	27.115	276	2437258	35.703	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091521\
 Data File : BP006981.D
 Acq On : 15 Sep 2021 18:30
 Operator : CG/JU
 Sample : PB138901BS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SLCS901

Manual Integrations
 APPROVED

mohammad
 9/16/2021 11:44:04 AM


Quant Time: Sep 15 18:56:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP090821.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 15 17:01:19 2021
 Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 13.441 ng/uL
RT: 3.387 min Scan# 51
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampleId :
SLCS901




Tgt Ion: 88 Resp: 152660
Ion Ratio Lower Upper
88 100
43 26.8 20.1 30.1
58 73.8 48.4 72.6#

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#5
Pyridine
Concen: 36.514 ng/uL
RT: 3.787 min Scan# 119
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion: 79 Resp: 974247
Ion Ratio Lower Upper
79 100
52 72.6 49.2 73.8
51 33.5 22.7 34.1





#6
Benzaldehyde
Concen: 39.801 ng/ul m
RT: 7.057 min Scan# 675
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampleId :
SLCS901




Tgt Ion: 77 Resp: 757130
Ion Ratio Lower Upper
77 100
105 100.6 74.4 111.6
106 97.2 70.3 105.5

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#8
Phenol
Concen: 40.213 ng/ul
RT: 7.110 min Scan# 684
Delta R.T. 0.006 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion: 94 Resp: 1240970
Ion Ratio Lower Upper
94 100
65 25.5 24.4 36.6
66 33.4 31.8 47.6





#10
Bis(2-Chloroethyl)ether
Concen: 39.724 ng/ul
RT: 7.345 min Scan# 724
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampleId :
SLCS901




Tgt Ion	Ratio	Lower	Upper
93	100		
63	73.8	59.7	89.5
95	32.8	25.3	37.9

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#12
2-Chlorophenol
Concen: 39.665 ng/ul
RT: 7.487 min Scan# 748
Delta R.T. 0.006 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30



Tgt Ion	Ratio	Lower	Upper
128	100		
64	43.3	40.6	60.8
130	32.6	26.4	39.6





#13

2-Methylphenol

Concen: 39.594 ng/ul

RT: 8.363 min Scan# 897

Delta R.T. 0.006 min

Lab File: BP006981.D

Acq: 15 Sep 2021 18:30

Tgt Ion: 108 Resp: 927904

Ion Ratio Lower Upper

108 100

107 87.6 72.2 108.2

Instrument :

BNA_P

ClientSampleId :





SLCS901

Manual Integrations

APPROVED

mohammad

9/16/2021 11:44:04 AM



#14

2,2'-oxybis(1-Chloropropane)

Concen: 42.187 ng/ul

RT: 8.451 min Scan# 912

Delta R.T. -0.000 min

Lab File: BP006981.D

Acq: 15 Sep 2021 18:30




Tgt Ion: 45 Resp: 1203905


Ion Ratio Lower Upper

45 100

77 15.4 13.8 20.8


79 12.0 10.9 16.3





#16
Acetophenone
Concen: 39.878 ng/ul
RT: 8.745 min Scan# 962
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampleId :
SLCS901




Tgt Ion:105 Resp: 1433704
Ion Ratio Lower Upper
105 100
77 73.7 67.0 100.4
51 23.9 19.0 28.6

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#17
N-Nitroso-di-n-propylamine
Concen: 40.670 ng/ul
RT: 8.739 min Scan# 961
Delta R.T. 0.006 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30



Tgt Ion: 70 Resp: 681276
Ion Ratio Lower Upper
70 100
42 51.3 40.4 60.6
101 10.6 7.2 10.8
130 23.8 14.9 22.3#









#18
4-Methylphenol
Concen: 39.846 ng/ul
RT: 8.692 min Scan# 953
Delta R.T. 0.006 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30

Instrument :
BNA_P
ClientSampled :
SLCS901

Tgt Ion:108 Resp: 1001536
Ion Ratio Lower Upper
108 100
107 108.8 88.6 133.0




Manual Integrations
APPROVED


mohammad
9/16/2021 11:44:04 AM



#19
Hexachloroethane
Concen: 38.509 ng/ul
RT: 9.004 min Scan# 1006
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30


Tgt Ion:117 Resp: 387726
Ion Ratio Lower Upper
117 100
201 102.2 68.2 102.2
199 66.1 43.5 65.3#





#22
Nitrobenzene
Concen: 41.484 ng/ul
RT: 9.122 min Scan# 1026
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampleId :
SLCS901




Tgt Ion: 77 Resp: 1033804
Ion Ratio Lower Upper
77 100
123 46.4 31.1 46.7
65 13.0 10.6 16.0

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#23
Isophorone
Concen: 38.503 ng/ul
RT: 9.657 min Scan# 1117
Delta R.T. 0.006 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion: 82 Resp: 1905681
Ion Ratio Lower Upper
82 100
95 6.8 5.8 8.8
138 19.0 12.5 18.7#





#25
2-Nitrophenol
Concen: 44.690 ng/ul
RT: 9.834 min Scan# 1147
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampled :
SLCS901




Tgt Ion:139 Resp: 506437
Ion Ratio Lower Upper
139 100
65 49.3 48.4 72.6
109 29.0 28.2 42.2

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#26
2,4-Dimethylphenol
Concen: 37.384 ng/ul
RT: 9.892 min Scan# 1157
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion:107 Resp: 1026356
Ion Ratio Lower Upper
107 100
121 51.1 38.1 57.1
122 87.4 64.2 96.4





#27
Bis(2-Chloroethoxy)methane
Concen: 38.857 ng/ul
RT: 10.134 min Scan# 1198
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampled :
SLCS901



Tgt Ion: 93 Resp: 1273073
Ion Ratio Lower Upper
93 100
95 32.5 25.6 38.4
123 11.6 8.2 12.4




Manual Integrations
APPROVED


mohammad
9/16/2021 11:44:04 AM



#29
2,4-Dichlorophenol
Concen: 38.837 ng/ul
RT: 10.369 min Scan# 1238
Delta R.T. 0.006 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30


Tgt Ion: 162 Resp: 919495
Ion Ratio Lower Upper
162 100
164 65.1 52.0 78.0
98 34.6 34.3 51.5





#30
Naphthalene
Concen: 36.517 ng/ul
RT: 10.775 min Scan# 1307
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampled :
SLCS901



Tgt Ion:128 Resp: 3011592
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 12.9 10.6 15.8




Manual Integrations
APPROVED


mohammad
9/16/2021 11:44:04 AM



#32
4-Chloroaniline
Concen: 35.813 ng/ul
RT: 10.881 min Scan# 1325
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30


Tgt Ion:127 Resp: 1205925
Ion Ratio Lower Upper
127 100
129 32.6 25.7 38.5





#33
Hexachlorobutadiene
Concen: 34.810 ng/ul
RT: 11.063 min Scan# 1356
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampleId :
SLCS901




Tgt Ion:225 Resp: 568542
Ion Ratio Lower Upper
225 100
223 62.8 48.2 72.2
227 63.8 49.6 74.4

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#34
Caprolactam
Concen: 33.313 ng/ul m
RT: 11.663 min Scan# 1458
Delta R.T. 0.011 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion:113 Resp: 235085
Ion Ratio Lower Upper
113 100
55 162.0 125.8 188.6
56 118.5 98.2 147.2





#35
4-Chloro-3-methylphenol
Concen: 38.606 ng/ul
RT: 11.998 min Scan# 1515
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampled :
SLCS901



Tgt Ion:107 Resp: 927061
Ion Ratio Lower Upper
107 100
144 26.8 18.8 28.2
142 83.0 58.4 87.6




Manual Integrations
APPROVED


mohammad
9/16/2021 11:44:04 AM



#36
2-Methylnaphthalene
Concen: 38.716 ng/ul
RT: 12.380 min Scan# 1580
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30


Tgt Ion:142 Resp: 2140757
Ion Ratio Lower Upper
142 100
141 87.9 72.3 108.5





#37
1-Methylnaphthalene
Concen: 36.150 ng/ul
RT: 12.598 min Scan# 1617
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampleId :
SLCS901




Tgt Ion:142 Resp: 2039047
Ion Ratio Lower Upper
142 100
141 91.4 74.9 112.3
116 3.6 3.3 4.9

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#39
1,2,4,5-Tetrachlorobenzene
Concen: 37.432 ng/ul
RT: 12.745 min Scan# 1642
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion:216 Resp: 1110717
Ion Ratio Lower Upper
216 100
214 78.1 64.3 96.5
179 20.3 19.3 28.9
108 16.7 19.8 29.6#





#40
Hexachlorocyclopentadiene
Concen: 32.457 ng/ul
RT: 12.728 min Scan# 1639
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampled :
SLCS901




Tgt Ion:237 Resp: 621573
Ion Ratio Lower Upper
237 100
235 63.9 50.4 75.6
272 12.9 9.7 14.5

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#41
2,4,6-Trichlorophenol
Concen: 40.718 ng/ul
RT: 12.980 min Scan# 1682
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion:196 Resp: 699329
Ion Ratio Lower Upper
196 100
198 95.3 76.2 114.4
200 31.2 24.6 36.8





#42
2,4,5-Trichlorophenol
Concen: 39.503 ng/ul
RT: 13.051 min Scan# 1694
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampleId :
SLCS901




Tgt Ion:196 Resp: 756328
Ion Ratio Lower Upper
196 100
198 94.8 79.3 118.9
200 30.9 25.8 38.8

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#43
1,1'-Biphenyl
Concen: 37.369 ng/ul
RT: 13.386 min Scan# 1751
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion:154 Resp: 2696223
Ion Ratio Lower Upper
154 100
153 40.1 32.4 48.6
76 12.3 12.9 19.3#





#44
2-Chloronaphthalene
Concen: 37.065 ng/ul
RT: 13.427 min Scan# 1758
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampleId :
SLCS901




Tgt Ion:162 Resp: 2061829
Ion Ratio Lower Upper
162 100
164 32.4 25.6 38.4
127 37.0 36.2 54.2

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#45
2-Nitroaniline
Concen: 45.392 ng/ul
RT: 13.627 min Scan# 1792
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion: 65 Resp: 513209
Ion Ratio Lower Upper
65 100
92 75.0 53.7 80.5
138 116.3 71.4 107.2#





#47
Dimethylphthalate
Concen: 36.388 ng/ul
RT: 14.010 min Scan# 1857
Delta R.T. 0.006 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampleId :
SLCS901




Tgt Ion:163 Resp: 2463115
Ion Ratio Lower Upper
163 100
194 4.3 3.0 4.4
164 9.9 7.8 11.8

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#48
2,6-Dinitrotoluene
Concen: 42.550 ng/ul
RT: 14.127 min Scan# 1877
Delta R.T. 0.006 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion:165 Resp: 481019
Ion Ratio Lower Upper
165 100
89 52.3 62.2 93.2#
121 20.5 22.1 33.1#





#50
Acenaphthylene
Concen: 34.528 ng/ul
RT: 14.269 min Scan# 1901
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampleId :
SLCS901




Tgt Ion:152 Resp: 3045271
Ion Ratio Lower Upper
152 100
151 20.2 15.9 23.9
153 13.4 10.4 15.6

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#51
3-Nitroaniline
Concen: 38.398 ng/ul
RT: 14.451 min Scan# 1932
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion:138 Resp: 497602
Ion Ratio Lower Upper
138 100
108 11.3 9.7 14.5
92 113.4 106.7 160.1





#52
Acenaphthene
Concen: 37.319 ng/ul
RT: 14.610 min Scan# 1959
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampleId :
SLCS901




Tgt Ion:153 Resp: 2179451
Ion Ratio Lower Upper
153 100
152 47.8 38.6 57.8
154 89.6 69.0 103.6

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#53
2,4-Dinitrophenol
Concen: 39.178 ng/ul
RT: 14.651 min Scan# 1966
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion:184 Resp: 248804
Ion Ratio Lower Upper
184 100
63 61.5 73.4 110.2#
154 57.4 59.4 89.0#





#55
4-Nitrophenol
Concen: 37.337 ng/ul
RT: 14.751 min Scan# 1983
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampleId :
SLCS901



Tgt Ion:109 Resp: 319999
Ion Ratio Lower Upper
109 100
139 145.0 90.3 135.5#
65 132.7 102.2 153.2




Manual Integrations
APPROVED


mohammad
9/16/2021 11:44:04 AM



#56
Dibenzofuran
Concen: 36.123 ng/ul
RT: 14.945 min Scan# 2016
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30


Tgt Ion:168 Resp: 3044298
Ion Ratio Lower Upper
168 100
139 35.7 35.3 52.9





#57
2,4-Dinitrotoluene
Concen: 41.556 ng/ul
RT: 14.910 min Scan# 2010
Delta R.T. 0.006 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampleId :
SLCS901




Tgt Ion:165 Resp: 671349
Ion Ratio Lower Upper
165 100
63 38.5 45.2 67.8#
182 2.7 1.9 2.9

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#58
2,3,4,6-Tetrachlorophenol
Concen: 38.503 ng/ul
RT: 15.169 min Scan# 2054
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion:232 Resp: 593996
Ion Ratio Lower Upper
232 100
131 45.8 51.8 77.6#
130 2.4 3.1 4.7#
166 31.4 31.0 46.4





#59
Diethylphthalate
Concen: 36.102 ng/ul
RT: 15.369 min Scan# 2088
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampled :
SLCS901




Tgt Ion:149 Resp: 2379398
Ion Ratio Lower Upper
149 100
177 22.8 15.5 23.3
150 12.4 9.9 14.9

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#61
Fluorene
Concen: 36.474 ng/ul
RT: 15.592 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion:166 Resp: 2449946
Ion Ratio Lower Upper
166 100
165 96.3 79.6 119.4
167 13.7 10.7 16.1





#62
4-Chlorophenyl-phenylether
Concen: 36.067 ng/ul
RT: 15.592 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampled :
SLCS901




Tgt Ion:204 Resp: 1235043
Ion Ratio Lower Upper
204 100
206 33.5 26.3 39.5
141 52.2 62.0 93.0#

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#63
4-Nitroaniline
Concen: 39.375 ng/ul
RT: 15.616 min Scan# 2130
Delta R.T. 0.006 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion:138 Resp: 509431
Ion Ratio Lower Upper
138 100
92 48.1 47.1 70.7
108 72.9 73.8 110.8#





#66
4,6-Dinitro-2-methylphenol
Concen: 42.169 ng/ul
RT: 15.669 min Scan# 2139
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampled :
SLCS901




Tgt Ion:198 Resp: 375629
Ion Ratio Lower Upper
198 100
182 4.0 4.5 6.7#
77 24.3 31.7 47.5#

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#67
N-Nitrosodiphenylamine
Concen: 39.727 ng/ul
RT: 15.804 min Scan# 2162
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion:169 Resp: 2069289
Ion Ratio Lower Upper
169 100
168 67.1 53.9 80.9
167 36.6 30.0 45.0





#68
4-Bromophenyl-phenylether
Concen: 38.662 ng/ul
RT: 16.486 min Scan# 2278
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampled :
SLCS901




Tgt Ion:248 Resp: 722891
Ion Ratio Lower Upper
248 100
250 97.4 78.4 117.6
141 68.0 80.6 121.0#

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#69
Hexachlorobenzene
Concen: 36.756 ng/ul
RT: 16.598 min Scan# 2297
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30



Tgt Ion:284 Resp: 796909
Ion Ratio Lower Upper
284 100
142 35.9 37.8 56.6#
249 31.5 27.1 40.7









#70
Atrazine
Concen: 34.969 ng/ul
RT: 16.757 min Scan# 2324
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30

Tgt Ion	Ratio	Lower	Upper
200	100		
173	26.4	22.2	33.4
215	48.7	35.4	53.0

Instrument :
BNA_P
ClientSampled :
SLCS901




Manual Integrations
APPROVED


mohammad
9/16/2021 11:44:04 AM



#71
Pentachlorophenol
Concen: 37.714 ng/ul
RT: 16.945 min Scan# 2356
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30


Tgt Ion	Ratio	Lower	Upper
266	100		
264	61.8	51.0	76.4
268	62.8	49.4	74.2





#72
Phenanthrene
Concen: 37.258 ng/ul
RT: 17.339 min Scan# 2423
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampleId :
SLCS901




Tgt Ion:178 Resp: 3670641
Ion Ratio Lower Upper
178 100
179 15.7 12.0 18.0
176 19.2 15.6 23.4

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#74
Anthracene
Concen: 37.682 ng/ul
RT: 17.433 min Scan# 2439
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30



Tgt Ion:178 Resp: 3690037
Ion Ratio Lower Upper
178 100
179 15.5 12.1 18.1
176 18.8 14.9 22.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 40.928 ng/uL
RT: 13.351 min Scan# 1745
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30

Instrument :
BNA_P
ClientSampled :
SLCS901

Tgt Ion:216 Resp: 1134513
Ion Ratio Lower Upper
216 100
214 77.9 62.9 94.3
218 47.9 37.8 56.8




Manual Integrations
APPROVED


mohammad
9/16/2021 11:44:04 AM



#76
Pentachlorobenzene
Concen: 37.252 ng/uL
RT: 14.869 min Scan# 2003
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30

Tgt Ion:250 Resp: 999553
Ion Ratio Lower Upper
250 100
248 63.2 51.0 76.6
252 64.6 51.0 76.4









#77
Carbazole
Concen: 36.827 ng/ul
RT: 17.698 min Scan# 2484
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30

Tgt Ion:167 Resp: 3205115
Ion Ratio Lower Upper
167 100
166 21.8 17.6 26.4
139 12.3 11.9 17.9

Instrument :
BNA_P
ClientSampleId :
SLCS901




Manual Integrations
APPROVED


mohammad
9/16/2021 11:44:04 AM



#78
Di-n-butylphthalate
Concen: 36.739 ng/ul
RT: 18.251 min Scan# 2578
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30


Tgt Ion:149 Resp: 3623902
Ion Ratio Lower Upper
149 100
150 9.4 7.3 10.9
104 4.8 4.8 7.2#





#80
Fluoranthene
Concen: 44.533 ng/ul
RT: 19.304 min Scan# 2757
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampled :
SLCS901




Tgt Ion:202 Resp: 3941032
Ion Ratio Lower Upper
202 100
101 10.7 11.0 16.4#
100 8.2 8.3 12.5#

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#82
Pyrene
Concen: 43.996 ng/ul
RT: 19.657 min Scan# 2817
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion:202 Resp: 3993646
Ion Ratio Lower Upper
202 100
101 12.5 12.6 19.0#
100 10.4 10.6 15.8#





#83
Butylbenzylphthalate
Concen: 42.009 ng/ul
RT: 20.527 min Scan# 2965
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampleId :
SLCS901




Tgt Ion:149 Resp: 1327801
Ion Ratio Lower Upper
149 100
91 69.5 63.0 94.6
206 21.6 14.1 21.1#

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#84
3,3'-Dichlorobenzidine
Concen: 34.611 ng/ul
RT: 21.292 min Scan# 3095
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion:252 Resp: 1091127
Ion Ratio Lower Upper
252 100
254 65.4 51.6 77.4
126 11.0 10.7 16.1





#85
Benzo(a)anthracene
Concen: 37.809 ng/ul
RT: 21.362 min Scan# 3107
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampled :
SLCS901




Tgt Ion:228 Resp: 3344492
Ion Ratio Lower Upper
228 100
229 19.8 15.8 23.8
226 27.6 21.6 32.4

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#86
Bis(2-ethylhexyl)phthalate
Concen: 40.424 ng/ul
RT: 21.286 min Scan# 3094
Delta R.T. -0.006 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion:149 Resp: 1904775
Ion Ratio Lower Upper
149 100
167 28.1 20.2 30.2
279 4.0 2.7 4.1





#87
Chrysene
Concen: 37.075 ng/ul
RT: 21.415 min Scan# 3116
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampleId :
SLCS901



Tgt Ion:228 Resp: 3245394
Ion Ratio Lower Upper
228 100
226 30.5 24.5 36.7
229 19.9 15.9 23.9




Manual Integrations
APPROVED


mohammad
9/16/2021 11:44:04 AM



#89
Di-n-octyl phthalate
Concen: 45.656 ng/ul
RT: 22.221 min Scan# 3253
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30


Tgt Ion:149 Resp: 2875817





#90
Benzo(b)fluoranthene
Concen: 40.463 ng/ul
RT: 23.056 min Scan# 3395
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampleId :
SLCS901




Tgt Ion:252 Resp: 2885574
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 10.3 10.1 15.1

Manual Integrations
APPROVED



mohammad
9/16/2021 11:44:04 AM




#91
Benzo(k)fluoranthene
Concen: 43.784 ng/ul
RT: 23.109 min Scan# 3404
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Tgt Ion:252 Resp: 2922722
Ion Ratio Lower Upper
252 100
253 21.9 17.2 25.8
125 9.5 9.4 14.0





#93
Benzo(a)pyrene
Concen: 37.385 ng/ul
RT: 23.692 min Scan# 3503
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampled :
SLCS901



Tgt Ion:252 Resp: 2565136
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 11.1 11.3 16.9#




Manual Integrations
APPROVED


mohammad
9/16/2021 11:44:04 AM



#94
Indeno(1,2,3-cd)pyrene
Concen: 37.101 ng/ul
RT: 26.339 min Scan# 3953
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30


Tgt Ion:276 Resp: 2953135
Ion Ratio Lower Upper
276 100
138 24.3 21.8 32.8
277 25.0 19.9 29.9





#95
Dibenzo(a,h)anthracene
Concen: 36.687 ng/ul
RT: 26.350 min Scan# 3955
Delta R.T. -0.006 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30




Instrument :
BNA_P
ClientSampled :
SLCS901




Tgt Ion:278 Resp: 2526569
Ion Ratio Lower Upper
278 100
139 16.3 15.4 23.2
279 23.7 19.0 28.4

Manual Integrations
APPROVED

mohammad
9/16/2021 11:44:04 AM



#96
Benzo(g,h,i)perylene
Concen: 35.703 ng/ul
RT: 27.115 min Scan# 4085
Delta R.T. -0.000 min
Lab File: BP006981.D
Acq: 15 Sep 2021 18:30



Tgt Ion:276 Resp: 2437258
Ion Ratio Lower Upper
276 100
138 22.0 20.7 31.1
277 24.0 19.0 28.4

