

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091525\
 Data File : BP025746.D
 Acq On : 15 Sep 2025 10:06
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 15 10:51:08 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 16:45:04 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	121	-0.01
2	1,4-Dioxane	40.000	41.155	-2.9	118	0.00
3	Pyridine	40.000	42.627	-6.6	118	0.00
4	n-Nitrosodimethylamine	40.000	42.615	-6.5	120	0.00
5 S	2-Fluorophenol	80.000	86.602	-8.3	119	0.00
6	Aniline	40.000	43.249	-8.1	118	0.00
7 S	Phenol-d6	80.000	87.689	-9.6	117	-0.01
8	2-Chlorophenol	40.000	43.380	-8.5	118	-0.01
9	Benzaldehyde	40.000	34.760	13.1	118	0.00
10 C	Phenol	40.000	43.628	-9.1	116	0.00
11	bis(2-Chloroethyl)ether	40.000	42.895	-7.2	116	0.00
12	1,3-Dichlorobenzene	40.000	41.890	-4.7	117	0.00
13 C	1,4-Dichlorobenzene	40.000	42.373	-5.9	119	0.00
14	1,2-Dichlorobenzene	40.000	42.227	-5.6	118	0.00
15	Benzyl Alcohol	40.000	42.192	-5.5	115	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	43.206	-8.0	119	0.00
17	2-Methylphenol	40.000	43.500	-8.8	116	0.00
18	Hexachloroethane	40.000	41.960	-4.9	118	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	44.479	-11.2	117	0.00
20	3+4-Methylphenols	40.000	42.426	-6.1	115	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	116	0.00
22	Acetophenone	40.000	43.708	-9.3	115	0.00
23 S	Nitrobenzene-d5	80.000	86.942	-8.7	115	0.00
24	Nitrobenzene	40.000	44.079	-10.2	116	-0.01
25	Isophorone	40.000	43.205	-8.0	116	0.00
26 C	2-Nitrophenol	40.000	44.682	-11.7	115	0.00
27	2,4-Dimethylphenol	40.000	43.989	-10.0	113	-0.01
28	bis(2-Chloroethoxy)methane	40.000	44.055	-10.1	118	0.00
29 C	2,4-Dichlorophenol	40.000	45.111	-12.8	118	0.00
30	1,2,4-Trichlorobenzene	40.000	42.725	-6.8	116	0.00
31	Naphthalene	40.000	42.618	-6.5	116	0.00
32	Benzoic acid	40.000	40.983	-2.5	107	0.00
33	4-Chloroaniline	40.000	45.107	-12.8	115	0.00
34 C	Hexachlorobutadiene	40.000	42.615	-6.5	117	0.00
35	Caprolactam	40.000	41.837	-4.6	112	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.284	-8.2	114	0.00
37	2-Methylnaphthalene	40.000	42.776	-6.9	115	0.00
38	1-Methylnaphthalene	40.000	42.254	-5.6	113	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	114	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	44.150	-10.4	116	0.00
41 P	Hexachlorocyclopentadiene	40.000	44.209	-10.5	110	0.00
42 S	2,4,6-Tribromophenol	80.000	86.862	-8.6	112	-0.02
43 C	2,4,6-Trichlorophenol	40.000	46.004	-15.0	117	0.00
44	2,4,5-Trichlorophenol	40.000	44.791	-12.0	113	0.00
45 S	2-Fluorobiphenyl	80.000	85.507	-6.9	115	0.00
46	1,1'-Biphenyl	40.000	44.426	-11.1	117	0.00
47	2-Chloronaphthalene	40.000	44.097	-10.2	117	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091525\
 Data File : BP025746.D
 Acq On : 15 Sep 2025 10:06
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 15 10:51:08 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 16:45:04 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.388	-11.0	111	0.00
49	Acenaphthylene	40.000	43.219	-8.0	115	0.00
50	Dimethylphthalate	40.000	42.955	-7.4	113	0.00
51	2,6-Dinitrotoluene	40.000	44.837	-12.1	116	0.00
52 C	Acenaphthene	40.000	42.413	-6.0	112	0.00
53	3-Nitroaniline	40.000	46.224	-15.6	114	0.00
54 P	2,4-Dinitrophenol	40.000	42.760	-6.9	116	0.00
55	Dibenzofuran	40.000	42.345	-5.9	113	0.00
56 P	4-Nitrophenol	40.000	43.992	-10.0	114	0.00
57	2,4-Dinitrotoluene	40.000	45.553	-13.9	115	0.00
58	Fluorene	40.000	41.935	-4.8	111	-0.02
59	2,3,4,6-Tetrachlorophenol	40.000	44.432	-11.1	113	0.00
60	Diethylphthalate	40.000	42.462	-6.2	114	0.00
61	4-Chlorophenyl-phenylether	40.000	41.883	-4.7	112	0.00
62	4-Nitroaniline	40.000	46.040	-15.1	113	-0.01
63	Azobenzene	40.000	44.060	-10.2	115	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	115	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.982	-12.5	114	-0.01
66 c	n-Nitrosodiphenylamine	40.000	43.162	-7.9	114	0.00
67	4-Bromophenyl-phenylether	40.000	43.480	-8.7	115	0.00
68	Hexachlorobenzene	40.000	42.835	-7.1	116	0.00
69	Atrazine	40.000	44.224	-10.6	114	-0.01
70 C	Pentachlorophenol	40.000	44.609	-11.5	116	0.00
71	Phenanthrene	40.000	42.390	-6.0	114	0.00
72	Anthracene	40.000	42.928	-7.3	115	0.00
73	Carbazole	40.000	43.453	-8.6	114	0.00
74	Di-n-butylphthalate	40.000	43.975	-9.9	115	0.00
75 C	Fluoranthene	40.000	43.094	-7.7	116	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	113	-0.02
77	Benzydine	40.000	47.046	-17.6	118	0.00
78	Pyrene	40.000	44.033	-10.1	116	0.00
79 S	Terphenyl-d14	80.000	84.624	-5.8	113	-0.01
80	Butylbenzylphthalate	40.000	44.864	-12.2	115	-0.02
81	Benzo(a)anthracene	40.000	43.507	-8.8	116	-0.01
82	3,3'-Dichlorobenzidine	40.000	46.162	-15.4	119	-0.02
83	Chrysene	40.000	43.011	-7.5	116	-0.02
84	Bis(2-ethylhexyl)phthalate	40.000	44.549	-11.4	114	-0.02
85 c	Di-n-octyl phthalate	40.000	44.655	-11.6	115	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	117	-0.03
87	Indeno(1,2,3-cd)pyrene	40.000	43.362	-8.4	115	-0.04
88	Benzo(b)fluoranthene	40.000	43.727	-9.3	117	-0.02
89	Benzo(k)fluoranthene	40.000	43.308	-8.3	118	-0.02
90 C	Benzo(a)pyrene	40.000	43.012	-7.5	115	-0.02
91	Dibenzo(a,h)anthracene	40.000	43.199	-8.0	116	-0.06
92	Benzo(g,h,i)perylene	40.000	43.124	-7.8	116	-0.06

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091525\
Data File : BP025746.D
Acq On : 15 Sep 2025 10:06
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Sep 15 10:51:08 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 11 16:45:04 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0