

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091621\
 Data File : BP006998.D
 Acq On : 16 Sep 2021 16:34
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 16 17:01:57 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 16 15:16:52 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	98	0.00
2	1,4-Dioxane	40.000	38.539	3.7	103	0.00
3	Pyridine	40.000	40.626	-1.6	105	0.00
4	n-Nitrosodimethylamine	40.000	40.757	-1.9	107	0.00
5 S	2-Fluorophenol	80.000	78.022	2.5	101	0.00
6	Aniline	40.000	38.485	3.8	98	0.00
7 S	Phenol-d6	80.000	77.577	3.0	100	0.00
8	2-Chlorophenol	40.000	38.092	4.8	98	0.00
9	Benzaldehyde	40.000	42.518	-6.3	101	0.00
10 C	Phenol	40.000	38.565	3.6	99	0.00
11	bis(2-Chloroethyl)ether	40.000	37.637	5.9	98	0.00
12	1,3-Dichlorobenzene	40.000	37.618	6.0	98	0.00
13 C	1,4-Dichlorobenzene	40.000	37.293	6.8	98	0.00
14	1,2-Dichlorobenzene	40.000	37.494	6.3	98	0.00
15	Benzyl Alcohol	40.000	38.723	3.2	98	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.701	3.2	101	0.00
17	2-Methylphenol	40.000	38.106	4.7	96	0.00
18	Hexachloroethane	40.000	38.164	4.6	99	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.250	4.4	97	0.00
20	3+4-Methylphenols	40.000	38.259	4.4	96	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	95	0.00
22	Acetophenone	40.000	38.444	3.9	96	0.00
23 S	Nitrobenzene-d5	80.000	79.652	0.4	98	0.00
24	Nitrobenzene	40.000	39.361	1.6	99	0.00
25	Isophorone	40.000	38.615	3.5	94	0.00
26 C	2-Nitrophenol	40.000	39.670	0.8	94	0.00
27	2,4-Dimethylphenol	40.000	38.768	3.1	96	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.087	4.8	95	0.00
29 C	2,4-Dichlorophenol	40.000	38.433	3.9	94	0.00
30	1,2,4-Trichlorobenzene	40.000	37.518	6.2	94	0.00
31	Naphthalene	40.000	37.735	5.7	96	0.00
32	Benzoic acid	40.000	34.050	14.9	90	0.00
33	4-Chloroaniline	40.000	38.079	4.8	94	0.00
34 C	Hexachlorobutadiene	40.000	37.388	6.5	94	0.00
35	Caprolactam	40.000	38.711	3.2	91	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.848	5.4	93	0.00
37	2-Methylnaphthalene	40.000	37.048	7.4	93	0.00
38	1-Methylnaphthalene	40.000	37.140	7.1	93	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	91	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.012	5.0	91	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.110	2.2	91	0.00
42 S	2,4,6-Tribromophenol	80.000	76.689	4.1	88	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.318	1.7	90	0.00
44	2,4,5-Trichlorophenol	40.000	39.185	2.0	92	0.00
45 S	2-Fluorobiphenyl	80.000	76.360	4.6	92	0.00
46	1,1'-Biphenyl	40.000	38.208	4.5	92	0.00
47	2-Chloronaphthalene	40.000	38.049	4.9	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091621\
 Data File : BP006998.D
 Acq On : 16 Sep 2021 16:34
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 16 17:01:57 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 16 15:16:52 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.761	-4.4	94	0.00
49	Acenaphthylene	40.000	38.521	3.7	91	0.00
50	Dimethylphthalate	40.000	37.538	6.2	90	0.00
51	2,6-Dinitrotoluene	40.000	38.886	2.8	89	0.00
52 C	Acenaphthene	40.000	37.864	5.3	91	0.00
53	3-Nitroaniline	40.000	40.072	-0.2	91	0.00
54 P	2,4-Dinitrophenol	40.000	37.244	6.9	88	0.00
55	Dibenzofuran	40.000	37.661	5.8	91	0.00
56 P	4-Nitrophenol	40.000	40.647	-1.6	89	0.00
57	2,4-Dinitrotoluene	40.000	40.032	-0.1	89	0.00
58	Fluorene	40.000	37.654	5.9	91	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.388	4.0	90	0.00
60	Diethylphthalate	40.000	37.294	6.8	89	0.00
61	4-Chlorophenyl-phenylether	40.000	37.308	6.7	89	0.00
62	4-Nitroaniline	40.000	40.895	-2.2	92	0.00
63	Azobenzene	40.000	39.373	1.6	92	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	88	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.772	3.1	88	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.862	2.8	90	0.00
67	4-Bromophenyl-phenylether	40.000	38.241	4.4	88	0.00
68	Hexachlorobenzene	40.000	38.099	4.8	89	0.00
69	Atrazine	40.000	41.124	-2.8	89	0.00
70 C	Pentachlorophenol	40.000	40.574	-1.4	89	0.00
71	Phenanthrene	40.000	38.282	4.3	90	0.00
72	Anthracene	40.000	38.633	3.4	89	0.00
73	Carbazole	40.000	39.075	2.3	90	0.00
74	Di-n-butylphthalate	40.000	39.476	1.3	89	0.00
75 C	Fluoranthene	40.000	38.488	3.8	90	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	92	0.00
77	Benzydine	40.000	37.120	7.2	84	0.00
78	Pyrene	40.000	38.057	4.9	91	0.00
79 S	Terphenyl-d14	80.000	76.143	4.8	90	0.00
80	Butylbenzylphthalate	40.000	39.101	2.2	89	0.00
81	Benzo(a)anthracene	40.000	38.172	4.6	92	0.00
82	3,3'-Dichlorobenzidine	40.000	40.008	-0.0	91	0.00
83	Chrysene	40.000	37.866	5.3	92	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.257	1.9	89	0.00
85 c	Di-n-octyl phthalate	40.000	39.902	0.2	90	0.00
86 I	Perylene-d12	20.000	20.000	0.0	94	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	39.264	1.8	97	-0.02
88	Benzo(b)fluoranthene	40.000	38.765	3.1	96	0.00
89	Benzo(k)fluoranthene	40.000	38.635	3.4	93	-0.01
90 C	Benzo(a)pyrene	40.000	39.130	2.2	94	-0.02
91	Dibenzo(a,h)anthracene	40.000	39.366	1.6	97	-0.02
92	Benzo(g,h,i)perylene	40.000	39.251	1.9	97	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091621\
Data File : BP006998.D
Acq On : 16 Sep 2021 16:34
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Sep 16 17:01:57 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 16 15:16:52 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0