

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091721\
 Data File : BP007024.D
 Acq On : 17 Sep 2021 13:50
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 17 14:32:36 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 17 10:50:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	86	0.00
2	1,4-Dioxane	40.000	37.282	6.8	88	0.00
3	Pyridine	40.000	39.013	2.5	89	0.00
4	n-Nitrosodimethylamine	40.000	38.893	2.8	90	0.00
5 S	2-Fluorophenol	80.000	76.377	4.5	87	0.00
6	Aniline	40.000	38.082	4.8	86	0.00
7 S	Phenol-d6	80.000	76.331	4.6	86	0.00
8	2-Chlorophenol	40.000	37.798	5.5	86	0.00
9	Benzaldehyde	40.000	42.164	-5.4	88	0.00
10 C	Phenol	40.000	37.735	5.7	86	0.00
11	bis(2-Chloroethyl)ether	40.000	37.298	6.8	86	0.00
12	1,3-Dichlorobenzene	40.000	37.392	6.5	86	0.00
13 C	1,4-Dichlorobenzene	40.000	37.326	6.7	86	0.00
14	1,2-Dichlorobenzene	40.000	37.360	6.6	86	0.00
15	Benzyl Alcohol	40.000	38.592	3.5	86	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.078	4.8	88	0.00
17	2-Methylphenol	40.000	38.407	4.0	86	0.00
18	Hexachloroethane	40.000	37.879	5.3	87	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.724	3.2	86	0.00
20	3+4-Methylphenols	40.000	38.674	3.3	85	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	86	0.00
22	Acetophenone	40.000	37.946	5.1	86	0.00
23 S	Nitrobenzene-d5	80.000	77.409	3.2	86	0.00
24	Nitrobenzene	40.000	38.113	4.7	86	0.00
25	Isophorone	40.000	38.847	2.9	86	0.00
26 C	2-Nitrophenol	40.000	40.937	-2.3	88	0.00
27	2,4-Dimethylphenol	40.000	38.122	4.7	85	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.923	5.2	86	0.00
29 C	2,4-Dichlorophenol	40.000	38.855	2.9	86	0.00
30	1,2,4-Trichlorobenzene	40.000	37.288	6.8	85	0.00
31	Naphthalene	40.000	37.209	7.0	85	0.00
32	Benzoic acid	40.000	36.334	9.2	88	0.00
33	4-Chloroaniline	40.000	38.710	3.2	86	0.00
34 C	Hexachlorobutadiene	40.000	37.837	5.4	86	0.00
35	Caprolactam	40.000	41.678	-4.2	89	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.994	2.5	86	0.00
37	2-Methylnaphthalene	40.000	37.630	5.9	85	0.00
38	1-Methylnaphthalene	40.000	37.446	6.4	85	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	86	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.912	5.2	86	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.980	7.6	81	0.00
42 S	2,4,6-Tribromophenol	80.000	81.306	-1.6	88	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.672	0.8	86	0.00
44	2,4,5-Trichlorophenol	40.000	39.282	1.8	87	0.00
45 S	2-Fluorobiphenyl	80.000	76.349	4.6	86	0.00
46	1,1'-Biphenyl	40.000	37.650	5.9	86	0.00
47	2-Chloronaphthalene	40.000	37.913	5.2	86	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091721\
 Data File : BP007024.D
 Acq On : 17 Sep 2021 13:50
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 17 14:32:36 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 17 10:50:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.540	-3.8	88	0.00
49	Acenaphthylene	40.000	38.932	2.7	86	0.00
50	Dimethylphthalate	40.000	38.921	2.7	88	0.00
51	2,6-Dinitrotoluene	40.000	40.307	-0.8	87	0.00
52 C	Acenaphthene	40.000	38.037	4.9	86	0.00
53	3-Nitroaniline	40.000	41.173	-2.9	88	0.00
54 P	2,4-Dinitrophenol	40.000	39.701	0.7	88	0.00
55	Dibenzofuran	40.000	38.208	4.5	87	0.00
56 P	4-Nitrophenol	40.000	42.222	-5.6	87	0.00
57	2,4-Dinitrotoluene	40.000	41.787	-4.5	88	0.00
58	Fluorene	40.000	38.685	3.3	88	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.761	0.6	88	0.00
60	Diethylphthalate	40.000	39.419	1.5	89	0.00
61	4-Chlorophenyl-phenylether	40.000	38.187	4.5	86	0.00
62	4-Nitroaniline	40.000	42.400	-6.0	89	0.00
63	Azobenzene	40.000	39.969	0.1	88	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	88	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.116	2.2	89	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.900	5.3	88	0.00
67	4-Bromophenyl-phenylether	40.000	37.546	6.1	86	0.00
68	Hexachlorobenzene	40.000	37.672	5.8	88	0.00
69	Atrazine	40.000	41.734	-4.3	90	0.00
70 C	Pentachlorophenol	40.000	41.303	-3.3	90	0.00
71	Phenanthrene	40.000	37.770	5.6	89	0.00
72	Anthracene	40.000	38.624	3.4	89	0.00
73	Carbazole	40.000	39.063	2.3	89	0.00
74	Di-n-butylphthalate	40.000	41.059	-2.6	92	0.00
75 C	Fluoranthene	40.000	39.051	2.4	91	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	95	0.00
77	Benzydine	40.000	39.590	1.0	91	0.00
78	Pyrene	40.000	37.315	6.7	92	0.00
79 S	Terphenyl-d14	80.000	75.628	5.5	92	0.00
80	Butylbenzylphthalate	40.000	40.764	-1.9	95	0.00
81	Benzo(a)anthracene	40.000	38.527	3.7	96	0.00
82	3,3'-Dichlorobenzidine	40.000	41.006	-2.5	96	0.00
83	Chrysene	40.000	37.804	5.5	94	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.471	-3.7	96	0.00
85 c	Di-n-octyl phthalate	40.000	42.955	-7.4	99	0.00
86 I	Perylene-d12	20.000	20.000	0.0	97	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.745	3.1	98	0.00
88	Benzo(b)fluoranthene	40.000	38.392	4.0	98	0.00
89	Benzo(k)fluoranthene	40.000	38.790	3.0	96	0.00
90 C	Benzo(a)pyrene	40.000	38.885	2.8	97	0.00
91	Dibenzo(a,h)anthracene	40.000	38.684	3.3	98	-0.01
92	Benzo(g,h,i)perylene	40.000	38.403	4.0	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091721\
Data File : BP007024.D
Acq On : 17 Sep 2021 13:50
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Sep 17 14:32:36 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 17 10:50:16 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0