

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091725\
 Data File : BP025777.D
 Acq On : 17 Sep 2025 22:37
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 18 04:27:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 16:45:04 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	144	0.00
2	1,4-Dioxane	0.520	0.543	-4.4	143	0.00
3	Pyridine	1.432	1.474	-2.9	136	0.00
4	n-Nitrosodimethylamine	0.676	0.701	-3.7	140	0.00
5 S	2-Fluorophenol	1.240	1.328	-7.1	141	0.00
6	Aniline	2.262	2.273	-0.5	131	0.00
7 S	Phenol-d6	1.648	1.699	-3.1	132	0.00
8	2-Chlorophenol	1.360	1.423	-4.6	136	0.00
9	Benzaldehyde	0.987	0.828	16.1	137	0.00
10 C	Phenol	1.737	1.781	-2.5	131	0.00
11	bis(2-Chloroethyl)ether	1.396	1.429	-2.4	133	0.00
12	1,3-Dichlorobenzene	1.548	1.599	-3.3	138	0.00
13 C	1,4-Dichlorobenzene	1.576	1.624	-3.0	139	0.00
14	1,2-Dichlorobenzene	1.531	1.559	-1.8	136	0.00
15	Benzyl Alcohol	1.361	1.356	0.4	130	0.00
16	2,2'-oxybis(1-Chloropropane	2.282	2.330	-2.1	135	0.00
17	2-Methylphenol	1.171	1.198	-2.3	131	0.00
18	Hexachloroethane	0.602	0.627	-4.2	140	0.00
19 P	n-Nitroso-di-n-propylamine	1.149	1.174	-2.2	128	0.00
20	3+4-Methylphenols	1.640	1.615	1.5	128	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	135	0.00
22	Acetophenone	0.534	0.553	-3.6	127	0.00
23 S	Nitrobenzene-d5	0.453	0.472	-4.2	128	0.00
24	Nitrobenzene	0.407	0.423	-3.9	127	0.00
25	Isophorone	0.797	0.813	-2.0	127	0.00
26 C	2-Nitrophenol	0.181	0.197	-8.8	131	0.00
27	2,4-Dimethylphenol	0.317	0.335	-5.7	126	0.00
28	bis(2-Chloroethoxy)methane	0.461	0.481	-4.3	130	0.00
29 C	2,4-Dichlorophenol	0.316	0.332	-5.1	128	0.00
30	1,2,4-Trichlorobenzene	0.356	0.372	-4.5	132	0.00
31	Naphthalene	1.078	1.090	-1.1	128	0.00
32	Benzoic acid	0.236	0.237	-0.4	122	0.01
33	4-Chloroaniline	0.449	0.472	-5.1	125	0.00
34 C	Hexachlorobutadiene	0.229	0.237	-3.5	132	0.00
35	Caprolactam	0.122	0.117	4.1	120	0.00
36 C	4-Chloro-3-methylphenol	0.377	0.379	-0.5	123	0.00
37	2-Methylnaphthalene	0.693	0.704	-1.6	127	0.00
38	1-Methylnaphthalene	0.739	0.749	-1.4	126	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	130	0.00
40	1,2,4,5-Tetrachlorobenzene	0.606	0.637	-5.1	126	0.00
41 P	Hexachlorocyclopentadiene	0.330	0.341	-3.3	117	0.00
42 S	2,4,6-Tribromophenol	0.249	0.256	-2.8	120	-0.01
43 C	2,4,6-Trichlorophenol	0.398	0.435	-9.3	127	0.00
44	2,4,5-Trichlorophenol	0.443	0.465	-5.0	121	0.00
45 S	2-Fluorobiphenyl	1.502	1.520	-1.2	124	0.00
46	1,1'-Biphenyl	1.468	1.514	-3.1	124	0.00
47	2-Chloronaphthalene	1.156	1.197	-3.5	126	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091725\
 Data File : BP025777.D
 Acq On : 17 Sep 2025 22:37
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 18 04:27:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 16:45:04 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.385	0.405	-5.2	120	0.00
49	Acenaphthylene	1.915	1.972	-3.0	125	0.00
50	Dimethylphthalate	1.527	1.534	-0.5	120	0.00
51	2,6-Dinitrotoluene	0.324	0.339	-4.6	123	0.00
52 C	Acenaphthene	1.104	1.124	-1.8	122	0.00
53	3-Nitroaniline	0.340	0.356	-4.7	118	0.00
54 P	2,4-Dinitrophenol	0.183	0.195	-6.6	121	0.00
55	Dibenzofuran	1.800	1.818	-1.0	123	-0.01
56 P	4-Nitrophenol	0.258	0.262	-1.6	111	0.00
57	2,4-Dinitrotoluene	0.460	0.480	-4.3	120	0.00
58	Fluorene	1.431	1.417	1.0	119	-0.01
59	2,3,4,6-Tetrachlorophenol	0.399	0.410	-2.8	120	0.00
60	Diethylphthalate	1.549	1.536	0.8	122	0.00
61	4-Chlorophenyl-phenylether	0.721	0.715	0.8	121	-0.01
62	4-Nitroaniline	0.349	0.376	-7.7	121	-0.01
63	Azobenzene	1.498	1.524	-1.7	121	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	125	0.00
65	4,6-Dinitro-2-methylphenol	0.133	0.142	-6.8	118	0.00
66 c	n-Nitrosodiphenylamine	0.612	0.632	-3.3	119	0.00
67	4-Bromophenyl-phenylether	0.220	0.232	-5.5	122	0.00
68	Hexachlorobenzene	0.245	0.252	-2.9	121	0.00
69	Atrazine	0.234	0.241	-3.0	116	0.00
70 C	Pentachlorophenol	0.162	0.169	-4.3	118	0.00
71	Phenanthrene	1.135	1.146	-1.0	118	0.00
72	Anthracene	1.151	1.175	-2.1	119	0.00
73	Carbazole	1.070	1.096	-2.4	117	0.00
74	Di-n-butylphthalate	1.313	1.363	-3.8	119	0.00
75 C	Fluoranthene	1.361	1.359	0.1	118	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	117	-0.02
77	Benzydine	0.576	0.663	-15.1	119	0.00
78	Pyrene	1.335	1.409	-5.5	116	0.00
79 S	Terphenyl-d14	1.112	1.109	0.3	111	0.00
80	Butylbenzylphthalate	0.585	0.638	-9.1	116	0.00
81	Benzo(a)anthracene	1.369	1.439	-5.1	116	-0.02
82	3,3'-Dichlorobenzidine	0.473	0.504	-6.6	114	-0.01
83	Chrysene	1.305	1.351	-3.5	115	-0.01
84	Bis(2-ethylhexyl)phthalate	0.856	0.929	-8.5	115	-0.01
85 c	Di-n-octyl phthalate	1.523	1.627	-6.8	114	-0.02
86 I	Perylene-d12	1.000	1.000	0.0	122	-0.01
87	Indeno(1,2,3-cd)pyrene	1.454	1.520	-4.5	116	-0.02
88	Benzo(b)fluoranthene	1.223	1.274	-4.2	116	-0.02
89	Benzo(k)fluoranthene	1.248	1.270	-1.8	115	-0.02
90 C	Benzo(a)pyrene	1.198	1.240	-3.5	116	-0.02
91	Dibenzo(a,h)anthracene	1.190	1.244	-4.5	117	-0.03
92	Benzo(g,h,i)perylene	1.170	1.233	-5.4	119	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091725\
Data File : BP025777.D
Acq On : 17 Sep 2025 22:37
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Sep 18 04:27:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 11 16:45:04 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0