

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091725\
 Data File : BP025777.D
 Acq On : 17 Sep 2025 22:37
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 18 04:27:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 16:45:04 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	144	0.00
2	1,4-Dioxane	40.000	41.713	-4.3	143	0.00
3	Pyridine	40.000	41.171	-2.9	136	0.00
4	n-Nitrosodimethylamine	40.000	41.486	-3.7	140	0.00
5 S	2-Fluorophenol	80.000	85.703	-7.1	141	0.00
6	Aniline	40.000	40.202	-0.5	131	0.00
7 S	Phenol-d6	80.000	82.490	-3.1	132	0.00
8	2-Chlorophenol	40.000	41.855	-4.6	136	0.00
9	Benzaldehyde	40.000	33.548	16.1	137	0.00
10 C	Phenol	40.000	41.016	-2.5	131	0.00
11	bis(2-Chloroethyl)ether	40.000	40.964	-2.4	133	0.00
12	1,3-Dichlorobenzene	40.000	41.319	-3.3	138	0.00
13 C	1,4-Dichlorobenzene	40.000	41.202	-3.0	139	0.00
14	1,2-Dichlorobenzene	40.000	40.727	-1.8	136	0.00
15	Benzyl Alcohol	40.000	39.850	0.4	130	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.832	-2.1	135	0.00
17	2-Methylphenol	40.000	40.925	-2.3	131	0.00
18	Hexachloroethane	40.000	41.623	-4.1	140	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.851	-2.1	128	0.00
20	3+4-Methylphenols	40.000	39.385	1.5	128	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	135	0.00
22	Acetophenone	40.000	41.396	-3.5	127	0.00
23 S	Nitrobenzene-d5	80.000	83.203	-4.0	128	0.00
24	Nitrobenzene	40.000	41.639	-4.1	127	0.00
25	Isophorone	40.000	40.785	-2.0	127	0.00
26 C	2-Nitrophenol	40.000	43.601	-9.0	131	0.00
27	2,4-Dimethylphenol	40.000	42.233	-5.6	126	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.799	-4.5	130	0.00
29 C	2,4-Dichlorophenol	40.000	42.051	-5.1	128	0.00
30	1,2,4-Trichlorobenzene	40.000	41.796	-4.5	132	0.00
31	Naphthalene	40.000	40.414	-1.0	128	0.00
32	Benzoic acid	40.000	40.251	-0.6	122	0.01
33	4-Chloroaniline	40.000	42.047	-5.1	125	0.00
34 C	Hexachlorobutadiene	40.000	41.372	-3.4	132	0.00
35	Caprolactam	40.000	38.574	3.6	120	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.260	-0.6	123	0.00
37	2-Methylnaphthalene	40.000	40.672	-1.7	127	0.00
38	1-Methylnaphthalene	40.000	40.560	-1.4	126	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	130	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.021	-5.1	126	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.268	-3.2	117	0.00
42 S	2,4,6-Tribromophenol	80.000	82.015	-2.5	120	-0.01
43 C	2,4,6-Trichlorophenol	40.000	43.642	-9.1	127	0.00
44	2,4,5-Trichlorophenol	40.000	41.963	-4.9	121	0.00
45 S	2-Fluorobiphenyl	80.000	80.967	-1.2	124	0.00
46	1,1'-Biphenyl	40.000	41.256	-3.1	124	0.00
47	2-Chloronaphthalene	40.000	41.425	-3.6	126	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091725\
 Data File : BP025777.D
 Acq On : 17 Sep 2025 22:37
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 18 04:27:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 16:45:04 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.047	-5.1	120	0.00
49	Acenaphthylene	40.000	41.179	-2.9	125	0.00
50	Dimethylphthalate	40.000	40.187	-0.5	120	0.00
51	2,6-Dinitrotoluene	40.000	41.914	-4.8	123	0.00
52 C	Acenaphthene	40.000	40.701	-1.8	122	0.00
53	3-Nitroaniline	40.000	41.850	-4.6	118	0.00
54 P	2,4-Dinitrophenol	40.000	39.629	0.9	121	0.00
55	Dibenzofuran	40.000	40.389	-1.0	123	-0.01
56 P	4-Nitrophenol	40.000	38.252	4.4	111	0.00
57	2,4-Dinitrotoluene	40.000	41.684	-4.2	120	0.00
58	Fluorene	40.000	39.586	1.0	119	-0.01
59	2,3,4,6-Tetrachlorophenol	40.000	41.084	-2.7	120	0.00
60	Diethylphthalate	40.000	39.668	0.8	122	0.00
61	4-Chlorophenyl-phenylether	40.000	39.641	0.9	121	-0.01
62	4-Nitroaniline	40.000	43.195	-8.0	121	-0.01
63	Azobenzene	40.000	40.697	-1.7	121	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	125	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.753	-6.9	118	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.294	-3.2	119	0.00
67	4-Bromophenyl-phenylether	40.000	42.237	-5.6	122	0.00
68	Hexachlorobenzene	40.000	41.140	-2.9	121	0.00
69	Atrazine	40.000	41.165	-2.9	116	0.00
70 C	Pentachlorophenol	40.000	41.893	-4.7	118	0.00
71	Phenanthrene	40.000	40.414	-1.0	118	0.00
72	Anthracene	40.000	40.820	-2.1	119	0.00
73	Carbazole	40.000	40.949	-2.4	117	0.00
74	Di-n-butylphthalate	40.000	41.514	-3.8	119	0.00
75 C	Fluoranthene	40.000	39.951	0.1	118	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	117	-0.02
77	Benzydine	40.000	46.060	-15.2	119	0.00
78	Pyrene	40.000	42.235	-5.6	116	0.00
79 S	Terphenyl-d14	80.000	79.787	0.3	111	0.00
80	Butylbenzylphthalate	40.000	43.589	-9.0	116	0.00
81	Benzo(a)anthracene	40.000	42.030	-5.1	116	-0.02
82	3,3'-Dichlorobenzidine	40.000	42.581	-6.5	114	-0.01
83	Chrysene	40.000	41.408	-3.5	115	-0.01
84	Bis(2-ethylhexyl)phthalate	40.000	43.388	-8.5	115	-0.01
85 c	Di-n-octyl phthalate	40.000	42.727	-6.8	114	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	122	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	41.807	-4.5	116	-0.02
88	Benzo(b)fluoranthene	40.000	41.666	-4.2	116	-0.02
89	Benzo(k)fluoranthene	40.000	40.694	-1.7	115	-0.02
90 C	Benzo(a)pyrene	40.000	41.425	-3.6	116	-0.02
91	Dibenzo(a,h)anthracene	40.000	41.808	-4.5	117	-0.03
92	Benzo(g,h,i)perylene	40.000	42.120	-5.3	119	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091725\
Data File : BP025777.D
Acq On : 17 Sep 2025 22:37
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Sep 18 04:27:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 11 16:45:04 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0