

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091924\
 Data File : BP021956.D
 Acq On : 19 Sep 2024 18:25
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP091924

Quant Time: Sep 20 01:25:53 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 20 01:22:00 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2	1,4-Dioxane	0.442	0.414	6.3	111	0.00
3	Pyridine	1.179	1.110	5.9	110	0.00
4	n-Nitrosodimethylamine	0.622	0.577	7.2	110	0.00
5 S	2-Fluorophenol	1.064	1.031	3.1	113	0.00
6	Aniline	1.214	1.190	2.0	114	0.00
7 S	Phenol-d6	1.382	1.350	2.3	115	0.00
8	2-Chlorophenol	1.138	1.113	2.2	113	0.00
9	Benzaldehyde	0.722	0.735	-1.8	116	0.00
10 C	Phenol	1.372	1.331	3.0	114	0.00
11	bis(2-Chloroethyl)ether	1.034	1.020	1.4	117	0.00
12	1,3-Dichlorobenzene	1.437	1.412	1.7	115	0.00
13 C	1,4-Dichlorobenzene	1.457	1.422	2.4	114	0.00
14	1,2-Dichlorobenzene	1.398	1.338	4.3	111	0.00
15	Benzyl Alcohol	1.060	1.067	-0.7	116	0.00
16	2,2'-oxybis(1-Chloropropane	1.102	1.103	-0.1	117	0.00
17	2-Methylphenol	0.927	0.936	-1.0	114	0.00
18	Hexachloroethane	0.536	0.523	2.4	113	0.00
19 P	n-Nitroso-di-n-propylamine	0.859	0.882	-2.7	116	0.00
20	3+4-Methylphenols	1.309	1.306	0.2	115	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
22	Acetophenone	0.488	0.480	1.6	114	0.00
23 S	Nitrobenzene-d5	0.378	0.380	-0.5	114	0.00
24	Nitrobenzene	0.384	0.385	-0.3	113	0.00
25	Isophorone	0.623	0.647	-3.9	116	0.00
26 C	2-Nitrophenol	0.168	0.173	-3.0	115	0.00
27	2,4-Dimethylphenol	0.194	0.200	-3.1	116	0.00
28	bis(2-Chloroethoxy)methane	0.374	0.376	-0.5	114	0.00
29 C	2,4-Dichlorophenol	0.330	0.335	-1.5	111	0.00
30	1,2,4-Trichlorobenzene	0.411	0.405	1.5	112	0.00
31	Naphthalene	0.998	0.989	0.9	113	0.00
32	Benzoic acid	0.227	0.236	-4.0	112	0.00
33	4-Chloroaniline	0.360	0.361	-0.3	111	0.00
34 C	Hexachlorobutadiene	0.321	0.319	0.6	114	0.01
35	Caprolactam	0.098	0.100	-2.0	112	0.00
36 C	4-Chloro-3-methylphenol	0.354	0.359	-1.4	113	0.00
37	2-Methylnaphthalene	0.730	0.732	-0.3	114	0.00
38	1-Methylnaphthalene	0.721	0.713	1.1	113	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.01
40	1,2,4,5-Tetrachlorobenzene	0.700	0.674	3.7	113	0.00
41 P	Hexachlorocyclopentadiene	0.260	0.255	1.9	116	0.00
42 S	2,4,6-Tribromophenol	0.371	0.380	-2.4	121	-0.01
43 C	2,4,6-Trichlorophenol	0.429	0.432	-0.7	115	0.00
44	2,4,5-Trichlorophenol	0.482	0.483	-0.2	116	0.00
45 S	2-Fluorobiphenyl	1.370	1.333	2.7	116	0.00
46	1,1'-Biphenyl	1.364	1.327	2.7	116	0.00
47	2-Chloronaphthalene	1.103	1.088	1.4	116	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091924\
 Data File : BP021956.D
 Acq On : 19 Sep 2024 18:25
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP091924

Quant Time: Sep 20 01:25:53 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 20 01:22:00 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.301	0.312	-3.7	114	0.00
49	Acenaphthylene	1.556	1.530	1.7	113	0.00
50	Dimethylphthalate	1.451	1.445	0.4	116	0.00
51	2,6-Dinitrotoluene	0.323	0.328	-1.5	116	0.01
52 C	Acenaphthene	1.013	0.989	2.4	115	0.00
53	3-Nitroaniline	0.288	0.292	-1.4	112	0.00
54 P	2,4-Dinitrophenol	0.189	0.198	-4.8	118	0.00
55	Dibenzofuran	1.702	1.674	1.6	116	0.00
56 P	4-Nitrophenol	0.251	0.258	-2.8	111	0.01
57	2,4-Dinitrotoluene	0.457	0.471	-3.1	116	0.00
58	Fluorene	1.421	1.393	2.0	115	0.00
59	2,3,4,6-Tetrachlorophenol	0.465	0.463	0.4	116	0.00
60	Diethylphthalate	1.474	1.488	-0.9	119	0.00
61	4-Chlorophenyl-phenylether	0.828	0.817	1.3	118	0.00
62	4-Nitroaniline	0.317	0.321	-1.3	112	0.00
63	Azobenzene	1.220	1.200	1.6	115	0.01
64 I	Phenanthrene-d10	1.000	1.000	0.0	115	-0.01
65	4,6-Dinitro-2-methylphenol	0.116	0.122	-5.2	118	-0.02
66 c	n-Nitrosodiphenylamine	0.498	0.500	-0.4	115	0.01
67	4-Bromophenyl-phenylether	0.233	0.236	-1.3	120	0.00
68	Hexachlorobenzene	0.291	0.292	-0.3	119	0.00
69	Atrazine	0.178	0.160	10.1	115	0.00
70 C	Pentachlorophenol	0.196	0.198	-1.0	115	0.00
71	Phenanthrene	0.960	0.938	2.3	115	-0.01
72	Anthracene	0.929	0.916	1.4	114	0.00
73	Carbazole	0.898	0.897	0.1	113	0.00
74	Di-n-butylphthalate	1.054	1.111	-5.4	125	0.00
75 C	Fluoranthene	1.340	1.362	-1.6	118	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
77	Benzidine	0.244	0.349	-43.0#	99	0.00
78	Pyrene	1.210	1.215	-0.4	114	0.00
79 S	Terphenyl-d14	1.065	1.087	-2.1	117	0.00
80	Butylbenzylphthalate	0.432	0.474	-9.7	125	0.00
81	Benzo(a)anthracene	1.285	1.301	-1.2	118	0.00
82	3,3'-Dichlorobenzidine	0.466	0.495	-6.2	114	0.00
83	Chrysene	1.210	1.197	1.1	115	0.00
84	Bis(2-ethylhexyl)phthalate	0.633	0.706	-11.5	132	0.00
85 c	Di-n-octyl phthalate	1.076	1.208	-12.3	132	0.00
86 I	Perylene-d12	1.000	1.000	0.0	115	0.00
87	Indeno(1,2,3-cd)pyrene	1.330	1.309	1.6	114	-0.03
88	Benzo(b)fluoranthene	1.183	1.166	1.4	112	0.00
89	Benzo(k)fluoranthene	1.119	1.114	0.4	117	0.00
90 C	Benzo(a)pyrene	1.019	1.031	-1.2	116	0.00
91	Dibenzo(a,h)anthracene	1.098	1.076	2.0	115	0.00
92	Benzo(g,h,i)perylene	1.125	1.114	1.0	114	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091924\
Data File : BP021956.D
Acq On : 19 Sep 2024 18:25
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP091924

Quant Time: Sep 20 01:25:53 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 20 01:22:00 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0