

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092220\
 Data File : BP003334.D
 Acq On : 22 Sep 2020 11:40
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 22 13:20:28 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 22 12:48:08 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	134	0.00
2	1,4-Dioxane	0.471	0.491	-4.2	149	0.00
3	Pyridine	1.253	1.302	-3.9	144	0.00
4	n-Nitrosodimethylamine	0.375	0.375	0.0	138	0.00
5 S	2-Fluorophenol	1.145	1.158	-1.1	141	0.00
6	Aniline	1.753	1.787	-1.9	142	0.00
7 S	Phenol-d6	1.527	1.571	-2.9	141	0.00
8	2-Chlorophenol	1.276	1.232	3.4	135	0.00
9	Benzaldehyde	0.851	0.824	3.2	137	0.00
10 C	Phenol	1.458	1.495	-2.5	141	0.00
11	bis(2-Chloroethyl)ether	1.156	1.190	-2.9	143	0.00
12	1,3-Dichlorobenzene	1.525	1.473	3.4	134	0.00
13 C	1,4-Dichlorobenzene	1.543	1.483	3.9	135	0.00
14	1,2-Dichlorobenzene	1.481	1.413	4.6	134	0.00
15	Benzyl Alcohol	1.015	1.036	-2.1	137	0.00
16	2,2'-oxybis(1-Chloropropane	0.840	0.941	-12.0	157#	0.00
17	2-Methylphenol	1.054	1.044	0.9	137	0.00
18	Hexachloroethane	0.575	0.558	3.0	136	0.00
19 P	n-Nitroso-di-n-propylamine	0.818	0.830	-1.5	139	0.00
20	3+4-Methylphenols	1.403	1.391	0.9	135	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	132	0.00
22	Acetophenone	0.491	0.478	2.6	133	0.00
23 S	Nitrobenzene-d5	0.340	0.344	-1.2	140	0.00
24	Nitrobenzene	0.341	0.345	-1.2	141	0.00
25	Isophorone	0.624	0.623	0.2	138	0.00
26 C	2-Nitrophenol	0.186	0.174	6.5	128	0.00
27	2,4-Dimethylphenol	0.263	0.253	3.8	133	0.00
28	bis(2-Chloroethoxy)methane	0.415	0.423	-1.9	142	0.00
29 C	2,4-Dichlorophenol	0.329	0.304	7.6	128	0.00
30	1,2,4-Trichlorobenzene	0.390	0.354	9.2	127	0.00
31	Naphthalene	1.056	1.009	4.5	134	0.00
32	Benzoic acid	0.159	0.136	14.5	113	0.00
33	4-Chloroaniline	0.449	0.431	4.0	132	0.00
34 C	Hexachlorobutadiene	0.266	0.230	13.5	121	0.00
35	Caprolactam	0.110	0.103	6.4	128	0.00
36 C	4-Chloro-3-methylphenol	0.354	0.331	6.5	129	0.00
37	2-Methylnaphthalene	0.765	0.709	7.3	131	0.00
38	1-Methylnaphthalene	0.727	0.669	8.0	130	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
40	1,2,4,5-Tetrachlorobenzene	0.655	0.614	6.3	121	0.00
41 P	Hexachlorocyclopentadiene	0.161	0.141	12.4	109	0.00
42 S	2,4,6-Tribromophenol	0.305	0.257	15.7	106	0.00
43 C	2,4,6-Trichlorophenol	0.423	0.393	7.1	119	0.00
44	2,4,5-Trichlorophenol	0.436	0.402	7.8	117	0.00
45 S	2-Fluorobiphenyl	1.367	1.308	4.3	126	0.00
46	1,1'-Biphenyl	1.492	1.462	2.0	128	0.00
47	2-Chloronaphthalene	1.235	1.191	3.6	125	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092220\
 Data File : BP003334.D
 Acq On : 22 Sep 2020 11:40
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 22 13:20:28 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 22 12:48:08 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.313	0.326	-4.2	130	0.00
49	Acenaphthylene	1.887	1.823	3.4	125	0.00
50	Dimethylphthalate	1.597	1.503	5.9	123	0.00
51	2,6-Dinitrotoluene	0.342	0.323	5.6	122	0.00
52 C	Acenaphthene	1.149	1.095	4.7	125	0.00
53	3-Nitroaniline	0.371	0.366	1.3	124	0.00
54 P	2,4-Dinitrophenol	0.127	0.111	12.6	102	0.00
55	Dibenzofuran	1.833	1.734	5.4	124	0.00
56 P	4-Nitrophenol	0.227	0.214	5.7	116	0.00
57	2,4-Dinitrotoluene	0.475	0.451	5.1	119	0.00
58	Fluorene	1.448	1.352	6.6	123	0.00
59	2,3,4,6-Tetrachlorophenol	0.409	0.355	13.2	111	0.00
60	Diethylphthalate	1.629	1.532	6.0	123	0.00
61	4-Chlorophenyl-phenylether	0.790	0.716	9.4	121	0.00
62	4-Nitroaniline	0.394	0.391	0.8	122	0.00
63	Azobenzene	1.242	1.271	-2.3	132	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
65	4,6-Dinitro-2-methylphenol	0.106	0.097	8.5	108	0.00
66 c	n-Nitrosodiphenylamine	0.587	0.569	3.1	122	0.00
67	4-Bromophenyl-phenylether	0.242	0.220	9.1	114	0.00
68	Hexachlorobenzene	0.287	0.255	11.1	112	0.00
69	Atrazine	0.230	0.211	8.3	114	0.00
70 C	Pentachlorophenol	0.104	0.086	17.3	99	0.00
71	Phenanthrene	1.088	1.034	5.0	119	0.00
72	Anthracene	1.072	1.014	5.4	117	0.00
73	Carbazole	1.019	0.959	5.9	117	0.00
74	Di-n-butylphthalate	1.302	1.225	5.9	118	0.00
75 C	Fluoranthene	1.368	1.245	9.0	113	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
77	Benzydine	0.620	0.644	-3.9	113	0.00
78	Pyrene	1.247	1.248	-0.1	114	0.00
79 S	Terphenyl-d14	0.960	0.923	3.9	112	0.00
80	Butylbenzylphthalate	0.571	0.571	0.0	114	0.00
81	Benzo(a)anthracene	1.293	1.227	5.1	108	0.00
82	3,3'-Dichlorobenzidine	0.500	0.482	3.6	109	0.00
83	Chrysene	1.242	1.186	4.5	110	0.00
84	Bis(2-ethylhexyl)phthalate	0.803	0.800	0.4	117	0.00
85 c	Di-n-octyl phthalate	1.468	1.430	2.6	111	0.00
86 I	Perylene-d12	1.000	1.000	0.0	107	0.00
87	Indeno(1,2,3-cd)pyrene	1.516	1.410	7.0	105	0.00
88	Benzo(b)fluoranthene	1.263	1.220	3.4	109	0.00
89	Benzo(k)fluoranthene	1.210	1.129	6.7	105	0.00
90 C	Benzo(a)pyrene	1.191	1.124	5.6	107	0.00
91	Dibenzo(a,h)anthracene	1.251	1.162	7.1	105	0.00
92	Benzo(g,h,i)perylene	1.249	1.147	8.2	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092220\
Data File : BP003334.D
Acq On : 22 Sep 2020 11:40
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Sep 22 13:20:28 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 22 12:48:08 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0