




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092321\
 Data File : BP007116.D
 Acq On : 23 Sep 2021 09:12
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 23 11:11:41 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 21 16:50:24 2021
 Response via : Initial Calibration



TIC: BP007116.D\data.ms

(70) Pentachlorophenol (C)




16.928min (-0.000) 1546174.70 ng

response 744374

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.50	64.15
264.00	62.90	63.61
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092321\
 Data File : BP007116.D
 Acq On : 23 Sep 2021 09:12
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 23 11:11:41 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 21 16:50:24 2021
 Response via : Initial Calibration



TIC: BP007116.D\data.ms

(77) Benzidine**19.469min (-0.000) 412798.49 ng****response 4034130**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.70	14.29
183.00	11.90	12.00
0.00	0.00	0.00