




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092321\
 Data File : BP007128.D
 Acq On : 23 Sep 2021 16:59
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 23 17:44:32 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 23 11:17:25 2021
 Response via : Initial Calibration



TIC: BP007128.D\data.ms

(70) Pentachlorophenol (C)




16.928min (+ 0.006) 634169.74 ng

response 796663

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.50	64.98
264.00	62.90	63.12
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092321\
 Data File : BP007128.D
 Acq On : 23 Sep 2021 16:59
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 23 17:44:32 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 23 11:17:25 2021
 Response via : Initial Calibration



TIC: BP007128.D\data.ms

(77) Benzidine

19.463min (-0.000) 238082.75 ng

response 4287560

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.70	14.48
183.00	11.90	12.41
0.00	0.00	0.00