

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092322\
 Data File : BP011779.D
 Acq On : 23 Sep 2022 10:49
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 24 01:32:41 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 02:45:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	93	0.00
2	1,4-Dioxane	40.000	40.177	-0.4	93	0.00
3	Pyridine	40.000	46.420	-16.1	101	0.00
4	n-Nitrosodimethylamine	40.000	40.790	-2.0	91	0.00
5 S	2-Fluorophenol	80.000	81.173	-1.5	93	0.00
6	Aniline	40.000	43.030	-7.6	96	0.00
7 S	Phenol-d6	80.000	84.285	-5.4	94	0.00
8	2-Chlorophenol	40.000	41.567	-3.9	94	0.00
9	Benzaldehyde	40.000	41.744	-4.4	97	0.00
10 C	Phenol	40.000	42.249	-5.6	95	0.00
11	bis(2-Chloroethyl)ether	40.000	41.614	-4.0	96	0.00
12	1,3-Dichlorobenzene	40.000	40.370	-0.9	93	0.00
13 C	1,4-Dichlorobenzene	40.000	40.792	-2.0	94	0.00
14	1,2-Dichlorobenzene	40.000	40.619	-1.5	93	0.00
15	Benzyl Alcohol	40.000	43.277	-8.2	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	42.245	-5.6	96	0.00
17	2-Methylphenol	40.000	42.631	-6.6	95	0.00
18	Hexachloroethane	40.000	40.344	-0.9	94	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	44.198	-10.5	97	0.00
20	3+4-Methylphenols	40.000	43.385	-8.5	96	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	96	0.00
22	Acetophenone	40.000	41.504	-3.8	97	0.00
23 S	Nitrobenzene-d5	80.000	83.132	-3.9	97	0.00
24	Nitrobenzene	40.000	41.263	-3.2	96	0.00
25	Isophorone	40.000	42.458	-6.1	98	0.00
26 C	2-Nitrophenol	40.000	37.873	5.3	92	0.00
27	2,4-Dimethylphenol	40.000	42.416	-6.0	97	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.941	-4.9	98	0.00
29 C	2,4-Dichlorophenol	40.000	42.545	-6.4	96	0.00
30	1,2,4-Trichlorobenzene	40.000	40.786	-2.0	95	0.00
31	Naphthalene	40.000	40.940	-2.3	96	0.00
32	Benzoic acid	40.000	38.642	3.4	91	0.00
33	4-Chloroaniline	40.000	43.033	-7.6	99	0.00
34 C	Hexachlorobutadiene	40.000	39.678	0.8	93	0.00
35	Caprolactam	40.000	44.288	-10.7	100	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.546	-8.9	98	0.00
37	2-Methylnaphthalene	40.000	41.951	-4.9	98	0.00
38	1-Methylnaphthalene	40.000	41.991	-5.0	97	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	98	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.869	-2.2	97	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.856	-2.1	94	0.00
42 S	2,4,6-Tribromophenol	80.000	85.503	-6.9	98	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.310	-5.8	96	0.00
44	2,4,5-Trichlorophenol	40.000	42.425	-6.1	98	0.00
45 S	2-Fluorobiphenyl	80.000	83.693	-4.6	99	0.00
46	1,1'-Biphenyl	40.000	41.105	-2.8	98	0.00
47	2-Chloronaphthalene	40.000	41.067	-2.7	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092322\
 Data File : BP011779.D
 Acq On : 23 Sep 2022 10:49
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 24 01:32:41 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 02:45:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.695	-9.2	98	0.00
49	Acenaphthylene	40.000	42.131	-5.3	99	0.00
50	Dimethylphthalate	40.000	41.937	-4.8	99	0.00
51	2,6-Dinitrotoluene	40.000	43.597	-9.0	100	0.00
52 C	Acenaphthene	40.000	41.665	-4.2	99	0.00
53	3-Nitroaniline	40.000	45.075	-12.7	102	0.00
54 P	2,4-Dinitrophenol	40.000	39.987	0.0	102	0.00
55	Dibenzofuran	40.000	42.091	-5.2	101	0.00
56 P	4-Nitrophenol	40.000	43.675	-9.2	102	0.00
57	2,4-Dinitrotoluene	40.000	41.318	-3.3	102	0.00
58	Fluorene	40.000	42.895	-7.2	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.655	-6.6	99	0.00
60	Diethylphthalate	40.000	42.767	-6.9	101	0.00
61	4-Chlorophenyl-phenylether	40.000	42.913	-7.3	100	0.00
62	4-Nitroaniline	40.000	42.468	-6.2	103	0.00
63	Azobenzene	40.000	42.551	-6.4	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.943	2.6	101	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.279	-3.2	101	0.00
67	4-Bromophenyl-phenylether	40.000	41.692	-4.2	102	0.00
68	Hexachlorobenzene	40.000	41.140	-2.9	100	0.00
69	Atrazine	40.000	42.877	-7.2	101	0.00
70 C	Pentachlorophenol	40.000	39.220	2.0	100	0.00
71	Phenanthrene	40.000	41.348	-3.4	102	0.00
72	Anthracene	40.000	42.074	-5.2	102	0.00
73	Carbazole	40.000	42.668	-6.7	104	0.00
74	Di-n-butylphthalate	40.000	43.701	-9.3	104	0.00
75 C	Fluoranthene	40.000	43.859	-9.6	105	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	107	0.00
77	Benzydine	40.000	44.241	-10.6	124	0.00
78	Pyrene	40.000	41.249	-3.1	106	0.00
79 S	Terphenyl-d14	80.000	82.652	-3.3	105	0.00
80	Butylbenzylphthalate	40.000	38.835	2.9	105	0.00
81	Benzo(a)anthracene	40.000	41.844	-4.6	107	0.00
82	3,3'-Dichlorobenzidine	40.000	43.068	-7.7	110	0.00
83	Chrysene	40.000	41.713	-4.3	108	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.989	-2.5	106	0.00
85 c	Di-n-octyl phthalate	40.000	42.700	-6.8	108	0.00
86 I	Perylene-d12	20.000	20.000	0.0	110	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.527	-6.3	109	0.01
88	Benzo(b)fluoranthene	40.000	41.862	-4.7	110	0.00
89	Benzo(k)fluoranthene	40.000	42.461	-6.2	109	0.00
90 C	Benzo(a)pyrene	40.000	42.764	-6.9	110	0.00
91	Dibenzo(a,h)anthracene	40.000	42.878	-7.2	110	0.01
92	Benzo(g,h,i)perylene	40.000	42.151	-5.4	110	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092322\
Data File : BP011779.D
Acq On : 23 Sep 2022 10:49
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Sep 24 01:32:41 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 22 02:45:14 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0