



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092921\
 Data File : BP007242.D
 Acq On : 29 Sep 2021 12:52
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Sep 29 14:45:49 2021






AutoFind: Scans 2484, 2485, 2486; Background Corrected with Scan 2475

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.6	71587	PASS
68	69	0.00	2	1.9	1443	PASS
69	198	0.00	100	27.5	76841	PASS
70	69	0.00	2	0.3	262	PASS
127	198	10	80	40.2	112213	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	279211	PASS
199	198	5	9	6.7	18772	PASS
275	198	10	60	29.2	81605	PASS
365	198	1	100	3.9	10774	PASS
441	198	0.01	100	15.8	44096	PASS
442	442	100	100	100.0	281477	PASS
443	442	15	24	19.7	55435	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092921\
Data File : BP007242.D
Acq On : 29 Sep 2021 12:52
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Sep 29 14:42:36 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 23 11:17:25 2021
Response via : Initial Calibration



TIC: BP007242.D\data.ms

(70) Pentachlorophenol (C)

16.916min (-0.006) 425260.32 ng




response 323094

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.50	64.09
264.00	62.90	63.19
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092921\
 Data File : BP007242.D
 Acq On : 29 Sep 2021 12:52
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Sep 29 14:42:36 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 23 11:17:25 2021
 Response via : Initial Calibration



TIC: BP007242.D\data.ms

(77) Benzidine

19.451min (-0.012) 2972951.78 ng

response 2192723

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.70	14.22
183.00	11.90	12.18
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
9/28/2021	BNA_P	<u>BP007242.D</u>
Compound Name	Response	Retention Time
DDT	1056967	20.65
DDD	6240	20.22
DDE	932	19.73
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
7172	1064139	0.67