



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100220\
 Data File : BP003453.D
 Acq On : 02 Oct 2020 10:02
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Oct 02 11:17:24 2020



Spectrum Information: Average of 17.404 to 17.416 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.3	21992	PASS
68	69	0.00	2	1.9	577	PASS
69	198	0.00	100	34.3	29781	PASS
70	69	0.00	2	0.5	163	PASS
127	198	10	80	46.8	40681	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	86877	PASS
199	198	5	9	7.0	6082	PASS
275	198	10	60	28.8	24997	PASS
365	198	1	100	3.9	3369	PASS
441	198	0.01	100	11.2	9692	PASS
442	198	50	100	71.4	62021	PASS
443	442	15	24	18.9	11734	PASS