

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100421\
 Data File : BP007300.D
 Acq On : 04 Oct 2021 11:29
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Oct 04 13:16:13 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 01 12:49:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.863	152	167377	20.000	ng	0.00
21) Naphthalene-d8	10.663	136	649685	20.000	ng	# 0.00
39) Acenaphthene-d10	14.498	164	421983	20.000	ng	0.00
64) Phenanthrene-d10	17.251	188	871893	20.000	ng	# 0.00
76) Chrysene-d12	21.339	240	876110	20.000	ng	# 0.00
86) Perylene-d12	23.745	264	921056	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.446	112	790539	79.062	ng	0.00
7) Phenol-d6	7.046	99	1085124	79.403	ng	0.00
23) Nitrobenzene-d5	9.028	82	917794	82.269	ng	0.00
42) 2,4,6-Tribromophenol	15.992	330	435773	79.654	ng	0.00
45) 2-Fluorobiphenyl	13.122	172	2282091	77.477	ng	0.00
79) Terphenyl-d14	19.810	244	3601876	78.782	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.334	88	162857	40.276	ng	# 88
3) Pyridine	3.746	79	441180	39.872	ng	# 86
4) n-Nitrosodimethylamine	3.652	42	169325	44.638	ng	# 88
6) Aniline	7.193	93	594885	39.491	ng	99
8) 2-Chlorophenol	7.434	128	428412	39.423	ng	98
9) Benzaldehyde	7.005	77	284292	36.930	ng	95
10) Phenol	7.069	94	516316	39.187	ng	92
11) bis(2-Chloroethyl)ether	7.293	93	404699	38.912	ng	95
12) 1,3-Dichlorobenzene	7.752	146	477440	38.950	ng	97
13) 1,4-Dichlorobenzene	7.899	146	481211	38.508	ng	98
14) 1,2-Dichlorobenzene	8.216	146	470536	38.856	ng	98
15) Benzyl Alcohol	8.111	79	360892	41.231	ng	95
16) 2,2'-oxybis(1-Chloropr...	8.393	45	459484	38.584	ng	91
17) 2-Methylphenol	8.316	107	353396	39.787	ng	94
18) Hexachloroethane	8.940	117	166354	39.650	ng	95
19) n-Nitroso-di-n-propyla...	8.675	70	295095	39.702	ng	99
20) 3+4-Methylphenols	8.646	107	486786	39.635	ng	91
22) Acetophenone	8.693	105	631056	40.312	ng	99
24) Nitrobenzene	9.069	77	436114	41.002	ng	99
25) Isophorone	9.599	82	798073	41.025	ng	99
26) 2-Nitrophenol	9.781	139	232816	41.977	ng	# 90
27) 2,4-Dimethylphenol	9.846	122	346603	39.690	ng	97
28) bis(2-Chloroethoxy)met...	10.075	93	531252	38.709	ng	99
29) 2,4-Dichlorophenol	10.322	162	410446	40.292	ng	99
30) 1,2,4-Trichlorobenzene	10.528	180	452902	39.026	ng	98
31) Naphthalene	10.716	128	1293902	39.023	ng	99
32) Benzoic acid	10.010	122	272342	40.577	ng	95
33) 4-Chloroaniline	10.834	127	548447	40.037	ng	99
34) Hexachlorobutadiene	10.993	225	273811	39.406	ng	99
35) Caprolactam	11.640	113	134444	42.900	ng	# 70
36) 4-Chloro-3-methylphenol	11.963	107	424799	41.026	ng	99
37) 2-Methylnaphthalene	12.328	142	905251	38.990	ng	96
38) 1-Methylnaphthalene	12.328	142	905251	39.905	ng	97
40) 1,2,4,5-Tetrachloroben...	12.693	216	498923	38.609	ng	99
41) Hexachlorocyclopentadiene	12.669	237	254342	35.512	ng	99
43) 2,4,6-Trichlorophenol	12.940	196	349250	39.608	ng	96


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100421\
 Data File : BP007300.D
 Acq On : 04 Oct 2021 11:29
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Oct 04 13:16:13 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 01 12:49:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.016	196	369571	38.921	ng	98
46) 1,1'-Biphenyl	13.334	154	1205114	38.394	ng	99
47) 2-Chloronaphthalene	13.375	162	938295	38.591	ng	98
48) 2-Nitroaniline	13.587	65	259972	43.855	ng	97
49) Acenaphthylene	14.222	152	1477753	39.461	ng	100
50) Dimethylphthalate	13.963	163	1189969	39.208	ng	99
51) 2,6-Dinitrotoluene	14.081	165	250043	40.332	ng	98
52) Acenaphthene	14.563	154	878219	38.707	ng	100
53) 3-Nitroaniline	14.416	138	277628	41.417	ng	96
54) 2,4-Dinitrophenol	14.628	184	140650	39.384	ng	93
55) Dibenzofuran	14.898	168	1459527	38.556	ng	96
56) 4-Nitrophenol	14.745	139	208872	38.382	ng	# 86
57) 2,4-Dinitrotoluene	14.875	165	343068	41.934	ng	95
58) Fluorene	15.551	166	1154725	39.485	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.051	232	307119	36.834	ng	97
60) Diethylphthalate	15.322	149	1179708	40.114	ng	98
61) 4-Chlorophenyl-phenyle...	15.540	204	601150	38.900	ng	93
62) 4-Nitroaniline	15.581	138	290883	43.148	ng	# 84
63) Azobenzene	15.834	77	1049231	41.171	ng	96
65) 4,6-Dinitro-2-methylph...	15.640	198	216391	41.357	ng	95
66) n-Nitrosodiphenylamine	15.757	169	1015624	39.149	ng	98
67) 4-Bromophenyl-phenylether	16.439	248	370245	38.745	ng	96
68) Hexachlorobenzene	16.557	284	408741	38.703	ng	99
69) Atrazine	16.716	200	359214	39.282	ng	96
70) Pentachlorophenol	16.910	266	221055	33.704	ng	99
71) Phenanthrene	17.292	178	1819207	38.913	ng	99
72) Anthracene	17.386	178	1809203	39.560	ng	99
73) Carbazole	17.663	167	1807618	40.335	ng	99
74) Di-n-butylphthalate	18.210	149	2055644	41.391	ng	99
75) Fluoranthene	19.263	202	2170390	40.295	ng	97
77) Benzidine	19.445	184	913370	33.924	ng	100
78) Pyrene	19.616	202	2226875	38.741	ng	98
80) Butylbenzylphthalate	20.486	149	956587	41.556	ng	99
81) Benzo(a)anthracene	21.327	228	2233947	39.283	ng	99
82) 3,3'-Dichlorobenzidine	21.257	252	785194	40.204	ng	99
83) Chrysene	21.380	228	2108641	38.797	ng	98
84) Bis(2-ethylhexyl)phtha...	21.245	149	1371958	41.455	ng	# 99
85) Di-n-octyl phthalate	22.169	149	2378687	42.217	ng	99
87) Indeno(1,2,3-cd)pyrene	26.274	276	2620818	39.594	ng	# 95
88) Benzo(b)fluoranthene	23.016	252	2094324	38.048	ng	99
89) Benzo(k)fluoranthene	23.063	252	2158110	38.724	ng	# 98
90) Benzo(a)pyrene	23.639	252	1841067	39.369	ng	# 97
91) Dibenzo(a,h)anthracene	26.286	278	2177116	39.248	ng	# 96
92) Benzo(g,h,i)perylene	27.045	276	2114705	39.065	ng	# 95





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.863 min Scan# 812
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 167377
Ion Ratio Lower Upper
152 100
150 153.2 125.6 188.4
115 57.1 48.5 72.7



#2
1,4-Dioxane
Concen: 40.276 ng
RT: 3.334 min Scan# 42
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion: 88 Resp: 162857
Ion Ratio Lower Upper
88 100
58 70.4 46.7 70.1#
43 24.8 20.3 30.5









#3
Pyridine
Concen: 39.872 ng
RT: 3.746 min Scan# 112
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 441180
Ion Ratio Lower Upper
79 100
52 65.3 41.7 62.5#
51 33.3 24.1 36.1



#4
n-Nitrosodimethylamine
Concen: 44.638 ng
RT: 3.652 min Scan# 96
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion: 42 Resp: 169325
Ion Ratio Lower Upper
42 100
74 145.7 105.4 158.0
44 7.4 4.2 6.4#









#5
2-Fluorophenol
Concen: 79.062 ng
RT: 5.446 min Scan# 401
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:112 Resp: 790539
Ion Ratio Lower Upper
112 100
64 52.6 41.4 62.2
63 27.0 23.8 35.8



#6
Aniline
Concen: 39.491 ng
RT: 7.193 min Scan# 698
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion: 93 Resp: 594885
Ion Ratio Lower Upper
93 100
66 36.5 29.7 44.5
65 18.6 15.3 22.9









#7
Phenol-d6
Concen: 79.403 ng
RT: 7.046 min Scan# 673
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 99 Resp: 1085124




Ion	Ratio	Lower	Upper
99	100		
42	14.8	13.4	20.0
71	30.7	30.6	46.0




#8
2-Chlorophenol
Concen: 39.423 ng
RT: 7.434 min Scan# 739
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion: 128 Resp: 428412

Ion	Ratio	Lower	Upper
128	100		
130	32.9	12.4	52.4
64	41.6	20.3	60.3









#9
Benzaldehyde
Concen: 36.930 ng
RT: 7.005 min Scan# 666
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 284292
Ion Ratio Lower Upper
77 100
105 103.5 76.7 116.7
106 99.6 76.2 116.2



#10
Phenol
Concen: 39.187 ng
RT: 7.069 min Scan# 677
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion: 94 Resp: 516316
Ion Ratio Lower Upper
94 100
65 26.3 8.8 48.8
66 34.6 20.7 60.7









#11
bis(2-Chloroethyl)ether
Concen: 38.912 ng
RT: 7.293 min Scan# 715
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 93 Resp: 404699
Ion Ratio Lower Upper
93 100
63 71.5 46.2 86.2
95 32.4 12.2 52.2



#12
1,3-Dichlorobenzene
Concen: 38.950 ng
RT: 7.752 min Scan# 793
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:146 Resp: 477440
Ion Ratio Lower Upper
146 100
148 63.9 51.5 77.3
75 28.2 25.7 38.5









#13
1,4-Dichlorobenzene
Concen: 38.508 ng
RT: 7.899 min Scan# 818
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:146 Resp: 481211
Ion Ratio Lower Upper
146 100
148 64.1 51.3 76.9
111 38.9 33.0 49.6



#14
1,2-Dichlorobenzene
Concen: 38.856 ng
RT: 8.216 min Scan# 872
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:146 Resp: 470536
Ion Ratio Lower Upper
146 100
148 64.2 50.3 75.5
111 40.7 33.7 50.5









#15
Benzyl Alcohol
Concen: 41.231 ng
RT: 8.111 min Scan# 854
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 79	Resp: 360892
Ion Ratio	Lower Upper
79 100	
108 83.3	62.5 93.7
77 63.9	52.9 79.3



#16
2,2'-oxybis(1-Chloropropane)
Concen: 38.584 ng
RT: 8.393 min Scan# 902
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion: 45	Resp: 459484
Ion Ratio	Lower Upper
45 100	
77 17.6	2.0 42.0
79 12.8	0.0 36.7









#17
2-Methylphenol
Concen: 39.787 ng
RT: 8.316 min Scan# 889
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	107	Resp:	353396
Ion	Ratio	Lower	Upper
107	100		
108	112.6	87.7	131.5
77	44.0	40.2	60.2
79	43.1	39.0	58.4



#18
Hexachloroethane
Concen: 39.650 ng
RT: 8.940 min Scan# 995
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:	117	Resp:	166354
Ion	Ratio	Lower	Upper
117	100		
119	97.1	76.2	114.2
201	112.2	83.4	125.0









#19
n-Nitroso-di-n-propylamine
Concen: 39.702 ng
RT: 8.675 min Scan# 950
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	70	Resp:	295095
Ion	Ratio	Lower	Upper
70	100		
42	51.6	40.8	61.2
101	10.8	8.8	13.2
130	23.3	18.9	28.3



#20
3+4-Methylphenols
Concen: 39.635 ng
RT: 8.646 min Scan# 945
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:	107	Resp:	486786
Ion	Ratio	Lower	Upper
107	100		
108	85.5	71.0	111.0
77	29.9	16.6	56.6
79	25.4	12.7	52.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.663 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	136	Resp:	649685
Ion Ratio	Lower	Upper	
136	100		
137	10.8	8.6	12.8
54	7.2	4.6	6.8#
68	5.6	3.6	5.4#



#22
Acetophenone
Concen: 40.312 ng
RT: 8.693 min Scan# 953
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:	105	Resp:	631056
Ion Ratio	Lower	Upper	
105	100		
71	2.0	1.9	2.9
51	23.9	18.4	27.6
120	22.7	17.9	26.9









#23
Nitrobenzene-d5
Concen: 82.269 ng
RT: 9.028 min Scan# 1010
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 917794
Ion Ratio Lower Upper
82 100
128 46.8 35.7 53.5
54 46.9 35.0 52.4



#24
Nitrobenzene
Concen: 41.002 ng
RT: 9.069 min Scan# 1017
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion: 77 Resp: 436114
Ion Ratio Lower Upper
77 100
123 48.3 38.2 57.2
65 13.3 11.4 17.2









#25
Isophorone
Concen: 41.025 ng
RT: 9.599 min Scan# 1107
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 798073
Ion Ratio Lower Upper
82 100
95 7.0 6.5 9.7
138 19.0 15.0 22.4



#26
2-Nitrophenol
Concen: 41.977 ng
RT: 9.781 min Scan# 1138
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion: 139 Resp: 232816
Ion Ratio Lower Upper
139 100
109 28.8 28.8 43.2#
65 45.8 41.0 61.6









#27
2,4-Dimethylphenol
Concen: 39.690 ng
RT: 9.846 min Scan# 1149
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:122 Resp: 346603
Ion Ratio Lower Upper
122 100
107 114.0 94.5 141.7
121 58.5 47.4 71.2



#28
bis(2-Chloroethoxy)methane
Concen: 38.709 ng
RT: 10.075 min Scan# 1188
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion: 93 Resp: 531252
Ion Ratio Lower Upper
93 100
95 32.2 25.9 38.9
123 11.6 10.2 15.2









#29
2,4-Dichlorophenol
Concen: 40.292 ng
RT: 10.322 min Scan# 1230
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 410446
Ion Ratio Lower Upper
162 100
164 64.1 44.9 84.9
98 33.0 14.3 54.3



#30
1,2,4-Trichlorobenzene
Concen: 39.026 ng
RT: 10.528 min Scan# 1265
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:180 Resp: 452902
Ion Ratio Lower Upper
180 100
182 95.6 77.9 116.9
145 28.8 24.7 37.1









#31
Naphthalene
Concen: 39.023 ng
RT: 10.716 min Scan# 1297
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:128 Resp: 1293902
Ion Ratio Lower Upper
128 100
129 11.0 9.0 13.6
127 13.1 10.7 16.1



#32
Benzoic acid
Concen: 40.577 ng
RT: 10.010 min Scan# 1177
Delta R.T. -0.011 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:122 Resp: 272342
Ion Ratio Lower Upper
122 100
105 120.8 103.2 143.2
77 86.7 76.2 116.2









#33
4-Chloroaniline
Concen: 40.037 ng
RT: 10.834 min Scan# 1317
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	127	Resp:	548447
Ion	Ratio	Lower	Upper
127	100		
129	32.4	26.1	39.1
65	27.3	22.5	33.7
92	17.7	15.0	22.6



#34
Hexachlorobutadiene
Concen: 39.406 ng
RT: 10.993 min Scan# 1344
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:	225	Resp:	273811
Ion	Ratio	Lower	Upper
225	100		
223	63.0	51.0	76.6
227	63.8	51.4	77.2









#35
Caprolactam
Concen: 42.900 ng
RT: 11.640 min Scan# 1454
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:113 Resp: 134444
Ion Ratio Lower Upper
113 100
55 157.0 98.9 138.9#
56 115.6 71.7 111.7#



#36
4-Chloro-3-methylphenol
Concen: 41.026 ng
RT: 11.963 min Scan# 1509
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:107 Resp: 424799
Ion Ratio Lower Upper
107 100
144 26.9 21.1 31.7
142 83.8 66.0 99.0









#37
2-Methylnaphthalene
Concen: 38.990 ng
RT: 12.328 min Scan# 1571
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:142 Resp: 905251
Ion Ratio Lower Upper
142 100
141 89.0 71.8 107.6
115 33.4 31.6 47.4



#38
1-Methylnaphthalene
Concen: 39.905 ng
RT: 12.328 min Scan# 1571
Delta R.T. -0.217 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:142 Resp: 905251
Ion Ratio Lower Upper
142 100
141 89.0 73.3 109.9
116 3.5 3.5 5.3









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.498 min Scan# 1940
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:164 Resp: 421983
Ion Ratio Lower Upper
164 100
162 99.1 80.2 120.2
160 44.6 34.5 51.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 38.609 ng
RT: 12.693 min Scan# 1633
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:216 Resp: 498923
Ion Ratio Lower Upper
216 100
214 78.1 61.9 92.9
179 20.2 16.8 25.2
108 17.2 14.9 22.3









#41
Hexachlorocyclopentadiene
Concen: 35.512 ng
RT: 12.669 min Scan# 1629
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	237	Resp:	254342
Ion Ratio	Lower	Upper	
237	100		
235	63.4	44.5	84.5
272	12.9	0.0	32.5



#42
2,4,6-Tribromophenol
Concen: 79.654 ng
RT: 15.992 min Scan# 2194
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:	330	Resp:	435773
Ion Ratio	Lower	Upper	
330	100		
332	96.8	77.8	116.8
141	33.0	25.4	38.2









#43
2,4,6-Trichlorophenol
Concen: 39.608 ng
RT: 12.940 min Scan# 1675
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:196 Resp: 349250
Ion Ratio Lower Upper
196 100
198 94.7 79.5 119.3
200 30.9 25.0 37.6



#44
2,4,5-Trichlorophenol
Concen: 38.921 ng
RT: 13.016 min Scan# 1688
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:196 Resp: 369571
Ion Ratio Lower Upper
196 100
198 95.5 74.7 112.1
97 44.4 37.4 56.0
132 25.0 20.0 30.0









#45
2-Fluorobiphenyl
Concen: 77.477 ng
RT: 13.122 min Scan# 1706
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:172 Resp: 2282091
Ion Ratio Lower Upper
172 100
171 35.2 27.9 41.9
170 23.5 18.6 27.8



#46
1,1'-Biphenyl
Concen: 38.394 ng
RT: 13.334 min Scan# 1742
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:154 Resp: 1205114
Ion Ratio Lower Upper
154 100
153 40.1 20.4 60.4
76 11.6 0.0 30.8









#47
2-Chloronaphthalene
Concen: 38.591 ng
RT: 13.375 min Scan# 1749
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 938295
Ion Ratio Lower Upper
162 100
127 37.2 31.4 47.0
164 32.7 26.6 39.8



#48
2-Nitroaniline
Concen: 43.855 ng
RT: 13.587 min Scan# 1785
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion: 65 Resp: 259972
Ion Ratio Lower Upper
65 100
92 73.5 57.7 86.5
138 120.0 92.1 138.1









#49
Acenaphthylene
Concen: 39.461 ng
RT: 14.222 min Scan# 1893
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 1477753
Ion Ratio Lower Upper
152 100
151 20.0 16.2 24.2
153 13.1 10.6 16.0



#50
Dimethylphthalate
Concen: 39.208 ng
RT: 13.963 min Scan# 1849
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:163 Resp: 1189969
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.2
164 10.0 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 40.332 ng
RT: 14.081 min Scan# 1869
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 250043
Ion Ratio Lower Upper
165 100
63 50.4 41.3 61.9
89 53.9 44.8 67.2



#52
Acenaphthene
Concen: 38.707 ng
RT: 14.563 min Scan# 1951
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:154 Resp: 878219
Ion Ratio Lower Upper
154 100
153 111.5 89.4 134.2
152 53.8 43.0 64.4









#53
3-Nitroaniline
Concen: 41.417 ng
RT: 14.416 min Scan# 1926
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:138 Resp: 277628
Ion Ratio Lower Upper
138 100
108 9.1 8.4 12.6
92 105.6 87.8 131.8



#54
2,4-Dinitrophenol
Concen: 39.384 ng
RT: 14.628 min Scan# 1962
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:184 Resp: 140650
Ion Ratio Lower Upper
184 100
63 63.3 47.0 70.6
154 59.1 52.3 78.5









#55
Dibenzenofuran
Concen: 38.556 ng
RT: 14.898 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:168 Resp: 1459527
Ion Ratio Lower Upper
168 100
139 35.9 31.0 46.6
169 13.6 10.8 16.2



#56
4-Nitrophenol
Concen: 38.382 ng
RT: 14.745 min Scan# 1982
Delta R.T. 0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:139 Resp: 208872
Ion Ratio Lower Upper
139 100
109 71.2 72.3 112.3#
65 87.0 72.5 112.5









#57
2,4-Dinitrotoluene
Concen: 41.934 ng
RT: 14.875 min Scan# 2004
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	165	Resp:	343068
Ion Ratio	Lower	Upper	
165	100		
63	39.0	32.3	48.5
89	64.3	55.9	83.9
182	2.9	2.2	3.4



#58
Fluorene
Concen: 39.485 ng
RT: 15.551 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:	166	Resp:	1154725
Ion Ratio	Lower	Upper	
166	100		
165	97.1	78.0	117.0
167	13.3	10.5	15.7









#59
2,3,4,6-Tetrachlorophenol
Concen: 36.834 ng
RT: 15.051 min Scan# 2034
Delta R.T. -0.082 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:232 Resp: 307119




Ion	Ratio	Lower	Upper
232	100		
131	37.2	32.2	48.4
130	2.3	2.0	3.0
166	23.8	19.4	29.2




#60
Diethylphthalate
Concen: 40.114 ng
RT: 15.322 min Scan# 2080
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:149 Resp: 1179708

Ion	Ratio	Lower	Upper
149	100		
177	23.1	17.5	26.3
150	12.5	9.9	14.9









#61
4-Chlorophenyl-phenylether
Concen: 38.900 ng
RT: 15.540 min Scan# 2117
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:204 Resp: 601150
Ion Ratio Lower Upper
204 100
206 32.9 26.6 40.0
141 53.0 48.6 73.0



#62
4-Nitroaniline
Concen: 43.148 ng
RT: 15.581 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:138 Resp: 290883
Ion Ratio Lower Upper
138 100
92 49.3 31.2 71.2
108 75.9 78.6 118.6#









#63
Azobenzene
Concen: 41.171 ng
RT: 15.834 min Scan# 2167
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 1049231




Ion	Ratio	Lower	Upper
77	100		
182	28.2	6.0	46.0
105	21.9	0.0	38.8
51	28.4	8.8	48.8




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.251 min Scan# 2408
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion: 188 Resp: 871893

Ion	Ratio	Lower	Upper
188	100		
94	8.5	5.6	8.4#
80	9.2	6.3	9.5









#65
4,6-Dinitro-2-methylphenol
Concen: 41.357 ng
RT: 15.640 min Scan# 2134
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:198 Resp: 216391
Ion Ratio Lower Upper
198 100
51 39.1 14.0 54.0
105 40.2 19.5 59.5



#66
n-Nitrosodiphenylamine
Concen: 39.149 ng
RT: 15.757 min Scan# 2154
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:169 Resp: 1015624
Ion Ratio Lower Upper
169 100
168 66.5 54.4 81.6
167 36.8 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 38.745 ng
RT: 16.439 min Scan# 2270
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:248 Resp: 370245
Ion Ratio Lower Upper
248 100
250 98.7 79.8 119.8
141 66.1 57.8 86.6



#68
Hexachlorobenzene
Concen: 38.703 ng
RT: 16.557 min Scan# 2290
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:284 Resp: 408741
Ion Ratio Lower Upper
284 100
142 32.5 26.9 40.3
249 30.8 24.5 36.7









#69
Atrazine
Concen: 39.282 ng
RT: 16.716 min Scan# 2317
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:200 Resp: 359214
Ion Ratio Lower Upper
200 100
173 26.7 4.1 44.1
215 49.6 27.7 67.7



#70
Pentachlorophenol
Concen: 33.704 ng
RT: 16.910 min Scan# 2350
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:266 Resp: 221055
Ion Ratio Lower Upper
266 100
268 63.7 51.6 77.4
264 63.0 50.3 75.5









#71
Phenanthrene
Concen: 38.913 ng
RT: 17.292 min Scan# 2415
Delta R.T. -0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:178 Resp: 1819207
Ion Ratio Lower Upper
178 100
176 19.4 15.1 22.7
179 15.7 12.4 18.6



#72
Anthracene
Concen: 39.560 ng
RT: 17.386 min Scan# 2431
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:178 Resp: 1809203
Ion Ratio Lower Upper
178 100
176 18.8 14.9 22.3
179 15.6 12.2 18.2









#73
 Carbazole
 Concen: 40.335 ng
 RT: 17.663 min Scan# 2478
 Delta R.T. 0.000 min
 Lab File: BP007300.D
 Acq: 04 Oct 2021 11:29




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:167 Resp: 1807618
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.1 25.7
 139 12.6 10.7 16.1



#74
 Di-n-butylphthalate
 Concen: 41.391 ng
 RT: 18.210 min Scan# 2571
 Delta R.T. 0.000 min
 Lab File: BP007300.D
 Acq: 04 Oct 2021 11:29

Tgt Ion:149 Resp: 2055644
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.3 10.9
 104 4.8 4.5 6.7









#75
Fluoranthene
Concen: 40.295 ng
RT: 19.263 min Scan# 2750
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:202 Resp: 2170390
Ion Ratio Lower Upper
202 100
101 10.1 0.0 27.6
203 17.4 0.0 37.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.339 min Scan# 3103
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:240 Resp: 876110
Ion Ratio Lower Upper
240 100
120 10.5 6.5 9.7#
236 25.4 20.6 31.0









#77
Benzidine
Concen: 33.924 ng
RT: 19.445 min Scan# 2781
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:184 Resp: 913370
Ion Ratio Lower Upper
184 100
185 14.0 11.0 16.4
183 11.8 9.5 14.3



#78
Pyrene
Concen: 38.741 ng
RT: 19.616 min Scan# 2810
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:202 Resp: 2226875
Ion Ratio Lower Upper
202 100
200 21.6 16.3 24.5
203 18.0 14.2 21.2









#79
Terphenyl-d14
Concen: 78.782 ng
RT: 19.810 min Scan# 2843
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:244 Resp: 3601876
Ion Ratio Lower Upper
244 100
212 7.2 6.0 9.0
122 10.0 7.1 10.7



#80
Butylbenzylphthalate
Concen: 41.556 ng
RT: 20.486 min Scan# 2958
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:149 Resp: 956587
Ion Ratio Lower Upper
149 100
91 67.4 54.6 81.8
206 22.1 16.7 25.1









#81
Benzo(a)anthracene
Concen: 39.283 ng
RT: 21.327 min Scan# 3101
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 2233947
Ion Ratio Lower Upper
228 100
226 27.6 21.6 32.4
229 19.9 15.8 23.6



#82
3,3'-Dichlorobenzidine
Concen: 40.204 ng
RT: 21.257 min Scan# 3089
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:252 Resp: 785194
Ion Ratio Lower Upper
252 100
254 63.8 50.6 76.0
126 10.4 7.0 10.6









#83
Chrysene
Concen: 38.797 ng
RT: 21.380 min Scan# 3110
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 2108641
Ion Ratio Lower Upper
228 100
226 30.6 23.8 35.6
229 20.0 15.4 23.2



#84
Bis(2-ethylhexyl)phthalate
Concen: 41.455 ng
RT: 21.245 min Scan# 3087
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:149 Resp: 1371958
Ion Ratio Lower Upper
149 100
167 27.9 21.9 32.9
279 4.3 4.3 6.5#









#85
Di-n-octyl phthalate
Concen: 42.217 ng
RT: 22.169 min Scan# 3244
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 2378687
Ion Ratio Lower Upper
149 100
167 1.6 1.4 2.0
43 7.3 5.6 8.4



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.745 min Scan# 3512
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:264 Resp: 921056
Ion Ratio Lower Upper
264 100
260 24.0 18.6 28.0
265 22.4 17.6 26.4









#87
Indeno(1,2,3-cd)pyrene
Concen: 39.594 ng
RT: 26.274 min Scan# 3942
Delta R.T. 0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:276 Resp: 2620818
Ion Ratio Lower Upper
276 100
138 23.7 15.3 22.9#
277 25.4 20.4 30.6



#88
Benzo(b)fluoranthene
Concen: 38.048 ng
RT: 23.016 min Scan# 3388
Delta R.T. 0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:252 Resp: 2094324
Ion Ratio Lower Upper
252 100
253 22.0 17.5 26.3
125 9.1 6.2 9.4









#89
Benzo(k)fluoranthene
Concen: 38.724 ng
RT: 23.063 min Scan# 3396
Delta R.T. 0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:252 Resp: 2158110
Ion Ratio Lower Upper
252 100
253 21.9 17.6 26.4
125 9.7 6.0 9.0#



#90
Benzo(a)pyrene
Concen: 39.369 ng
RT: 23.639 min Scan# 3494
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:252 Resp: 1841067
Ion Ratio Lower Upper
252 100
253 22.2 17.9 26.9
125 11.1 6.5 9.7#









#91
Dibenzo(a,h)anthracene
Concen: 39.248 ng
RT: 26.286 min Scan# 3944
Delta R.T. 0.006 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 2177116
Ion Ratio Lower Upper
278 100
139 15.8 9.3 13.9#
279 23.7 18.8 28.2



#92
Benzo(g,h,i)perylene
Concen: 39.065 ng
RT: 27.045 min Scan# 4073
Delta R.T. 0.000 min
Lab File: BP007300.D
Acq: 04 Oct 2021 11:29

Tgt Ion:276 Resp: 2114705
Ion Ratio Lower Upper
276 100
277 24.1 18.9 28.3
138 20.6 12.6 18.8#

