

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100520\
 Data File : BP003505.D
 Acq On : 06 Oct 2020 01:17
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Oct 06 04:21:33 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 05 14:33:14 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.499	152	54	20.00	ng	# 0.00
21) Naphthalene-d8	10.287	136	29	20.00	ng	# 0.01
39) Acenaphthene-d10	14.169	164	49	20.00	ng	# 0.01
64) Phenanthrene-d10	16.928	188	464	20.00	ng	# 0.01
76) Chrysene-d12	21.033	240	984	20.00	ng	# 0.00
86) Perylene-d12	23.210	264	792	20.00	ng	# 0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.117	112	38	12.29	ng	0.00
7) Phenol-d6	6.699	99	30	7.28	ng	0.00
23) Nitrobenzene-d5	8.563	82	54	109.39	ng	-0.09
42) 2,4,6-Tribromophenol	15.722	330	103	137.82	ng	0.06
45) 2-Fluorobiphenyl	12.804	172	528	157.60	ng	0.03
79) Terphenyl-d14	19.498	244	4376	92.66	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.970	88	33	25.97	ng	# 1
3) Pyridine	3.393	79	119	35.17	ng	# 38
4) n-Nitrosodimethylamine	3.334	42	121	119.54	ng	# 1
6) Aniline	6.846	93	177	37.39	ng	# 73
8) 2-Chlorophenol	6.905	128	34	9.87	ng	92
9) Benzaldehyde	6.687	77	47	20.45	ng	# 23
10) Phenol	6.734	94	13	3.30	ng	# 1
11) bis(2-Chloroethyl)ether	6.846	93	177	56.73	ng	75
12) 1,3-Dichlorobenzene	7.393	146	20	4.86	ng	# 60
13) 1,4-Dichlorobenzene	7.522	146	29	6.96	ng	# 71
14) 1,2-Dichlorobenzene	7.858	146	32	8.00	ng	# 87
15) Benzyl Alcohol	7.746	79	137	50.01	ng	# 45
16) 2,2'-oxybis(1-Chloropr...	8.040	45	53	23.36	ng	# 1
17) 2-Methylphenol	7.969	107	19	6.67	ng	# 1
18) Hexachloroethane	8.587	117	69	44.45	ng	# 14
19) n-Nitroso-di-n-propyla...	8.322	70	34	15.40	ng	# 1
20) 3+4-Methylphenols	8.293	107	44	11.61	ng	# 9
22) Acetophenone	8.322	105	23	32.29	ng	# 18
24) Nitrobenzene	8.722	77	68	137.67	ng	# 64
25) Isophorone	9.240	82	23	25.42	ng	# 9
26) 2-Nitrophenol	9.422	139	15	55.60	ng	# 1
27) 2,4-Dimethylphenol	9.699	122	25	65.53	ng	93
28) bis(2-Chloroethoxy)met...	9.728	93	82	136.18	ng	# 85
29) 2,4-Dichlorophenol	9.969	162	10	20.98	ng	91
30) 1,2,4-Trichlorobenzene	10.193	180	13	22.98	ng	99
31) Naphthalene	10.322	128	68	44.39	ng	# 48
32) Benzoic acid	9.699	122	25	83.26	ng	# 1
33) 4-Chloroaniline	10.475	127	28	43.02	ng	# 1
34) Hexachlorobutadiene	10.675	225	8	20.72	ng	88
35) Caprolactam	11.252	113	19	119.43	ng	# 1
36) 4-Chloro-3-methylphenol	11.628	107	27	52.53	ng	# 22
37) 2-Methylnaphthalene	11.969	142	24	21.63	ng	# 1
38) 1-Methylnaphthalene	12.199	142	28	26.57	ng	# 55
40) 1,2,4,5-Tetrachloroben...	12.340	216	9	5.61	ng	# 1
41) Hexachlorocyclopentadiene	12.257	237	19	44.14	ng	74
43) 2,4,6-Trichlorophenol	12.610	196	26	25.07	ng	# 53

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100520\
 Data File : BP003505.D
 Acq On : 06 Oct 2020 01:17
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Oct 06 04:21:33 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 05 14:33:14 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.710	196	26	24.37	ng	# 11
46) 1,1'-Biphenyl	12.987	154	16	4.38	ng	# 1
47) 2-Chloronaphthalene	13.022	162	70	23.13	ng	# 92
48) 2-Nitroaniline	13.251	65	38	49.62	ng	# 44
49) Acenaphthylene	13.898	152	431	93.21	ng	# 95
50) Dimethylphthalate	13.646	163	100	25.56	ng	# 48
51) 2,6-Dinitrotoluene	13.904	165	34	40.54	ng	# 93
53) 3-Nitroaniline	14.075	138	20	22.01	ng	# 1
54) 2,4-Dinitrophenol	14.345	184	14	39.33	ng	# 17
55) Dibenzofuran	14.581	168	42	9.35	ng	# 1
56) 4-Nitrophenol	14.363	139	31	55.81	ng	# 90
57) 2,4-Dinitrotoluene	14.569	165	18	15.47	ng	# 1
58) Fluorene	15.234	166	103	29.04	ng	# 60
59) 2,3,4,6-Tetrachlorophenol	14.804	232	131	130.57	ng	# 54
60) Diethylphthalate	15.016	149	1058	265.09	ng	# 92
61) 4-Chlorophenyl-phenyle...	15.222	204	26	13.43	ng	# 22
62) 4-Nitroaniline	15.387	138	31	32.14	ng	# 84
63) Azobenzene	15.534	77	155	50.95	ng	# 61
65) 4,6-Dinitro-2-methylph...	15.334	198	19	7.76	ng	# 1
66) n-Nitrosodiphenylamine	15.457	169	253	18.56	ng	# 70
67) 4-Bromophenyl-phenylether	16.134	248	72	12.84	ng	# 79
68) Hexachlorobenzene	16.245	284	159	23.86	ng	# 59
69) Atrazine	16.587	200	37079	6935.50	ng	# 36
70) Pentachlorophenol	16.592	266	187760	60099.78	ng	# 99
71) Phenanthrene	16.963	178	922	36.52	ng	# 64
72) Anthracene	17.069	178	1013	40.72	ng	# 72
73) Carbazole	17.386	167	7814	330.65	ng	# 80
74) Di-n-butylphthalate	17.898	149	2938	97.25	ng	# 91
75) Fluoranthene	18.963	202	1742	54.88	ng	# 88
77) Benzidine	19.133	184	501594	16438.87	ng	# 100
78) Pyrene	19.310	202	1923	31.34	ng	# 90
80) Butylbenzylphthalate	20.186	149	855	30.45	ng	# 53
81) Benzo(a)anthracene	21.022	228	2404	37.79	ng	# 79
82) 3,3'-Dichlorobenzidine	20.957	252	5188	211.03	ng	# 97
83) Chrysene	21.069	228	2247	36.77	ng	# 84
84) Bis(2-ethylhexyl)phtha...	20.945	149	2512	63.55	ng	# 88
85) Di-n-octyl phthalate	21.792	149	2456	34.00	ng	# 89
87) Indeno(1,2,3-cd)pyrene	25.415	276	3312	55.17	ng	# 60
88) Benzo(b)fluoranthene	22.569	252	2554	51.07	ng	# 57
89) Benzo(k)fluoranthene	22.598	252	2072	43.23	ng	# 51
90) Benzo(a)pyrene	23.127	252	2815	59.69	ng	# 31
91) Dibenzo(a,h)anthracene	25.415	278	2859	57.72	ng	# 60
92) Benzo(g,h,i)perylene	26.098	276	2442	49.38	ng	# 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Oct 06 04:21:33 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 05 14:33:14 2020
Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100520\
Data File : BP003505.D
Acq On : 06 Oct 2020 01:17
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Oct 06 04:21:33 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 05 14:33:14 2020
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 7.499 min Scan# 784
 Delta R.T. 0.001 min
 Lab File: BP003505.D
 Acq: 06 Oct 2020 01:17




Instrument :
 BNA_P
 ClientSampleId :
 DFTPP


Tgt Ion:152 Resp: 54
 Ion Ratio Lower Upper
 152 100
 150 30.3 123.8 185.6#
 115 615.2 43.8 65.6#



#2
 1,4-Dioxane
 Concen: 25.97 ng
 RT: 2.970 min Scan# 14
 Delta R.T. -0.035 min
 Lab File: BP003505.D
 Acq: 06 Oct 2020 01:17

Tgt Ion: 88 Resp: 33
 Ion Ratio Lower Upper
 88 100
 58 0.0 44.2 66.4#
 43 136.4 15.2 22.8#









#3
Pyridine
Concen: 35.17 ng
RT: 3.393 min Scan# 86
Delta R.T. -0.006 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17



Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion: 79 Resp: 119
Ion Ratio Lower Upper
79 100
52 15.1 41.3 61.9#
51 68.8 21.2 31.8#



#4
n-Nitrosodimethylamine
Concen: 119.54 ng
RT: 3.334 min Scan# 76
Delta R.T. 0.012 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion: 42 Resp: 121
Ion Ratio Lower Upper
42 100
74 19.2 148.6 223.0#
44 130.4 5.3 7.9#









#5
2-Fluorophenol
Concen: 12.29 ng
RT: 5.117 min Scan# 379
Delta R.T. 0.006 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	112	Resp:	38
Ion	Ratio	Lower	Upper
112	100		
64	41.7	32.8	49.2
63	104.2	17.0	25.4



#6
Aniline
Concen: 37.39 ng
RT: 6.846 min Scan# 673
Delta R.T. 0.012 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:	93	Resp:	177
Ion	Ratio	Lower	Upper
93	100		
66	39.5	25.4	38.0
65	37.7	12.5	18.7









#7
Phenol-d6
Concen: 7.28 ng
RT: 6.699 min Scan# 648
Delta R.T. -0.006 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion: 99 Resp: 30
Ion Ratio Lower Upper
99 100
42 284.0 8.6 13.0#
71 72.0 23.4 35.2#



#8
2-Chlorophenol
Concen: 9.87 ng
RT: 6.905 min Scan# 683
Delta R.T. -0.170 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion: 128 Resp: 34
Ion Ratio Lower Upper
128 100
130 41.2 13.0 53.0
64 29.4 10.4 50.4









#9
Benzaldehyde
Concen: 20.45 ng
RT: 6.687 min Scan# 646
Delta R.T. 0.041 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampled :
DFTPP


Tgt Ion: 77 Resp: 47
Ion Ratio Lower Upper
77 100
105 54.5 97.1 137.1#
106 8.1 93.7 133.7#



#10
Phenol
Concen: 3.30 ng
RT: 6.734 min Scan# 654
Delta R.T. 0.000 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion: 94 Resp: 13
Ion Ratio Lower Upper
94 100
65 63.3 2.3 42.3#
66 113.3 10.9 50.9#









#11
bis(2-Chloroethyl)ether
Concen: 56.73 ng
RT: 6.846 min Scan# 673
Delta R.T. -0.088 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampled :
DFTPP


Tgt Ion: 93	Resp:	177	
Ion	Ratio	Lower	Upper
93	100		
63	40.4	39.0	79.0
95	46.5	12.8	52.8



#12
1,3-Dichlorobenzene
Concen: 4.86 ng
RT: 7.393 min Scan# 766
Delta R.T. 0.000 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:146	Resp:	20	
Ion	Ratio	Lower	Upper
146	100		
148	73.3	52.5	78.7
75	83.3	18.1	27.1#









#13
1,4-Dichlorobenzene
Concen: 6.96 ng
RT: 7.522 min Scan# 788
Delta R.T. -0.011 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	146	Resp:	29
Ion	Ratio	Lower	Upper
146	100		
148	48.3	51.3	76.9#
111	10.3	28.9	43.3#



#14
1,2-Dichlorobenzene
Concen: 8.00 ng
RT: 7.858 min Scan# 845
Delta R.T. 0.012 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:	146	Resp:	32
Ion	Ratio	Lower	Upper
146	100		
148	68.0	51.6	77.4
111	56.0	31.0	46.6#









#15
Benzyl Alcohol
Concen: 50.01 ng
RT: 7.746 min Scan# 826
Delta R.T. -0.006 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampled :
DFTPP


Tgt Ion: 79 Resp: 137
Ion Ratio Lower Upper
79 100
108 12.4 73.5 110.3#
77 51.8 50.3 75.5



#16
2,2'-oxybis(1-Chloropropane)
Concen: 23.36 ng
RT: 8.040 min Scan# 876
Delta R.T. 0.000 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17


Tgt Ion: 45 Resp: 53
Ion Ratio Lower Upper
45 100
77 88.6 2.3 42.3#
79 100.0 0.0 36.6#








#17
2-Methylphenol
Concen: 6.67 ng
RT: 7.969 min Scan# 864
Delta R.T. 0.000 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Instrument :
BNA_P
ClientSampleId :
DFTPP







Tgt Ion:	107	Resp:	19
Ion	Ratio	Lower	Upper
107	100		
108	100.0	89.6	134.4
77	313.3	31.8	47.6#
79	206.7	31.1	46.7#



#18
Hexachloroethane
Concen: 44.45 ng
RT: 8.587 min Scan# 969
Delta R.T. 0.024 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:	117	Resp:	69
Ion	Ratio	Lower	Upper
117	100		
119	21.3	78.8	118.2#
201	15.0	92.3	138.5#









#19
n-Nitroso-di-n-propylamine
Concen: 15.40 ng
RT: 8.322 min Scan# 924
Delta R.T. 0.012 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion: 70	Resp: 34
Ion Ratio	Lower Upper
70 100	
42 185.0	32.8 49.2#
101 70.0	9.6 14.4#
130 25.0	24.3 36.5



#20
3+4-Methylphenols
Concen: 11.61 ng
RT: 8.293 min Scan# 919
Delta R.T. -0.011 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt	Ion:107	Resp:	44
Ion	Ratio	Lower	Upper
107	100		
108	43.8	69.0	109.0#
77	95.8	8.5	48.5#
79	125.0	4.8	44.8#









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.287 min Scan# 1258
Delta R.T. 0.012 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampled :
DFTPP


Tgt Ion:	136	Resp:	29
Ion	Ratio	Lower	Upper
136	100		
137	0.0	8.6	13.0#
54	76.0	3.6	5.4#
68	120.0	3.3	4.9#



#22
Acetophenone
Concen: 32.29 ng
RT: 8.322 min Scan# 924
Delta R.T. 0.000 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:	105	Resp:	23
Ion	Ratio	Lower	Upper
105	100		
71	26.8	2.6	3.8#
51	98.2	13.5	20.3#
120	19.6	19.6	29.4









#23
Nitrobenzene-d5
Concen: 109.39 ng
RT: 8.563 min Scan# 965
Delta R.T. -0.088 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	82	Resp:	54
Ion	Ratio	Lower	Upper
82	100		
128	56.4	45.6	68.4
54	34.5	30.1	45.1



#24
Nitrobenzene
Concen: 137.67 ng
RT: 8.722 min Scan# 992
Delta R.T. 0.030 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:	77	Resp:	68
Ion	Ratio	Lower	Upper
77	100		
123	34.3	48.6	72.8#
65	29.4	10.0	15.0#









#25
Isophorone
Concen: 25.42 ng
RT: 9.240 min Scan# 1080
Delta R.T. 0.024 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion: 82 Resp: 23
Ion Ratio Lower Upper
82 100
95 100.0 6.4 9.6#
138 7.5 18.7 28.1#



#26
2-Nitrophenol
Concen: 55.60 ng
RT: 9.422 min Scan# 1111
Delta R.T. 0.012 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion: 139 Resp: 15
Ion Ratio Lower Upper
139 100
109 192.9 22.1 33.1#
65 100.0 26.5 39.7#









#27
2,4-Dimethylphenol
Concen: 65.53 ng
RT: 9.699 min Scan# 1158
Delta R.T. 0.212 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	122	Resp:	25
Ion	Ratio	Lower	Upper
122	100		
107	115.4	87.3	130.9
121	65.4	47.3	70.9



#28
bis(2-Chloroethoxy)methane
Concen: 136.18 ng
RT: 9.728 min Scan# 1163
Delta R.T. 0.024 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:	93	Resp:	82
Ion	Ratio	Lower	Upper
93	100		
95	30.4	26.7	40.1
123	29.0	10.5	15.7









#29
2,4-Dichlorophenol
Concen: 20.98 ng
RT: 9.969 min Scan# 1204
Delta R.T. 0.012 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:162	Resp:	10
Ion Ratio	Lower	Upper
162	100	
164	65.0	45.8 85.8
98	45.0	10.2 50.2



#30
1,2,4-Trichlorobenzene
Concen: 22.98 ng
RT: 10.193 min Scan# 1242
Delta R.T. 0.041 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:180	Resp:	13
Ion Ratio	Lower	Upper
180	100	
182	95.2	76.2 114.2
145	28.6	23.8 35.8









#31
Naphthalene
Concen: 44.39 ng
RT: 10.322 min Scan# 1264
Delta R.T. 0.000 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:128 Resp: 68
Ion Ratio Lower Upper
128 100
129 0.0 8.8 13.2#
127 41.9 10.6 15.8#



#32
Benzoic acid
Concen: 83.26 ng
RT: 9.699 min Scan# 1158
Delta R.T. 0.006 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:122 Resp: 25
Ion Ratio Lower Upper
122 100
105 196.2 94.1 134.1#
77 365.4 50.1 90.1#









#33
4-Chloroaniline
Concen: 43.02 ng
RT: 10.475 min Scan# 1290
Delta R.T. 0.024 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampled :
DFTPP


Tgt Ion:	127	Resp:	28
Ion	Ratio	Lower	Upper
127	100		
129	139.3	25.7	38.5#
65	128.6	17.3	25.9#
92	60.7	12.9	19.3#



#34
Hexachlorobutadiene
Concen: 20.72 ng
RT: 10.675 min Scan# 1324
Delta R.T. 0.059 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:	225	Resp:	8
Ion	Ratio	Lower	Upper
225	100		
223	53.3	50.1	75.1
227	73.3	51.4	77.2









#35
Caprolactam
Concen: 119.43 ng
RT: 11.252 min Scan# 1422
Delta R.T. 0.036 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:113 Resp: 19
Ion Ratio Lower Upper
113 100
55 448.1 77.3 117.3#
56 325.9 55.9 95.9#



#36
4-Chloro-3-methylphenol
Concen: 52.53 ng
RT: 11.628 min Scan# 1486
Delta R.T. 0.018 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:107 Resp: 27
Ion Ratio Lower Upper
107 100
144 41.7 22.7 34.1#
142 0.0 70.7 106.1#









#37
2-Methylnaphthalene
Concen: 21.63 ng
RT: 11.969 min Scan# 1544
Delta R.T. 0.018 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:142 Resp: 24
Ion Ratio Lower Upper
142 100
141 95.7 71.4 107.2
115 708.7 25.8 38.6#



#38
1-Methylnaphthalene
Concen: 26.57 ng
RT: 12.199 min Scan# 1583
Delta R.T. 0.024 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:142 Resp: 28
Ion Ratio Lower Upper
142 100
141 100.0 74.4 111.6
116 337.0 2.9 4.3#









#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.169 min Scan# 1918
Delta R.T. 0.012 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampled :
DFTPP


Tgt Ion:164 Resp: 49
Ion Ratio Lower Upper
164 100
162 34.7 79.4 119.2#
160 30.6 36.1 54.1#



#40
1,2,4,5-Tetrachlorobenzene
Concen: 5.61 ng
RT: 12.340 min Scan# 1607
Delta R.T. 0.000 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:216 Resp: 9
Ion Ratio Lower Upper
216 100
214 333.3 63.0 94.6#
179 133.3 16.9 25.3#
108 244.4 12.6 19.0#









#41

Hexachlorocyclopentadiene
Concen: 44.14 ng
RT: 12.257 min Scan# 1593
Delta R.T. -0.059 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Instrument :
BNA_P
ClientSampleId :
DFTPP




Tgt Ion:	237	Resp:	19
Ion	Ratio	Lower	Upper
237	100		
235	44.4	43.2	83.2
272	0.0	0.0	33.3




#42

2,4,6-Tribromophenol
Concen: 137.82 ng
RT: 15.722 min Scan# 2182
Delta R.T. 0.059 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:	330	Resp:	103
Ion	Ratio	Lower	Upper
330	100		
332	134.0	77.1	115.7#
141	106.8	23.4	35.2#









#43
2,4,6-Trichlorophenol
Concen: 25.07 ng
RT: 12.610 min Scan# 1653
Delta R.T. 0.018 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	196	Resp:	26
Ion	Ratio	Lower	Upper
196	100		
198	40.9	77.0	115.6#
200	40.9	25.1	37.7#



#44
2,4,5-Trichlorophenol
Concen: 24.37 ng
RT: 12.710 min Scan# 1670
Delta R.T. 0.024 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:	196	Resp:	26
Ion	Ratio	Lower	Upper
196	100		
198	40.0	77.3	115.9#
97	144.0	31.0	46.4#
132	68.0	20.9	31.3#









#45
2-Fluorobiphenyl
Concen: 157.60 ng
RT: 12.804 min Scan# 1686
Delta R.T. 0.030 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	172	Resp:	528
Ion Ratio	Lower	Upper	
172	100		
171	33.7	28.5	42.7
170	10.5	19.2	28.8#



#46
1,1'-Biphenyl
Concen: 4.38 ng
RT: 12.987 min Scan# 1717
Delta R.T. 0.000 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:	154	Resp:	16
Ion Ratio	Lower	Upper	
154	100		
153	0.0	20.4	60.4#
76	119.0	0.0	29.3#









#47
2-Chloronaphthalene
Concen: 23.13 ng
RT: 13.022 min Scan# 1723
Delta R.T. 0.000 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:162 Resp: 70
Ion Ratio Lower Upper
162 100
127 38.9 29.9 44.9
164 25.0 26.1 39.1#



#48
2-Nitroaniline
Concen: 49.62 ng
RT: 13.251 min Scan# 1762
Delta R.T. 0.006 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion: 65 Resp: 38
Ion Ratio Lower Upper
65 100
92 69.0 70.6 106.0#
138 65.5 133.6 200.4#









#49
Acenaphthylene
Concen: 93.21 ng
RT: 13.898 min Scan# 1872
Delta R.T. 0.024 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:152 Resp: 431
Ion Ratio Lower Upper
152 100
151 22.2 15.9 23.9
153 11.1 10.5 15.7



#50
Dimethylphthalate
Concen: 25.56 ng
RT: 13.646 min Scan# 1829
Delta R.T. 0.018 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:163 Resp: 100
Ion Ratio Lower Upper
163 100
194 17.4 3.3 4.9#
164 31.5 8.2 12.2#









#51
2,6-Dinitrotoluene
Concen: 40.54 ng
RT: 13.904 min Scan# 1873
Delta R.T. 0.159 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	165	Resp:	34
Ion	Ratio	Lower	Upper
165	100		
63	28.8	27.8	41.6
89	40.4	34.6	52.0



#53
3-Nitroaniline
Concen: 22.01 ng
RT: 14.075 min Scan# 1902
Delta R.T. -0.006 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:	138	Resp:	20
Ion	Ratio	Lower	Upper
138	100		
108	166.7	7.5	11.3#
92	209.5	75.0	112.6#









#54
2,4-Dinitrophenol
Concen: 39.33 ng
RT: 14.345 min Scan# 1948
Delta R.T. 0.018 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	184	Resp:	14
Ion	Ratio	Lower	Upper
184	100		
63	115.8	32.3	48.5#
154	105.3	49.1	73.7#



#55
Dibenzofuran
Concen: 9.35 ng
RT: 14.581 min Scan# 1988
Delta R.T. 0.018 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:	168	Resp:	42
Ion	Ratio	Lower	Upper
168	100		
139	27.5	29.1	43.7#
169	150.0	10.4	15.6#









#56
4-Nitrophenol
Concen: 55.81 ng
RT: 14.363 min Scan# 1951
Delta R.T. -0.088 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:	139	Resp:	31
Ion	Ratio	Lower	Upper
139	100		
109	54.8	44.6	84.6
65	66.7	41.6	81.6



#57
2,4-Dinitrotoluene
Concen: 15.47 ng
RT: 14.569 min Scan# 1986
Delta R.T. 0.012 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:	165	Resp:	18
Ion	Ratio	Lower	Upper
165	100		
63	96.7	20.6	31.0#
89	206.7	42.6	63.8#
182	70.0	2.4	3.6#





#58

Fluorene

Concen: 29.04 ng

RT: 15.234 min Scan# 2099

Delta R.T. 0.018 min

Lab File: BP003505.D

Acq: 06 Oct 2020 01:17





Instrument :

BNA_P

ClientSampled :

DFTPP

Tgt Ion:166	Resp:	103
Ion Ratio	Lower	Upper
166	100	
165	72.0	77.8 116.8#
167	72.0	11.0 16.4#



#59

2,3,4,6-Tetrachlorophenol

Concen: 130.57 ng




RT: 14.804 min Scan# 2026


Delta R.T. -0.006 min

Lab File: BP003505.D

Acq: 06 Oct 2020 01:17

Tgt Ion:232	Resp:	131
Ion Ratio	Lower	Upper
232	100	
131	67.9	34.1 51.1#
130	12.2	1.8 2.8#
166	0.0	24.6 36.8#









#60
Diethylphthalate
Concen: 265.09 ng
RT: 15.016 min Scan# 2062
Delta R.T. 0.012 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampled :
DFTPP


Tgt Ion:149 Resp: 1058
Ion Ratio Lower Upper
149 100
177 19.5 18.0 27.0
150 16.4 10.0 15.0#



#61
4-Chlorophenyl-phenylether
Concen: 13.43 ng
RT: 15.222 min Scan# 2097
Delta R.T. 0.006 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:204 Resp: 26
Ion Ratio Lower Upper
204 100
206 77.8 25.9 38.9#
141 114.8 46.5 69.7#









#62
4-Nitroaniline
Concen: 32.14 ng
RT: 15.387 min Scan# 2125
Delta R.T. 0.136 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampled :
DFTPP


Tgt Ion:138 Resp: 31
Ion Ratio Lower Upper
138 100
92 50.0 20.7 60.7
108 81.3 47.8 87.8



#63
Azobenzene
Concen: 50.95 ng
RT: 15.534 min Scan# 2150
Delta R.T. 0.024 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion: 77 Resp: 155
Ion Ratio Lower Upper
77 100
182 45.9 16.9 56.9
105 37.7 5.6 45.6
51 68.6 2.9 42.9#





#64

Phenanthrene-d10

Concen: 20.00 ng

RT: 16.928 min Scan# 2387

Delta R.T. 0.012 min

Lab File: BP003505.D

Acq: 06 Oct 2020 01:17





Instrument :

BNA_P

ClientSampleId :

DFTPP

Tgt Ion:188	Resp:	464
Ion Ratio	Lower	Upper
188	100	
94	7.9	6.2 9.2
80	17.5	6.2 9.2#



#65

4,6-Dinitro-2-methylphenol

Concen: 7.76 ng




RT: 15.334 min Scan# 2116


Delta R.T. -0.006 min

Lab File: BP003505.D

Acq: 06 Oct 2020 01:17

Tgt Ion:198	Resp:	19
Ion Ratio	Lower	Upper
198	100	
51	290.9	6.4 46.4#
105	254.5	18.9 58.9#









#66
n-Nitrosodiphenylamine
Concen: 18.56 ng
RT: 15.457 min Scan# 2137
Delta R.T. 0.024 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampled :
DFTPP


Tgt Ion:	169	Resp:	253
Ion Ratio	Lower	Upper	
169	100		
168	72.9	53.8	80.6
167	78.3	28.9	43.3#



#67
4-Bromophenyl-phenylether
Concen: 12.84 ng
RT: 16.134 min Scan# 2252
Delta R.T. 0.024 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:	248	Resp:	72
Ion Ratio	Lower	Upper	
248	100		
250	83.8	76.7	115.1
141	40.5	53.0	79.6#









#68
Hexachlorobenzene
Concen: 23.86 ng
RT: 16.245 min Scan# 2271
Delta R.T. 0.012 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:284 Resp: 159
Ion Ratio Lower Upper
284 100
142 64.6 24.2 36.4#
249 39.6 23.8 35.8#



#69
Atrazine
Concen: 6935.50 ng
RT: 16.587 min Scan# 2329
Delta R.T. 0.188 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:200 Resp: 37079
Ion Ratio Lower Upper
200 100
173 0.1 5.9 45.9#
215 0.4 28.4 68.4#









#70
Pentachlorophenol
Concen: 60099.78 ng
RT: 16.592 min Scan# 2330
Delta R.T. 0.000 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:266 Resp: 187760
Ion Ratio Lower Upper
266 100
268 64.2 51.1 76.7
264 63.9 50.6 76.0



#71
Phenanthrene
Concen: 36.52 ng
RT: 16.963 min Scan# 2393
Delta R.T. 0.006 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:178 Resp: 922
Ion Ratio Lower Upper
178 100
176 36.4 15.3 22.9#
179 29.2 12.5 18.7#









#72
 Anthracene
 Concen: 40.72 ng
 RT: 17.069 min Scan# 2411
 Delta R.T. 0.018 min
 Lab File: BP003505.D
 Acq: 06 Oct 2020 01:17




Instrument :
 BNA_P
 ClientSampleId :
 DFTPP


Tgt Ion:178	Resp:	1013
Ion Ratio	Lower	Upper
178	100	
176	9.2	15.0 22.4#
179	30.4	12.5 18.7#



#73
 Carbazole
 Concen: 330.65 ng
 RT: 17.386 min Scan# 2465
 Delta R.T. 0.059 min
 Lab File: BP003505.D
 Acq: 06 Oct 2020 01:17

Tgt Ion:167	Resp:	7814
Ion Ratio	Lower	Upper
167	100	
166	13.8	17.6 26.4#
139	3.1	10.2 15.4#









#74
Di-n-butylphthalate
Concen: 97.25 ng
RT: 17.898 min Scan# 2552
Delta R.T. 0.000 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:149 Resp: 2938
Ion Ratio Lower Upper
149 100
150 11.5 7.3 10.9#
104 8.8 3.4 5.0#



#75
Fluoranthene
Concen: 54.88 ng
RT: 18.963 min Scan# 2733
Delta R.T. 0.018 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:202 Resp: 1742
Ion Ratio Lower Upper
202 100
101 15.3 0.0 29.4
203 22.4 0.0 37.8





#76

Chrysene-d12

Concen: 20.00 ng

RT: 21.033 min Scan# 3085

Delta R.T. 0.006 min

Lab File: BP003505.D

Acq: 06 Oct 2020 01:17





Instrument :

BNA_P

ClientSampleId :

DFTPP

Tgt Ion:240	Resp:	984
Ion Ratio	Lower	Upper
240	100	
120	43.3	7.8 11.6#
236	51.6	20.8 31.2#



#77

Benzidine

Concen: 16438.87 ng




RT: 19.133 min Scan# 2762


Delta R.T. 0.000 min

Lab File: BP003505.D

Acq: 06 Oct 2020 01:17

Tgt Ion:184	Resp:	501594
Ion Ratio	Lower	Upper
184	100	
185	14.1	11.5 17.3
183	12.0	9.6 14.4





#78

Pyrene

Concen: 31.34 ng

RT: 19.310 min Scan# 2792

Delta R.T. 0.012 min

Lab File: BP003505.D

Acq: 06 Oct 2020 01:17





Instrument :

BNA_P

ClientSampleId :

DFTPP

Tgt Ion:202	Resp:	1923
Ion Ratio	Lower	Upper
202	100	
200	24.9	17.1 25.7
203	23.6	14.2 21.2#



#79

Terphenyl-d14

Concen: 92.66 ng




RT: 19.498 min Scan# 2824


Delta R.T. -0.006 min

Lab File: BP003505.D

Acq: 06 Oct 2020 01:17

Tgt Ion:244	Resp:	4376
Ion Ratio	Lower	Upper
244	100	
212	8.7	5.8 8.6#
122	13.5	7.8 11.8#









#80
Butylbenzylphthalate
Concen: 30.45 ng
RT: 20.186 min Scan# 2941
Delta R.T. 0.006 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:149 Resp: 855
Ion Ratio Lower Upper
149 100
91 102.3 45.4 68.0#
206 24.2 17.1 25.7



#81
Benzo(a)anthracene
Concen: 37.79 ng
RT: 21.022 min Scan# 3083
Delta R.T. 0.012 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:228 Resp: 2404
Ion Ratio Lower Upper
228 100
226 29.4 22.1 33.1
229 40.6 15.9 23.9#









#82
3,3'-Dichlorobenzidine
Concen: 211.03 ng
RT: 20.957 min Scan# 3072
Delta R.T. 0.012 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampled :
DFTPP


Tgt Ion:252 Resp: 5188
Ion Ratio Lower Upper
252 100
254 63.2 51.0 76.4
126 16.8 7.9 11.9#



#83
Chrysene
Concen: 36.77 ng
RT: 21.069 min Scan# 3091
Delta R.T. 0.006 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:228 Resp: 2247
Ion Ratio Lower Upper
228 100
226 28.7 24.4 36.6
229 35.6 15.8 23.6#









#84

Bis(2-ethylhexyl)phthalate
Concen: 63.55 ng
RT: 20.945 min Scan# 3070
Delta R.T. -0.006 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Instrument :
BNA_P
ClientSampleId :
DFTPP




Tgt Ion:149 Resp: 2512
Ion Ratio Lower Upper
149 100
167 35.1 22.6 34.0#
279 7.6 4.2 6.2#




#85

Di-n-octyl phthalate
Concen: 34.00 ng
RT: 21.792 min Scan# 3214
Delta R.T. -0.006 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:149 Resp: 2456
Ion Ratio Lower Upper
149 100
167 2.2 1.3 1.9#
43 0.0 3.7 5.5#





#86

Perylene-d12

Concen: 20.00 ng

RT: 23.210 min Scan# 3455

Delta R.T. 0.012 min

Lab File: BP003505.D

Acq: 06 Oct 2020 01:17





Instrument :

BNA_P

ClientSampled :

DFTPP

Tgt Ion:264	Resp:	792
Ion Ratio	Lower	Upper
264	100	
260	32.7	18.8 28.2#
265	144.4	17.4 26.2#



#87

Indeno(1,2,3-cd)pyrene

Concen: 55.17 ng




RT: 25.415 min Scan# 3830


Delta R.T. 0.018 min

Lab File: BP003505.D

Acq: 06 Oct 2020 01:17

Tgt Ion:276	Resp:	3312
Ion Ratio	Lower	Upper
276	100	
138	0.0	18.7 28.1#
277	9.0	20.4 30.6#









#88
Benzo(b)fluoranthene
Concen: 51.07 ng
RT: 22.569 min Scan# 3346
Delta R.T. 0.024 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17




Instrument :
BNA_P
ClientSampleId :
DFTPP


Tgt Ion:252 Resp: 2554
Ion Ratio Lower Upper
252 100
253 34.7 17.6 26.4#
125 38.6 7.3 10.9#



#89
Benzo(k)fluoranthene
Concen: 43.23 ng
RT: 22.598 min Scan# 3351
Delta R.T. 0.006 min
Lab File: BP003505.D
Acq: 06 Oct 2020 01:17

Tgt Ion:252 Resp: 2072
Ion Ratio Lower Upper
252 100
253 40.8 17.8 26.8#
125 36.1 7.4 11.2#





#90

Benzo(a)pyrene

Concen: 59.69 ng

RT: 23.127 min Scan# 3441

Delta R.T. 0.024 min

Lab File: BP003505.D

Acq: 06 Oct 2020 01:17





Instrument :

BNA_P

ClientSampleId :

DFTPP

Tgt Ion:	252	Resp:	2815
Ion Ratio	Lower	Upper	
252	100		
253	43.0	17.4	26.0#
125	54.5	8.2	12.4#



#91

Dibenzo(a,h)anthracene

Concen: 57.72 ng




RT: 25.415 min Scan# 3830


Delta R.T. 0.012 min

Lab File: BP003505.D

Acq: 06 Oct 2020 01:17

Tgt Ion:	278	Resp:	2859
Ion Ratio	Lower	Upper	
278	100		
139	33.7	12.4	18.6#
279	42.3	19.0	28.6#





#92
 Benzo(g,h,i)perylene
 Concen: 49.38 ng
 RT: 26.098 min Scan# 3946
 Delta R.T. 0.030 min
 Lab File: BP003505.D
 Acq: 06 Oct 2020 01:17

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Tgt Ion:276 Resp: 2442
 Ion Ratio Lower Upper
 276 100
 277 32.3 19.0 28.6#
 138 33.3 16.6 24.8#

