

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100621\
 Data File : BP007366.D
 Acq On : 06 Oct 2021 21:59
 Operator : CG/JU
 Sample : M4067-01DL 10X
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SU-02-100421DL

Manual Integrations
 APPROVED

mohammad
 10/8/2021 8:34:32 AM

Quant Time: Oct 07 01:48:41 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 01 12:49:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.858	152	162521	20.00	ng	0.00
21) Naphthalene-d8	10.658	136	606045	20.00	ng	# 0.00
39) Acenaphthene-d10	14.493	164	374450	20.00	ng	0.00
64) Phenanthrene-d10	17.245	188	795884	20.00	ng	# 0.00
76) Chrysene-d12	21.339	240	828031	20.00	ng	0.00
86) Perylene-d12	23.745	264	900030	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.446	112	88310	9.10	ng	0.00
7) Phenol-d6	7.046	99	94888	7.15	ng	0.00
23) Nitrobenzene-d5	9.028	82	57563	5.53	ng	0.00
42) 2,4,6-Tribromophenol	15.993	330	37677	7.76	ng	0.00
45) 2-Fluorobiphenyl	13.122	172	140237	5.37	ng	0.00
79) Terphenyl-d14	19.804	244	222697	5.15	ng	0.00
Target Compounds						Qvalue
49) Acenaphthylene	14.216	152	247063	7.43	ng	100
71) Phenanthrene	17.292	178	492090	11.53	ng	99
72) Anthracene	17.381	178	228249	5.47	ng	99
75) Fluoranthene	19.263	202	1244837	25.32	ng	98
78) Pyrene	19.616	202	1154726	21.26	ng	99
81) Benzo(a)anthracene	21.322	228	675750	12.57	ng	96
83) Chrysene	21.375	228	586975	11.43	ng	96
87) Indeno(1,2,3-cd)pyrene	26.257	276	502858	7.77	ng	98
88) Benzo(b)fluoranthene	23.010	252	940222	17.48	ng	# 96
89) Benzo(k)fluoranthene	23.051	252	342657m	6.29	ng	
90) Benzo(a)pyrene	23.639	252	720592	15.77	ng	# 97
91) Dibenzo(a,h)anthracene	26.263	278	133608	2.46	ng	# 84
92) Benzo(g,h,i)perylene	27.033	276	474662	8.97	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100621\
Data File : BP007366.D
Acq On : 06 Oct 2021 21:59
Operator : CG/JU
Sample : M4067-01DL 10X
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SU-02-100421DL

Manual Integrations
APPROVED

mohammad
10/8/2021 8:34:32 AM

Quant Time: Oct 07 01:48:41 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 01 12:49:05 2021
Response via : Initial Calibration

