

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP100625\  
 Data File : BP025929.D  
 Acq On : 06 Oct 2025 19:04  
 Operator : CG/JU  
 Sample : Q3269-03MSD  
 Misc :  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 EG1-TP1MSD

Manual Integrations  
 APPROVED

Reviewed By :Rahul Chavli 10/07/2025  
 Supervised By :Jagrut Upadhyay 10/07/2025

Quant Time: Oct 07 02:06:34 2025  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP091225.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Oct 06 19:16:54 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-------------------------------|--------|------|----------|---------|-------|----------|
| -----                         |        |      |          |         |       |          |
| Internal Standards            |        |      |          |         |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.737  | 152  | 156742   | 20.000  | ng    | -0.02    |
| 21) Naphthalene-d8            | 10.496 | 136  | 597901   | 20.000  | ng    | -0.03    |
| 39) Acenaphthene-d10          | 14.343 | 164  | 371539   | 20.000  | ng    | -0.04    |
| 64) Phenanthrene-d10          | 17.154 | 188  | 743213   | 20.000  | ng    | -0.03    |
| 76) Chrysene-d12              | 21.601 | 240  | 809368   | 20.000  | ng    | -0.04    |
| 86) Perylene-d12              | 24.948 | 264  | 926256   | 20.000  | ng    | -0.05    |
|                               |        |      |          |         |       |          |
| System Monitoring Compounds   |        |      |          |         |       |          |
| 5) 2-Fluorophenol             | 5.378  | 112  | 1210923  | 124.656 | ng    | 0.00     |
| 7) Phenol-d6                  | 6.943  | 99   | 1459555  | 113.040 | ng    | 0.00     |
| 23) Nitrobenzene-d5           | 8.884  | 82   | 1033296  | 76.233  | ng    | -0.02    |
| 42) 2,4,6-Tribromophenol      | 15.872 | 330  | 692292   | 149.387 | ng    | -0.04    |
| 45) 2-Fluorobiphenyl          | 12.954 | 172  | 2311618  | 82.843  | ng    | -0.04    |
| 79) Terphenyl-d14             | 19.872 | 244  | 3786551  | 84.164  | ng    | -0.04    |
|                               |        |      |          |         |       |          |
| Target Compounds              |        |      |          |         |       | Qvalue   |
| 2) 1,4-Dioxane                | 3.314  | 88   | 145261   | 35.629  | ng    | # 81     |
| 3) Pyridine                   | 3.714  | 79   | 378846   | 33.746  | ng    | 95       |
| 4) n-Nitrosodimethylamine     | 3.614  | 42   | 197296   | 37.244  | ng    | 87       |
| 6) Aniline                    | 7.078  | 93   | 344388   | 19.430  | ng    | 99       |
| 8) 2-Chlorophenol             | 7.319  | 128  | 491712   | 46.141  | ng    | 99       |
| 9) Benzaldehyde               | 6.902  | 77   | 118015   | 15.249  | ng    | 96       |
| 10) Phenol                    | 6.972  | 94   | 562139   | 41.289  | ng    | 97       |
| 11) bis(2-Chloroethyl)ether   | 7.166  | 93   | 484900   | 44.330  | ng    | 94       |
| 12) 1,3-Dichlorobenzene       | 7.625  | 146  | 517268   | 42.648  | ng    | 98       |
| 13) 1,4-Dichlorobenzene       | 7.772  | 146  | 533081   | 43.148  | ng    | 100      |
| 14) 1,2-Dichlorobenzene       | 8.084  | 146  | 509061   | 42.414  | ng    | 100      |
| 15) Benzyl Alcohol            | 7.984  | 79   | 448977   | 42.081  | ng    | 98       |
| 16) 2,2'-oxybis(1-Chloropr... | 8.255  | 45   | 691013   | 38.630  | ng    | 97       |
| 17) 2-Methylphenol            | 8.196  | 107  | 406552   | 44.291  | ng    | 99       |
| 18) Hexachloroethane          | 8.796  | 117  | 199468   | 42.268  | ng    | 98       |
| 19) n-Nitroso-di-n-propyla... | 8.531  | 70   | 373122   | 41.432  | ng    | 95       |
| 20) 3+4-Methylphenols         | 8.519  | 107  | 547868   | 42.618  | ng    | 98       |
| 22) Acetophenone              | 8.555  | 105  | 816012   | 51.071  | ng    | # 98     |
| 24) Nitrobenzene              | 8.925  | 77   | 548632   | 45.119  | ng    | 98       |
| 25) Isophorone                | 9.449  | 82   | 1058219  | 44.417  | ng    | 99       |
| 26) 2-Nitrophenol             | 9.631  | 139  | 268817   | 49.793  | ng    | 90       |
| 27) 2,4-Dimethylphenol        | 9.696  | 122  | 458076   | 48.365  | ng    | 98       |
| 28) bis(2-Chloroethoxy)met... | 9.913  | 93   | 640007   | 46.473  | ng    | 99       |
| 29) 2,4-Dichlorophenol        | 10.190 | 162  | 468930   | 49.658  | ng    | 98       |
| 30) 1,2,4-Trichlorobenzene    | 10.366 | 180  | 501060   | 47.081  | ng    | 98       |
| 31) Naphthalene               | 10.549 | 128  | 1464195  | 45.417  | ng    | 99       |
| 32) Benzoic acid              | 9.890  | 122  | 326440   | 46.359  | ng    | 95       |
| 33) 4-Chloroaniline           | 10.678 | 127  | 182724   | 13.598  | ng    | 98       |
| 34) Hexachlorobutadiene       | 10.831 | 225  | 325289   | 47.480  | ng    | 100      |
| 35) Caprolactam               | 11.484 | 113  | 136901   | 37.608  | ng    | 95       |
| 36) 4-Chloro-3-methylphenol   | 11.813 | 107  | 534367   | 47.464  | ng    | 95       |
| 37) 2-Methylnaphthalene       | 12.154 | 142  | 948924   | 45.827  | ng    | 98       |
| 38) 1-Methylnaphthalene       | 12.378 | 142  | 991064   | 44.883  | ng    | 100      |
| 40) 1,2,4,5-Tetrachloroben... | 12.531 | 216  | 646366   | 57.411  | ng    | 99       |
| 41) Hexachlorocyclopentadiene | 12.501 | 237  | 506678   | 82.630  | ng    | 99       |
| 43) 2,4,6-Trichlorophenol     | 12.784 | 196  | 395647   | 53.456  | ng    | 99       |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP100625\  
 Data File : BP025929.D  
 Acq On : 06 Oct 2025 19:04  
 Operator : CG/JU  
 Sample : Q3269-03MSD  
 Misc :  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 EG1-TP1MSD

Manual Integrations  
 APPROVED

Reviewed By :Rahul Chavli 10/07/2025  
 Supervised By :Jagrut Upadhyay 10/07/2025

Quant Time: Oct 07 02:06:34 2025  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP091225.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Oct 06 19:16:54 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-------------------------------|--------|------|----------|---------|-------|----------|
| 44) 2,4,5-Trichlorophenol     | 12.878 | 196  | 423781   | 51.447  | ng    | 95       |
| 46) 1,1'-Biphenyl             | 13.172 | 154  | 1497069  | 54.896  | ng    | 99       |
| 47) 2-Chloronaphthalene       | 13.219 | 162  | 1056352  | 49.194  | ng    | 98       |
| 48) 2-Nitroaniline            | 13.431 | 65   | 359079   | 50.165  | ng    | 96       |
| 49) Acenaphthylene            | 14.066 | 152  | 1800125  | 50.595  | ng    | 100      |
| 50) Dimethylphthalate         | 13.801 | 163  | 1412063  | 49.780  | ng    | 99       |
| 51) 2,6-Dinitrotoluene        | 13.925 | 165  | 305445   | 50.825  | ng    | 93       |
| 52) Acenaphthene              | 14.407 | 154  | 984501   | 47.982  | ng    | 99       |
| 53) 3-Nitroaniline            | 14.272 | 138  | 206392   | 32.658  | ng    | 97       |
| 54) 2,4-Dinitrophenol         | 14.490 | 184  | 395942   | 100.045 | ng    | 98       |
| 55) Dibenzofuran              | 14.754 | 168  | 1631056  | 48.775  | ng    | 98       |
| 56) 4-Nitrophenol             | 14.637 | 139  | 394130   | 72.844  | ng    | 95       |
| 57) 2,4-Dinitrotoluene        | 14.731 | 165  | 440756   | 51.534  | ng    | 96       |
| 58) Fluorene                  | 15.413 | 166  | 1308291  | 49.200  | ng    | 99       |
| 59) 2,3,4,6-Tetrachlorophenol | 14.990 | 232  | 376785   | 50.778  | ng    | 96       |
| 60) Diethylphthalate          | 15.184 | 149  | 1423905  | 49.488  | ng    | 98       |
| 61) 4-Chlorophenyl-phenyle... | 15.401 | 204  | 672776   | 50.202  | ng    | 97       |
| 62) 4-Nitroaniline            | 15.454 | 138  | 300951   | 46.484  | ng    | 97       |
| 63) Azobenzene                | 15.701 | 77   | 1325034  | 47.627  | ng    | 96       |
| 65) 4,6-Dinitro-2-methylph... | 15.519 | 198  | 269731   | 54.572  | ng    | 93       |
| 66) n-Nitrosodiphenylamine    | 15.619 | 169  | 1175664  | 51.691  | ng    | 99       |
| 67) 4-Bromophenyl-phenylether | 16.313 | 248  | 440472   | 53.947  | ng    | 98       |
| 68) Hexachlorobenzene         | 16.437 | 284  | 498850   | 54.787  | ng    | 92       |
| 69) Atrazine                  | 16.607 | 200  | 419034   | 48.199  | ng    | 97       |
| 70) Pentachlorophenol         | 16.807 | 266  | 621905   | 103.489 | ng    | 98       |
| 71) Phenanthrene              | 17.195 | 178  | 2114400  | 50.153  | ng    | 99       |
| 72) Anthracene                | 17.289 | 178  | 2167142  | 50.662  | ng    | 99       |
| 73) Carbazole                 | 17.572 | 167  | 2043964  | 51.396  | ng    | 100      |
| 74) Di-n-butylphthalate       | 18.148 | 149  | 2602269  | 53.321  | ng    | 99       |
| 75) Fluoranthene              | 19.283 | 202  | 2572890  | 50.885  | ng    | 100      |
| 77) Benzidine                 | 19.489 | 184  | 828816   | 35.550  | ng    | 99       |
| 78) Pyrene                    | 19.660 | 202  | 2702893  | 50.049  | ng    | 99       |
| 80) Butylbenzylphthalate      | 20.601 | 149  | 1222347  | 51.621  | ng    | 100      |
| 81) Benzo(a)anthracene        | 21.583 | 228  | 2754458  | 49.710  | ng    | 100      |
| 82) 3,3'-Dichlorobenzidine    | 21.501 | 252  | 790776   | 41.274  | ng    | 99       |
| 83) Chrysene                  | 21.648 | 228  | 2595782  | 49.143  | ng    | 100      |
| 84) Bis(2-ethylhexyl)phtha... | 21.501 | 149  | 1871340  | 54.017  | ng    | 100      |
| 85) Di-n-octyl phthalate      | 22.766 | 149  | 3287060  | 53.340  | ng    | 97       |
| 87) Indeno(1,2,3-cd)pyrene    | 28.759 | 276  | 3568053m | 52.971  | ng    |          |
| 88) Benzo(b)fluoranthene      | 23.889 | 252  | 2795296  | 49.350  | ng    | 99       |
| 89) Benzo(k)fluoranthene      | 23.960 | 252  | 2946691  | 50.981  | ng    | 98       |
| 90) Benzo(a)pyrene            | 24.783 | 252  | 2751550  | 49.604  | ng    | 99       |
| 91) Dibenzo(a,h)anthracene    | 28.818 | 278  | 2949384  | 53.506  | ng    | 97       |
| 92) Benzo(g,h,i)perylene      | 29.936 | 276  | 2862470  | 52.804  | ng    | 98       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP100625\  
 Data File : BP025929.D  
 Acq On : 06 Oct 2025 19:04  
 Operator : CG/JU  
 Sample : Q3269-03MSD  
 Misc :  
 ALS Vial : 7 Sample Multiplier: 1

## Instrument :

BNA\_P

## ClientSampleId :

EG1-TP1MSD

## Manual Integrations

## APPROVED

Reviewed By :Rahul Chavli 10/07/2025

Supervised By :Jagrut Upadhyay 10/07/2025


Quant Time: Oct 07 02:06:34 2025


Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP091225.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Mon Oct 06 19:16:54 2025


Response via : Initial Calibration





#1  
1,4-Dichlorobenzene-d4  
Concen: 20.000 ng  
RT: 7.737 min Scan# 81  
Delta R.T. -0.023 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD




Tgt Ion:152 Resp: 156742  
Ion Ratio Lower Upper  
152 100  
150 156.4 124.3 186.5  
115 61.2 50.3 75.5

Manual Integrations  
APPROVED



Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025




#2  
1,4-Dioxane  
Concen: 35.629 ng  
RT: 3.314 min Scan# 64  
Delta R.T. 0.006 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Tgt Ion: 88 Resp: 145261  
Ion Ratio Lower Upper  
88 100  
58 78.0 65.5 98.3  
43 63.1 25.2 37.8





#3  
Pyridine  
Concen: 33.746 ng  
RT: 3.714 min Scan# 131  
Delta R.T. 0.006 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD




Tgt Ion: 79 Resp: 378846  
Ion Ratio Lower Upper  
79 100  
52 75.2 63.7 95.5  
51 37.0 32.4 48.6

Manual Integrations  
APPROVED



Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025




#4  
n-Nitrosodimethylamine  
Concen: 37.244 ng  
RT: 3.614 min Scan# 115  
Delta R.T. 0.000 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Tgt Ion: 42 Resp: 197296  
Ion Ratio Lower Upper  
42 100  
74 139.4 99.2 148.8  
44 8.9 6.4 9.6





#5  
2-Fluorophenol  
Concen: 124.656 ng  
RT: 5.378 min Scan# 416  
Delta R.T. -0.006 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD



Tgt Ion:112 Resp: 1210923  
Ion Ratio Lower Upper  
112 100  
64 61.0 50.2 75.4  
63 31.1 26.7 40.1




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#6  
Aniline  
Concen: 19.430 ng  
RT: 7.078 min Scan# 704  
Delta R.T. -0.018 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04


Tgt Ion: 93 Resp: 344388  
Ion Ratio Lower Upper  
93 100  
66 40.5 32.6 49.0  
65 21.0 16.2 24.2





#7  
Phenol-d6  
Concen: 113.040 ng  
RT: 6.943 min Scan# 681  
Delta R.T. -0.006 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD



Tgt Ion: 99 Resp: 145955  
Ion Ratio Lower Upper  
99 100  
42 17.5 15.6 23.4  
71 35.3 28.6 43.0




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#8  
2-Chlorophenol  
Concen: 46.141 ng  
RT: 7.319 min Scan# 745  
Delta R.T. -0.018 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04


Tgt Ion:128 Resp: 491712  
Ion Ratio Lower Upper  
128 100  
130 32.4 12.6 52.6  
64 50.7 31.6 71.6





#9  
Benzaldehyde  
Concen: 15.249 ng  
RT: 6.902 min Scan# 61  
Delta R.T. -0.012 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD




Tgt Ion: 77 Resp: 118015  
Ion Ratio Lower Upper  
77 100  
105 99.2 75.2 115.2  
106 94.2 71.3 111.3

Manual Integrations  
APPROVED



Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025




#10  
Phenol  
Concen: 41.289 ng  
RT: 6.972 min Scan# 686  
Delta R.T. 0.000 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Tgt Ion: 94 Resp: 562139  
Ion Ratio Lower Upper  
94 100  
65 28.7 10.2 50.2  
66 39.0 21.4 61.4





#11  
bis(2-Chloroethyl)ether  
Concen: 44.330 ng  
RT: 7.166 min Scan# 71  
Delta R.T. -0.018 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD




Tgt Ion: 93 Resp: 484900  
Ion Ratio Lower Upper  
93 100  
63 76.2 62.6 102.6  
95 31.2 13.2 53.2

Manual Integrations  
APPROVED



Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025




#12  
1,3-Dichlorobenzene  
Concen: 42.648 ng  
RT: 7.625 min Scan# 797  
Delta R.T. -0.023 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04



Tgt Ion:146 Resp: 517268  
Ion Ratio Lower Upper  
146 100  
148 64.3 52.7 79.1  
75 32.1 26.4 39.6









#13  
1,4-Dichlorobenzene  
Concen: 43.148 ng  
RT: 7.772 min Scan# 81  
Delta R.T. -0.018 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD

Tgt Ion:146 Resp: 533081  
Ion Ratio Lower Upper  
146 100  
148 64.8 52.0 78.0  
111 42.7 33.9 50.9




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#14  
1,2-Dichlorobenzene  
Concen: 42.414 ng  
RT: 8.084 min Scan# 875  
Delta R.T. -0.023 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04


Tgt Ion:146 Resp: 509061  
Ion Ratio Lower Upper  
146 100  
148 64.7 51.7 77.5  
111 44.6 35.5 53.3





#15  
Benzyl Alcohol  
Concen: 42.081 ng  
RT: 7.984 min Scan# 81  
Delta R.T. -0.018 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD




Tgt Ion: 79 Resp: 448972  
Ion Ratio Lower Upper  
79 100  
108 74.9 58.7 88.1  
77 60.7 50.2 75.2

Manual Integrations  
APPROVED



Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#16  
2,2'-oxybis(1-Chloropropane)  
Concen: 38.630 ng  
RT: 8.255 min Scan# 904  
Delta R.T. -0.023 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04



Tgt Ion: 45 Resp: 691013  
Ion Ratio Lower Upper  
45 100  
77 13.8 0.0 33.1  
79 11.4 0.0 30.0





#17  
2-Methylphenol  
Concen: 44.291 ng  
RT: 8.196 min Scan# 895  
Delta R.T. -0.006 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Instrument :

BNA\_P

ClientSampleId :





EG1-TP1MSD

Tgt Ion:107 Resp: 406552  
Ion Ratio Lower Upper  
107 100  
108 112.3 89.6 134.4  
77 46.6 39.3 58.9  
79 46.6 37.8 56.6

Manual Integrations




APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#18  
Hexachloroethane  
Concen: 42.268 ng  
RT: 8.796 min Scan# 996  
Delta R.T. -0.029 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Tgt Ion:117 Resp: 199468  
Ion Ratio Lower Upper  
117 100  
119 94.2 77.1 115.7  
201 104.5 85.6 128.4





#19

n-Nitroso-di-n-propylamine

Concen: 41.432 ng

RT: 8.531 min Scan# 91

Delta R.T. -0.023 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04

Instrument :

BNA\_P

ClientSampleId :

EG1-TP1MSD

Tgt Ion: 70 Resp: 37312

Ion Ratio Lower Upper

70 100

42 55.7 49.0 73.4

101 10.6 7.9 11.9





130 20.5 16.6 25.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/07/2025

Supervised By :Jagrut Upadhyay 10/07/2025



#20

3+4-Methylphenols

Concen: 42.618 ng

RT: 8.519 min Scan# 949

Delta R.T. -0.006 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04

Tgt Ion:107 Resp: 547868




Ion Ratio Lower Upper


107 100

108 90.2 67.8 107.8

77 34.7 16.1 56.1

79 29.4 10.5 50.5





#21  
Naphthalene-d8  
Concen: 20.000 ng  
RT: 10.496 min Scan# 1290  
Delta R.T. -0.029 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Instrument :

BNA\_P

ClientSampleId :

EG1-TP1MSD





Tgt Ion:136 Resp: 597901  
Ion Ratio Lower Upper  
136 100  
137 10.8 8.7 13.1  
54 9.4 7.5 11.3  
68 5.7 5.0 7.6

Manual Integrations

APPROVED




Reviewed By :Rahul Chavli 10/07/2025


Supervised By :Jagrut Upadhyay 10/07/2025



#22  
Acetophenone  
Concen: 51.071 ng  
RT: 8.555 min Scan# 955  
Delta R.T. -0.018 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04


Tgt Ion:105 Resp: 816012  
Ion Ratio Lower Upper  
105 100  
71 1.0 1.8 2.8#  
51 27.6 23.5 35.3  
120 21.9 17.7 26.5





#23  
Nitrobenzene-d5  
Concen: 76.233 ng  
RT: 8.884 min Scan# 1011  
Delta R.T. -0.018 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD



Tgt Ion: 82 Resp: 1033296  
Ion Ratio Lower Upper  
82 100  
128 42.6 32.9 49.3  
54 49.8 43.4 65.2




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#24  
Nitrobenzene  
Concen: 45.119 ng  
RT: 8.925 min Scan# 1018  
Delta R.T. -0.023 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04


Tgt Ion: 77 Resp: 548632  
Ion Ratio Lower Upper  
77 100  
123 44.6 34.4 51.6  
65 14.0 10.8 16.2





#25  
Isophorone  
Concen: 44.417 ng  
RT: 9.449 min Scan# 1107  
Delta R.T. -0.018 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD




Tgt Ion: 82 Resp: 1058219  
Ion Ratio Lower Upper  
82 100  
95 7.6 5.8 8.8  
138 17.3 13.5 20.3

Manual Integrations  
APPROVED



Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025




#26  
2-Nitrophenol  
Concen: 49.793 ng  
RT: 9.631 min Scan# 1138  
Delta R.T. -0.018 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04



Tgt Ion:139 Resp: 268817  
Ion Ratio Lower Upper  
139 100  
109 32.1 29.0 43.4  
65 49.5 46.9 70.3





#27

2,4-Dimethylphenol

Concen: 48.365 ng

RT: 9.696 min Scan# 1149

Delta R.T. -0.018 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04

Instrument :

BNA\_P

ClientSampleId :

EG1-TP1MSD





|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 122   | Resp: | 458076 |
| Ion Ratio | Lower | Upper |        |
| 122       | 100   |       |        |
| 107       | 119.9 | 97.9  | 146.9  |
| 121       | 58.2  | 47.4  | 71.0   |

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/07/2025

Supervised By :Jagrut Upadhyay 10/07/2025



#28

bis(2-Chloroethoxy)methane

Concen: 46.473 ng




RT: 9.913 min Scan# 1186


Delta R.T. -0.029 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04


|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 93    | Resp: | 640007 |
| Ion Ratio | Lower | Upper |        |
| 93        | 100   |       |        |
| 95        | 32.9  | 25.6  | 38.4   |
| 123       | 11.6  | 9.2   | 13.8   |





#29  
2,4-Dichlorophenol  
Concen: 49.658 ng  
RT: 10.190 min Scan# 1232  
Delta R.T. 0.006 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD



Tgt Ion:162 Resp: 468930  
Ion Ratio Lower Upper  
162 100  
164 65.9 44.7 84.7  
98 36.5 18.1 58.1




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#30  
1,2,4-Trichlorobenzene  
Concen: 47.081 ng  
RT: 10.366 min Scan# 1263  
Delta R.T. -0.023 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04


Tgt Ion:180 Resp: 501060  
Ion Ratio Lower Upper  
180 100  
182 96.8 76.0 114.0  
145 30.0 24.7 37.1





#31  
Naphthalene  
Concen: 45.417 ng  
RT: 10.549 min Scan# 1299  
Delta R.T. -0.029 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD



Tgt Ion:128 Resp: 1464195  
Ion Ratio Lower Upper  
128 100  
129 11.0 8.6 12.8  
127 13.0 10.2 15.4




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#32  
Benzoic acid  
Concen: 46.359 ng  
RT: 9.890 min Scan# 1182  
Delta R.T. 0.012 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04


Tgt Ion:122 Resp: 326440  
Ion Ratio Lower Upper  
122 100  
105 124.9 111.5 151.5  
77 97.1 81.8 121.8





#33  
4-Chloroaniline  
Concen: 13.598 ng  
RT: 10.678 min Scan# 1316  
Delta R.T. -0.012 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD



|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 127   | Resp: | 182724 |
| Ion      | Ratio | Lower | Upper  |
| 127      | 100   |       |        |
| 129      | 29.7  | 25.1  | 37.7   |
| 65       | 32.2  | 26.4  | 39.6   |
| 92       | 20.5  | 15.9  | 23.9   |




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#34  
Hexachlorobutadiene  
Concen: 47.480 ng  
RT: 10.831 min Scan# 1342  
Delta R.T. -0.029 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 225   | Resp: | 325289 |
| Ion      | Ratio | Lower | Upper  |
| 225      | 100   |       |        |
| 223      | 62.2  | 50.1  | 75.1   |
| 227      | 64.2  | 51.2  | 76.8   |





#35  
Caprolactam  
Concen: 37.608 ng  
RT: 11.484 min Scan# 1453  
Delta R.T. 0.000 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD



Tgt Ion:113 Resp: 136901  
Ion Ratio Lower Upper  
113 100  
55 180.7 165.5 205.5  
56 133.4 121.1 161.1




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#36  
4-Chloro-3-methylphenol  
Concen: 47.464 ng  
RT: 11.813 min Scan# 1509  
Delta R.T. -0.006 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Tgt Ion:107 Resp: 534367  
Ion Ratio Lower Upper  
107 100  
144 25.0 19.4 29.2  
142 74.7 63.8 95.6









#37  
2-Methylnaphthalene  
Concen: 45.827 ng  
RT: 12.154 min Scan# 1567  
Delta R.T. -0.029 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD

Tgt Ion:142 Resp: 948924  
Ion Ratio Lower Upper  
142 100  
141 86.0 69.6 104.4  
115 35.7 30.0 45.0




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#38  
1-Methylnaphthalene  
Concen: 44.883 ng  
RT: 12.378 min Scan# 1605  
Delta R.T. -0.023 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Tgt Ion:142 Resp: 991064  
Ion Ratio Lower Upper  
142 100  
141 91.1 72.7 109.1  
116 3.9 3.5 5.3






#39  
Acenaphthene-d10  
Concen: 20.000 ng  
RT: 14.343 min Scan# 1939  
Delta R.T. -0.035 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Instrument :

BNA\_P

ClientSampleId :

EG1-TP1MSD






Tgt Ion:164 Resp: 371539  
Ion Ratio Lower Upper  
164 100  
162 98.8 80.6 120.8  
160 43.3 35.0 52.6

Manual Integrations

APPROVED




Reviewed By :Rahul Chavli 10/07/2025


Supervised By :Jagrut Upadhyay 10/07/2025



#40  
1,2,4,5-Tetrachlorobenzene  
Concen: 57.411 ng  
RT: 12.531 min Scan# 1631  
Delta R.T. -0.023 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04


Tgt Ion:216 Resp: 646366  
Ion Ratio Lower Upper  
216 100  
214 78.0 62.4 93.6  
179 20.8 17.0 25.6  
108 17.7 15.2 22.8





#41  
Hexachlorocyclopentadiene  
Concen: 82.630 ng  
RT: 12.501 min Scan# 1626  
Delta R.T. -0.035 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD



Tgt Ion: 237 Resp: 506678  
Ion Ratio Lower Upper  
237 100  
235 62.8 41.9 81.9  
272 12.3 0.0 31.8




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#42  
2,4,6-Tribromophenol  
Concen: 149.387 ng  
RT: 15.872 min Scan# 2199  
Delta R.T. -0.035 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Tgt Ion: 330 Resp: 692292  
Ion Ratio Lower Upper  
330 100  
332 96.6 76.9 115.3  
141 32.4 31.0 46.4





#43

2,4,6-Trichlorophenol

Concen: 53.456 ng

RT: 12.784 min Scan# 1674

Delta R.T. -0.023 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04

Instrument :

BNA\_P

ClientSampleId :

EG1-TP1MSD





Tgt Ion:196 Resp: 39564  
 Ion Ratio Lower Upper  
 196 100  
 198 95.2 77.2 115.8  
 200 31.3 24.2 36.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/07/2025

Supervised By :Jagrut Upadhyay 10/07/2025



#44

2,4,5-Trichlorophenol

Concen: 51.447 ng




RT: 12.878 min Scan# 1690


Delta R.T. -0.006 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04

Tgt Ion:196 Resp: 423781  
 Ion Ratio Lower Upper  
 196 100  
 198 100.6 77.8 116.8  
 97 47.6 43.3 64.9  
 132 27.4 22.3 33.5





#45  
2-Fluorobiphenyl  
Concen: 82.843 ng  
RT: 12.954 min Scan# 1703  
Delta R.T. -0.035 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Instrument :

BNA\_P

ClientSampleId :





EG1-TP1MSD

Tgt Ion:172 Resp: 2311618  
Ion Ratio Lower Upper  
172 100  
171 35.2 27.8 41.6  
170 22.9 18.6 27.8

Manual Integrations




APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#46  
1,1'-Biphenyl  
Concen: 54.896 ng  
RT: 13.172 min Scan# 1740  
Delta R.T. -0.029 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Tgt Ion:154 Resp: 1497069  
Ion Ratio Lower Upper  
154 100  
153 39.9 19.7 59.7  
76 12.7 0.0 33.3









#47  
2-Chloronaphthalene  
Concen: 49.194 ng  
RT: 13.219 min Scan# 1753  
Delta R.T. -0.029 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD

Tgt Ion:162 Resp: 1056352  
Ion Ratio Lower Upper  
162 100  
127 39.0 32.2 48.4  
164 32.6 26.5 39.7




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#48  
2-Nitroaniline  
Concen: 50.165 ng  
RT: 13.431 min Scan# 1784  
Delta R.T. -0.023 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04


Tgt Ion: 65 Resp: 359079  
Ion Ratio Lower Upper  
65 100  
92 69.2 53.1 79.7  
138 101.8 77.8 116.8





#49  
Acenaphthylene  
Concen: 50.595 ng  
RT: 14.066 min Scan# 1892  
Delta R.T. -0.035 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD




Tgt Ion:152 Resp: 1800125  
Ion Ratio Lower Upper  
152 100  
151 19.6 15.8 23.6  
153 13.0 10.5 15.7

Manual Integrations  
APPROVED



Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025




#50  
Dimethylphthalate  
Concen: 49.780 ng  
RT: 13.801 min Scan# 1847  
Delta R.T. -0.035 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Tgt Ion:163 Resp: 1412063  
Ion Ratio Lower Upper  
163 100  
194 4.1 3.4 5.2  
164 10.0 8.2 12.2





#51  
2,6-Dinitrotoluene  
Concen: 50.825 ng  
RT: 13.925 min Scan# 1873  
Delta R.T. -0.029 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD




Tgt Ion:165 Resp: 305445  
Ion Ratio Lower Upper  
165 100  
63 53.8 48.1 72.1  
89 53.0 45.4 68.0

Manual Integrations  
APPROVED



Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025




#52  
Acenaphthene  
Concen: 47.982 ng  
RT: 14.407 min Scan# 1950  
Delta R.T. -0.047 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Tgt Ion:154 Resp: 984501  
Ion Ratio Lower Upper  
154 100  
153 112.1 89.4 134.0  
152 52.3 42.5 63.7





#53  
3-Nitroaniline  
Concen: 32.658 ng  
RT: 14.272 min Scan# 1930  
Delta R.T. -0.018 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD



Tgt Ion:138 Resp: 206392  
Ion Ratio Lower Upper  
138 100  
108 11.4 8.7 13.1  
92 115.0 94.6 142.0




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#54  
2,4-Dinitrophenol  
Concen: 100.045 ng  
RT: 14.490 min Scan# 1964  
Delta R.T. -0.017 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04


Tgt Ion:184 Resp: 395942  
Ion Ratio Lower Upper  
184 100  
63 66.0 52.5 78.7  
154 66.7 51.5 77.3





#55  
Dibenzofuran  
Concen: 48.775 ng  
RT: 14.754 min Scan# 2015  
Delta R.T. -0.035 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD



Tgt Ion:168 Resp: 1631056  
Ion Ratio Lower Upper  
168 100  
139 38.3 31.7 47.5  
169 13.3 10.8 16.2




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#56  
4-Nitrophenol  
Concen: 72.844 ng  
RT: 14.637 min Scan# 1989  
Delta R.T. 0.018 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04


Tgt Ion:139 Resp: 394130  
Ion Ratio Lower Upper  
139 100  
109 75.3 62.9 102.9  
65 106.3 83.4 123.4





#57  
2,4-Dinitrotoluene  
Concen: 51.534 ng  
RT: 14.731 min Scan# 2009  
Delta R.T. -0.023 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD



Tgt Ion:165 Resp: 440756  
Ion Ratio Lower Upper  
165 100  
63 46.4 37.7 56.5  
89 71.5 61.0 91.6  
182 3.1 2.4 3.6




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#58  
Fluorene  
Concen: 49.200 ng  
RT: 15.413 min Scan# 2121  
Delta R.T. -0.041 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04


Tgt Ion:166 Resp: 1308291  
Ion Ratio Lower Upper  
166 100  
165 96.5 77.7 116.5  
167 13.3 10.6 16.0





#59  
2,3,4,6-Tetrachlorophenol  
Concen: 50.778 ng  
RT: 14.990 min Scan# 2055  
Delta R.T. -0.035 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD



Tgt Ion:232 Resp: 376785  
Ion Ratio Lower Upper  
232 100  
131 42.3 36.7 55.1  
130 2.2 2.0 3.0  
166 30.8 25.4 38.2




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#60  
Diethylphthalate  
Concen: 49.488 ng  
RT: 15.184 min Scan# 2082  
Delta R.T. -0.035 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04


Tgt Ion:149 Resp: 1423905  
Ion Ratio Lower Upper  
149 100  
177 23.3 17.7 26.5  
150 12.2 10.0 15.0





#61  
4-Chlorophenyl-phenylether  
Concen: 50.202 ng  
RT: 15.401 min Scan# 2119  
Delta R.T. -0.041 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD




Tgt Ion:204 Resp: 672776  
Ion Ratio Lower Upper  
204 100  
206 33.3 26.0 39.0  
141 59.1 49.7 74.5

Manual Integrations  
APPROVED



Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025




#62  
4-Nitroaniline  
Concen: 46.484 ng  
RT: 15.454 min Scan# 2128  
Delta R.T. -0.023 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Tgt Ion:138 Resp: 300951  
Ion Ratio Lower Upper  
138 100  
92 53.9 33.3 73.3  
108 82.0 66.7 106.7





#63  
Azobenzene  
Concen: 47.627 ng  
RT: 15.701 min Scan# 2176  
Delta R.T. -0.035 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD



Tgt Ion: 77 Resp: 1325034  
Ion Ratio Lower Upper  
77 100  
182 26.9 5.2 45.2  
105 21.6 1.1 41.1  
51 28.3 11.6 51.6




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#64  
Phenanthrene-d10  
Concen: 20.000 ng  
RT: 17.154 min Scan# 2417  
Delta R.T. -0.029 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04


Tgt Ion:188 Resp: 743213  
Ion Ratio Lower Upper  
188 100  
94 8.6 7.4 11.0  
80 9.5 7.7 11.5





#65  
4,6-Dinitro-2-methylphenol  
Concen: 54.572 ng  
RT: 15.519 min Scan# 2115  
Delta R.T. -0.018 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD



Tgt Ion:198 Resp: 269731  
Ion Ratio Lower Upper  
198 100  
51 44.4 33.4 73.4  
105 44.2 24.3 64.3




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#66  
n-Nitrosodiphenylamine  
Concen: 51.691 ng  
RT: 15.619 min Scan# 2156  
Delta R.T. -0.035 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04


Tgt Ion:169 Resp: 1175664  
Ion Ratio Lower Upper  
169 100  
168 66.8 52.6 78.8  
167 35.0 28.5 42.7





#67  
4-Bromophenyl-phenylether  
Concen: 53.947 ng  
RT: 16.313 min Scan# 21  
Delta R.T. -0.041 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD



Tgt Ion:248 Resp: 440472  
Ion Ratio Lower Upper  
248 100  
250 96.8 77.5 116.3  
141 71.9 60.2 90.4




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#68  
Hexachlorobenzene  
Concen: 54.787 ng  
RT: 16.437 min Scan# 2295  
Delta R.T. -0.035 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Tgt Ion:284 Resp: 498850  
Ion Ratio Lower Upper  
284 100  
142 32.6 31.3 46.9  
249 31.6 27.3 40.9









#69  
Atrazine  
Concen: 48.199 ng  
RT: 16.607 min Scan# 2319  
Delta R.T. -0.029 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD

Tgt Ion:200 Resp: 419034  
Ion Ratio Lower Upper  
200 100  
173 24.9 4.2 44.2  
215 50.2 27.3 67.3




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#70  
Pentachlorophenol  
Concen: 103.489 ng  
RT: 16.807 min Scan# 2358  
Delta R.T. -0.023 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Tgt Ion:266 Resp: 621905  
Ion Ratio Lower Upper  
266 100  
268 63.8 49.8 74.8  
264 63.8 50.2 75.2





#71

Phenanthrene

Concen: 50.153 ng

RT: 17.195 min Scan# 2429

Delta R.T. -0.029 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04

Instrument :

BNA\_P

ClientSampleId :

EG1-TP1MSD

Tgt Ion:178 Resp: 2114400

Ion Ratio Lower Upper

178 100

176 18.8 15.4 23.2





179 15.2 12.2 18.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/07/2025

Supervised By :Jagrut Upadhyay 10/07/2025



#72

Anthracene

Concen: 50.662 ng

RT: 17.289 min Scan# 2440

Delta R.T. -0.029 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04




Tgt Ion:178 Resp: 2167142


Ion Ratio Lower Upper

178 100

176 18.1 14.8 22.2

179 15.4 12.2 18.4





#73

Carbazole

Concen: 51.396 ng

RT: 17.572 min Scan# 24

Delta R.T. -0.029 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04

Instrument :

BNA\_P

ClientSampleId :

EG1-TP1MSD

Tgt Ion:167 Resp: 2043964

Ion Ratio Lower Upper

167 100

166 21.4 17.0 25.4





139 13.1 10.6 16.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/07/2025

Supervised By :Jagrut Upadhyay 10/07/2025



#74

Di-n-butylphthalate

Concen: 53.321 ng

RT: 18.148 min Scan# 2586

Delta R.T. -0.035 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04




Tgt Ion:149 Resp: 2602269


Ion Ratio Lower Upper

149 100

150 9.2 7.4 11.0


104 4.8 4.3 6.5





#75  
Fluoranthene  
Concen: 50.885 ng  
RT: 19.283 min Scan# 21  
Delta R.T. -0.029 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD



Tgt Ion:202 Resp: 2572890  
Ion Ratio Lower Upper  
202 100  
101 10.4 0.0 30.4  
203 17.6 0.0 37.4




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#76  
Chrysene-d12  
Concen: 20.000 ng  
RT: 21.601 min Scan# 3173  
Delta R.T. -0.041 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Tgt Ion:240 Resp: 809368  
Ion Ratio Lower Upper  
240 100  
120 10.1 8.0 12.0  
236 25.2 20.7 31.1





#77

Benzidine

Concen: 35.550 ng

RT: 19.489 min Scan# 2814

Delta R.T. -0.018 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04

Instrument :

BNA\_P

ClientSampleId :

EG1-TP1MSD

Tgt Ion:184 Resp: 828816

Ion Ratio Lower Upper

184 100

185 13.9 11.5 17.3





183 11.1 9.2 13.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/07/2025

Supervised By :Jagrut Upadhyay 10/07/2025



#78

Pyrene

Concen: 50.049 ng

RT: 19.660 min Scan# 2843

Delta R.T. -0.029 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04




Tgt Ion:202 Resp: 2702893


Ion Ratio Lower Upper

202 100

200 20.8 17.0 25.6

203 17.5 13.9 20.9





#79  
Terphenyl-d14  
Concen: 84.164 ng  
RT: 19.872 min Scan# 2886  
Delta R.T. -0.041 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Instrument :

BNA\_P

ClientSampleId :





EG1-TP1MSD

Tgt Ion:244 Resp: 378655  
Ion Ratio Lower Upper  
244 100  
212 7.1 5.9 8.9  
122 9.3 7.1 10.7

Manual Integrations




APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#80  
Butylbenzylphthalate  
Concen: 51.621 ng  
RT: 20.601 min Scan# 3003  
Delta R.T. -0.041 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Tgt Ion:149 Resp: 1222347  
Ion Ratio Lower Upper  
149 100  
91 70.7 56.6 85.0  
206 22.0 17.4 26.0





#81

Benzo(a)anthracene

Concen: 49.710 ng

RT: 21.583 min Scan# 31

Delta R.T. -0.035 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04

Instrument :

BNA\_P

ClientSampleId :

EG1-TP1MSD

Tgt Ion:228 Resp: 2754458

Ion Ratio Lower Upper

228 100

226 26.7 21.6 32.4





229 19.6 15.7 23.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/07/2025

Supervised By :Jagrut Upadhyay 10/07/2025



#82

3,3'-Dichlorobenzidine

Concen: 41.274 ng

RT: 21.501 min Scan# 3156

Delta R.T. -0.035 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04




Tgt Ion:252 Resp: 790776


Ion Ratio Lower Upper

252 100

254 63.4 51.4 77.0

126 9.9 8.3 12.5





#83

Chrysene

Concen: 49.143 ng

RT: 21.648 min Scan# 31

Delta R.T. -0.041 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04

Instrument :

BNA\_P

ClientSampleId :

EG1-TP1MSD

Tgt Ion:228 Resp: 2595782

Ion Ratio Lower Upper

228 100

226 29.4 23.8 35.6





229 19.6 15.6 23.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/07/2025

Supervised By :Jagrut Upadhyay 10/07/2025



#84

Bis(2-ethylhexyl)phthalate

Concen: 54.017 ng

RT: 21.501 min Scan# 3156

Delta R.T. -0.047 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04




Tgt Ion:149 Resp: 1871340


Ion Ratio Lower Upper

149 100

167 28.5 22.9 34.3

279 4.8 3.8 5.8





#85

Di-n-octyl phthalate

Concen: 53.340 ng

RT: 22.766 min Scan# 3382

Delta R.T. -0.065 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04

Instrument :

BNA\_P

ClientSampleId :

EG1-TP1MSD

Tgt Ion:149 Resp: 3287060

Ion Ratio Lower Upper

149 100

167 1.7 1.3 1.9





43 8.9 8.0 12.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/07/2025

Supervised By :Jagrut Upadhyay 10/07/2025



#86

Perylene-d12

Concen: 20.000 ng

RT: 24.948 min Scan# 3742

Delta R.T. -0.053 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04




Tgt Ion:264 Resp: 926256


Ion Ratio Lower Upper

264 100

260 23.2 18.3 27.5

265 21.4 17.3 25.9





#87

Indeno(1,2,3-cd)pyrene

Concen: 52.971 ng m

RT: 28.759 min Scan# 4390

Delta R.T. -0.076 min

Lab File: BP025929.D


Acq: 06 Oct 2025 19:04

Instrument :

BNA\_P

ClientSampleId :

EG1-TP1MSD



Tgt Ion:276 Resp: 356805

Ion Ratio Lower Upper

276 100

138 9.6 13.4 20.2




277 12.4 9.1 13.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/07/2025

Supervised By :Jagrut Upadhyay 10/07/2025



#88

Benzo(b)fluoranthene

Concen: 49.350 ng

RT: 23.889 min Scan# 3562

Delta R.T. -0.041 min

Lab File: BP025929.D

Acq: 06 Oct 2025 19:04




Tgt Ion:252 Resp: 2795296


Ion Ratio Lower Upper

252 100

253 22.3 17.9 26.9

125 8.6 7.3 10.9









#89  
Benzo(k)fluoranthene  
Concen: 50.981 ng  
RT: 23.960 min Scan# 31  
Delta R.T. -0.041 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD

Tgt Ion:252 Resp: 294669  
Ion Ratio Lower Upper  
252 100  
253 21.2 17.5 26.3  
125 8.8 7.5 11.3




Manual Integrations  
APPROVED


Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#90  
Benzo(a)pyrene  
Concen: 49.604 ng  
RT: 24.783 min Scan# 3714  
Delta R.T. -0.059 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04


Tgt Ion:252 Resp: 2751550  
Ion Ratio Lower Upper  
252 100  
253 22.0 17.2 25.8  
125 10.8 8.3 12.5





#91  
Dibenzo(a,h)anthracene  
Concen: 53.506 ng  
RT: 28.818 min Scan# 4417  
Delta R.T. -0.100 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04




Instrument :  
BNA\_P  
ClientSampleId :  
EG1-TP1MSD



Tgt Ion:278 Resp: 2949384  
Ion Ratio Lower Upper  
278 100  
139 13.8 12.1 18.1  
279 23.1 19.6 29.4

Manual Integrations  
APPROVED

Reviewed By :Rahul Chavli 10/07/2025  
Supervised By :Jagrut Upadhyay 10/07/2025



#92  
Benzo(g,h,i)perylene  
Concen: 52.804 ng  
RT: 29.936 min Scan# 4590  
Delta R.T. -0.076 min  
Lab File: BP025929.D  
Acq: 06 Oct 2025 19:04

Tgt Ion:276 Resp: 2862470  
Ion Ratio Lower Upper  
276 100  
277 24.4 18.7 28.1  
138 19.6 16.0 24.0

