

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100820\
 Data File : BP003572.D
 Acq On : 08 Oct 2020 14:45
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 08 16:03:46 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 08 15:55:58 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	62	0.00
2	1,4-Dioxane	0.471	0.411	12.7	58	0.00
3	Pyridine	1.253	1.135	9.4	58	0.00
4	n-Nitrosodimethylamine	0.375	0.405	-8.0	69	0.00
5 S	2-Fluorophenol	1.145	1.064	7.1	60	0.00
6	Aniline	1.753	1.702	2.9	62	0.00
7 S	Phenol-d6	1.527	1.498	1.9	62	0.00
8	2-Chlorophenol	1.276	1.209	5.3	61	0.00
9	Benzaldehyde	0.851	0.764	10.2	59	0.00
10 C	Phenol	1.458	1.427	2.1	62	0.00
11	bis(2-Chloroethyl)ether	1.156	1.119	3.2	62	0.00
12	1,3-Dichlorobenzene	1.525	1.475	3.3	62	0.00
13 C	1,4-Dichlorobenzene	1.543	1.508	2.3	63	0.00
14	1,2-Dichlorobenzene	1.481	1.424	3.8	62	0.00
15	Benzyl Alcohol	1.015	1.249	-23.1	76	0.00
16	2,2'-oxybis(1-Chloropropane	0.840	0.886	-5.5	68	0.00
17	2-Methylphenol	1.054	1.080	-2.5	66	0.00
18	Hexachloroethane	0.575	0.587	-2.1	66	0.00
19 P	n-Nitroso-di-n-propylamine	0.818	0.905	-10.6	70	0.00
20	3+4-Methylphenols	1.403	1.398	0.4	62	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	62	0.00
22	Acetophenone	0.491	0.486	1.0	64	0.00
23 S	Nitrobenzene-d5	0.340	0.382	-12.4	73	0.00
24	Nitrobenzene	0.341	0.386	-13.2	74	0.00
25	Isophorone	0.624	0.671	-7.5	70	0.00
26 C	2-Nitrophenol	0.186	0.187	-0.5	65	0.00
27	2,4-Dimethylphenol	0.263	0.255	3.0	63	0.00
28	bis(2-Chloroethoxy)methane	0.415	0.429	-3.4	68	0.00
29 C	2,4-Dichlorophenol	0.329	0.317	3.6	63	0.00
30	1,2,4-Trichlorobenzene	0.390	0.367	5.9	62	0.00
31	Naphthalene	1.056	1.041	1.4	65	0.00
32	Benzoic acid	0.159	0.081	49.1#	32#	0.00
33	4-Chloroaniline	0.449	0.441	1.8	64	0.00
34 C	Hexachlorobutadiene	0.266	0.252	5.3	62	0.00
35	Caprolactam	0.110	0.107	2.7	63	0.00
36 C	4-Chloro-3-methylphenol	0.354	0.356	-0.6	65	0.00
37	2-Methylnaphthalene	0.765	0.736	3.8	64	0.00
38	1-Methylnaphthalene	0.727	0.699	3.9	64	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	59	0.00
40	1,2,4,5-Tetrachlorobenzene	0.655	0.642	2.0	61	0.00
41 P	Hexachlorocyclopentadiene	0.161	0.170	-5.6	64	0.00
42 S	2,4,6-Tribromophenol	0.305	0.300	1.6	60	0.00
43 C	2,4,6-Trichlorophenol	0.423	0.409	3.3	60	0.00
44	2,4,5-Trichlorophenol	0.436	0.418	4.1	59	0.00
45 S	2-Fluorobiphenyl	1.367	1.367	0.0	64	0.00
46	1,1'-Biphenyl	1.492	1.503	-0.7	63	0.00
47	2-Chloronaphthalene	1.235	1.225	0.8	62	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100820\
 Data File : BP003572.D
 Acq On : 08 Oct 2020 14:45
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 08 16:03:46 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 08 15:55:58 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.313	0.388	-24.0	75	0.00
49	Acenaphthylene	1.887	1.868	1.0	62	0.00
50	Dimethylphthalate	1.597	1.583	0.9	62	0.00
51	2,6-Dinitrotoluene	0.342	0.340	0.6	62	0.00
52 C	Acenaphthene	1.149	1.138	1.0	62	0.00
53	3-Nitroaniline	0.371	0.379	-2.2	62	0.00
54 P	2,4-Dinitrophenol	0.127	0.164	-29.1#	72	0.00
55	Dibenzofuran	1.833	1.793	2.2	62	0.00
56 P	4-Nitrophenol	0.227	0.265	-16.7	69	0.00
57	2,4-Dinitrotoluene	0.475	0.477	-0.4	61	0.00
58	Fluorene	1.448	1.436	0.8	63	0.00
59	2,3,4,6-Tetrachlorophenol	0.409	0.386	5.6	58	0.00
60	Diethylphthalate	1.629	1.621	0.5	63	0.00
61	4-Chlorophenyl-phenylether	0.790	0.777	1.6	63	0.00
62	4-Nitroaniline	0.394	0.386	2.0	58	0.00
63	Azobenzene	1.242	1.439	-15.9	72	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	58	0.00
65	4,6-Dinitro-2-methylphenol	0.106	0.099	6.6	54	0.00
66 c	n-Nitrosodiphenylamine	0.587	0.588	-0.2	62	0.00
67	4-Bromophenyl-phenylether	0.242	0.240	0.8	61	0.00
68	Hexachlorobenzene	0.287	0.283	1.4	61	0.00
69	Atrazine	0.230	0.226	1.7	60	0.00
70 C	Pentachlorophenol	0.104	0.146	-40.4#	83	0.00
71	Phenanthrene	1.088	1.095	-0.6	62	0.00
72	Anthracene	1.072	1.079	-0.7	61	0.00
73	Carbazole	1.019	1.031	-1.2	61	0.00
74	Di-n-butylphthalate	1.302	1.343	-3.1	63	0.00
75 C	Fluoranthene	1.368	1.368	0.0	61	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	58	0.00
77	Benzydine	0.620	0.325	47.6#	31#	0.00
78	Pyrene	1.247	1.245	0.2	61	0.00
79 S	Terphenyl-d14	0.960	0.949	1.1	62	0.00
80	Butylbenzylphthalate	0.571	0.579	-1.4	62	0.00
81	Benzo(a)anthracene	1.293	1.271	1.7	61	0.00
82	3,3'-Dichlorobenzidine	0.500	0.479	4.2	59	0.00
83	Chrysene	1.242	1.232	0.8	62	0.00
84	Bis(2-ethylhexyl)phthalate	0.803	0.775	3.5	61	0.00
85 c	Di-n-octyl phthalate	1.468	1.495	-1.8	63	0.00
86 I	Perylene-d12	1.000	1.000	0.0	60	0.00
87	Indeno(1,2,3-cd)pyrene	1.516	1.521	-0.3	64	0.00
88	Benzo(b)fluoranthene	1.263	1.275	-1.0	64	0.00
89	Benzo(k)fluoranthene	1.210	1.160	4.1	61	0.00
90 C	Benzo(a)pyrene	1.191	1.170	1.8	63	0.00
91	Dibenzo(a,h)anthracene	1.251	1.263	-1.0	64	0.00
92	Benzo(g,h,i)perylene	1.249	1.247	0.2	64	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100820\
Data File : BP003572.D
Acq On : 08 Oct 2020 14:45
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Oct 08 16:03:46 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 08 15:55:58 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range SPCC's out = 0 CCC's out = 1