

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100820\
 Data File : BP003572.D
 Acq On : 08 Oct 2020 14:45
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 08 16:03:46 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 08 15:55:58 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	62	0.00
2	1,4-Dioxane	40.000	34.941	12.6	58	0.00
3	Pyridine	40.000	36.225	9.4	58	0.00
4	n-Nitrosodimethylamine	40.000	43.234	-8.1	69	0.00
5 S	2-Fluorophenol	80.000	74.284	7.1	60	0.00
6	Aniline	40.000	38.828	2.9	62	0.00
7 S	Phenol-d6	80.000	78.482	1.9	62	0.00
8	2-Chlorophenol	40.000	37.900	5.3	61	0.00
9	Benzaldehyde	40.000	35.876	10.3	59	0.00
10 C	Phenol	40.000	39.130	2.2	62	0.00
11	bis(2-Chloroethyl)ether	40.000	38.743	3.1	62	0.00
12	1,3-Dichlorobenzene	40.000	38.692	3.3	62	0.00
13 C	1,4-Dichlorobenzene	40.000	39.089	2.3	63	0.00
14	1,2-Dichlorobenzene	40.000	38.473	3.8	62	0.00
15	Benzyl Alcohol	40.000	49.255	-23.1	76	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	42.196	-5.5	68	0.00
17	2-Methylphenol	40.000	40.977	-2.4	66	0.00
18	Hexachloroethane	40.000	40.835	-2.1	66	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	44.243	-10.6	70	0.00
20	3+4-Methylphenols	40.000	39.845	0.4	62	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	62	0.00
22	Acetophenone	40.000	39.585	1.0	64	0.00
23 S	Nitrobenzene-d5	80.000	89.731	-12.2	73	0.00
24	Nitrobenzene	40.000	45.356	-13.4	74	0.00
25	Isophorone	40.000	43.000	-7.5	70	0.00
26 C	2-Nitrophenol	40.000	40.187	-0.5	65	0.00
27	2,4-Dimethylphenol	40.000	38.744	3.1	63	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.305	-3.3	68	0.00
29 C	2,4-Dichlorophenol	40.000	38.516	3.7	63	0.00
30	1,2,4-Trichlorobenzene	40.000	37.679	5.8	62	0.00
31	Naphthalene	40.000	39.408	1.5	65	0.00
32	Benzoic acid	40.000	23.163	42.1#	32	0.00
33	4-Chloroaniline	40.000	39.331	1.7	64	0.00
34 C	Hexachlorobutadiene	40.000	37.790	5.5	62	0.00
35	Caprolactam	40.000	38.990	2.5	63	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.165	-0.4	65	0.00
37	2-Methylnaphthalene	40.000	38.466	3.8	64	0.00
38	1-Methylnaphthalene	40.000	38.481	3.8	64	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	59	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.220	2.0	61	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.427	-1.1	64	0.00
42 S	2,4,6-Tribromophenol	80.000	78.636	1.7	60	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.623	3.4	60	0.00
44	2,4,5-Trichlorophenol	40.000	38.369	4.1	59	0.00
45 S	2-Fluorobiphenyl	80.000	79.974	0.0	64	0.00
46	1,1'-Biphenyl	40.000	40.292	-0.7	63	0.00
47	2-Chloronaphthalene	40.000	39.682	0.8	62	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100820\
 Data File : BP003572.D
 Acq On : 08 Oct 2020 14:45
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 08 16:03:46 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 08 15:55:58 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	49.682	-24.2	75	0.00
49	Acenaphthylene	40.000	39.592	1.0	62	0.00
50	Dimethylphthalate	40.000	39.641	0.9	62	0.00
51	2,6-Dinitrotoluene	40.000	39.757	0.6	62	0.00
52 C	Acenaphthene	40.000	39.631	0.9	62	0.00
53	3-Nitroaniline	40.000	40.912	-2.3	62	0.00
54 P	2,4-Dinitrophenol	40.000	44.044	-10.1	72	0.00
55	Dibenzofuran	40.000	39.116	2.2	62	0.00
56 P	4-Nitrophenol	40.000	46.705	-16.8	69	0.00
57	2,4-Dinitrotoluene	40.000	40.180	-0.4	61	0.00
58	Fluorene	40.000	39.661	0.8	63	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.742	5.6	58	0.00
60	Diethylphthalate	40.000	39.804	0.5	63	0.00
61	4-Chlorophenyl-phenylether	40.000	39.306	1.7	63	0.00
62	4-Nitroaniline	40.000	39.213	2.0	58	0.00
63	Azobenzene	40.000	46.347	-15.9	72	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	58	0.00
65	4,6-Dinitro-2-methylphenol	40.000	37.607	6.0	54	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.030	-0.1	62	0.00
67	4-Bromophenyl-phenylether	40.000	39.634	0.9	61	0.00
68	Hexachlorobenzene	40.000	39.408	1.5	61	0.00
69	Atrazine	40.000	39.204	2.0	60	0.00
70 C	Pentachlorophenol	40.000	50.174	-25.4#	83	0.00
71	Phenanthrene	40.000	40.265	-0.7	62	0.00
72	Anthracene	40.000	40.237	-0.6	61	0.00
73	Carbazole	40.000	40.483	-1.2	61	0.00
74	Di-n-butylphthalate	40.000	41.245	-3.1	63	0.00
75 C	Fluoranthene	40.000	39.982	0.0	61	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	58	0.00
77	Benzydine	40.000	20.980	47.5#	31	0.00
78	Pyrene	40.000	39.923	0.2	61	0.00
79 S	Terphenyl-d14	80.000	79.098	1.1	62	0.00
80	Butylbenzylphthalate	40.000	40.579	-1.4	62	0.00
81	Benzo(a)anthracene	40.000	39.314	1.7	61	0.00
82	3,3'-Dichlorobenzidine	40.000	38.329	4.2	59	0.00
83	Chrysene	40.000	39.662	0.8	62	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.573	3.6	61	0.00
85 c	Di-n-octyl phthalate	40.000	40.731	-1.8	63	0.00
86 I	Perylene-d12	20.000	20.000	0.0	60	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.135	-0.3	64	0.00
88	Benzo(b)fluoranthene	40.000	40.381	-1.0	64	0.00
89	Benzo(k)fluoranthene	40.000	38.328	4.2	61	0.00
90 C	Benzo(a)pyrene	40.000	39.301	1.7	63	0.00
91	Dibenzo(a,h)anthracene	40.000	40.377	-0.9	64	0.00
92	Benzo(g,h,i)perylene	40.000	39.935	0.2	64	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100820\
Data File : BP003572.D
Acq On : 08 Oct 2020 14:45
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Oct 08 16:03:46 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 08 15:55:58 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 1