

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100821\
 Data File : BP007369.D
 Acq On : 08 Oct 2021 10:48
 Operator : CG/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010

Quant Time: Oct 08 14:01:13 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP100821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 08 13:59:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.781	152	168148	20.000	ng	0.00
21) Naphthalene-d8	10.581	136	711045	20.000	ng	# 0.00
39) Acenaphthene-d10	14.428	164	492457	20.000	ng	0.00
64) Phenanthrene-d10	17.180	188	705327	20.000	ng	0.00
76) Chrysene-d12	21.274	240	799969	20.000	ng	0.00
86) Perylene-d12	23.639	264	831618	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.357	112	213727	21.277	ng	0.00
7) Phenol-d6	6.963	99	292932	21.337	ng	0.00
23) Nitrobenzene-d5	8.945	82	262930	21.535	ng	0.00
42) 2,4,6-Tribromophenol	15.922	330	136551	21.388	ng	0.00
45) 2-Fluorobiphenyl	13.051	172	735429	21.395	ng	0.00
79) Terphenyl-d14	19.745	244	918238	21.996	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.228	88	39957	9.836	ng	# 88
3) Pyridine	3.652	79	111231	10.007	ng	# 86
4) n-Nitrosodimethylamine	3.558	42	43742	11.479	ng	# 86
6) Aniline	7.110	93	162908	10.765	ng	99
8) 2-Chlorophenol	7.346	128	118731	10.876	ng	98
9) Benzaldehyde	6.928	77	89802	11.612	ng	96
10) Phenol	6.987	94	140262	10.597	ng	93
11) bis(2-Chloroethyl)ether	7.210	93	111562	10.678	ng	94
12) 1,3-Dichlorobenzene	7.669	146	132097	10.727	ng	97
13) 1,4-Dichlorobenzene	7.816	146	135113	10.762	ng	99
14) 1,2-Dichlorobenzene	8.134	146	133999	11.015	ng	96
15) Benzyl Alcohol	8.034	79	94661	10.765	ng	97
16) 2,2'-oxybis(1-Chloropr...	8.322	45	136686	11.425	ng	92
17) 2-Methylphenol	8.240	107	97444	10.920	ng	96
18) Hexachloroethane	8.857	117	46125	10.943	ng	96
19) n-Nitroso-di-n-propyla...	8.593	70	87210	11.679	ng	98
20) 3+4-Methylphenols	8.569	107	137347	11.132	ng	93
22) Acetophenone	8.610	105	185640	10.835	ng	# 97
24) Nitrobenzene	8.987	77	123655	10.622	ng	99
25) Isophorone	9.516	82	236957	11.130	ng	99
26) 2-Nitrophenol	9.704	139	62521	10.300	ng	# 91
27) 2,4-Dimethylphenol	9.769	122	100014	10.465	ng	95
28) bis(2-Chloroethoxy)met...	9.998	93	163483	10.884	ng	99
29) 2,4-Dichlorophenol	10.245	162	115999	10.404	ng	98
30) 1,2,4-Trichlorobenzene	10.445	180	132767	10.453	ng	98
31) Naphthalene	10.628	128	383347	10.564	ng	100
32) Benzoic acid	9.904	122	64699	8.808	ng	95
33) 4-Chloroaniline	10.757	127	159151	10.615	ng	99
34) Hexachlorobutadiene	10.910	225	81793	10.755	ng	97
35) Caprolactam	11.545	113	39150	11.414	ng	# 67
36) 4-Chloro-3-methylphenol	11.886	107	126030	11.121	ng	99
37) 2-Methylnaphthalene	12.245	142	278515	10.961	ng	96
38) 1-Methylnaphthalene	12.463	142	272700	10.984	ng	100
40) 1,2,4,5-Tetrachloroben...	12.616	216	154282	10.230	ng	99
41) Hexachlorocyclopentadiene	12.592	237	26338	3.151	ng	99
43) 2,4,6-Trichlorophenol	12.869	196	105051	10.209	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100821\
 Data File : BP007369.D
 Acq On : 08 Oct 2021 10:48
 Operator : CG/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010

Quant Time: Oct 08 14:01:13 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP100821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 08 13:59:37 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.945	196	113975	10.286	ng	98
46) 1,1'-Biphenyl	13.263	154	386269	10.545	ng	98
47) 2-Chloronaphthalene	13.304	162	298020	10.503	ng	96
48) 2-Nitroaniline	13.516	65	73740	10.659	ng	96
49) Acenaphthylene	14.151	152	466455	10.673	ng	100
50) Dimethylphthalate	13.892	163	392732	11.088	ng	99
51) 2,6-Dinitrotoluene	14.010	165	78083	10.792	ng	99
52) Acenaphthene	14.492	154	277936	10.497	ng	99
53) 3-Nitroaniline	14.345	138	81011	10.356	ng	100
54) 2,4-Dinitrophenol	14.569	184	32985	7.914	ng	94
55) Dibenzofuran	14.827	168	472862	10.704	ng	96
56) 4-Nitrophenol	14.675	139	57474	9.050	ng	# 88
57) 2,4-Dinitrotoluene	14.810	165	101451	10.626	ng	98
58) Fluorene	15.480	166	369605	10.830	ng	99
59) 2,3,4,6-Tetrachlorophenol	14.980	232	95259	9.790	ng	96
60) Diethylphthalate	15.257	149	385996	11.247	ng	98
61) 4-Chlorophenyl-phenyle...	15.475	204	198260	10.993	ng	92
62) 4-Nitroaniline	15.510	138	82184	10.446	ng	# 86
63) Azobenzene	15.769	77	328648	11.050	ng	95
65) 4,6-Dinitro-2-methylph...	15.575	198	60670	14.334	ng	93
66) n-Nitrosodiphenylamine	15.692	169	328893	15.672	ng	99
67) 4-Bromophenyl-phenylether	16.369	248	121304	15.692	ng	95
68) Hexachlorobenzene	16.486	284	134793	15.777	ng	97
69) Atrazine	16.651	200	116824	15.792	ng	97
70) Pentachlorophenol	16.839	266	67487	12.720	ng	97
71) Phenanthrene	17.221	178	394178	10.423	ng	99
72) Anthracene	17.316	178	385714	10.426	ng	100
73) Carbazole	17.592	167	373071	10.291	ng	98
74) Di-n-butylphthalate	18.145	149	426928	10.626	ng	# 98
75) Fluoranthene	19.198	202	491429	11.278	ng	98
77) Benzidine	19.386	184	202701	8.245	ng	100
78) Pyrene	19.551	202	507518	9.670	ng	99
80) Butylbenzylphthalate	20.427	149	201121	9.569	ng	91
81) Benzo(a)anthracene	21.262	228	534233	10.288	ng	100
82) 3,3'-Dichlorobenzidine	21.198	252	197220	11.059	ng	98
83) Chrysene	21.315	228	511746	10.312	ng	100
84) Bis(2-ethylhexyl)phtha...	21.186	149	295049	9.764	ng	98
85) Di-n-octyl phthalate	22.092	149	505824	9.832	ng	98
87) Indeno(1,2,3-cd)pyrene	26.103	276	612166	10.243	ng	98
88) Benzo(b)fluoranthene	22.915	252	525455	10.573	ng	99
89) Benzo(k)fluoranthene	22.962	252	520305	10.340	ng	99
90) Benzo(a)pyrene	23.527	252	434508	10.291	ng	99
91) Dibenzo(a,h)anthracene	26.109	278	507878	10.140	ng	99
92) Benzo(g,h,i)perylene	26.850	276	492901	10.085	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100821\
 Data File : BP007369.D
 Acq On : 08 Oct 2021 10:48
 Operator : CG/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010

Quant Time: Oct 08 14:01:13 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP100821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 08 13:59:37 2021
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.781 min Scan# 798
 Delta R.T. -0.000 min
 Lab File: BP007369.D
 Acq: 08 Oct 2021 10:48




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010


Tgt Ion:152 Resp: 168148
 Ion Ratio Lower Upper
 152 100
 150 152.8 125.6 188.4
 115 55.1 48.5 72.7



#2
 1,4-Dioxane
 Concen: 9.836 ng
 RT: 3.228 min Scan# 24
 Delta R.T. 0.006 min
 Lab File: BP007369.D
 Acq: 08 Oct 2021 10:48

Tgt Ion: 88 Resp: 39957
 Ion Ratio Lower Upper
 88 100
 58 70.8 46.7 70.1#
 43 25.2 20.3 30.5









#3
Pyridine
Concen: 10.007 ng
RT: 3.652 min Scan# 96
Delta R.T. 0.012 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion: 79 Resp: 111231
Ion Ratio Lower Upper
79 100
52 63.9 41.7 62.5#
51 35.2 24.1 36.1



#4
n-Nitrosodimethylamine
Concen: 11.479 ng
RT: 3.558 min Scan# 80
Delta R.T. 0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion: 42 Resp: 43742
Ion Ratio Lower Upper
42 100
74 147.8 105.4 158.0
44 9.6 4.2 6.4#









#5
2-Fluorophenol
Concen: 21.277 ng
RT: 5.357 min Scan# 386
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:112 Resp: 213727
Ion Ratio Lower Upper
112 100
64 52.2 41.4 62.2
63 27.9 23.8 35.8



#6
Aniline
Concen: 10.765 ng
RT: 7.110 min Scan# 684
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion: 93 Resp: 162908
Ion Ratio Lower Upper
93 100
66 36.5 29.7 44.5
65 19.3 15.3 22.9









#7
Phenol-d6
Concen: 21.337 ng
RT: 6.963 min Scan# 659
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	99	Resp:	292932
Ion	Ratio	Lower	Upper
99	100		
42	15.2	13.4	20.0
71	31.4	30.6	46.0



#8
2-Chlorophenol
Concen: 10.876 ng
RT: 7.346 min Scan# 724
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:	128	Resp:	118731
Ion	Ratio	Lower	Upper
128	100		
130	32.4	12.4	52.4
64	42.0	20.3	60.3









#9
Benzaldehyde
Concen: 11.612 ng
RT: 6.928 min Scan# 653
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion: 77 Resp: 89802
Ion Ratio Lower Upper
77 100
105 100.3 76.7 116.7
106 101.0 76.2 116.2



#10
Phenol
Concen: 10.597 ng
RT: 6.987 min Scan# 663
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion: 94 Resp: 140262
Ion Ratio Lower Upper
94 100
65 27.0 8.8 48.8
66 35.1 20.7 60.7









#11
bis(2-Chloroethyl)ether
Concen: 10.678 ng
RT: 7.210 min Scan# 701
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion: 93 Resp: 111562
Ion Ratio Lower Upper
93 100
63 72.0 46.2 86.2
95 33.3 12.2 52.2



#12
1,3-Dichlorobenzene
Concen: 10.727 ng
RT: 7.669 min Scan# 779
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:146 Resp: 132097
Ion Ratio Lower Upper
146 100
148 63.6 51.5 77.3
75 28.2 25.7 38.5









#13
1,4-Dichlorobenzene
Concen: 10.762 ng
RT: 7.816 min Scan# 804
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:146 Resp: 135113
Ion Ratio Lower Upper
146 100
148 64.3 51.3 76.9
111 39.6 33.0 49.6



#14
1,2-Dichlorobenzene
Concen: 11.015 ng
RT: 8.134 min Scan# 858
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:146 Resp: 133999
Ion Ratio Lower Upper
146 100
148 65.1 50.3 75.5
111 38.7 33.7 50.5









#15
Benzyl Alcohol
Concen: 10.765 ng
RT: 8.034 min Scan# 841
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	79	Resp:	94661
Ion	Ratio	Lower	Upper
79	100		
108	82.4	62.5	93.7
77	64.8	52.9	79.3



#16
2,2'-oxybis(1-Chloropropane)
Concen: 11.425 ng
RT: 8.322 min Scan# 890
Delta R.T. 0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:	45	Resp:	136686
Ion	Ratio	Lower	Upper
45	100		
77	17.5	2.0	42.0
79	13.6	0.0	36.7









#17
2-Methylphenol
Concen: 10.920 ng
RT: 8.240 min Scan# 876
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	107	Resp:	97444
Ion	Ratio	Lower	Upper
107	100		
108	109.7	87.7	131.5
77	43.7	40.2	60.2
79	43.6	39.0	58.4



#18
Hexachloroethane
Concen: 10.943 ng
RT: 8.857 min Scan# 981
Delta R.T. 0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:	117	Resp:	46125
Ion	Ratio	Lower	Upper
117	100		
119	93.9	76.2	114.2
201	111.0	83.4	125.0









#19
n-Nitroso-di-n-propylamine
Concen: 11.679 ng
RT: 8.593 min Scan# 936
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	70	Resp:	87210
Ion	Ratio	Lower	Upper
70	100		
42	53.1	40.8	61.2
101	10.5	8.8	13.2
130	23.5	18.9	28.3



#20
3+4-Methylphenols
Concen: 11.132 ng
RT: 8.569 min Scan# 932
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:	107	Resp:	137347
Ion	Ratio	Lower	Upper
107	100		
108	87.0	71.0	111.0
77	31.2	16.6	56.6
79	26.2	12.7	52.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.581 min Scan# 1274
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	136	Resp:	711045
Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.6	12.8
54	6.9	4.6	6.8#
68	5.3	3.6	5.4



#22
Acetophenone
Concen: 10.835 ng
RT: 8.610 min Scan# 939
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:	105	Resp:	185640
Ion	Ratio	Lower	Upper
105	100		
71	3.0	1.9	2.9#
51	25.5	18.4	27.6
120	22.2	17.9	26.9









#23
Nitrobenzene-d5
Concen: 21.535 ng
RT: 8.945 min Scan# 996
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion: 82 Resp: 262930
Ion Ratio Lower Upper
82 100
128 45.7 35.7 53.5
54 48.1 35.0 52.4



#24
Nitrobenzene
Concen: 10.622 ng
RT: 8.987 min Scan# 1003
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion: 77 Resp: 123655
Ion Ratio Lower Upper
77 100
123 48.5 38.2 57.2
65 13.8 11.4 17.2









#25
Isophorone
Concen: 11.130 ng
RT: 9.516 min Scan# 1093
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion: 82 Resp: 236957
Ion Ratio Lower Upper
82 100
95 7.2 6.5 9.7
138 18.4 15.0 22.4



#26
2-Nitrophenol
Concen: 10.300 ng
RT: 9.704 min Scan# 1125
Delta R.T. 0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion: 139 Resp: 62521
Ion Ratio Lower Upper
139 100
109 28.1 28.8 43.2#
65 46.8 41.0 61.6









#27
2,4-Dimethylphenol
Concen: 10.465 ng
RT: 9.769 min Scan# 1136
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:122 Resp: 100014
Ion Ratio Lower Upper
122 100
107 110.1 94.5 141.7
121 58.6 47.4 71.2



#28
bis(2-Chloroethoxy)methane
Concen: 10.884 ng
RT: 9.998 min Scan# 1175
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion: 93 Resp: 163483
Ion Ratio Lower Upper
93 100
95 32.7 25.9 38.9
123 11.6 10.2 15.2









#29
2,4-Dichlorophenol
Concen: 10.404 ng
RT: 10.245 min Scan# 1217
Delta R.T. 0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:162 Resp: 115999
Ion Ratio Lower Upper
162 100
164 63.2 44.9 84.9
98 33.2 14.3 54.3



#30
1,2,4-Trichlorobenzene
Concen: 10.453 ng
RT: 10.445 min Scan# 1251
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:180 Resp: 132767
Ion Ratio Lower Upper
180 100
182 95.5 77.9 116.9
145 28.7 24.7 37.1









#31
Naphthalene
Concen: 10.564 ng
RT: 10.628 min Scan# 1282
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:128 Resp: 383347
Ion Ratio Lower Upper
128 100
129 11.3 9.0 13.6
127 13.2 10.7 16.1



#32
Benzoic acid
Concen: 8.808 ng
RT: 9.904 min Scan# 1159
Delta R.T. -0.059 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:122 Resp: 64699
Ion Ratio Lower Upper
122 100
105 125.9 103.2 143.2
77 87.1 76.2 116.2









#33
4-Chloroaniline
Concen: 10.615 ng
RT: 10.757 min Scan# 1304
Delta R.T. 0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	127	Resp:	159151
Ion	Ratio	Lower	Upper
127	100		
129	32.5	26.1	39.1
65	27.4	22.5	33.7
92	18.0	15.0	22.6



#34
Hexachlorobutadiene
Concen: 10.755 ng
RT: 10.910 min Scan# 1330
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:	225	Resp:	81793
Ion	Ratio	Lower	Upper
225	100		
223	63.3	51.0	76.6
227	60.4	51.4	77.2









#35
Caprolactam
Concen: 11.414 ng
RT: 11.545 min Scan# 1438
Delta R.T. -0.024 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:113 Resp: 39150
Ion Ratio Lower Upper
113 100
55 160.5 98.9 138.9#
56 116.7 71.7 111.7#



#36
4-Chloro-3-methylphenol
Concen: 11.121 ng
RT: 11.886 min Scan# 1496
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:107 Resp: 126030
Ion Ratio Lower Upper
107 100
144 27.7 21.1 31.7
142 83.3 66.0 99.0









#37
2-Methylnaphthalene
Concen: 10.961 ng
RT: 12.245 min Scan# 1557
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:142 Resp: 278515
Ion Ratio Lower Upper
142 100
141 87.8 71.8 107.6
115 33.4 31.6 47.4



#38
1-Methylnaphthalene
Concen: 10.984 ng
RT: 12.463 min Scan# 1594
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:142 Resp: 272700
Ion Ratio Lower Upper
142 100
141 91.6 73.3 109.9
116 3.8 3.5 5.3









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.428 min Scan# 1928
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:164 Resp: 492457




Ion	Ratio	Lower	Upper
164	100		
162	101.4	80.2	120.2
160	44.4	34.5	51.7




#40
1,2,4,5-Tetrachlorobenzene
Concen: 10.230 ng
RT: 12.616 min Scan# 1620
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:216 Resp: 154282

Ion	Ratio	Lower	Upper
216	100		
214	77.6	61.9	92.9
179	19.8	16.8	25.2
108	16.2	14.9	22.3









#41
Hexachlorocyclopentadiene
Concen: 3.151 ng
RT: 12.592 min Scan# 1616
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:237	Resp:	26338
Ion Ratio	Lower	Upper
237	100	
235	64.0	44.5 84.5
272	12.2	0.0 32.5



#42
2,4,6-Tribromophenol
Concen: 21.388 ng
RT: 15.922 min Scan# 2182
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:330	Resp:	136551
Ion Ratio	Lower	Upper
330	100	
332	97.7	77.8 116.8
141	32.3	25.4 38.2









#43
2,4,6-Trichlorophenol
Concen: 10.209 ng
RT: 12.869 min Scan# 1663
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:196 Resp: 105051
Ion Ratio Lower Upper
196 100
198 96.6 79.5 119.3
200 30.9 25.0 37.6



#44
2,4,5-Trichlorophenol
Concen: 10.286 ng
RT: 12.945 min Scan# 1676
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:196 Resp: 113975
Ion Ratio Lower Upper
196 100
198 94.4 74.7 112.1
97 43.2 37.4 56.0
132 25.4 20.0 30.0









#45
2-Fluorobiphenyl
Concen: 21.395 ng
RT: 13.051 min Scan# 1694
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	172	Resp:	735429
Ion Ratio	Lower	Upper	
172	100		
171	35.6	27.9	41.9
170	23.5	18.6	27.8



#46
1,1'-Biphenyl
Concen: 10.545 ng
RT: 13.263 min Scan# 1730
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:	154	Resp:	386269
Ion Ratio	Lower	Upper	
154	100		
153	39.5	20.4	60.4
76	11.6	0.0	30.8









#47
2-Chloronaphthalene
Concen: 10.503 ng
RT: 13.304 min Scan# 1737
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:162 Resp: 298020
Ion Ratio Lower Upper
162 100
127 35.5 31.4 47.0
164 32.3 26.6 39.8



#48
2-Nitroaniline
Concen: 10.659 ng
RT: 13.516 min Scan# 1773
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion: 65 Resp: 73740
Ion Ratio Lower Upper
65 100
92 73.7 57.7 86.5
138 120.1 92.1 138.1





#49

Acenaphthylene

Concen: 10.673 ng

RT: 14.151 min Scan# 1881

Delta R.T. -0.000 min

Lab File: BP007369.D

Acq: 08 Oct 2021 10:48





Instrument :

BNA_P

ClientSampleId :

SSTDIC010

Tgt Ion:152	Resp: 466455
Ion Ratio	Lower Upper
152	100
151	20.2 16.2 24.2
153	12.9 10.6 16.0



#50

Dimethylphthalate

Concen: 11.088 ng




RT: 13.892 min Scan# 1837


Delta R.T. -0.006 min

Lab File: BP007369.D

Acq: 08 Oct 2021 10:48

Tgt Ion:163	Resp: 392732
Ion Ratio	Lower Upper
163	100
194	4.1 3.4 5.2
164	9.8 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 10.792 ng
RT: 14.010 min Scan# 1857
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:165 Resp: 78083
Ion Ratio Lower Upper
165 100
63 52.2 41.3 61.9
89 55.1 44.8 67.2



#52
Acenaphthene
Concen: 10.497 ng
RT: 14.492 min Scan# 1939
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:154 Resp: 277936
Ion Ratio Lower Upper
154 100
153 112.4 89.4 134.2
152 54.7 43.0 64.4









#53
3-Nitroaniline
Concen: 10.356 ng
RT: 14.345 min Scan# 1914
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:138 Resp: 81011
Ion Ratio Lower Upper
138 100
108 9.1 8.4 12.6
92 109.7 87.8 131.8



#54
2,4-Dinitrophenol
Concen: 7.914 ng
RT: 14.569 min Scan# 1952
Delta R.T. 0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:184 Resp: 32985
Ion Ratio Lower Upper
184 100
63 63.1 47.0 70.6
154 60.4 52.3 78.5









#55
Dibenzenofuran
Concen: 10.704 ng
RT: 14.827 min Scan# 1996
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:168 Resp: 472862
Ion Ratio Lower Upper
168 100
139 35.4 31.0 46.6
169 13.2 10.8 16.2



#56
4-Nitrophenol
Concen: 9.050 ng
RT: 14.675 min Scan# 1970
Delta R.T. 0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:139 Resp: 57474
Ion Ratio Lower Upper
139 100
109 71.5 72.3 112.3#
65 89.8 72.5 112.5









#57
2,4-Dinitrotoluene
Concen: 10.626 ng
RT: 14.810 min Scan# 1993
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	165	Resp:	101451
Ion Ratio	Lower	Upper	
165	100		
63	42.9	32.3	48.5
89	69.9	55.9	83.9
182	3.0	2.2	3.4



#58
Fluorene
Concen: 10.830 ng
RT: 15.480 min Scan# 2107
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:	166	Resp:	369605
Ion Ratio	Lower	Upper	
166	100		
165	98.6	78.0	117.0
167	12.7	10.5	15.7









#59
2,3,4,6-Tetrachlorophenol
Concen: 9.790 ng
RT: 14.980 min Scan# 2022
Delta R.T. -0.083 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:232	Resp:	95259
Ion Ratio	Lower	Upper
232	100	
131	36.6	32.2 48.4
130	2.2	2.0 3.0
166	23.9	19.4 29.2



#60
Diethylphthalate
Concen: 11.247 ng
RT: 15.257 min Scan# 2069
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:149	Resp:	385996
Ion Ratio	Lower	Upper
149	100	
177	22.8	17.5 26.3
150	12.0	9.9 14.9









#61
4-Chlorophenyl-phenylether
Concen: 10.993 ng
RT: 15.475 min Scan# 2106
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:204 Resp: 198260
Ion Ratio Lower Upper
204 100
206 32.7 26.6 40.0
141 51.8 48.6 73.0



#62
4-Nitroaniline
Concen: 10.446 ng
RT: 15.510 min Scan# 2112
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:138 Resp: 82184
Ion Ratio Lower Upper
138 100
92 51.1 31.2 71.2
108 77.1 78.6 118.6#









#63
Azobenzene
Concen: 11.050 ng
RT: 15.769 min Scan# 2156
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion: 77	Resp: 328648
Ion Ratio	Lower Upper
77 100	
182 30.2	6.0 46.0
105 21.8	0.0 38.8
51 27.8	8.8 48.8



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.180 min Scan# 2396
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion: 188	Resp: 705327
Ion Ratio	Lower Upper
188 100	
94 7.5	5.6 8.4
80 8.4	6.3 9.5









#65
4,6-Dinitro-2-methylphenol
Concen: 14.334 ng
RT: 15.575 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:198 Resp: 60670
Ion Ratio Lower Upper
198 100
51 39.4 14.0 54.0
105 42.0 19.5 59.5



#66
n-Nitrosodiphenylamine
Concen: 15.672 ng
RT: 15.692 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:169 Resp: 328893
Ion Ratio Lower Upper
169 100
168 66.7 54.4 81.6
167 35.7 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 15.692 ng
RT: 16.369 min Scan# 2258
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:248 Resp: 121304
Ion Ratio Lower Upper
248 100
250 95.4 79.8 119.8
141 67.8 57.8 86.6



#68
Hexachlorobenzene
Concen: 15.777 ng
RT: 16.486 min Scan# 2278
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:284 Resp: 134793
Ion Ratio Lower Upper
284 100
142 31.2 26.9 40.3
249 30.2 24.5 36.7





#69

Atrazine

Concen: 15.792 ng

RT: 16.651 min Scan# 2306

Delta R.T. -0.006 min

Lab File: BP007369.D

Acq: 08 Oct 2021 10:48

Tgt Ion:200 Resp: 116824

Ion Ratio Lower Upper

200 100

173 26.0 4.1 44.1





215 49.7 27.7 67.7

Instrument :

BNA_P

ClientSampleId :

SSTDIC010



#70

Pentachlorophenol

Concen: 12.720 ng

RT: 16.839 min Scan# 2338

Delta R.T. -0.000 min

Lab File: BP007369.D

Acq: 08 Oct 2021 10:48




Tgt Ion:266 Resp: 67487


Ion Ratio Lower Upper

266 100

268 60.7 51.6 77.4

264 61.8 50.3 75.5









#71
Phenanthrene
Concen: 10.423 ng
RT: 17.221 min Scan# 2403
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:178 Resp: 394178
Ion Ratio Lower Upper
178 100
176 19.1 15.1 22.7
179 14.9 12.4 18.6



#72
Anthracene
Concen: 10.426 ng
RT: 17.316 min Scan# 2419
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:178 Resp: 385714
Ion Ratio Lower Upper
178 100
176 18.4 14.9 22.3
179 15.3 12.2 18.2









#73
 Carbazole
 Concen: 10.291 ng
 RT: 17.592 min Scan# 2466
 Delta R.T. -0.000 min
 Lab File: BP007369.D
 Acq: 08 Oct 2021 10:48




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010


Tgt Ion:	167	Resp:	373071
Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.1	25.7
139	11.8	10.7	16.1



#74
 Di-n-butylphthalate
 Concen: 10.626 ng
 RT: 18.145 min Scan# 2560
 Delta R.T. -0.000 min
 Lab File: BP007369.D
 Acq: 08 Oct 2021 10:48

Tgt Ion:	149	Resp:	426928
Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.3	10.9
104	4.3	4.5	6.7









#75
Fluoranthene
Concen: 11.278 ng
RT: 19.198 min Scan# 2739
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	202	Resp:	491429
Ion Ratio	Lower	Upper	
202	100		
101	8.6	0.0	27.6
203	17.2	0.0	37.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.274 min Scan# 3092
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:	240	Resp:	799969
Ion Ratio	Lower	Upper	
240	100		
120	7.7	6.5	9.7
236	25.0	20.6	31.0









#77
Benzidine
Concen: 8.245 ng
RT: 19.386 min Scan# 2771
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	184	Resp:	202701
Ion Ratio	Lower	Upper	
184	100		
185	14.0	11.0	16.4
183	12.0	9.5	14.3



#78
Pyrene
Concen: 9.670 ng
RT: 19.551 min Scan# 2799
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:	202	Resp:	507518
Ion Ratio	Lower	Upper	
202	100		
200	20.8	16.3	24.5
203	17.1	14.2	21.2





#79

Terphenyl-d14

Concen: 21.996 ng

RT: 19.745 min Scan# 2832

Delta R.T. -0.006 min

Lab File: BP007369.D

Acq: 08 Oct 2021 10:48

Instrument :

BNA_P

ClientSampleId :

SSTDIC010





Tgt Ion:244 Resp: 918238

Ion Ratio Lower Upper

244 100

212 6.8 6.0 9.0

122 8.1 7.1 10.7



#80

Butylbenzylphthalate

Concen: 9.569 ng

RT: 20.427 min Scan# 2948

Delta R.T. -0.000 min

Lab File: BP007369.D

Acq: 08 Oct 2021 10:48




Tgt Ion:149 Resp: 201121


Ion Ratio Lower Upper

149 100

91 60.6 54.6 81.8

206 24.1 16.7 25.1









#81
Benzo(a)anthracene
Concen: 10.288 ng
RT: 21.262 min Scan# 3090
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:228	Resp: 534233
Ion Ratio	Lower Upper
228 100	
226 26.9	21.6 32.4
229 19.8	15.8 23.6



#82
3,3'-Dichlorobenzidine
Concen: 11.059 ng
RT: 21.198 min Scan# 3079
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:252	Resp: 197220
Ion Ratio	Lower Upper
252 100	
254 65.0	50.6 76.0
126 8.0	7.0 10.6









#83
Chrysene
Concen: 10.312 ng
RT: 21.315 min Scan# 3099
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:228 Resp: 511746
Ion Ratio Lower Upper
228 100
226 30.1 23.8 35.6
229 19.2 15.4 23.2



#84
Bis(2-ethylhexyl)phthalate
Concen: 9.764 ng
RT: 21.186 min Scan# 3077
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:149 Resp: 295049
Ion Ratio Lower Upper
149 100
167 28.8 21.9 32.9
279 5.5 4.3 6.5









#85
Di-n-octyl phthalate
Concen: 9.832 ng
RT: 22.092 min Scan# 3231
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	149	Resp:	505824
Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.4	2.0
43	6.3	5.6	8.4



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.639 min Scan# 3494
Delta R.T. -0.000 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:	264	Resp:	831618
Ion	Ratio	Lower	Upper
264	100		
260	24.2	18.6	28.0
265	22.5	17.6	26.4









#87
Indeno(1,2,3-cd)pyrene
Concen: 10.243 ng
RT: 26.103 min Scan# 3913
Delta R.T. -0.012 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:276 Resp: 612166
Ion Ratio Lower Upper
276 100
138 17.1 15.3 22.9
277 25.6 20.4 30.6



#88
Benzo(b)fluoranthene
Concen: 10.573 ng
RT: 22.915 min Scan# 3371
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:252 Resp: 525455
Ion Ratio Lower Upper
252 100
253 21.5 17.5 26.3
125 7.2 6.2 9.4









#89
Benzo(k)fluoranthene
Concen: 10.340 ng
RT: 22.962 min Scan# 3379
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:252 Resp: 520305
Ion Ratio Lower Upper
252 100
253 21.8 17.6 26.4
125 7.1 6.0 9.0



#90
Benzo(a)pyrene
Concen: 10.291 ng
RT: 23.527 min Scan# 3475
Delta R.T. -0.006 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:252 Resp: 434508
Ion Ratio Lower Upper
252 100
253 22.1 17.9 26.9
125 8.5 6.5 9.7









#91
Dibenzo(a,h)anthracene
Concen: 10.140 ng
RT: 26.109 min Scan# 3914
Delta R.T. -0.018 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Instrument :
BNA_P
ClientSampleId :
SSTDIC010

Tgt Ion:278 Resp: 507878
Ion Ratio Lower Upper
278 100
139 11.2 9.3 13.9
279 23.8 18.8 28.2



#92
Benzo(g,h,i)perylene
Concen: 10.085 ng
RT: 26.850 min Scan# 4040
Delta R.T. -0.018 min
Lab File: BP007369.D
Acq: 08 Oct 2021 10:48

Tgt Ion:276 Resp: 492901
Ion Ratio Lower Upper
276 100
277 24.0 18.9 28.3
138 13.8 12.6 18.8

