

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100821\
 Data File : BP007381.D
 Acq On : 08 Oct 2021 19:17
 Operator : CG/JU
 Sample : M4077-09
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 CC-33

Quant Time: Oct 09 06:18:28 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP100821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 08 15:35:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	108244	20.00	ng	0.00
21) Naphthalene-d8	10.575	136	477584	20.00	ng	0.00
39) Acenaphthene-d10	14.422	164	226326	20.00	ng	0.00
64) Phenanthrene-d10	17.174	188	435460	20.00	ng	-0.01
76) Chrysene-d12	21.268	240	517743	20.00	ng	-0.01
86) Perylene-d12	23.621	264	587193	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.357	112	525929	80.06	ng	0.00
7) Phenol-d6	6.957	99	583153	66.12	ng	0.00
23) Nitrobenzene-d5	8.940	82	232408	28.66	ng	-0.01
42) 2,4,6-Tribromophenol	15.916	330	197210	53.12	ng	-0.01
45) 2-Fluorobiphenyl	13.045	172	608570	37.57	ng	0.00
79) Terphenyl-d14	19.739	244	926886	31.58	ng	-0.01
Target Compounds						Qvalue
74) Di-n-butylphthalate	18.139	149	874240	31.82	ng	98
80) Butylbenzylphthalate	20.421	149	438519	31.84	ng	89
84) Bis(2-ethylhexyl)phtha...	21.180	149	646962	33.75	ng	97
85) Di-n-octyl phthalate	22.092	149	74887	2.29	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100821\
Data File : BP007381.D
Acq On : 08 Oct 2021 19:17
Operator : CG/JU
Sample : M4077-09
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
CC-33

Quant Time: Oct 09 06:18:28 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP100821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 08 15:35:55 2021
Response via : Initial Calibration

