

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100821\
 Data File : BP007386.D
 Acq On : 08 Oct 2021 22:08
 Operator : CG/JU
 Sample : M4077-10
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 CC-34

Quant Time: Oct 09 06:19:36 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP100821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 08 15:35:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.781	152	24999	20.00	ng	0.00
21) Naphthalene-d8	10.575	136	88616	20.00	ng	0.00
39) Acenaphthene-d10	14.422	164	54868	20.00	ng	0.00
64) Phenanthrene-d10	17.175	188	179710	20.00	ng	#-0.01
76) Chrysene-d12	21.268	240	128935	20.00	ng	-0.01
86) Perylene-d12	23.621	264	137316	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.358	112	202083	133.20	ng	0.00
7) Phenol-d6	6.963	99	115538	56.72	ng	0.00
23) Nitrobenzene-d5	8.940	82	70370	46.77	ng	-0.01
42) 2,4,6-Tribromophenol	15.916	330	69934	77.70	ng	-0.01
45) 2-Fluorobiphenyl	13.045	172	196139	49.95	ng	0.00
79) Terphenyl-d14	19.739	244	329180	45.03	ng	-0.01
Target Compounds						
74) Di-n-butylphthalate	18.139	149	366801	32.35	ng	Qvalue 99
80) Butylbenzylphthalate	20.421	149	75018	21.87	ng	91
84) Bis(2-ethylhexyl)phtha...	21.180	149	370188	77.55	ng	96
85) Di-n-octyl phthalate	22.092	149	70449	8.65	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100821\
Data File : BP007386.D
Acq On : 08 Oct 2021 22:08
Operator : CG/JU
Sample : M4077-10
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
CC-34

Quant Time: Oct 09 06:19:36 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP100821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 08 15:35:55 2021
Response via : Initial Calibration

