


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100821\
 Data File : BP007377.D
 Acq On : 08 Oct 2021 16:26
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP100821.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Oct 08 15:35:55 2021






AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2464

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	21.4	60787	PASS
68	69	0.00	2	1.2	823	PASS
69	198	0.00	100	24.5	69645	PASS
70	69	0.00	2	0.6	389	PASS
127	198	10	80	38.1	108507	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	284701	PASS
199	198	5	9	6.9	19633	PASS
275	198	10	60	32.0	90970	PASS
365	198	1	100	4.2	11839	PASS
441	198	0.01	100	18.5	52707	PASS
442	442	100	100	100.0	337165	PASS
443	442	15	24	19.8	66779	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100821\
 Data File : BP007377.D
 Acq On : 08 Oct 2021 16:26
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Oct 08 18:24:19 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP100821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 08 15:35:55 2021
 Response via : Initial Calibration



TIC: BP007377.D\data.ms

(70) Pentachlorophenol (C)

16.834min (-0.005) 776393.54 ng




response 387381

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.50	64.84
264.00	62.90	61.95
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100821\
Data File : BP007377.D
Acq On : 08 Oct 2021 16:26
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Oct 08 18:24:19 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP100821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 08 15:35:55 2021
Response via : Initial Calibration



TIC: BP007377.D\data.ms

(77) Benzidine

19.386min (+ 0.000) 2508528.06 ng

response 2038508

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.70	14.01
183.00	11.90	12.38
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/8/2021	BNA_P	BP007377.D
Compound Name	Response	Retention Time
DDT	1141337	20.58
DDD	14732	20.15
DDE	984	20.15
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
15716	1157053	1.36

Instrument :
BNA_P
ClientSampleId :
DFTPP