

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100825\
 Data File : BP025953.D
 Acq On : 08 Oct 2025 14:45
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 08 16:22:26 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 08 16:22:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	36.635	8.4	100	0.00
3	Pyridine	40.000	38.832	2.9	100	0.00
4	n-Nitrosodimethylamine	40.000	36.092	9.8	100	0.00
5 S	2-Fluorophenol	80.000	82.446	-3.1	100	0.00
6	Aniline	40.000	39.201	2.0	100	0.00
7 S	Phenol-d6	80.000	81.842	-2.3	100	0.00
8	2-Chlorophenol	40.000	41.643	-4.1	100	0.00
9	Benzaldehyde	40.000	48.788	-22.0	100	0.00
10 C	Phenol	40.000	40.379	-0.9	100	0.00
11	bis(2-Chloroethyl)ether	40.000	39.292	1.8	100	0.00
12	1,3-Dichlorobenzene	40.000	40.486	-1.2	100	0.00
13 C	1,4-Dichlorobenzene	40.000	40.316	-0.8	100	0.00
14	1,2-Dichlorobenzene	40.000	39.840	0.4	100	0.00
15	Benzyl Alcohol	40.000	40.243	-0.6	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.418	11.5	100	0.00
17	2-Methylphenol	40.000	40.518	-1.3	100	0.00
18	Hexachloroethane	40.000	39.795	0.5	100	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.496	3.8	100	0.00
20	3+4-Methylphenols	40.000	39.094	2.3	100	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	100	0.00
22	Acetophenone	40.000	41.064	-2.7	100	0.00
23 S	Nitrobenzene-d5	80.000	80.616	-0.8	100	0.00
24	Nitrobenzene	40.000	40.268	-0.7	100	0.00
25	Isophorone	40.000	39.340	1.6	100	0.00
26 C	2-Nitrophenol	40.000	44.507	-11.3	100	0.00
27	2,4-Dimethylphenol	40.000	40.484	-1.2	100	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.456	-1.1	100	0.00
29 C	2,4-Dichlorophenol	40.000	42.175	-5.4	100	0.00
30	1,2,4-Trichlorobenzene	40.000	41.195	-3.0	100	0.00
31	Naphthalene	40.000	39.982	0.0	100	0.00
32	Benzoic acid	40.000	40.238	-0.6	100	0.00
33	4-Chloroaniline	40.000	41.698	-4.2	100	0.00
34 C	Hexachlorobutadiene	40.000	42.095	-5.2	100	0.00
35	Caprolactam	40.000	46.406	-16.0	100	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.523	-1.3	100	0.00
37	2-Methylnaphthalene	40.000	39.936	0.2	100	0.00
38	1-Methylnaphthalene	40.000	39.570	1.1	100	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	43.242	-8.1	100	0.00
41 P	Hexachlorocyclopentadiene	40.000	48.428	-21.1	100	0.00
42 S	2,4,6-Tribromophenol	80.000	89.275	-11.6	100	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.162	-7.9	100	0.00
44	2,4,5-Trichlorophenol	40.000	42.715	-6.8	100	0.00
45 S	2-Fluorobiphenyl	80.000	80.784	-1.0	100	0.00
46	1,1'-Biphenyl	40.000	40.649	-1.6	100	0.00
47	2-Chloronaphthalene	40.000	40.959	-2.4	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100825\
 Data File : BP025953.D
 Acq On : 08 Oct 2025 14:45
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 08 16:22:26 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 08 16:22:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.033	-2.6	100	0.00
49	Acenaphthylene	40.000	40.087	-0.2	100	0.00
50	Dimethylphthalate	40.000	40.285	-0.7	100	0.00
51	2,6-Dinitrotoluene	40.000	42.261	-5.7	100	0.00
52 C	Acenaphthene	40.000	40.076	-0.2	100	0.00
53	3-Nitroaniline	40.000	43.163	-7.9	100	0.00
54 P	2,4-Dinitrophenol	40.000	42.201	-5.5	100	0.00
55	Dibenzofuran	40.000	39.382	1.5	100	0.00
56 P	4-Nitrophenol	40.000	42.881	-7.2	100	0.00
57	2,4-Dinitrotoluene	40.000	42.175	-5.4	100	0.00
58	Fluorene	40.000	39.680	0.8	100	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.576	-3.9	100	0.00
60	Diethylphthalate	40.000	39.944	0.1	100	0.00
61	4-Chlorophenyl-phenylether	40.000	40.863	-2.2	100	0.00
62	4-Nitroaniline	40.000	40.523	-1.3	100	0.00
63	Azobenzene	40.000	38.833	2.9	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.347	-15.9	100	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.202	-3.0	100	0.00
67	4-Bromophenyl-phenylether	40.000	42.973	-7.4	100	0.00
68	Hexachlorobenzene	40.000	43.358	-8.4	100	0.00
69	Atrazine	40.000	41.541	-3.9	100	0.00
70 C	Pentachlorophenol	40.000	38.372	4.1	100	0.00
71	Phenanthrene	40.000	39.764	0.6	100	0.00
72	Anthracene	40.000	40.293	-0.7	100	0.00
73	Carbazole	40.000	39.380	1.5	100	0.00
74	Di-n-butylphthalate	40.000	40.683	-1.7	100	0.00
75 C	Fluoranthene	40.000	39.256	1.9	100	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzidine	40.000	45.337	-13.3	100	0.00
78	Pyrene	40.000	41.200	-3.0	100	0.00
79 S	Terphenyl-d14	80.000	83.601	-4.5	100	0.00
80	Butylbenzylphthalate	40.000	41.521	-3.8	100	0.00
81	Benzo(a)anthracene	40.000	39.737	0.7	100	0.00
82	3,3'-Dichlorobenzidine	40.000	42.943	-7.4	100	0.00
83	Chrysene	40.000	39.820	0.4	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.946	0.1	100	0.00
85 c	Di-n-octyl phthalate	40.000	41.594	-4.0	100	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.035	-2.6	100	0.00
88	Benzo(b)fluoranthene	40.000	39.747	0.6	100	0.00
89	Benzo(k)fluoranthene	40.000	39.125	2.2	100	0.00
90 C	Benzo(a)pyrene	40.000	40.233	-0.6	100	0.00
91	Dibenzo(a,h)anthracene	40.000	41.091	-2.7	100	0.00
92	Benzo(g,h,i)perylene	40.000	40.870	-2.2	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100825\
Data File : BP025953.D
Acq On : 08 Oct 2025 14:45
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Oct 08 16:22:26 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 08 16:22:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0