

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100924\
 Data File : BP022301.D
 Acq On : 09 Oct 2024 23:11
 Operator : RC/JU
 Sample : PB163728BL
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK728

Quant Time: Oct 10 02:09:37 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 08 02:14:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.711	152	135700	20.000	ng/u1	0.00
20) Naphthalene-d8	10.463	136	513324	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.316	164	399920	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.110	188	988322	20.000	ng/u1	0.00
79) Chrysene-d12	21.545	240	977949	20.000	ng/u1	0.00
88) Perylene-d12	24.833	264	1142553	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.228	96	14204	4.068	ng/uL	0.00
4) Pyridine-d5	3.640	84	207029	21.596	ng/u1	0.00
7) Phenol-d5	6.899	99	279921	24.505	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.046	67	160103	23.847	ng/u1	0.00
11) 2-Chlorophenol-d4	7.252	132	209544	25.853	ng/u1	0.00
15) 4-Methylphenol-d8	8.411	113	225360	24.721	ng/u1	0.00
21) Nitrobenzene-d5	8.846	128	103338	26.182	ng/u1	0.00
24) 2-Nitrophenol-d4	9.557	143	124975	27.349	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.099	165	251341	25.879	ng/u1	0.00
31) 4-Chloroaniline-d4	10.605	131	290195	25.287	ng/u1	0.01
46) Dimethylphthalate-d6	13.722	166	799409	25.159	ng/u1	0.00
49) Acenaphthylene-d8	14.010	160	871969	25.837	ng/u1	0.01
54) 4-Nitrophenol-d4	14.540	143	127804	23.976	ng/u1	0.01
60) Fluorene-d10	15.316	176	697200	25.297	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.457	200	109991	16.922	ng/u1	0.00
73) Anthracene-d10	17.216	188	1242449	26.723	ng/u1	0.00
81) Pyrene-d10	19.586	212	1572653	30.410	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.604	264	1667295	27.325	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100924\
Data File : BP022301.D
Acq On : 09 Oct 2024 23:11
Operator : RC/JU
Sample : PB163728BL
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK728

Quant Time: Oct 10 02:09:37 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 08 02:14:47 2024
Response via : Initial Calibration

