

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100925\
 Data File : BP025962.D
 Acq On : 09 Oct 2025 12:23
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 09 13:19:05 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 08 16:22:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	112	0.00
2	1,4-Dioxane	40.000	36.733	8.2	113	0.00
3	Pyridine	40.000	38.521	3.7	111	0.00
4	n-Nitrosodimethylamine	40.000	35.898	10.3	112	0.00
5 S	2-Fluorophenol	80.000	83.340	-4.2	114	0.00
6	Aniline	40.000	39.247	1.9	112	0.00
7 S	Phenol-d6	80.000	81.994	-2.5	113	0.00
8	2-Chlorophenol	40.000	41.231	-3.1	111	0.00
9	Benzaldehyde	40.000	48.193	-20.5	111	0.00
10 C	Phenol	40.000	40.466	-1.2	113	0.00
11	bis(2-Chloroethyl)ether	40.000	39.488	1.3	113	0.00
12	1,3-Dichlorobenzene	40.000	40.872	-2.2	113	0.00
13 C	1,4-Dichlorobenzene	40.000	40.290	-0.7	112	0.00
14	1,2-Dichlorobenzene	40.000	40.146	-0.4	113	0.00
15	Benzyl Alcohol	40.000	39.034	2.4	109	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.597	11.0	113	0.00
17	2-Methylphenol	40.000	39.797	0.5	110	0.00
18	Hexachloroethane	40.000	39.775	0.6	112	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.803	5.5	110	0.00
20	3+4-Methylphenols	40.000	38.553	3.6	111	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	111	0.00
22	Acetophenone	40.000	40.523	-1.3	109	0.00
23 S	Nitrobenzene-d5	80.000	79.618	0.5	109	0.00
24	Nitrobenzene	40.000	40.240	-0.6	111	0.00
25	Isophorone	40.000	39.007	2.5	110	0.00
26 C	2-Nitrophenol	40.000	44.241	-10.6	110	0.00
27	2,4-Dimethylphenol	40.000	40.421	-1.1	110	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.180	-0.4	110	0.00
29 C	2,4-Dichlorophenol	40.000	41.843	-4.6	110	0.00
30	1,2,4-Trichlorobenzene	40.000	41.595	-4.0	112	0.00
31	Naphthalene	40.000	40.138	-0.3	111	0.00
32	Benzoic acid	40.000	39.275	1.8	108	0.00
33	4-Chloroaniline	40.000	41.239	-3.1	109	0.00
34 C	Hexachlorobutadiene	40.000	42.216	-5.5	111	0.00
35	Caprolactam	40.000	44.543	-11.4	106	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.883	0.3	109	0.00
37	2-Methylnaphthalene	40.000	39.707	0.7	110	0.00
38	1-Methylnaphthalene	40.000	38.310	4.2	107	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	109	0.01
40	1,2,4,5-Tetrachlorobenzene	40.000	43.452	-8.6	109	0.00
41 P	Hexachlorocyclopentadiene	40.000	54.171	-35.4#	122	0.00
42 S	2,4,6-Tribromophenol	80.000	89.940	-12.4	109	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.205	-8.0	109	0.00
44	2,4,5-Trichlorophenol	40.000	42.311	-5.8	108	0.00
45 S	2-Fluorobiphenyl	80.000	81.083	-1.4	109	0.00
46	1,1'-Biphenyl	40.000	41.251	-3.1	110	0.00
47	2-Chloronaphthalene	40.000	41.074	-2.7	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100925\
 Data File : BP025962.D
 Acq On : 09 Oct 2025 12:23
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 09 13:19:05 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 08 16:22:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.847	0.4	105	0.00
49	Acenaphthylene	40.000	39.624	0.9	107	0.00
50	Dimethylphthalate	40.000	39.457	1.4	106	0.00
51	2,6-Dinitrotoluene	40.000	41.602	-4.0	107	0.00
52 C	Acenaphthene	40.000	38.949	2.6	106	0.01
53	3-Nitroaniline	40.000	41.656	-4.1	105	0.00
54 P	2,4-Dinitrophenol	40.000	41.624	-4.1	107	0.00
55	Dibenzofuran	40.000	39.136	2.2	108	0.01
56 P	4-Nitrophenol	40.000	42.285	-5.7	107	0.00
57	2,4-Dinitrotoluene	40.000	41.658	-4.1	107	0.00
58	Fluorene	40.000	39.440	1.4	108	0.01
59	2,3,4,6-Tetrachlorophenol	40.000	42.002	-5.0	110	0.00
60	Diethylphthalate	40.000	38.182	4.5	104	0.01
61	4-Chlorophenyl-phenylether	40.000	40.800	-2.0	108	0.01
62	4-Nitroaniline	40.000	39.687	0.8	106	0.00
63	Azobenzene	40.000	38.338	4.2	107	0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	45.997	-15.0	104	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.039	-5.1	107	0.00
67	4-Bromophenyl-phenylether	40.000	44.029	-10.1	107	0.00
68	Hexachlorobenzene	40.000	44.167	-10.4	106	0.00
69	Atrazine	40.000	41.271	-3.2	104	0.00
70 C	Pentachlorophenol	40.000	40.223	-0.6	109	0.00
71	Phenanthrene	40.000	40.623	-1.6	107	0.00
72	Anthracene	40.000	40.848	-2.1	106	0.01
73	Carbazole	40.000	40.230	-0.6	107	0.00
74	Di-n-butylphthalate	40.000	41.513	-3.8	107	0.00
75 C	Fluoranthene	40.000	39.778	0.6	106	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	106	0.00
77	Benzidine	40.000	45.901	-14.8	107	0.00
78	Pyrene	40.000	41.000	-2.5	105	0.00
79 S	Terphenyl-d14	80.000	83.835	-4.8	106	0.01
80	Butylbenzylphthalate	40.000	41.329	-3.3	105	0.01
81	Benzo(a)anthracene	40.000	40.054	-0.1	107	0.00
82	3,3'-Dichlorobenzidine	40.000	42.573	-6.4	105	0.01
83	Chrysene	40.000	40.304	-0.8	107	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.314	-0.8	107	0.01
85 c	Di-n-octyl phthalate	40.000	40.101	-0.3	102	0.00
86 I	Perylene-d12	20.000	20.000	0.0	105	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	41.438	-3.6	106	-0.01
88	Benzo(b)fluoranthene	40.000	41.058	-2.6	108	0.01
89	Benzo(k)fluoranthene	40.000	39.601	1.0	106	0.01
90 C	Benzo(a)pyrene	40.000	41.008	-2.5	107	0.01
91	Dibenzo(a,h)anthracene	40.000	41.544	-3.9	106	0.00
92	Benzo(g,h,i)perylene	40.000	41.025	-2.6	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100925\
Data File : BP025962.D
Acq On : 09 Oct 2025 12:23
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Oct 09 13:19:05 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 08 16:22:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0