

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BP012035.D
 Acq On : 11 Oct 2022 12:01
 Operator : CG/JU
 Sample : SSTD01069
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010632

Quant Time: Oct 11 23:04:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 11 23:01:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.857	152	226020	20.000	ng/u1	0.00
20) Naphthalene-d8	10.663	136	922192	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.504	164	507067	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.263	188	958575	20.000	ng/u1	0.00
79) Chrysene-d12	21.351	240	896564	20.000	ng/u1	0.00
88) Perylene-d12	23.774	264	927515	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	23106	9.655	ng/uL	0.00
4) Pyridine-d5	3.711	84	149706	19.260	ng/u1	0.00
7) Phenol-d5	7.028	99	169442	8.819	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.193	67	104652	9.214	ng/u1	0.00
11) 2-Chlorophenol-d4	7.387	132	140381	9.519	ng/u1	0.00
15) 4-Methylphenol-d8	8.569	113	129009	8.932	ng/u1	0.00
21) Nitrobenzene-d5	9.028	128	59763	8.944	ng/u1	0.00
24) 2-Nitrophenol-d4	9.751	143	56597	8.352	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.287	165	119786	9.397	ng/u1	0.00
31) 4-Chloroaniline-d4	10.810	131	178076	9.011	ng/u1	0.00
46) Dimethylphthalate-d6	13.916	166	313999	9.335	ng/u1	0.00
49) Acenaphthylene-d8	14.198	160	385322	9.341	ng/u1	0.00
54) 4-Nitrophenol-d4	14.716	143	48410	6.702	ng/u1	0.00
60) Fluorene-d10	15.504	176	291425	9.628	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.628	200	34596	6.661	ng/u1	0.00
73) Anthracene-d10	17.363	188	410255	9.447	ng/u1	0.00
81) Pyrene-d10	19.598	212	451823	10.254	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.615	264	434615	9.495	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.311	88	24863	9.971	ng/uL	99
5) Pyridine	3.728	79	136619	17.517	ng/u1	94
6) Benzaldehyde	7.005	77	88442	9.929	ng/u1	97
8) Phenol	7.052	94	184708	9.129	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.281	93	152139	9.620	ng/u1	99
12) 2-Chlorophenol	7.422	128	155116	9.694	ng/u1	97
13) 2-Methylphenol	8.304	108	133604	9.046	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.393	45	160934	8.689	ng/u1	100
16) Acetophenone	8.687	105	212891	9.245	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.669	70	92607	9.118	ng/u1	97
18) 4-Methylphenol	8.634	108	144654	9.064	ng/u1	96
19) Hexachloroethane	8.940	117	58553	8.855	ng/u1	97
22) Nitrobenzene	9.069	77	147124	8.916	ng/u1	100
23) Isophorone	9.599	82	250593	8.473	ng/u1	99
25) 2-Nitrophenol	9.781	139	71083	9.004	ng/u1	99
26) 2,4-Dimethylphenol	9.840	107	150430	9.608	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.081	93	180963	9.575	ng/u1	100
29) 2,4-Dichlorophenol	10.316	162	129946	9.746	ng/u1	98
30) Naphthalene	10.716	128	496757	9.980	ng/u1	99
32) 4-Chloroaniline	10.834	127	186076	9.062	ng/u1	100
33) Hexachlorobutadiene	10.998	225	81372	10.053	ng/u1	97
34) Caprolactam	11.628	113	26396	6.048	ng/u1	90
35) 4-Chloro-3-methylphenol	11.957	107	119284	8.728	ng/u1	99
36) 2-Methylnaphthalene	12.328	142	307183	9.682	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BP012035.D
 Acq On : 11 Oct 2022 12:01
 Operator : CG/JU
 Sample : SSTD01069
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD010632


Quant Time: Oct 11 23:04:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 11 23:01:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.545	142	311882	9.777	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.692	216	152929	10.454	ng/ul#	97
40) Hexachlorocyclopentadiene	12.675	237	52943	7.808	ng/ul	98
41) 2,4,6-Trichlorophenol	12.940	196	84680	9.175	ng/ul	97
42) 2,4,5-Trichlorophenol	13.010	196	95678	9.732	ng/ul	98
43) 1,1'-Biphenyl	13.339	154	389352	10.185	ng/ul	100
44) 2-Chloronaphthalene	13.381	162	313909	10.237	ng/ul	99
45) 2-Nitroaniline	13.592	65	56989	8.154	ng/ul	94
47) Dimethylphthalate	13.963	163	334425	9.347	ng/ul	98
48) 2,6-Dinitrotoluene	14.086	165	49726	8.371	ng/ul	94
50) Acenaphthylene	14.228	152	448602	9.613	ng/ul	99
51) 3-Nitroaniline	14.422	138	52680	7.185	ng/ul#	94
52) Acenaphthene	14.569	153	325328	9.856	ng/ul	99
53) 2,4-Dinitrophenol	14.628	184	22271	5.348	ng/ul	95
55) 4-Nitrophenol	14.728	109	36635	6.722	ng/ul	97
56) Dibenzofuran	14.904	168	441223	9.916	ng/ul	99
57) 2,4-Dinitrotoluene	14.875	165	75675	7.931	ng/ul#	94
58) 2,3,4,6-Tetrachlorophenol	15.134	232	71768	8.724	ng/ul	97
59) Diethylphthalate	15.334	149	301720	8.623	ng/ul	100
61) Fluorene	15.557	166	350965	9.724	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.551	204	164555	9.900	ng/ul	99
63) 4-Nitroaniline	15.586	138	42111	5.950	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.639	198	39482	6.974	ng/ul#	96
67) N-Nitrosodiphenylamine	15.769	169	282145	10.271	ng/ul	99
68) 4-Bromophenyl-phenylether	16.451	248	87588	9.860	ng/ul	99
69) Hexachlorobenzene	16.563	284	101958	9.873	ng/ul	98
70) Atrazine	16.728	200	75687	8.281	ng/ul	98
71) Pentachlorophenol	16.916	266	47099	7.334	ng/ul	97
72) Phenanthrene	17.310	178	528467	9.837	ng/ul	99
74) Anthracene	17.398	178	518779	9.592	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.304	216	150745	11.199	ng/ul	99
76) Pentachlorobenzene	14.822	250	146708	10.836	ng/ul	99
77) Carbazole	17.675	167	460291	9.110	ng/ul	99
78) Di-n-butylphthalate	18.222	149	414819	7.778	ng/ul	100
80) Fluoranthene	19.274	202	556602	10.294	ng/ul	97
82) Pyrene	19.627	202	604168	10.428	ng/ul	96
83) Butylbenzylphthalate	20.498	149	175387	8.117	ng/ul#	87
84) 3,3'-Dichlorobenzidine	21.274	252	177396	9.141	ng/ul#	97
85) Benzo(a)anthracene	21.339	228	566451	9.759	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.257	149	238878	7.734	ng/ul#	99
87) Chrysene	21.392	228	549963	9.880	ng/ul	99
89) Di-n-octyl phthalate	22.186	149	412124	7.326	ng/ul	100
90) Benzo(b)fluoranthene	23.033	252	583806	9.833	ng/ul#	98
91) Benzo(k)fluoranthene	23.080	252	574436	9.856	ng/ul	98
93) Benzo(a)pyrene	23.662	252	519721	9.770	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	26.286	276	635576	9.460	ng/ul#	95
95) Dibenzo(a,h)anthracene	26.304	278	575787	9.572	ng/ul#	97
96) Benzo(g,h,i)perylene	27.062	276	570998	9.443	ng/ul#	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 11 23:04:18 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 11 23:01:01 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 9.971 ng/uL
RT: 3.311 min Scan# 51
Delta R.T. -0.006 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion: 88 Resp: 24863
Ion Ratio Lower Upper
88 100
43 21.5 16.6 25.0
58 66.9 52.6 79.0



#5
Pyridine
Concen: 17.517 ng/uL
RT: 3.728 min Scan# 126
Delta R.T. 0.006 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion: 79 Resp: 136619
Ion Ratio Lower Upper
79 100
52 69.0 51.2 76.8
51 30.6 22.8 34.2









#6
Benzaldehyde
Concen: 9.929 ng/ul
RT: 7.005 min Scan# 61
Delta R.T. 0.006 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion: 77 Resp: 88442
Ion Ratio Lower Upper
77 100
105 103.9 85.2 127.8
106 100.6 83.1 124.7



#8
Phenol
Concen: 9.129 ng/ul
RT: 7.052 min Scan# 691
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion: 94 Resp: 184708
Ion Ratio Lower Upper
94 100
65 24.1 19.4 29.0
66 31.6 24.6 37.0









#10
Bis(2-Chloroethyl)ether
Concen: 9.620 ng/ul
RT: 7.281 min Scan# 730
Delta R.T. -0.006 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion: 93 Resp: 152139
Ion Ratio Lower Upper
93 100
63 71.4 56.7 85.1
95 32.1 26.2 39.2



#12
2-Chlorophenol
Concen: 9.694 ng/ul
RT: 7.422 min Scan# 754
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:128 Resp: 155116
Ion Ratio Lower Upper
128 100
64 42.3 32.7 49.1
130 31.1 26.3 39.5





#13

2-Methylphenol

Concen: 9.046 ng/ul

RT: 8.304 min Scan# 904

Delta R.T. -0.000 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01

Instrument :

BNA_P

ClientSampleId :





SSTD010632

Tgt Ion:108 Resp: 133604

Ion Ratio Lower Upper

108 100

107 86.3 71.9 107.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 8.689 ng/ul

RT: 8.393 min Scan# 919

Delta R.T. -0.000 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01




Tgt Ion: 45 Resp: 160934


Ion Ratio Lower Upper

45 100

77 18.2 14.4 21.6

79 13.9 11.3 16.9









#16
Acetophenone
Concen: 9.245 ng/ul
RT: 8.687 min Scan# 969
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:105 Resp: 212891
Ion Ratio Lower Upper
105 100
77 73.1 57.0 85.4
51 21.9 16.4 24.6



#17
N-Nitroso-di-n-propylamine
Concen: 9.118 ng/ul
RT: 8.669 min Scan# 966
Delta R.T. -0.006 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion: 70 Resp: 92607
Ion Ratio Lower Upper
70 100
42 45.0 34.6 52.0
101 9.2 8.0 12.0
130 21.5 18.3 27.5









#18
4-Methylphenol
Concen: 9.064 ng/ul
RT: 8.634 min Scan# 90
Delta R.T. -0.006 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:108 Resp: 144654
Ion Ratio Lower Upper
108 100
107 111.9 86.3 129.5



#19
Hexachloroethane
Concen: 8.855 ng/ul
RT: 8.940 min Scan# 1012
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:117 Resp: 58553
Ion Ratio Lower Upper
117 100
201 98.9 82.3 123.5
199 63.5 50.0 75.0









#22
Nitrobenzene
Concen: 8.916 ng/ul
RT: 9.069 min Scan# 1034
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion: 77 Resp: 147124
Ion Ratio Lower Upper
77 100
123 49.1 39.4 59.0
65 12.4 9.9 14.9



#23
Isophorone
Concen: 8.473 ng/ul
RT: 9.599 min Scan# 1124
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion: 82 Resp: 250593
Ion Ratio Lower Upper
82 100
95 6.3 5.4 8.0
138 19.5 15.2 22.8









#25
2-Nitrophenol
Concen: 9.004 ng/ul
RT: 9.781 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:139 Resp: 71083
Ion Ratio Lower Upper
139 100
65 45.7 36.3 54.5
109 25.8 21.2 31.8



#26
2,4-Dimethylphenol
Concen: 9.608 ng/ul
RT: 9.840 min Scan# 1165
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:107 Resp: 150430
Ion Ratio Lower Upper
107 100
121 50.0 43.0 64.4
122 88.0 71.4 107.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 9.575 ng/ul
 RT: 10.081 min Scan# 1206
 Delta R.T. -0.000 min
 Lab File: BP012035.D
 Acq: 11 Oct 2022 12:01




Instrument :
 BNA_P
 ClientSampleId :
 SSTD010632


Tgt Ion: 93 Resp: 180963
 Ion Ratio Lower Upper
 93 100
 95 31.9 25.7 38.5
 123 11.6 9.2 13.8



#29
 2,4-Dichlorophenol
 Concen: 9.746 ng/ul
 RT: 10.316 min Scan# 1246
 Delta R.T. -0.000 min
 Lab File: BP012035.D
 Acq: 11 Oct 2022 12:01

Tgt Ion: 162 Resp: 129946
 Ion Ratio Lower Upper
 162 100
 164 66.5 51.5 77.3
 98 33.7 26.3 39.5









#30
Naphthalene
Concen: 9.980 ng/ul
RT: 10.716 min Scan# 1314
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:128 Resp: 496757
Ion Ratio Lower Upper
128 100
129 10.7 8.9 13.3
127 12.9 10.2 15.4



#32
4-Chloroaniline
Concen: 9.062 ng/ul
RT: 10.834 min Scan# 1334
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:127 Resp: 186076
Ion Ratio Lower Upper
127 100
129 32.2 25.6 38.4









#33
Hexachlorobutadiene
Concen: 10.053 ng/ul
RT: 10.998 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:225 Resp: 81372
Ion Ratio Lower Upper
225 100
223 61.0 50.9 76.3
227 61.9 51.1 76.7



#34
Caprolactam
Concen: 6.048 ng/ul
RT: 11.628 min Scan# 1469
Delta R.T. -0.006 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:113 Resp: 26396
Ion Ratio Lower Upper
113 100
55 157.5 115.2 172.8
56 115.0 85.6 128.4









#35
4-Chloro-3-methylphenol
Concen: 8.728 ng/ul
RT: 11.957 min Scan# 1525
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:107 Resp: 119284
Ion Ratio Lower Upper
107 100
144 27.1 23.0 34.4
142 87.2 69.8 104.8



#36
2-Methylnaphthalene
Concen: 9.682 ng/ul
RT: 12.328 min Scan# 1588
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:142 Resp: 307183
Ion Ratio Lower Upper
142 100
141 88.0 70.7 106.1





#37

1-Methylnaphthalene

Concen: 9.777 ng/ul

RT: 12.545 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:142 Resp: 311882

Ion Ratio Lower Upper

142 100

141 91.4 72.9 109.3

116 3.6 2.8 4.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 10.454 ng/ul

RT: 12.692 min Scan# 1650

Delta R.T. -0.006 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01

Tgt Ion:216 Resp: 152929




Ion Ratio Lower Upper


216 100

214 80.4 63.0 94.4

179 21.3 16.3 24.5

108 19.3 12.5 18.7#









#40
Hexachlorocyclopentadiene
Concen: 7.808 ng/ul
RT: 12.675 min Scan# 1647
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:237	Resp:	52943
Ion Ratio	Lower	Upper
237	100	
235	60.1	49.8 74.8
272	12.6	9.9 14.9



#41
2,4,6-Trichlorophenol
Concen: 9.175 ng/ul
RT: 12.940 min Scan# 1692
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:196	Resp:	84680
Ion Ratio	Lower	Upper
196	100	
198	98.1	76.2 114.2
200	31.9	24.2 36.2









#42
2,4,5-Trichlorophenol
Concen: 9.732 ng/ul
RT: 13.010 min Scan# 1704
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:196 Resp: 95678
Ion Ratio Lower Upper
196 100
198 92.2 75.4 113.2
200 30.5 24.8 37.2



#43
1,1'-Biphenyl
Concen: 10.185 ng/ul
RT: 13.339 min Scan# 1760
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:154 Resp: 389352
Ion Ratio Lower Upper
154 100
153 39.9 32.2 48.2
76 12.9 10.3 15.5









#44
2-Chloronaphthalene
Concen: 10.237 ng/ul
RT: 13.381 min Scan# 1767
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:162 Resp: 313909
Ion Ratio Lower Upper
162 100
164 31.7 26.0 39.0
127 36.6 29.4 44.2



#45
2-Nitroaniline
Concen: 8.154 ng/ul
RT: 13.592 min Scan# 1803
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion: 65 Resp: 56989
Ion Ratio Lower Upper
65 100
92 73.5 61.6 92.4
138 112.8 96.5 144.7









#47
Dimethylphthalate
Concen: 9.347 ng/ul
RT: 13.963 min Scan# 1866
Delta R.T. -0.006 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:163 Resp: 334425
Ion Ratio Lower Upper
163 100
194 3.8 3.1 4.7
164 10.7 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 8.371 ng/ul
RT: 14.086 min Scan# 1887
Delta R.T. -0.006 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:165 Resp: 49726
Ion Ratio Lower Upper
165 100
89 58.3 43.4 65.0
121 25.6 18.1 27.1









#50
Acenaphthylene
Concen: 9.613 ng/ul
RT: 14.228 min Scan# 1911
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:152 Resp: 448602
Ion Ratio Lower Upper
152 100
151 19.9 15.7 23.5
153 12.9 10.4 15.6



#51
3-Nitroaniline
Concen: 7.185 ng/ul
RT: 14.422 min Scan# 1944
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:138 Resp: 52680
Ion Ratio Lower Upper
138 100
108 13.4 8.5 12.7#
92 113.9 85.8 128.8









#52
Acenaphthene
Concen: 9.856 ng/ul
RT: 14.569 min Scan# 1969
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:153 Resp: 325328
Ion Ratio Lower Upper
153 100
152 47.3 38.2 57.2
154 86.6 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 5.348 ng/ul
RT: 14.628 min Scan# 1979
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:184 Resp: 22271
Ion Ratio Lower Upper
184 100
63 62.4 45.5 68.3
154 61.3 47.0 70.6









#55
4-Nitrophenol
Concen: 6.722 ng/ul
RT: 14.728 min Scan# 1996
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:109 Resp: 36635
Ion Ratio Lower Upper
109 100
139 153.0 124.1 186.1
65 135.4 104.2 156.4



#56
Dibenzofuran
Concen: 9.916 ng/ul
RT: 14.904 min Scan# 2026
Delta R.T. -0.006 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:168 Resp: 441223
Ion Ratio Lower Upper
168 100
139 35.7 28.2 42.4





#57

2,4-Dinitrotoluene

Concen: 7.931 ng/ul

RT: 14.875 min Scan# 2021

Delta R.T. -0.000 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:165 Resp: 75675

Ion Ratio Lower Upper

165 100

63 42.6 31.1 46.7

182 2.8 1.8 2.6#



#58

2,3,4,6-Tetrachlorophenol

Concen: 8.724 ng/ul

RT: 15.134 min Scan# 2065

Delta R.T. -0.000 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01

Tgt Ion:232 Resp: 71768




Ion Ratio Lower Upper


232 100

131 49.5 37.8 56.6

130 2.4 1.8 2.6

166 34.8 26.6 39.8









#59
Diethylphthalate
Concen: 8.623 ng/ul
RT: 15.334 min Scan# 2099
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:149 Resp: 301720
Ion Ratio Lower Upper
149 100
177 22.2 17.5 26.3
150 12.4 9.8 14.8



#61
Fluorene
Concen: 9.724 ng/ul
RT: 15.557 min Scan# 2137
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:166 Resp: 350965
Ion Ratio Lower Upper
166 100
165 97.2 79.5 119.3
167 13.6 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 9.900 ng/ul
RT: 15.551 min Scan# 2136
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:204 Resp: 164555
Ion Ratio Lower Upper
204 100
206 32.4 25.4 38.2
141 56.9 46.2 69.2



#63
4-Nitroaniline
Concen: 5.950 ng/ul
RT: 15.586 min Scan# 2142
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:138 Resp: 42111
Ion Ratio Lower Upper
138 100
92 55.0 39.2 58.8
108 74.3 55.4 83.0









#66
4,6-Dinitro-2-methylphenol
Concen: 6.974 ng/ul
RT: 15.639 min Scan# 2151
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:198 Resp: 39482
Ion Ratio Lower Upper
198 100
182 9.7 4.9 7.3#
77 26.1 20.0 30.0



#67
N-Nitrosodiphenylamine
Concen: 10.271 ng/ul
RT: 15.769 min Scan# 2173
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:169 Resp: 282145
Ion Ratio Lower Upper
169 100
168 65.8 53.5 80.3
167 35.9 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 9.860 ng/ul
RT: 16.451 min Scan# 2289
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:248 Resp: 87588
Ion Ratio Lower Upper
248 100
250 96.7 77.9 116.9
141 75.9 59.5 89.3



#69
Hexachlorobenzene
Concen: 9.873 ng/ul
RT: 16.563 min Scan# 2308
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:284 Resp: 101958
Ion Ratio Lower Upper
284 100
142 39.5 30.5 45.7
249 32.5 25.4 38.0





#70

Atrazine

Concen: 8.281 ng/ul

RT: 16.728 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:200 Resp: 75687

Ion Ratio Lower Upper

200 100

173 26.1 22.1 33.1

215 49.2 38.8 58.2



#71

Pentachlorophenol

Concen: 7.334 ng/ul

RT: 16.916 min Scan# 2368

Delta R.T. -0.000 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01




Tgt Ion:266 Resp: 47099


Ion Ratio Lower Upper

266 100

264 62.3 50.6 75.8

268 61.3 51.8 77.6





#72

Phenanthrene

Concen: 9.837 ng/ul

RT: 17.310 min Scan# 2435

Delta R.T. -0.000 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:178 Resp: 528467

Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.4

176 18.8 15.8 23.6



#74

Anthracene

Concen: 9.592 ng/ul

RT: 17.398 min Scan# 2450

Delta R.T. -0.000 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01




Tgt Ion:178 Resp: 518779


Ion Ratio Lower Upper

178 100

179 15.6 12.2 18.4

176 18.6 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 11.199 ng/uL
RT: 13.304 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:216 Resp: 150745
Ion Ratio Lower Upper
216 100
214 78.2 62.6 94.0
218 46.9 38.9 58.3



#76
Pentachlorobenzene
Concen: 10.836 ng/uL
RT: 14.822 min Scan# 2012
Delta R.T. -0.006 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:250 Resp: 146708
Ion Ratio Lower Upper
250 100
248 63.6 51.4 77.2
252 64.2 50.7 76.1





#77

Carbazole

Concen: 9.110 ng/ul

RT: 17.675 min Scan# 2497

Delta R.T. -0.000 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:167 Resp: 460291

Ion Ratio Lower Upper

167 100

166 21.6 17.0 25.6

139 12.8 10.1 15.1



#78

Di-n-butylphthalate

Concen: 7.778 ng/ul

RT: 18.222 min Scan# 2590

Delta R.T. -0.000 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01




Tgt Ion:149 Resp: 414819


Ion Ratio Lower Upper

149 100

150 9.0 7.2 10.8

104 4.5 3.4 5.2









#80
Fluoranthene
Concen: 10.294 ng/ul
RT: 19.274 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:202 Resp: 556602
Ion Ratio Lower Upper
202 100
101 13.0 9.0 13.6
100 9.5 7.0 10.6



#82
Pyrene
Concen: 10.428 ng/ul
RT: 19.627 min Scan# 2829
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:202 Resp: 604168
Ion Ratio Lower Upper
202 100
101 14.5 10.4 15.6
100 12.1 8.5 12.7









#83
Butylbenzylphthalate
Concen: 8.117 ng/ul
RT: 20.498 min Scan# 2978
Delta R.T. -0.006 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:149 Resp: 175387
Ion Ratio Lower Upper
149 100
91 71.0 48.5 72.7
206 17.8 17.9 26.9#



#84
3,3'-Dichlorobenzidine
Concen: 9.141 ng/ul
RT: 21.274 min Scan# 3109
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:252 Resp: 177396
Ion Ratio Lower Upper
252 100
254 62.5 51.4 77.2
126 12.1 7.8 11.8#





#85

Benzo(a)anthracene

Concen: 9.759 ng/ul

RT: 21.339 min Scan# 3120

Delta R.T. -0.000 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:228 Resp: 566451

Ion Ratio Lower Upper

228 100

229 20.1 15.7 23.5

226 27.3 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 7.734 ng/ul

RT: 21.257 min Scan# 3106

Delta R.T. -0.000 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01




Tgt Ion:149 Resp: 238878


Ion Ratio Lower Upper

149 100

167 26.7 21.9 32.9

279 3.4 3.4 5.2#





#87

Chrysene

Concen: 9.880 ng/ul

RT: 21.392 min Scan# 3129

Delta R.T. -0.000 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:228 Resp: 549963

Ion Ratio Lower Upper

228 100

226 30.9 24.0 36.0

229 19.7 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 7.326 ng/ul




RT: 22.186 min Scan# 3264


Delta R.T. -0.000 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01

Tgt Ion:149 Resp: 412124









#90
Benzo(b)fluoranthene
Concen: 9.833 ng/ul
RT: 23.033 min Scan# 3408
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:252 Resp: 583806
Ion Ratio Lower Upper
252 100
253 21.7 17.6 26.4
125 11.9 7.8 11.8#



#91
Benzo(k)fluoranthene
Concen: 9.856 ng/ul
RT: 23.080 min Scan# 3416
Delta R.T. -0.000 min
Lab File: BP012035.D
Acq: 11 Oct 2022 12:01

Tgt Ion:252 Resp: 574436
Ion Ratio Lower Upper
252 100
253 22.5 17.6 26.4
125 11.9 8.2 12.2





#93

Benzo(a)pyrene

Concen: 9.770 ng/ul

RT: 23.662 min Scan# 3515

Delta R.T. -0.000 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:252 Resp: 519721

Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4

125 13.3 8.9 13.3#



#94

Indeno(1,2,3-cd)pyrene

Concen: 9.460 ng/ul

RT: 26.286 min Scan# 3961

Delta R.T. -0.006 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01




Tgt Ion:276 Resp: 635576


Ion Ratio Lower Upper

276 100

138 29.3 19.1 28.7#

277 24.9 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 9.572 ng/ul

RT: 26.304 min Scan# 3964

Delta R.T. -0.006 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:278 Resp: 575787

Ion Ratio Lower Upper

278 100

139 19.1 12.5 18.7#

279 23.8 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 9.443 ng/ul

RT: 27.062 min Scan# 4093

Delta R.T. -0.000 min

Lab File: BP012035.D

Acq: 11 Oct 2022 12:01

Tgt Ion:276 Resp: 570998

Ion Ratio Lower Upper

276 100

138 25.6 16.5 24.7#

277 24.5 19.0 28.4

