

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BP012036.D
 Acq On : 11 Oct 2022 12:35
 Operator : CG/JU
 Sample : SST02070
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST020633

Quant Time: Oct 11 23:04:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 11 23:01:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.858	152	271188	20.000	ng/ul	0.00
20) Naphthalene-d8	10.669	136	1171866	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.510	164	673759	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.269	188	1197570	20.000	ng/ul	0.00
79) Chrysene-d12	21.351	240	1230784	20.000	ng/ul	0.00
88) Perylene-d12	23.769	264	1284406	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	54466	18.968	ng/uL	0.00
4) Pyridine-d5	3.705	84	366791	39.330	ng/ul	0.00
7) Phenol-d5	7.028	99	452056	19.609	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.193	67	267909	19.660	ng/ul	0.00
11) 2-Chlorophenol-d4	7.387	132	367818	20.788	ng/ul	0.00
15) 4-Methylphenol-d8	8.575	113	355299	20.501	ng/ul	0.00
21) Nitrobenzene-d5	9.028	128	176370	20.772	ng/ul	0.00
24) 2-Nitrophenol-d4	9.752	143	178971	20.783	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.287	165	344243	21.251	ng/ul	0.00
31) 4-Chloroaniline-d4	10.810	131	505169	20.117	ng/ul	0.00
46) Dimethylphthalate-d6	13.922	166	890731	19.929	ng/ul	0.00
49) Acenaphthylene-d8	14.198	160	1124620	20.517	ng/ul	0.00
54) 4-Nitrophenol-d4	14.716	143	157199	16.379	ng/ul	0.00
60) Fluorene-d10	15.504	176	797987	19.841	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.628	200	118627	18.283	ng/ul	0.00
73) Anthracene-d10	17.363	188	1097348	20.226	ng/ul	0.00
81) Pyrene-d10	19.598	212	1360604	22.494	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.616	264	1291718	20.378	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.317	88	58498	19.553	ng/uL	100
5) Pyridine	3.723	79	358844	38.346	ng/ul	100
6) Benzaldehyde	6.999	77	258750	24.210	ng/ul	100
8) Phenol	7.052	94	481315	19.827	ng/ul	100
10) Bis(2-Chloroethyl)ether	7.287	93	387253	20.407	ng/ul	100
12) 2-Chlorophenol	7.422	128	396408	20.646	ng/ul	100
13) 2-Methylphenol	8.305	108	359433	20.283	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.393	45	412654	18.569	ng/ul	100
16) Acetophenone	8.687	105	575317	20.822	ng/ul	100
17) N-Nitroso-di-n-propyla...	8.675	70	264775	21.727	ng/ul	100
18) 4-Methylphenol	8.640	108	395458	20.652	ng/ul	100
19) Hexachloroethane	8.940	117	149312	18.820	ng/ul	100
22) Nitrobenzene	9.069	77	403720	19.254	ng/ul	100
23) Isophorone	9.599	82	741554	19.731	ng/ul	100
25) 2-Nitrophenol	9.781	139	213332	21.265	ng/ul	100
26) 2,4-Dimethylphenol	9.840	107	410330	20.624	ng/ul	100
27) Bis(2-Chloroethoxy)met...	10.081	93	499106	20.781	ng/ul	100
29) 2,4-Dichlorophenol	10.316	162	363514	21.455	ng/ul	100
30) Naphthalene	10.716	128	1288576	20.372	ng/ul	100
32) 4-Chloroaniline	10.834	127	523693	20.070	ng/ul	100
33) Hexachlorobutadiene	10.999	225	213764	20.783	ng/ul	100
34) Caprolactam	11.634	113	91649	16.524	ng/ul	100
35) 4-Chloro-3-methylphenol	11.957	107	356770	20.543	ng/ul	100
36) 2-Methylnaphthalene	12.328	142	848953	21.057	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BP012036.D
 Acq On : 11 Oct 2022 12:35
 Operator : CG/JU
 Sample : SSTD02070
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD020633


Quant Time: Oct 11 23:04:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 11 23:01:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.546	142	845083	20.848	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.698	216	418334	21.521	ng/ul	100
40) Hexachlorocyclopentadiene	12.675	237	180616	20.048	ng/ul	100
41) 2,4,6-Trichlorophenol	12.940	196	263406	21.479	ng/ul	100
42) 2,4,5-Trichlorophenol	13.010	196	285987	21.893	ng/ul	100
43) 1,1'-Biphenyl	13.340	154	1067203	21.010	ng/ul	100
44) 2-Chloronaphthalene	13.381	162	846005	20.763	ng/ul	100
45) 2-Nitroaniline	13.593	65	187024	20.140	ng/ul	100
47) Dimethylphthalate	13.969	163	938769	19.746	ng/ul	100
48) 2,6-Dinitrotoluene	14.093	165	170264	21.572	ng/ul	100
50) Acenaphthylene	14.228	152	1264697	20.395	ng/ul	100
51) 3-Nitroaniline	14.422	138	191082	19.614	ng/ul	100
52) Acenaphthene	14.569	153	883859	20.152	ng/ul	100
53) 2,4-Dinitrophenol	14.628	184	83068	15.012	ng/ul	100
55) 4-Nitrophenol	14.728	109	110546	15.265	ng/ul	100
56) Dibenzofuran	14.910	168	1190537	20.136	ng/ul	100
57) 2,4-Dinitrotoluene	14.875	165	242472	19.124	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.134	232	217541	19.901	ng/ul	100
59) Diethylphthalate	15.334	149	856790	18.428	ng/ul	100
61) Fluorene	15.557	166	940795	19.617	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.551	204	454343	20.571	ng/ul	100
63) 4-Nitroaniline	15.587	138	174103	18.514	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.640	198	131515	18.595	ng/ul	100
67) N-Nitrosodiphenylamine	15.769	169	776178	22.618	ng/ul	100
68) 4-Bromophenyl-phenylether	16.451	248	247111	22.267	ng/ul	100
69) Hexachlorobenzene	16.563	284	278460	21.584	ng/ul	100
70) Atrazine	16.728	200	214410	18.777	ng/ul	100
71) Pentachlorophenol	16.916	266	145029	18.077	ng/ul	100
72) Phenanthrene	17.310	178	1361554	20.286	ng/ul	100
74) Anthracene	17.398	178	1367862	20.243	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.304	216	416106	24.744	ng/ul	100
76) Pentachlorobenzene	14.828	250	407071	24.065	ng/ul	100
77) Carbazole	17.675	167	1178158	18.664	ng/ul	100
78) Di-n-butylphthalate	18.222	149	1322672	19.850	ng/ul	100
80) Fluoranthene	19.275	202	1664275	22.423	ng/ul	100
82) Pyrene	19.628	202	1741147	21.891	ng/ul	100
83) Butylbenzylphthalate	20.504	149	494907	16.684	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.275	252	538577	20.217	ng/ul	100
85) Benzo(a)anthracene	21.339	228	1579953	19.829	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.257	149	653599	15.415	ng/ul	100
87) Chrysene	21.392	228	1522970	19.930	ng/ul	100
89) Di-n-octyl phthalate	22.186	149	1098137	14.097	ng/ul	100
90) Benzo(b)fluoranthene	23.033	252	1682423	20.463	ng/ul	100
91) Benzo(k)fluoranthene	23.080	252	1619511	20.066	ng/ul	100
93) Benzo(a)pyrene	23.663	252	1485574	20.168	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.292	276	1932343	20.770	ng/ul	100
95) Dibenzo(a,h)anthracene	26.310	278	1745091	20.951	ng/ul	100
96) Benzo(g,h,i)perylene	27.062	276	1744319	20.831	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 11 23:04:52 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 11 23:01:01 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 19.553 ng/uL
 RT: 3.317 min Scan# 50
 Delta R.T. 0.000 min
 Lab File: BP012036.D
 Acq: 11 Oct 2022 12:35




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020633


Tgt Ion	88	Ratio	100	Lower	Upper
88	100				
43	20.8		16.6	25.0	
58	65.8		52.6	79.0	



#5
 Pyridine
 Concen: 38.346 ng/uL
 RT: 3.723 min Scan# 125
 Delta R.T. 0.000 min
 Lab File: BP012036.D
 Acq: 11 Oct 2022 12:35

Tgt Ion	79	Ratio	100	Lower	Upper
79	100				
52	64.0		51.2	76.8	
51	28.5		22.8	34.2	









#6
Benzaldehyde
Concen: 24.210 ng/ul
RT: 6.999 min Scan# 682
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion: 77 Resp: 258750
Ion Ratio Lower Upper
77 100
105 106.5 85.2 127.8
106 103.9 83.1 124.7



#8
Phenol
Concen: 19.827 ng/ul
RT: 7.052 min Scan# 691
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion: 94 Resp: 481315
Ion Ratio Lower Upper
94 100
65 24.2 19.4 29.0
66 30.8 24.6 37.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 20.407 ng/ul
 RT: 7.287 min Scan# 731
 Delta R.T. 0.000 min
 Lab File: BP012036.D
 Acq: 11 Oct 2022 12:35




Instrument :
 BNA_P
 ClientSampleId :
 SST020633


Tgt Ion: 93 Resp: 387253
 Ion Ratio Lower Upper
 93 100
 63 70.9 56.7 85.1
 95 32.7 26.2 39.2



#12
 2-Chlorophenol
 Concen: 20.646 ng/ul
 RT: 7.422 min Scan# 754
 Delta R.T. 0.000 min
 Lab File: BP012036.D
 Acq: 11 Oct 2022 12:35

Tgt Ion:128 Resp: 396408
 Ion Ratio Lower Upper
 128 100
 64 40.9 32.7 49.1
 130 32.9 26.3 39.5





#13

2-Methylphenol

Concen: 20.283 ng/ul

RT: 8.305 min Scan# 904

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35

Instrument :

BNA_P

ClientSampleId :





SSTD020633

Tgt Ion:108 Resp: 359433

Ion Ratio Lower Upper

108 100

107 89.9 71.9 107.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 18.569 ng/ul

RT: 8.393 min Scan# 919

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35




Tgt Ion: 45 Resp: 412654


Ion Ratio Lower Upper

45 100

77 18.0 14.4 21.6

79 14.1 11.3 16.9









#16
Acetophenone
Concen: 20.822 ng/ul
RT: 8.687 min Scan# 969
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:105 Resp: 575317
Ion Ratio Lower Upper
105 100
77 71.2 57.0 85.4
51 20.5 16.4 24.6



#17
N-Nitroso-di-n-propylamine
Concen: 21.727 ng/ul
RT: 8.675 min Scan# 967
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion: 70 Resp: 264775
Ion Ratio Lower Upper
70 100
42 43.3 34.6 52.0
101 10.0 8.0 12.0
130 22.9 18.3 27.5









#18
4-Methylphenol
Concen: 20.652 ng/ul
RT: 8.640 min Scan# 961
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:108 Resp: 395458
Ion Ratio Lower Upper
108 100
107 107.9 86.3 129.5



#19
Hexachloroethane
Concen: 18.820 ng/ul
RT: 8.940 min Scan# 1012
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:117 Resp: 149312
Ion Ratio Lower Upper
117 100
201 102.9 82.3 123.5
199 62.5 50.0 75.0









#22
Nitrobenzene
Concen: 19.254 ng/ul
RT: 9.069 min Scan# 1034
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion: 77 Resp: 403720
Ion Ratio Lower Upper
77 100
123 49.2 39.4 59.0
65 12.4 9.9 14.9



#23
Isophorone
Concen: 19.731 ng/ul
RT: 9.599 min Scan# 1124
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion: 82 Resp: 741554
Ion Ratio Lower Upper
82 100
95 6.7 5.4 8.0
138 19.0 15.2 22.8









#25
2-Nitrophenol
Concen: 21.265 ng/ul
RT: 9.781 min Scan# 1155
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:139 Resp: 213332
Ion Ratio Lower Upper
139 100
65 45.4 36.3 54.5
109 26.5 21.2 31.8



#26
2,4-Dimethylphenol
Concen: 20.624 ng/ul
RT: 9.840 min Scan# 1165
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:107 Resp: 410330
Ion Ratio Lower Upper
107 100
121 53.7 43.0 64.4
122 89.3 71.4 107.2





#27

Bis(2-Chloroethoxy)methane

Concen: 20.781 ng/ul

RT: 10.081 min Scan# 1206

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion: 93 Resp: 499106

Ion Ratio Lower Upper

93 100

95 32.1 25.7 38.5

123 11.5 9.2 13.8



#29

2,4-Dichlorophenol

Concen: 21.455 ng/ul

RT: 10.316 min Scan# 1246

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35




Tgt Ion:162 Resp: 363514


Ion Ratio Lower Upper

162 100

164 64.4 51.5 77.3

98 32.9 26.3 39.5









#30
Naphthalene
Concen: 20.372 ng/ul
RT: 10.716 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:128 Resp: 1288576
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 12.8 10.2 15.4



#32
4-Chloroaniline
Concen: 20.070 ng/ul
RT: 10.834 min Scan# 1334
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:127 Resp: 523693
Ion Ratio Lower Upper
127 100
129 32.0 25.6 38.4









#33
Hexachlorobutadiene
Concen: 20.783 ng/ul
RT: 10.999 min Scan# 1362
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:225 Resp: 213764
Ion Ratio Lower Upper
225 100
223 63.6 50.9 76.3
227 63.9 51.1 76.7



#34
Caprolactam
Concen: 16.524 ng/ul
RT: 11.634 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:113 Resp: 91649
Ion Ratio Lower Upper
113 100
55 144.0 115.2 172.8
56 107.0 85.6 128.4









#35
4-Chloro-3-methylphenol
Concen: 20.543 ng/ul
RT: 11.957 min Scan# 1525
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:107 Resp: 356770
Ion Ratio Lower Upper
107 100
144 28.7 23.0 34.4
142 87.3 69.8 104.8



#36
2-Methylnaphthalene
Concen: 21.057 ng/ul
RT: 12.328 min Scan# 1588
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:142 Resp: 848953
Ion Ratio Lower Upper
142 100
141 88.4 70.7 106.1









#37
1-Methylnaphthalene
Concen: 20.848 ng/ul
RT: 12.546 min Scan# 1625
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:142 Resp: 845083
Ion Ratio Lower Upper
142 100
141 91.1 72.9 109.3
116 3.5 2.8 4.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.521 ng/ul
RT: 12.698 min Scan# 1651
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:216 Resp: 418334
Ion Ratio Lower Upper
216 100
214 78.7 63.0 94.4
179 20.4 16.3 24.5
108 15.6 12.5 18.7









#40
Hexachlorocyclopentadiene
Concen: 20.048 ng/ul
RT: 12.675 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:237 Resp: 180616
Ion Ratio Lower Upper
237 100
235 62.3 49.8 74.8
272 12.4 9.9 14.9



#41
2,4,6-Trichlorophenol
Concen: 21.479 ng/ul
RT: 12.940 min Scan# 1692
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:196 Resp: 263406
Ion Ratio Lower Upper
196 100
198 95.2 76.2 114.2
200 30.2 24.2 36.2





#42

2,4,5-Trichlorophenol

Concen: 21.893 ng/ul

RT: 13.010 min Scan# 1704

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:196 Resp: 285987

Ion Ratio Lower Upper

196 100

198 94.3 75.4 113.2

200 31.0 24.8 37.2



#43

1,1'-Biphenyl

Concen: 21.010 ng/ul

RT: 13.340 min Scan# 1760

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35




Tgt Ion:154 Resp: 1067203


Ion Ratio Lower Upper

154 100

153 40.2 32.2 48.2

76 12.9 10.3 15.5





#44

2-Chloronaphthalene

Concen: 20.763 ng/ul

RT: 13.381 min Scan# 1767

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:162 Resp: 846005

Ion Ratio Lower Upper

162 100

164 32.5 26.0 39.0

127 36.8 29.4 44.2



#45

2-Nitroaniline

Concen: 20.140 ng/ul

RT: 13.593 min Scan# 1803

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35




Tgt Ion: 65 Resp: 187024


Ion Ratio Lower Upper

65 100

92 77.0 61.6 92.4

138 120.6 96.5 144.7









#47
Dimethylphthalate
Concen: 19.746 ng/ul
RT: 13.969 min Scan# 1867
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:163 Resp: 938769
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 9.9 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 21.572 ng/ul
RT: 14.093 min Scan# 1888
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:165 Resp: 170264
Ion Ratio Lower Upper
165 100
89 54.2 43.4 65.0
121 22.6 18.1 27.1





#50

Acenaphthylene

Concen: 20.395 ng/ul

RT: 14.228 min Scan# 1911

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:152 Resp: 1264697

Ion Ratio Lower Upper

152 100

151 19.6 15.7 23.5

153 13.0 10.4 15.6



#51

3-Nitroaniline

Concen: 19.614 ng/ul

RT: 14.422 min Scan# 1944

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35




Tgt Ion:138 Resp: 191082


Ion Ratio Lower Upper

138 100

108 10.6 8.5 12.7

92 107.3 85.8 128.8









#52
Acenaphthene
Concen: 20.152 ng/ul
RT: 14.569 min Scan# 1969
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:153 Resp: 883859
Ion Ratio Lower Upper
153 100
152 47.7 38.2 57.2
154 88.2 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 15.012 ng/ul
RT: 14.628 min Scan# 1979
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:184 Resp: 83068
Ion Ratio Lower Upper
184 100
63 56.9 45.5 68.3
154 58.8 47.0 70.6









#55
4-Nitrophenol
Concen: 15.265 ng/ul
RT: 14.728 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:109 Resp: 110546
Ion Ratio Lower Upper
109 100
139 155.1 124.1 186.1
65 130.3 104.2 156.4



#56
Dibenzofuran
Concen: 20.136 ng/ul
RT: 14.910 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:168 Resp: 1190537
Ion Ratio Lower Upper
168 100
139 35.3 28.2 42.4









#57
2,4-Dinitrotoluene
Concen: 19.124 ng/ul
RT: 14.875 min Scan# 2021
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:165 Resp: 242472
Ion Ratio Lower Upper
165 100
63 38.9 31.1 46.7
182 2.2 1.8 2.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.901 ng/ul
RT: 15.134 min Scan# 2065
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:232 Resp: 217541
Ion Ratio Lower Upper
232 100
131 47.2 37.8 56.6
130 2.2 1.8 2.6
166 33.2 26.6 39.8









#59
Diethylphthalate
Concen: 18.428 ng/ul
RT: 15.334 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:149 Resp: 856790
Ion Ratio Lower Upper
149 100
177 21.9 17.5 26.3
150 12.3 9.8 14.8



#61
Fluorene
Concen: 19.617 ng/ul
RT: 15.557 min Scan# 2137
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:166 Resp: 940795
Ion Ratio Lower Upper
166 100
165 99.4 79.5 119.3
167 13.6 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 20.571 ng/ul
RT: 15.551 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:204 Resp: 454343
Ion Ratio Lower Upper
204 100
206 31.8 25.4 38.2
141 57.7 46.2 69.2



#63
4-Nitroaniline
Concen: 18.514 ng/ul
RT: 15.587 min Scan# 2142
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:138 Resp: 174103
Ion Ratio Lower Upper
138 100
92 49.0 39.2 58.8
108 69.2 55.4 83.0









#66
4,6-Dinitro-2-methylphenol
Concen: 18.595 ng/ul
RT: 15.640 min Scan# 2151
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:198 Resp: 131515
Ion Ratio Lower Upper
198 100
182 6.1 4.9 7.3
77 25.0 20.0 30.0



#67
N-Nitrosodiphenylamine
Concen: 22.618 ng/ul
RT: 15.769 min Scan# 2173
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:169 Resp: 776178
Ion Ratio Lower Upper
169 100
168 66.9 53.5 80.3
167 36.9 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 22.267 ng/ul
RT: 16.451 min Scan# 2289
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:248 Resp: 247111
Ion Ratio Lower Upper
248 100
250 97.4 77.9 116.9
141 74.4 59.5 89.3



#69
Hexachlorobenzene
Concen: 21.584 ng/ul
RT: 16.563 min Scan# 2308
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:284 Resp: 278460
Ion Ratio Lower Upper
284 100
142 38.1 30.5 45.7
249 31.7 25.4 38.0





#70

Atrazine

Concen: 18.777 ng/ul

RT: 16.728 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:200 Resp: 214410

Ion Ratio Lower Upper

200 100

173 27.6 22.1 33.1

215 48.5 38.8 58.2



#71

Pentachlorophenol

Concen: 18.077 ng/ul

RT: 16.916 min Scan# 2368

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35




Tgt Ion:266 Resp: 145029


Ion Ratio Lower Upper

266 100

264 63.2 50.6 75.8

268 64.7 51.8 77.6





#72

Phenanthrene

Concen: 20.286 ng/ul

RT: 17.310 min Scan# 2435

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:178 Resp: 1361554

Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 19.7 15.8 23.6



#74

Anthracene

Concen: 20.243 ng/ul

RT: 17.398 min Scan# 2450

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35




Tgt Ion:178 Resp: 1367862


Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 18.5 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 24.744 ng/uL
RT: 13.304 min Scan# 1754
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:216 Resp: 416106
Ion Ratio Lower Upper
216 100
214 78.3 62.6 94.0
218 48.6 38.9 58.3



#76
Pentachlorobenzene
Concen: 24.065 ng/uL
RT: 14.828 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:250 Resp: 407071
Ion Ratio Lower Upper
250 100
248 64.3 51.4 77.2
252 63.4 50.7 76.1









#77
 Carbazole
 Concen: 18.664 ng/ul
 RT: 17.675 min Scan# 2497
 Delta R.T. 0.000 min
 Lab File: BP012036.D
 Acq: 11 Oct 2022 12:35




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020633


Tgt Ion:167 Resp: 1178158
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.0 25.6
 139 12.6 10.1 15.1



#78
 Di-n-butylphthalate
 Concen: 19.850 ng/ul
 RT: 18.222 min Scan# 2590
 Delta R.T. 0.000 min
 Lab File: BP012036.D
 Acq: 11 Oct 2022 12:35

Tgt Ion:149 Resp: 1322672
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.2 10.8
 104 4.3 3.4 5.2









#80
Fluoranthene
Concen: 22.423 ng/ul
RT: 19.275 min Scan# 2769
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:202 Resp: 1664275
Ion Ratio Lower Upper
202 100
101 11.3 9.0 13.6
100 8.8 7.0 10.6



#82
Pyrene
Concen: 21.891 ng/ul
RT: 19.628 min Scan# 2829
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:202 Resp: 1741147
Ion Ratio Lower Upper
202 100
101 13.0 10.4 15.6
100 10.6 8.5 12.7





#83

Butylbenzylphthalate

Concen: 16.684 ng/ul

RT: 20.504 min Scan# 2978

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:149 Resp: 494907

Ion Ratio Lower Upper

149 100

91 60.6 48.5 72.7

206 22.4 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 20.217 ng/ul

RT: 21.275 min Scan# 3109

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35




Tgt Ion:252 Resp: 538577


Ion Ratio Lower Upper

252 100

254 64.3 51.4 77.2

126 9.8 7.8 11.8









#85
Benzo(a)anthracene
Concen: 19.829 ng/ul
RT: 21.339 min Scan# 3120
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:228 Resp: 1579953
Ion Ratio Lower Upper
228 100
229 19.6 15.7 23.5
226 27.4 21.9 32.9



#86
Bis(2-ethylhexyl)phthalate
Concen: 15.415 ng/ul
RT: 21.257 min Scan# 3106
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:149 Resp: 653599
Ion Ratio Lower Upper
149 100
167 27.4 21.9 32.9
279 4.3 3.4 5.2





#87

Chrysene

Concen: 19.930 ng/ul

RT: 21.392 min Scan# 3129

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:228 Resp: 1522970

Ion Ratio Lower Upper

228 100

226 30.0 24.0 36.0

229 19.4 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 14.097 ng/ul




RT: 22.186 min Scan# 3264


Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35

Tgt Ion:149 Resp: 1098137









#90
Benzo(b)fluoranthene
Concen: 20.463 ng/ul
RT: 23.033 min Scan# 3408
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:252 Resp: 1682423
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 9.8 7.8 11.8



#91
Benzo(k)fluoranthene
Concen: 20.066 ng/ul
RT: 23.080 min Scan# 3416
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:252 Resp: 1619511
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 10.2 8.2 12.2





#93

Benzo(a)pyrene

Concen: 20.168 ng/ul

RT: 23.663 min Scan# 3515

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:252 Resp: 1485574

Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4

125 11.1 8.9 13.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.770 ng/ul

RT: 26.292 min Scan# 3962

Delta R.T. 0.000 min

Lab File: BP012036.D

Acq: 11 Oct 2022 12:35




Tgt Ion:276 Resp: 1932343


Ion Ratio Lower Upper

276 100

138 23.9 19.1 28.7

277 24.9 19.9 29.9









#95
Dibenzo(a,h)anthracene
Concen: 20.951 ng/ul
RT: 26.310 min Scan# 3965
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Instrument :
BNA_P
ClientSampleId :
SSTD020633

Tgt Ion:278 Resp: 1745091
Ion Ratio Lower Upper
278 100
139 15.6 12.5 18.7
279 23.7 19.0 28.4



#96
Benzo(g,h,i)perylene
Concen: 20.831 ng/ul
RT: 27.062 min Scan# 4093
Delta R.T. 0.000 min
Lab File: BP012036.D
Acq: 11 Oct 2022 12:35

Tgt Ion:276 Resp: 1744319
Ion Ratio Lower Upper
276 100
138 20.6 16.5 24.7
277 23.7 19.0 28.4

