

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BP012040.D
 Acq On : 11 Oct 2022 15:03
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV637

Quant Time: Oct 12 00:00:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 11 23:54:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.863	152	213846	20.000	ng/u1	0.00
20) Naphthalene-d8	10.675	136	891041	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.510	164	541042	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.269	188	1144142	20.000	ng/u1	0.00
79) Chrysene-d12	21.357	240	1232068	20.000	ng/u1	0.00
88) Perylene-d12	23.780	264	1291281	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	44640	8.230	ng/uL	0.00
4) Pyridine-d5	3.705	84	275845	19.123	ng/u1	0.00
7) Phenol-d5	7.028	99	351374	20.430	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.199	67	200874	20.577	ng/u1	0.00
11) 2-Chlorophenol-d4	7.393	132	286496	20.986	ng/u1	0.00
15) 4-Methylphenol-d8	8.581	113	273249	20.685	ng/u1	0.00
21) Nitrobenzene-d5	9.034	128	135133	21.474	ng/u1	0.00
24) 2-Nitrophenol-d4	9.757	143	138088	21.641	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.293	165	266277	21.532	ng/u1	0.00
31) 4-Chloroaniline-d4	10.816	131	391634	21.490	ng/u1	0.00
46) Dimethylphthalate-d6	13.928	166	749582	21.876	ng/u1	0.00
49) Acenaphthylene-d8	14.204	160	898479	21.205	ng/u1	0.00
54) 4-Nitrophenol-d4	14.722	143	153485	22.816	ng/u1	0.00
60) Fluorene-d10	15.510	176	665366	21.462	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.633	200	121004	20.029	ng/u1	0.00
73) Anthracene-d10	17.369	188	1042962	20.992	ng/u1	0.00
81) Pyrene-d10	19.604	212	1248199	19.775	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.621	264	1315146	20.951	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.316	88	48469	8.492	ng/uL	99
5) Pyridine	3.728	79	287965	20.836	ng/u1	98
6) Benzaldehyde	7.005	77	197205	24.439	ng/u1	99
8) Phenol	7.057	94	372003	20.603	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.287	93	290231	20.373	ng/u1	99
12) 2-Chlorophenol	7.422	128	307572	20.759	ng/u1	100
13) 2-Methylphenol	8.310	108	277194	20.584	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.399	45	291676	19.976	ng/u1	98
16) Acetophenone	8.693	105	433773	21.298	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.681	70	193775	20.890	ng/u1	98
18) 4-Methylphenol	8.640	108	303987	20.943	ng/u1	97
19) Hexachloroethane	8.946	117	115947	20.273	ng/u1	98
22) Nitrobenzene	9.075	77	304375	21.099	ng/u1	99
23) Isophorone	9.604	82	553007	21.612	ng/u1	98
25) 2-Nitrophenol	9.787	139	164666	21.965	ng/u1	96
26) 2,4-Dimethylphenol	9.846	107	313558	21.476	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.087	93	366838	20.891	ng/u1	99
29) 2,4-Dichlorophenol	10.322	162	279927	21.540	ng/u1	99
30) Naphthalene	10.722	128	982433	20.865	ng/u1	99
32) 4-Chloroaniline	10.840	127	407783	21.704	ng/u1	100
33) Hexachlorobutadiene	11.004	225	166134	20.615	ng/u1	99
34) Caprolactam	11.640	113	81877	22.927	ng/u1	99
35) 4-Chloro-3-methylphenol	11.963	107	279418	22.695	ng/u1	98
36) 2-Methylnaphthalene	12.334	142	647402	21.353	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BP012040.D
 Acq On : 11 Oct 2022 15:03
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SICV637


Quant Time: Oct 12 00:00:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 11 23:54:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.551	142	649999	21.407	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.698	216	326009	19.712	ng/ul	98
40) Hexachlorocyclopentadiene	12.675	237	139850	17.409	ng/ul	99
41) 2,4,6-Trichlorophenol	12.945	196	210427	21.122	ng/ul	97
42) 2,4,5-Trichlorophenol	13.016	196	226700	20.938	ng/ul	99
43) 1,1'-Biphenyl	13.345	154	827457	20.258	ng/ul	99
44) 2-Chloronaphthalene	13.387	162	663756	20.350	ng/ul	99
45) 2-Nitroaniline	13.598	65	158066	22.718	ng/ul	98
47) Dimethylphthalate	13.975	163	787597	21.816	ng/ul	99
48) 2,6-Dinitrotoluene	14.098	165	146467	22.916	ng/ul	96
50) Acenaphthylene	14.234	152	1008657	21.038	ng/ul	99
51) 3-Nitroaniline	14.428	138	171787	23.082	ng/ul	100
52) Acenaphthene	14.575	153	708334	20.754	ng/ul	99
53) 2,4-Dinitrophenol	14.634	184	81944	19.845	ng/ul	98
55) 4-Nitrophenol	14.734	109	107619	23.140	ng/ul	96
56) Dibenzofuran	14.910	168	966328	20.949	ng/ul	98
57) 2,4-Dinitrotoluene	14.881	165	223375	24.317	ng/ul#	99
58) 2,3,4,6-Tetrachlorophenol	15.139	232	191662	22.711	ng/ul	95
59) Diethylphthalate	15.339	149	762677	23.068	ng/ul	99
61) Fluorene	15.563	166	789565	21.806	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.557	204	374099	21.357	ng/ul	98
63) 4-Nitroaniline	15.592	138	173756	25.568	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.645	198	134609	20.610	ng/ul	97
67) N-Nitrosodiphenylamine	15.775	169	675915	19.959	ng/ul	99
68) 4-Bromophenyl-phenylether	16.457	248	219419	19.708	ng/ul	98
69) Hexachlorobenzene	16.569	284	257911	19.959	ng/ul	99
70) Atrazine	16.733	200	224983	21.604	ng/ul	99
71) Pentachlorophenol	16.922	266	151761	20.045	ng/ul	99
72) Phenanthrene	17.316	178	1288570	20.709	ng/ul	100
74) Anthracene	17.404	178	1296747	21.060	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.310	216	326576	17.664	ng/uL	99
76) Pentachlorobenzene	14.834	250	338621	18.649	ng/uL	99
77) Carbazole	17.680	167	1217441	22.168	ng/ul	99
78) Di-n-butylphthalate	18.227	149	1277990	23.450	ng/ul	100
80) Fluoranthene	19.280	202	1530105	19.783	ng/ul	99
82) Pyrene	19.633	202	1604299	19.600	ng/ul	97
83) Butylbenzylphthalate	20.510	149	605046	23.319	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.280	252	589394	21.709	ng/ul	99
85) Benzo(a)anthracene	21.345	228	1641089	21.161	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.263	149	860734	24.608	ng/ul	99
87) Chrysene	21.398	228	1562220	20.927	ng/ul	99
89) Di-n-octyl phthalate	22.192	149	1490977	23.063	ng/ul	100
90) Benzo(b)fluoranthene	23.039	252	1684551	20.610	ng/ul	98
91) Benzo(k)fluoranthene	23.092	252	1707710	21.241	ng/ul	99
93) Benzo(a)pyrene	23.674	252	1520242	20.847	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.303	276	1854175	20.109	ng/ul	96
95) Dibenzo(a,h)anthracene	26.321	278	1674765	20.247	ng/ul	97
96) Benzo(g,h,i)perylene	27.080	276	1664905	20.027	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 12 00:00:58 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 11 23:54:43 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 8.492 ng/uL
 RT: 3.316 min Scan# 50
 Delta R.T. -0.001 min
 Lab File: BP012040.D
 Acq: 11 Oct 2022 15:03




Instrument :
 BNA_P
 ClientSampleId :
 SICV637


Tgt Ion:	88	Resp:	48469
Ion	Ratio	Lower	Upper
88	100		
43	19.1	16.6	25.0
58	65.2	52.6	79.0



#5
 Pyridine
 Concen: 20.836 ng/uL
 RT: 3.728 min Scan# 126
 Delta R.T. 0.005 min
 Lab File: BP012040.D
 Acq: 11 Oct 2022 15:03

Tgt Ion:	79	Resp:	287965
Ion	Ratio	Lower	Upper
79	100		
52	62.4	51.2	76.8
51	29.4	22.8	34.2









#6
Benzaldehyde
Concen: 24.439 ng/ul
RT: 7.005 min Scan# 61
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion: 77 Resp: 197205
Ion Ratio Lower Upper
77 100
105 106.6 85.2 127.8
106 102.6 83.1 124.7



#8
Phenol
Concen: 20.603 ng/ul
RT: 7.057 min Scan# 692
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion: 94 Resp: 372003
Ion Ratio Lower Upper
94 100
65 25.0 19.4 29.0
66 30.9 24.6 37.0









#10
Bis(2-Chloroethyl)ether
Concen: 20.373 ng/ul
RT: 7.287 min Scan# 731
Delta R.T. -0.000 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion: 93 Resp: 290231
Ion Ratio Lower Upper
93 100
63 72.4 56.7 85.1
95 32.5 26.2 39.2



#12
2-Chlorophenol
Concen: 20.759 ng/ul
RT: 7.422 min Scan# 754
Delta R.T. -0.000 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:128 Resp: 307572
Ion Ratio Lower Upper
128 100
64 41.1 32.7 49.1
130 32.7 26.3 39.5





#13

2-Methylphenol

Concen: 20.584 ng/ul

RT: 8.310 min Scan# 904

Delta R.T. 0.006 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03

Instrument :

BNA_P

ClientSampleId :





SICV637

Tgt Ion:108 Resp: 277194

Ion Ratio Lower Upper

108 100

107 87.8 71.9 107.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.976 ng/ul

RT: 8.399 min Scan# 920

Delta R.T. 0.006 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03




Tgt Ion: 45 Resp: 291676


Ion Ratio Lower Upper

45 100

77 18.7 14.4 21.6

79 14.9 11.3 16.9









#16
Acetophenone
Concen: 21.298 ng/ul
RT: 8.693 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:105 Resp: 433773
Ion Ratio Lower Upper
105 100
77 71.8 57.0 85.4
51 20.4 16.4 24.6



#17
N-Nitroso-di-n-propylamine
Concen: 20.890 ng/ul
RT: 8.681 min Scan# 968
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion: 70 Resp: 193775
Ion Ratio Lower Upper
70 100
42 41.5 34.6 52.0
101 9.9 8.0 12.0
130 23.9 18.3 27.5









#18
4-Methylphenol
Concen: 20.943 ng/ul
RT: 8.640 min Scan# 90
Delta R.T. -0.000 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:108 Resp: 303987
Ion Ratio Lower Upper
108 100
107 111.0 86.3 129.5



#19
Hexachloroethane
Concen: 20.273 ng/ul
RT: 8.946 min Scan# 1013
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:117 Resp: 115947
Ion Ratio Lower Upper
117 100
201 104.4 82.3 123.5
199 63.9 50.0 75.0









#22
Nitrobenzene
Concen: 21.099 ng/ul
RT: 9.075 min Scan# 1035
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion: 77 Resp: 304375
Ion Ratio Lower Upper
77 100
123 50.2 39.4 59.0
65 12.1 9.9 14.9



#23
Isophorone
Concen: 21.612 ng/ul
RT: 9.604 min Scan# 1125
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion: 82 Resp: 553007
Ion Ratio Lower Upper
82 100
95 6.7 5.4 8.0
138 20.4 15.2 22.8









#25
2-Nitrophenol
Concen: 21.965 ng/ul
RT: 9.787 min Scan# 1155
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:139 Resp: 164666
Ion Ratio Lower Upper
139 100
65 41.7 36.3 54.5
109 25.3 21.2 31.8



#26
2,4-Dimethylphenol
Concen: 21.476 ng/ul
RT: 9.846 min Scan# 1166
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:107 Resp: 313558
Ion Ratio Lower Upper
107 100
121 53.3 43.0 64.4
122 89.8 71.4 107.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.891 ng/ul
 RT: 10.087 min Scan# 1206
 Delta R.T. 0.006 min
 Lab File: BP012040.D
 Acq: 11 Oct 2022 15:03




Instrument :
 BNA_P
 ClientSampleId :
 SICV637


Tgt Ion: 93 Resp: 366838
 Ion Ratio Lower Upper
 93 100
 95 31.5 25.7 38.5
 123 11.9 9.2 13.8



#29
 2,4-Dichlorophenol
 Concen: 21.540 ng/ul
 RT: 10.322 min Scan# 1247
 Delta R.T. 0.006 min
 Lab File: BP012040.D
 Acq: 11 Oct 2022 15:03

Tgt Ion: 162 Resp: 279927
 Ion Ratio Lower Upper
 162 100
 164 65.2 51.5 77.3
 98 32.0 26.3 39.5









#30
Naphthalene
Concen: 20.865 ng/ul
RT: 10.722 min Scan# 1315
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:128 Resp: 982433
Ion Ratio Lower Upper
128 100
129 10.8 8.9 13.3
127 13.0 10.2 15.4



#32
4-Chloroaniline
Concen: 21.704 ng/ul
RT: 10.840 min Scan# 1335
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:127 Resp: 407783
Ion Ratio Lower Upper
127 100
129 32.0 25.6 38.4









#33
Hexachlorobutadiene
Concen: 20.615 ng/ul
RT: 11.004 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:225 Resp: 166134
Ion Ratio Lower Upper
225 100
223 63.2 50.9 76.3
227 62.7 51.1 76.7



#34
Caprolactam
Concen: 22.927 ng/ul
RT: 11.640 min Scan# 1471
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:113 Resp: 81877
Ion Ratio Lower Upper
113 100
55 142.8 115.2 172.8
56 104.9 85.6 128.4









#35
4-Chloro-3-methylphenol
Concen: 22.695 ng/ul
RT: 11.963 min Scan# 1525
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:107 Resp: 279418
Ion Ratio Lower Upper
107 100
144 27.9 23.0 34.4
142 85.7 69.8 104.8



#36
2-Methylnaphthalene
Concen: 21.353 ng/ul
RT: 12.334 min Scan# 1589
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:142 Resp: 647402
Ion Ratio Lower Upper
142 100
141 88.1 70.7 106.1









#37
1-Methylnaphthalene
Concen: 21.407 ng/ul
RT: 12.551 min Scan# 1625
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:142 Resp: 649999
Ion Ratio Lower Upper
142 100
141 91.7 72.9 109.3
116 3.5 2.8 4.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.712 ng/ul
RT: 12.698 min Scan# 1651
Delta R.T. -0.000 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:216 Resp: 326009
Ion Ratio Lower Upper
216 100
214 80.5 63.0 94.4
179 20.8 16.3 24.5
108 17.1 12.5 18.7









#40
Hexachlorocyclopentadiene
Concen: 17.409 ng/ul
RT: 12.675 min Scan# 1647
Delta R.T. -0.000 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:237 Resp: 139850
Ion Ratio Lower Upper
237 100
235 63.1 49.8 74.8
272 13.1 9.9 14.9



#41
2,4,6-Trichlorophenol
Concen: 21.122 ng/ul
RT: 12.945 min Scan# 1693
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:196 Resp: 210427
Ion Ratio Lower Upper
196 100
198 98.0 76.2 114.2
200 31.4 24.2 36.2









#42
2,4,5-Trichlorophenol
Concen: 20.938 ng/ul
RT: 13.016 min Scan# 1705
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:196 Resp: 226700
Ion Ratio Lower Upper
196 100
198 93.8 75.4 113.2
200 30.3 24.8 37.2



#43
1,1'-Biphenyl
Concen: 20.258 ng/ul
RT: 13.345 min Scan# 1761
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:154 Resp: 827457
Ion Ratio Lower Upper
154 100
153 39.9 32.2 48.2
76 12.2 10.3 15.5





#44

2-Chloronaphthalene

Concen: 20.350 ng/ul

RT: 13.387 min Scan# 11

Delta R.T. 0.006 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03

Instrument :

BNA_P

ClientSampleId :

SICV637





Tgt Ion:162 Resp: 663756

Ion Ratio Lower Upper

162 100

164 32.6 26.0 39.0

127 38.3 29.4 44.2



#45

2-Nitroaniline

Concen: 22.718 ng/ul

RT: 13.598 min Scan# 1804

Delta R.T. 0.006 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03




Tgt Ion: 65 Resp: 158066


Ion Ratio Lower Upper

65 100

92 74.2 61.6 92.4

138 118.4 96.5 144.7









#47
Dimethylphthalate
Concen: 21.816 ng/ul
RT: 13.975 min Scan# 1868
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:163 Resp: 787597
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 9.6 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 22.916 ng/ul
RT: 14.098 min Scan# 1889
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:165 Resp: 146467
Ion Ratio Lower Upper
165 100
89 50.8 43.4 65.0
121 20.7 18.1 27.1









#50
Acenaphthylene
Concen: 21.038 ng/ul
RT: 14.234 min Scan# 1912
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:152 Resp: 1008657
Ion Ratio Lower Upper
152 100
151 20.0 15.7 23.5
153 13.2 10.4 15.6



#51
3-Nitroaniline
Concen: 23.082 ng/ul
RT: 14.428 min Scan# 1945
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:138 Resp: 171787
Ion Ratio Lower Upper
138 100
108 10.6 8.5 12.7
92 107.1 85.8 128.8









#52
Acenaphthene
Concen: 20.754 ng/ul
RT: 14.575 min Scan# 1969
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:153 Resp: 708334
Ion Ratio Lower Upper
153 100
152 47.5 38.2 57.2
154 88.9 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 19.845 ng/ul
RT: 14.634 min Scan# 1980
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:184 Resp: 81944
Ion Ratio Lower Upper
184 100
63 55.1 45.5 68.3
154 57.0 47.0 70.6









#55
4-Nitrophenol
Concen: 23.140 ng/ul
RT: 14.734 min Scan# 1996
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:109 Resp: 107619
Ion Ratio Lower Upper
109 100
139 159.6 124.1 186.1
65 134.6 104.2 156.4



#56
Dibenzofuran
Concen: 20.949 ng/ul
RT: 14.910 min Scan# 2027
Delta R.T. -0.000 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:168 Resp: 966328
Ion Ratio Lower Upper
168 100
139 36.5 28.2 42.4





#57

2,4-Dinitrotoluene

Concen: 24.317 ng/ul

RT: 14.881 min Scan# 2021

Delta R.T. 0.006 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03

Instrument :

BNA_P

ClientSampleId :

SICV637





Tgt Ion:165 Resp: 223375

Ion Ratio Lower Upper

165 100

63 38.7 31.1 46.7

182 3.0 1.8 2.6#



#58

2,3,4,6-Tetrachlorophenol

Concen: 22.711 ng/ul

RT: 15.139 min Scan# 2066

Delta R.T. 0.006 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03

Tgt Ion:232 Resp: 191662




Ion Ratio Lower Upper


232 100

131 44.1 37.8 56.6

130 1.9 1.8 2.6

166 30.0 26.6 39.8









#59
Diethylphthalate
Concen: 23.068 ng/ul
RT: 15.339 min Scan# 2100
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:149 Resp: 762677
Ion Ratio Lower Upper
149 100
177 22.3 17.5 26.3
150 12.1 9.8 14.8



#61
Fluorene
Concen: 21.806 ng/ul
RT: 15.563 min Scan# 2138
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:166 Resp: 789565
Ion Ratio Lower Upper
166 100
165 97.2 79.5 119.3
167 13.4 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 21.357 ng/ul
RT: 15.557 min Scan# 2136
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:204 Resp: 374099
Ion Ratio Lower Upper
204 100
206 33.1 25.4 38.2
141 56.5 46.2 69.2



#63
4-Nitroaniline
Concen: 25.568 ng/ul
RT: 15.592 min Scan# 2143
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:138 Resp: 173756
Ion Ratio Lower Upper
138 100
92 48.5 39.2 58.8
108 65.3 55.4 83.0









#66
4,6-Dinitro-2-methylphenol
Concen: 20.610 ng/ul
RT: 15.645 min Scan# 2151
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:198 Resp: 134609
Ion Ratio Lower Upper
198 100
182 5.1 4.9 7.3
77 23.4 20.0 30.0



#67
N-Nitrosodiphenylamine
Concen: 19.959 ng/ul
RT: 15.775 min Scan# 2174
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:169 Resp: 675915
Ion Ratio Lower Upper
169 100
168 67.5 53.5 80.3
167 35.7 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 19.708 ng/ul
RT: 16.457 min Scan# 2194
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:248 Resp: 219419
Ion Ratio Lower Upper
248 100
250 99.3 77.9 116.9
141 73.4 59.5 89.3



#69
Hexachlorobenzene
Concen: 19.959 ng/ul
RT: 16.569 min Scan# 2309
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:284 Resp: 257911
Ion Ratio Lower Upper
284 100
142 37.8 30.5 45.7
249 32.5 25.4 38.0









#70
Atrazine
Concen: 21.604 ng/ul
RT: 16.733 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:200 Resp: 224983
Ion Ratio Lower Upper
200 100
173 27.7 22.1 33.1
215 49.0 38.8 58.2



#71
Pentachlorophenol
Concen: 20.045 ng/ul
RT: 16.922 min Scan# 2369
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:266 Resp: 151761
Ion Ratio Lower Upper
266 100
264 63.9 50.6 75.8
268 64.8 51.8 77.6





#72

Phenanthrene

Concen: 20.709 ng/ul

RT: 17.316 min Scan# 2435

Delta R.T. 0.006 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03

Instrument :

BNA_P

ClientSampleId :

SICV637





Tgt Ion:178 Resp: 1288570

Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 19.5 15.8 23.6



#74

Anthracene

Concen: 21.060 ng/ul

RT: 17.404 min Scan# 2451

Delta R.T. 0.006 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03




Tgt Ion:178 Resp: 1296747


Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.4

176 18.5 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 17.664 ng/uL
RT: 13.310 min Scan# 1755
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:216 Resp: 326576
Ion Ratio Lower Upper
216 100
214 78.6 62.6 94.0
218 47.7 38.9 58.3



#76
Pentachlorobenzene
Concen: 18.649 ng/uL
RT: 14.834 min Scan# 2014
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:250 Resp: 338621
Ion Ratio Lower Upper
250 100
248 63.6 51.4 77.2
252 64.1 50.7 76.1









#77
Carbazole
Concen: 22.168 ng/ul
RT: 17.680 min Scan# 2497
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:167 Resp: 1217441
Ion Ratio Lower Upper
167 100
166 22.1 17.0 25.6
139 12.6 10.1 15.1



#78
Di-n-butylphthalate
Concen: 23.450 ng/ul
RT: 18.227 min Scan# 2591
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:149 Resp: 1277990
Ion Ratio Lower Upper
149 100
150 8.9 7.2 10.8
104 4.5 3.4 5.2









#80
Fluoranthene
Concen: 19.783 ng/ul
RT: 19.280 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03




Instrument :
BNA_P
ClientSampleId :
SICV637


Tgt Ion:202 Resp: 1530105
Ion Ratio Lower Upper
202 100
101 12.0 9.0 13.6
100 9.1 7.0 10.6



#82
Pyrene
Concen: 19.600 ng/ul
RT: 19.633 min Scan# 2830
Delta R.T. 0.006 min
Lab File: BP012040.D
Acq: 11 Oct 2022 15:03

Tgt Ion:202 Resp: 1604299
Ion Ratio Lower Upper
202 100
101 14.3 10.4 15.6
100 11.7 8.5 12.7





#83

Butylbenzylphthalate

Concen: 23.319 ng/ul

RT: 20.510 min Scan# 2978

Delta R.T. 0.006 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03

Instrument :

BNA_P

ClientSampleId :

SICV637





Tgt Ion:149 Resp: 605046

Ion Ratio Lower Upper

149 100

91 62.7 48.5 72.7

206 20.8 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 21.709 ng/ul

RT: 21.280 min Scan# 3110

Delta R.T. 0.006 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03




Tgt Ion:252 Resp: 589394


Ion Ratio Lower Upper

252 100

254 64.0 51.4 77.2

126 11.2 7.8 11.8





#85

Benzo(a)anthracene

Concen: 21.161 ng/ul

RT: 21.345 min Scan# 3120

Delta R.T. 0.006 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03

Instrument :

BNA_P

ClientSampleId :

SICV637





Tgt Ion:228 Resp: 1641089

Ion Ratio Lower Upper

228 100

229 19.6 15.7 23.5

226 26.9 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 24.608 ng/ul

RT: 21.263 min Scan# 3107

Delta R.T. 0.006 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03




Tgt Ion:149 Resp: 860734


Ion Ratio Lower Upper

149 100

167 27.2 21.9 32.9

279 3.7 3.4 5.2





#87

Chrysene

Concen: 20.927 ng/ul

RT: 21.398 min Scan# 3129

Delta R.T. 0.006 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03

Instrument :

BNA_P

ClientSampleId :

SICV637





Tgt Ion:228 Resp: 1562220

Ion Ratio Lower Upper

228 100

226 30.3 24.0 36.0

229 19.9 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 23.063 ng/ul




RT: 22.192 min Scan# 3265


Delta R.T. 0.006 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03

Tgt Ion:149 Resp: 1490977





#90

Benzo(b)fluoranthene

Concen: 20.610 ng/ul

RT: 23.039 min Scan# 3408

Delta R.T. 0.006 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03

Instrument :

BNA_P

ClientSampleId :

SICV637





Tgt Ion:252 Resp: 1684551

Ion Ratio Lower Upper

252 100

253 22.5 17.6 26.4

125 11.5 7.8 11.8



#91

Benzo(k)fluoranthene

Concen: 21.241 ng/ul

RT: 23.092 min Scan# 3418

Delta R.T. 0.012 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03




Tgt Ion:252 Resp: 1707710


Ion Ratio Lower Upper

252 100

253 21.5 17.6 26.4

125 10.8 8.2 12.2





#93

Benzo(a)pyrene

Concen: 20.847 ng/ul

RT: 23.674 min Scan# 3515

Delta R.T. 0.012 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03

Instrument :

BNA_P

ClientSampleId :

SICV637





Tgt Ion:252 Resp: 1520242

Ion Ratio Lower Upper

252 100

253 21.9 17.6 26.4

125 11.7 8.9 13.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.109 ng/ul

RT: 26.303 min Scan# 3964

Delta R.T. 0.012 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03




Tgt Ion:276 Resp: 1854175


Ion Ratio Lower Upper

276 100

138 27.4 19.1 28.7

277 25.0 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 20.247 ng/ul

RT: 26.321 min Scan# 3967

Delta R.T. 0.012 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03

Instrument :

BNA_P

ClientSampleId :

SICV637





Tgt Ion:278 Resp: 1674765

Ion Ratio Lower Upper

278 100

139 18.0 12.5 18.7

279 24.0 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 20.027 ng/ul

RT: 27.080 min Scan# 4096

Delta R.T. 0.017 min

Lab File: BP012040.D

Acq: 11 Oct 2022 15:03

Tgt Ion:276 Resp: 1664905

Ion Ratio Lower Upper

276 100

138 23.8 16.5 24.7

277 23.6 19.0 28.4

