

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BP012043.D
 Acq On : 11 Oct 2022 16:48
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020773

Quant Time: Oct 12 00:22:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 11 23:54:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.857	152	280786	20.000	ng/u1	0.00
20) Naphthalene-d8	10.669	136	1214256	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.510	164	708333	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.263	188	1308257	20.000	ng/u1	0.00
79) Chrysene-d12	21.351	240	942034	20.000	ng/u1	0.00
88) Perylene-d12	23.768	264	892774	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	54387	7.637	ng/uL	0.00
4) Pyridine-d5	3.705	84	367368	19.397	ng/u1	0.00
7) Phenol-d5	7.028	99	458477	20.302	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.193	67	267724	20.886	ng/u1	0.00
11) 2-Chlorophenol-d4	7.387	132	372018	20.754	ng/u1	0.00
15) 4-Methylphenol-d8	8.575	113	363322	20.946	ng/u1	0.00
21) Nitrobenzene-d5	9.028	128	178442	20.808	ng/u1	0.00
24) 2-Nitrophenol-d4	9.751	143	189694	21.815	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.287	165	357289	21.201	ng/u1	0.00
31) 4-Chloroaniline-d4	10.810	131	514844	20.731	ng/u1	0.00
46) Dimethylphthalate-d6	13.922	166	943588	21.034	ng/u1	0.00
49) Acenaphthylene-d8	14.198	160	1162242	20.952	ng/u1	0.00
54) 4-Nitrophenol-d4	14.710	143	171279	19.448	ng/u1	0.00
60) Fluorene-d10	15.504	176	837511	20.635	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.628	200	134094	19.411	ng/u1	0.00
73) Anthracene-d10	17.363	188	1176134	20.703	ng/u1	0.00
81) Pyrene-d10	19.598	212	1136887	23.557	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.615	264	882685	20.338	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.316	88	55778	7.443	ng/uL	98
5) Pyridine	3.722	79	356472	19.644	ng/u1	99
6) Benzaldehyde	7.005	77	246845	23.298	ng/u1	100
8) Phenol	7.052	94	480272	20.258	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.287	93	387759	20.730	ng/u1	99
12) 2-Chlorophenol	7.422	128	399151	20.518	ng/u1	99
13) 2-Methylphenol	8.310	108	363031	20.531	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.393	45	408249	21.294	ng/u1	99
16) Acetophenone	8.693	105	582564	21.785	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.675	70	268019	22.005	ng/u1	98
18) 4-Methylphenol	8.640	108	400226	21.000	ng/u1	98
19) Hexachloroethane	8.940	117	150463	20.036	ng/u1	98
22) Nitrobenzene	9.069	77	413946	21.057	ng/u1	99
23) Isophorone	9.599	82	758346	21.748	ng/u1	100
25) 2-Nitrophenol	9.781	139	218786	21.416	ng/u1	96
26) 2,4-Dimethylphenol	9.840	107	416192	20.918	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.081	93	505072	21.107	ng/u1	100
29) 2,4-Dichlorophenol	10.316	162	369017	20.837	ng/u1	99
30) Naphthalene	10.716	128	1304802	20.335	ng/u1	100
32) 4-Chloroaniline	10.834	127	533678	20.844	ng/u1	99
33) Hexachlorobutadiene	10.998	225	219849	20.019	ng/u1	98
34) Caprolactam	11.628	113	112122	23.039	ng/u1	98
35) 4-Chloro-3-methylphenol	11.957	107	371114	22.119	ng/u1	97
36) 2-Methylnaphthalene	12.328	142	864472	20.922	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BP012043.D
 Acq On : 11 Oct 2022 16:48
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD020773


Quant Time: Oct 12 00:22:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 11 23:54:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.545	142	868152	20.981	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.692	216	437445	20.203	ng/ul	99
40) Hexachlorocyclopentadiene	12.669	237	221864	21.095	ng/ul	100
41) 2,4,6-Trichlorophenol	12.940	196	275406	21.115	ng/ul	98
42) 2,4,5-Trichlorophenol	13.010	196	294942	20.807	ng/ul	99
43) 1,1'-Biphenyl	13.339	154	1095557	20.487	ng/ul	99
44) 2-Chloronaphthalene	13.381	162	873334	20.452	ng/ul	99
45) 2-Nitroaniline	13.592	65	201383	22.108	ng/ul	98
47) Dimethylphthalate	13.969	163	993703	21.025	ng/ul	99
48) 2,6-Dinitrotoluene	14.092	165	187267	22.379	ng/ul	97
50) Acenaphthylene	14.228	152	1312934	20.917	ng/ul	99
51) 3-Nitroaniline	14.422	138	201412	20.671	ng/ul	97
52) Acenaphthene	14.569	153	916223	20.505	ng/ul	99
53) 2,4-Dinitrophenol	14.628	184	111183	20.567	ng/ul	99
55) 4-Nitrophenol	14.728	109	119551	19.635	ng/ul	96
56) Dibenzofuran	14.904	168	1221934	20.234	ng/ul	97
57) 2,4-Dinitrotoluene	14.875	165	263904	21.944	ng/ul#	98
58) 2,3,4,6-Tetrachlorophenol	15.134	232	233595	21.142	ng/ul	99
59) Diethylphthalate	15.334	149	926164	21.397	ng/ul	100
61) Fluorene	15.557	166	988348	20.850	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.551	204	475123	20.719	ng/ul	99
63) 4-Nitroaniline	15.586	138	181694	20.422	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.639	198	147448	19.744	ng/ul	96
67) N-Nitrosodiphenylamine	15.769	169	821791	21.222	ng/ul	100
68) 4-Bromophenyl-phenylether	16.451	248	267619	21.022	ng/ul	98
69) Hexachlorobenzene	16.563	284	300721	20.352	ng/ul	99
70) Atrazine	16.728	200	240200	20.172	ng/ul	100
71) Pentachlorophenol	16.916	266	163858	18.928	ng/ul	99
72) Phenanthrene	17.304	178	1454146	20.439	ng/ul	100
74) Anthracene	17.398	178	1454501	20.659	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.304	216	434120	20.535	ng/ul	99
76) Pentachlorobenzene	14.822	250	429815	20.702	ng/ul	99
77) Carbazole	17.675	167	1264813	20.142	ng/ul	100
78) Di-n-butylphthalate	18.222	149	1269395	20.371	ng/ul	100
80) Fluoranthene	19.274	202	1429036	24.165	ng/ul	99
82) Pyrene	19.627	202	1486704	23.756	ng/ul	97
83) Butylbenzylphthalate	20.498	149	440494	22.204	ng/ul	89
84) 3,3'-Dichlorobenzidine	21.274	252	385953	18.592	ng/ul	99
85) Benzo(a)anthracene	21.339	228	1212262	20.444	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.257	149	556302	20.801	ng/ul	99
87) Chrysene	21.392	228	1160340	20.330	ng/ul	100
89) Di-n-octyl phthalate	22.186	149	896178	20.050	ng/ul	100
90) Benzo(b)fluoranthene	23.033	252	1176395	20.817	ng/ul	97
91) Benzo(k)fluoranthene	23.080	252	1142718	20.558	ng/ul	99
93) Benzo(a)pyrene	23.662	252	1040767	20.643	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.292	276	1315305	20.633	ng/ul#	94
95) Dibenzo(a,h)anthracene	26.304	278	1178129	20.600	ng/ul	97
96) Benzo(g,h,i)perylene	27.062	276	1181284	20.552	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 12 00:22:42 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 11 23:54:43 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 7.443 ng/uL
 RT: 3.316 min Scan# 50
 Delta R.T. -0.001 min
 Lab File: BP012043.D
 Acq: 11 Oct 2022 16:48




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020773


Tgt Ion: 88 Resp: 55778
 Ion Ratio Lower Upper
 88 100
 43 20.9 16.6 25.0
 58 68.3 52.6 79.0



#5
 Pyridine
 Concen: 19.644 ng/uL
 RT: 3.722 min Scan# 125
 Delta R.T. -0.001 min
 Lab File: BP012043.D
 Acq: 11 Oct 2022 16:48

Tgt Ion: 79 Resp: 356472
 Ion Ratio Lower Upper
 79 100
 52 64.8 51.2 76.8
 51 28.1 22.8 34.2









#6
Benzaldehyde
Concen: 23.298 ng/ul
RT: 7.005 min Scan# 61
Delta R.T. 0.006 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion: 77 Resp: 246845
Ion Ratio Lower Upper
77 100
105 106.5 85.2 127.8
106 103.4 83.1 124.7



#8
Phenol
Concen: 20.258 ng/ul
RT: 7.052 min Scan# 691
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion: 94 Resp: 480272
Ion Ratio Lower Upper
94 100
65 25.4 19.4 29.0
66 31.6 24.6 37.0









#10
Bis(2-Chloroethyl)ether
Concen: 20.730 ng/ul
RT: 7.287 min Scan# 731
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion: 93 Resp: 387759
Ion Ratio Lower Upper
93 100
63 71.9 56.7 85.1
95 33.3 26.2 39.2



#12
2-Chlorophenol
Concen: 20.518 ng/ul
RT: 7.422 min Scan# 754
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion:128 Resp: 399151
Ion Ratio Lower Upper
128 100
64 40.0 32.7 49.1
130 32.4 26.3 39.5









#13
2-Methylphenol
Concen: 20.531 ng/ul
RT: 8.310 min Scan# 904
Delta R.T. 0.006 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:108 Resp: 363031
Ion Ratio Lower Upper
108 100
107 88.0 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 21.294 ng/ul
RT: 8.393 min Scan# 919
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion: 45 Resp: 408249
Ion Ratio Lower Upper
45 100
77 18.7 14.4 21.6
79 13.8 11.3 16.9









#16
Acetophenone
Concen: 21.785 ng/ul
RT: 8.693 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:105 Resp: 582564
Ion Ratio Lower Upper
105 100
77 69.2 57.0 85.4
51 19.7 16.4 24.6



#17
N-Nitroso-di-n-propylamine
Concen: 22.005 ng/ul
RT: 8.675 min Scan# 967
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion: 70 Resp: 268019
Ion Ratio Lower Upper
70 100
42 42.5 34.6 52.0
101 9.7 8.0 12.0
130 24.5 18.3 27.5









#18
4-Methylphenol
Concen: 21.000 ng/ul
RT: 8.640 min Scan# 961
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:108 Resp: 400226
Ion Ratio Lower Upper
108 100
107 109.8 86.3 129.5



#19
Hexachloroethane
Concen: 20.036 ng/ul
RT: 8.940 min Scan# 1012
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion:117 Resp: 150463
Ion Ratio Lower Upper
117 100
201 99.5 82.3 123.5
199 62.9 50.0 75.0









#22
Nitrobenzene
Concen: 21.057 ng/ul
RT: 9.069 min Scan# 1034
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion: 77 Resp: 413946
Ion Ratio Lower Upper
77 100
123 50.3 39.4 59.0
65 12.1 9.9 14.9



#23
Isophorone
Concen: 21.748 ng/ul
RT: 9.599 min Scan# 1124
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion: 82 Resp: 758346
Ion Ratio Lower Upper
82 100
95 6.7 5.4 8.0
138 18.9 15.2 22.8









#25
2-Nitrophenol
Concen: 21.416 ng/ul
RT: 9.781 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:139 Resp: 218786
Ion Ratio Lower Upper
139 100
65 42.3 36.3 54.5
109 25.4 21.2 31.8



#26
2,4-Dimethylphenol
Concen: 20.918 ng/ul
RT: 9.840 min Scan# 1165
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion:107 Resp: 416192
Ion Ratio Lower Upper
107 100
121 52.7 43.0 64.4
122 90.9 71.4 107.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.107 ng/ul
 RT: 10.081 min Scan# 1206
 Delta R.T. -0.000 min
 Lab File: BP012043.D
 Acq: 11 Oct 2022 16:48




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020773


Tgt Ion: 93 Resp: 505072
 Ion Ratio Lower Upper
 93 100
 95 32.4 25.7 38.5
 123 11.5 9.2 13.8



#29
 2,4-Dichlorophenol
 Concen: 20.837 ng/ul
 RT: 10.316 min Scan# 1246
 Delta R.T. -0.000 min
 Lab File: BP012043.D
 Acq: 11 Oct 2022 16:48

Tgt Ion:162 Resp: 369017
 Ion Ratio Lower Upper
 162 100
 164 65.3 51.5 77.3
 98 33.9 26.3 39.5









#30
Naphthalene
Concen: 20.335 ng/ul
RT: 10.716 min Scan# 1314
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:128 Resp: 1304802
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 13.0 10.2 15.4



#32
4-Chloroaniline
Concen: 20.844 ng/ul
RT: 10.834 min Scan# 1334
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion:127 Resp: 533678
Ion Ratio Lower Upper
127 100
129 32.8 25.6 38.4









#33
Hexachlorobutadiene
Concen: 20.019 ng/ul
RT: 10.998 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:225 Resp: 219849
Ion Ratio Lower Upper
225 100
223 66.0 50.9 76.3
227 65.0 51.1 76.7



#34
Caprolactam
Concen: 23.039 ng/ul
RT: 11.628 min Scan# 1469
Delta R.T. -0.006 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion:113 Resp: 112122
Ion Ratio Lower Upper
113 100
55 140.9 115.2 172.8
56 108.7 85.6 128.4









#35
4-Chloro-3-methylphenol
Concen: 22.119 ng/ul
RT: 11.957 min Scan# 1525
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:107 Resp: 371114
Ion Ratio Lower Upper
107 100
144 27.8 23.0 34.4
142 84.2 69.8 104.8



#36
2-Methylnaphthalene
Concen: 20.922 ng/ul
RT: 12.328 min Scan# 1588
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion:142 Resp: 864472
Ion Ratio Lower Upper
142 100
141 89.6 70.7 106.1









#37
1-Methylnaphthalene
Concen: 20.981 ng/ul
RT: 12.545 min Scan# 1625
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:142 Resp: 868152
Ion Ratio Lower Upper
142 100
141 92.9 72.9 109.3
116 3.6 2.8 4.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.203 ng/ul
RT: 12.692 min Scan# 1650
Delta R.T. -0.006 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion:216 Resp: 437445
Ion Ratio Lower Upper
216 100
214 79.2 63.0 94.4
179 21.1 16.3 24.5
108 17.0 12.5 18.7









#40
Hexachlorocyclopentadiene
Concen: 21.095 ng/ul
RT: 12.669 min Scan# 1646
Delta R.T. -0.006 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:237 Resp: 221864
Ion Ratio Lower Upper
237 100
235 62.6 49.8 74.8
272 12.6 9.9 14.9



#41
2,4,6-Trichlorophenol
Concen: 21.115 ng/ul
RT: 12.940 min Scan# 1692
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion:196 Resp: 275406
Ion Ratio Lower Upper
196 100
198 97.4 76.2 114.2
200 30.7 24.2 36.2





#42

2,4,5-Trichlorophenol

Concen: 20.807 ng/ul

RT: 13.010 min Scan# 1704

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48

Instrument :

BNA_P

ClientSampleId :

SSTD020773





Tgt Ion:196 Resp: 294942

Ion Ratio Lower Upper

196 100

198 95.8 75.4 113.2

200 31.0 24.8 37.2



#43

1,1'-Biphenyl

Concen: 20.487 ng/ul

RT: 13.339 min Scan# 1760

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48




Tgt Ion:154 Resp: 1095557


Ion Ratio Lower Upper

154 100

153 39.7 32.2 48.2

76 12.6 10.3 15.5









#44
2-Chloronaphthalene
Concen: 20.452 ng/ul
RT: 13.381 min Scan# 1767
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:162 Resp: 873334
Ion Ratio Lower Upper
162 100
164 32.2 26.0 39.0
127 38.1 29.4 44.2



#45
2-Nitroaniline
Concen: 22.108 ng/ul
RT: 13.592 min Scan# 1803
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion: 65 Resp: 201383
Ion Ratio Lower Upper
65 100
92 73.7 61.6 92.4
138 119.8 96.5 144.7









#47
Dimethylphthalate
Concen: 21.025 ng/ul
RT: 13.969 min Scan# 1867
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:163 Resp: 993703
Ion Ratio Lower Upper
163 100
194 4.1 3.1 4.7
164 10.4 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 22.379 ng/ul
RT: 14.092 min Scan# 1888
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion:165 Resp: 187267
Ion Ratio Lower Upper
165 100
89 52.4 43.4 65.0
121 20.8 18.1 27.1









#50
Acenaphthylene
Concen: 20.917 ng/ul
RT: 14.228 min Scan# 1911
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:152 Resp: 1312934
Ion Ratio Lower Upper
152 100
151 19.9 15.7 23.5
153 13.2 10.4 15.6



#51
3-Nitroaniline
Concen: 20.671 ng/ul
RT: 14.422 min Scan# 1944
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion:138 Resp: 201412
Ion Ratio Lower Upper
138 100
108 10.2 8.5 12.7
92 104.1 85.8 128.8





#52

Acenaphthene

Concen: 20.505 ng/ul

RT: 14.569 min Scan# 1969

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48

Instrument :

BNA_P

ClientSampleId :

SSTD020773





Tgt Ion:153 Resp: 916223

Ion Ratio Lower Upper

153 100

152 47.5 38.2 57.2

154 87.5 70.6 105.8



#53

2,4-Dinitrophenol

Concen: 20.567 ng/ul

RT: 14.628 min Scan# 1979

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48




Tgt Ion:184 Resp: 111183


Ion Ratio Lower Upper

184 100

63 56.5 45.5 68.3

154 57.2 47.0 70.6









#55
4-Nitrophenol
Concen: 19.635 ng/ul
RT: 14.728 min Scan# 1996
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:109 Resp: 119551
Ion Ratio Lower Upper
109 100
139 163.2 124.1 186.1
65 132.1 104.2 156.4



#56
Dibenzofuran
Concen: 20.234 ng/ul
RT: 14.904 min Scan# 2026
Delta R.T. -0.006 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion:168 Resp: 1221934
Ion Ratio Lower Upper
168 100
139 37.0 28.2 42.4





#57

2,4-Dinitrotoluene

Concen: 21.944 ng/ul

RT: 14.875 min Scan# 2021

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48

Instrument :

BNA_P

ClientSampleId :

SSTD020773





Tgt Ion:165 Resp: 263904

Ion Ratio Lower Upper

165 100

63 37.9 31.1 46.7

182 3.0 1.8 2.6#



#58

2,3,4,6-Tetrachlorophenol

Concen: 21.142 ng/ul

RT: 15.134 min Scan# 2065

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48

Tgt Ion:232 Resp: 233595




Ion Ratio Lower Upper


232 100

131 47.0 37.8 56.6

130 2.5 1.8 2.6

166 31.8 26.6 39.8









#59
Diethylphthalate
Concen: 21.397 ng/ul
RT: 15.334 min Scan# 2099
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:149 Resp: 926164
Ion Ratio Lower Upper
149 100
177 21.9 17.5 26.3
150 12.2 9.8 14.8



#61
Fluorene
Concen: 20.850 ng/ul
RT: 15.557 min Scan# 2137
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion:166 Resp: 988348
Ion Ratio Lower Upper
166 100
165 97.7 79.5 119.3
167 13.3 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 20.719 ng/ul
RT: 15.551 min Scan# 2136
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:204 Resp: 475123
Ion Ratio Lower Upper
204 100
206 32.8 25.4 38.2
141 57.8 46.2 69.2



#63
4-Nitroaniline
Concen: 20.422 ng/ul
RT: 15.586 min Scan# 2142
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion:138 Resp: 181694
Ion Ratio Lower Upper
138 100
92 49.0 39.2 58.8
108 66.2 55.4 83.0









#66
4,6-Dinitro-2-methylphenol
Concen: 19.744 ng/ul
RT: 15.639 min Scan# 2151
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:198 Resp: 147448
Ion Ratio Lower Upper
198 100
182 5.4 4.9 7.3
77 22.5 20.0 30.0



#67
N-Nitrosodiphenylamine
Concen: 21.222 ng/ul
RT: 15.769 min Scan# 2173
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion:169 Resp: 821791
Ion Ratio Lower Upper
169 100
168 67.2 53.5 80.3
167 36.5 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 21.022 ng/ul
RT: 16.451 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:248 Resp: 267619
Ion Ratio Lower Upper
248 100
250 97.0 77.9 116.9
141 71.2 59.5 89.3



#69
Hexachlorobenzene
Concen: 20.352 ng/ul
RT: 16.563 min Scan# 2308
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion:284 Resp: 300721
Ion Ratio Lower Upper
284 100
142 36.9 30.5 45.7
249 31.5 25.4 38.0





#70

Atrazine

Concen: 20.172 ng/ul

RT: 16.728 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48

Instrument :

BNA_P

ClientSampleId :

SSTD020773





Tgt Ion:200 Resp: 240200

Ion Ratio Lower Upper

200 100

173 27.5 22.1 33.1

215 48.1 38.8 58.2



#71

Pentachlorophenol

Concen: 18.928 ng/ul

RT: 16.916 min Scan# 2368

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48




Tgt Ion:266 Resp: 163858


Ion Ratio Lower Upper

266 100

264 62.6 50.6 75.8

268 63.8 51.8 77.6





#72

Phenanthrene

Concen: 20.439 ng/ul

RT: 17.304 min Scan# 2435

Delta R.T. -0.006 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48

Instrument :

BNA_P

ClientSampleId :

SSTD020773





Tgt Ion:178 Resp: 1454146

Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 19.5 15.8 23.6



#74

Anthracene

Concen: 20.659 ng/ul

RT: 17.398 min Scan# 2450

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48




Tgt Ion:178 Resp: 1454501


Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.4

176 18.9 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.535 ng/uL
RT: 13.304 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:216 Resp: 434120
Ion Ratio Lower Upper
216 100
214 78.9 62.6 94.0
218 47.7 38.9 58.3



#76
Pentachlorobenzene
Concen: 20.702 ng/uL
RT: 14.822 min Scan# 2012
Delta R.T. -0.006 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion:250 Resp: 429815
Ion Ratio Lower Upper
250 100
248 63.4 51.4 77.2
252 62.5 50.7 76.1





#77

Carbazole

Concen: 20.142 ng/ul

RT: 17.675 min Scan# 2497

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48

Instrument :

BNA_P

ClientSampleId :

SSTD020773





Tgt Ion:167 Resp: 1264813

Ion Ratio Lower Upper

167 100

166 21.4 17.0 25.6

139 12.6 10.1 15.1



#78

Di-n-butylphthalate

Concen: 20.371 ng/ul

RT: 18.222 min Scan# 2590

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48




Tgt Ion:149 Resp: 1269395


Ion Ratio Lower Upper

149 100

150 9.2 7.2 10.8

104 4.5 3.4 5.2





#80

Fluoranthene

Concen: 24.165 ng/ul

RT: 19.274 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48

Instrument :

BNA_P

ClientSampleId :

SSTD020773





Tgt Ion:202 Resp: 1429036

Ion Ratio Lower Upper

202 100

101 12.0 9.0 13.6

100 9.1 7.0 10.6



#82

Pyrene

Concen: 23.756 ng/ul

RT: 19.627 min Scan# 2829

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48




Tgt Ion:202 Resp: 1486704


Ion Ratio Lower Upper

202 100

101 14.2 10.4 15.6

100 11.4 8.5 12.7





#83

Butylbenzylphthalate

Concen: 22.204 ng/ul

RT: 20.498 min Scan# 2977

Delta R.T. -0.006 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48

Instrument :

BNA_P

ClientSampleId :

SSTD020773





Tgt Ion:149 Resp: 440494

Ion Ratio Lower Upper

149 100

91 70.4 48.5 72.7

206 20.0 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 18.592 ng/ul

RT: 21.274 min Scan# 3109

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48




Tgt Ion:252 Resp: 385953


Ion Ratio Lower Upper

252 100

254 64.4 51.4 77.2

126 11.7 7.8 11.8





#85

Benzo(a)anthracene

Concen: 20.444 ng/ul

RT: 21.339 min Scan# 3120

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48

Instrument :

BNA_P

ClientSampleId :

SSTD020773





Tgt Ion:228 Resp: 1212262

Ion Ratio Lower Upper

228 100

229 19.6 15.7 23.5

226 27.1 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.801 ng/ul

RT: 21.257 min Scan# 3106

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48




Tgt Ion:149 Resp: 556302


Ion Ratio Lower Upper

149 100

167 27.1 21.9 32.9

279 3.8 3.4 5.2









#87
Chrysene
Concen: 20.330 ng/ul
RT: 21.392 min Scan# 3129
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48




Instrument :
BNA_P
ClientSampleId :
SSTD020773


Tgt Ion:228 Resp: 1160340
Ion Ratio Lower Upper
228 100
226 30.1 24.0 36.0
229 19.7 15.5 23.3



#89
Di-n-octyl phthalate
Concen: 20.050 ng/ul
RT: 22.186 min Scan# 3264
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion:149 Resp: 896178





#90

Benzo(b)fluoranthene

Concen: 20.817 ng/ul

RT: 23.033 min Scan# 3408

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48

Instrument :

BNA_P

ClientSampleId :

SSTD020773





Tgt Ion:252 Resp: 1176395

Ion Ratio Lower Upper

252 100

253 21.3 17.6 26.4

125 11.7 7.8 11.8



#91

Benzo(k)fluoranthene

Concen: 20.558 ng/ul

RT: 23.080 min Scan# 3416

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48




Tgt Ion:252 Resp: 1142718


Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4

125 11.0 8.2 12.2





#93

Benzo(a)pyrene

Concen: 20.643 ng/ul

RT: 23.662 min Scan# 3515

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48

Instrument :

BNA_P

ClientSampleId :

SSTD020773





Tgt Ion:252 Resp: 1040767

Ion Ratio Lower Upper

252 100

253 21.6 17.6 26.4

125 12.6 8.9 13.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.633 ng/ul

RT: 26.292 min Scan# 3962

Delta R.T. -0.000 min

Lab File: BP012043.D

Acq: 11 Oct 2022 16:48




Tgt Ion:276 Resp: 1315305


Ion Ratio Lower Upper

276 100

138 29.2 19.1 28.7#

277 25.2 19.9 29.9









#95
Dibenzo(a,h)anthracene
Concen: 20.600 ng/ul
RT: 26.304 min Scan# 3964
Delta R.T. -0.006 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Instrument :
BNA_P
ClientSampleId :
SSTD020773

Tgt Ion:278 Resp: 1178129
Ion Ratio Lower Upper
278 100
139 18.2 12.5 18.7
279 24.1 19.0 28.4



#96
Benzo(g,h,i)perylene
Concen: 20.552 ng/ul
RT: 27.062 min Scan# 4093
Delta R.T. -0.000 min
Lab File: BP012043.D
Acq: 11 Oct 2022 16:48

Tgt Ion:276 Resp: 1181284
Ion Ratio Lower Upper
276 100
138 24.1 16.5 24.7
277 23.5 19.0 28.4

