

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BP012071.D
 Acq On : 12 Oct 2022 09:24
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020775

Quant Time: Oct 12 23:05:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 23:03:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.852	152	251883	20.000	ng/u1	0.00
20) Naphthalene-d8	10.657	136	1008239	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.498	164	580178	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.257	188	1191886	20.000	ng/u1	0.00
79) Chrysene-d12	21.345	240	1298729	20.000	ng/u1	0.00
88) Perylene-d12	23.763	264	1375986	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	52146	8.162	ng/uL	0.00
4) Pyridine-d5	3.699	84	350669	20.640	ng/u1	0.00
7) Phenol-d5	7.016	99	402587	19.873	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.187	67	230246	20.024	ng/u1	0.00
11) 2-Chlorophenol-d4	7.381	132	329238	20.475	ng/u1	0.00
15) 4-Methylphenol-d8	8.563	113	305400	19.627	ng/u1	0.00
21) Nitrobenzene-d5	9.022	128	147813	20.758	ng/u1	0.00
24) 2-Nitrophenol-d4	9.746	143	151914	21.040	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.281	165	293469	20.972	ng/u1	0.00
31) 4-Chloroaniline-d4	10.804	131	424863	20.604	ng/u1	0.00
46) Dimethylphthalate-d6	13.916	166	748562	20.373	ng/u1	0.00
49) Acenaphthylene-d8	14.193	160	935825	20.596	ng/u1	0.00
54) 4-Nitrophenol-d4	14.710	143	158090	21.916	ng/u1	0.00
60) Fluorene-d10	15.498	176	695147	20.911	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.622	200	116017	18.434	ng/u1	0.00
73) Anthracene-d10	17.357	188	1073651	20.744	ng/u1	0.00
81) Pyrene-d10	19.592	212	1289748	19.385	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.610	264	1348805	20.165	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.311	88	54131	8.052	ng/uL	98
5) Pyridine	3.717	79	339460	20.853	ng/u1	98
6) Benzaldehyde	6.993	77	213935	22.508	ng/u1	95
8) Phenol	7.046	94	423631	19.919	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.281	93	334172	19.916	ng/u1	98
12) 2-Chlorophenol	7.416	128	358057	20.517	ng/u1	98
13) 2-Methylphenol	8.299	108	311818	19.658	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.381	45	343215	19.956	ng/u1	99
16) Acetophenone	8.681	105	475833	19.835	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.663	70	213377	19.529	ng/u1	100
18) 4-Methylphenol	8.628	108	338171	19.780	ng/u1	98
19) Hexachloroethane	8.934	117	132877	19.725	ng/u1	99
22) Nitrobenzene	9.063	77	343178	21.024	ng/u1	100
23) Isophorone	9.593	82	586800	20.267	ng/u1	100
25) 2-Nitrophenol	9.775	139	177177	20.887	ng/u1	99
26) 2,4-Dimethylphenol	9.834	107	342965	20.760	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.075	93	400805	20.173	ng/u1	99
29) 2,4-Dichlorophenol	10.310	162	305867	20.800	ng/u1	98
30) Naphthalene	10.710	128	1093585	20.526	ng/u1	99
32) 4-Chloroaniline	10.828	127	440960	20.741	ng/u1	100
33) Hexachlorobutadiene	10.993	225	179457	19.680	ng/u1	98
34) Caprolactam	11.616	113	91473	22.637	ng/u1	97
35) 4-Chloro-3-methylphenol	11.951	107	300558	21.574	ng/u1	99
36) 2-Methylnaphthalene	12.322	142	700000	20.404	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BP012071.D
 Acq On : 12 Oct 2022 09:24
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020775

Quant Time: Oct 12 23:05:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 23:03:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.540	142	707029	20.579	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.687	216	351979	19.846	ng/ul	99
40) Hexachlorocyclopentadiene	12.663	237	155544	18.056	ng/ul	99
41) 2,4,6-Trichlorophenol	12.934	196	221498	20.733	ng/ul	98
42) 2,4,5-Trichlorophenol	13.004	196	242196	20.860	ng/ul	99
43) 1,1'-Biphenyl	13.334	154	874304	19.961	ng/ul	100
44) 2-Chloronaphthalene	13.375	162	709087	20.273	ng/ul	100
45) 2-Nitroaniline	13.587	65	163831	21.958	ng/ul	97
47) Dimethylphthalate	13.963	163	796880	20.585	ng/ul	100
48) 2,6-Dinitrotoluene	14.087	165	148122	21.611	ng/ul	97
50) Acenaphthylene	14.222	152	1075275	20.915	ng/ul	100
51) 3-Nitroaniline	14.416	138	173186	21.700	ng/ul	100
52) Acenaphthene	14.563	153	748563	20.453	ng/ul	100
53) 2,4-Dinitrophenol	14.622	184	92967	20.996	ng/ul	98
55) 4-Nitrophenol	14.722	109	112388	22.536	ng/ul	97
56) Dibenzofuran	14.898	168	1016129	20.543	ng/ul	98
57) 2,4-Dinitrotoluene	14.875	165	223777	22.718	ng/ul#	96
58) 2,3,4,6-Tetrachlorophenol	15.128	232	197021	21.771	ng/ul	98
59) Diethylphthalate	15.328	149	758622	21.398	ng/ul	99
61) Fluorene	15.551	166	829784	21.371	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.545	204	389473	20.735	ng/ul	99
63) 4-Nitroaniline	15.581	138	186653	25.613	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.634	198	127884	18.796	ng/ul	100
67) N-Nitrosodiphenylamine	15.763	169	694162	19.677	ng/ul	99
68) 4-Bromophenyl-phenylether	16.445	248	222655	19.198	ng/ul	98
69) Hexachlorobenzene	16.557	284	260060	19.319	ng/ul	100
70) Atrazine	16.722	200	217337	20.034	ng/ul	100
71) Pentachlorophenol	16.910	266	157622	19.985	ng/ul	98
72) Phenanthrene	17.304	178	1337397	20.633	ng/ul	99
74) Anthracene	17.392	178	1342336	20.927	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.298	216	349342	18.138	ng/ul	99
76) Pentachlorobenzene	14.816	250	350081	18.508	ng/ul	100
77) Carbazole	17.669	167	1251579	21.877	ng/ul	100
78) Di-n-butylphthalate	18.216	149	1220484	21.498	ng/ul	100
80) Fluoranthene	19.269	202	1561177	19.149	ng/ul	98
82) Pyrene	19.622	202	1678957	19.460	ng/ul	96
83) Butylbenzylphthalate	20.498	149	573234	20.959	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.269	252	523450	18.291	ng/ul#	99
85) Benzo(a)anthracene	21.333	228	1704264	20.848	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.251	149	801895	21.749	ng/ul	98
87) Chrysene	21.386	228	1611527	20.480	ng/ul	99
89) Di-n-octyl phthalate	22.180	149	1398423	20.300	ng/ul	100
90) Benzo(b)fluoranthene	23.027	252	1790644	20.559	ng/ul	99
91) Benzo(k)fluoranthene	23.074	252	1714172	20.009	ng/ul	99
93) Benzo(a)pyrene	23.657	252	1590339	20.466	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.280	276	1958122	19.930	ng/ul	95
95) Dibenzo(a,h)anthracene	26.298	278	1755104	19.912	ng/ul	97
96) Benzo(g,h,i)perylene	27.057	276	1719419	19.409	ng/ul#	95





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.052 ng/uL
RT: 3.311 min Scan# 51
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion: 88 Resp: 54131
Ion Ratio Lower Upper
88 100
43 21.1 16.6 25.0
58 67.8 52.6 79.0



#5
Pyridine
Concen: 20.853 ng/uL
RT: 3.717 min Scan# 124
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion: 79 Resp: 339460
Ion Ratio Lower Upper
79 100
52 66.0 51.2 76.8
51 28.5 22.8 34.2









#6
Benzaldehyde
Concen: 22.508 ng/ul
RT: 6.993 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion: 77 Resp: 213935
Ion Ratio Lower Upper
77 100
105 100.7 85.2 127.8
106 100.4 83.1 124.7



#8
Phenol
Concen: 19.919 ng/ul
RT: 7.046 min Scan# 690
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion: 94 Resp: 423631
Ion Ratio Lower Upper
94 100
65 25.1 19.4 29.0
66 31.7 24.6 37.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 19.916 ng/ul
 RT: 7.281 min Scan# 71
 Delta R.T. 0.000 min
 Lab File: BP012071.D
 Acq: 12 Oct 2022 09:24




Instrument :
 BNA_P
 ClientSampleId :
 SST020775


Tgt Ion: 93 Resp: 334172
 Ion Ratio Lower Upper
 93 100
 63 68.9 56.7 85.1
 95 32.3 26.2 39.2



#12
 2-Chlorophenol
 Concen: 20.517 ng/ul
 RT: 7.416 min Scan# 753
 Delta R.T. 0.000 min
 Lab File: BP012071.D
 Acq: 12 Oct 2022 09:24

Tgt Ion:128 Resp: 358057
 Ion Ratio Lower Upper
 128 100
 64 39.6 32.7 49.1
 130 32.1 26.3 39.5





#13

2-Methylphenol

Concen: 19.658 ng/ul

RT: 8.299 min Scan# 905

Delta R.T. 0.000 min

Lab File: BP012071.D

Acq: 12 Oct 2022 09:24

Instrument :

BNA_P

ClientSampleId :





SSTD020775

Tgt Ion:108 Resp: 311818

Ion Ratio Lower Upper

108 100

107 88.0 71.9 107.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.956 ng/ul

RT: 8.381 min Scan# 917

Delta R.T. 0.000 min

Lab File: BP012071.D

Acq: 12 Oct 2022 09:24




Tgt Ion: 45 Resp: 343215


Ion Ratio Lower Upper

45 100

77 17.7 14.4 21.6

79 13.4 11.3 16.9









#16
Acetophenone
Concen: 19.835 ng/ul
RT: 8.681 min Scan# 90
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:105 Resp: 475833
Ion Ratio Lower Upper
105 100
77 72.8 57.0 85.4
51 20.9 16.4 24.6



#17
N-Nitroso-di-n-propylamine
Concen: 19.529 ng/ul
RT: 8.663 min Scan# 965
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion: 70 Resp: 213377
Ion Ratio Lower Upper
70 100
42 43.1 34.6 52.0
101 10.0 8.0 12.0
130 22.7 18.3 27.5









#18
4-Methylphenol
Concen: 19.780 ng/ul
RT: 8.628 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:108 Resp: 338171
Ion Ratio Lower Upper
108 100
107 110.0 86.3 129.5



#19
Hexachloroethane
Concen: 19.725 ng/ul
RT: 8.934 min Scan# 1011
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:117 Resp: 132877
Ion Ratio Lower Upper
117 100
201 102.1 82.3 123.5
199 62.4 50.0 75.0









#22
Nitrobenzene
Concen: 21.024 ng/ul
RT: 9.063 min Scan# 1033
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion: 77 Resp: 343178
Ion Ratio Lower Upper
77 100
123 49.2 39.4 59.0
65 12.1 9.9 14.9



#23
Isophorone
Concen: 20.267 ng/ul
RT: 9.593 min Scan# 1123
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion: 82 Resp: 586800
Ion Ratio Lower Upper
82 100
95 6.6 5.4 8.0
138 19.1 15.2 22.8









#25
2-Nitrophenol
Concen: 20.887 ng/ul
RT: 9.775 min Scan# 1154
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:139 Resp: 177177
Ion Ratio Lower Upper
139 100
65 44.8 36.3 54.5
109 25.2 21.2 31.8



#26
2,4-Dimethylphenol
Concen: 20.760 ng/ul
RT: 9.834 min Scan# 1164
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:107 Resp: 342965
Ion Ratio Lower Upper
107 100
121 53.5 43.0 64.4
122 90.9 71.4 107.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.173 ng/ul
 RT: 10.075 min Scan# 1205
 Delta R.T. 0.000 min
 Lab File: BP012071.D
 Acq: 12 Oct 2022 09:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020775


Tgt Ion: 93 Resp: 400805
 Ion Ratio Lower Upper
 93 100
 95 32.8 25.7 38.5
 123 11.2 9.2 13.8



#29
 2,4-Dichlorophenol
 Concen: 20.800 ng/ul
 RT: 10.310 min Scan# 1245
 Delta R.T. 0.000 min
 Lab File: BP012071.D
 Acq: 12 Oct 2022 09:24

Tgt Ion: 162 Resp: 305867
 Ion Ratio Lower Upper
 162 100
 164 65.8 51.5 77.3
 98 33.4 26.3 39.5









#30
Naphthalene
Concen: 20.526 ng/ul
RT: 10.710 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:128 Resp: 1093585
Ion Ratio Lower Upper
128 100
129 10.8 8.9 13.3
127 13.1 10.2 15.4



#32
4-Chloroaniline
Concen: 20.741 ng/ul
RT: 10.828 min Scan# 1333
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:127 Resp: 440960
Ion Ratio Lower Upper
127 100
129 32.0 25.6 38.4









#33
Hexachlorobutadiene
Concen: 19.680 ng/ul
RT: 10.993 min Scan# 1361
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:225 Resp: 179457
Ion Ratio Lower Upper
225 100
223 61.8 50.9 76.3
227 65.4 51.1 76.7



#34
Caprolactam
Concen: 22.637 ng/ul
RT: 11.616 min Scan# 1467
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:113 Resp: 91473
Ion Ratio Lower Upper
113 100
55 149.3 115.2 172.8
56 107.8 85.6 128.4









#35
4-Chloro-3-methylphenol
Concen: 21.574 ng/ul
RT: 11.951 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:107 Resp: 300558
Ion Ratio Lower Upper
107 100
144 28.4 23.0 34.4
142 86.4 69.8 104.8



#36
2-Methylnaphthalene
Concen: 20.404 ng/ul
RT: 12.322 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:142 Resp: 700000
Ion Ratio Lower Upper
142 100
141 88.2 70.7 106.1









#37
1-Methylnaphthalene
Concen: 20.579 ng/ul
RT: 12.540 min Scan# 1625
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:142 Resp: 707029
Ion Ratio Lower Upper
142 100
141 91.5 72.9 109.3
116 3.5 2.8 4.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.846 ng/ul
RT: 12.687 min Scan# 1649
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:216 Resp: 351979
Ion Ratio Lower Upper
216 100
214 78.9 63.0 94.4
179 21.0 16.3 24.5
108 17.2 12.5 18.7









#40
Hexachlorocyclopentadiene
Concen: 18.056 ng/ul
RT: 12.663 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:237 Resp: 155544
Ion Ratio Lower Upper
237 100
235 63.4 49.8 74.8
272 12.7 9.9 14.9



#41
2,4,6-Trichlorophenol
Concen: 20.733 ng/ul
RT: 12.934 min Scan# 1691
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:196 Resp: 221498
Ion Ratio Lower Upper
196 100
198 97.4 76.2 114.2
200 31.0 24.2 36.2









#42
2,4,5-Trichlorophenol
Concen: 20.860 ng/ul
RT: 13.004 min Scan# 1703
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:196 Resp: 242196
Ion Ratio Lower Upper
196 100
198 95.1 75.4 113.2
200 30.1 24.8 37.2



#43
1,1'-Biphenyl
Concen: 19.961 ng/ul
RT: 13.334 min Scan# 1759
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:154 Resp: 874304
Ion Ratio Lower Upper
154 100
153 39.9 32.2 48.2
76 12.9 10.3 15.5









#44
2-Chloronaphthalene
Concen: 20.273 ng/ul
RT: 13.375 min Scan# 1766
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:162 Resp: 709087
Ion Ratio Lower Upper
162 100
164 32.5 26.0 39.0
127 37.3 29.4 44.2



#45
2-Nitroaniline
Concen: 21.958 ng/ul
RT: 13.587 min Scan# 1802
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion: 65 Resp: 163831
Ion Ratio Lower Upper
65 100
92 72.4 61.6 92.4
138 119.1 96.5 144.7









#47
Dimethylphthalate
Concen: 20.585 ng/ul
RT: 13.963 min Scan# 1866
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:163 Resp: 796880
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.0 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 21.611 ng/ul
RT: 14.087 min Scan# 1887
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:165 Resp: 148122
Ion Ratio Lower Upper
165 100
89 52.0 43.4 65.0
121 21.3 18.1 27.1









#50
Acenaphthylene
Concen: 20.915 ng/ul
RT: 14.222 min Scan# 1910
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:152 Resp: 1075275
Ion Ratio Lower Upper
152 100
151 19.6 15.7 23.5
153 13.0 10.4 15.6



#51
3-Nitroaniline
Concen: 21.700 ng/ul
RT: 14.416 min Scan# 1943
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:138 Resp: 173186
Ion Ratio Lower Upper
138 100
108 10.7 8.5 12.7
92 107.6 85.8 128.8









#52
Acenaphthene
Concen: 20.453 ng/ul
RT: 14.563 min Scan# 1968
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:153 Resp: 748563
Ion Ratio Lower Upper
153 100
152 47.7 38.2 57.2
154 88.8 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 20.996 ng/ul
RT: 14.622 min Scan# 1978
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:184 Resp: 92967
Ion Ratio Lower Upper
184 100
63 55.6 45.5 68.3
154 56.7 47.0 70.6









#55
4-Nitrophenol
Concen: 22.536 ng/ul
RT: 14.722 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:109 Resp: 112388
Ion Ratio Lower Upper
109 100
139 157.9 124.1 186.1
65 134.0 104.2 156.4



#56
Dibenzofuran
Concen: 20.543 ng/ul
RT: 14.898 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:168 Resp: 1016129
Ion Ratio Lower Upper
168 100
139 36.7 28.2 42.4





#57

2,4-Dinitrotoluene

Concen: 22.718 ng/ul

RT: 14.875 min Scan# 2021

Delta R.T. 0.000 min

Lab File: BP012071.D

Acq: 12 Oct 2022 09:24

Instrument :

BNA_P

ClientSampleId :

SSTD020775





Tgt Ion:165 Resp: 223777

Ion Ratio Lower Upper

165 100

63 36.4 31.1 46.7

182 2.8 1.8 2.6#



#58

2,3,4,6-Tetrachlorophenol

Concen: 21.771 ng/ul

RT: 15.128 min Scan# 2064

Delta R.T. 0.000 min

Lab File: BP012071.D

Acq: 12 Oct 2022 09:24

Tgt Ion:232 Resp: 197021




Ion Ratio Lower Upper


232 100

131 46.6 37.8 56.6

130 2.3 1.8 2.6

166 31.6 26.6 39.8









#59
Diethylphthalate
Concen: 21.398 ng/ul
RT: 15.328 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:149 Resp: 758622
Ion Ratio Lower Upper
149 100
177 22.6 17.5 26.3
150 12.8 9.8 14.8



#61
Fluorene
Concen: 21.371 ng/ul
RT: 15.551 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:166 Resp: 829784
Ion Ratio Lower Upper
166 100
165 98.2 79.5 119.3
167 13.3 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 20.735 ng/ul
RT: 15.545 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:204 Resp: 389473
Ion Ratio Lower Upper
204 100
206 32.3 25.4 38.2
141 57.4 46.2 69.2



#63
4-Nitroaniline
Concen: 25.613 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:138 Resp: 186653
Ion Ratio Lower Upper
138 100
92 47.3 39.2 58.8
108 65.2 55.4 83.0









#66
4,6-Dinitro-2-methylphenol
Concen: 18.796 ng/ul
RT: 15.634 min Scan# 2151
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:198 Resp: 127884
Ion Ratio Lower Upper
198 100
182 6.2 4.9 7.3
77 25.0 20.0 30.0



#67
N-Nitrosodiphenylamine
Concen: 19.677 ng/ul
RT: 15.763 min Scan# 2172
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:169 Resp: 694162
Ion Ratio Lower Upper
169 100
168 67.3 53.5 80.3
167 36.2 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 19.198 ng/ul
RT: 16.445 min Scan# 211
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:248 Resp: 222655
Ion Ratio Lower Upper
248 100
250 99.0 77.9 116.9
141 73.3 59.5 89.3



#69
Hexachlorobenzene
Concen: 19.319 ng/ul
RT: 16.557 min Scan# 2307
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:284 Resp: 260060
Ion Ratio Lower Upper
284 100
142 38.3 30.5 45.7
249 31.5 25.4 38.0





#70

Atrazine

Concen: 20.034 ng/ul

RT: 16.722 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP012071.D

Acq: 12 Oct 2022 09:24

Instrument :

BNA_P

ClientSampleId :

SSTD020775





Tgt Ion:200 Resp: 217337

Ion Ratio Lower Upper

200 100

173 27.6 22.1 33.1

215 48.8 38.8 58.2



#71

Pentachlorophenol

Concen: 19.985 ng/ul

RT: 16.910 min Scan# 2367

Delta R.T. 0.000 min

Lab File: BP012071.D

Acq: 12 Oct 2022 09:24




Tgt Ion:266 Resp: 157622


Ion Ratio Lower Upper

266 100

264 61.4 50.6 75.8

268 63.4 51.8 77.6









#72
Phenanthrene
Concen: 20.633 ng/ul
RT: 17.304 min Scan# 2434
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:178 Resp: 1337397
Ion Ratio Lower Upper
178 100
179 15.6 12.2 18.4
176 19.3 15.8 23.6



#74
Anthracene
Concen: 20.927 ng/ul
RT: 17.392 min Scan# 2449
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:178 Resp: 1342336
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.4
176 18.7 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.138 ng/uL
RT: 13.298 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:216 Resp: 349342
Ion Ratio Lower Upper
216 100
214 79.6 62.6 94.0
218 48.9 38.9 58.3



#76
Pentachlorobenzene
Concen: 18.508 ng/uL
RT: 14.816 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:250 Resp: 350081
Ion Ratio Lower Upper
250 100
248 64.0 51.4 77.2
252 63.6 50.7 76.1









#77
 Carbazole
 Concen: 21.877 ng/ul
 RT: 17.669 min Scan# 2497
 Delta R.T. 0.000 min
 Lab File: BP012071.D
 Acq: 12 Oct 2022 09:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020775


Tgt Ion:167 Resp: 1251579
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.0 25.6
 139 12.4 10.1 15.1



#78
 Di-n-butylphthalate
 Concen: 21.498 ng/ul
 RT: 18.216 min Scan# 2589
 Delta R.T. 0.000 min
 Lab File: BP012071.D
 Acq: 12 Oct 2022 09:24

Tgt Ion:149 Resp: 1220484
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.2 10.8
 104 4.5 3.4 5.2









#80
Fluoranthene
Concen: 19.149 ng/ul
RT: 19.269 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:202 Resp: 1561177
Ion Ratio Lower Upper
202 100
101 12.5 9.0 13.6
100 9.4 7.0 10.6



#82
Pyrene
Concen: 19.460 ng/ul
RT: 19.622 min Scan# 2828
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:202 Resp: 1678957
Ion Ratio Lower Upper
202 100
101 14.7 10.4 15.6
100 12.0 8.5 12.7





#83

Butylbenzylphthalate

Concen: 20.959 ng/ul

RT: 20.498 min Scan# 2977

Delta R.T. 0.000 min

Lab File: BP012071.D

Acq: 12 Oct 2022 09:24

Instrument :

BNA_P

ClientSampleId :

SSTD020775





Tgt Ion:149 Resp: 573234

Ion Ratio Lower Upper

149 100

91 65.4 48.5 72.7

206 20.2 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 18.291 ng/ul

RT: 21.269 min Scan# 3108

Delta R.T. 0.000 min

Lab File: BP012071.D

Acq: 12 Oct 2022 09:24




Tgt Ion:252 Resp: 523450


Ion Ratio Lower Upper

252 100

254 63.9 51.4 77.2

126 11.8 7.8 11.8#





#85

Benzo(a)anthracene

Concen: 20.848 ng/ul

RT: 21.333 min Scan# 3119

Delta R.T. 0.000 min

Lab File: BP012071.D

Acq: 12 Oct 2022 09:24

Instrument :

BNA_P

ClientSampleId :

SSTD020775





Tgt Ion:228 Resp: 1704264

Ion Ratio Lower Upper

228 100

229 19.6 15.7 23.5

226 27.5 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.749 ng/ul

RT: 21.251 min Scan# 3105

Delta R.T. 0.000 min

Lab File: BP012071.D

Acq: 12 Oct 2022 09:24




Tgt Ion:149 Resp: 801895


Ion Ratio Lower Upper

149 100

167 26.4 21.9 32.9

279 3.5 3.4 5.2





#87

Chrysene

Concen: 20.480 ng/ul

RT: 21.386 min Scan# 31

Delta R.T. 0.000 min

Lab File: BP012071.D

Acq: 12 Oct 2022 09:24

Instrument :

BNA_P

ClientSampleId :

SSTD020775





Tgt Ion:228 Resp: 1611527

Ion Ratio Lower Upper

228 100

226 30.4 24.0 36.0

229 19.9 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 20.300 ng/ul




RT: 22.180 min Scan# 3263


Delta R.T. 0.000 min

Lab File: BP012071.D

Acq: 12 Oct 2022 09:24

Tgt Ion:149 Resp: 1398423









#90
Benzo(b)fluoranthene
Concen: 20.559 ng/ul
RT: 23.027 min Scan# 3407
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:252 Resp: 1790644
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 11.3 7.8 11.8



#91
Benzo(k)fluoranthene
Concen: 20.009 ng/ul
RT: 23.074 min Scan# 3415
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:252 Resp: 1714172
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 11.2 8.2 12.2









#93
Benzo(a)pyrene
Concen: 20.466 ng/ul
RT: 23.657 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24




Instrument :
BNA_P
ClientSampleId :
SSTD020775


Tgt Ion:252 Resp: 1590339
Ion Ratio Lower Upper
252 100
253 22.5 17.6 26.4
125 12.8 8.9 13.3



#94
Indeno(1,2,3-cd)pyrene
Concen: 19.930 ng/ul
RT: 26.280 min Scan# 3960
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:276 Resp: 1958122
Ion Ratio Lower Upper
276 100
138 28.5 19.1 28.7
277 25.0 19.9 29.9









#95
Dibenzo(a,h)anthracene
Concen: 19.912 ng/ul
RT: 26.298 min Scan# 3963
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Instrument :
BNA_P
ClientSampleId :
SSTD020775

Tgt Ion:278 Resp: 1755104
Ion Ratio Lower Upper
278 100
139 18.5 12.5 18.7
279 24.0 19.0 28.4



#96
Benzo(g,h,i)perylene
Concen: 19.409 ng/ul
RT: 27.057 min Scan# 4092
Delta R.T. 0.000 min
Lab File: BP012071.D
Acq: 12 Oct 2022 09:24

Tgt Ion:276 Resp: 1719419
Ion Ratio Lower Upper
276 100
138 25.1 16.5 24.7#
277 23.9 19.0 28.4

