

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BPCCC089.D
 Acq On : 12 Oct 2022 20:22
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020776

Quant Time: Oct 12 23:11:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 23:03:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.852	152	392076	20.000	ng/u1	0.00
20) Naphthalene-d8	10.663	136	1639306	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.504	164	938445	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.257	188	1634549	20.000	ng/u1	0.00
79) Chrysene-d12	21.351	240	1223008	20.000	ng/u1	0.00
88) Perylene-d12	23.768	264	1425807	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	71518	7.192	ng/uL	0.00
4) Pyridine-d5	3.699	84	506382	19.147	ng/u1	0.00
7) Phenol-d5	7.022	99	629365	19.959	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.187	67	355938	19.886	ng/u1	0.00
11) 2-Chlorophenol-d4	7.387	132	522970	20.894	ng/u1	0.00
15) 4-Methylphenol-d8	8.569	113	501834	20.720	ng/u1	0.00
21) Nitrobenzene-d5	9.022	128	256292	22.137	ng/u1	0.00
24) 2-Nitrophenol-d4	9.746	143	279678	23.824	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.281	165	506022	22.241	ng/u1	0.00
31) 4-Chloroaniline-d4	10.804	131	688113	20.524	ng/u1	0.00
46) Dimethylphthalate-d6	13.916	166	1267111	21.320	ng/u1	0.00
49) Acenaphthylene-d8	14.192	160	1565700	21.304	ng/u1	0.00
54) 4-Nitrophenol-d4	14.710	143	220705	18.915	ng/u1	0.00
60) Fluorene-d10	15.498	176	1093361	20.333	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.622	200	169992	19.696	ng/u1	0.00
73) Anthracene-d10	17.357	188	1468348	20.687	ng/u1	0.00
81) Pyrene-d10	19.592	212	1316368	21.010	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.610	264	1418144	20.460	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.311	88	74999	7.167	ng/uL	98
5) Pyridine	3.722	79	494052	19.497	ng/u1	95
6) Benzaldehyde	6.999	77	328928	22.233	ng/u1	99
8) Phenol	7.052	94	656634	19.835	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.281	93	524907	20.097	ng/u1	98
12) 2-Chlorophenol	7.416	128	560076	20.618	ng/u1	97
13) 2-Methylphenol	8.304	108	507387	20.550	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.387	45	525604	19.633	ng/u1	98
16) Acetophenone	8.687	105	770948	20.646	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.669	70	355181	20.884	ng/u1	98
18) 4-Methylphenol	8.634	108	549405	20.645	ng/u1	96
19) Hexachloroethane	8.934	117	213914	20.400	ng/u1	97
22) Nitrobenzene	9.063	77	547339	20.623	ng/u1	97
23) Isophorone	9.598	82	1030729	21.895	ng/u1	99
25) 2-Nitrophenol	9.775	139	314385	22.794	ng/u1	95
26) 2,4-Dimethylphenol	9.840	107	568090	21.149	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.075	93	676944	20.955	ng/u1	100
29) 2,4-Dichlorophenol	10.310	162	515908	21.578	ng/u1	99
30) Naphthalene	10.710	128	1738429	20.068	ng/u1	100
32) 4-Chloroaniline	10.828	127	702572	20.325	ng/u1	98
33) Hexachlorobutadiene	10.993	225	313817	21.166	ng/u1	99
34) Caprolactam	11.628	113	160568	24.439	ng/u1	97
35) 4-Chloro-3-methylphenol	11.957	107	502499	22.184	ng/u1	99
36) 2-Methylnaphthalene	12.322	142	1157987	20.759	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BPCCC089.D
 Acq On : 12 Oct 2022 20:22
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020776


Quant Time: Oct 12 23:11:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 23:03:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.540	142	1154849	20.673	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.692	216	599309	20.892	ng/ul	99
40) Hexachlorocyclopentadiene	12.663	237	272486	19.556	ng/ul	98
41) 2,4,6-Trichlorophenol	12.934	196	391307	22.645	ng/ul	99
42) 2,4,5-Trichlorophenol	13.004	196	419017	22.312	ng/ul	98
43) 1,1'-Biphenyl	13.334	154	1450300	20.471	ng/ul	100
44) 2-Chloronaphthalene	13.375	162	1155031	20.416	ng/ul	99
45) 2-Nitroaniline	13.587	65	275684	22.843	ng/ul	97
47) Dimethylphthalate	13.963	163	1333363	21.294	ng/ul	100
48) 2,6-Dinitrotoluene	14.086	165	270943	24.439	ng/ul	94
50) Acenaphthylene	14.222	152	1721833	20.705	ng/ul	98
51) 3-Nitroaniline	14.416	138	291673	22.594	ng/ul	98
52) Acenaphthene	14.569	153	1192240	20.139	ng/ul	99
53) 2,4-Dinitrophenol	14.628	184	140028	19.551	ng/ul	92
55) 4-Nitrophenol	14.728	109	152186	18.866	ng/ul	94
56) Dibenzofuran	14.904	168	1606715	20.082	ng/ul	100
57) 2,4-Dinitrotoluene	14.875	165	362497	22.751	ng/ul#	95
58) 2,3,4,6-Tetrachlorophenol	15.128	232	319712	21.841	ng/ul	97
59) Diethylphthalate	15.328	149	1266466	22.084	ng/ul	99
61) Fluorene	15.551	166	1271734	20.249	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.545	204	631604	20.789	ng/ul	98
63) 4-Nitroaniline	15.581	138	277297	23.525	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.639	198	179796	19.269	ng/ul#	93
67) N-Nitrosodiphenylamine	15.763	169	1074212	22.203	ng/ul	99
68) 4-Bromophenyl-phenylether	16.445	248	368221	23.151	ng/ul	97
69) Hexachlorobenzene	16.557	284	399158	21.622	ng/ul	99
70) Atrazine	16.722	200	342400	23.015	ng/ul	99
71) Pentachlorophenol	16.910	266	215979	19.968	ng/ul	99
72) Phenanthrene	17.304	178	1794736	20.190	ng/ul	99
74) Anthracene	17.392	178	1798974	20.451	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.298	216	590600	22.360	ng/ul	100
76) Pentachlorobenzene	14.822	250	583144	22.480	ng/ul	99
77) Carbazole	17.669	167	1535150	19.567	ng/ul	100
78) Di-n-butylphthalate	18.216	149	1856952	23.851	ng/ul	100
80) Fluoranthene	19.269	202	1663966	21.674	ng/ul	98
82) Pyrene	19.627	202	1707299	21.013	ng/ul	100
83) Butylbenzylphthalate	20.498	149	612248	23.771	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.268	252	460992	17.105	ng/ul#	98
85) Benzo(a)anthracene	21.333	228	1602395	20.815	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.251	149	822311	23.683	ng/ul	99
87) Chrysene	21.386	228	1504340	20.301	ng/ul	99
89) Di-n-octyl phthalate	22.180	149	1386036	19.417	ng/ul	100
90) Benzo(b)fluoranthene	23.027	252	1795756	19.897	ng/ul	98
91) Benzo(k)fluoranthene	23.074	252	1744435	19.650	ng/ul	99
93) Benzo(a)pyrene	23.662	252	1638387	20.348	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.286	276	2064795	20.281	ng/ul	97
95) Dibenzo(a,h)anthracene	26.304	278	1846020	20.211	ng/ul	98
96) Benzo(g,h,i)perylene	27.056	276	1830996	19.947	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 12 23:11:04 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 12 23:03:21 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.167 ng/uL
RT: 3.311 min Scan# 51
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion: 88 Resp: 74999
Ion Ratio Lower Upper
88 100
43 20.3 16.6 25.0
58 67.4 52.6 79.0



#5
Pyridine
Concen: 19.497 ng/uL
RT: 3.722 min Scan# 125
Delta R.T. 0.006 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion: 79 Resp: 494052
Ion Ratio Lower Upper
79 100
52 59.5 51.2 76.8
51 26.3 22.8 34.2









#6
Benzaldehyde
Concen: 22.233 ng/ul
RT: 6.999 min Scan# 682
Delta R.T. 0.006 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion: 77 Resp: 328928
Ion Ratio Lower Upper
77 100
105 108.7 85.2 127.8
106 104.3 83.1 124.7



#8
Phenol
Concen: 19.835 ng/ul
RT: 7.052 min Scan# 691
Delta R.T. 0.006 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion: 94 Resp: 656634
Ion Ratio Lower Upper
94 100
65 24.1 19.4 29.0
66 31.4 24.6 37.0









#10
Bis(2-Chloroethyl)ether
Concen: 20.097 ng/ul
RT: 7.281 min Scan# 71
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion: 93 Resp: 524907
Ion Ratio Lower Upper
93 100
63 69.1 56.7 85.1
95 32.8 26.2 39.2



#12
2-Chlorophenol
Concen: 20.618 ng/ul
RT: 7.416 min Scan# 753
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:128 Resp: 560076
Ion Ratio Lower Upper
128 100
64 39.2 32.7 49.1
130 31.6 26.3 39.5





#13

2-Methylphenol

Concen: 20.550 ng/ul

RT: 8.304 min Scan# 905

Delta R.T. 0.006 min

Lab File: BPCCC089.D

Acq: 12 Oct 2022 20:22

Instrument :

BNA_P

ClientSampleId :





SSTD020776

Tgt Ion:108 Resp: 507387

Ion Ratio Lower Upper

108 100

107 89.9 71.9 107.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.633 ng/ul

RT: 8.387 min Scan# 918

Delta R.T. 0.006 min

Lab File: BPCCC089.D

Acq: 12 Oct 2022 20:22




Tgt Ion: 45 Resp: 525604


Ion Ratio Lower Upper

45 100

77 18.8 14.4 21.6

79 14.9 11.3 16.9









#16
Acetophenone
Concen: 20.646 ng/ul
RT: 8.687 min Scan# 90
Delta R.T. 0.006 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:105 Resp: 770948
Ion Ratio Lower Upper
105 100
77 70.4 57.0 85.4
51 19.2 16.4 24.6



#17
N-Nitroso-di-n-propylamine
Concen: 20.884 ng/ul
RT: 8.669 min Scan# 966
Delta R.T. 0.006 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion: 70 Resp: 355181
Ion Ratio Lower Upper
70 100
42 42.0 34.6 52.0
101 10.0 8.0 12.0
130 24.8 18.3 27.5









#18
4-Methylphenol
Concen: 20.645 ng/ul
RT: 8.634 min Scan# 90
Delta R.T. 0.006 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:108 Resp: 549405
Ion Ratio Lower Upper
108 100
107 111.6 86.3 129.5



#19
Hexachloroethane
Concen: 20.400 ng/ul
RT: 8.934 min Scan# 1011
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:117 Resp: 213914
Ion Ratio Lower Upper
117 100
201 105.8 82.3 123.5
199 65.8 50.0 75.0









#22
Nitrobenzene
Concen: 20.623 ng/ul
RT: 9.063 min Scan# 1033
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion: 77 Resp: 547339
Ion Ratio Lower Upper
77 100
123 51.4 39.4 59.0
65 12.8 9.9 14.9



#23
Isophorone
Concen: 21.895 ng/ul
RT: 9.598 min Scan# 1124
Delta R.T. 0.006 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion: 82 Resp: 1030729
Ion Ratio Lower Upper
82 100
95 6.7 5.4 8.0
138 19.7 15.2 22.8









#25
2-Nitrophenol
Concen: 22.794 ng/ul
RT: 9.775 min Scan# 1154
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:139 Resp: 314385
Ion Ratio Lower Upper
139 100
65 41.4 36.3 54.5
109 24.7 21.2 31.8



#26
2,4-Dimethylphenol
Concen: 21.149 ng/ul
RT: 9.840 min Scan# 1165
Delta R.T. 0.006 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:107 Resp: 568090
Ion Ratio Lower Upper
107 100
121 53.5 43.0 64.4
122 90.8 71.4 107.2





#27

Bis(2-Chloroethoxy)methane

Concen: 20.955 ng/ul

RT: 10.075 min Scan# 1205

Delta R.T. -0.000 min

Lab File: BPCCC089.D

Acq: 12 Oct 2022 20:22

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion: 93 Resp: 676944

Ion Ratio Lower Upper

93 100

95 32.0 25.7 38.5

123 11.7 9.2 13.8



#29

2,4-Dichlorophenol

Concen: 21.578 ng/ul

RT: 10.310 min Scan# 1245

Delta R.T. -0.000 min

Lab File: BPCCC089.D

Acq: 12 Oct 2022 20:22




Tgt Ion:162 Resp: 515908


Ion Ratio Lower Upper

162 100

164 65.2 51.5 77.3

98 32.8 26.3 39.5









#30
Naphthalene
Concen: 20.068 ng/ul
RT: 10.710 min Scan# 1313
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:128 Resp: 1738429
Ion Ratio Lower Upper
128 100
129 11.2 8.9 13.3
127 13.0 10.2 15.4



#32
4-Chloroaniline
Concen: 20.325 ng/ul
RT: 10.828 min Scan# 1333
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:127 Resp: 702572
Ion Ratio Lower Upper
127 100
129 33.0 25.6 38.4









#33
Hexachlorobutadiene
Concen: 21.166 ng/ul
RT: 10.993 min Scan# 11
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:225 Resp: 313817
Ion Ratio Lower Upper
225 100
223 62.6 50.9 76.3
227 64.6 51.1 76.7



#34
Caprolactam
Concen: 24.439 ng/ul
RT: 11.628 min Scan# 1469
Delta R.T. 0.012 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:113 Resp: 160568
Ion Ratio Lower Upper
113 100
55 140.0 115.2 172.8
56 103.5 85.6 128.4









#35
4-Chloro-3-methylphenol
Concen: 22.184 ng/ul
RT: 11.957 min Scan# 1525
Delta R.T. 0.006 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:107 Resp: 502499
Ion Ratio Lower Upper
107 100
144 28.1 23.0 34.4
142 87.8 69.8 104.8



#36
2-Methylnaphthalene
Concen: 20.759 ng/ul
RT: 12.322 min Scan# 1587
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:142 Resp: 1157987
Ion Ratio Lower Upper
142 100
141 88.5 70.7 106.1









#37
1-Methylnaphthalene
Concen: 20.673 ng/ul
RT: 12.540 min Scan# 1625
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:142 Resp: 1154849
Ion Ratio Lower Upper
142 100
141 91.4 72.9 109.3
116 3.6 2.8 4.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.892 ng/ul
RT: 12.692 min Scan# 1650
Delta R.T. 0.006 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:216 Resp: 599309
Ion Ratio Lower Upper
216 100
214 77.4 63.0 94.4
179 20.3 16.3 24.5
108 15.6 12.5 18.7









#40
Hexachlorocyclopentadiene
Concen: 19.556 ng/ul
RT: 12.663 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:237 Resp: 272486
Ion Ratio Lower Upper
237 100
235 64.3 49.8 74.8
272 12.7 9.9 14.9



#41
2,4,6-Trichlorophenol
Concen: 22.645 ng/ul
RT: 12.934 min Scan# 1691
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:196 Resp: 391307
Ion Ratio Lower Upper
196 100
198 96.0 76.2 114.2
200 31.2 24.2 36.2









#42
2,4,5-Trichlorophenol
Concen: 22.312 ng/ul
RT: 13.004 min Scan# 1704
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:196 Resp: 419017
Ion Ratio Lower Upper
196 100
198 96.7 75.4 113.2
200 30.4 24.8 37.2



#43
1,1'-Biphenyl
Concen: 20.471 ng/ul
RT: 13.334 min Scan# 1759
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:154 Resp: 1450300
Ion Ratio Lower Upper
154 100
153 40.2 32.2 48.2
76 12.3 10.3 15.5









#44
2-Chloronaphthalene
Concen: 20.416 ng/ul
RT: 13.375 min Scan# 11
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:162 Resp: 1155031
Ion Ratio Lower Upper
162 100
164 32.6 26.0 39.0
127 37.3 29.4 44.2



#45
2-Nitroaniline
Concen: 22.843 ng/ul
RT: 13.587 min Scan# 1802
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion: 65 Resp: 275684
Ion Ratio Lower Upper
65 100
92 75.7 61.6 92.4
138 125.1 96.5 144.7









#47
Dimethylphthalate
Concen: 21.294 ng/ul
RT: 13.963 min Scan# 1866
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:163 Resp: 1333363
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.0 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 24.439 ng/ul
RT: 14.086 min Scan# 1887
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:165 Resp: 270943
Ion Ratio Lower Upper
165 100
89 49.2 43.4 65.0
121 20.7 18.1 27.1









#50
Acenaphthylene
Concen: 20.705 ng/ul
RT: 14.222 min Scan# 1910
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:152 Resp: 1721833
Ion Ratio Lower Upper
152 100
151 20.4 15.7 23.5
153 13.5 10.4 15.6



#51
3-Nitroaniline
Concen: 22.594 ng/ul
RT: 14.416 min Scan# 1943
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:138 Resp: 291673
Ion Ratio Lower Upper
138 100
108 10.6 8.5 12.7
92 105.6 85.8 128.8









#52
Acenaphthene
Concen: 20.139 ng/ul
RT: 14.569 min Scan# 1969
Delta R.T. 0.006 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:153 Resp: 1192240
Ion Ratio Lower Upper
153 100
152 47.5 38.2 57.2
154 89.2 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 19.551 ng/ul
RT: 14.628 min Scan# 1979
Delta R.T. 0.006 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:184 Resp: 140028
Ion Ratio Lower Upper
184 100
63 49.1 45.5 68.3
154 55.1 47.0 70.6









#55
4-Nitrophenol
Concen: 18.866 ng/ul
RT: 14.728 min Scan# 1996
Delta R.T. 0.006 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:109 Resp: 152186
Ion Ratio Lower Upper
109 100
139 164.7 124.1 186.1
65 126.4 104.2 156.4



#56
Dibenzofuran
Concen: 20.082 ng/ul
RT: 14.904 min Scan# 2026
Delta R.T. 0.006 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:168 Resp: 1606715
Ion Ratio Lower Upper
168 100
139 35.1 28.2 42.4









#57
2,4-Dinitrotoluene
Concen: 22.751 ng/ul
RT: 14.875 min Scan# 2021
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:165 Resp: 362497
Ion Ratio Lower Upper
165 100
63 35.5 31.1 46.7
182 2.7 1.8 2.6#



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.841 ng/ul
RT: 15.128 min Scan# 2064
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:232 Resp: 319712
Ion Ratio Lower Upper
232 100
131 45.7 37.8 56.6
130 2.2 1.8 2.6
166 30.3 26.6 39.8









#59
Diethylphthalate
Concen: 22.084 ng/ul
RT: 15.328 min Scan# 2098
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:149 Resp: 1266466
Ion Ratio Lower Upper
149 100
177 22.5 17.5 26.3
150 12.1 9.8 14.8



#61
Fluorene
Concen: 20.249 ng/ul
RT: 15.551 min Scan# 2136
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:166 Resp: 1271734
Ion Ratio Lower Upper
166 100
165 97.1 79.5 119.3
167 13.6 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 20.789 ng/ul
RT: 15.545 min Scan# 2135
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:204 Resp: 631604
Ion Ratio Lower Upper
204 100
206 33.2 25.4 38.2
141 56.9 46.2 69.2



#63
4-Nitroaniline
Concen: 23.525 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:138 Resp: 277297
Ion Ratio Lower Upper
138 100
92 47.4 39.2 58.8
108 66.0 55.4 83.0









#66
4,6-Dinitro-2-methylphenol
Concen: 19.269 ng/ul
RT: 15.639 min Scan# 2151
Delta R.T. 0.006 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:198 Resp: 179796
Ion Ratio Lower Upper
198 100
182 4.4 4.9 7.3#
77 21.1 20.0 30.0



#67
N-Nitrosodiphenylamine
Concen: 22.203 ng/ul
RT: 15.763 min Scan# 2172
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:169 Resp: 1074212
Ion Ratio Lower Upper
169 100
168 67.0 53.5 80.3
167 36.1 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 23.151 ng/ul
RT: 16.445 min Scan# 21
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:248 Resp: 368221
Ion Ratio Lower Upper
248 100
250 95.9 77.9 116.9
141 69.2 59.5 89.3



#69
Hexachlorobenzene
Concen: 21.622 ng/ul
RT: 16.557 min Scan# 2307
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:284 Resp: 399158
Ion Ratio Lower Upper
284 100
142 36.8 30.5 45.7
249 31.6 25.4 38.0





#70

Atrazine

Concen: 23.015 ng/ul

RT: 16.722 min Scan# 21

Delta R.T. -0.000 min

Lab File: BPCCC089.D

Acq: 12 Oct 2022 20:22

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:200 Resp: 342400

Ion Ratio Lower Upper

200 100

173 27.9 22.1 33.1

215 48.0 38.8 58.2



#71

Pentachlorophenol

Concen: 19.968 ng/ul

RT: 16.910 min Scan# 2367

Delta R.T. -0.000 min

Lab File: BPCCC089.D

Acq: 12 Oct 2022 20:22




Tgt Ion:266 Resp: 215979


Ion Ratio Lower Upper

266 100

264 64.0 50.6 75.8

268 64.7 51.8 77.6









#72
Phenanthrene
Concen: 20.190 ng/ul
RT: 17.304 min Scan# 2434
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:178 Resp: 1794736
Ion Ratio Lower Upper
178 100
179 15.6 12.2 18.4
176 19.3 15.8 23.6



#74
Anthracene
Concen: 20.451 ng/ul
RT: 17.392 min Scan# 2449
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:178 Resp: 1798974
Ion Ratio Lower Upper
178 100
179 15.3 12.2 18.4
176 18.6 14.8 22.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 22.360 ng/uL

RT: 13.298 min Scan# 1

Delta R.T. -0.000 min

Lab File: BPCCC089.D

Acq: 12 Oct 2022 20:22

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:216 Resp: 590600

Ion Ratio Lower Upper

216 100

214 78.7 62.6 94.0

218 48.6 38.9 58.3



#76

Pentachlorobenzene

Concen: 22.480 ng/uL

RT: 14.822 min Scan# 2012

Delta R.T. 0.006 min

Lab File: BPCCC089.D

Acq: 12 Oct 2022 20:22




Tgt Ion:250 Resp: 583144


Ion Ratio Lower Upper

250 100

248 62.9 51.4 77.2

252 63.9 50.7 76.1









#77
Carbazole
Concen: 19.567 ng/ul
RT: 17.669 min Scan# 2497
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:167 Resp: 1535150
Ion Ratio Lower Upper
167 100
166 21.3 17.0 25.6
139 12.5 10.1 15.1



#78
Di-n-butylphthalate
Concen: 23.851 ng/ul
RT: 18.216 min Scan# 2589
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:149 Resp: 1856952
Ion Ratio Lower Upper
149 100
150 9.0 7.2 10.8
104 4.3 3.4 5.2





#80

Fluoranthene

Concen: 21.674 ng/ul

RT: 19.269 min Scan# 21

Delta R.T. -0.000 min

Lab File: BPCCC089.D

Acq: 12 Oct 2022 20:22

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:202 Resp: 1663966

Ion Ratio Lower Upper

202 100

101 12.3 9.0 13.6

100 9.1 7.0 10.6



#82

Pyrene

Concen: 21.013 ng/ul

RT: 19.627 min Scan# 2829

Delta R.T. 0.006 min

Lab File: BPCCC089.D

Acq: 12 Oct 2022 20:22




Tgt Ion:202 Resp: 1707299


Ion Ratio Lower Upper

202 100

101 13.0 10.4 15.6

100 10.8 8.5 12.7





#83

Butylbenzylphthalate

Concen: 23.771 ng/ul

RT: 20.498 min Scan# 2977

Delta R.T. -0.000 min

Lab File: BPCCC089.D

Acq: 12 Oct 2022 20:22

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:149 Resp: 612248

Ion Ratio Lower Upper

149 100

91 63.7 48.5 72.7

206 21.3 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 17.105 ng/ul

RT: 21.268 min Scan# 3108

Delta R.T. -0.000 min

Lab File: BPCCC089.D

Acq: 12 Oct 2022 20:22




Tgt Ion:252 Resp: 460992


Ion Ratio Lower Upper

252 100

254 65.2 51.4 77.2

126 12.0 7.8 11.8#









#85
Benzo(a)anthracene
Concen: 20.815 ng/ul
RT: 21.333 min Scan# 3119
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:228 Resp: 1602395
Ion Ratio Lower Upper
228 100
229 19.6 15.7 23.5
226 27.1 21.9 32.9



#86
Bis(2-ethylhexyl)phthalate
Concen: 23.683 ng/ul
RT: 21.251 min Scan# 3105
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:149 Resp: 822311
Ion Ratio Lower Upper
149 100
167 26.6 21.9 32.9
279 3.9 3.4 5.2





#87

Chrysene

Concen: 20.301 ng/ul

RT: 21.386 min Scan# 3129

Delta R.T. -0.000 min

Lab File: BPCCC089.D

Acq: 12 Oct 2022 20:22

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:228 Resp: 1504340

Ion Ratio Lower Upper

228 100

226 30.8 24.0 36.0

229 19.9 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 19.417 ng/ul




RT: 22.180 min Scan# 3263


Delta R.T. -0.000 min

Lab File: BPCCC089.D

Acq: 12 Oct 2022 20:22

Tgt Ion:149 Resp: 1386036









#90
Benzo(b)fluoranthene
Concen: 19.897 ng/ul
RT: 23.027 min Scan# 3407
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:252 Resp: 1795756
Ion Ratio Lower Upper
252 100
253 22.3 17.6 26.4
125 11.0 7.8 11.8



#91
Benzo(k)fluoranthene
Concen: 19.650 ng/ul
RT: 23.074 min Scan# 3415
Delta R.T. -0.000 min
Lab File: BPCCC089.D
Acq: 12 Oct 2022 20:22

Tgt Ion:252 Resp: 1744435
Ion Ratio Lower Upper
252 100
253 22.3 17.6 26.4
125 11.3 8.2 12.2





#93

Benzo(a)pyrene

Concen: 20.348 ng/ul

RT: 23.662 min Scan# 3515

Delta R.T. 0.006 min

Lab File: BPCCC089.D

Acq: 12 Oct 2022 20:22

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:252 Resp: 1638387

Ion Ratio Lower Upper

252 100

253 21.7 17.6 26.4

125 11.7 8.9 13.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.281 ng/ul

RT: 26.286 min Scan# 3961

Delta R.T. 0.006 min

Lab File: BPCCC089.D

Acq: 12 Oct 2022 20:22




Tgt Ion:276 Resp: 2064795


Ion Ratio Lower Upper

276 100

138 27.1 19.1 28.7

277 24.9 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 20.211 ng/ul

RT: 26.304 min Scan# 3964

Delta R.T. 0.006 min

Lab File: BPCCC089.D

Acq: 12 Oct 2022 20:22

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:278 Resp: 1846020

Ion Ratio Lower Upper

278 100

139 17.6 12.5 18.7

279 23.8 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 19.947 ng/ul

RT: 27.056 min Scan# 4092

Delta R.T. -0.000 min

Lab File: BPCCC089.D

Acq: 12 Oct 2022 20:22

Tgt Ion:276 Resp: 1830996

Ion Ratio Lower Upper

276 100

138 24.3 16.5 24.7

277 23.1 19.0 28.4

