

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BP012057.D
 Acq On : 12 Oct 2022 00:52
 Operator : CG/JU
 Sample : N4991-09
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0BG1

Quant Time: Oct 12 01:25:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 11 23:54:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.857	152	234226	20.000	ng/u1	0.00
20) Naphthalene-d8	10.663	136	928064	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.504	164	515462	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.263	188	1072676	20.000	ng/u1	0.00
79) Chrysene-d12	21.345	240	1202538	20.000	ng/u1	0.00
88) Perylene-d12	23.762	264	1301730	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	33225	5.593	ng/uL	0.00
4) Pyridine-d5	3.728	84	28204	1.785	ng/u1	0.02
7) Phenol-d5	7.022	99	219707	11.663	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.187	67	318215	29.760	ng/u1	0.00
11) 2-Chlorophenol-d4	7.387	132	328053	21.939	ng/u1	0.00
15) 4-Methylphenol-d8	8.569	113	127716	8.827	ng/u1	0.00
21) Nitrobenzene-d5	9.022	128	206098	31.444	ng/u1	0.00
24) 2-Nitrophenol-d4	9.746	143	205535	30.926	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.287	165	295700	22.957	ng/u1	0.00
31) 4-Chloroaniline-d4	10.810	131	252470	13.301	ng/u1	0.00
46) Dimethylphthalate-d6	13.916	166	973895	29.833	ng/u1	0.00
49) Acenaphthylene-d8	14.198	160	1192488	29.540	ng/u1	0.00
54) 4-Nitrophenol-d4	14.728	143	15654	2.443	ng/u1	0.01
60) Fluorene-d10	15.498	176	921417	31.197	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.622	200	61884	10.926	ng/u1	0.00
73) Anthracene-d10	17.363	188	1335689	28.675	ng/u1	0.00
81) Pyrene-d10	19.598	212	1585646	25.738	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.610	264	968059	15.298	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
Data File : BP012057.D
Acq On : 12 Oct 2022 00:52
Operator : CG/JU
Sample : N4991-09
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0BG1

Quant Time: Oct 12 01:25:21 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 11 23:54:43 2022
Response via : Initial Calibration

