

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BP012097.D
 Acq On : 13 Oct 2022 00:59
 Operator : CG/JU
 Sample : PB148211BS
 Misc :
 ALS Vial : 67 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS211

Quant Time: Oct 13 01:44:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 23:03:21 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 10/13/2022
 Supervised By :mohammad ahmed 10/14/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.852	152	304076	20.000	ng/u1	0.00
20) Naphthalene-d8	10.658	136	1202216	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.499	164	634934	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.257	188	1134366	20.000	ng/u1	0.00
79) Chrysene-d12	21.351	240	1213671	20.000	ng/u1	0.00
88) Perylene-d12	23.763	264	1369434	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.276	96	41469	5.377	ng/uL	0.00
4) Pyridine-d5	3.699	84	542355	26.443	ng/u1	0.00
7) Phenol-d5	7.022	99	760001	31.077	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.187	67	413499	29.788	ng/u1	0.00
11) 2-Chlorophenol-d4	7.387	132	628716	32.388	ng/u1	0.00
15) 4-Methylphenol-d8	8.569	113	581105	30.936	ng/u1	0.00
21) Nitrobenzene-d5	9.022	128	299108	35.228	ng/u1	0.00
24) 2-Nitrophenol-d4	9.740	143	326926	37.973	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.281	165	564849	33.853	ng/u1	0.00
31) 4-Chloroaniline-d4	10.805	131	707309	28.766	ng/u1	0.00
46) Dimethylphthalate-d6	13.916	166	1286400	31.991	ng/u1	0.00
49) Acenaphthylene-d8	14.193	160	1656819	33.320	ng/u1	0.00
54) 4-Nitrophenol-d4	14.716	143	241054	30.535	ng/u1	0.00
60) Fluorene-d10	15.498	176	1134842	31.193	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.622	200	193796	32.354	ng/u1	0.00
73) Anthracene-d10	17.357	188	1584547	32.168	ng/u1	0.00
81) Pyrene-d10	19.592	212	1831670	29.459	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.610	264	2193361	32.947	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.311	88	85595	10.547	ng/uL	95
5) Pyridine	3.717	79	548837	27.928	ng/u1	97
6) Benzaldehyde	6.993	77	451467m	39.347	ng/u1	
8) Phenol	7.052	94	749324	29.186	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.281	93	587854	29.021	ng/u1	99
12) 2-Chlorophenol	7.417	128	638108	30.289	ng/u1	98
13) 2-Methylphenol	8.305	108	562438	29.372	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.387	45	556992	26.827	ng/u1	98
16) Acetophenone	8.681	105	848494	29.299	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.669	70	380110	28.818	ng/u1	98
18) 4-Methylphenol	8.634	108	602758	29.204	ng/u1	96
19) Hexachloroethane	8.928	117	243004	29.881	ng/u1	99
22) Nitrobenzene	9.063	77	608602	31.269	ng/u1	95
23) Isophorone	9.593	82	1066184	30.882	ng/u1	99
25) 2-Nitrophenol	9.775	139	345343	34.142	ng/u1	93
26) 2,4-Dimethylphenol	9.840	107	569146	28.892	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.075	93	719839	30.384	ng/u1	99
29) 2,4-Dichlorophenol	10.310	162	552450	31.506	ng/u1	98
30) Naphthalene	10.710	128	1899383	29.898	ng/u1	99
32) 4-Chloroaniline	10.828	127	689463	27.198	ng/u1	99
33) Hexachlorobutadiene	10.987	225	340534	31.319	ng/u1	99
34) Caprolactam	11.628	113	144806	30.053	ng/u1	95
35) 4-Chloro-3-methylphenol	11.957	107	513868	30.934	ng/u1	99
36) 2-Methylnaphthalene	12.322	142	1209131	29.557	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BP012097.D
 Acq On : 13 Oct 2022 00:59
 Operator : CG/JU
 Sample : PB148211BS
 Misc :
 ALS Vial : 67 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS211


Quant Time: Oct 13 01:44:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 23:03:21 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 10/13/2022
 Supervised By :mohammad ahmed 10/14/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.540	142	1198175	29.247	ng/u1	99
39) 1,2,4,5-Tetrachloroben...	12.687	216	619764	31.932	ng/u1	99
40) Hexachlorocyclopentadiene	12.663	237	266767	28.297	ng/u1	98
41) 2,4,6-Trichlorophenol	12.934	196	393558	33.662	ng/u1	99
42) 2,4,5-Trichlorophenol	13.004	196	420799	33.117	ng/u1	99
43) 1,1'-Biphenyl	13.334	154	1501556	31.326	ng/u1	99
44) 2-Chloronaphthalene	13.375	162	1195657	31.237	ng/u1	99
45) 2-Nitroaniline	13.587	65	273246	33.464	ng/u1	95
47) Dimethylphthalate	13.957	163	1297499	30.626	ng/u1	99
48) 2,6-Dinitrotoluene	14.087	165	265788	35.435	ng/u1	94
50) Acenaphthylene	14.222	152	1767023	31.406	ng/u1	99
51) 3-Nitroaniline	14.416	138	283988	32.515	ng/u1	97
52) Acenaphthene	14.563	153	1207185	30.140	ng/u1	99
53) 2,4-Dinitrophenol	14.622	184	132175	27.277	ng/u1	99
55) 4-Nitrophenol	14.728	109	157317	28.824	ng/u1	97
56) Dibenzofuran	14.898	168	1613240	29.802	ng/u1	99
57) 2,4-Dinitrotoluene	14.875	165	363364	33.707	ng/u1#	92
58) 2,3,4,6-Tetrachlorophenol	15.128	232	316563	31.964	ng/u1	97
59) Diethylphthalate	15.328	149	1187157	30.597	ng/u1	99
61) Fluorene	15.551	166	1266586	29.808	ng/u1	98
62) 4-Chlorophenyl-phenyle...	15.546	204	620020	30.163	ng/u1	97
63) 4-Nitroaniline	15.581	138	293242	36.770	ng/u1	96
66) 4,6-Dinitro-2-methylph...	15.640	198	200199	30.917	ng/u1#	90
67) N-Nitrosodiphenylamine	15.763	169	1044970	31.122	ng/u1	99
68) 4-Bromophenyl-phenylether	16.445	248	356414	32.289	ng/u1	97
69) Hexachlorobenzene	16.557	284	399750	31.201	ng/u1	95
70) Atrazine	16.722	200	339391	32.871	ng/u1	99
71) Pentachlorophenol	16.910	266	241101	32.120	ng/u1	99
72) Phenanthrene	17.304	178	1875300	30.399	ng/u1	99
74) Anthracene	17.392	178	1849357	30.293	ng/u1	100
75) 1,2,3,4-Tetrachloroben...	13.299	216	632948	34.530	ng/uL	98
76) Pentachlorobenzene	14.816	250	534881	29.712	ng/uL	98
77) Carbazole	17.669	167	1716964	31.534	ng/u1	100
78) Di-n-butylphthalate	18.216	149	1798629	33.288	ng/u1	100
80) Fluoranthene	19.269	202	2123154	27.867	ng/u1	99
82) Pyrene	19.622	202	2266694	28.113	ng/u1	97
83) Butylbenzylphthalate	20.492	149	862007	33.726	ng/u1	94
84) 3,3'-Dichlorobenzidine	21.269	252	840560	31.429	ng/u1	99
85) Benzo(a)anthracene	21.333	228	2384245	31.210	ng/u1	99
86) Bis(2-ethylhexyl)phtha...	21.251	149	1193051	34.625	ng/u1	100
87) Chrysene	21.386	228	2252958	30.638	ng/u1	100
89) Di-n-octyl phthalate	22.180	149	2200001	32.089	ng/u1	100
90) Benzo(b)fluoranthene	23.027	252	2664631	30.740	ng/u1	98
91) Benzo(k)fluoranthene	23.075	252	2567343	30.111	ng/u1	99
93) Benzo(a)pyrene	23.657	252	2398878	31.019	ng/u1	99
94) Indeno(1,2,3-cd)pyrene	26.286	276	3182693	32.548	ng/u1	97
95) Dibenzo(a,h)anthracene	26.304	278	2714247	30.941	ng/u1	98
96) Benzo(g,h,i)perylene	27.057	276	2704130	30.671	ng/u1	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 10.547 ng/uL
RT: 3.311 min Scan# 51
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211



Tgt Ion: 88 Resp: 85595
Ion Ratio Lower Upper
88 100
43 20.0 16.6 25.0
58 61.4 52.6 79.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#5
Pyridine
Concen: 27.928 ng/uL
RT: 3.717 min Scan# 124
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59

Tgt Ion: 79 Resp: 548837
Ion Ratio Lower Upper
79 100
52 61.2 51.2 76.8
51 27.0 22.8 34.2









#6
Benzaldehyde
Concen: 39.347 ng/ul m
RT: 6.993 min Scan# 681
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS211

Tgt Ion: 77 Resp: 451462
Ion Ratio Lower Upper
77 100
105 105.2 85.2 127.8
106 102.6 83.1 124.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#8
Phenol
Concen: 29.186 ng/ul
RT: 7.052 min Scan# 691
Delta R.T. 0.006 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59


Tgt Ion: 94 Resp: 749324
Ion Ratio Lower Upper
94 100
65 24.5 19.4 29.0
66 32.4 24.6 37.0





#10
Bis(2-Chloroethyl)ether
Concen: 29.021 ng/ul
RT: 7.281 min Scan# 730
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211




Tgt Ion: 93 Resp: 587854
Ion Ratio Lower Upper
93 100
63 70.1 56.7 85.1
95 33.3 26.2 39.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022




#12
2-Chlorophenol
Concen: 30.289 ng/ul
RT: 7.417 min Scan# 753
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59



Tgt Ion:128 Resp: 638108
Ion Ratio Lower Upper
128 100
64 39.3 32.7 49.1
130 32.3 26.3 39.5





#13

2-Methylphenol

Concen: 29.372 ng/ul

RT: 8.305 min Scan# 904

Delta R.T. 0.006 min

Lab File: BP012097.D

Acq: 13 Oct 2022 00:59

Instrument :

BNA_P

ClientSampleId :

SLCS211

Tgt Ion:108 Resp: 562438

Ion Ratio Lower Upper





108 100

107 87.1 71.9 107.9

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 26.827 ng/ul

RT: 8.387 min Scan# 918

Delta R.T. 0.006 min

Lab File: BP012097.D

Acq: 13 Oct 2022 00:59




Tgt Ion: 45 Resp: 556992


Ion Ratio Lower Upper

45 100

77 19.1 14.4 21.6

79 14.8 11.3 16.9









#16
Acetophenone
Concen: 29.299 ng/ul
RT: 8.681 min Scan# 969
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS211

Tgt Ion:105 Resp: 848494
Ion Ratio Lower Upper
105 100
77 70.2 57.0 85.4
51 19.6 16.4 24.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#17
N-Nitroso-di-n-propylamine
Concen: 28.818 ng/ul
RT: 8.669 min Scan# 966
Delta R.T. 0.006 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59

Tgt Ion: 70 Resp: 380110
Ion Ratio Lower Upper
70 100
42 42.2 34.6 52.0
101 10.3 8.0 12.0
130 23.9 18.3 27.5









#18
4-Methylphenol
Concen: 29.204 ng/ul
RT: 8.634 min Scan# 90
Delta R.T. 0.006 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS211

Tgt Ion:108 Resp: 602758
Ion Ratio Lower Upper
108 100
107 112.2 86.3 129.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#19
Hexachloroethane
Concen: 29.881 ng/ul
RT: 8.928 min Scan# 1010
Delta R.T. -0.006 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59

Tgt Ion:117 Resp: 243004
Ion Ratio Lower Upper
117 100
201 104.0 82.3 123.5
199 63.0 50.0 75.0









#22
Nitrobenzene
Concen: 31.269 ng/ul
RT: 9.063 min Scan# 1033
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS211

Tgt Ion: 77 Resp: 608602
Ion Ratio Lower Upper
77 100
123 53.1 39.4 59.0
65 12.7 9.9 14.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#23
Isophorone
Concen: 30.882 ng/ul
RT: 9.593 min Scan# 1123
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59

Tgt Ion: 82 Resp: 1066184
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.0
138 19.7 15.2 22.8









#25
2-Nitrophenol
Concen: 34.142 ng/ul
RT: 9.775 min Scan# 1154
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS211

Tgt Ion:139 Resp: 345343
Ion Ratio Lower Upper
139 100
65 39.7 36.3 54.5
109 24.7 21.2 31.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#26
2,4-Dimethylphenol
Concen: 28.892 ng/ul
RT: 9.840 min Scan# 1165
Delta R.T. 0.006 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59

Tgt Ion:107 Resp: 569146
Ion Ratio Lower Upper
107 100
121 53.5 43.0 64.4
122 91.9 71.4 107.2









#27
Bis(2-Chloroethoxy)methane
Concen: 30.384 ng/ul
RT: 10.075 min Scan# 1205
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS211

Tgt Ion: 93 Resp: 719839
Ion Ratio Lower Upper
93 100
95 32.5 25.7 38.5
123 11.8 9.2 13.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#29
2,4-Dichlorophenol
Concen: 31.506 ng/ul
RT: 10.310 min Scan# 1245
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59


Tgt Ion:162 Resp: 552450
Ion Ratio Lower Upper
162 100
164 65.9 51.5 77.3
98 33.4 26.3 39.5





#30
Naphthalene
Concen: 29.898 ng/ul
RT: 10.710 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211



Tgt Ion:128 Resp: 1899383
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 13.2 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#32
4-Chloroaniline
Concen: 27.198 ng/ul
RT: 10.828 min Scan# 1333
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59


Tgt Ion:127 Resp: 689463
Ion Ratio Lower Upper
127 100
129 31.7 25.6 38.4





#33
Hexachlorobutadiene
Concen: 31.319 ng/ul
RT: 10.987 min Scan# 1360
Delta R.T. -0.006 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211




Tgt Ion:225 Resp: 340534
Ion Ratio Lower Upper
225 100
223 62.9 50.9 76.3
227 63.2 51.1 76.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022




#34
Caprolactam
Concen: 30.053 ng/ul
RT: 11.628 min Scan# 1469
Delta R.T. 0.012 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Tgt Ion:113 Resp: 144806
Ion Ratio Lower Upper
113 100
55 138.0 115.2 172.8
56 100.5 85.6 128.4





#35
4-Chloro-3-methylphenol
Concen: 30.934 ng/ul
RT: 11.957 min Scan# 1525
Delta R.T. 0.006 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211



Tgt Ion:107 Resp: 513868
Ion Ratio Lower Upper
107 100
144 28.2 23.0 34.4
142 86.4 69.8 104.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#36
2-Methylnaphthalene
Concen: 29.557 ng/ul
RT: 12.322 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59


Tgt Ion:142 Resp: 1209131
Ion Ratio Lower Upper
142 100
141 89.2 70.7 106.1





#37
1-Methylnaphthalene
Concen: 29.247 ng/ul
RT: 12.540 min Scan# 1624
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211




Tgt Ion:142 Resp: 1198175
Ion Ratio Lower Upper
142 100
141 91.8 72.9 109.3
116 3.6 2.8 4.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 31.932 ng/ul
RT: 12.687 min Scan# 1649
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Tgt Ion:216 Resp: 619764
Ion Ratio Lower Upper
216 100
214 78.2 63.0 94.4
179 20.5 16.3 24.5
108 15.8 12.5 18.7





#40
Hexachlorocyclopentadiene
Concen: 28.297 ng/ul
RT: 12.663 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211




Tgt Ion:237 Resp: 266762
Ion Ratio Lower Upper
237 100
235 63.8 49.8 74.8
272 12.8 9.9 14.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022




#41
2,4,6-Trichlorophenol
Concen: 33.662 ng/ul
RT: 12.934 min Scan# 1691
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59



Tgt Ion:196 Resp: 393558
Ion Ratio Lower Upper
196 100
198 94.7 76.2 114.2
200 30.5 24.2 36.2









#42
2,4,5-Trichlorophenol
Concen: 33.117 ng/ul
RT: 13.004 min Scan# 1703
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS211

Tgt Ion:196 Resp: 420799
Ion Ratio Lower Upper
196 100
198 96.1 75.4 113.2
200 31.3 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#43
1,1'-Biphenyl
Concen: 31.326 ng/ul
RT: 13.334 min Scan# 1759
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59


Tgt Ion:154 Resp: 1501556
Ion Ratio Lower Upper
154 100
153 40.1 32.2 48.2
76 11.7 10.3 15.5





#44
2-Chloronaphthalene
Concen: 31.237 ng/ul
RT: 13.375 min Scan# 1766
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211



Tgt Ion:162 Resp: 119565
Ion Ratio Lower Upper
162 100
164 33.2 26.0 39.0
127 37.7 29.4 44.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#45
2-Nitroaniline
Concen: 33.464 ng/ul
RT: 13.587 min Scan# 1802
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59


Tgt Ion: 65 Resp: 273246
Ion Ratio Lower Upper
65 100
92 77.3 61.6 92.4
138 129.6 96.5 144.7





#47
Dimethylphthalate
Concen: 30.626 ng/ul
RT: 13.957 min Scan# 1865
Delta R.T. -0.006 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211




Tgt Ion:163 Resp: 1297499
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.5 7.9 11.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022




#48
2,6-Dinitrotoluene
Concen: 35.435 ng/ul
RT: 14.087 min Scan# 1887
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Tgt Ion:165 Resp: 265788
Ion Ratio Lower Upper
165 100
89 48.9 43.4 65.0
121 20.8 18.1 27.1





#50
Acenaphthylene
Concen: 31.406 ng/ul
RT: 14.222 min Scan# 1910
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211



Tgt Ion:152 Resp: 1767023
Ion Ratio Lower Upper
152 100
151 20.1 15.7 23.5
153 13.1 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#51
3-Nitroaniline
Concen: 32.515 ng/ul
RT: 14.416 min Scan# 1943
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59


Tgt Ion:138 Resp: 283988
Ion Ratio Lower Upper
138 100
108 9.9 8.5 12.7
92 104.6 85.8 128.8





#52
Acenaphthene
Concen: 30.140 ng/ul
RT: 14.563 min Scan# 1969
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211



Tgt Ion:153 Resp: 1207185
Ion Ratio Lower Upper
153 100
152 47.0 38.2 57.2
154 88.5 70.6 105.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#53
2,4-Dinitrophenol
Concen: 27.277 ng/ul
RT: 14.622 min Scan# 1978
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59

Tgt Ion:184 Resp: 132175
Ion Ratio Lower Upper
184 100
63 56.1 45.5 68.3
154 57.6 47.0 70.6









#55
4-Nitrophenol
Concen: 28.824 ng/ul
RT: 14.728 min Scan# 1996
Delta R.T. 0.006 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS211

Tgt Ion:109 Resp: 157317
Ion Ratio Lower Upper
109 100
139 160.0 124.1 186.1
65 132.3 104.2 156.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#56
Dibenzofuran
Concen: 29.802 ng/ul
RT: 14.898 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59


Tgt Ion:168 Resp: 1613240
Ion Ratio Lower Upper
168 100
139 36.0 28.2 42.4





#57
2,4-Dinitrotoluene
Concen: 33.707 ng/ul
RT: 14.875 min Scan# 2021
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211




Tgt Ion:165 Resp: 363364
Ion Ratio Lower Upper
165 100
63 33.7 31.1 46.7
182 2.7 1.8 2.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 31.964 ng/ul
RT: 15.128 min Scan# 2064
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Tgt Ion:232 Resp: 316563
Ion Ratio Lower Upper
232 100
131 44.9 37.8 56.6
130 2.3 1.8 2.6
166 31.5 26.6 39.8





#59
Diethylphthalate
Concen: 30.597 ng/ul
RT: 15.328 min Scan# 2098
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211




Tgt Ion:149 Resp: 118715
Ion Ratio Lower Upper
149 100
177 22.6 17.5 26.3
150 12.3 9.8 14.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022




#61
Fluorene
Concen: 29.808 ng/ul
RT: 15.551 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Tgt Ion:166 Resp: 1266586
Ion Ratio Lower Upper
166 100
165 97.6 79.5 119.3
167 13.6 10.9 16.3





#62
4-Chlorophenyl-phenylether
Concen: 30.163 ng/ul
RT: 15.546 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211




Tgt Ion:204 Resp: 620020
Ion Ratio Lower Upper
204 100
206 33.1 25.4 38.2
141 55.7 46.2 69.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022




#63
4-Nitroaniline
Concen: 36.770 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59



Tgt Ion:138 Resp: 293242
Ion Ratio Lower Upper
138 100
92 47.1 39.2 58.8
108 65.2 55.4 83.0









#66
4,6-Dinitro-2-methylphenol
Concen: 30.917 ng/ul
RT: 15.640 min Scan# 2151
Delta R.T. 0.006 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS211

Tgt Ion:198 Resp: 200199
Ion Ratio Lower Upper
198 100
182 4.4 4.9 7.3
77 19.5 20.0 30.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#67
N-Nitrosodiphenylamine
Concen: 31.122 ng/ul
RT: 15.763 min Scan# 2172
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59


Tgt Ion:169 Resp: 1044970
Ion Ratio Lower Upper
169 100
168 67.4 53.5 80.3
167 36.6 29.5 44.3





#68
4-Bromophenyl-phenylether
Concen: 32.289 ng/ul
RT: 16.445 min Scan# 2288
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211




Tgt Ion:248 Resp: 356414
Ion Ratio Lower Upper
248 100
250 98.1 77.9 116.9
141 69.7 59.5 89.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022




#69
Hexachlorobenzene
Concen: 31.201 ng/ul
RT: 16.557 min Scan# 2307
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59



Tgt Ion:284 Resp: 399750
Ion Ratio Lower Upper
284 100
142 34.1 30.5 45.7
249 30.6 25.4 38.0





#70

Atrazine

Concen: 32.871 ng/ul

RT: 16.722 min Scan# 2335

Delta R.T. 0.000 min

Lab File: BP012097.D

Acq: 13 Oct 2022 00:59

Instrument :

BNA_P

ClientSampleId :





SLCS211

Tgt Ion:200 Resp: 339391
 Ion Ratio Lower Upper
 200 100
 173 26.8 22.1 33.1
 215 47.8 38.8 58.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/13/2022
 Supervised By :mohammad ahmed 10/14/2022



#71

Pentachlorophenol

Concen: 32.120 ng/ul




RT: 16.910 min Scan# 2367


Delta R.T. 0.000 min

Lab File: BP012097.D

Acq: 13 Oct 2022 00:59


Tgt Ion:266 Resp: 241101
 Ion Ratio Lower Upper
 266 100
 264 62.5 50.6 75.8
 268 63.9 51.8 77.6





#72
Phenanthrene
Concen: 30.399 ng/u1
RT: 17.304 min Scan# 2434
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211



Tgt Ion:178 Resp: 1875300
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.4
176 19.1 15.8 23.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#74
Anthracene
Concen: 30.293 ng/u1
RT: 17.392 min Scan# 2449
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59


Tgt Ion:178 Resp: 1849357
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.4
176 18.7 14.8 22.2





#75
1,2,3,4-Tetrachlorobenzene
Concen: 34.530 ng/uL
RT: 13.299 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211



Tgt Ion:216 Resp: 632948
Ion Ratio Lower Upper
216 100
214 76.8 62.6 94.0
218 47.1 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#76
Pentachlorobenzene
Concen: 29.712 ng/uL
RT: 14.816 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59


Tgt Ion:250 Resp: 534881
Ion Ratio Lower Upper
250 100
248 61.9 51.4 77.2
252 63.2 50.7 76.1





#77
Carbazole
Concen: 31.534 ng/ul
RT: 17.669 min Scan# 2496
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211




Tgt Ion:167 Resp: 1716964
Ion Ratio Lower Upper
167 100
166 21.5 17.0 25.6
139 12.6 10.1 15.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022




#78
Di-n-butylphthalate
Concen: 33.288 ng/ul
RT: 18.216 min Scan# 2589
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Tgt Ion:149 Resp: 1798629
Ion Ratio Lower Upper
149 100
150 9.1 7.2 10.8
104 4.3 3.4 5.2





#80
Fluoranthene
Concen: 27.867 ng/ul
RT: 19.269 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211




Tgt Ion:202 Resp: 2123154
Ion Ratio Lower Upper
202 100
101 11.9 9.0 13.6
100 8.9 7.0 10.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022




#82
Pyrene
Concen: 28.113 ng/ul
RT: 19.622 min Scan# 2828
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59



Tgt Ion:202 Resp: 2266694
Ion Ratio Lower Upper
202 100
101 14.3 10.4 15.6
100 11.5 8.5 12.7





#83

Butylbenzylphthalate

Concen: 33.726 ng/ul

RT: 20.492 min Scan# 2977

Delta R.T. -0.006 min

Lab File: BP012097.D

Acq: 13 Oct 2022 00:59

Instrument :

BNA_P

ClientSampleId :





SLCS211

Tgt Ion:149	Resp: 86200
Ion Ratio	Lower Upper
149 100	
91 65.7	48.5 72.7
206 20.3	17.9 26.9

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#84

3,3'-Dichlorobenzidine

Concen: 31.429 ng/ul




RT: 21.269 min Scan# 3108


Delta R.T. 0.000 min

Lab File: BP012097.D

Acq: 13 Oct 2022 00:59


Tgt Ion:252	Resp: 840560
Ion Ratio	Lower Upper
252 100	
254 65.0	51.4 77.2
126 11.2	7.8 11.8





#85
Benzo(a)anthracene
Concen: 31.210 ng/ul
RT: 21.333 min Scan# 3119
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211




Tgt Ion:228 Resp: 2384245
Ion Ratio Lower Upper
228 100
229 19.9 15.7 23.5
226 27.8 21.9 32.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022




#86
Bis(2-ethylhexyl)phthalate
Concen: 34.625 ng/ul
RT: 21.251 min Scan# 3105
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59



Tgt Ion:149 Resp: 1193051
Ion Ratio Lower Upper
149 100
167 27.2 21.9 32.9
279 4.1 3.4 5.2





#87

Chrysene

Concen: 30.638 ng/ul

RT: 21.386 min Scan# 3128

Delta R.T. 0.000 min

Lab File: BP012097.D

Acq: 13 Oct 2022 00:59

Instrument :

BNA_P

ClientSampleId :

SLCS211

Tgt Ion:228 Resp: 2252958

Ion Ratio Lower Upper

228 100





226 30.2 24.0 36.0

229 19.5 15.5 23.3

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#89

Di-n-octyl phthalate

Concen: 32.089 ng/ul




RT: 22.180 min Scan# 3263


Delta R.T. 0.000 min

Lab File: BP012097.D

Acq: 13 Oct 2022 00:59


Tgt Ion:149 Resp: 2200001





#90
Benzo(b)fluoranthene
Concen: 30.740 ng/ul
RT: 23.027 min Scan# 3407
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211




Tgt Ion:252 Resp: 2664631
Ion Ratio Lower Upper
252 100
253 22.4 17.6 26.4
125 10.9 7.8 11.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022




#91
Benzo(k)fluoranthene
Concen: 30.111 ng/ul
RT: 23.075 min Scan# 3415
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59



Tgt Ion:252 Resp: 2567343
Ion Ratio Lower Upper
252 100
253 22.1 17.6 26.4
125 11.0 8.2 12.2





#93

Benzo(a)pyrene

Concen: 31.019 ng/ul

RT: 23.657 min Scan# 3514

Delta R.T. 0.000 min

Lab File: BP012097.D

Acq: 13 Oct 2022 00:59

Instrument :

BNA_P

ClientSampleId :

SLCS211

Tgt Ion:252 Resp: 2398878

Ion Ratio Lower Upper

252 100





253 22.1 17.6 26.4

125 12.0 8.9 13.3

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#94

Indeno(1,2,3-cd)pyrene

Concen: 32.548 ng/ul

RT: 26.286 min Scan# 3961

Delta R.T. 0.006 min

Lab File: BP012097.D

Acq: 13 Oct 2022 00:59




Tgt Ion:276 Resp: 3182693


Ion Ratio Lower Upper

276 100

138 26.7 19.1 28.7


277 25.4 19.9 29.9





#95
Dibenzo(a,h)anthracene
Concen: 30.941 ng/ul
RT: 26.304 min Scan# 3964
Delta R.T. 0.006 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS211




Tgt Ion:278 Resp: 271424
Ion Ratio Lower Upper
278 100
139 17.5 12.5 18.7
279 24.1 19.0 28.4

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/13/2022
Supervised By :mohammad ahmed 10/14/2022



#96
Benzo(g,h,i)perylene
Concen: 30.671 ng/ul
RT: 27.057 min Scan# 4092
Delta R.T. 0.000 min
Lab File: BP012097.D
Acq: 13 Oct 2022 00:59



Tgt Ion:276 Resp: 2704130
Ion Ratio Lower Upper
276 100
138 24.1 16.5 24.7
277 24.0 19.0 28.4

