

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BP012114.D
 Acq On : 13 Oct 2022 11:16
 Operator : CG/JU
 Sample : PB148251BS
 Misc :
 ALS Vial : 85 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS251

Quant Time: Oct 13 23:31:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 23:03:21 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2022
 Supervised By :mohammad ahmed 10/14/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.852	152	239782	20.000	ng/u1	0.00
20) Naphthalene-d8	10.657	136	976451	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.498	164	571594	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.257	188	1240425	20.000	ng/u1	0.00
79) Chrysene-d12	21.345	240	1366162	20.000	ng/u1	0.00
88) Perylene-d12	23.768	264	1440296	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	37560	6.176	ng/uL	0.00
4) Pyridine-d5	3.699	84	504609	31.199	ng/u1	0.00
7) Phenol-d5	7.016	99	655556	33.994	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.181	67	358399	32.742	ng/u1	0.00
11) 2-Chlorophenol-d4	7.381	132	538691	35.191	ng/u1	0.00
15) 4-Methylphenol-d8	8.569	113	496290	33.505	ng/u1	0.00
21) Nitrobenzene-d5	9.016	128	251097	36.411	ng/u1	0.00
24) 2-Nitrophenol-d4	9.740	143	267956	38.320	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.281	165	487067	35.941	ng/u1	0.00
31) 4-Chloroaniline-d4	10.799	131	630708	31.582	ng/u1	0.00
46) Dimethylphthalate-d6	13.910	166	1262067	34.864	ng/u1	0.00
49) Acenaphthylene-d8	14.193	160	1570660	35.088	ng/u1	0.00
54) 4-Nitrophenol-d4	14.710	143	295172	41.533	ng/u1	0.00
60) Fluorene-d10	15.492	176	1161172	35.453	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.622	200	232776	35.539	ng/u1	0.00
73) Anthracene-d10	17.357	188	1903881	35.346	ng/u1	0.00
81) Pyrene-d10	19.598	212	2339937	33.433	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.610	264	2489579	35.557	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.311	88	78657	12.291	ng/uL	99
5) Pyridine	3.717	79	490718	31.666	ng/u1	96
6) Benzaldehyde	6.993	77	383107m	42.342	ng/u1	
8) Phenol	7.046	94	641655	31.694	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.275	93	496570	31.088	ng/u1	99
12) 2-Chlorophenol	7.410	128	540136	32.513	ng/u1	99
13) 2-Methylphenol	8.299	108	476400	31.550	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.381	45	478193	29.207	ng/u1	99
16) Acetophenone	8.681	105	722487	31.637	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.669	70	320377	30.802	ng/u1	97
18) 4-Methylphenol	8.628	108	516911	31.760	ng/u1	96
19) Hexachloroethane	8.928	117	206496	32.200	ng/u1	98
22) Nitrobenzene	9.057	77	520772	32.943	ng/u1	99
23) Isophorone	9.593	82	906985	32.345	ng/u1	99
25) 2-Nitrophenol	9.769	139	286800	34.910	ng/u1	97
26) 2,4-Dimethylphenol	9.834	107	508686	31.793	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.069	93	609912	31.696	ng/u1	99
29) 2,4-Dichlorophenol	10.304	162	473552	33.251	ng/u1	99
30) Naphthalene	10.704	128	1631425	31.617	ng/u1	100
32) 4-Chloroaniline	10.822	127	606413	29.452	ng/u1	100
33) Hexachlorobutadiene	10.987	225	276717	31.334	ng/u1	99
34) Caprolactam	11.622	113	142882m	36.510	ng/u1	
35) 4-Chloro-3-methylphenol	11.951	107	473084	35.064	ng/u1	98
36) 2-Methylnaphthalene	12.316	142	1050352	31.612	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BP012114.D
 Acq On : 13 Oct 2022 11:16
 Operator : CG/JU
 Sample : PB148251BS
 Misc :
 ALS Vial : 85 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS251


Quant Time: Oct 13 23:31:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 23:03:21 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2022
 Supervised By :mohammad ahmed 10/14/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.540	142	1053421	31.659	ng/u1	99
39) 1,2,4,5-Tetrachloroben...	12.687	216	539909	30.900	ng/u1	98
40) Hexachlorocyclopentadiene	12.663	237	225588	26.581	ng/u1	99
41) 2,4,6-Trichlorophenol	12.928	196	355504	33.777	ng/u1	98
42) 2,4,5-Trichlorophenol	13.004	196	395723	34.595	ng/u1	98
43) 1,1'-Biphenyl	13.334	154	1347713	31.232	ng/u1	99
44) 2-Chloronaphthalene	13.375	162	1085621	31.505	ng/u1	99
45) 2-Nitroaniline	13.587	65	272388	37.056	ng/u1	94
47) Dimethylphthalate	13.957	163	1263982	33.141	ng/u1	99
48) 2,6-Dinitrotoluene	14.081	165	259485	38.428	ng/u1	98
50) Acenaphthylene	14.222	152	1673007	33.030	ng/u1	99
51) 3-Nitroaniline	14.410	138	289569	36.828	ng/u1	99
52) Acenaphthene	14.563	153	1149736	31.886	ng/u1	100
53) 2,4-Dinitrophenol	14.622	184	150843	34.579	ng/u1	99
55) 4-Nitrophenol	14.728	109	188814	38.429	ng/u1	95
56) Dibenzofuran	14.898	168	1578540	32.393	ng/u1	98
57) 2,4-Dinitrotoluene	14.875	165	397389	40.948	ng/u1	92
58) 2,3,4,6-Tetrachlorophenol	15.128	232	337065	37.805	ng/u1	98
59) Diethylphthalate	15.328	149	1224929	35.069	ng/u1	100
61) Fluorene	15.551	166	1294046	33.829	ng/u1	98
62) 4-Chlorophenyl-phenyle...	15.545	204	610764	33.005	ng/u1	97
63) 4-Nitroaniline	15.581	138	341314	47.540	ng/u1	96
66) 4,6-Dinitro-2-methylph...	15.639	198	234491	33.116	ng/u1#	92
67) N-Nitrosodiphenylamine	15.763	169	1098070	29.908	ng/u1	99
68) 4-Bromophenyl-phenylether	16.445	248	372035	30.822	ng/u1	97
69) Hexachlorobenzene	16.557	284	437770	31.247	ng/u1	97
70) Atrazine	16.722	200	386431	34.227	ng/u1	99
71) Pentachlorophenol	16.904	266	287882	35.073	ng/u1	98
72) Phenanthrene	17.298	178	2177501	32.279	ng/u1	99
74) Anthracene	17.392	178	2203115	33.002	ng/u1	99
75) 1,2,3,4-Tetrachloroben...	13.293	216	556695	27.773	ng/uL	99
76) Pentachlorobenzene	14.816	250	515769	26.201	ng/uL	99
77) Carbazole	17.669	167	2104861	35.352	ng/u1	99
78) Di-n-butylphthalate	18.216	149	2129359	36.039	ng/u1	100
80) Fluoranthene	19.269	202	2662828	31.050	ng/u1	98
82) Pyrene	19.622	202	2827839	31.158	ng/u1	97
83) Butylbenzylphthalate	20.492	149	1072399	37.274	ng/u1	92
84) 3,3'-Dichlorobenzidine	21.269	252	963697	32.012	ng/u1	99
85) Benzo(a)anthracene	21.333	228	2857316	33.227	ng/u1	98
86) Bis(2-ethylhexyl)phtha...	21.251	149	1485307	38.296	ng/u1	99
87) Chrysene	21.386	228	2683725	32.422	ng/u1	99
89) Di-n-octyl phthalate	22.180	149	2644507	36.674	ng/u1	100
90) Benzo(b)fluoranthene	23.027	252	3004070	32.951	ng/u1	97
91) Benzo(k)fluoranthene	23.074	252	2951952	32.918	ng/u1	99
93) Benzo(a)pyrene	23.663	252	2695648	33.142	ng/u1	99
94) Indeno(1,2,3-cd)pyrene	26.292	276	3494564	33.979	ng/u1	95
95) Dibenzo(a,h)anthracene	26.304	278	2973996	32.234	ng/u1	97
96) Benzo(g,h,i)perylene	27.057	276	2923275	31.526	ng/u1	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 12.291 ng/uL
RT: 3.311 min Scan# 55
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
Client Sample Id :
SLCS251




Tgt Ion: 88 Resp: 78652
Ion Ratio Lower Upper
88 100
43 20.8 16.6 25.0
58 66.7 52.6 79.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022




#5
Pyridine
Concen: 31.666 ng/uL
RT: 3.717 min Scan# 124
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Tgt Ion: 79 Resp: 490718
Ion Ratio Lower Upper
79 100
52 60.6 51.2 76.8
51 27.0 22.8 34.2





#6
Benzaldehyde
Concen: 42.342 ng/ul m
RT: 6.993 min Scan# 681
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251



Tgt Ion: 77 Resp: 38310
Ion Ratio Lower Upper
77 100
105 106.3 85.2 127.8
106 103.1 83.1 124.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#8
Phenol
Concen: 31.694 ng/ul
RT: 7.046 min Scan# 690
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16


Tgt Ion: 94 Resp: 641655
Ion Ratio Lower Upper
94 100
65 25.2 19.4 29.0
66 31.9 24.6 37.0





#10
Bis(2-Chloroethyl)ether
Concen: 31.088 ng/ul
RT: 7.275 min Scan# 729
Delta R.T. -0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251




Tgt Ion: 93 Resp: 496576
Ion Ratio Lower Upper
93 100
63 71.5 56.7 85.1
95 32.9 26.2 39.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022




#12
2-Chlorophenol
Concen: 32.513 ng/ul
RT: 7.410 min Scan# 752
Delta R.T. -0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16



Tgt Ion:128 Resp: 540136
Ion Ratio Lower Upper
128 100
64 40.4 32.7 49.1
130 32.3 26.3 39.5





#13

2-Methylphenol

Concen: 31.550 ng/ul

RT: 8.299 min Scan# 903

Delta R.T. -0.000 min

Lab File: BP012114.D

Acq: 13 Oct 2022 11:16

Instrument :

BNA_P

ClientSampleId :

SLCS251

Tgt Ion:108 Resp: 476400

Ion Ratio Lower Upper





108 100

107 88.5 71.9 107.9

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 29.207 ng/ul

RT: 8.381 min Scan# 917

Delta R.T. -0.000 min

Lab File: BP012114.D

Acq: 13 Oct 2022 11:16




Tgt Ion: 45 Resp: 478193


Ion Ratio Lower Upper

45 100

77 18.5 14.4 21.6


79 14.6 11.3 16.9





#16
Acetophenone
Concen: 31.637 ng/ul
RT: 8.681 min Scan# 968
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
Client SampleId :
SLCS251




Tgt Ion: 105 Resp: 722487
Ion Ratio Lower Upper
105 100
77 70.4 57.0 85.4
51 19.7 16.4 24.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022




#17
N-Nitroso-di-n-propylamine
Concen: 30.802 ng/ul
RT: 8.669 min Scan# 966
Delta R.T. 0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16



Tgt Ion: 70 Resp: 320377
Ion Ratio Lower Upper
70 100
42 42.1 34.6 52.0
101 10.8 8.0 12.0
130 25.2 18.3 27.5









#18
4-Methylphenol
Concen: 31.760 ng/ul
RT: 8.628 min Scan# 959
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16

Instrument :
BNA_P
ClientSampleId :
SLCS251

Tgt Ion:108 Resp: 516911
Ion Ratio Lower Upper
108 100
107 112.3 86.3 129.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#19
Hexachloroethane
Concen: 32.200 ng/ul
RT: 8.928 min Scan# 1010
Delta R.T. -0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16


Tgt Ion:117 Resp: 206496
Ion Ratio Lower Upper
117 100
201 103.1 82.3 123.5
199 66.3 50.0 75.0





#22
Nitrobenzene
Concen: 32.943 ng/ul
RT: 9.057 min Scan# 1032
Delta R.T. -0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251




Tgt Ion: 77 Resp: 520772
Ion Ratio Lower Upper
77 100
123 49.9 39.4 59.0
65 12.6 9.9 14.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022




#23
Isophorone
Concen: 32.345 ng/ul
RT: 9.593 min Scan# 1123
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16



Tgt Ion: 82 Resp: 906985
Ion Ratio Lower Upper
82 100
95 6.7 5.4 8.0
138 19.6 15.2 22.8









#25
2-Nitrophenol
Concen: 34.910 ng/ul
RT: 9.769 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16

Instrument :
BNA_P
ClientSampleId :
SLCS251

Tgt Ion:139 Resp: 286800
Ion Ratio Lower Upper
139 100
65 43.2 36.3 54.5
109 25.9 21.2 31.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#26
2,4-Dimethylphenol
Concen: 31.793 ng/ul
RT: 9.834 min Scan# 1164
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16

Tgt Ion:107 Resp: 508686
Ion Ratio Lower Upper
107 100
121 54.0 43.0 64.4
122 92.2 71.4 107.2









#27
Bis(2-Chloroethoxy)methane
Concen: 31.696 ng/ul
RT: 10.069 min Scan# 1204
Delta R.T. -0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16

Instrument :
BNA_P
ClientSampleId :
SLCS251

Tgt Ion: 93 Resp: 609912
Ion Ratio Lower Upper
93 100
95 32.5 25.7 38.5
123 11.8 9.2 13.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#29
2,4-Dichlorophenol
Concen: 33.251 ng/ul
RT: 10.304 min Scan# 1244
Delta R.T. -0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16


Tgt Ion:162 Resp: 473552
Ion Ratio Lower Upper
162 100
164 64.9 51.5 77.3
98 33.3 26.3 39.5





#30
Naphthalene
Concen: 31.617 ng/ul
RT: 10.704 min Scan# 1312
Delta R.T. -0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251



Tgt Ion:128 Resp: 1631425
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 12.8 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#32
4-Chloroaniline
Concen: 29.452 ng/ul
RT: 10.822 min Scan# 1332
Delta R.T. -0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16

Tgt Ion:127 Resp: 606413
Ion Ratio Lower Upper
127 100
129 31.8 25.6 38.4









#33
Hexachlorobutadiene
Concen: 31.334 ng/ul
RT: 10.987 min Scan# 1360
Delta R.T. -0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16

Instrument :
BNA_P
ClientSampleId :
SLCS251

Tgt Ion:225 Resp: 27671
Ion Ratio Lower Upper
225 100
223 62.5 50.9 76.3
227 63.6 51.1 76.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#34
Caprolactam
Concen: 36.510 ng/ul m
RT: 11.622 min Scan# 1468
Delta R.T. 0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16


Tgt Ion:113 Resp: 142882
Ion Ratio Lower Upper
113 100
55 142.3 115.2 172.8
56 106.3 85.6 128.4





#35
4-Chloro-3-methylphenol
Concen: 35.064 ng/ul
RT: 11.951 min Scan# 1524
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251




Tgt Ion:107 Resp: 473084
Ion Ratio Lower Upper
107 100
144 27.4 23.0 34.4
142 85.2 69.8 104.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022




#36
2-Methylnaphthalene
Concen: 31.612 ng/ul
RT: 12.316 min Scan# 1586
Delta R.T. -0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Tgt Ion:142 Resp: 1050352
Ion Ratio Lower Upper
142 100
141 88.6 70.7 106.1





#37
1-Methylnaphthalene
Concen: 31.659 ng/ul
RT: 12.540 min Scan# 1623
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251




Tgt Ion:142 Resp: 105342
Ion Ratio Lower Upper
142 100
141 91.8 72.9 109.3
116 3.2 2.8 4.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 30.900 ng/ul
RT: 12.687 min Scan# 1649
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Tgt Ion:216 Resp: 539909
Ion Ratio Lower Upper
216 100
214 77.0 63.0 94.4
179 19.8 16.3 24.5
108 16.0 12.5 18.7





#40
Hexachlorocyclopentadiene
Concen: 26.581 ng/ul
RT: 12.663 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251




Tgt Ion:237 Resp: 225588
Ion Ratio Lower Upper
237 100
235 61.2 49.8 74.8
272 12.7 9.9 14.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022




#41
2,4,6-Trichlorophenol
Concen: 33.777 ng/ul
RT: 12.928 min Scan# 1690
Delta R.T. -0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Tgt Ion:196 Resp: 355504
Ion Ratio Lower Upper
196 100
198 96.6 76.2 114.2
200 31.7 24.2 36.2





#42
2,4,5-Trichlorophenol
Concen: 34.595 ng/ul
RT: 13.004 min Scan# 1703
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251



Tgt Ion:196 Resp: 395723
Ion Ratio Lower Upper
196 100
198 96.2 75.4 113.2
200 31.9 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#43
1,1'-Biphenyl
Concen: 31.232 ng/ul
RT: 13.334 min Scan# 1759
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16


Tgt Ion:154 Resp: 1347713
Ion Ratio Lower Upper
154 100
153 40.3 32.2 48.2
76 11.7 10.3 15.5





#44
2-Chloronaphthalene
Concen: 31.505 ng/ul
RT: 13.375 min Scan# 1765
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251




Tgt Ion:162 Resp: 108562
Ion Ratio Lower Upper
162 100
164 33.1 26.0 39.0
127 36.9 29.4 44.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022




#45
2-Nitroaniline
Concen: 37.056 ng/ul
RT: 13.587 min Scan# 1802
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Tgt Ion: 65 Resp: 272388
Ion Ratio Lower Upper
65 100
92 73.7 61.6 92.4
138 128.4 96.5 144.7





#47
Dimethylphthalate
Concen: 33.141 ng/ul
RT: 13.957 min Scan# 1865
Delta R.T. -0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251



Tgt Ion:163 Resp: 1263982
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.4 7.9 11.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#48
2,6-Dinitrotoluene
Concen: 38.428 ng/ul
RT: 14.081 min Scan# 1886
Delta R.T. -0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16

Tgt Ion:165 Resp: 259485
Ion Ratio Lower Upper
165 100
89 53.2 43.4 65.0
121 21.7 18.1 27.1









#50
Acenaphthylene
Concen: 33.030 ng/ul
RT: 14.222 min Scan# 1910
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16

Instrument :
BNA_P
ClientSampleId :
SLCS251

Tgt Ion:152 Resp: 167300
Ion Ratio Lower Upper
152 100
151 20.1 15.7 23.5
153 13.3 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#51
3-Nitroaniline
Concen: 36.828 ng/ul
RT: 14.410 min Scan# 1942
Delta R.T. -0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16


Tgt Ion:138 Resp: 289569
Ion Ratio Lower Upper
138 100
108 9.6 8.5 12.7
92 106.5 85.8 128.8





#52
Acenaphthene
Concen: 31.886 ng/ul
RT: 14.563 min Scan# 1968
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251




Tgt Ion:153 Resp: 1149736
Ion Ratio Lower Upper
153 100
152 47.8 38.2 57.2
154 87.7 70.6 105.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022




#53
2,4-Dinitrophenol
Concen: 34.579 ng/ul
RT: 14.622 min Scan# 1978
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Tgt Ion:184 Resp: 150843
Ion Ratio Lower Upper
184 100
63 56.3 45.5 68.3
154 57.4 47.0 70.6





#55
4-Nitrophenol
Concen: 38.429 ng/ul
RT: 14.728 min Scan# 1995
Delta R.T. 0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251



Tgt Ion:109 Resp: 188814
Ion Ratio Lower Upper
109 100
139 161.8 124.1 186.1
65 124.8 104.2 156.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#56
Dibenzofuran
Concen: 32.393 ng/ul
RT: 14.898 min Scan# 2025
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16


Tgt Ion:168 Resp: 1578540
Ion Ratio Lower Upper
168 100
139 36.3 28.2 42.4





#57
2,4-Dinitrotoluene
Concen: 40.948 ng/ul
RT: 14.875 min Scan# 2021
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251




Tgt Ion:165 Resp: 397389
Ion Ratio Lower Upper
165 100
63 34.0 31.1 46.7
182 2.6 1.8 2.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 37.805 ng/ul
RT: 15.128 min Scan# 2064
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Tgt Ion:232 Resp: 337065
Ion Ratio Lower Upper
232 100
131 45.5 37.8 56.6
130 2.2 1.8 2.6
166 31.9 26.6 39.8





#59
Diethylphthalate
Concen: 35.069 ng/ul
RT: 15.328 min Scan# 2097
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251



Tgt Ion:149 Resp: 1224929
Ion Ratio Lower Upper
149 100
177 22.2 17.5 26.3
150 12.2 9.8 14.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#61
Fluorene
Concen: 33.829 ng/ul
RT: 15.551 min Scan# 2136
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16


Tgt Ion:166 Resp: 1294046
Ion Ratio Lower Upper
166 100
165 97.3 79.5 119.3
167 13.6 10.9 16.3





#62
4-Chlorophenyl-phenylether
Concen: 33.005 ng/ul
RT: 15.545 min Scan# 2135
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251



Tgt Ion:204 Resp: 610764
Ion Ratio Lower Upper
204 100
206 32.8 25.4 38.2
141 55.2 46.2 69.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#63
4-Nitroaniline
Concen: 47.540 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16


Tgt Ion:138 Resp: 341314
Ion Ratio Lower Upper
138 100
92 46.6 39.2 58.8
108 65.3 55.4 83.0





#66
4,6-Dinitro-2-methylphenol
Concen: 33.116 ng/ul
RT: 15.639 min Scan# 2150
Delta R.T. 0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251



Tgt Ion:198 Resp: 234491
Ion Ratio Lower Upper
198 100
182 4.5 4.9 7.3
77 20.5 20.0 30.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#67
N-Nitrosodiphenylamine
Concen: 29.908 ng/ul
RT: 15.763 min Scan# 2172
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16


Tgt Ion:169 Resp: 1098070
Ion Ratio Lower Upper
169 100
168 67.4 53.5 80.3
167 36.4 29.5 44.3





#68
4-Bromophenyl-phenylether
Concen: 30.822 ng/ul
RT: 16.445 min Scan# 2288
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251



Tgt Ion:248 Resp: 372035
Ion Ratio Lower Upper
248 100
250 98.4 77.9 116.9
141 69.8 59.5 89.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#69
Hexachlorobenzene
Concen: 31.247 ng/ul
RT: 16.557 min Scan# 2307
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16


Tgt Ion:284 Resp: 437770
Ion Ratio Lower Upper
284 100
142 35.6 30.5 45.7
249 30.3 25.4 38.0





#70
Atrazine
Concen: 34.227 ng/ul
RT: 16.722 min Scan# 2335
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251



Tgt Ion:200 Resp: 386431
Ion Ratio Lower Upper
200 100
173 26.9 22.1 33.1
215 48.9 38.8 58.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#71
Pentachlorophenol
Concen: 35.073 ng/ul
RT: 16.904 min Scan# 2366
Delta R.T. -0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16


Tgt Ion:266 Resp: 287882
Ion Ratio Lower Upper
266 100
264 64.5 50.6 75.8
268 63.0 51.8 77.6





#72
Phenanthrene
Concen: 32.279 ng/ul
RT: 17.298 min Scan# 2433
Delta R.T. -0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument : BNA_P
ClientSampleId : SLCS251




Tgt Ion:178 Resp: 2177501
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.4
176 19.2 15.8 23.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022




#74
Anthracene
Concen: 33.002 ng/ul
RT: 17.392 min Scan# 2449
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16



Tgt Ion:178 Resp: 2203115
Ion Ratio Lower Upper
178 100
179 15.7 12.2 18.4
176 18.8 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 27.773 ng/uL
RT: 13.293 min Scan# 1752
Delta R.T. -0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16

Instrument :
BNA_P
ClientSampleId :
SLCS251

Tgt Ion:216 Resp: 556695
Ion Ratio Lower Upper
216 100
214 78.0 62.6 94.0
218 47.8 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#76
Pentachlorobenzene
Concen: 26.201 ng/uL
RT: 14.816 min Scan# 2011
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16


Tgt Ion:250 Resp: 515769
Ion Ratio Lower Upper
250 100
248 63.2 51.4 77.2
252 64.3 50.7 76.1





#77
 Carbazole
 Concen: 35.352 ng/ul
 RT: 17.669 min Scan# 2496
 Delta R.T. -0.000 min
 Lab File: BP012114.D
 Acq: 13 Oct 2022 11:16




Instrument :
 BNA_P
 ClientSampleId :
 SLCS251



Tgt Ion:167 Resp: 2104861
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.0 25.6
 139 12.7 10.1 15.1




Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
 Supervised By :mohammad ahmed 10/14/2022



#78
 Di-n-butylphthalate
 Concen: 36.039 ng/ul
 RT: 18.216 min Scan# 2589
 Delta R.T. -0.000 min
 Lab File: BP012114.D
 Acq: 13 Oct 2022 11:16


Tgt Ion:149 Resp: 2129359
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.2 10.8
 104 4.4 3.4 5.2





#80
Fluoranthene
Concen: 31.050 ng/ul
RT: 19.269 min Scan# 2768
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251



Tgt Ion:202 Resp: 2662828
Ion Ratio Lower Upper
202 100
101 12.2 9.0 13.6
100 9.1 7.0 10.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#82
Pyrene
Concen: 31.158 ng/ul
RT: 19.622 min Scan# 2828
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16

Tgt Ion:202 Resp: 2827839
Ion Ratio Lower Upper
202 100
101 14.3 10.4 15.6
100 11.5 8.5 12.7





#83

Butylbenzylphthalate

Concen: 37.274 ng/ul

RT: 20.492 min Scan# 2976

Delta R.T. -0.006 min

Lab File: BP012114.D

Acq: 13 Oct 2022 11:16

Instrument :

BNA_P

ClientSampleId :

SLCS251

Tgt Ion:149 Resp: 1072399

Ion Ratio Lower Upper

149 100





91 67.3 48.5 72.7

206 19.9 17.9 26.9

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#84

3,3'-Dichlorobenzidine

Concen: 32.012 ng/ul

RT: 21.269 min Scan# 3108

Delta R.T. -0.000 min

Lab File: BP012114.D

Acq: 13 Oct 2022 11:16




Tgt Ion:252 Resp: 963697


Ion Ratio Lower Upper

252 100

254 63.9 51.4 77.2


126 11.5 7.8 11.8





#85
Benzo(a)anthracene
Concen: 33.227 ng/ul
RT: 21.333 min Scan# 3119
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251



Tgt Ion:228 Resp: 2857316
Ion Ratio Lower Upper
228 100
229 20.2 15.7 23.5
226 28.5 21.9 32.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 38.296 ng/ul
RT: 21.251 min Scan# 3105
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16


Tgt Ion:149 Resp: 1485307
Ion Ratio Lower Upper
149 100
167 27.7 21.9 32.9
279 3.9 3.4 5.2





#87
Chrysene
Concen: 32.422 ng/ul
RT: 21.386 min Scan# 3127
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251



Tgt Ion:228 Resp: 2683725
Ion Ratio Lower Upper
228 100
226 30.7 24.0 36.0
229 19.9 15.5 23.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#89
Di-n-octyl phthalate
Concen: 36.674 ng/ul
RT: 22.180 min Scan# 3263
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16


Tgt Ion:149 Resp: 2644507





#90
Benzo(b)fluoranthene
Concen: 32.951 ng/ul
RT: 23.027 min Scan# 3406
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
ClientSampleId :
SLCS251



Tgt Ion:252 Resp: 3004070
Ion Ratio Lower Upper
252 100
253 22.7 17.6 26.4
125 11.5 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#91
Benzo(k)fluoranthene
Concen: 32.918 ng/ul
RT: 23.074 min Scan# 3415
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16

Tgt Ion:252 Resp: 2951952
Ion Ratio Lower Upper
252 100
253 22.5 17.6 26.4
125 11.0 8.2 12.2





#93

Benzo(a)pyrene

Concen: 33.142 ng/ul

RT: 23.663 min Scan# 3962

Delta R.T. 0.006 min

Lab File: BP012114.D

Acq: 13 Oct 2022 11:16

Instrument :

BNA_P

ClientSampleId :

SLCS251

Tgt Ion:252 Resp: 2695648

Ion Ratio Lower Upper

252 100





253 22.0 17.6 26.4

125 11.9 8.9 13.3

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2022
Supervised By :mohammad ahmed 10/14/2022



#94

Indeno(1,2,3-cd)pyrene

Concen: 33.979 ng/ul

RT: 26.292 min Scan# 3962

Delta R.T. 0.012 min

Lab File: BP012114.D

Acq: 13 Oct 2022 11:16




Tgt Ion:276 Resp: 3494564


Ion Ratio Lower Upper

276 100

138 28.1 19.1 28.7


277 25.7 19.9 29.9





#95
Dibenzo(a,h)anthracene
Concen: 32.234 ng/ul
RT: 26.304 min Scan# 3964
Delta R.T. 0.006 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16




Instrument :
BNA_P
Client Sample Id :
SLCS251




Tgt Ion: 278 Resp: 2973996
Ion Ratio Lower Upper
278 100
139 18.1 12.5 18.7
279 23.9 19.0 28.4

Manual Integrations
APPROVED

Reviewed By : Jagrut Upadhyay 10/14/2022
Supervised By : mohammad ahmed 10/14/2022



#96
Benzo(g,h,i)perylene
Concen: 31.526 ng/ul
RT: 27.057 min Scan# 4092
Delta R.T. -0.000 min
Lab File: BP012114.D
Acq: 13 Oct 2022 11:16



Tgt Ion: 276 Resp: 2923275
Ion Ratio Lower Upper
276 100
138 24.2 16.5 24.7
277 23.5 19.0 28.4

