

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BP012115.D
 Acq On : 13 Oct 2022 11:51
 Operator : CG/JU
 Sample : PB148249BS
 Misc :
 ALS Vial : 86 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS249

Quant Time: Oct 13 23:31:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 23:03:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.852	152	266657	20.000	ng/u1	0.00
20) Naphthalene-d8	10.657	136	1059643	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.498	164	577936	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.257	188	1134245	20.000	ng/u1	0.00
79) Chrysene-d12	21.351	240	1272350	20.000	ng/u1	0.00
88) Perylene-d12	23.763	264	1344148	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	38452	5.685	ng/uL	0.00
4) Pyridine-d5	3.693	84	494194	27.476	ng/u1	0.00
7) Phenol-d5	7.016	99	653574	30.475	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.181	67	359453	29.528	ng/u1	0.00
11) 2-Chlorophenol-d4	7.381	132	537130	31.552	ng/u1	0.00
15) 4-Methylphenol-d8	8.563	113	487390	29.588	ng/u1	0.00
21) Nitrobenzene-d5	9.016	128	251113	33.554	ng/u1	0.00
24) 2-Nitrophenol-d4	9.740	143	266508	35.121	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.275	165	474286	32.250	ng/u1	0.00
31) 4-Chloroaniline-d4	10.799	131	612052	28.242	ng/u1	0.00
46) Dimethylphthalate-d6	13.910	166	1141092	31.176	ng/u1	0.00
49) Acenaphthylene-d8	14.192	160	1441130	31.841	ng/u1	0.00
54) 4-Nitrophenol-d4	14.710	143	242295	33.719	ng/u1	0.00
60) Fluorene-d10	15.492	176	1037422	31.327	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.622	200	189340	31.614	ng/u1	0.00
73) Anthracene-d10	17.357	188	1571319	31.903	ng/u1	0.00
81) Pyrene-d10	19.592	212	1916995	29.409	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.610	264	2132391	32.634	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.311	88	76764	10.786	ng/uL	97
5) Pyridine	3.717	79	484696	28.125	ng/u1	95
6) Benzaldehyde	6.993	77	386119	38.374	ng/u1	100
8) Phenol	7.046	94	633869	28.154	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.275	93	500635	28.183	ng/u1	99
12) 2-Chlorophenol	7.410	128	541458	29.308	ng/u1	99
13) 2-Methylphenol	8.299	108	470938	28.045	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.387	45	479869	26.356	ng/u1	97
16) Acetophenone	8.681	105	717761	28.263	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.663	70	317880	27.482	ng/u1	98
18) 4-Methylphenol	8.628	108	506286	27.972	ng/u1	98
19) Hexachloroethane	8.928	117	209505	29.376	ng/u1	97
22) Nitrobenzene	9.057	77	521702	30.410	ng/u1	100
23) Isophorone	9.587	82	881476	28.967	ng/u1	99
25) 2-Nitrophenol	9.769	139	284896	31.956	ng/u1	96
26) 2,4-Dimethylphenol	9.834	107	500448	28.823	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.069	93	602849	28.870	ng/u1	99
29) 2,4-Dichlorophenol	10.304	162	460735	29.811	ng/u1	99
30) Naphthalene	10.704	128	1608210	28.720	ng/u1	100
32) 4-Chloroaniline	10.822	127	575779	25.769	ng/u1	99
33) Hexachlorobutadiene	10.987	225	278554	29.065	ng/u1	99
34) Caprolactam	11.616	113	123249	29.021	ng/u1	98
35) 4-Chloro-3-methylphenol	11.951	107	432905	29.567	ng/u1	100
36) 2-Methylnaphthalene	12.316	142	1013198	28.100	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BP012115.D
 Acq On : 13 Oct 2022 11:51
 Operator : CG/JU
 Sample : PB148249BS
 Misc :
 ALS Vial : 86 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS249

Quant Time: Oct 13 23:31:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 23:03:21 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.540	142	1013075	28.056	ng/u1	99
39) 1,2,4,5-Tetrachloroben...	12.687	216	519577	29.410	ng/u1	99
40) Hexachlorocyclopentadiene	12.657	237	219686	25.601	ng/u1	100
41) 2,4,6-Trichlorophenol	12.928	196	328894	30.906	ng/u1	99
42) 2,4,5-Trichlorophenol	13.004	196	353782	30.589	ng/u1	98
43) 1,1'-Biphenyl	13.328	154	1275158	29.226	ng/u1	99
44) 2-Chloronaphthalene	13.369	162	1022910	29.359	ng/u1	98
45) 2-Nitroaniline	13.587	65	240112	32.306	ng/u1	94
47) Dimethylphthalate	13.957	163	1131892	29.352	ng/u1	99
48) 2,6-Dinitrotoluene	14.081	165	227770	33.361	ng/u1	98
50) Acenaphthylene	14.222	152	1540972	30.090	ng/u1	100
51) 3-Nitroaniline	14.416	138	246453	31.000	ng/u1	97
52) Acenaphthene	14.563	153	1060195	29.080	ng/u1	99
53) 2,4-Dinitrophenol	14.622	184	118605	26.890	ng/u1	97
55) 4-Nitrophenol	14.722	109	156131	31.428	ng/u1	98
56) Dibenzofuran	14.898	168	1434179	29.107	ng/u1	100
57) 2,4-Dinitrotoluene	14.869	165	331156	33.749	ng/u1	97
58) 2,3,4,6-Tetrachlorophenol	15.128	232	288746	32.030	ng/u1	95
59) Diethylphthalate	15.322	149	1068705	30.261	ng/u1	99
61) Fluorene	15.551	166	1155702	29.881	ng/u1	98
62) 4-Chlorophenyl-phenyle...	15.545	204	537612	28.733	ng/u1	97
63) 4-Nitroaniline	15.581	138	277856	38.276	ng/u1	97
66) 4,6-Dinitro-2-methylph...	15.634	198	190721	29.456	ng/u1#	96
67) N-Nitrosodiphenylamine	15.763	169	952814	28.381	ng/u1	99
68) 4-Bromophenyl-phenylether	16.445	248	315489	28.584	ng/u1	97
69) Hexachlorobenzene	16.557	284	368787	28.788	ng/u1	95
70) Atrazine	16.722	200	314541	30.468	ng/u1	100
71) Pentachlorophenol	16.904	266	235511	31.378	ng/u1	99
72) Phenanthrene	17.298	178	1808342	29.316	ng/u1	99
74) Anthracene	17.392	178	1832206	30.016	ng/u1	99
75) 1,2,3,4-Tetrachloroben...	13.293	216	530558	28.947	ng/uL	99
76) Pentachlorobenzene	14.816	250	464619	25.812	ng/uL	98
77) Carbazole	17.669	167	1714803	31.497	ng/u1	99
78) Di-n-butylphthalate	18.216	149	1694462	31.363	ng/u1	100
80) Fluoranthene	19.269	202	2182789	27.329	ng/u1	99
82) Pyrene	19.622	202	2338060	27.661	ng/u1	97
83) Butylbenzylphthalate	20.498	149	840272	31.360	ng/u1	96
84) 3,3'-Dichlorobenzidine	21.269	252	809176	28.861	ng/u1	99
85) Benzo(a)anthracene	21.333	228	2412215	30.120	ng/u1	100
86) Bis(2-ethylhexyl)phtha...	21.251	149	1169228	32.369	ng/u1	100
87) Chrysene	21.386	228	2280264	29.579	ng/u1	99
89) Di-n-octyl phthalate	22.180	149	2138210	31.774	ng/u1	100
90) Benzo(b)fluoranthene	23.027	252	2584113	30.372	ng/u1	97
91) Benzo(k)fluoranthene	23.074	252	2499885	29.871	ng/u1	98
93) Benzo(a)pyrene	23.663	252	2310042	30.432	ng/u1	98
94) Indeno(1,2,3-cd)pyrene	26.286	276	3005443	31.314	ng/u1	97
95) Dibenzo(a,h)anthracene	26.309	278	2572623	29.878	ng/u1	98
96) Benzo(g,h,i)perylene	27.057	276	2524925	29.177	ng/u1	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101122\
 Data File : BP012115.D
 Acq On : 13 Oct 2022 11:51
 Operator : CG/JU
 Sample : PB148249BS
 Misc :
 ALS Vial : 86 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS249

Quant Time: Oct 13 23:31:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 23:03:21 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 10.786 ng/uL
RT: 3.311 min Scan# 51
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion: 88	Resp: 76764
Ion Ratio	Lower Upper
88	100
43	18.4 16.6 25.0
58	63.9 52.6 79.0



#5
Pyridine
Concen: 28.125 ng/uL
RT: 3.717 min Scan# 124
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion: 79	Resp: 484696
Ion Ratio	Lower Upper
79	100
52	58.8 51.2 76.8
51	27.0 22.8 34.2









#6
Benzaldehyde
Concen: 38.374 ng/ul
RT: 6.993 min Scan# 681
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion: 77 Resp: 386119
Ion Ratio Lower Upper
77 100
105 106.9 85.2 127.8
106 104.1 83.1 124.7



#8
Phenol
Concen: 28.154 ng/ul
RT: 7.046 min Scan# 690
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion: 94 Resp: 633869
Ion Ratio Lower Upper
94 100
65 24.3 19.4 29.0
66 31.6 24.6 37.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 28.183 ng/ul
 RT: 7.275 min Scan# 729
 Delta R.T. -0.006 min
 Lab File: BP012115.D
 Acq: 13 Oct 2022 11:51




Instrument :
 BNA_P
 ClientSampleId :
 SLCS249


Tgt Ion:	93	Resp:	500635
Ion	Ratio	Lower	Upper
93	100		
63	69.6	56.7	85.1
95	32.9	26.2	39.2



#12
 2-Chlorophenol
 Concen: 29.308 ng/ul
 RT: 7.410 min Scan# 752
 Delta R.T. -0.006 min
 Lab File: BP012115.D
 Acq: 13 Oct 2022 11:51

Tgt Ion:	128	Resp:	541458
Ion	Ratio	Lower	Upper
128	100		
64	40.4	32.7	49.1
130	33.2	26.3	39.5









#13
2-Methylphenol
Concen: 28.045 ng/ul
RT: 8.299 min Scan# 903
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:108 Resp: 470938
Ion Ratio Lower Upper
108 100
107 87.7 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 26.356 ng/ul
RT: 8.387 min Scan# 918
Delta R.T. 0.006 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion: 45 Resp: 479869
Ion Ratio Lower Upper
45 100
77 19.3 14.4 21.6
79 15.1 11.3 16.9









#16
Acetophenone
Concen: 28.263 ng/ul
RT: 8.681 min Scan# 968
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:105 Resp: 717761
Ion Ratio Lower Upper
105 100
77 71.7 57.0 85.4
51 19.6 16.4 24.6



#17
N-Nitroso-di-n-propylamine
Concen: 27.482 ng/ul
RT: 8.663 min Scan# 965
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion: 70 Resp: 317880
Ion Ratio Lower Upper
70 100
42 42.5 34.6 52.0
101 10.2 8.0 12.0
130 24.2 18.3 27.5









#18
4-Methylphenol
Concen: 27.972 ng/ul
RT: 8.628 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:108 Resp: 506286
Ion Ratio Lower Upper
108 100
107 110.2 86.3 129.5



#19
Hexachloroethane
Concen: 29.376 ng/ul
RT: 8.928 min Scan# 1010
Delta R.T. -0.006 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion:117 Resp: 209505
Ion Ratio Lower Upper
117 100
201 104.3 82.3 123.5
199 66.0 50.0 75.0









#22
Nitrobenzene
Concen: 30.410 ng/ul
RT: 9.057 min Scan# 1032
Delta R.T. -0.006 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion: 77 Resp: 521702
Ion Ratio Lower Upper
77 100
123 49.4 39.4 59.0
65 12.4 9.9 14.9



#23
Isophorone
Concen: 28.967 ng/ul
RT: 9.587 min Scan# 1122
Delta R.T. -0.006 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion: 82 Resp: 881476
Ion Ratio Lower Upper
82 100
95 6.7 5.4 8.0
138 19.5 15.2 22.8









#25
2-Nitrophenol
Concen: 31.956 ng/ul
RT: 9.769 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:139 Resp: 284896
Ion Ratio Lower Upper
139 100
65 41.9 36.3 54.5
109 25.4 21.2 31.8



#26
2,4-Dimethylphenol
Concen: 28.823 ng/ul
RT: 9.834 min Scan# 1164
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion:107 Resp: 500448
Ion Ratio Lower Upper
107 100
121 53.8 43.0 64.4
122 89.9 71.4 107.2





#27

Bis(2-Chloroethoxy)methane

Concen: 28.870 ng/ul

RT: 10.069 min Scan# 1204

Delta R.T. -0.006 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51

Instrument :

BNA_P

ClientSampleId :

SLCS249





Tgt Ion: 93 Resp: 602849

Ion Ratio Lower Upper

93 100

95 32.3 25.7 38.5

123 12.2 9.2 13.8



#29

2,4-Dichlorophenol

Concen: 29.811 ng/ul

RT: 10.304 min Scan# 1244

Delta R.T. -0.006 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51




Tgt Ion:162 Resp: 460735


Ion Ratio Lower Upper

162 100

164 65.1 51.5 77.3

98 33.0 26.3 39.5









#30
Naphthalene
Concen: 28.720 ng/ul
RT: 10.704 min Scan# 1312
Delta R.T. -0.006 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:128 Resp: 1608210
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 12.8 10.2 15.4



#32
4-Chloroaniline
Concen: 25.769 ng/ul
RT: 10.822 min Scan# 1332
Delta R.T. -0.006 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion:127 Resp: 575779
Ion Ratio Lower Upper
127 100
129 32.6 25.6 38.4





#33

Hexachlorobutadiene

Concen: 29.065 ng/ul

RT: 10.987 min Scan# 1360

Delta R.T. -0.006 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51

Instrument :

BNA_P

ClientSampleId :

SLCS249





Tgt Ion:225 Resp: 278554

Ion Ratio Lower Upper

225 100

223 63.0 50.9 76.3

227 65.4 51.1 76.7



#34

Caprolactam

Concen: 29.021 ng/ul

RT: 11.616 min Scan# 1467

Delta R.T. -0.000 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51




Tgt Ion:113 Resp: 123249


Ion Ratio Lower Upper

113 100

55 141.0 115.2 172.8

56 104.4 85.6 128.4









#35
4-Chloro-3-methylphenol
Concen: 29.567 ng/ul
RT: 11.951 min Scan# 1524
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:107 Resp: 432905
Ion Ratio Lower Upper
107 100
144 27.8 23.0 34.4
142 87.3 69.8 104.8



#36
2-Methylnaphthalene
Concen: 28.100 ng/ul
RT: 12.316 min Scan# 1586
Delta R.T. -0.006 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion:142 Resp: 1013198
Ion Ratio Lower Upper
142 100
141 88.9 70.7 106.1









#37
1-Methylnaphthalene
Concen: 28.056 ng/ul
RT: 12.540 min Scan# 1623
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:142 Resp: 1013075
Ion Ratio Lower Upper
142 100
141 91.7 72.9 109.3
116 3.4 2.8 4.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 29.410 ng/ul
RT: 12.687 min Scan# 1649
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion:216 Resp: 519577
Ion Ratio Lower Upper
216 100
214 77.5 63.0 94.4
179 20.3 16.3 24.5
108 15.7 12.5 18.7









#40
Hexachlorocyclopentadiene
Concen: 25.601 ng/ul
RT: 12.657 min Scan# 1644
Delta R.T. -0.006 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:237	Resp: 219686
Ion Ratio	Lower Upper
237 100	
235 62.6	49.8 74.8
272 12.8	9.9 14.9



#41
2,4,6-Trichlorophenol
Concen: 30.906 ng/ul
RT: 12.928 min Scan# 1690
Delta R.T. -0.006 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion:196	Resp: 328894
Ion Ratio	Lower Upper
196 100	
198 94.6	76.2 114.2
200 31.3	24.2 36.2





#42

2,4,5-Trichlorophenol

Concen: 30.589 ng/ul

RT: 13.004 min Scan# 1703

Delta R.T. -0.000 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51

Instrument :

BNA_P

ClientSampleId :

SLCS249





Tgt Ion:196 Resp: 353782

Ion Ratio Lower Upper

196 100

198 97.1 75.4 113.2

200 31.1 24.8 37.2



#43

1,1'-Biphenyl

Concen: 29.226 ng/ul

RT: 13.328 min Scan# 1758

Delta R.T. -0.006 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51




Tgt Ion:154 Resp: 1275158


Ion Ratio Lower Upper

154 100

153 39.4 32.2 48.2

76 12.2 10.3 15.5





#44

2-Chloronaphthalene

Concen: 29.359 ng/ul

RT: 13.369 min Scan# 1765

Delta R.T. -0.006 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51

Instrument :

BNA_P

ClientSampleId :

SLCS249





Tgt Ion:162 Resp: 1022910

Ion Ratio Lower Upper

162 100

164 32.7 26.0 39.0

127 38.4 29.4 44.2



#45

2-Nitroaniline

Concen: 32.306 ng/ul

RT: 13.587 min Scan# 1802

Delta R.T. -0.000 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51




Tgt Ion: 65 Resp: 240112


Ion Ratio Lower Upper

65 100

92 77.6 61.6 92.4

138 130.4 96.5 144.7









#47
Dimethylphthalate
Concen: 29.352 ng/ul
RT: 13.957 min Scan# 1865
Delta R.T. -0.006 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:163 Resp: 1131892
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.2 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 33.361 ng/ul
RT: 14.081 min Scan# 1886
Delta R.T. -0.006 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion:165 Resp: 227770
Ion Ratio Lower Upper
165 100
89 53.2 43.4 65.0
121 21.6 18.1 27.1









#50
Acenaphthylene
Concen: 30.090 ng/ul
RT: 14.222 min Scan# 1910
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:152 Resp: 1540972
Ion Ratio Lower Upper
152 100
151 19.7 15.7 23.5
153 13.1 10.4 15.6



#51
3-Nitroaniline
Concen: 31.000 ng/ul
RT: 14.416 min Scan# 1943
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion:138 Resp: 246453
Ion Ratio Lower Upper
138 100
108 9.5 8.5 12.7
92 104.5 85.8 128.8









#52
Acenaphthene
Concen: 29.080 ng/ul
RT: 14.563 min Scan# 1968
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:153 Resp: 1060195
Ion Ratio Lower Upper
153 100
152 47.0 38.2 57.2
154 87.6 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 26.890 ng/ul
RT: 14.622 min Scan# 1978
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion:184 Resp: 118605
Ion Ratio Lower Upper
184 100
63 55.8 45.5 68.3
154 55.2 47.0 70.6









#55
4-Nitrophenol
Concen: 31.428 ng/ul
RT: 14.722 min Scan# 1995
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:109 Resp: 156131
Ion Ratio Lower Upper
109 100
139 158.8 124.1 186.1
65 131.3 104.2 156.4



#56
Dibenzofuran
Concen: 29.107 ng/ul
RT: 14.898 min Scan# 2025
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion:168 Resp: 1434179
Ion Ratio Lower Upper
168 100
139 35.1 28.2 42.4









#57
2,4-Dinitrotoluene
Concen: 33.749 ng/ul
RT: 14.869 min Scan# 2020
Delta R.T. -0.006 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:165 Resp: 331156
Ion Ratio Lower Upper
165 100
63 36.8 31.1 46.7
182 2.4 1.8 2.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 32.030 ng/ul
RT: 15.128 min Scan# 2064
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion:232 Resp: 288746
Ion Ratio Lower Upper
232 100
131 43.6 37.8 56.6
130 2.0 1.8 2.6
166 30.7 26.6 39.8





#59

Diethylphthalate

Concen: 30.261 ng/ul

RT: 15.322 min Scan# 2097

Delta R.T. -0.006 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51

Instrument :

BNA_P

ClientSampleId :

SLCS249





Tgt Ion:149 Resp: 1068705

Ion Ratio Lower Upper

149 100

177 21.5 17.5 26.3

150 12.2 9.8 14.8



#61

Fluorene

Concen: 29.881 ng/ul

RT: 15.551 min Scan# 2136

Delta R.T. -0.000 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51




Tgt Ion:166 Resp: 1155702


Ion Ratio Lower Upper

166 100

165 97.3 79.5 119.3

167 13.4 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 28.733 ng/ul
RT: 15.545 min Scan# 2135
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:204 Resp: 537612
Ion Ratio Lower Upper
204 100
206 32.9 25.4 38.2
141 54.6 46.2 69.2



#63
4-Nitroaniline
Concen: 38.276 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion:138 Resp: 277856
Ion Ratio Lower Upper
138 100
92 47.1 39.2 58.8
108 66.2 55.4 83.0









#66
4,6-Dinitro-2-methylphenol
Concen: 29.456 ng/ul
RT: 15.634 min Scan# 2150
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:198 Resp: 190721
Ion Ratio Lower Upper
198 100
182 4.8 4.9 7.3#
77 23.1 20.0 30.0



#67
N-Nitrosodiphenylamine
Concen: 28.381 ng/ul
RT: 15.763 min Scan# 2172
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion:169 Resp: 952814
Ion Ratio Lower Upper
169 100
168 67.3 53.5 80.3
167 37.4 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 28.584 ng/ul
RT: 16.445 min Scan# 211
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:248 Resp: 315489
Ion Ratio Lower Upper
248 100
250 98.6 77.9 116.9
141 69.1 59.5 89.3



#69
Hexachlorobenzene
Concen: 28.788 ng/ul
RT: 16.557 min Scan# 2307
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion:284 Resp: 368787
Ion Ratio Lower Upper
284 100
142 33.7 30.5 45.7
249 30.5 25.4 38.0









#70
Atrazine
Concen: 30.468 ng/ul
RT: 16.722 min Scan# 2335
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:200 Resp: 314541
Ion Ratio Lower Upper
200 100
173 27.1 22.1 33.1
215 48.4 38.8 58.2



#71
Pentachlorophenol
Concen: 31.378 ng/ul
RT: 16.904 min Scan# 2366
Delta R.T. -0.006 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion:266 Resp: 235511
Ion Ratio Lower Upper
266 100
264 62.8 50.6 75.8
268 63.1 51.8 77.6





#72

Phenanthrene

Concen: 29.316 ng/ul

RT: 17.298 min Scan# 2433

Delta R.T. -0.006 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51

Instrument :

BNA_P

ClientSampleId :

SLCS249





Tgt Ion:178 Resp: 1808342

Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.4

176 19.3 15.8 23.6



#74

Anthracene

Concen: 30.016 ng/ul

RT: 17.392 min Scan# 2449

Delta R.T. -0.000 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51




Tgt Ion:178 Resp: 1832206


Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.4

176 18.7 14.8 22.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 28.947 ng/uL

RT: 13.293 min Scan# 1752

Delta R.T. -0.006 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51

Instrument :

BNA_P

ClientSampleId :

SLCS249





Tgt Ion:216 Resp: 530558

Ion Ratio Lower Upper

216 100

214 77.9 62.6 94.0

218 47.1 38.9 58.3



#76

Pentachlorobenzene

Concen: 25.812 ng/uL

RT: 14.816 min Scan# 2011

Delta R.T. -0.000 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51




Tgt Ion:250 Resp: 464619


Ion Ratio Lower Upper

250 100

248 62.2 51.4 77.2

252 63.8 50.7 76.1





#77

Carbazole

Concen: 31.497 ng/ul

RT: 17.669 min Scan# 2496

Delta R.T. -0.000 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51

Instrument :

BNA_P

ClientSampleId :

SLCS249





Tgt Ion:167 Resp: 1714803

Ion Ratio Lower Upper

167 100

166 21.7 17.0 25.6

139 12.7 10.1 15.1



#78

Di-n-butylphthalate

Concen: 31.363 ng/ul

RT: 18.216 min Scan# 2589

Delta R.T. -0.000 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51




Tgt Ion:149 Resp: 1694462


Ion Ratio Lower Upper

149 100

150 9.1 7.2 10.8

104 4.3 3.4 5.2









#80
Fluoranthene
Concen: 27.329 ng/ul
RT: 19.269 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:202 Resp: 2182789
Ion Ratio Lower Upper
202 100
101 11.9 9.0 13.6
100 9.2 7.0 10.6



#82
Pyrene
Concen: 27.661 ng/ul
RT: 19.622 min Scan# 2828
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion:202 Resp: 2338060
Ion Ratio Lower Upper
202 100
101 14.2 10.4 15.6
100 11.4 8.5 12.7





#83

Butylbenzylphthalate

Concen: 31.360 ng/ul

RT: 20.498 min Scan# 2976

Delta R.T. -0.000 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51

Instrument :

BNA_P

ClientSampleId :

SLCS249





Tgt Ion:149 Resp: 840272

Ion Ratio Lower Upper

149 100

91 64.4 48.5 72.7

206 21.1 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 28.861 ng/ul

RT: 21.269 min Scan# 3108

Delta R.T. -0.000 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51




Tgt Ion:252 Resp: 809176


Ion Ratio Lower Upper

252 100

254 63.9 51.4 77.2

126 11.4 7.8 11.8





#85

Benzo(a)anthracene

Concen: 30.120 ng/ul

RT: 21.333 min Scan# 3118

Delta R.T. -0.000 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51

Instrument :

BNA_P

ClientSampleId :

SLCS249





Tgt Ion:228 Resp: 2412215

Ion Ratio Lower Upper

228 100

229 19.9 15.7 23.5

226 27.5 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 32.369 ng/ul

RT: 21.251 min Scan# 3105

Delta R.T. -0.000 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51




Tgt Ion:149 Resp: 1169228


Ion Ratio Lower Upper

149 100

167 27.4 21.9 32.9

279 3.7 3.4 5.2





#87

Chrysene

Concen: 29.579 ng/ul

RT: 21.386 min Scan# 3127

Delta R.T. -0.000 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51

Instrument :

BNA_P

ClientSampleId :

SLCS249





Tgt Ion:228 Resp: 2280264

Ion Ratio Lower Upper

228 100

226 30.8 24.0 36.0

229 19.9 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 31.774 ng/ul




RT: 22.180 min Scan# 3263


Delta R.T. -0.000 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51

Tgt Ion:149 Resp: 2138210









#90
Benzo(b)fluoranthene
Concen: 30.372 ng/ul
RT: 23.027 min Scan# 3406
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51




Instrument :
BNA_P
ClientSampleId :
SLCS249


Tgt Ion:252 Resp: 2584113
Ion Ratio Lower Upper
252 100
253 22.8 17.6 26.4
125 11.5 7.8 11.8



#91
Benzo(k)fluoranthene
Concen: 29.871 ng/ul
RT: 23.074 min Scan# 3415
Delta R.T. -0.000 min
Lab File: BP012115.D
Acq: 13 Oct 2022 11:51

Tgt Ion:252 Resp: 2499885
Ion Ratio Lower Upper
252 100
253 22.7 17.6 26.4
125 11.2 8.2 12.2





#93

Benzo(a)pyrene

Concen: 30.432 ng/ul

RT: 23.663 min Scan# 3911

Delta R.T. 0.006 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51

Instrument :

BNA_P

ClientSampleId :

SLCS249





Tgt Ion:252 Resp: 2310042

Ion Ratio Lower Upper

252 100

253 21.1 17.6 26.4

125 12.0 8.9 13.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 31.314 ng/ul

RT: 26.286 min Scan# 3961

Delta R.T. 0.006 min

Lab File: BP012115.D

Acq: 13 Oct 2022 11:51




Tgt Ion:276 Resp: 3005443


Ion Ratio Lower Upper

276 100

138 27.0 19.1 28.7

277 25.1 19.9 29.9









#95
 Dibenzo(a,h)anthracene
 Concen: 29.878 ng/ul
 RT: 26.309 min Scan# 3965
 Delta R.T. 0.012 min
 Lab File: BP012115.D
 Acq: 13 Oct 2022 11:51

Instrument :
 BNA_P
 ClientSampleId :
 SLCS249

Tgt Ion:278 Resp: 2572623
 Ion Ratio Lower Upper
 278 100
 139 17.9 12.5 18.7
 279 23.7 19.0 28.4



#96
 Benzo(g,h,i)perylene
 Concen: 29.177 ng/ul
 RT: 27.057 min Scan# 4092
 Delta R.T. -0.000 min
 Lab File: BP012115.D
 Acq: 13 Oct 2022 11:51

Tgt Ion:276 Resp: 2524925
 Ion Ratio Lower Upper
 276 100
 138 24.1 16.5 24.7
 277 23.6 19.0 28.4

