

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101124\
 Data File : BP022330.D
 Acq On : 11 Oct 2024 11:42
 Operator : RC/JU
 Sample : PB163830BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK830

Quant Time: Oct 12 01:25:19 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 08 02:14:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.704	152	108590	20.000	ng/u1	0.00
20) Naphthalene-d8	10.463	136	394354	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.310	164	305930	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.104	188	745952	20.000	ng/u1	0.00
79) Chrysene-d12	21.545	240	831830	20.000	ng/u1	0.00
88) Perylene-d12	24.827	264	921195	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.222	96	10358	3.707	ng/uL	0.00
4) Pyridine-d5	3.640	84	151771	19.784	ng/u1	0.00
7) Phenol-d5	6.893	99	186391	20.391	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.040	67	110320	20.534	ng/u1	0.00
11) 2-Chlorophenol-d4	7.240	132	141437	21.806	ng/u1	0.00
15) 4-Methylphenol-d8	8.404	113	147672	20.243	ng/u1	0.00
21) Nitrobenzene-d5	8.846	128	67220	22.169	ng/u1	0.00
24) 2-Nitrophenol-d4	9.557	143	80826	23.024	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.092	165	160293	21.484	ng/u1	0.00
31) 4-Chloroaniline-d4	10.604	131	189398	21.483	ng/u1	0.01
46) Dimethylphthalate-d6	13.710	166	524263	21.569	ng/u1	0.00
49) Acenaphthylene-d8	13.992	160	569644	22.065	ng/u1	0.00
54) 4-Nitrophenol-d4	14.522	143	72705	17.830	ng/u1	0.00
60) Fluorene-d10	15.310	176	461702	21.899	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.439	200	81791	16.672	ng/u1	-0.01
73) Anthracene-d10	17.204	188	814169	23.201	ng/u1	-0.01
81) Pyrene-d10	19.580	212	1044553	23.746	ng/u1	-0.01
92) Benzo(a)pyrene-d12	24.604	264	1134794	23.067	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101124\
Data File : BP022330.D
Acq On : 11 Oct 2024 11:42
Operator : RC/JU
Sample : PB163830BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK830

Quant Time: Oct 12 01:25:19 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 08 02:14:47 2024
Response via : Initial Calibration

