

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101321\
 Data File : BP007400.D
 Acq On : 13 Oct 2021 11:50
 Operator : CG/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Quant Time: Oct 13 13:09:08 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP101321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 13:04:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.969	152	195377	20.000	ng	0.00
21) Naphthalene-d8	10.775	136	757295	20.000	ng	0.00
39) Acenaphthene-d10	14.598	164	489320	20.000	ng	0.00
64) Phenanthrene-d10	17.351	188	1033220	20.000	ng	0.00
76) Chrysene-d12	21.433	240	1024646	20.000	ng	0.00
86) Perylene-d12	23.886	264	1109975	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	481936	40.645	ng	0.00
7) Phenol-d6	7.122	99	645048	40.521	ng	0.00
23) Nitrobenzene-d5	9.122	82	491555	38.229	ng	0.00
42) 2,4,6-Tribromophenol	16.086	330	205395	25.590	ng	0.00
45) 2-Fluorobiphenyl	13.228	172	1325060	37.839	ng	0.00
79) Terphenyl-d14	19.910	244	2084673	35.884	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.422	88	105054	23.030	ng	99
3) Pyridine	3.816	79	288765	22.377	ng	99
4) n-Nitrosodimethylamine	3.728	42	119641	24.797	ng	# 99
6) Aniline	7.287	93	371011	21.010	ng	100
8) 2-Chlorophenol	7.528	128	251289	19.580	ng	99
9) Benzaldehyde	7.099	77	208268	22.608	ng	98
10) Phenol	7.146	94	311592	20.470	ng	99
11) bis(2-Chloroethyl)ether	7.387	93	256412	21.016	ng	99
12) 1,3-Dichlorobenzene	7.857	146	287582	19.873	ng	100
13) 1,4-Dichlorobenzene	8.004	146	293218	20.178	ng	99
14) 1,2-Dichlorobenzene	8.322	146	279830	19.888	ng	98
15) Benzyl Alcohol	8.199	79	237715	23.009	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.499	45	333954	23.789	ng	99
17) 2-Methylphenol	8.404	107	215112	20.472	ng	98
18) Hexachloroethane	9.057	117	102936	20.961	ng	99
19) n-Nitroso-di-n-propyla...	8.775	70	193131	21.395	ng	100
20) 3+4-Methylphenols	8.728	107	293405	19.922	ng	98
22) Acetophenone	8.787	105	384931	21.609	ng	# 99
24) Nitrobenzene	9.163	77	249555	20.570	ng	100
25) Isophorone	9.693	82	523597	21.578	ng	100
26) 2-Nitrophenol	9.881	139	85386	12.599	ng	96
27) 2,4-Dimethylphenol	9.940	122	213775	21.241	ng	99
28) bis(2-Chloroethoxy)met...	10.181	93	342866	21.036	ng	99
29) 2,4-Dichlorophenol	10.410	162	229658	18.974	ng	98
30) 1,2,4-Trichlorobenzene	10.640	180	266769	19.564	ng	98
31) Naphthalene	10.822	128	773285	19.953	ng	100
32) Benzoic acid	10.022	122	100293	12.351	ng	96
33) 4-Chloroaniline	10.922	127	332395	20.727	ng	99
34) Hexachlorobutadiene	11.128	225	168773	20.746	ng	97
35) Caprolactam	11.669	113	43809	11.447	ng	97
36) 4-Chloro-3-methylphenol	12.045	107	258422	22.809	ng	98
37) 2-Methylnaphthalene	12.428	142	547194	22.554	ng	98
38) 1-Methylnaphthalene	12.651	142	531074	22.296	ng	99
40) 1,2,4,5-Tetrachloroben...	12.798	216	293028	18.434	ng	99
41) Hexachlorocyclopentadiene	12.787	237	155468	33.306	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101321\
 Data File : BP007400.D
 Acq On : 13 Oct 2021 11:50
 Operator : CG/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Quant Time: Oct 13 13:09:08 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP101321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 13:04:34 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	13.034	196	176388	16.213	ng	99
44) 2,4,5-Trichlorophenol	13.098	196	174625	14.825	ng	99
46) 1,1'-Biphenyl	13.439	154	725085	19.737	ng	99
47) 2-Chloronaphthalene	13.475	162	554570	19.350	ng	100
48) 2-Nitroaniline	13.669	65	109226	14.719	ng	96
49) Acenaphthylene	14.322	152	887339	20.059	ng	100
50) Dimethylphthalate	14.057	163	719286	19.076	ng	100
51) 2,6-Dinitrotoluene	14.169	165	117464	14.965	ng	99
52) Acenaphthene	14.663	154	545162	20.878	ng	99
53) 3-Nitroaniline	14.486	138	130462	16.205	ng	99
54) 2,4-Dinitrophenol	14.692	184	35872	8.155	ng	# 90
55) Dibenzofuran	14.998	168	878449	19.380	ng	99
56) 4-Nitrophenol	14.786	139	103819	16.443	ng	98
57) 2,4-Dinitrotoluene	14.951	165	144257	13.433	ng	95
58) Fluorene	15.645	166	666790	19.004	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.139	232	154123	14.317	ng	97
60) Diethylphthalate	15.428	149	735859	19.634	ng	100
61) 4-Chlorophenyl-phenyle...	15.645	204	345822	17.692	ng	99
62) 4-Nitroaniline	15.651	138	130775	15.180	ng	99
63) Azobenzene	15.939	77	702882	22.911	ng	99
65) 4,6-Dinitro-2-methylph...	15.716	198	68275	9.592	ng	94
66) n-Nitrosodiphenylamine	15.857	169	616455	18.819	ng	99
67) 4-Bromophenyl-phenylether	16.545	248	216409	16.228	ng	98
68) Hexachlorobenzene	16.657	284	242449	16.568	ng	98
69) Atrazine	16.810	200	222679	18.474	ng	100
70) Pentachlorophenol	16.998	266	124739	15.486	ng	96
71) Phenanthrene	17.392	178	1102561	20.207	ng	99
72) Anthracene	17.486	178	1105050	19.165	ng	100
73) Carbazole	17.751	167	1118575	19.910	ng	100
74) Di-n-butylphthalate	18.316	149	1287847	19.753	ng	100
75) Fluoranthene	19.357	202	1345349	19.119	ng	99
77) Benzidine	19.527	184	540849	18.187	ng	99
78) Pyrene	19.710	202	1400728	19.852	ng	99
80) Butylbenzylphthalate	20.592	149	541860	19.879	ng	98
81) Benzo(a)anthracene	21.416	228	1377433	20.809	ng	100
82) 3,3'-Dichlorobenzidine	21.345	252	453802	19.052	ng	99
83) Chrysene	21.474	228	1294129	20.771	ng	99
84) Bis(2-ethylhexyl)phtha...	21.363	149	838001	22.090	ng	98
85) Di-n-octyl phthalate	22.310	149	1392887	21.533	ng	99
87) Indeno(1,2,3-cd)pyrene	26.451	276	1593672	20.977	ng	100
88) Benzo(b)fluoranthene	23.139	252	1341893	18.568	ng	99
89) Benzo(k)fluoranthene	23.186	252	1326848	18.547	ng	99
90) Benzo(a)pyrene	23.774	252	1142081	20.353	ng	100
91) Dibenzo(a,h)anthracene	26.480	278	1321802	20.782	ng	98
92) Benzo(g,h,i)perylene	27.245	276	1321725	22.213	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101321\
 Data File : BP007400.D
 Acq On : 13 Oct 2021 11:50
 Operator : CG/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Quant Time: Oct 13 13:09:08 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP101321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 13:04:34 2021
 Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.969 min Scan# 830
 Delta R.T. 0.000 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion:	152	Resp:	195377
Ion	Ratio	Lower	Upper
152	100		
150	152.2	124.1	186.1
115	58.9	48.1	72.1



1,4-Dioxane
 Concen: 23.030 ng
 RT: 3.422 min Scan# 57
 Delta R.T. 0.000 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50

Tgt Ion:	88	Resp:	105054
Ion	Ratio	Lower	Upper
88	100		
58	72.4	58.1	87.1
43	29.6	22.2	33.4









Pyridine
 Concen: 22.377 ng
 RT: 3.816 min Scan# 124
 Delta R.T. 0.000 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC020


Tgt Ion:	79	Resp:	288765
Ion	Ratio	Lower	Upper
79	100		
52	68.6	55.7	83.5
51	35.0	28.5	42.7



n-Nitrosodimethylamine
 Concen: 24.797 ng
 RT: 3.728 min Scan# 109
 Delta R.T. 0.000 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50

Tgt Ion:	42	Resp:	119641
Ion	Ratio	Lower	Upper
42	100		
74	133.9	107.0	160.4
44	12.0	7.4	11.2#









2-Fluorophenol
 Concen: 40.645 ng
 RT: 5.528 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion:	112	Resp:	481936
Ion	Ratio	Lower	Upper
112	100		
64	58.5	46.6	70.0
63	30.5	25.0	37.6



Aniline
 Concen: 21.010 ng
 RT: 7.287 min Scan# 714
 Delta R.T. 0.000 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50

Tgt Ion:	93	Resp:	371011
Ion	Ratio	Lower	Upper
93	100		
66	38.3	30.5	45.7
65	19.7	15.8	23.8









Phenol-d6
 Concen: 40.521 ng
 RT: 7.122 min Scan# 686
 Delta R.T. 0.000 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion:	99	Resp:	645048
Ion	Ratio	Lower	Upper
99	100		
42	17.0	14.9	22.3
71	35.0	29.2	43.8



2-Chlorophenol
 Concen: 19.580 ng
 RT: 7.528 min Scan# 755
 Delta R.T. 0.000 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50

Tgt Ion:	128	Resp:	251289
Ion	Ratio	Lower	Upper
128	100		
130	31.6	12.0	52.0
64	46.1	27.0	67.0









Scan 682 (7.098 min): BP007401.D\data.ms (-673)
 Benzaldehyde
 Concen: 22.608 ng
 RT: 7.099 min Scan# 682
 Delta R.T. 0.000 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Tgt Ion:	77	Resp:	208268
Ion	Ratio	Lower	Upper
77	100		
105	98.1	77.5	117.5
106	95.9	72.1	112.1



Scan 691 (7.151 min): BP007401.D\data.ms (-679)
 Phenol
 Concen: 20.470 ng
 RT: 7.146 min Scan# 690
 Delta R.T. -0.006 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50


Tgt Ion:	94	Resp:	311592
Ion	Ratio	Lower	Upper
94	100		
65	27.9	8.2	48.2
66	37.9	18.8	58.8






Abundance Scan 731 (7.387 min): BP007401.D\data.ms (-724) #11

bis(2-Chloroethyl)ether
Concen: 21.016 ng
RT: 7.387 min Scan# 731
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020




Tgt Ion:	93	Resp:	256412
Ion	Ratio	Lower	Upper
93	100		
63	74.7	55.7	95.7
95	32.8	12.7	52.7







Abundance Scan 811 (7.857 min): BP007401.D\data.ms (-802) #12

1,3-Dichlorobenzene
Concen: 19.873 ng
RT: 7.857 min Scan# 811
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50



Tgt Ion:	146	Resp:	287582
Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.4	77.0
75	32.0	26.1	39.1









1,4-Dichlorobenzene
 Concen: 20.178 ng
 RT: 8.004 min Scan# 836
 Delta R.T. 0.000 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion:	146	Resp:	293218
Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.0	76.6
111	41.6	32.8	49.2



1,2-Dichlorobenzene
 Concen: 19.888 ng
 RT: 8.322 min Scan# 890
 Delta R.T. 0.000 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50

Tgt Ion:	146	Resp:	279830
Ion	Ratio	Lower	Upper
146	100		
148	62.7	51.5	77.3
111	43.4	35.1	52.7









Benzyl Alcohol
Concen: 23.009 ng
RT: 8.199 min Scan# 869
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	79	Resp:	237715
Ion	Ratio	Lower	Upper
79	100		
108	77.2	59.9	89.9
77	63.9	51.1	76.7



2,2'-oxybis(1-Chloropropane)
Concen: 23.789 ng
RT: 8.499 min Scan# 920
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Tgt Ion:	45	Resp:	333954
Ion	Ratio	Lower	Upper
45	100		
77	15.2	0.0	34.9
79	10.7	0.0	31.7









2-Methylphenol
Concen: 20.472 ng
RT: 8.404 min Scan# 904
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50




Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion	Ratio	Lower	Upper
107	100		
108	109.7	89.2	133.8
77	45.4	38.5	57.7
79	48.0	38.4	57.6



Hexachloroethane
Concen: 20.961 ng
RT: 9.057 min Scan# 1015
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50


Tgt Ion	Ratio	Lower	Upper
117	100		
119	95.8	77.6	116.4
201	103.1	82.9	124.3





Abundance Scan 967 (8.775 min): BP007401.D\data.ms (-960) #19

n-Nitroso-di-n-propylamine
Concen: 21.395 ng
RT: 8.775 min Scan# 967
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50


Instrument :
BNA_P
ClientSampleId :
SSTDIC020



Tgt Ion:	70	Resp:	193131
Ion	Ratio	Lower	Upper
70	100		
42	57.9	46.2	69.2
101	10.3	7.7	11.5
130	20.3	16.2	24.4




Abundance Scan 967 (8.775 min): BP007400.D\data.ms (-916) (-)





Abundance Scan 960 (8.734 min): BP007401.D\data.ms (-950) #20


3+4-Methylphenols
Concen: 19.922 ng
RT: 8.728 min Scan# 959
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50



Tgt Ion:	107	Resp:	293405
Ion	Ratio	Lower	Upper
107	100		
108	88.0	69.2	109.2
77	34.3	13.8	53.8
79	28.6	10.2	50.2




Abundance Scan 959 (8.728 min): BP007400.D\data.ms (-909) (-)






Abundance Scan 1307 (10.775 min): BP007401.D\data.ms (-1256) (-)

Naphthalene-d8
Concen: 20.000 ng
RT: 10.775 min Scan# 1307
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020






Tgt Ion:	136	Resp:	757295
Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.6	13.0
54	7.7	6.2	9.4
68	5.8	4.7	7.1



Abundance Scan 969 (8.787 min): BP007401.D\data.ms (-962) (-)

Acetophenone
Concen: 21.609 ng
RT: 8.787 min Scan# 969
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50


Tgt Ion:	105	Resp:	384931
Ion	Ratio	Lower	Upper
105	100		
71	3.6	4.5	6.7#
51	27.5	21.8	32.8
120	22.4	17.6	26.4






Abundance Scan 1026 (9.122 min): BP007401.D\data.ms (-10193)

Nitrobenzene-d5
Concen: 38.229 ng
RT: 9.122 min Scan# 1026
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020




Tgt Ion	Ratio	Lower	Upper
82	100		
128	41.1	32.6	48.8
54	49.9	40.4	60.6






Abundance Scan 1033 (9.163 min): BP007401.D\data.ms (-10194)


Nitrobenzene
Concen: 20.570 ng
RT: 9.163 min Scan# 1033
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50



Tgt Ion	Ratio	Lower	Upper
77	100		
123	43.8	35.3	52.9
65	13.4	10.9	16.3






Abundance Scan 1124 (9.698 min): BP007401.D\data.ms (-1125-)




Isophorone
Concen: 21.578 ng
RT: 9.693 min Scan# 1123
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	82	Resp:	523597
Ion	Ratio	Lower	Upper
82	100		
95	7.8	6.2	9.2
138	17.7	14.0	21.0






Abundance Scan 1155 (9.881 min): BP007401.D\data.ms (-1142-)



2-Nitrophenol
Concen: 12.599 ng
RT: 9.881 min Scan# 1155
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50


Tgt Ion:	139	Resp:	85386
Ion	Ratio	Lower	Upper
139	100		
109	30.8	26.2	39.2
65	47.9	40.3	60.5






Abundance Scan 1166 (9.945 min): BP007401.D\data.ms (-1155) (-)

2,4-Dimethylphenol
Concen: 21.241 ng
RT: 9.940 min Scan# 1165
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020




Tgt Ion:	122	Resp:	213775
Ion	Ratio	Lower	Upper
122	100		
107	117.4	93.8	140.6
121	60.0	47.0	70.4






Abundance Scan 1206 (10.181 min): BP007401.D\data.ms (-1155) (-)

bis(2-Chloroethoxy)methane
Concen: 21.036 ng
RT: 10.181 min Scan# 1206
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50




Tgt Ion:	93	Resp:	342866
Ion	Ratio	Lower	Upper
93	100		
95	31.8	26.1	39.1
123	11.1	8.5	12.7





Abundance Scan 1246 (10.416 min): BP007401.D\data.ms (-1229) (-)

2,4-Dichlorophenol
Concen: 18.974 ng
RT: 10.410 min Scan# 1245
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50


Instrument :
BNA_P
ClientSampleId :
SSTDIC020



Tgt Ion	Ratio	Lower	Upper
162	100		
164	61.8	44.0	84.0
98	36.5	16.3	56.3




Abundance Scan 1245 (10.410 min): BP007400.D\data.ms (-1195) (-)





Abundance Scan 1284 (10.639 min): BP007401.D\data.ms (-1230) (-)


1,2,4-Trichlorobenzene
Concen: 19.564 ng
RT: 10.640 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50




Tgt Ion	Ratio	Lower	Upper
180	100		
182	95.0	78.4	117.6
145	29.9	23.8	35.8



Abundance Scan 1284 (10.640 min): BP007400.D\data.ms (-1233) (-)









Naphthalene
 Concen: 19.953 ng
 RT: 10.822 min Scan# 1315
 Delta R.T. 0.000 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion:	128	Resp:	773285
Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.9	13.3
127	13.0	10.2	15.4



Benzoic acid
 Concen: 12.351 ng
 RT: 10.022 min Scan# 1179
 Delta R.T. -0.041 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50

Tgt Ion:	122	Resp:	100293
Ion	Ratio	Lower	Upper
122	100		
105	121.3	105.9	145.9
77	90.6	74.6	114.6









Scan 1332 (10.922 min): BP007401.D\data.ms (-128) (-)
 4-Chloroaniline
 Concen: 20.727 ng
 RT: 10.922 min Scan# 1332
 Delta R.T. 0.000 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion	Ratio	Lower	Upper
127	100		
129	31.9	25.1	37.7
65	31.2	25.5	38.3
92	20.1	16.5	24.7



Scan 1367 (11.128 min): BP007401.D\data.ms (-131) (-)
 Hexachlorobutadiene
 Concen: 20.746 ng
 RT: 11.128 min Scan# 1367
 Delta R.T. 0.000 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50

Tgt Ion	Ratio	Lower	Upper
225	100		
223	64.4	49.4	74.0
227	62.3	51.0	76.6









Caprolactam
Concen: 11.447 ng
RT: 11.669 min Scan# 1459
Delta R.T. -0.017 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	113	Resp:	43809
Ion	Ratio	Lower	Upper
113	100		
55	165.0	143.9	183.9
56	129.3	101.6	141.6



4-Chloro-3-methylphenol
Concen: 22.809 ng
RT: 12.045 min Scan# 1523
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Tgt Ion:	107	Resp:	258422
Ion	Ratio	Lower	Upper
107	100		
144	24.5	20.5	30.7
142	76.9	63.0	94.4









2-Methylnaphthalene
Concen: 22.554 ng
RT: 12.428 min Scan# 1588
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50




Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	142	Resp:	547194
Ion	Ratio	Lower	Upper
142	100		
141	87.2	71.4	107.0
115	36.3	29.5	44.3




1-Methylnaphthalene
Concen: 22.296 ng
RT: 12.651 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Tgt Ion:	142	Resp:	531074
Ion	Ratio	Lower	Upper
142	100		
141	91.8	74.4	111.6
116	3.8	3.4	5.0






Abundance Scan 1958 (14.604 min): BP007401.D\data.ms (-1907) (-)




Acenaphthene-d10
Concen: 20.000 ng
RT: 14.598 min Scan# 1957
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleID :
SSTDIC020

Tgt Ion:	164	Resp:	489320
Ion Ratio	Lower	Upper	
164	100		
162	100.5	81.1	121.7
160	44.6	36.0	54.0







Abundance Scan 1651 (12.798 min): BP007401.D\data.ms (-1600) (-)



1,2,4,5-Tetrachlorobenzene
Concen: 18.434 ng
RT: 12.798 min Scan# 1651
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Tgt Ion:	216	Resp:	293028
Ion Ratio	Lower	Upper	
216	100		
214	78.6	62.2	93.4
179	21.2	17.1	25.7
108	19.0	15.4	23.0









Hexachlorocyclopentadiene
Concen: 33.306 ng
RT: 12.787 min Scan# 1649
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50




Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	237	Resp:	155468
Ion	Ratio	Lower	Upper
237	100		
235	61.8	41.2	81.2
272	11.5	0.0	31.9




2,4,6-Tribromophenol
Concen: 25.590 ng
RT: 16.086 min Scan# 2210
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Tgt Ion:	330	Resp:	205395
Ion	Ratio	Lower	Upper
330	100		
332	98.2	78.9	118.3
141	37.8	30.7	46.1






Abundance Scan 1691 (13.033 min): BP007401.D\data.ms (-1640) (-)




2,4,6-Trichlorophenol
Concen: 16.213 ng
RT: 13.034 min Scan# 1691
Delta R.T. 0.001 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	196	Resp:	176388
Ion	Ratio	Lower	Upper
196	100		
198	95.9	76.4	114.6
200	30.2	25.0	37.4






Abundance Scan 1702 (13.098 min): BP007401.D\data.ms (-1651) (-)




2,4,5-Trichlorophenol
Concen: 14.825 ng
RT: 13.098 min Scan# 1702
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Tgt Ion:	196	Resp:	174625
Ion	Ratio	Lower	Upper
196	100		
198	96.6	77.1	115.7
97	51.1	40.9	61.3
132	27.8	21.2	31.8






Abundance Scan 1725 (13.233 min): BP007401.D\data.ms (-1715) (-)




2-Fluorobiphenyl
Concen: 37.839 ng
RT: 13.228 min Scan# 1724
Delta R.T. -0.005 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:172 Resp: 1325060
Ion Ratio Lower Upper
172 100
171 34.7 28.2 42.2
170 23.2 19.0 28.4







Abundance Scan 1760 (13.439 min): BP007401.D\data.ms (-1756) (-)



1,1'-Biphenyl
Concen: 19.737 ng
RT: 13.439 min Scan# 1760
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Tgt Ion:154 Resp: 725085
Ion Ratio Lower Upper
154 100
153 40.1 19.9 59.9
76 12.9 0.0 32.4









2-Chloronaphthalene
Concen: 19.350 ng
RT: 13.475 min Scan# 1766
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50




Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	162	Resp:	554570
Ion	Ratio	Lower	Upper
162	100		
127	38.8	31.0	46.6
164	32.6	26.4	39.6



2-Nitroaniline
Concen: 14.719 ng
RT: 13.669 min Scan# 1799
Delta R.T. -0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50


Tgt Ion:	65	Resp:	109226
Ion	Ratio	Lower	Upper
65	100		
92	73.9	55.9	83.9
138	111.5	86.3	129.5






Abundance Scan 1910 (14.322 min): BP007401.D\data.ms (-1909) (-)

Acenaphthylene
Concen: 20.059 ng
RT: 14.322 min Scan# 1910
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDICC020




Tgt Ion:	152	Resp:	887339
Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.8	23.6
153	12.7	10.2	15.2






Abundance Scan 1866 (14.063 min): BP007401.D\data.ms (-1859) (-)

Dimethylphthalate
Concen: 19.076 ng
RT: 14.057 min Scan# 1865
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50




Tgt Ion:	163	Resp:	719286
Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.3	4.9
164	10.2	8.2	12.2





Abundance Scan 1884 (14.169 min): BP007401.D\data.ms (-1852) (-)

2,6-Dinitrotoluene
Concen: 14.965 ng
RT: 14.169 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50


Instrument :
BNA_P
ClientSampleId :
SSTDIC020



Tgt Ion:	165	Resp:	117464
Ion	Ratio	Lower	Upper
165	100		
63	56.6	44.2	66.2
89	59.0	46.6	70.0




Abundance Scan 1884 (14.169 min): BP007400.D\data.ms (-1833) (-)





Abundance Scan 1968 (14.663 min): BP007401.D\data.ms (-1952) (-)


Acenaphthene
Concen: 20.878 ng
RT: 14.663 min Scan# 1968
Delta R.T. -0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50




Tgt Ion:	154	Resp:	545162
Ion	Ratio	Lower	Upper
154	100		
153	114.4	90.4	135.6
152	54.0	42.3	63.5



Abundance Scan 1968 (14.663 min): BP007400.D\data.ms (-1917) (-)









3-Nitroaniline
Concen: 16.205 ng
RT: 14.486 min Scan# 1938
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50




Instrument :
BNA_P
ClientSampleId :
SSTDICC020


Tgt Ion:	138	Resp:	130462
Ion	Ratio	Lower	Upper
138	100		
108	18.0	12.3	18.5
92	117.7	94.9	142.3



2,4-Dinitrophenol
Concen: 8.155 ng
RT: 14.692 min Scan# 1973
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Tgt Ion:	184	Resp:	35872
Ion	Ratio	Lower	Upper
184	100		
63	55.9	44.0	66.0
154	74.6	48.9	73.3#









Dibenzofuran
Concen: 19.380 ng
RT: 14.998 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50




Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	168	Resp:	878449
Ion	Ratio	Lower	Upper
168	100		
139	37.9	30.8	46.2
169	13.2	11.0	16.6



4-Nitrophenol
Concen: 16.443 ng
RT: 14.786 min Scan# 1989
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50


Tgt Ion:	139	Resp:	103819
Ion	Ratio	Lower	Upper
139	100		
109	81.5	60.9	100.9
65	99.4	76.5	116.5






Abundance Scan 2017 (14.951 min): BP007401.D\data.ms (-2057) (-)

2,4-Dinitrotoluene
Concen: 13.433 ng
RT: 14.951 min Scan# 2017
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020




Tgt Ion:	165	Resp:	144257
Ion	Ratio	Lower	Upper
165	100		
63	42.6	37.4	56.0
89	73.5	61.8	92.6
182	2.5	2.1	3.1






Abundance Scan 2136 (15.651 min): BP007401.D\data.ms (-2158) (-)


Fluorene
Concen: 19.004 ng
RT: 15.645 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50



Tgt Ion:	166	Resp:	666790
Ion	Ratio	Lower	Upper
166	100		
165	99.0	78.9	118.3
167	13.4	10.9	16.3





Abundance Scan 2063 (15.221 min): BP007401.D\data.ms (-2059) (-)




2,3,4,6-Tetrachlorophenol
Concen: 14.317 ng
RT: 15.139 min Scan# 2049
Delta R.T. -0.082 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion	Ratio	Lower	Upper
232	100		
131	40.0	30.3	45.5
130	2.5	1.8	2.8
166	25.9	19.4	29.0



Abundance Scan 2049 (15.139 min): BP007400.D\data.ms (-2012) (-)





Abundance Scan 2098 (15.427 min): BP007401.D\data.ms (-2059) (-)





Diethylphthalate
Concen: 19.634 ng
RT: 15.428 min Scan# 2098
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Tgt Ion	Ratio	Lower	Upper
149	100		
177	22.0	17.8	26.8
150	12.7	10.2	15.2



Abundance Scan 2098 (15.428 min): BP007400.D\data.ms (-2047) (-)









4-Chlorophenyl-phenylether
Concen: 17.692 ng
RT: 15.645 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50




Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	204	Resp:	345822
Ion	Ratio	Lower	Upper
204	100		
206	32.6	25.8	38.8
141	60.0	48.8	73.2



4-Nitroaniline
Concen: 15.180 ng
RT: 15.651 min Scan# 2136
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50


Tgt Ion:	138	Resp:	130775
Ion	Ratio	Lower	Upper
138	100		
92	50.4	30.7	70.7
108	90.3	71.2	111.2






Abundance Scan 2185 (15.939 min): BP007401.D\data.ms (-2185) (-)

Azobenzene
Concen: 22.911 ng
RT: 15.939 min Scan# 2185
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020






Tgt Ion:	77	Resp:	702882
Ion	Ratio	Lower	Upper
77	100		
182	26.1	6.5	46.5
105	21.7	1.3	41.3
51	29.3	10.5	50.5



Abundance Scan 2425 (17.351 min): BP007401.D\data.ms (-2425) (-)

Phenanthrene-d10
Concen: 20.000 ng
RT: 17.351 min Scan# 2425
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50


Tgt Ion:	188	Resp:	1033220
Ion	Ratio	Lower	Upper
188	100		
94	8.9	7.1	10.7
80	10.0	7.8	11.6





Abundance Scan 2147 (15.716 min): BP007401.D\data.ms (-2168) (-)

4,6-Dinitro-2-methylphenol
Concen: 9.592 ng
RT: 15.716 min Scan# 2147
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50


Instrument :
BNA_P
ClientSampleId :
SSTDIC020



Tgt Ion:	198	Resp:	68275
Ion	Ratio	Lower	Upper
198	100		
51	38.9	23.8	63.8
105	42.3	24.6	64.6




Abundance Scan 2147 (15.716 min): BP007400.D\data.ms (-2096) (-)





Abundance Scan 2171 (15.857 min): BP007401.D\data.ms (-2168) (-)

n-Nitrosodiphenylamine
Concen: 18.819 ng
RT: 15.857 min Scan# 2171
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Tgt Ion:	169	Resp:	616455
Ion	Ratio	Lower	Upper
169	100		
168	66.5	52.7	79.1
167	36.3	29.0	43.6




Abundance Scan 2171 (15.857 min): BP007400.D\data.ms (-2120) (-)






Abundance Scan 2288 (16.545 min): BP007401.D\data.ms (-2288) (-)

4-Bromophenyl-phenylether
Concen: 16.228 ng
RT: 16.545 min Scan# 2288
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020




Tgt Ion:	248	Resp:	216409
Ion Ratio	Lower	Upper	
248	100		
250	99.0	78.2	117.2
141	73.7	60.8	91.2






Abundance Scan 2307 (16.657 min): BP007401.D\data.ms (-2307) (-)

Hexachlorobenzene
Concen: 16.568 ng
RT: 16.657 min Scan# 2307
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50




Tgt Ion:	284	Resp:	242449
Ion Ratio	Lower	Upper	
284	100		
142	36.2	30.6	46.0
249	33.3	26.6	40.0






Abundance Scan 2333 (16.810 min): BP007401.D\data.ms (-2359) (-)

Atrazine
Concen: 18.474 ng
RT: 16.810 min Scan# 2333
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020




Tgt Ion	Ratio	Lower	Upper
200	100		
173	26.1	6.2	46.2
215	48.0	28.2	68.2






Abundance Scan 2365 (16.998 min): BP007401.D\data.ms (-2370) (-)


Pentachlorophenol
Concen: 15.486 ng
RT: 16.998 min Scan# 2365
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50



Tgt Ion	Ratio	Lower	Upper
266	100		
268	63.5	49.3	73.9
264	64.6	48.7	73.1





Abundance Scan 2432 (17.392 min): BP007401.D\data.ms (-2470) (-)




Phenanthrene
Concen: 20.207 ng
RT: 17.392 min Scan# 2432
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	178	Resp:	1102561
Ion	Ratio	Lower	Upper
178	100		
176	19.8	15.4	23.0
179	15.3	12.3	18.5



Abundance Scan 2432 (17.392 min): BP007400.D\data.ms (-2381) (-)





Abundance Scan 2448 (17.486 min): BP007401.D\data.ms (-2470) (-)





Anthracene
Concen: 19.165 ng
RT: 17.486 min Scan# 2448
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Tgt Ion:	178	Resp:	1105050
Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.0	22.6
179	15.0	12.2	18.4



Abundance Scan 2448 (17.486 min): BP007400.D\data.ms (-2397) (-)









Carbazole
Concen: 19.910 ng
RT: 17.751 min Scan# 2493
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50




Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	167	Resp:	1118575
Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.3	25.9
139	12.9	10.4	15.6




Di-n-butylphthalate
Concen: 19.753 ng
RT: 18.316 min Scan# 2589
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Tgt Ion:	149	Resp:	1287847
Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.6	11.4
104	5.3	4.2	6.4






Abundance Scan 2766 (19.357 min): BP007401.D\data.ms (-2766) (-)




Fluoranthene
Concen: 19.119 ng
RT: 19.357 min Scan# 2766
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:202 Resp: 1345349
Ion Ratio Lower Upper
202 100
101 10.4 0.0 30.4
203 17.1 0.0 37.6






Abundance Scan 3120 (21.439 min): BP007401.D\data.ms (-3109) (-)



Chrysene-d12
Concen: 20.000 ng
RT: 21.433 min Scan# 3119
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50


Tgt Ion:240 Resp: 1024646
Ion Ratio Lower Upper
240 100
120 10.5 7.8 11.8
236 26.1 20.6 30.8






Abundance Scan 2795 (19.527 min): BP007401.D\data.ms (-2777) (-)

Benzidine
Concen: 18.187 ng
RT: 19.527 min Scan# 2795
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020




Tgt Ion	Ratio	Lower	Upper
184	100		
185	14.0	11.2	16.8
183	12.1	9.1	13.7







Abundance Scan 2826 (19.709 min): BP007401.D\data.ms (-2774) (-)

Pyrene
Concen: 19.852 ng
RT: 19.710 min Scan# 2826
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50



Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.6	16.9	25.3
203	17.3	13.8	20.6









Terphenyl-d14
 Concen: 35.884 ng
 RT: 19.910 min Scan# 2860
 Delta R.T. 0.000 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Tgt Ion:	244	Resp:	2084673
Ion	Ratio	Lower	Upper
244	100		
212	7.5	6.2	9.2
122	10.0	7.8	11.8




Butylbenzylphthalate
 Concen: 19.879 ng
 RT: 20.592 min Scan# 2976
 Delta R.T. 0.000 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50

Tgt Ion:	149	Resp:	541860
Ion	Ratio	Lower	Upper
149	100		
91	69.0	56.6	85.0
206	21.7	17.4	26.2






Abundance Scan 3117 (21.421 min): BP007401.D\data.ms (-3117) (-)




Benzo(a)anthracene
Concen: 20.809 ng
RT: 21.416 min Scan# 3116
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:228 Resp: 1377433
Ion Ratio Lower Upper
228 100
226 27.3 21.8 32.6
229 19.8 15.7 23.5







Abundance Scan 3105 (21.351 min): BP007401.D\data.ms (-3105) (-)



3,3'-Dichlorobenzidine
Concen: 19.052 ng
RT: 21.345 min Scan# 3104
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Tgt Ion:252 Resp: 453802
Ion Ratio Lower Upper
252 100
254 64.4 51.5 77.3
126 12.1 8.5 12.7









#28 (-)
Chrysene
Concen: 20.771 ng
RT: 21.474 min Scan# 3126
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	228	Resp:	1294129
Ion	Ratio	Lower	Upper
228	100		
226	30.1	24.6	37.0
229	19.5	15.7	23.5



#37 (-)
Bis(2-ethylhexyl)phthalate
Concen: 22.090 ng
RT: 21.363 min Scan# 3107
Delta R.T. 0.000 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Tgt Ion:	149	Resp:	838001
Ion	Ratio	Lower	Upper
149	100		
167	28.2	21.6	32.4
279	4.4	3.3	4.9









Scan 3269 (22.315 min): BP007401.D\data.ms (-3269) (-)
 Di-n-octyl phthalate
 Concen: 21.533 ng
 RT: 22.310 min Scan# 3268
 Delta R.T. -0.006 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC020

Tgt Ion:	149	Resp:	1392887
Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	9.5	7.3	10.9




Scan 3537 (23.892 min): BP007401.D\data.ms (-3537) (-)
 Perylene-d12
 Concen: 20.000 ng
 RT: 23.886 min Scan# 3536
 Delta R.T. -0.006 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50

Tgt Ion:	264	Resp:	1109975
Ion	Ratio	Lower	Upper
264	100		
260	24.5	19.1	28.7
265	22.0	16.7	25.1






Abundance Scan 3975 (26.468 min): BP007401.D\data.ms (-3957) (-)




Indeno(1,2,3-cd)pyrene
Concen: 20.977 ng
RT: 26.451 min Scan# 3972
Delta R.T. -0.017 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	276	Resp:	1593672
Ion	Ratio	Lower	Upper
276	100		
138	24.5	19.4	29.0
277	25.2	20.2	30.2







Abundance Scan 3410 (23.145 min): BP007401.D\data.ms (-3408) (-)



Benzo(b)fluoranthene
Concen: 18.568 ng
RT: 23.139 min Scan# 3409
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Tgt Ion:	252	Resp:	1341893
Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.4	26.2
125	9.1	7.4	11.0









Benzo(k)fluoranthene
 Concen: 18.547 ng
 RT: 23.186 min Scan# 3417
 Delta R.T. -0.006 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Tgt Ion:	252	Resp:	1326848
Ion	Ratio	Lower	Upper
252	100		
253	21.3	17.6	26.4
125	9.8	7.4	11.0




Benzo(a)pyrene
 Concen: 20.353 ng
 RT: 23.774 min Scan# 3517
 Delta R.T. -0.006 min
 Lab File: BP007400.D
 Acq: 13 Oct 2021 11:50

Tgt Ion:	252	Resp:	1142081
Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	10.6	8.4	12.6






Abundance Scan 3979 (26.491 min): BP007401.D\data.ms (-3998) (-)




Dibenzo(a,h)anthracene
Concen: 20.782 ng
RT: 26.480 min Scan# 3977
Delta R.T. -0.012 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	278	Resp:	1321802
Ion	Ratio	Lower	Upper
278	100		
139	15.9	11.8	17.6
279	22.8	18.6	27.8



Abundance Scan 4108 (27.250 min): BP007401.D\data.ms (-4098) (-)



Benzo(g,h,i)perylene
Concen: 22.213 ng
RT: 27.245 min Scan# 4107
Delta R.T. -0.006 min
Lab File: BP007400.D
Acq: 13 Oct 2021 11:50

Tgt Ion:	276	Resp:	1321725
Ion	Ratio	Lower	Upper
276	100		
277	22.7	19.4	29.0
138	20.0	16.2	24.4

