

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101321\
 Data File : BP007401.D
 Acq On : 13 Oct 2021 12:24
 Operator : CG/JU
 Sample : SSTDICC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC040

Quant Time: Oct 13 13:07:17 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP101321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 13:04:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.969	152	208283	20.000	ng	0.00
21) Naphthalene-d8	10.775	136	803518	20.000	ng	0.00
39) Acenaphthene-d10	14.604	164	505498	20.000	ng	0.00
64) Phenanthrene-d10	17.351	188	1063382	20.000	ng	0.00
76) Chrysene-d12	21.439	240	1019845	20.000	ng	0.00
86) Perylene-d12	23.892	264	1149124	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	952549	75.357	ng	0.00
7) Phenol-d6	7.122	99	1285753	75.765	ng	0.00
23) Nitrobenzene-d5	9.122	82	1052306	77.131	ng	0.00
42) 2,4,6-Tribromophenol	16.086	330	409910	49.435	ng	0.00
45) 2-Fluorobiphenyl	13.233	172	2473605	68.376	ng	0.00
79) Terphenyl-d14	19.909	244	3598985	62.243	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.422	88	209455	43.072	ng	100
3) Pyridine	3.816	79	569192	41.375	ng	100
4) n-Nitrosodimethylamine	3.728	42	248296	48.273	ng	100
6) Aniline	7.287	93	738220	39.214	ng	100
8) 2-Chlorophenol	7.528	128	489792	35.799	ng	100
9) Benzaldehyde	7.098	77	377458	38.434	ng	100
10) Phenol	7.151	94	626223	38.591	ng	100
11) bis(2-Chloroethyl)ether	7.387	93	509748	39.191	ng	100
12) 1,3-Dichlorobenzene	7.857	146	572016	37.078	ng	100
13) 1,4-Dichlorobenzene	8.004	146	576047	37.185	ng	100
14) 1,2-Dichlorobenzene	8.322	146	544918	36.328	ng	100
15) Benzyl Alcohol	8.198	79	481828	43.748	ng	100
16) 2,2'-oxybis(1-Chloropr...	8.504	45	667409	44.597	ng	100
17) 2-Methylphenol	8.404	107	424284	37.876	ng	100
18) Hexachloroethane	9.057	117	208770	39.878	ng	100
19) n-Nitroso-di-n-propyla...	8.775	70	372952	38.756	ng	100
20) 3+4-Methylphenols	8.734	107	585846	37.315	ng	100
22) Acetophenone	8.787	105	748612	39.608	ng	100
24) Nitrobenzene	9.163	77	523409	40.661	ng	100
25) Isophorone	9.698	82	1034032	40.163	ng	100
26) 2-Nitrophenol	9.881	139	201751	28.056	ng	100
27) 2,4-Dimethylphenol	9.945	122	420774	39.403	ng	100
28) bis(2-Chloroethoxy)met...	10.181	93	674568	39.006	ng	100
29) 2,4-Dichlorophenol	10.416	162	468768	36.502	ng	100
30) 1,2,4-Trichlorobenzene	10.639	180	531602	36.744	ng	100
31) Naphthalene	10.822	128	1506577	36.638	ng	100
32) Benzoic acid	10.063	122	244582	28.387	ng	100
33) 4-Chloroaniline	10.922	127	653617	38.413	ng	100
34) Hexachlorobutadiene	11.128	225	336708	39.008	ng	100
35) Caprolactam	11.686	113	92563	22.795	ng	100
36) 4-Chloro-3-methylphenol	12.045	107	508597	42.308	ng	100
37) 2-Methylnaphthalene	12.433	142	1062546	41.277	ng	100
38) 1-Methylnaphthalene	12.651	142	1031887	40.830	ng	100
40) 1,2,4,5-Tetrachloroben...	12.798	216	565351	34.427	ng	100
41) Hexachlorocyclopentadiene	12.786	237	325418	67.484	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101321\
 Data File : BP007401.D
 Acq On : 13 Oct 2021 12:24
 Operator : CG/JU
 Sample : SSTDICC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC040

Quant Time: Oct 13 13:07:17 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP101321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 13:04:34 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	13.033	196	347084	30.882	ng	100
44) 2,4,5-Trichlorophenol	13.098	196	338369	27.806	ng	100
46) 1,1'-Biphenyl	13.439	154	1378551	36.324	ng	100
47) 2-Chloronaphthalene	13.480	162	1060853	35.831	ng	100
48) 2-Nitroaniline	13.669	65	256232	33.425	ng	100
49) Acenaphthylene	14.322	152	1699670	37.192	ng	100
50) Dimethylphthalate	14.063	163	1387720	35.626	ng	100
51) 2,6-Dinitrotoluene	14.169	165	262094	32.322	ng	100
52) Acenaphthene	14.663	154	1040255	38.563	ng	100
53) 3-Nitroaniline	14.492	138	286842	34.488	ng	100
54) 2,4-Dinitrophenol	14.698	184	83270	18.325	ng	100
55) Dibenzofuran	14.998	168	1660529	35.462	ng	100
56) 4-Nitrophenol	14.792	139	231417	35.480	ng	100
57) 2,4-Dinitrotoluene	14.951	165	334996	30.197	ng	100
58) Fluorene	15.651	166	1205074	33.246	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.139	232	327396	29.439	ng	97
60) Diethylphthalate	15.427	149	1398687	36.125	ng	100
61) 4-Chlorophenyl-phenyle...	15.645	204	630653	31.231	ng	100
62) 4-Nitroaniline	15.657	138	269158	30.243	ng	100
63) Azobenzene	15.939	77	1329746	41.957	ng	100
65) 4,6-Dinitro-2-methylph...	15.716	198	163263	22.287	ng	100
66) n-Nitrosodiphenylamine	15.857	169	1144070	33.935	ng	100
67) 4-Bromophenyl-phenylether	16.545	248	408100	29.735	ng	100
68) Hexachlorobenzene	16.657	284	449977	29.877	ng	100
69) Atrazine	16.810	200	424067	34.184	ng	100
70) Pentachlorophenol	16.998	266	272572	32.879	ng	100
71) Phenanthrene	17.392	178	2068143	36.829	ng	100
72) Anthracene	17.486	178	2052164	34.582	ng	100
73) Carbazole	17.751	167	2053633	35.517	ng	100
74) Di-n-butylphthalate	18.315	149	2426912	36.168	ng	100
75) Fluoranthene	19.357	202	2493001	34.424	ng	100
77) Benzidine	19.527	184	1199205	40.516	ng	100
78) Pyrene	19.709	202	2564059	36.510	ng	100
80) Butylbenzylphthalate	20.592	149	1077856	39.729	ng	100
81) Benzo(a)anthracene	21.421	228	2549954	38.704	ng	100
82) 3,3'-Dichlorobenzidine	21.351	252	865922	36.524	ng	100
83) Chrysene	21.474	228	2389062	38.526	ng	100
84) Bis(2-ethylhexyl)phtha...	21.362	149	1574446	41.698	ng	100
85) Di-n-octyl phthalate	22.315	149	2769055	43.008	ng	100
87) Indeno(1,2,3-cd)pyrene	26.468	276	3078892	39.145	ng	100
88) Benzo(b)fluoranthene	23.145	252	2571431	34.369	ng	100
89) Benzo(k)fluoranthene	23.192	252	2489072	33.607	ng	100
90) Benzo(a)pyrene	23.780	252	2204884	37.954	ng	100
91) Dibenzo(a,h)anthracene	26.491	278	2541576	38.598	ng	100
92) Benzo(g,h,i)perylene	27.250	276	2570362	41.725	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101321\
 Data File : BP007401.D
 Acq On : 13 Oct 2021 12:24
 Operator : CG/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICCC040

Quant Time: Oct 13 13:07:17 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP101321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 13:04:34 2021
 Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.969 min Scan# 830
 Delta R.T. -0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC040


Tgt Ion:	152	Resp:	208283
Ion	Ratio	Lower	Upper
152	100		
150	155.1	124.1	186.1
115	60.1	48.1	72.1



1,4-Dioxane
 Concen: 43.072 ng
 RT: 3.422 min Scan# 57
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24

Tgt Ion:	88	Resp:	209455
Ion	Ratio	Lower	Upper
88	100		
58	72.6	58.1	87.1
43	27.8	22.2	33.4









Pyridine
 Concen: 41.375 ng
 RT: 3.816 min Scan# 124
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC040


Tgt Ion:	79	Resp:	569192
Ion	Ratio	Lower	Upper
79	100		
52	69.6	55.7	83.5
51	35.6	28.5	42.7



n-Nitrosodimethylamine
 Concen: 48.273 ng
 RT: 3.728 min Scan# 109
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24

Tgt Ion:	42	Resp:	248296
Ion	Ratio	Lower	Upper
42	100		
74	133.7	107.0	160.4
44	9.3	7.4	11.2









2-Fluorophenol
 Concen: 75.357 ng
 RT: 5.528 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC040


Tgt Ion:	112	Resp:	952549
Ion	Ratio	Lower	Upper
112	100		
64	58.3	46.6	70.0
63	31.3	25.0	37.6



Aniline
 Concen: 39.214 ng
 RT: 7.287 min Scan# 714
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24

Tgt Ion:	93	Resp:	738220
Ion	Ratio	Lower	Upper
93	100		
66	38.1	30.5	45.7
65	19.8	15.8	23.8









Phenol-d6
 Concen: 75.765 ng
 RT: 7.122 min Scan# 686
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC040


Tgt Ion:	99	Resp:	1285753
Ion	Ratio	Lower	Upper
99	100		
42	18.6	14.9	22.3
71	36.5	29.2	43.8



2-Chlorophenol
 Concen: 35.799 ng
 RT: 7.528 min Scan# 755
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24

Tgt Ion:	128	Resp:	489792
Ion	Ratio	Lower	Upper
128	100		
130	32.0	12.0	52.0
64	47.0	27.0	67.0









Scan 682 (7.098 min): BP007401.D\data.ms (-673) #9
 Benzaldehyde
 Concen: 38.434 ng
 RT: 7.098 min Scan# 682
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC040

Tgt Ion:	77	Resp:	377458
Ion	Ratio	Lower	Upper
77	100		
105	97.5	77.5	117.5
106	92.1	72.1	112.1



Scan 691 (7.151 min): BP007401.D\data.ms (-679) #10
 Phenol
 Concen: 38.591 ng
 RT: 7.151 min Scan# 691
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24


Tgt Ion:	94	Resp:	626223
Ion	Ratio	Lower	Upper
94	100		
65	28.2	8.2	48.2
66	38.8	18.8	58.8






Abundance Scan 731 (7.387 min): BP007401.D\data.ms (-724) #11

bis(2-Chloroethyl)ether
Concen: 39.191 ng
RT: 7.387 min Scan# 731
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040




Tgt Ion	Ratio	Lower	Upper
93	100		
63	75.7	55.7	95.7
95	32.7	12.7	52.7






Abundance Scan 811 (7.857 min): BP007401.D\data.ms (-802) #12


1,3-Dichlorobenzene
Concen: 37.078 ng
RT: 7.857 min Scan# 811
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24



Tgt Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.4	77.0
75	32.6	26.1	39.1






Abundance Scan 836 (8.004 min): BP007401.D\data.ms (-825) #13




1,4-Dichlorobenzene
Concen: 37.185 ng
RT: 8.004 min Scan# 836
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:	146	Resp:	576047
Ion Ratio	Lower	Upper	
146	100		
148	63.8	51.0	76.6
111	41.0	32.8	49.2







Abundance Scan 890 (8.322 min): BP007401.D\data.ms (-879) #14



1,2-Dichlorobenzene
Concen: 36.328 ng
RT: 8.322 min Scan# 890
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Tgt Ion:	146	Resp:	544918
Ion Ratio	Lower	Upper	
146	100		
148	64.4	51.5	77.3
111	43.9	35.1	52.7









Benzyl Alcohol
 Concen: 43.748 ng
 RT: 8.198 min Scan# 869
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC040


Tgt Ion:	79	Resp:	481828
Ion	Ratio	Lower	Upper
79	100		
108	74.9	59.9	89.9
77	63.9	51.1	76.7



2,2'-oxybis(1-Chloropropane)
 Concen: 44.597 ng
 RT: 8.504 min Scan# 921
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24

Tgt Ion:	45	Resp:	667409
Ion	Ratio	Lower	Upper
45	100		
77	14.9	0.0	34.9
79	11.7	0.0	31.7









2-Methylphenol
Concen: 37.876 ng
RT: 8.404 min Scan# 904
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:	107	Resp:	424284
Ion	Ratio	Lower	Upper
107	100		
108	111.5	89.2	133.8
77	48.1	38.5	57.7
79	48.0	38.4	57.6



Hexachloroethane
Concen: 39.878 ng
RT: 9.057 min Scan# 1015
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24


Tgt Ion:	117	Resp:	208770
Ion	Ratio	Lower	Upper
117	100		
119	97.0	77.6	116.4
201	103.6	82.9	124.3






Abundance Scan 967 (8.775 min): BP007401.D\data.ms (-960) #19

n-Nitroso-di-n-propylamine
Concen: 38.756 ng
RT: 8.775 min Scan# 967
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040




Tgt Ion:	70	Resp:	372952
Ion	Ratio	Lower	Upper
70	100		
42	57.7	46.2	69.2
101	9.6	7.7	11.5
130	20.3	16.2	24.4







Abundance Scan 960 (8.734 min): BP007401.D\data.ms (-950) #20

3+4-Methylphenols
Concen: 37.315 ng
RT: 8.734 min Scan# 960
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24



Tgt Ion:	107	Resp:	585846
Ion	Ratio	Lower	Upper
107	100		
108	89.2	69.2	109.2
77	33.8	13.8	53.8
79	30.2	10.2	50.2









Scan 1307 (10.775 min): BP007401.D\data.ms (-1256) (-)
 Naphthalene-d8
 Concen: 20.000 ng
 RT: 10.775 min Scan# 1307
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC040

Tgt Ion:	136	Resp:	803518
Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.6	13.0
54	7.8	6.2	9.4
68	5.9	4.7	7.1




Scan 969 (8.787 min): BP007401.D\data.ms (-962) (-)
 Acetophenone
 Concen: 39.608 ng
 RT: 8.787 min Scan# 969
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24

Tgt Ion:	105	Resp:	748612
Ion	Ratio	Lower	Upper
105	100		
71	5.6	4.5	6.7
51	27.3	21.8	32.8
120	22.0	17.6	26.4






Abundance Scan 1026 (9.122 min): BP007401.D\data.ms (-1023)




Nitrobenzene-d5
Concen: 77.131 ng
RT: 9.122 min Scan# 1026
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:	82	Resp:	1052306
Ion	Ratio	Lower	Upper
82	100		
128	40.7	32.6	48.8
54	50.5	40.4	60.6






Abundance Scan 1033 (9.163 min): BP007401.D\data.ms (-1024)




Nitrobenzene
Concen: 40.661 ng
RT: 9.163 min Scan# 1033
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Tgt Ion:	77	Resp:	523409
Ion	Ratio	Lower	Upper
77	100		
123	44.1	35.3	52.9
65	13.6	10.9	16.3






Abundance Scan 1124 (9.698 min): BP007401.D\data.ms (-1125)




Isophorone
Concen: 40.163 ng
RT: 9.698 min Scan# 1124
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:	82	Resp:	1034032
Ion	Ratio	Lower	Upper
82	100		
95	7.7	6.2	9.2
138	17.5	14.0	21.0






Abundance Scan 1155 (9.881 min): BP007401.D\data.ms (-1126)



2-Nitrophenol
Concen: 28.056 ng
RT: 9.881 min Scan# 1155
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24


Tgt Ion:	139	Resp:	201751
Ion	Ratio	Lower	Upper
139	100		
109	32.7	26.2	39.2
65	50.4	40.3	60.5





Abundance Scan 1166 (9.945 min): BP007401.D\data.ms (-1157) (-)

2,4-Dimethylphenol
Concen: 39.403 ng
RT: 9.945 min Scan# 1166
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24


Instrument :
BNA_P
ClientSampleId :
SSTDICCC040



Tgt Ion:	122	Resp:	420774
Ion	Ratio	Lower	Upper
122	100		
107	117.2	93.8	140.6
121	58.7	47.0	70.4





Abundance Scan 1166 (9.945 min): BP007401.D\data.ms (-1115) (-)




Abundance Scan 1206 (10.181 min): BP007401.D\data.ms (-1123) (-)

bis(2-Chloroethoxy)methane
Concen: 39.006 ng
RT: 10.181 min Scan# 1206
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Tgt Ion:	93	Resp:	674568
Ion	Ratio	Lower	Upper
93	100		
95	32.6	26.1	39.1
123	10.6	8.5	12.7




Abundance Scan 1206 (10.181 min): BP007401.D\data.ms (-1155) (-)






Abundance Scan 1246 (10.416 min): BP007401.D\data.ms (-1229) (-)

2,4-Dichlorophenol
Concen: 36.502 ng
RT: 10.416 min Scan# 1246
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040




Tgt Ion	Ratio	Lower	Upper
162	100		
164	64.0	44.0	84.0
98	36.3	16.3	56.3







Abundance Scan 1284 (10.639 min): BP007401.D\data.ms (-1230) (-)

1,2,4-Trichlorobenzene
Concen: 36.744 ng
RT: 10.639 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24



Tgt Ion	Ratio	Lower	Upper
180	100		
182	98.0	78.4	117.6
145	29.8	23.8	35.8









Naphthalene
 Concen: 36.638 ng
 RT: 10.822 min Scan# 1315
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC040


Tgt Ion:	128	Resp:	1506577
Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.9	13.3
127	12.8	10.2	15.4



Benzoic acid
 Concen: 28.387 ng
 RT: 10.063 min Scan# 1186
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24

Tgt Ion:	122	Resp:	244582
Ion	Ratio	Lower	Upper
122	100		
105	125.9	105.9	145.9
77	94.6	74.6	114.6









Scan 1332 (10.922 min): BP007401.D\data.ms (-128) (-)
 4-Chloroaniline
 Concen: 38.413 ng
 RT: 10.922 min Scan# 1332
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC040


Tgt Ion	Ratio	Lower	Upper
127	100		
129	31.4	25.1	37.7
65	31.9	25.5	38.3
92	20.6	16.5	24.7



Scan 1367 (11.128 min): BP007401.D\data.ms (-134) (-)
 Hexachlorobutadiene
 Concen: 39.008 ng
 RT: 11.128 min Scan# 1367
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24

Tgt Ion	Ratio	Lower	Upper
225	100		
223	61.7	49.4	74.0
227	63.8	51.0	76.6









Caprolactam
Concen: 22.795 ng
RT: 11.686 min Scan# 1462
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:	113	Resp:	92563
Ion	Ratio	Lower	Upper
113	100		
55	163.9	143.9	183.9
56	121.6	101.6	141.6



4-Chloro-3-methylphenol
Concen: 42.308 ng
RT: 12.045 min Scan# 1523
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Tgt Ion:	107	Resp:	508597
Ion	Ratio	Lower	Upper
107	100		
144	25.6	20.5	30.7
142	78.7	63.0	94.4









2-Methylnaphthalene
Concen: 41.277 ng
RT: 12.433 min Scan# 1589
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:	142	Resp:	1062546
Ion	Ratio	Lower	Upper
142	100		
141	89.2	71.4	107.0
115	36.9	29.5	44.3




1-Methylnaphthalene
Concen: 40.830 ng
RT: 12.651 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Tgt Ion:	142	Resp:	1031887
Ion	Ratio	Lower	Upper
142	100		
141	93.0	74.4	111.6
116	4.2	3.4	5.0






Abundance Scan 1958 (14.604 min): BP007401.D\data.ms (-1907) (-)




Acenaphthene-d10
Concen: 20.000 ng
RT: 14.604 min Scan# 1958
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:	164	Resp:	505498
Ion Ratio	Lower	Upper	
164	100		
162	101.4	81.1	121.7
160	45.0	36.0	54.0







Abundance Scan 1651 (12.798 min): BP007401.D\data.ms (-1600) (-)



1,2,4,5-Tetrachlorobenzene
Concen: 34.427 ng
RT: 12.798 min Scan# 1651
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Tgt Ion:	216	Resp:	565351
Ion Ratio	Lower	Upper	
216	100		
214	77.8	62.2	93.4
179	21.4	17.1	25.7
108	19.2	15.4	23.0









Hexachlorocyclopentadiene
Concen: 67.484 ng
RT: 12.786 min Scan# 1649
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:	237	Resp:	325418
Ion	Ratio	Lower	Upper
237	100		
235	61.2	41.2	81.2
272	11.9	0.0	31.9



2,4,6-Tribromophenol
Concen: 49.435 ng
RT: 16.086 min Scan# 2210
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24


Tgt Ion:	330	Resp:	409910
Ion	Ratio	Lower	Upper
330	100		
332	98.6	78.9	118.3
141	38.4	30.7	46.1





Abundance Scan 1691 (13.033 min): BP007401.D\data.ms (-1640) (-)

2,4,6-Trichlorophenol
Concen: 30.882 ng
RT: 13.033 min Scan# 1691
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24


Instrument :
BNA_P
ClientSampleId :
SSTDICCC040



Tgt Ion:	196	Resp:	347084
Ion	Ratio	Lower	Upper
196	100		
198	95.5	76.4	114.6
200	31.2	25.0	37.4




Abundance Scan 1691 (13.033 min): BP007401.D\data.ms (-1640) (-)





Abundance Scan 1702 (13.098 min): BP007401.D\data.ms (-1651) (-)

2,4,5-Trichlorophenol
Concen: 27.806 ng
RT: 13.098 min Scan# 1702
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24


Tgt Ion:	196	Resp:	338369
Ion	Ratio	Lower	Upper
196	100		
198	96.4	77.1	115.7
97	51.1	40.9	61.3
132	26.5	21.2	31.8



Abundance Scan 1702 (13.098 min): BP007401.D\data.ms (-1651) (-)






Abundance Scan 1725 (13.233 min): BP007401.D\data.ms (-1715)




2-Fluorobiphenyl
Concen: 68.376 ng
RT: 13.233 min Scan# 1725
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:	172	Resp:	2473605
Ion	Ratio	Lower	Upper
172	100		
171	35.2	28.2	42.2
170	23.7	19.0	28.4







Abundance Scan 1760 (13.439 min): BP007401.D\data.ms (-17516)



1,1'-Biphenyl
Concen: 36.324 ng
RT: 13.439 min Scan# 1760
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Tgt Ion:	154	Resp:	1378551
Ion	Ratio	Lower	Upper
154	100		
153	39.9	19.9	59.9
76	12.4	0.0	32.4









2-Chloronaphthalene
Concen: 35.831 ng
RT: 13.480 min Scan# 1767
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:	162	Resp:	1060853
Ion	Ratio	Lower	Upper
162	100		
127	38.8	31.0	46.6
164	33.0	26.4	39.6




2-Nitroaniline
Concen: 33.425 ng
RT: 13.669 min Scan# 1799
Delta R.T. -0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Tgt Ion:	65	Resp:	256232
Ion	Ratio	Lower	Upper
65	100		
92	69.9	55.9	83.9
138	107.9	86.3	129.5






Abundance Scan 1910 (14.322 min): BP007401.D\data.ms (-1859) (-)




Acenaphthylene
Concen: 37.192 ng
RT: 14.322 min Scan# 1910
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:	152	Resp:	1699670
Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.8	23.6
153	12.7	10.2	15.2






Abundance Scan 1866 (14.063 min): BP007401.D\data.ms (-1859) (-)



Dimethylphthalate
Concen: 35.626 ng
RT: 14.063 min Scan# 1866
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24


Tgt Ion:	163	Resp:	1387720
Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.3	4.9
164	10.2	8.2	12.2






Abundance Scan 1884 (14.169 min): BP007401.D\data.ms (-1833) (-)

2,6-Dinitrotoluene
Concen: 32.322 ng
RT: 14.169 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040




Tgt Ion:	165	Resp:	262094
Ion Ratio	Lower	Upper	
165	100		
63	55.2	44.2	66.2
89	58.3	46.6	70.0






Abundance Scan 1968 (14.663 min): BP007401.D\data.ms (-1917) (-)

Acenaphthene
Concen: 38.563 ng
RT: 14.663 min Scan# 1968
Delta R.T. -0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24




Tgt Ion:	154	Resp:	1040255
Ion Ratio	Lower	Upper	
154	100		
153	113.0	90.4	135.6
152	52.9	42.3	63.5






Abundance Scan 1939 (14.492 min): BP007401.D\data.ms (-1958)

3-Nitroaniline
Concen: 34.488 ng
RT: 14.492 min Scan# 1939
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040




Tgt Ion:	138	Resp:	286842
Ion	Ratio	Lower	Upper
138	100		
108	15.4	12.3	18.5
92	118.6	94.9	142.3







Abundance Scan 1974 (14.698 min): BP007401.D\data.ms (-1958)

2,4-Dinitrophenol
Concen: 18.325 ng
RT: 14.698 min Scan# 1974
Delta R.T. -0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24



Tgt Ion:	184	Resp:	83270
Ion	Ratio	Lower	Upper
184	100		
63	55.0	44.0	66.0
154	61.1	48.9	73.3









Dibenzenofuran
Concen: 35.462 ng
RT: 14.998 min Scan# 2025
Delta R.T. -0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:	168	Resp:	1660529
Ion	Ratio	Lower	Upper
168	100		
139	38.5	30.8	46.2
169	13.8	11.0	16.6



4-Nitrophenol
Concen: 35.480 ng
RT: 14.792 min Scan# 1990
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Tgt Ion:	139	Resp:	231417
Ion	Ratio	Lower	Upper
139	100		
109	80.9	60.9	100.9
65	96.5	76.5	116.5









Scan 2017 (14.951 min): BP007401.D\data.ms (-2057) (-)
 2,4-Dinitrotoluene
 Concen: 30.197 ng
 RT: 14.951 min Scan# 2017
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC040

Tgt Ion:	165	Resp:	334996
Ion	Ratio	Lower	Upper
165	100		
63	46.7	37.4	56.0
89	77.2	61.8	92.6
182	2.6	2.1	3.1




Scan 2136 (15.651 min): BP007401.D\data.ms (-2158) (-)
 Fluorene
 Concen: 33.246 ng
 RT: 15.651 min Scan# 2136
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24

Tgt Ion:	166	Resp:	1205074
Ion	Ratio	Lower	Upper
166	100		
165	98.6	78.9	118.3
167	13.6	10.9	16.3






Abundance Scan 2063 (15.221 min): BP007401.D\data.ms (-2059) (-)




2,3,4,6-Tetrachlorophenol
Concen: 29.439 ng
RT: 15.139 min Scan# 2049
Delta R.T. -0.082 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion	Ratio	Lower	Upper
232	100		
131	40.1	30.3	45.5
130	2.4	1.8	2.8
166	25.5	19.4	29.0






Abundance Scan 2098 (15.427 min): BP007401.D\data.ms (-2047) (-)



Diethylphthalate
Concen: 36.125 ng
RT: 15.427 min Scan# 2098
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24


Tgt Ion	Ratio	Lower	Upper
149	100		
177	22.3	17.8	26.8
150	12.7	10.2	15.2





Abundance Scan 2135 (15.645 min): BP007401.D\data.ms (-2184) (-)

4-Chlorophenyl-phenylether
Concen: 31.231 ng
RT: 15.645 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24


Instrument :
BNA_P
ClientSampleId :
SSTDICCC040



Tgt Ion	Ratio	Lower	Upper
204	100		
206	32.3	25.8	38.8
141	61.0	48.8	73.2






Abundance Scan 2135 (15.645 min): BP007401.D\data.ms (-2084) (-)




Abundance Scan 2137 (15.657 min): BP007401.D\data.ms (-2184) (-)

4-Nitroaniline
Concen: 30.243 ng
RT: 15.657 min Scan# 2137
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Tgt Ion	Ratio	Lower	Upper
138	100		
92	50.7	30.7	70.7
108	91.2	71.2	111.2




Abundance Scan 2137 (15.657 min): BP007401.D\data.ms (-2086) (-)






Abundance Scan 2185 (15.939 min): BP007401.D\data.ms (-213461) (-)

Azobenzene
Concen: 41.957 ng
RT: 15.939 min Scan# 2185
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040






Tgt Ion:	77	Resp:	1329746
Ion	Ratio	Lower	Upper
77	100		
182	26.5	6.5	46.5
105	21.3	1.3	41.3
51	30.5	10.5	50.5




Abundance Scan 2425 (17.351 min): BP007401.D\data.ms (-2374) (-)

Phenanthrene-d10
Concen: 20.000 ng
RT: 17.351 min Scan# 2425
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Tgt Ion:	188	Resp:	1063382
Ion	Ratio	Lower	Upper
188	100		
94	8.9	7.1	10.7
80	9.7	7.8	11.6









4,6-Dinitro-2-methylphenol
 Concen: 22.287 ng
 RT: 15.716 min Scan# 2147
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC040

Tgt Ion:	198	Resp:	163263
Ion Ratio	Lower	Upper	
198	100		
51	43.8	23.8	63.8
105	44.6	24.6	64.6



n-Nitrosodiphenylamine
 Concen: 33.935 ng
 RT: 15.857 min Scan# 2171
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24


Tgt Ion:	169	Resp:	1144070
Ion Ratio	Lower	Upper	
169	100		
168	65.9	52.7	79.1
167	36.3	29.0	43.6






Abundance Scan 2288 (16.545 min): BP007401.D\data.ms (-2287) (-)

4-Bromophenyl-phenylether
Concen: 29.735 ng
RT: 16.545 min Scan# 2288
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040




Tgt Ion:	248	Resp:	408100
Ion Ratio	Lower	Upper	
248	100		
250	97.7	78.2	117.2
141	76.0	60.8	91.2






Abundance Scan 2307 (16.657 min): BP007401.D\data.ms (-2298) (-)

Hexachlorobenzene
Concen: 29.877 ng
RT: 16.657 min Scan# 2307
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24




Tgt Ion:	284	Resp:	449977
Ion Ratio	Lower	Upper	
284	100		
142	38.3	30.6	46.0
249	33.3	26.6	40.0





Abundance Scan 2333 (16.810 min): BP007401.D\data.ms (-2333) (-)

Atrazine
Concen: 34.184 ng
RT: 16.810 min Scan# 2333
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24


Instrument :
BNA_P
ClientSampleId :
SSTDICCC040



Tgt Ion	Ratio	Lower	Upper
200	100		
173	26.2	6.2	46.2
215	48.2	28.2	68.2




Abundance Scan 2333 (16.810 min): BP007401.D\data.ms (-2333) (-)





Abundance Scan 2365 (16.998 min): BP007401.D\data.ms (-2365) (-)


Pentachlorophenol
Concen: 32.879 ng
RT: 16.998 min Scan# 2365
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24




Tgt Ion	Ratio	Lower	Upper
266	100		
268	61.6	49.3	73.9
264	60.9	48.7	73.1



Abundance Scan 2365 (16.998 min): BP007401.D\data.ms (-2365) (-)





Abundance Scan 2432 (17.392 min): BP007401.D\data.ms (-2470) (-)




Phenanthrene
Concen: 36.829 ng
RT: 17.392 min Scan# 2432
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:178 Resp: 2068143
Ion Ratio Lower Upper
178 100
176 19.2 15.4 23.0
179 15.4 12.3 18.5



Abundance Scan 2432 (17.392 min): BP007401.D\data.ms (-2381) (-)





Abundance Scan 2448 (17.486 min): BP007401.D\data.ms (-2470) (-)




Anthracene
Concen: 34.582 ng
RT: 17.486 min Scan# 2448
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24


Tgt Ion:178 Resp: 2052164
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 15.3 12.2 18.4



Abundance Scan 2448 (17.486 min): BP007401.D\data.ms (-2397) (-)





Abundance Scan 2493 (17.751 min): BP007401.D\data.ms (-2442) (-)




Carbazole
Concen: 35.517 ng
RT: 17.751 min Scan# 2493
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:	167	Resp:	2053633
Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.3	25.9
139	13.0	10.4	15.6



Abundance Scan 2493 (17.751 min): BP007401.D\data.ms (-2442) (-)





Abundance Scan 2589 (18.315 min): BP007401.D\data.ms (-2538) (-)




Di-n-butylphthalate
Concen: 36.168 ng
RT: 18.315 min Scan# 2589
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24


Tgt Ion:	149	Resp:	2426912
Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.6	11.4
104	5.3	4.2	6.4



Abundance Scan 2589 (18.315 min): BP007401.D\data.ms (-2538) (-)






Abundance Scan 2766 (19.357 min): BP007401.D\data.ms (-2766) (-)




Fluoranthene
Concen: 34.424 ng
RT: 19.357 min Scan# 2766
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:	202	Resp:	2493001
Ion	Ratio	Lower	Upper
202	100		
101	10.4	0.0	30.4
203	17.6	0.0	37.6






Abundance Scan 3120 (21.439 min): BP007401.D\data.ms (-3120) (-)



Chrysene-d12
Concen: 20.000 ng
RT: 21.439 min Scan# 3120
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24


Tgt Ion:	240	Resp:	1019845
Ion	Ratio	Lower	Upper
240	100		
120	9.8	7.8	11.8
236	25.7	20.6	30.8





Abundance Scan 2795 (19.527 min): BP007401.D\data.ms (-2777) (-)

Benzidine
Concen: 40.516 ng
RT: 19.527 min Scan# 2795
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24


Instrument :
BNA_P
ClientSampleId :
SSTDICCC040



Tgt Ion	Ratio	Lower	Upper
184	100		
185	14.0	11.2	16.8
183	11.4	9.1	13.7





Abundance Scan 2795 (19.527 min): BP007401.D\data.ms (-2744) (-)




Abundance Scan 2826 (19.709 min): BP007401.D\data.ms (-2874) (-)

Pyrene
Concen: 36.510 ng
RT: 19.709 min Scan# 2826
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Tgt Ion	Ratio	Lower	Upper
202	100		
200	21.1	16.9	25.3
203	17.2	13.8	20.6




Abundance Scan 2826 (19.709 min): BP007401.D\data.ms (-2775) (-)






Abundance Scan 2860 (19.909 min): BP007401.D\data.ms (-2809) (-)

Terphenyl-d14
Concen: 62.243 ng
RT: 19.909 min Scan# 2860
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040






Tgt Ion:	244	Resp:	3598985
Ion	Ratio	Lower	Upper
244	100		
212	7.7	6.2	9.2
122	9.8	7.8	11.8




Abundance Scan 2976 (20.592 min): BP007401.D\data.ms (-2925) (-)

Butylbenzylphthalate
Concen: 39.729 ng
RT: 20.592 min Scan# 2976
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Tgt Ion:	149	Resp:	1077856
Ion	Ratio	Lower	Upper
149	100		
91	70.8	56.6	85.0
206	21.8	17.4	26.2






Abundance Scan 3117 (21.421 min): BP007401.D\data.ms (-3181) (-)




Benzo(a)anthracene
Concen: 38.704 ng
RT: 21.421 min Scan# 3117
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:	228	Resp:	2549954
Ion	Ratio	Lower	Upper
228	100		
226	27.2	21.8	32.6
229	19.6	15.7	23.5







Abundance Scan 3105 (21.351 min): BP007401.D\data.ms (-3082) (-)



3,3'-Dichlorobenzidine
Concen: 36.524 ng
RT: 21.351 min Scan# 3105
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Tgt Ion:	252	Resp:	865922
Ion	Ratio	Lower	Upper
252	100		
254	64.4	51.5	77.3
126	10.6	8.5	12.7









Chrysene
 Concen: 38.526 ng
 RT: 21.474 min Scan# 3126
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC040

Tgt Ion:	228	Resp:	2389062
Ion	Ratio	Lower	Upper
228	100		
226	30.8	24.6	37.0
229	19.6	15.7	23.5




Bis(2-ethylhexyl)phthalate
 Concen: 41.698 ng
 RT: 21.362 min Scan# 3107
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24

Tgt Ion:	149	Resp:	1574446
Ion	Ratio	Lower	Upper
149	100		
167	27.0	21.6	32.4
279	4.1	3.3	4.9





Abundance Scan 3269 (22.315 min): BP007401.D\data.ms (-3269) (-)




Di-n-octyl phthalate
Concen: 43.008 ng
RT: 22.315 min Scan# 3269
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:	149	Resp:	2769055
Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	9.1	7.3	10.9



Abundance Scan 3269 (22.315 min): BP007401.D\data.ms (-3218) (-)





Abundance Scan 3537 (23.892 min): BP007401.D\data.ms (-3537) (-)




Perylene-d12
Concen: 20.000 ng
RT: 23.892 min Scan# 3537
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24


Tgt Ion:	264	Resp:	1149124
Ion	Ratio	Lower	Upper
264	100		
260	23.9	19.1	28.7
265	20.9	16.7	25.1



Abundance Scan 3537 (23.892 min): BP007401.D\data.ms (-3486) (-)






Abundance Scan 3975 (26.468 min): BP007401.D\data.ms (-3927) (-)




Indeno(1,2,3-cd)pyrene
Concen: 39.145 ng
RT: 26.468 min Scan# 3975
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion	Ratio	Lower	Upper
276	100		
138	24.2	19.4	29.0
277	25.2	20.2	30.2







Abundance Scan 3410 (23.145 min): BP007401.D\data.ms (-3482) (-)



Benzo(b)fluoranthene
Concen: 34.369 ng
RT: 23.145 min Scan# 3410
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.4	26.2
125	9.2	7.4	11.0









Benzo(k)fluoranthene
 Concen: 33.607 ng
 RT: 23.192 min Scan# 3418
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC040

Tgt Ion:252 Resp: 2489072




Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	9.2	7.4	11.0




Benzo(a)pyrene
 Concen: 37.954 ng
 RT: 23.780 min Scan# 3518
 Delta R.T. 0.000 min
 Lab File: BP007401.D
 Acq: 13 Oct 2021 12:24

Tgt Ion:252 Resp: 2204884

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	10.5	8.4	12.6






Abundance Scan 3979 (26.491 min): BP007401.D\data.ms (-3928) (-)




Dibenzo(a,h)anthracene
Concen: 38.598 ng
RT: 26.491 min Scan# 3979
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:	278	Resp:	2541576
Ion	Ratio	Lower	Upper
278	100		
139	14.7	11.8	17.6
279	23.2	18.6	27.8



Abundance Scan 4108 (27.250 min): BP007401.D\data.ms (-4057) (-)



Benzo(g,h,i)perylene
Concen: 41.725 ng
RT: 27.250 min Scan# 4108
Delta R.T. 0.000 min
Lab File: BP007401.D
Acq: 13 Oct 2021 12:24

Tgt Ion:	276	Resp:	2570362
Ion	Ratio	Lower	Upper
276	100		
277	24.2	19.4	29.0
138	20.3	16.2	24.4

