

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101321\
 Data File : BP007402.D
 Acq On : 13 Oct 2021 12:58
 Operator : CG/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050

Quant Time: Oct 13 13:39:42 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP101321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 13:04:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.969	152	224763	20.000	ng	0.00
21) Naphthalene-d8	10.775	136	852325	20.000	ng	0.00
39) Acenaphthene-d10	14.604	164	531827	20.000	ng	0.00
64) Phenanthrene-d10	17.351	188	1106937	20.000	ng	0.00
76) Chrysene-d12	21.439	240	1053328	20.000	ng	0.00
86) Perylene-d12	23.892	264	1192798	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.534	112	1249234	91.582	ng	0.00
7) Phenol-d6	7.122	99	1677047	91.577	ng	0.00
23) Nitrobenzene-d5	9.128	82	1404695	97.064	ng	0.00
42) 2,4,6-Tribromophenol	16.092	330	537863	61.655	ng	0.00
45) 2-Fluorobiphenyl	13.234	172	3145575	82.646	ng	0.00
79) Terphenyl-d14	19.910	244	4502621	75.395	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.423	88	269278	51.314	ng	99
3) Pyridine	3.817	79	746782	50.304	ng	98
4) n-Nitrosodimethylamine	3.728	42	333693	60.119	ng	94
6) Aniline	7.287	93	960584	47.284	ng	98
8) 2-Chlorophenol	7.528	128	647849	43.880	ng	100
9) Benzaldehyde	7.099	77	490823	46.313	ng	99
10) Phenol	7.152	94	817078	46.661	ng	99
11) bis(2-Chloroethyl)ether	7.387	93	655228	46.682	ng	99
12) 1,3-Dichlorobenzene	7.858	146	745700	44.793	ng	99
13) 1,4-Dichlorobenzene	8.005	146	751079	44.929	ng	99
14) 1,2-Dichlorobenzene	8.322	146	707616	43.716	ng	99
15) Benzyl Alcohol	8.205	79	626159	52.684	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.505	45	846066	52.390	ng	99
17) 2-Methylphenol	8.405	107	553402	45.780	ng	99
18) Hexachloroethane	9.058	117	275342	48.738	ng	99
19) n-Nitroso-di-n-propyla...	8.781	70	478158	46.045	ng	99
20) 3+4-Methylphenols	8.734	107	768671	45.370	ng	100
22) Acetophenone	8.787	105	959849	47.876	ng	99
24) Nitrobenzene	9.169	77	687751	50.368	ng	97
25) Isophorone	9.699	82	1354434	49.595	ng	100
26) 2-Nitrophenol	9.881	139	289471	37.949	ng	99
27) 2,4-Dimethylphenol	9.946	122	551435	48.681	ng	97
28) bis(2-Chloroethoxy)met...	10.187	93	875383	47.719	ng	99
29) 2,4-Dichlorophenol	10.416	162	620562	45.555	ng	99
30) 1,2,4-Trichlorobenzene	10.640	180	694833	45.276	ng	98
31) Naphthalene	10.822	128	1963621	45.019	ng	99
32) Benzoic acid	10.081	122	361765	39.583	ng	98
33) 4-Chloroaniline	10.922	127	848813	47.028	ng	99
34) Hexachlorobutadiene	11.122	225	439525	48.004	ng	99
35) Caprolactam	11.693	113	128289	29.784	ng	96
36) 4-Chloro-3-methylphenol	12.051	107	669932	52.537	ng	98
37) 2-Methylnaphthalene	12.434	142	1375786	50.385	ng	100
38) 1-Methylnaphthalene	12.651	142	1331790	49.679	ng	99
40) 1,2,4,5-Tetrachloroben...	12.799	216	725693	42.003	ng	99
41) Hexachlorocyclopentadiene	12.787	237	425453	83.861	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101321\
 Data File : BP007402.D
 Acq On : 13 Oct 2021 12:58
 Operator : CG/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050

Quant Time: Oct 13 13:39:42 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP101321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 13:04:34 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	13.034	196	452877	38.300	ng	99
44) 2,4,5-Trichlorophenol	13.104	196	443949	34.676	ng	99
46) 1,1'-Biphenyl	13.440	154	1760591	44.094	ng	99
47) 2-Chloronaphthalene	13.481	162	1359302	43.638	ng	99
48) 2-Nitroaniline	13.669	65	363765	45.103	ng	98
49) Acenaphthylene	14.322	152	2200607	45.769	ng	99
50) Dimethylphthalate	14.063	163	1782894	43.505	ng	100
51) 2,6-Dinitrotoluene	14.169	165	354328	41.533	ng	97
52) Acenaphthene	14.663	154	1354714	47.734	ng	99
53) 3-Nitroaniline	14.493	138	393715	44.994	ng	98
54) 2,4-Dinitrophenol	14.698	184	120821	25.273	ng	95
55) Dibenzofuran	14.998	168	2132074	43.278	ng	99
56) 4-Nitrophenol	14.793	139	314835	45.879	ng	96
57) 2,4-Dinitrotoluene	14.957	165	463903	39.746	ng	95
58) Fluorene	15.651	166	1512821	39.669	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.140	232	437924	37.429	ng	97
60) Diethylphthalate	15.434	149	1798348	44.148	ng	99
61) 4-Chlorophenyl-phenyle...	15.645	204	786755	37.033	ng	97
62) 4-Nitroaniline	15.657	138	353963	37.803	ng	95
63) Azobenzene	15.940	77	1687771	50.617	ng	98
65) 4,6-Dinitro-2-methylph...	15.722	198	237720	31.175	ng	95
66) n-Nitrosodiphenylamine	15.857	169	1468374	41.840	ng	99
67) 4-Bromophenyl-phenylether	16.545	248	528366	36.983	ng	100
68) Hexachlorobenzene	16.663	284	582996	37.186	ng	96
69) Atrazine	16.816	200	536603	41.553	ng	99
70) Pentachlorophenol	16.998	266	367401	42.574	ng	97
71) Phenanthrene	17.398	178	2640262	45.167	ng	100
72) Anthracene	17.486	178	2630180	42.578	ng	100
73) Carbazole	17.751	167	2634494	43.770	ng	100
74) Di-n-butylphthalate	18.316	149	3126991	44.768	ng	99
75) Fluoranthene	19.357	202	3189648	42.311	ng	99
77) Benzidine	19.533	184	1343163	43.937	ng	100
78) Pyrene	19.710	202	3266086	45.028	ng	99
80) Butylbenzylphthalate	20.592	149	1410464	50.336	ng	100
81) Benzo(a)anthracene	21.422	228	3227964	47.438	ng	100
82) 3,3'-Dichlorobenzidine	21.351	252	1103394	45.062	ng	100
83) Chrysene	21.475	228	3015067	47.075	ng	99
84) Bis(2-ethylhexyl)phtha...	21.363	149	1975811	50.664	ng	99
85) Di-n-octyl phthalate	22.316	149	3587857	53.955	ng	100
87) Indeno(1,2,3-cd)pyrene	26.468	276	3994247	48.924	ng	99
88) Benzo(b)fluoranthene	23.145	252	3257343	41.943	ng	100
89) Benzo(k)fluoranthene	23.198	252	3230737	42.024	ng	100
90) Benzo(a)pyrene	23.786	252	2854289	47.334	ng	99
91) Dibenzo(a,h)anthracene	26.492	278	3275051	47.916	ng	99
92) Benzo(g,h,i)perylene	27.257	276	3358096	52.517	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101321\
 Data File : BP007402.D
 Acq On : 13 Oct 2021 12:58
 Operator : CG/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 Client Sample Id :
 SSTDICC050

Quant Time: Oct 13 13:39:42 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP101321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 13:04:34 2021
 Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.969 min Scan# 830
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC050


Tgt Ion:	152	Resp:	224763
Ion	Ratio	Lower	Upper
152	100		
150	156.0	124.1	186.1
115	60.7	48.1	72.1



1,4-Dioxane
 Concen: 51.314 ng
 RT: 3.423 min Scan# 57
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58

Tgt Ion:	88	Resp:	269278
Ion	Ratio	Lower	Upper
88	100		
58	72.9	58.1	87.1
43	29.1	22.2	33.4









Pyridine
 Concen: 50.304 ng
 RT: 3.817 min Scan# 124
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050


Tgt Ion:	79	Resp:	746782
Ion	Ratio	Lower	Upper
79	100		
52	71.7	55.7	83.5
51	36.5	28.5	42.7



n-Nitrosodimethylamine
 Concen: 60.119 ng
 RT: 3.728 min Scan# 109
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58

Tgt Ion:	42	Resp:	333693
Ion	Ratio	Lower	Upper
42	100		
74	126.8	107.0	160.4
44	7.9	7.4	11.2









Scan 415 (5.528 min): BP007401.D\data.ms (-404)
 2-Fluorophenol
 Concen: 91.582 ng
 RT: 5.534 min Scan# 416
 Delta R.T. 0.006 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050


Tgt Ion:	112	Resp:	1249234
Ion	Ratio	Lower	Upper
112	100		
64	56.4	46.6	70.0
63	30.3	25.0	37.6



Scan 714 (7.287 min): BP007401.D\data.ms (-702)
 Aniline
 Concen: 47.284 ng
 RT: 7.287 min Scan# 714
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58

Tgt Ion:	93	Resp:	960584
Ion	Ratio	Lower	Upper
93	100		
66	40.1	30.5	45.7
65	19.7	15.8	23.8









Phenol-d6
 Concen: 91.577 ng
 RT: 7.122 min Scan# 686
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050


Tgt Ion:	99	Resp:	1677047
Ion	Ratio	Lower	Upper
99	100		
42	18.1	14.9	22.3
71	36.0	29.2	43.8



2-Chlorophenol
 Concen: 43.880 ng
 RT: 7.528 min Scan# 755
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58

Tgt Ion:	128	Resp:	647849
Ion	Ratio	Lower	Upper
128	100		
130	31.6	12.0	52.0
64	46.9	27.0	67.0









Benzaldehyde
 Concen: 46.313 ng
 RT: 7.099 min Scan# 682
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050

Tgt Ion:	77	Resp:	490823
Ion	Ratio	Lower	Upper
77	100		
105	96.4	77.5	117.5
106	93.5	72.1	112.1



Phenol
 Concen: 46.661 ng
 RT: 7.152 min Scan# 691
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58


Tgt Ion:	94	Resp:	817078
Ion	Ratio	Lower	Upper
94	100		
65	28.9	8.2	48.2
66	39.4	18.8	58.8






Abundance Scan 731 (7.387 min): BP007401.D\data.ms (-724) #11

bis(2-Chloroethyl)ether
Concen: 46.682 ng
RT: 7.387 min Scan# 731
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050




Tgt Ion:	93	Resp:	655228
Ion	Ratio	Lower	Upper
93	100		
63	76.3	55.7	95.7
95	32.3	12.7	52.7






Abundance Scan 811 (7.857 min): BP007401.D\data.ms (-802) #12


1,3-Dichlorobenzene
Concen: 44.793 ng
RT: 7.858 min Scan# 811
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58



Tgt Ion:	146	Resp:	745700
Ion	Ratio	Lower	Upper
146	100		
148	64.9	51.4	77.0
75	32.6	26.1	39.1






Abundance Scan 836 (8.004 min): BP007401.D\data.ms (-825) #13




1,4-Dichlorobenzene
Concen: 44.929 ng
RT: 8.005 min Scan# 836
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	146	Resp:	751079
Ion	Ratio	Lower	Upper
146	100		
148	64.4	51.0	76.6
111	42.2	32.8	49.2







Abundance Scan 890 (8.322 min): BP007401.D\data.ms (-879) #14



1,2-Dichlorobenzene
Concen: 43.716 ng
RT: 8.322 min Scan# 890
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion:	146	Resp:	707616
Ion	Ratio	Lower	Upper
146	100		
148	65.2	51.5	77.3
111	44.3	35.1	52.7









Benzyl Alcohol
Concen: 52.684 ng
RT: 8.205 min Scan# 870
Delta R.T. 0.006 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58




Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:	79	Resp:	626159
Ion	Ratio	Lower	Upper
79	100		
108	75.6	59.9	89.9
77	63.0	51.1	76.7



2,2'-oxybis(1-Chloropropane)
Concen: 52.390 ng
RT: 8.505 min Scan# 921
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58


Tgt Ion:	45	Resp:	846066
Ion	Ratio	Lower	Upper
45	100		
77	15.3	0.0	34.9
79	11.7	0.0	31.7






Abundance Scan 904 (8.404 min): BP007401.D\data.ms (-896) #17

2-Methylphenol
Concen: 45.780 ng
RT: 8.405 min Scan# 904
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050




Tgt Ion	Ratio	Lower	Upper
107	100		
108	111.0	89.2	133.8
77	49.2	38.5	57.7
79	47.1	38.4	57.6






Abundance Scan 1015 (9.057 min): BP007401.D\data.ms (-1063) #18

Hexachloroethane
Concen: 48.738 ng
RT: 9.058 min Scan# 1015
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58




Tgt Ion	Ratio	Lower	Upper
117	100		
119	94.5	77.6	116.4
201	103.2	82.9	124.3






Abundance Scan 967 (8.775 min): BP007401.D\data.ms (-960) #19

n-Nitroso-di-n-propylamine
Concen: 46.045 ng
RT: 8.781 min Scan# 968
Delta R.T. 0.006 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050




Tgt Ion:	70	Resp:	478158
Ion	Ratio	Lower	Upper
70	100		
42	58.3	46.2	69.2
101	10.3	7.7	11.5
130	20.8	16.2	24.4






Abundance Scan 960 (8.734 min): BP007401.D\data.ms (-950) #20

3+4-Methylphenols
Concen: 45.370 ng
RT: 8.734 min Scan# 960
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58




Tgt Ion:	107	Resp:	768671
Ion	Ratio	Lower	Upper
107	100		
108	88.9	69.2	109.2
77	33.9	13.8	53.8
79	30.0	10.2	50.2






Abundance Scan 1307 (10.775 min): BP007401.D\data.ms (-1256) (-)

Naphthalene-d8
Concen: 20.000 ng
RT: 10.775 min Scan# 1307
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050







Tgt Ion:	136	Resp:	852325
Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.6	13.0
54	7.8	6.2	9.4
68	5.5	4.7	7.1




Abundance Scan 969 (8.787 min): BP007401.D\data.ms (-962) (-)

Acetophenone
Concen: 47.876 ng
RT: 8.787 min Scan# 969
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion:	105	Resp:	959849
Ion	Ratio	Lower	Upper
105	100		
71	6.6	4.5	6.7
51	27.9	21.8	32.8
120	22.2	17.6	26.4






Abundance Scan 1026 (9.122 min): BP007401.D\data.ms (-1023-)




Nitrobenzene-d5
Concen: 97.064 ng
RT: 9.128 min Scan# 1027
Delta R.T. 0.006 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	82	Resp:	1404695
Ion	Ratio	Lower	Upper
82	100		
128	42.0	32.6	48.8
54	50.3	40.4	60.6






Abundance Scan 1033 (9.163 min): BP007401.D\data.ms (-1024-)




Nitrobenzene
Concen: 50.368 ng
RT: 9.169 min Scan# 1034
Delta R.T. 0.006 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion:	77	Resp:	687751
Ion	Ratio	Lower	Upper
77	100		
123	46.2	35.3	52.9
65	13.6	10.9	16.3





Abundance Scan 1124 (9.698 min): BP007401.D\data.ms (-1125-)




Isophorone
Concen: 49.595 ng
RT: 9.699 min Scan# 1124
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDICC050


Tgt Ion:	82	Resp:	1354434
Ion	Ratio	Lower	Upper
82	100		
95	7.9	6.2	9.2
138	17.5	14.0	21.0



Abundance Scan 1124 (9.699 min): BP007402.D\data.ms (-1073-)





Abundance Scan 1155 (9.881 min): BP007401.D\data.ms (-11426-)





2-Nitrophenol
Concen: 37.949 ng
RT: 9.881 min Scan# 1155
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion:	139	Resp:	289471
Ion	Ratio	Lower	Upper
139	100		
109	33.7	26.2	39.2
65	51.0	40.3	60.5



Abundance Scan 1155 (9.881 min): BP007402.D\data.ms (-1104-)









Scan 1166 (9.945 min): BP007401.D\data.ms (-1157) (-)
 2,4-Dimethylphenol
 Concen: 48.681 ng
 RT: 9.946 min Scan# 1166
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050

Tgt Ion:	122	Resp:	551435
Ion	Ratio	Lower	Upper
122	100		
107	121.2	93.8	140.6
121	58.9	47.0	70.4




Scan 1206 (10.181 min): BP007401.D\data.ms (-1123) (-)
 bis(2-Chloroethoxy)methane
 Concen: 47.719 ng
 RT: 10.187 min Scan# 1207
 Delta R.T. 0.006 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58

Tgt Ion:	93	Resp:	875383
Ion	Ratio	Lower	Upper
93	100		
95	32.3	26.1	39.1
123	11.7	8.5	12.7





Abundance Scan 1246 (10.416 min): BP007401.D\data.ms (-1229) (-)




2,4-Dichlorophenol
Concen: 45.555 ng
RT: 10.416 min Scan# 1246
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050


Tgt Ion:	162	Resp:	620562
Ion	Ratio	Lower	Upper
162	100		
164	64.3	44.0	84.0
98	37.7	16.3	56.3



Abundance Scan 1246 (10.416 min): BP007402.D\data.ms (-1195) (-)





Abundance Scan 1284 (10.639 min): BP007401.D\data.ms (-1230) (-)





1,2,4-Trichlorobenzene
Concen: 45.276 ng
RT: 10.640 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion:	180	Resp:	694833
Ion	Ratio	Lower	Upper
180	100		
182	95.7	78.4	117.6
145	29.9	23.8	35.8



Abundance Scan 1284 (10.640 min): BP007402.D\data.ms (-1233) (-)









Scan 1315 (10.822 min): BP007401.D\data.ms (-1264) (-)
 Naphthalene
 Concen: 45.019 ng
 RT: 10.822 min Scan# 1315
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050


Tgt Ion:	128	Resp:	1963621
Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.9	13.3
127	13.1	10.2	15.4



Scan 1186 (10.063 min): BP007401.D\data.ms (-1135) (-)
 Benzoic acid
 Concen: 39.583 ng
 RT: 10.081 min Scan# 1189
 Delta R.T. 0.018 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58

Tgt Ion:	122	Resp:	361765
Ion	Ratio	Lower	Upper
122	100		
105	128.4	105.9	145.9
77	95.7	74.6	114.6









Scan 1332 (10.922 min): BP007401.D\data.ms (-128) (-)
 4-Chloroaniline
 Concen: 47.028 ng
 RT: 10.922 min Scan# 1332
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC050


Tgt Ion:	127	Resp:	848813
Ion	Ratio	Lower	Upper
127	100		
129	31.8	25.1	37.7
65	32.2	25.5	38.3
92	20.5	16.5	24.7



Scan 1367 (11.128 min): BP007401.D\data.ms (-131) (-)
 Hexachlorobutadiene
 Concen: 48.004 ng
 RT: 11.122 min Scan# 1366
 Delta R.T. -0.005 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58

Tgt Ion:	225	Resp:	439525
Ion	Ratio	Lower	Upper
225	100		
223	63.5	49.4	74.0
227	63.6	51.0	76.6









Caprolactam
Concen: 29.784 ng
RT: 11.693 min Scan# 1463
Delta R.T. 0.006 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58




Instrument :
BNA_P
ClientSampleId :
SSTDIC050


Tgt Ion:113 Resp: 128289
Ion Ratio Lower Upper
113 100
55 167.8 143.9 183.9
56 126.6 101.6 141.6



4-Chloro-3-methylphenol
Concen: 52.537 ng
RT: 12.051 min Scan# 1524
Delta R.T. 0.006 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion:107 Resp: 669932
Ion Ratio Lower Upper
107 100
144 25.5 20.5 30.7
142 76.8 63.0 94.4









2-Methylnaphthalene
Concen: 50.385 ng
RT: 12.434 min Scan# 1589
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58




Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	142	Resp:	1375786
Ion	Ratio	Lower	Upper
142	100		
141	89.1	71.4	107.0
115	37.5	29.5	44.3




1-Methylnaphthalene
Concen: 49.679 ng
RT: 12.651 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion:	142	Resp:	1331790
Ion	Ratio	Lower	Upper
142	100		
141	92.1	74.4	111.6
116	4.1	3.4	5.0






Abundance Scan 1958 (14.604 min): BP007401.D\data.ms (-1907) (-)




Acenaphthene-d10
Concen: 20.000 ng
RT: 14.604 min Scan# 1958
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion	Ratio	Lower	Upper
164	100		
162	100.3	81.1	121.7
160	44.5	36.0	54.0







Abundance Scan 1651 (12.798 min): BP007401.D\data.ms (-1649) (-)



1,2,4,5-Tetrachlorobenzene
Concen: 42.003 ng
RT: 12.799 min Scan# 1651
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion	Ratio	Lower	Upper
216	100		
214	79.0	62.2	93.4
179	21.3	17.1	25.7
108	19.5	15.4	23.0









Hexachlorocyclopentadiene
Concen: 83.861 ng
RT: 12.787 min Scan# 1649
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58




Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	237	Resp:	425453
Ion	Ratio	Lower	Upper
237	100		
235	62.5	41.2	81.2
272	12.5	0.0	31.9



2,4,6-Tribromophenol
Concen: 61.655 ng
RT: 16.092 min Scan# 2211
Delta R.T. 0.006 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58


Tgt Ion:	330	Resp:	537863
Ion	Ratio	Lower	Upper
330	100		
332	96.6	78.9	118.3
141	37.5	30.7	46.1





Abundance Scan 1691 (13.033 min): BP007401.D\data.ms (-1640) (-)

2,4,6-Trichlorophenol
Concen: 38.300 ng
RT: 13.034 min Scan# 1691
Delta R.T. 0.001 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58


Instrument :
BNA_P
ClientSampleId :
SSTDIC050



Tgt Ion:	196	Resp:	452877
Ion	Ratio	Lower	Upper
196	100		
198	96.6	76.4	114.6
200	31.1	25.0	37.4




Abundance Scan 1691 (13.034 min): BP007402.D\data.ms (-1640) (-)





Abundance Scan 1702 (13.098 min): BP007401.D\data.ms (-1651) (-)


2,4,5-Trichlorophenol
Concen: 34.676 ng
RT: 13.104 min Scan# 1703
Delta R.T. 0.006 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58




Tgt Ion:	196	Resp:	443949
Ion	Ratio	Lower	Upper
196	100		
198	97.2	77.1	115.7
97	49.1	40.9	61.3
132	25.9	21.2	31.8



Abundance Scan 1703 (13.104 min): BP007402.D\data.ms (-1651) (-)





Abundance Scan 1725 (13.233 min): BP007401.D\data.ms (-1715) (-)




2-Fluorobiphenyl
Concen: 82.646 ng
RT: 13.234 min Scan# 1725
Delta R.T. 0.001 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050


Tgt Ion:	172	Resp:	3145575
Ion	Ratio	Lower	Upper
172	100		
171	35.2	28.2	42.2
170	23.3	19.0	28.4



Abundance Scan 1725 (13.234 min): BP007402.D\data.ms (-1674) (-)





Abundance Scan 1760 (13.439 min): BP007401.D\data.ms (-1756) (-)





1,1'-Biphenyl
Concen: 44.094 ng
RT: 13.440 min Scan# 1760
Delta R.T. 0.001 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion:	154	Resp:	1760591
Ion	Ratio	Lower	Upper
154	100		
153	40.5	19.9	59.9
76	12.4	0.0	32.4



Abundance Scan 1760 (13.440 min): BP007402.D\data.ms (-1709) (-)









2-Chloronaphthalene
Concen: 43.638 ng
RT: 13.481 min Scan# 1767
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58




Instrument :
BNA_P
ClientSampleId :
SSTDIC050


Tgt Ion:	162	Resp:	1359302
Ion	Ratio	Lower	Upper
162	100		
127	39.6	31.0	46.6
164	32.9	26.4	39.6



2-Nitroaniline
Concen: 45.103 ng
RT: 13.669 min Scan# 1799
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion:	65	Resp:	363765
Ion	Ratio	Lower	Upper
65	100		
92	70.5	55.9	83.9
138	105.2	86.3	129.5









Acenaphthylene
Concen: 45.769 ng
RT: 14.322 min Scan# 1910
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58




Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	152	Resp:	2200607
Ion	Ratio	Lower	Upper
152	100		
151	20.2	15.8	23.6
153	13.1	10.2	15.2



Dimethylphthalate
Concen: 43.505 ng
RT: 14.063 min Scan# 1866
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58


Tgt Ion:	163	Resp:	1782894
Ion	Ratio	Lower	Upper
163	100		
194	4.0	3.3	4.9
164	10.2	8.2	12.2






Abundance Scan 1884 (14.169 min): BP007401.D\data.ms (-1833) (-)

2,6-Dinitrotoluene
Concen: 41.533 ng
RT: 14.169 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050




Tgt Ion:	165	Resp:	354328
Ion	Ratio	Lower	Upper
165	100		
63	58.0	44.2	66.2
89	59.7	46.6	70.0






Abundance Scan 1968 (14.663 min): BP007401.D\data.ms (-1917) (-)


Acenaphthene
Concen: 47.734 ng
RT: 14.663 min Scan# 1968
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58



Tgt Ion:	154	Resp:	1354714
Ion	Ratio	Lower	Upper
154	100		
153	112.2	90.4	135.6
152	53.0	42.3	63.5






Abundance Scan 1939 (14.492 min): BP007401.D\data.ms (-1953)




3-Nitroaniline
Concen: 44.994 ng
RT: 14.493 min Scan# 1939
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:138 Resp: 393715
Ion Ratio Lower Upper
138 100
108 14.4 12.3 18.5
92 121.0 94.9 142.3





Abundance Scan 1974 (14.698 min): BP007401.D\data.ms (-1954)





2,4-Dinitrophenol
Concen: 25.273 ng
RT: 14.698 min Scan# 1974
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion:184 Resp: 120821
Ion Ratio Lower Upper
184 100
63 57.4 44.0 66.0
154 65.9 48.9 73.3



Abundance Scan 1974 (14.698 min): BP007402.D\data.ms (-1923)









Dibenzenofuran
Concen: 43.278 ng
RT: 14.998 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58




Instrument :
BNA_P
ClientSampleId :
SSTDIC050


Tgt Ion:	168	Resp:	2132074
Ion	Ratio	Lower	Upper
168	100		
139	39.0	30.8	46.2
169	13.4	11.0	16.6



4-Nitrophenol
Concen: 45.879 ng
RT: 14.793 min Scan# 1990
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion:	139	Resp:	314835
Ion	Ratio	Lower	Upper
139	100		
109	82.4	60.9	100.9
65	101.5	76.5	116.5









Scan 2017 (14.951 min): BP007401.D\data.ms (-2057) (-)
 2,4-Dinitrotoluene
 Concen: 39.746 ng
 RT: 14.957 min Scan# 2018
 Delta R.T. 0.006 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC050

Tgt Ion:	165	Resp:	463903
Ion	Ratio	Lower	Upper
165	100		
63	43.2	37.4	56.0
89	72.9	61.8	92.6
182	2.9	2.1	3.1




Scan 2136 (15.651 min): BP007401.D\data.ms (-2158) (-)
 Fluorene
 Concen: 39.669 ng
 RT: 15.651 min Scan# 2136
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58

Tgt Ion:	166	Resp:	1512821
Ion	Ratio	Lower	Upper
166	100		
165	97.6	78.9	118.3
167	13.3	10.9	16.3






Abundance Scan 2063 (15.221 min): BP007401.D\data.ms (-2059) (-)




2,3,4,6-Tetrachlorophenol
Concen: 37.429 ng
RT: 15.140 min Scan# 2049
Delta R.T. -0.082 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleID :
SSTDIC050

Tgt Ion	Ratio	Lower	Upper
232	100		
131	40.1	30.3	45.5
130	2.4	1.8	2.8
166	25.6	19.4	29.0






Abundance Scan 2098 (15.427 min): BP007401.D\data.ms (-2059) (-)



Diethylphthalate
Concen: 44.148 ng
RT: 15.434 min Scan# 2099
Delta R.T. 0.006 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58


Tgt Ion	Ratio	Lower	Upper
149	100		
177	22.6	17.8	26.8
150	12.4	10.2	15.2






Abundance Scan 2135 (15.645 min): BP007401.D\data.ms (-2184) (-)

4-Chlorophenyl-phenylether
Concen: 37.033 ng
RT: 15.645 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050




Tgt Ion	Ratio	Lower	Upper
204	100		
206	33.0	25.8	38.8
141	64.1	48.8	73.2






Abundance Scan 2137 (15.657 min): BP007401.D\data.ms (-2184) (-)

4-Nitroaniline
Concen: 37.803 ng
RT: 15.657 min Scan# 2137
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58




Tgt Ion	Ratio	Lower	Upper
138	100		
92	54.0	30.7	70.7
108	96.0	71.2	111.2





Abundance Scan 2185 (15.939 min): BP007401.D\data.ms (-2185) (-)

Azobenzene
Concen: 50.617 ng
RT: 15.940 min Scan# 2185
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58


Instrument :
BNA_P
ClientSampleId :
SSTDIC050



Tgt Ion:	77	Resp:	1687771
Ion	Ratio	Lower	Upper
77	100		
182	25.3	6.5	46.5
105	20.8	1.3	41.3
51	31.8	10.5	50.5





Abundance Scan 2185 (15.940 min): BP007402.D\data.ms (-2134) (-)




Abundance Scan 2425 (17.351 min): BP007401.D\data.ms (-2425) (-)

Phenanthrene-d10
Concen: 20.000 ng
RT: 17.351 min Scan# 2425
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58


Tgt Ion:	188	Resp:	1106937
Ion	Ratio	Lower	Upper
188	100		
94	9.2	7.1	10.7
80	9.9	7.8	11.6



Abundance Scan 2425 (17.351 min): BP007402.D\data.ms (-2374) (-)






Abundance Scan 2147 (15.716 min): BP007401.D\data.ms (-2168) (-)




4,6-Dinitro-2-methylphenol
Concen: 31.175 ng
RT: 15.722 min Scan# 2148
Delta R.T. 0.006 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	198	Resp:	237720
Ion	Ratio	Lower	Upper
198	100		
51	41.1	23.8	63.8
105	41.0	24.6	64.6






Abundance Scan 2171 (15.857 min): BP007401.D\data.ms (-2168) (-)



n-Nitrosodiphenylamine
Concen: 41.840 ng
RT: 15.857 min Scan# 2171
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58


Tgt Ion:	169	Resp:	1468374
Ion	Ratio	Lower	Upper
169	100		
168	65.4	52.7	79.1
167	35.6	29.0	43.6





Abundance Scan 2288 (16.545 min): BP007401.D\data.ms (-2287) (-)

4-Bromophenyl-phenylether
Concen: 36.983 ng
RT: 16.545 min Scan# 2288
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58


Instrument :
BNA_P
ClientSampleId :
SSTDIC050



Tgt Ion:	248	Resp:	528366
Ion	Ratio	Lower	Upper
248	100		
250	97.7	78.2	117.2
141	75.7	60.8	91.2




Abundance Scan 2288 (16.545 min): BP007402.D\data.ms (-2237) (-)





Abundance Scan 2307 (16.657 min): BP007401.D\data.ms (-2298) (-)


Hexachlorobenzene
Concen: 37.186 ng
RT: 16.663 min Scan# 2308
Delta R.T. 0.006 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58




Tgt Ion:	284	Resp:	582996
Ion	Ratio	Lower	Upper
284	100		
142	34.6	30.6	46.0
249	32.3	26.6	40.0



Abundance Scan 2308 (16.663 min): BP007402.D\data.ms (-2256) (-)









Atrazine
Concen: 41.553 ng
RT: 16.816 min Scan# 2334
Delta R.T. 0.006 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58




Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion	Ratio	Lower	Upper
200	100		
173	25.6	6.2	46.2
215	48.5	28.2	68.2




Pentachlorophenol
Concen: 42.574 ng
RT: 16.998 min Scan# 2365
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion	Ratio	Lower	Upper
266	100		
268	63.0	49.3	73.9
264	63.6	48.7	73.1





Abundance Scan 2432 (17.392 min): BP007401.D\data.ms (-2470) (-)




Phenanthrene
Concen: 45.167 ng
RT: 17.398 min Scan# 2433
Delta R.T. 0.006 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050


Tgt Ion:	178	Resp:	2640262
Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.4	23.0
179	15.4	12.3	18.5



Abundance Scan 2433 (17.398 min): BP007402.D\data.ms (-2381) (-)





Abundance Scan 2448 (17.486 min): BP007401.D\data.ms (-2470) (-)





Anthracene
Concen: 42.578 ng
RT: 17.486 min Scan# 2448
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion:	178	Resp:	2630180
Ion	Ratio	Lower	Upper
178	100		
176	18.6	15.0	22.6
179	15.2	12.2	18.4



Abundance Scan 2448 (17.486 min): BP007402.D\data.ms (-2397) (-)









Carbazole
 Concen: 43.770 ng
 RT: 17.751 min Scan# 2493
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050

Tgt Ion:	167	Resp:	2634494
Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.3	25.9
139	13.2	10.4	15.6



Di-n-butylphthalate
 Concen: 44.768 ng
 RT: 18.316 min Scan# 2589
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58


Tgt Ion:	149	Resp:	3126991
Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.6	11.4
104	5.6	4.2	6.4





Abundance Scan 2766 (19.357 min): BP007401.D\data.ms (-2766) (-)

Fluoranthene
Concen: 42.311 ng
RT: 19.357 min Scan# 2766
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58


Instrument :
BNA_P
ClientSampleId :
SSTDIC050



Tgt Ion:	202	Resp:	3189648
Ion	Ratio	Lower	Upper
202	100		
101	10.6	0.0	30.4
203	17.8	0.0	37.6





Abundance Scan 2766 (19.357 min): BP007402.D\data.ms (-2766) (-)




Abundance Scan 3120 (21.439 min): BP007401.D\data.ms (-3120) (-)

Chrysene-d12
Concen: 20.000 ng
RT: 21.439 min Scan# 3120
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion:	240	Resp:	1053328
Ion	Ratio	Lower	Upper
240	100		
120	10.2	7.8	11.8
236	25.9	20.6	30.8




Abundance Scan 3120 (21.439 min): BP007402.D\data.ms (-3069) (-)






Abundance Scan 2795 (19.527 min): BP007401.D\data.ms (-2777) (-)

Benzidine
Concen: 43.937 ng
RT: 19.533 min Scan# 2796
Delta R.T. 0.006 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050







Tgt Ion:	184	Resp:	1343163
Ion	Ratio	Lower	Upper
184	100		
185	13.9	11.2	16.8
183	11.6	9.1	13.7




Abundance Scan 2826 (19.709 min): BP007401.D\data.ms (-2874) (-)

Pyrene
Concen: 45.028 ng
RT: 19.710 min Scan# 2826
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion:	202	Resp:	3266086
Ion	Ratio	Lower	Upper
202	100		
200	21.4	16.9	25.3
203	17.6	13.8	20.6









Scan 2860 (19.909 min): BP007401.D\data.ms (-2860) (-)
 Terphenyl-d14
 Concen: 75.395 ng
 RT: 19.910 min Scan# 2860
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC050

Tgt Ion:	244	Resp:	4502621
Ion	Ratio	Lower	Upper
244	100		
212	8.0	6.2	9.2
122	9.8	7.8	11.8




Scan 2976 (20.592 min): BP007401.D\data.ms (-2976) (-)
 Butylbenzylphthalate
 Concen: 50.336 ng
 RT: 20.592 min Scan# 2976
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58

Tgt Ion:	149	Resp:	1410464
Ion	Ratio	Lower	Upper
149	100		
91	70.5	56.6	85.0
206	21.7	17.4	26.2






Abundance Scan 3117 (21.421 min): BP007401.D\data.ms (-3117) (-)




Benzo(a)anthracene
Concen: 47.438 ng
RT: 21.422 min Scan# 3117
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	228	Resp:	3227964
Ion	Ratio	Lower	Upper
228	100		
226	27.2	21.8	32.6
229	19.8	15.7	23.5







Abundance Scan 3105 (21.351 min): BP007401.D\data.ms (-3105) (-)



3,3'-Dichlorobenzidine
Concen: 45.062 ng
RT: 21.351 min Scan# 3105
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion:	252	Resp:	1103394
Ion	Ratio	Lower	Upper
252	100		
254	64.7	51.5	77.3
126	10.8	8.5	12.7









Chrysene
 Concen: 47.075 ng
 RT: 21.475 min Scan# 3126
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050

Tgt Ion:	228	Resp:	3015067
Ion	Ratio	Lower	Upper
228	100		
226	30.3	24.6	37.0
229	19.9	15.7	23.5




Bis(2-ethylhexyl)phthalate
 Concen: 50.664 ng
 RT: 21.363 min Scan# 3107
 Delta R.T. 0.000 min
 Lab File: BP007402.D
 Acq: 13 Oct 2021 12:58

Tgt Ion:	149	Resp:	1975811
Ion	Ratio	Lower	Upper
149	100		
167	27.8	21.6	32.4
279	4.1	3.3	4.9






Abundance Scan 3269 (22.315 min): BP007401.D\data.ms (-3235) (-)




Di-n-octyl phthalate
Concen: 53.955 ng
RT: 22.316 min Scan# 3269
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	149	Resp:	3587857
Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	9.2	7.3	10.9






Abundance Scan 3537 (23.892 min): BP007401.D\data.ms (-3528) (-)




Perylene-d12
Concen: 20.000 ng
RT: 23.892 min Scan# 3537
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion:	264	Resp:	1192798
Ion	Ratio	Lower	Upper
264	100		
260	23.5	19.1	28.7
265	21.8	16.7	25.1






Abundance Scan 3975 (26.468 min): BP007401.D\data.ms (-3927) (-)




Indeno(1,2,3-cd)pyrene
Concen: 48.924 ng
RT: 26.468 min Scan# 3975
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	276	Resp:	3994247
Ion	Ratio	Lower	Upper
276	100		
138	23.6	19.4	29.0
277	25.3	20.2	30.2







Abundance Scan 3410 (23.145 min): BP007401.D\data.ms (-3408) (-)



Benzo(b)fluoranthene
Concen: 41.943 ng
RT: 23.145 min Scan# 3410
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Tgt Ion:	252	Resp:	3257343
Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.4	26.2
125	9.2	7.4	11.0









Benzo(k)fluoranthene
Concen: 42.024 ng
RT: 23.198 min Scan# 3419
Delta R.T. 0.006 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58




Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:252 Resp: 3230737
Ion Ratio Lower Upper
252 100
253 21.9 17.6 26.4
125 9.0 7.4 11.0



Benzo(a)pyrene
Concen: 47.334 ng
RT: 23.786 min Scan# 3519
Delta R.T. 0.006 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58


Tgt Ion:252 Resp: 2854289
Ion Ratio Lower Upper
252 100
253 21.6 17.6 26.4
125 10.1 8.4 12.6






Abundance Scan 3979 (26.491 min): BP007401.D\data.ms (-3998) (-)

Dibenzo(a,h)anthracene
Concen: 47.916 ng
RT: 26.492 min Scan# 3979
Delta R.T. 0.000 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050




Tgt Ion:	278	Resp:	3275051
Ion	Ratio	Lower	Upper
278	100		
139	14.8	11.8	17.6
279	24.0	18.6	27.8



Abundance Scan 4108 (27.250 min): BP007401.D\data.ms (-4098) (-)

Benzo(g,h,i)perylene
Concen: 52.517 ng
RT: 27.257 min Scan# 4109
Delta R.T. 0.006 min
Lab File: BP007402.D
Acq: 13 Oct 2021 12:58



Tgt Ion:	276	Resp:	3358096
Ion	Ratio	Lower	Upper
276	100		
277	23.7	19.4	29.0
138	20.8	16.2	24.4

