

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101321\
 Data File : BP007405.D
 Acq On : 13 Oct 2021 14:56
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP101321

Quant Time: Oct 13 15:27:07 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP101321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 14:54:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.969	152	192716	20.000	ng	0.00
21) Naphthalene-d8	10.775	136	749034	20.000	ng	0.00
39) Acenaphthene-d10	14.604	164	480434	20.000	ng	0.00
64) Phenanthrene-d10	17.357	188	1021796	20.000	ng	0.00
76) Chrysene-d12	21.439	240	1002679	20.000	ng	0.00
86) Perylene-d12	23.892	264	1137545	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	892209	76.734	ng	0.00
7) Phenol-d6	7.122	99	1208856	77.168	ng	0.00
23) Nitrobenzene-d5	9.128	82	1032673	85.083	ng	0.00
42) 2,4,6-Tribromophenol	16.092	330	402367	80.686	ng	0.00
45) 2-Fluorobiphenyl	13.234	172	2316565	73.702	ng	0.00
79) Terphenyl-d14	19.916	244	3519490	71.093	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.422	88	191517	37.542	ng	99
3) Pyridine	3.816	79	528999	38.401	ng	99
4) n-Nitrosodimethylamine	3.728	42	237304	39.434	ng	98
6) Aniline	7.287	93	680766	37.835	ng	99
8) 2-Chlorophenol	7.528	128	461693	38.424	ng	98
9) Benzaldehyde	7.099	77	354300	41.254	ng	96
10) Phenol	7.152	94	582875	38.337	ng	100
11) bis(2-Chloroethyl)ether	7.387	93	472180	38.464	ng	98
12) 1,3-Dichlorobenzene	7.857	146	529510	37.857	ng	100
13) 1,4-Dichlorobenzene	8.004	146	533702	37.700	ng	100
14) 1,2-Dichlorobenzene	8.322	146	502889	37.352	ng	99
15) Benzyl Alcohol	8.204	79	449781	39.069	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.504	45	615763	38.371	ng	99
17) 2-Methylphenol	8.404	107	396589	38.531	ng	99
18) Hexachloroethane	9.057	117	194113	38.441	ng	99
19) n-Nitroso-di-n-propyla...	8.781	70	351797	38.557	ng	99
20) 3+4-Methylphenols	8.734	107	546548	39.012	ng	99
22) Acetophenone	8.787	105	701310	37.815	ng	99
24) Nitrobenzene	9.169	77	495165	40.839	ng	97
25) Isophorone	9.698	82	972278	38.459	ng	100
26) 2-Nitrophenol	9.881	139	211075	41.123	ng	98
27) 2,4-Dimethylphenol	9.945	122	396521	38.618	ng	99
28) bis(2-Chloroethoxy)met...	10.187	93	622308	37.762	ng	99
29) 2,4-Dichlorophenol	10.416	162	444860	39.920	ng	99
30) 1,2,4-Trichlorobenzene	10.640	180	498739	38.269	ng	99
31) Naphthalene	10.828	128	1404816	37.595	ng	99
32) Benzoic acid	10.063	122	255184	40.696	ng	98
33) 4-Chloroaniline	10.928	127	607294	37.984	ng	98
34) Hexachlorobutadiene	11.128	225	308526	37.524	ng	97
35) Caprolactam	11.687	113	96381	42.514	ng	96
36) 4-Chloro-3-methylphenol	12.051	107	487521	39.521	ng	100
37) 2-Methylnaphthalene	12.434	142	992868	37.826	ng	99
38) 1-Methylnaphthalene	12.651	142	959758	37.637	ng	99
40) 1,2,4,5-Tetrachloroben...	12.804	216	527727	37.428	ng	100
41) Hexachlorocyclopentadiene	12.792	237	303952	39.542	ng	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101321\
 Data File : BP007405.D
 Acq On : 13 Oct 2021 14:56
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP101321

Quant Time: Oct 13 15:27:07 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP101321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 14:54:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	13.034	196	342056	40.363	ng	99
44) 2,4,5-Trichlorophenol	13.104	196	339516	40.285	ng	98
46) 1,1'-Biphenyl	13.439	154	1295108	37.332	ng	100
47) 2-Chloronaphthalene	13.481	162	999759	37.479	ng	100
48) 2-Nitroaniline	13.675	65	275380	41.987	ng	100
49) Acenaphthylene	14.328	152	1609172	37.907	ng	100
50) Dimethylphthalate	14.063	163	1319250	38.117	ng	100
51) 2,6-Dinitrotoluene	14.175	165	263610	41.735	ng	97
52) Acenaphthene	14.669	154	1000452	37.929	ng	100
53) 3-Nitroaniline	14.498	138	293473	45.035	ng	96
54) 2,4-Dinitrophenol	14.698	184	93847	47.506	ng	91
55) Dibenzofuran	15.004	168	1581798	37.684	ng	99
56) 4-Nitrophenol	14.792	139	241704	46.209	ng	99
57) 2,4-Dinitrotoluene	14.957	165	346310	41.435	ng	94
58) Fluorene	15.651	166	1144913	41.526	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.145	232	320200	38.843	ng	96
60) Diethylphthalate	15.433	149	1328758	37.993	ng	99
61) 4-Chlorophenyl-phenyle...	15.651	204	594678	41.203	ng	100
62) 4-Nitroaniline	15.657	138	274882	43.934	ng	99
63) Azobenzene	15.945	77	1257054	37.562	ng	100
65) 4,6-Dinitro-2-methylph...	15.722	198	176102	39.995	ng	95
66) n-Nitrosodiphenylamine	15.863	169	1106494	37.885	ng	99
67) 4-Bromophenyl-phenylether	16.545	248	385510	36.962	ng	98
68) Hexachlorobenzene	16.663	284	433712	37.728	ng	98
69) Atrazine	16.816	200	407305	38.565	ng	99
70) Pentachlorophenol	17.004	266	270822	42.279	ng	98
71) Phenanthrene	17.398	178	1982842	37.551	ng	100
72) Anthracene	17.492	178	1969698	37.362	ng	99
73) Carbazole	17.757	167	1993597	37.787	ng	99
74) Di-n-butylphthalate	18.322	149	2335803	38.381	ng	100
75) Fluoranthene	19.363	202	2433718	38.021	ng	100
77) Benzidine	19.533	184	1166655	40.412	ng	100
78) Pyrene	19.710	202	2484737	37.401	ng	99
80) Butylbenzylphthalate	20.592	149	1065192	40.528	ng	98
81) Benzo(a)anthracene	21.421	228	2467542	37.503	ng	99
82) 3,3'-Dichlorobenzidine	21.351	252	861625	39.913	ng	100
83) Chrysene	21.480	228	2331807	37.513	ng	99
84) Bis(2-ethylhexyl)phtha...	21.368	149	1536752	38.979	ng	98
85) Di-n-octyl phthalate	22.315	149	2746686	40.373	ng	100
87) Indeno(1,2,3-cd)pyrene	26.468	276	3045635	37.927	ng	100
88) Benzo(b)fluoranthene	23.145	252	2562896	37.996	ng	99
89) Benzo(k)fluoranthene	23.198	252	2404317	36.403	ng	99
90) Benzo(a)pyrene	23.786	252	2169124	37.812	ng	99
91) Dibenzo(a,h)anthracene	26.498	278	2505214	37.817	ng	98
92) Benzo(g,h,i)perylene	27.262	276	2547881	38.076	ng	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.969 min Scan# 830
 Delta R.T. 0.000 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56




Instrument :
 BNA_P
 ClientSampleId :
 ICVBP101321


Tgt Ion:	152	Resp:	192716
Ion	Ratio	Lower	Upper
152	100		
150	159.6	124.1	186.1
115	62.4	48.1	72.1



1,4-Dioxane
 Concen: 37.542 ng
 RT: 3.422 min Scan# 57
 Delta R.T. 0.000 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56

Tgt Ion:	88	Resp:	191517
Ion	Ratio	Lower	Upper
88	100		
58	71.8	58.1	87.1
43	28.7	22.2	33.4









Pyridine
Concen: 38.401 ng
RT: 3.816 min Scan# 124
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56




Instrument :
BNA_P
ClientSampleId :
ICVBP101321


Tgt Ion:	79	Resp:	528999
Ion	Ratio	Lower	Upper
79	100		
52	68.8	55.7	83.5
51	35.5	28.5	42.7



n-Nitrosodimethylamine
Concen: 39.434 ng
RT: 3.728 min Scan# 109
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion:	42	Resp:	237304
Ion	Ratio	Lower	Upper
42	100		
74	131.4	107.0	160.4
44	7.8	7.4	11.2









Scan 415 (5.528 min): BP007401.D\data.ms (-404)
 2-Fluorophenol
 Concen: 76.734 ng
 RT: 5.528 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56




Instrument :
 BNA_P
 ClientSampleId :
 ICVBP101321


Tgt Ion:	112	Resp:	892209
Ion	Ratio	Lower	Upper
112	100		
64	57.9	46.6	70.0
63	30.8	25.0	37.6



Scan 714 (7.287 min): BP007401.D\data.ms (-702)
 Aniline
 Concen: 37.835 ng
 RT: 7.287 min Scan# 714
 Delta R.T. 0.000 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56

Tgt Ion:	93	Resp:	680766
Ion	Ratio	Lower	Upper
93	100		
66	39.2	30.5	45.7
65	19.8	15.8	23.8









Phenol-d6
 Concen: 77.168 ng
 RT: 7.122 min Scan# 686
 Delta R.T. 0.000 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56




Instrument :
 BNA_P
 ClientSampleId :
 ICVBP101321


Tgt Ion:	99	Resp:	1208856
Ion	Ratio	Lower	Upper
99	100		
42	18.3	14.9	22.3
71	36.1	29.2	43.8



2-Chlorophenol
 Concen: 38.424 ng
 RT: 7.528 min Scan# 755
 Delta R.T. 0.000 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56

Tgt Ion:	128	Resp:	461693
Ion	Ratio	Lower	Upper
128	100		
130	31.2	12.0	52.0
64	48.5	27.0	67.0









Scan 682 (7.098 min): BP007401.D\data.ms (-673)
 Benzaldehyde
 Concen: 41.254 ng
 RT: 7.099 min Scan# 682
 Delta R.T. 0.001 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56




Instrument :
 BNA_P
 ClientSampleId :
 ICVBP101321

Tgt Ion:	77	Resp:	354300
Ion	Ratio	Lower	Upper
77	100		
105	102.3	77.5	117.5
106	94.6	72.1	112.1




Scan 691 (7.151 min): BP007401.D\data.ms (-679)
 Phenol
 Concen: 38.337 ng
 RT: 7.152 min Scan# 691
 Delta R.T. 0.001 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56

Tgt Ion:	94	Resp:	582875
Ion	Ratio	Lower	Upper
94	100		
65	28.1	8.2	48.2
66	39.1	18.8	58.8






Abundance Scan 731 (7.387 min): BP007401.D\data.ms (-724) #11




bis(2-Chloroethyl)ether
Concen: 38.464 ng
RT: 7.387 min Scan# 731
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321

Tgt Ion	Ratio	Lower	Upper
93	100		
63	77.5	55.7	95.7
95	31.8	12.7	52.7






Abundance Scan 811 (7.857 min): BP007401.D\data.ms (-802) #12




1,3-Dichlorobenzene
Concen: 37.857 ng
RT: 7.857 min Scan# 811
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion	Ratio	Lower	Upper
146	100		
148	64.3	51.4	77.0
75	32.6	26.1	39.1






Abundance Scan 836 (8.004 min): BP007401.D\data.ms (-825) #13




1,4-Dichlorobenzene
Concen: 37.700 ng
RT: 8.004 min Scan# 836
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321

Tgt Ion	Ratio	Lower	Upper
146	100		
148	63.9	51.0	76.6
111	41.3	32.8	49.2







Abundance Scan 890 (8.322 min): BP007401.D\data.ms (-879) #14



1,2-Dichlorobenzene
Concen: 37.352 ng
RT: 8.322 min Scan# 890
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion	Ratio	Lower	Upper
146	100		
148	64.1	51.5	77.3
111	43.4	35.1	52.7









Benzyl Alcohol
Concen: 39.069 ng
RT: 8.204 min Scan# 870
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56




Instrument :
BNA_P
ClientSampleId :
ICVBP101321


Tgt Ion:	79	Resp:	449781
Ion	Ratio	Lower	Upper
79	100		
108	75.6	59.9	89.9
77	64.3	51.1	76.7



2,2'-oxybis(1-Chloropropane)
Concen: 38.371 ng
RT: 8.504 min Scan# 921
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion:	45	Resp:	615763
Ion	Ratio	Lower	Upper
45	100		
77	15.4	0.0	34.9
79	11.5	0.0	31.7









2-Methylphenol
Concen: 38.531 ng
RT: 8.404 min Scan# 904
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56




Instrument :
BNA_P
ClientSampleId :
ICVBP101321

Tgt Ion:	107	Resp:	396589
Ion	Ratio	Lower	Upper
107	100		
108	112.0	89.2	133.8
77	46.9	38.5	57.7
79	48.1	38.4	57.6



Hexachloroethane
Concen: 38.441 ng
RT: 9.057 min Scan# 1015
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56


Tgt Ion:	117	Resp:	194113
Ion	Ratio	Lower	Upper
117	100		
119	95.4	77.6	116.4
201	103.3	82.9	124.3





Abundance Scan 967 (8.775 min): BP007401.D\data.ms (-960) #19

n-Nitroso-di-n-propylamine
Concen: 38.557 ng
RT: 8.781 min Scan# 968
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56


Instrument :
BNA_P
ClientSampleId :
ICVBP101321



Tgt Ion:	70	Resp:	351797
Ion	Ratio	Lower	Upper
70	100		
42	57.9	46.2	69.2
101	10.4	7.7	11.5
130	21.0	16.2	24.4




Abundance Scan 968 (8.781 min): BP007405.D\data.ms (-916) (-)





Abundance Scan 960 (8.734 min): BP007401.D\data.ms (-950) #20


3+4-Methylphenols
Concen: 39.012 ng
RT: 8.734 min Scan# 960
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56



Tgt Ion:	107	Resp:	546548
Ion	Ratio	Lower	Upper
107	100		
108	89.3	69.2	109.2
77	35.8	13.8	53.8
79	30.1	10.2	50.2




Abundance Scan 960 (8.734 min): BP007405.D\data.ms (-909) (-)





Abundance Scan 1307 (10.775 min): BP007401.D\data.ms (-1256) (-)

Naphthalene-d8
Concen: 20.000 ng
RT: 10.775 min Scan# 1307
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56


Instrument :
BNA_P
ClientSampleId :
ICVBP101321



Tgt Ion	Ratio	Lower	Upper
136	100		
137	11.2	8.6	13.0
54	7.9	6.2	9.4
68	6.0	4.7	7.1




Abundance Scan 1307 (10.775 min): BP007405.D\data.ms (-1256) (-)





Abundance Scan 969 (8.787 min): BP007401.D\data.ms (-962) (-)


Acetophenone
Concen: 37.815 ng
RT: 8.787 min Scan# 969
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56




Tgt Ion	Ratio	Lower	Upper
105	100		
71	6.6	4.5	6.7
51	27.6	21.8	32.8
120	22.3	17.6	26.4



Abundance Scan 969 (8.787 min): BP007405.D\data.ms (-918) (-)






Abundance Scan 1026 (9.122 min): BP007401.D\data.ms (-1023)




Nitrobenzene-d5
Concen: 85.083 ng
RT: 9.128 min Scan# 1027
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321

Tgt Ion:	82	Resp:	1032673
Ion	Ratio	Lower	Upper
82	100		
128	41.5	32.6	48.8
54	49.6	40.4	60.6






Abundance Scan 1033 (9.163 min): BP007401.D\data.ms (-1024)




Nitrobenzene
Concen: 40.839 ng
RT: 9.169 min Scan# 1034
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion:	77	Resp:	495165
Ion	Ratio	Lower	Upper
77	100		
123	46.8	35.3	52.9
65	14.0	10.9	16.3





Abundance Scan 1124 (9.698 min): BP007401.D\data.ms (-1125)




Isophorone
Concen: 38.459 ng
RT: 9.698 min Scan# 1124
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321


Tgt Ion:	82	Resp:	972278
Ion	Ratio	Lower	Upper
82	100		
95	7.5	6.2	9.2
138	17.3	14.0	21.0



Abundance Scan 1124 (9.698 min): BP007405.D\data.ms (-1073) (-)





Abundance Scan 1155 (9.881 min): BP007401.D\data.ms (-11426)




2-Nitrophenol
Concen: 41.123 ng
RT: 9.881 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion:	139	Resp:	211075
Ion	Ratio	Lower	Upper
139	100		
109	32.3	26.2	39.2
65	52.1	40.3	60.5




Abundance Scan 1155 (9.881 min): BP007405.D\data.ms (-1104) (-)






Abundance Scan 1166 (9.945 min): BP007401.D\data.ms (-1157) (-)

2,4-Dimethylphenol
Concen: 38.618 ng
RT: 9.945 min Scan# 1166
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321




Tgt Ion:	122	Resp:	396521
Ion	Ratio	Lower	Upper
122	100		
107	118.4	93.8	140.6
121	59.2	47.0	70.4






Abundance Scan 1206 (10.181 min): BP007401.D\data.ms (-1123) (-)


bis(2-Chloroethoxy)methane
Concen: 37.762 ng
RT: 10.187 min Scan# 1207
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56



Tgt Ion:	93	Resp:	622308
Ion	Ratio	Lower	Upper
93	100		
95	33.0	26.1	39.1
123	11.6	8.5	12.7






Abundance Scan 1246 (10.416 min): BP007401.D\data.ms (-1229) (-)




2,4-Dichlorophenol
Concen: 39.920 ng
RT: 10.416 min Scan# 1246
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321

Tgt Ion:	162	Resp:	444860
Ion	Ratio	Lower	Upper
162	100		
164	63.8	44.0	84.0
98	38.1	16.3	56.3







Abundance Scan 1284 (10.639 min): BP007401.D\data.ms (-1230) (-)



1,2,4-Trichlorobenzene
Concen: 38.269 ng
RT: 10.640 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion:	180	Resp:	498739
Ion	Ratio	Lower	Upper
180	100		
182	97.2	78.4	117.6
145	30.7	23.8	35.8









Naphthalene
 Concen: 37.595 ng
 RT: 10.828 min Scan# 1316
 Delta R.T. 0.006 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56




Instrument :
 BNA_P
 ClientSampleId :
 ICVBP101321


Tgt Ion:	128	Resp:	1404816
Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.9	13.3
127	13.1	10.2	15.4



Benzoic acid
 Concen: 40.696 ng
 RT: 10.063 min Scan# 1186
 Delta R.T. 0.000 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56

Tgt Ion:	122	Resp:	255184
Ion	Ratio	Lower	Upper
122	100		
105	129.6	105.9	145.9
77	93.6	74.6	114.6









Scan 1332 (10.922 min): BP007401.D\data.ms (-128) (-)
 4-Chloroaniline
 Concen: 37.984 ng
 RT: 10.928 min Scan# 1333
 Delta R.T. 0.006 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56




Instrument :
 BNA_P
 ClientSampleId :
 ICVBP101321


Tgt Ion	Ratio	Lower	Upper
127	100		
129	31.9	25.1	37.7
65	30.6	25.5	38.3
92	20.1	16.5	24.7



Scan 1367 (11.128 min): BP007401.D\data.ms (-131) (-)
 Hexachlorobutadiene
 Concen: 37.524 ng
 RT: 11.128 min Scan# 1367
 Delta R.T. 0.000 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56

Tgt Ion	Ratio	Lower	Upper
225	100		
223	64.3	49.4	74.0
227	65.2	51.0	76.6









Caprolactam
Concen: 42.514 ng
RT: 11.687 min Scan# 1462
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56




Instrument :
BNA_P
ClientSampleId :
ICVBP101321


Tgt Ion:	113	Resp:	96381
Ion	Ratio	Lower	Upper
113	100		
55	156.6	143.9	183.9
56	120.3	101.6	141.6



4-Chloro-3-methylphenol
Concen: 39.521 ng
RT: 12.051 min Scan# 1524
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion:	107	Resp:	487521
Ion	Ratio	Lower	Upper
107	100		
144	25.8	20.5	30.7
142	78.8	63.0	94.4









2-Methylnaphthalene
Concen: 37.826 ng
RT: 12.434 min Scan# 1589
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56




Instrument :
BNA_P
ClientSampleId :
ICVBP101321


Tgt Ion:	142	Resp:	992868
Ion	Ratio	Lower	Upper
142	100		
141	89.3	71.4	107.0
115	37.8	29.5	44.3



1-Methylnaphthalene
Concen: 37.637 ng
RT: 12.651 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion:	142	Resp:	959758
Ion	Ratio	Lower	Upper
142	100		
141	91.5	74.4	111.6
116	4.0	3.4	5.0









Scan 1958 (14.604 min): BP007401.D\data.ms (-1907) (-)
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.604 min Scan# 1958
 Delta R.T. 0.000 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56




Instrument :
 BNA_P
 ClientSampleId :
 ICVBP101321


Tgt Ion:	164	Resp:	480434
Ion Ratio	Lower	Upper	
164	100		
162	98.9	81.1	121.7
160	45.2	36.0	54.0



Scan 1651 (12.798 min): BP007401.D\data.ms (-1640) (-)
 1,2,4,5-Tetrachlorobenzene
 Concen: 37.428 ng
 RT: 12.804 min Scan# 1652
 Delta R.T. 0.006 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56

Tgt Ion:	216	Resp:	527727
Ion Ratio	Lower	Upper	
216	100		
214	78.3	62.2	93.4
179	21.2	17.1	25.7
108	19.2	15.4	23.0









Hexachlorocyclopentadiene
Concen: 39.542 ng
RT: 12.792 min Scan# 1650
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56




Instrument :
BNA_P
ClientSampleId :
ICVBP101321

Tgt Ion	Ratio	Lower	Upper
237	100		
235	60.9	41.2	81.2
272	11.7	0.0	31.9



2,4,6-Tribromophenol
Concen: 80.686 ng
RT: 16.092 min Scan# 2211
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56


Tgt Ion	Ratio	Lower	Upper
330	100		
332	96.9	78.9	118.3
141	37.8	30.7	46.1





Abundance Scan 1691 (13.033 min): BP007401.D\data.ms (-1640) (-)

2,4,6-Trichlorophenol
Concen: 40.363 ng
RT: 13.034 min Scan# 1691
Delta R.T. 0.001 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56


Instrument :
BNA_P
ClientSampleId :
ICVBP101321



Tgt Ion:	196	Resp:	342056
Ion	Ratio	Lower	Upper
196	100		
198	94.9	76.4	114.6
200	30.9	25.0	37.4




Abundance Scan 1691 (13.034 min): BP007405.D\data.ms (-1640) (-)





Abundance Scan 1702 (13.098 min): BP007401.D\data.ms (-1641) (-)


2,4,5-Trichlorophenol
Concen: 40.285 ng
RT: 13.104 min Scan# 1703
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56




Tgt Ion:	196	Resp:	339516
Ion	Ratio	Lower	Upper
196	100		
198	98.3	77.1	115.7
97	49.8	40.9	61.3
132	27.3	21.2	31.8



Abundance Scan 1703 (13.104 min): BP007405.D\data.ms (-1651) (-)









2-Fluorobiphenyl
 Concen: 73.702 ng
 RT: 13.234 min Scan# 1725
 Delta R.T. 0.001 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56




Instrument :
 BNA_P
 ClientSampleId :
 ICVBP101321


Tgt Ion:	172	Resp:	2316565
Ion	Ratio	Lower	Upper
172	100		
171	35.3	28.2	42.2
170	23.6	19.0	28.4



1,1'-Biphenyl
 Concen: 37.332 ng
 RT: 13.439 min Scan# 1760
 Delta R.T. 0.000 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56

Tgt Ion:	154	Resp:	1295108
Ion	Ratio	Lower	Upper
154	100		
153	39.8	19.9	59.9
76	12.5	0.0	32.4









2-Chloronaphthalene
Concen: 37.479 ng
RT: 13.481 min Scan# 1767
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56




Instrument :
BNA_P
ClientSampleId :
ICVBP101321

Tgt Ion:	162	Resp:	999759
Ion	Ratio	Lower	Upper
162	100		
127	39.3	31.0	46.6
164	33.0	26.4	39.6




2-Nitroaniline
Concen: 41.987 ng
RT: 13.675 min Scan# 1800
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion:	65	Resp:	275380
Ion	Ratio	Lower	Upper
65	100		
92	69.7	55.9	83.9
138	107.4	86.3	129.5






Abundance Scan 1910 (14.322 min): BP007401.D\data.ms (-1859) (-)




Acenaphthylene
Concen: 37.907 ng
RT: 14.328 min Scan# 1911
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion:	152	Resp:	1609172
Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.8	23.6
153	12.7	10.2	15.2

Instrument :
BNA_P
ClientSampleId :
ICVBP101321







Abundance Scan 1866 (14.063 min): BP007401.D\data.ms (-1859) (-)



Dimethylphthalate
Concen: 38.117 ng
RT: 14.063 min Scan# 1866
Delta R.T. -0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion:	163	Resp:	1319250
Ion	Ratio	Lower	Upper
163	100		
194	3.8	3.3	4.9
164	10.2	8.2	12.2









2,6-Dinitrotoluene
Concen: 41.735 ng
RT: 14.175 min Scan# 1885
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56




Instrument :
BNA_P
ClientSampleId :
ICVBP101321

Tgt Ion:	165	Resp:	263610
Ion	Ratio	Lower	Upper
165	100		
63	53.7	44.2	66.2
89	56.0	46.6	70.0




Acenaphthene
Concen: 37.929 ng
RT: 14.669 min Scan# 1969
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion:	154	Resp:	1000452
Ion	Ratio	Lower	Upper
154	100		
153	112.7	90.4	135.6
152	52.7	42.3	63.5






Abundance Scan 1939 (14.492 min): BP007401.D\data.ms (-1950)




3-Nitroaniline
Concen: 45.035 ng
RT: 14.498 min Scan# 1940
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321

Tgt Ion:	138	Resp:	293473
Ion	Ratio	Lower	Upper
138	100		
108	13.3	12.3	18.5
92	113.9	94.9	142.3







Abundance Scan 1974 (14.698 min): BP007401.D\data.ms (-1950)



2,4-Dinitrophenol
Concen: 47.506 ng
RT: 14.698 min Scan# 1974
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion:	184	Resp:	93847
Ion	Ratio	Lower	Upper
184	100		
63	59.1	44.0	66.0
154	69.9	48.9	73.3









Dibenzenofuran
Concen: 37.684 ng
RT: 15.004 min Scan# 2026
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56




Instrument :
BNA_P
ClientSampleId :
ICVBP101321

Tgt Ion:	168	Resp:	1581798
Ion	Ratio	Lower	Upper
168	100		
139	38.5	30.8	46.2
169	13.1	11.0	16.6



4-Nitrophenol
Concen: 46.209 ng
RT: 14.792 min Scan# 1990
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56


Tgt Ion:	139	Resp:	241704
Ion	Ratio	Lower	Upper
139	100		
109	80.2	60.9	100.9
65	98.3	76.5	116.5






Abundance Scan 2017 (14.951 min): BP007401.D\data.ms (-2057) (-)

2,4-Dinitrotoluene
Concen: 41.435 ng
RT: 14.957 min Scan# 2018
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321




Tgt Ion:	165	Resp:	346310
Ion	Ratio	Lower	Upper
165	100		
63	43.2	37.4	56.0
89	71.8	61.8	92.6
182	3.0	2.1	3.1






Abundance Scan 2136 (15.651 min): BP007401.D\data.ms (-2158) (-)


Fluorene
Concen: 41.526 ng
RT: 15.651 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56



Tgt Ion:	166	Resp:	1144913
Ion	Ratio	Lower	Upper
166	100		
165	97.7	78.9	118.3
167	13.6	10.9	16.3





Abundance Scan 2063 (15.221 min): BP007401.D\data.ms (-2050) (-)




2,3,4,6-Tetrachlorophenol
Concen: 38.843 ng
RT: 15.145 min Scan# 2050
Delta R.T. -0.076 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321


Tgt Ion	Ratio	Lower	Upper
232	100		
131	40.4	30.3	45.5
130	2.5	1.8	2.8
166	25.7	19.4	29.0



Abundance Scan 2050 (15.145 min): BP007405.D\data.ms (-2012) (-)





Abundance Scan 2098 (15.427 min): BP007401.D\data.ms (-2050) (-)




Diethylphthalate
Concen: 37.993 ng
RT: 15.433 min Scan# 2099
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion	Ratio	Lower	Upper
149	100		
177	21.9	17.8	26.8
150	12.5	10.2	15.2





Abundance Scan 2099 (15.433 min): BP007405.D\data.ms (-2047) (-)





Abundance Scan 2135 (15.645 min): BP007401.D\data.ms (-2185) (-)

4-Chlorophenyl-phenylether
Concen: 41.203 ng
RT: 15.651 min Scan# 2136
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321





Tgt Ion	Ratio	Lower	Upper
204	100		
206	32.9	25.8	38.8
141	61.0	48.8	73.2





Abundance Scan 2137 (15.657 min): BP007401.D\data.ms (-2186) (-)


4-Nitroaniline
Concen: 43.934 ng
RT: 15.657 min Scan# 2137
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56



Tgt Ion	Ratio	Lower	Upper
138	100		
92	52.3	30.7	70.7
108	91.4	71.2	111.2



Abundance Scan 2185 (15.939 min): BP007401.D\data.ms (-2185) (-)





Azobenzene
Concen: 37.562 ng
RT: 15.945 min Scan# 2186
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56


Instrument :
BNA_P
ClientSampleId :
ICVBP101321

Tgt Ion: 77 Resp: 1257054


Ion	Ratio	Lower	Upper
77	100		
182	26.3	6.5	46.5
105	21.3	1.3	41.3
51	30.2	10.5	50.5



Abundance Scan 2186 (15.945 min): BP007405.D\data.ms (-2186) (-)





Abundance Scan 2425 (17.351 min): BP007401.D\data.ms (-2425) (-)




Phenanthrene-d10
Concen: 20.000 ng
RT: 17.357 min Scan# 2426
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion: 188 Resp: 1021796


Ion	Ratio	Lower	Upper
188	100		
94	8.7	7.1	10.7
80	9.7	7.8	11.6



Abundance Scan 2426 (17.357 min): BP007405.D\data.ms (-2426) (-)





Abundance Scan 2147 (15.716 min): BP007401.D\data.ms (-2168) (-)




4,6-Dinitro-2-methylphenol
Concen: 39.995 ng
RT: 15.722 min Scan# 2148
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321


Tgt Ion:	198	Resp:	176102
Ion	Ratio	Lower	Upper
198	100		
51	41.3	23.8	63.8
105	41.0	24.6	64.6



Abundance Scan 2148 (15.722 min): BP007405.D\data.ms (-2096) (-)





Abundance Scan 2171 (15.857 min): BP007401.D\data.ms (-2168) (-)




n-Nitrosodiphenylamine
Concen: 37.885 ng
RT: 15.863 min Scan# 2172
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion:	169	Resp:	1106494
Ion	Ratio	Lower	Upper
169	100		
168	65.3	52.7	79.1
167	35.6	29.0	43.6




Abundance Scan 2172 (15.863 min): BP007405.D\data.ms (-2120) (-)






Abundance Scan 2288 (16.545 min): BP007401.D\data.ms (-2287) (-)

4-Bromophenyl-phenylether
Concen: 36.962 ng
RT: 16.545 min Scan# 2288
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321




Tgt Ion:	248	Resp:	385510
Ion	Ratio	Lower	Upper
248	100		
250	98.5	78.2	117.2
141	78.2	60.8	91.2






Abundance Scan 2307 (16.657 min): BP007401.D\data.ms (-2298) (-)


Hexachlorobenzene
Concen: 37.728 ng
RT: 16.663 min Scan# 2308
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56



Tgt Ion:	284	Resp:	433712
Ion	Ratio	Lower	Upper
284	100		
142	37.1	30.6	46.0
249	32.2	26.6	40.0






Abundance Scan 2333 (16.810 min): BP007401.D\data.ms (-2333) (-)




Atrazine
Concen: 38.565 ng
RT: 16.816 min Scan# 2334
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321

Tgt Ion	Ratio	Lower	Upper
200	100		
173	26.5	6.2	46.2
215	47.6	28.2	68.2






Abundance Scan 2365 (16.998 min): BP007401.D\data.ms (-2365) (-)




Pentachlorophenol
Concen: 42.279 ng
RT: 17.004 min Scan# 2366
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion	Ratio	Lower	Upper
266	100		
268	62.7	49.3	73.9
264	62.2	48.7	73.1





Abundance Scan 2432 (17.392 min): BP007401.D\data.ms (-2470) (-)




Phenanthrene
Concen: 37.551 ng
RT: 17.398 min Scan# 2433
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321


Tgt Ion:	178	Resp:	1982842
Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.4	23.0
179	15.4	12.3	18.5



Abundance Scan 2433 (17.398 min): BP007405.D\data.ms (-2381) (-)





Abundance Scan 2448 (17.486 min): BP007401.D\data.ms (-2470) (-)




Anthracene
Concen: 37.362 ng
RT: 17.492 min Scan# 2449
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56


Tgt Ion:	178	Resp:	1969698
Ion	Ratio	Lower	Upper
178	100		
176	18.6	15.0	22.6
179	15.0	12.2	18.4



Abundance Scan 2449 (17.492 min): BP007405.D\data.ms (-2397) (-)





Abundance Scan 2493 (17.751 min): BP007401.D\data.ms (-2442) (-)




Carbazole
Concen: 37.787 ng
RT: 17.757 min Scan# 2494
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321


Tgt Ion:	167	Resp:	1993597
Ion	Ratio	Lower	Upper
167	100		
166	21.3	17.3	25.9
139	12.7	10.4	15.6



Abundance Scan 2494 (17.757 min): BP007405.D\data.ms (-2442) (-)





Abundance Scan 2589 (18.315 min): BP007401.D\data.ms (-2571) (-)




Di-n-butylphthalate
Concen: 38.381 ng
RT: 18.322 min Scan# 2590
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56


Tgt Ion:	149	Resp:	2335803
Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.6	11.4
104	5.3	4.2	6.4



Abundance Scan 2590 (18.322 min): BP007405.D\data.ms (-2538) (-)






Abundance Scan 2766 (19.357 min): BP007401.D\data.ms (-2766) (-)




Fluoranthene
Concen: 38.021 ng
RT: 19.363 min Scan# 2767
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321

Tgt Ion:202 Resp: 2433718
Ion Ratio Lower Upper
202 100
101 10.1 0.0 30.4
203 17.5 0.0 37.6






Abundance Scan 3120 (21.439 min): BP007401.D\data.ms (-3120) (-)




Chrysene-d12
Concen: 20.000 ng
RT: 21.439 min Scan# 3120
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion:240 Resp: 1002679
Ion Ratio Lower Upper
240 100
120 10.4 7.8 11.8
236 25.8 20.6 30.8





Abundance Scan 2795 (19.527 min): BP007401.D\data.ms (-2777) (-)




Benzidine
Concen: 40.412 ng
RT: 19.533 min Scan# 2796
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321


Tgt Ion:184 Resp: 1166655
Ion Ratio Lower Upper
184 100
185 14.2 11.2 16.8
183 11.5 9.1 13.7



Abundance Scan 2796 (19.533 min): BP007405.D\data.ms (-2744) (-)





Abundance Scan 2826 (19.709 min): BP007401.D\data.ms (-2874) (-)




Pyrene
Concen: 37.401 ng
RT: 19.710 min Scan# 2826
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56


Tgt Ion:202 Resp: 2484737
Ion Ratio Lower Upper
202 100
200 21.5 16.9 25.3
203 18.1 13.8 20.6



Abundance Scan 2826 (19.710 min): BP007405.D\data.ms (-2775) (-)






Abundance Scan 2860 (19.909 min): BP007401.D\data.ms (-2809) (-)




Terphenyl-d14
Concen: 71.093 ng
RT: 19.916 min Scan# 2861
Delta R.T. 0.007 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321

Tgt Ion:244 Resp: 3519490
Ion Ratio Lower Upper
244 100
212 7.8 6.2 9.2
122 9.7 7.8 11.8






Abundance Scan 2976 (20.592 min): BP007401.D\data.ms (-2969) (-)




Butylbenzylphthalate
Concen: 40.528 ng
RT: 20.592 min Scan# 2976
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion:149 Resp: 1065192
Ion Ratio Lower Upper
149 100
91 72.4 56.6 85.0
206 21.0 17.4 26.2






Abundance Scan 3117 (21.421 min): BP007401.D\data.ms (-3181) (-)




Benzo(a)anthracene
Concen: 37.503 ng
RT: 21.421 min Scan# 3117
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321

Tgt Ion:	228	Resp:	2467542
Ion	Ratio	Lower	Upper
228	100		
226	27.7	21.8	32.6
229	20.1	15.7	23.5







Abundance Scan 3105 (21.351 min): BP007401.D\data.ms (-3082) (-)



3,3'-Dichlorobenzidine
Concen: 39.913 ng
RT: 21.351 min Scan# 3105
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion:	252	Resp:	861625
Ion	Ratio	Lower	Upper
252	100		
254	64.2	51.5	77.3
126	10.9	8.5	12.7









Chrysene
 Concen: 37.513 ng
 RT: 21.480 min Scan# 3127
 Delta R.T. 0.006 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56




Instrument :
 BNA_P
 ClientSampleId :
 ICVBP101321

Tgt Ion:	228	Resp:	2331807
Ion	Ratio	Lower	Upper
228	100		
226	30.4	24.6	37.0
229	19.6	15.7	23.5




Bis(2-ethylhexyl)phthalate
 Concen: 38.979 ng
 RT: 21.368 min Scan# 3108
 Delta R.T. 0.006 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56

Tgt Ion:	149	Resp:	1536752
Ion	Ratio	Lower	Upper
149	100		
167	28.3	21.6	32.4
279	4.5	3.3	4.9





Abundance Scan 3269 (22.315 min): BP007401.D\data.ms (-3235) (-)




Di-n-octyl phthalate
Concen: 40.373 ng
RT: 22.315 min Scan# 3269
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321


Tgt Ion:	149	Resp:	2746686
Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	9.0	7.3	10.9



Abundance Scan 3269 (22.315 min): BP007405.D\data.ms (-3218) (-)





Abundance Scan 3537 (23.892 min): BP007401.D\data.ms (-3528) (-)




Perylene-d12
Concen: 20.000 ng
RT: 23.892 min Scan# 3537
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56


Tgt Ion:	264	Resp:	1137545
Ion	Ratio	Lower	Upper
264	100		
260	24.4	19.1	28.7
265	21.8	16.7	25.1



Abundance Scan 3537 (23.892 min): BP007405.D\data.ms (-3486) (-)





Abundance Scan 3975 (26.468 min): BP007401.D\data.ms (-3927) (-)




Indeno(1,2,3-cd)pyrene
Concen: 37.927 ng
RT: 26.468 min Scan# 3975
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321


Tgt Ion	Ratio	Lower	Upper
276	100		
138	23.9	19.4	29.0
277	25.2	20.2	30.2



Abundance Scan 3975 (26.468 min): BP007405.D\data.ms (-3924) (-)





Abundance Scan 3410 (23.145 min): BP007401.D\data.ms (-3408) (-)





Benzo(b)fluoranthene
Concen: 37.996 ng
RT: 23.145 min Scan# 3410
Delta R.T. 0.000 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.7	17.4	26.2
125	9.2	7.4	11.0



Abundance Scan 3410 (23.145 min): BP007405.D\data.ms (-3359) (-)









Benzo(k)fluoranthene
 Concen: 36.403 ng
 RT: 23.198 min Scan# 3419
 Delta R.T. 0.006 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56




Instrument :
 BNA_P
 ClientSampleId :
 ICVBP101321

Tgt Ion:	252	Resp:	2404317
Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.6	26.4
125	8.9	7.4	11.0




Benzo(a)pyrene
 Concen: 37.812 ng
 RT: 23.786 min Scan# 3519
 Delta R.T. 0.006 min
 Lab File: BP007405.D
 Acq: 13 Oct 2021 14:56

Tgt Ion:	252	Resp:	2169124
Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.6	26.4
125	10.0	8.4	12.6





Abundance Scan 3979 (26.491 min): BP007401.D\data.ms (-3998) (-)




Dibenzo(a,h)anthracene
Concen: 37.817 ng
RT: 26.498 min Scan# 3980
Delta R.T. 0.006 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Instrument :
BNA_P
ClientSampleId :
ICVBP101321


Tgt Ion:	278	Resp:	2505214
Ion	Ratio	Lower	Upper
278	100		
139	15.5	11.8	17.6
279	24.6	18.6	27.8



Abundance Scan 3980 (26.498 min): BP007405.D\data.ms (-3928) (-)





Abundance Scan 4108 (27.250 min): BP007401.D\data.ms (-4098) (-)



Benzo(g,h,i)perylene
Concen: 38.076 ng
RT: 27.262 min Scan# 4110
Delta R.T. 0.012 min
Lab File: BP007405.D
Acq: 13 Oct 2021 14:56

Tgt Ion:	276	Resp:	2547881
Ion	Ratio	Lower	Upper
276	100		
277	23.9	19.4	29.0
138	21.2	16.2	24.4



Abundance Scan 4110 (27.262 min): BP007405.D\data.ms (-4057) (-)

