

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101321\
 Data File : BP007406.D
 Acq On : 13 Oct 2021 15:30
 Operator : CG/JU
 Sample : PB139656BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB139656BL

Quant Time: Oct 13 16:28:39 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP101321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 14:54:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.963	152	185049	20.00	ng	0.00
21) Naphthalene-d8	10.769	136	713519	20.00	ng	0.00
39) Acenaphthene-d10	14.598	164	466030	20.00	ng	0.00
64) Phenanthrene-d10	17.351	188	1032423	20.00	ng	0.00
76) Chrysene-d12	21.433	240	1133524	20.00	ng	0.00
86) Perylene-d12	23.886	264	1192307	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.534	112	1516010	135.79	ng	0.00
7) Phenol-d6	7.122	99	1885660	125.36	ng	0.00
23) Nitrobenzene-d5	9.122	82	1141733	98.75	ng	0.00
42) 2,4,6-Tribromophenol	16.087	330	670199	138.55	ng	0.00
45) 2-Fluorobiphenyl	13.228	172	2624037	86.06	ng	0.00
79) Terphenyl-d14	19.910	244	4264517	76.20	ng	0.00

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101321\
Data File : BP007406.D
Acq On : 13 Oct 2021 15:30
Operator : CG/JU
Sample : PB139656BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PB139656BL

Quant Time: Oct 13 16:28:39 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP101321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 13 14:54:20 2021
Response via : Initial Calibration

