


2,4-Dinitrotoluene

Response Ratio



$$\text{Response} = 4.655\text{e-}001 * \text{Amt} - 8.281\text{e-}002$$


Coef of Det (r^2) = 0.997154 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP101420.M

Calibration Table Last Updated: Wed Oct 14 17:03:23 2020

Benzaldehyde

Response Ratio




$$\text{Response} = 7.883\text{e-}001 * \text{Amt} + 1.965\text{e-}001$$

Coef of Det (r^2) = 0.992336 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP101420.M

Calibration Table Last Updated: Wed Oct 14 17:03:23 2020


Benzoic acid



$\text{Response} = 1.655\text{e-}001 * \text{Amt} - 5.310\text{e-}002$
 Coef of Det (r^2) = 0.995508 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP101420.M
 Calibration Table Last Updated: Wed Oct 14 17:03:23 2020

2-Nitroaniline

Response Ratio




$$\text{Response} = 4.008\text{e-}001 * \text{Amt} - 8.383\text{e-}002$$

Coef of Det (r^2) = 0.995541 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP101420.M

Calibration Table Last Updated: Wed Oct 14 17:03:23 2020


2,4-Dinitrophenol



$R = 1.680e-002 A^2 + 8.602e-002 A - 9.232e-003$
 Coef of Det (r^2) = 0.998581 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP101420.M
 Calibration Table Last Updated: Wed Oct 14 17:03:23 2020

2,4-Dinitrotoluene

Response Ratio



$$\text{Response} = 4.655\text{e-}001 * \text{Amt} - 8.281\text{e-}002$$

Coef of Det (r^2) = 0.997154 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP101420.M

Calibration Table Last Updated: Wed Oct 14 17:03:23 2020