

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101421\
 Data File : BP007435.D
 Acq On : 15 Oct 2021 08:39
 Operator : CG/JU
 Sample : PB139790BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK790

Quant Time: Oct 18 00:59:03 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP101421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 14 13:04:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.963	152	237067	20.00	ng/ul	0.00
20) Naphthalene-d8	10.769	136	951656	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.604	164	595032	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.357	188	1225539	20.00	ng/ul	0.00
79) Chrysene-d12	21.445	240	1245715	20.00	ng/ul	0.00
88) Perylene-d12	23.904	264	1264061	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.387	96	39341	6.55	ng/uL	0.00
4) Pyridine-d5	3.799	84	547436	33.32	ng/ul	0.00
7) Phenol-d5	7.116	99	661887	35.75	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.287	67	392806	35.57	ng/ul	0.00
11) 2-Chlorophenol-d4	7.493	132	524666	36.02	ng/ul	0.00
15) 4-Methylphenol-d8	8.663	113	526678	36.93	ng/ul	0.00
21) Nitrobenzene-d5	9.116	128	251351	41.74	ng/ul	0.00
24) 2-Nitrophenol-d4	9.846	143	252112	44.82	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.387	165	500220	36.46	ng/ul	0.00
31) 4-Chloroaniline-d4	10.899	131	713474	38.45	ng/ul	0.00
46) Dimethylphthalate-d6	14.010	166	1500985	36.46	ng/ul	0.00
49) Acenaphthylene-d8	14.293	160	1897751	36.68	ng/ul	0.00
54) 4-Nitrophenol-d4	14.781	143	239328	36.23	ng/ul	0.00
60) Fluorene-d10	15.598	176	1273806	36.97	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.704	200	138146	28.13	ng/ul	0.00
73) Anthracene-d10	17.457	188	1980716	38.06	ng/ul	0.00
81) Pyrene-d10	19.686	212	2245217	36.05	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.745	264	2222673	35.34	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101421\
Data File : BP007435.D
Acq On : 15 Oct 2021 08:39
Operator : CG/JU
Sample : PB139790BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK790

Quant Time: Oct 18 00:59:03 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP101421.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 14 13:04:52 2021
Response via : Initial Calibration

