

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101422\
 Data File : BP012127.D
 Acq On : 14 Oct 2022 08:57
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Oct 14 23:00:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 23:03:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.846	152	319033	20.000	ng/ul	0.00
20) Naphthalene-d8	10.652	136	1402609	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.498	164	813231	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.257	188	1434353	20.000	ng/ul	0.00
79) Chrysene-d12	21.345	240	1315208	20.000	ng/ul	0.00
88) Perylene-d12	23.769	264	1420338	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.276	96	57262	7.077	ng/uL	0.00
4) Pyridine-d5	3.693	84	408306	18.974	ng/ul	0.00
7) Phenol-d5	7.017	99	494841	19.286	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.181	67	287533	19.742	ng/ul	0.00
11) 2-Chlorophenol-d4	7.375	132	403348	19.804	ng/ul	0.00
15) 4-Methylphenol-d8	8.563	113	386562	19.614	ng/ul	0.00
21) Nitrobenzene-d5	9.016	128	198224	20.011	ng/ul	0.00
24) 2-Nitrophenol-d4	9.734	143	207040	20.612	ng/ul	-0.01
28) 2,4-Dichlorophenol-d3	10.275	165	383326	19.692	ng/ul	0.00
31) 4-Chloroaniline-d4	10.799	131	535962	18.683	ng/ul	0.00
46) Dimethylphthalate-d6	13.910	166	962116	18.681	ng/ul	0.00
49) Acenaphthylene-d8	14.187	160	1206362	18.942	ng/ul	0.00
54) 4-Nitrophenol-d4	14.710	143	159022	15.727	ng/ul	0.00
60) Fluorene-d10	15.493	176	850082	18.243	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.616	200	124339	16.417	ng/ul	0.00
73) Anthracene-d10	17.357	188	1161504	18.648	ng/ul	0.00
81) Pyrene-d10	19.592	212	1192022	17.691	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.610	264	1277434	18.501	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.311	88	59645	7.005	ng/uL	95
5) Pyridine	3.717	79	393867	19.102	ng/ul	96
6) Benzaldehyde	6.987	77	267920	22.255	ng/ul	99
8) Phenol	7.040	94	519521	19.287	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.275	93	419882	19.757	ng/ul	99
12) 2-Chlorophenol	7.411	128	434072	19.638	ng/ul	99
13) 2-Methylphenol	8.299	108	394209	19.621	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.369	45	441314	20.259	ng/ul	98
16) Acetophenone	8.675	105	612409	20.155	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.664	70	281749	20.359	ng/ul	98
18) 4-Methylphenol	8.622	108	430059	19.860	ng/ul	95
19) Hexachloroethane	8.922	117	169397	19.853	ng/ul	96
22) Nitrobenzene	9.058	77	443972	19.551	ng/ul	97
23) Isophorone	9.587	82	799009	19.837	ng/ul	99
25) 2-Nitrophenol	9.769	139	237858	20.156	ng/ul	97
26) 2,4-Dimethylphenol	9.828	107	441588	19.214	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.069	93	534547	19.339	ng/ul	99
29) 2,4-Dichlorophenol	10.305	162	399854	19.546	ng/ul	99
30) Naphthalene	10.705	128	1383577	18.667	ng/ul	100
32) 4-Chloroaniline	10.822	127	555674	18.788	ng/ul	98
33) Hexachlorobutadiene	10.981	225	238479	18.799	ng/ul	98
34) Caprolactam	11.616	113	95770	17.036	ng/ul	98
35) 4-Chloro-3-methylphenol	11.946	107	386741	19.955	ng/ul	97
36) 2-Methylnaphthalene	12.316	142	908632	19.038	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101422\
 Data File : BP012127.D
 Acq On : 14 Oct 2022 08:57
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Oct 14 23:00:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 23:03:21 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.534	142	910370	19.047	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.681	216	461567	18.567	ng/ul	100
40) Hexachlorocyclopentadiene	12.657	237	182044	15.077	ng/ul	99
41) 2,4,6-Trichlorophenol	12.928	196	285083	19.038	ng/ul	98
42) 2,4,5-Trichlorophenol	12.999	196	304609	18.717	ng/ul	98
43) 1,1'-Biphenyl	13.328	154	1134714	18.483	ng/ul	100
44) 2-Chloronaphthalene	13.369	162	912086	18.604	ng/ul	99
45) 2-Nitroaniline	13.581	65	202933	19.404	ng/ul	95
47) Dimethylphthalate	13.957	163	1019750	18.793	ng/ul	100
48) 2,6-Dinitrotoluene	14.081	165	191161	19.898	ng/ul	99
50) Acenaphthylene	14.216	152	1348138	18.708	ng/ul	99
51) 3-Nitroaniline	14.410	138	201882	18.047	ng/ul	98
52) Acenaphthene	14.563	153	945492	18.430	ng/ul	99
53) 2,4-Dinitrophenol	14.622	184	76291	12.292	ng/ul	97
55) 4-Nitrophenol	14.722	109	111858	16.002	ng/ul	98
56) Dibenzofuran	14.898	168	1263705	18.227	ng/ul	99
57) 2,4-Dinitrotoluene	14.869	165	264297	19.142	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.128	232	235085	18.533	ng/ul	95
59) Diethylphthalate	15.322	149	953128	19.179	ng/ul	99
61) Fluorene	15.545	166	1008852	18.537	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.545	204	491573	18.671	ng/ul	97
63) 4-Nitroaniline	15.575	138	183314	17.946	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.634	198	137717	16.820	ng/ul	96
67) N-Nitrosodiphenylamine	15.763	169	821419	19.348	ng/ul	100
68) 4-Bromophenyl-phenylether	16.440	248	270790	19.401	ng/ul	99
69) Hexachlorobenzene	16.551	284	301842	18.632	ng/ul	98
70) Atrazine	16.722	200	241998	18.536	ng/ul	99
71) Pentachlorophenol	16.904	266	161980	17.066	ng/ul	99
72) Phenanthrene	17.298	178	1452504	18.621	ng/ul	99
74) Anthracene	17.392	178	1442698	18.690	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.293	216	454839	19.624	ng/ul	100
76) Pentachlorobenzene	14.816	250	442802	19.453	ng/ul	98
77) Carbazole	17.663	167	1241633	18.034	ng/ul	100
78) Di-n-butylphthalate	18.210	149	1343699	19.667	ng/ul	99
80) Fluoranthene	19.269	202	1478181	17.904	ng/ul	99
82) Pyrene	19.622	202	1553313	17.778	ng/ul	97
83) Butylbenzylphthalate	20.492	149	528848	19.094	ng/ul	91
84) 3,3'-Dichlorobenzidine	21.269	252	509630	17.584	ng/ul	99
85) Benzo(a)anthracene	21.333	228	1560282	18.847	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.251	149	732405	19.615	ng/ul	99
87) Chrysene	21.386	228	1468497	18.428	ng/ul	99
89) Di-n-octyl phthalate	22.180	149	1283679	18.052	ng/ul	100
90) Benzo(b)fluoranthene	23.022	252	1687902	18.774	ng/ul	98
91) Benzo(k)fluoranthene	23.075	252	1589630	17.975	ng/ul	99
93) Benzo(a)pyrene	23.657	252	1473385	18.369	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.280	276	1790123	17.651	ng/ul	95
95) Dibenzo(a,h)anthracene	26.292	278	1618295	17.786	ng/ul	96
96) Benzo(g,h,i)perylene	27.051	276	1589102	17.378	ng/ul#	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP010422\
 Data File : BP012127.D
 Acq On : 14 Oct 2022 08:57
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :


Quant Time: Oct 14 23:00:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP010122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 23:03:21 2022
 Response via : Initial Calibration








#2
1,4-Dioxane
Concen: 7.005 ng/uL
RT: 3.311 min Scan# 51
Delta R.T. 0.000 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Instrument :
BNA_P
ClientSampleId :







Tgt Ion: 88 Resp: 59645
Ion Ratio Lower Upper
88 100
43 19.0 16.6 25.0
58 62.0 52.6 79.0



#5
Pyridine
Concen: 19.102 ng/uL
RT: 3.717 min Scan# 124
Delta R.T. 0.000 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion: 79 Resp: 393867
Ion Ratio Lower Upper
79 100
52 60.0 51.2 76.8
51 27.4 22.8 34.2









#6
Benzaldehyde
Concen: 22.255 ng/ul
RT: 6.987 min Scan# 680
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 267920
Ion Ratio Lower Upper
77 100
105 105.1 85.2 127.8
106 102.6 83.1 124.7



#8
Phenol
Concen: 19.287 ng/ul
RT: 7.040 min Scan# 689
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57


Tgt Ion: 94 Resp: 519521
Ion Ratio Lower Upper
94 100
65 25.2 19.4 29.0
66 31.9 24.6 37.0








#10
 Bis(2-Chloroethyl)ether
 Concen: 19.757 ng/ul
 RT: 7.275 min Scan# 729
 Delta R.T. -0.006 min
 Lab File: BP012127.D
 Acq: 14 Oct 2022 08:57

Instrument :
 BNA_P
 ClientSampleId :







Tgt Ion: 93 Resp: 419882
 Ion Ratio Lower Upper
 93 100
 63 69.4 56.7 85.1
 95 32.3 26.2 39.2



#12
 2-Chlorophenol
 Concen: 19.638 ng/ul
 RT: 7.411 min Scan# 752
 Delta R.T. -0.006 min
 Lab File: BP012127.D
 Acq: 14 Oct 2022 08:57

Tgt Ion:128 Resp: 434072
 Ion Ratio Lower Upper
 128 100
 64 40.2 32.7 49.1
 130 32.1 26.3 39.5









#13
2-Methylphenol
Concen: 19.621 ng/ul
RT: 8.299 min Scan# 903
Delta R.T. 0.000 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 394209
Ion Ratio Lower Upper
108 100
107 88.5 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.259 ng/ul
RT: 8.369 min Scan# 915
Delta R.T. -0.012 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion: 45 Resp: 441314
Ion Ratio Lower Upper
45 100
77 17.2 14.4 21.6
79 13.0 11.3 16.9









#16
Acetophenone
Concen: 20.155 ng/ul
RT: 8.675 min Scan# 967
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:105 Resp: 612409
Ion Ratio Lower Upper
105 100
77 72.1 57.0 85.4
51 20.9 16.4 24.6



#17
N-Nitroso-di-n-propylamine
Concen: 20.359 ng/ul
RT: 8.664 min Scan# 965
Delta R.T. 0.000 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion: 70 Resp: 281749
Ion Ratio Lower Upper
70 100
42 44.5 34.6 52.0
101 10.2 8.0 12.0
130 23.9 18.3 27.5









#18
4-Methylphenol
Concen: 19.860 ng/ul
RT: 8.622 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 430059
Ion Ratio Lower Upper
108 100
107 113.4 86.3 129.5



#19
Hexachloroethane
Concen: 19.853 ng/ul
RT: 8.922 min Scan# 1009
Delta R.T. -0.012 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion:117 Resp: 169397
Ion Ratio Lower Upper
117 100
201 97.5 82.3 123.5
199 60.9 50.0 75.0









#22
Nitrobenzene
Concen: 19.551 ng/ul
RT: 9.058 min Scan# 1032
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 443972
Ion Ratio Lower Upper
77 100
123 51.4 39.4 59.0
65 12.7 9.9 14.9



#23
Isophorone
Concen: 19.837 ng/ul
RT: 9.587 min Scan# 1122
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion: 82 Resp: 799009
Ion Ratio Lower Upper
82 100
95 6.5 5.4 8.0
138 19.6 15.2 22.8









#25
2-Nitrophenol
Concen: 20.156 ng/ul
RT: 9.769 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:139 Resp: 237858
Ion Ratio Lower Upper
139 100
65 43.8 36.3 54.5
109 24.8 21.2 31.8



#26
2,4-Dimethylphenol
Concen: 19.214 ng/ul
RT: 9.828 min Scan# 1163
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion:107 Resp: 441588
Ion Ratio Lower Upper
107 100
121 53.7 43.0 64.4
122 91.3 71.4 107.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.339 ng/ul
 RT: 10.069 min Scan# 1204
 Delta R.T. -0.006 min
 Lab File: BP012127.D
 Acq: 14 Oct 2022 08:57




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion: 93 Resp: 534547
 Ion Ratio Lower Upper
 93 100
 95 32.5 25.7 38.5
 123 11.7 9.2 13.8



#29
 2,4-Dichlorophenol
 Concen: 19.546 ng/ul
 RT: 10.305 min Scan# 1244
 Delta R.T. -0.006 min
 Lab File: BP012127.D
 Acq: 14 Oct 2022 08:57

Tgt Ion: 162 Resp: 399854
 Ion Ratio Lower Upper
 162 100
 164 64.8 51.5 77.3
 98 33.3 26.3 39.5









#30
Naphthalene
Concen: 18.667 ng/ul
RT: 10.705 min Scan# 1312
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:128 Resp: 1383577
Ion Ratio Lower Upper
128 100
129 10.9 8.9 13.3
127 12.9 10.2 15.4



#32
4-Chloroaniline
Concen: 18.788 ng/ul
RT: 10.822 min Scan# 1332
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion:127 Resp: 555674
Ion Ratio Lower Upper
127 100
129 33.1 25.6 38.4









#33
Hexachlorobutadiene
Concen: 18.799 ng/ul
RT: 10.981 min Scan# 1359
Delta R.T. -0.012 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:225 Resp: 238479
Ion Ratio Lower Upper
225 100
223 65.5 50.9 76.3
227 65.0 51.1 76.7



#34
Caprolactam
Concen: 17.036 ng/ul
RT: 11.616 min Scan# 1467
Delta R.T. 0.000 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion:113 Resp: 95770
Ion Ratio Lower Upper
113 100
55 147.7 115.2 172.8
56 107.7 85.6 128.4









#35
4-Chloro-3-methylphenol
Concen: 19.955 ng/ul
RT: 11.946 min Scan# 1523
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:107 Resp: 386741
Ion Ratio Lower Upper
107 100
144 27.0 23.0 34.4
142 85.2 69.8 104.8



#36
2-Methylnaphthalene
Concen: 19.038 ng/ul
RT: 12.316 min Scan# 1586
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion:142 Resp: 908632
Ion Ratio Lower Upper
142 100
141 89.6 70.7 106.1









#37
1-Methylnaphthalene
Concen: 19.047 ng/ul
RT: 12.534 min Scan# 1623
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:142 Resp: 910370
Ion Ratio Lower Upper
142 100
141 90.8 72.9 109.3
116 3.4 2.8 4.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 18.567 ng/ul
RT: 12.681 min Scan# 1648
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion:216 Resp: 461567
Ion Ratio Lower Upper
216 100
214 78.5 63.0 94.4
179 20.2 16.3 24.5
108 16.5 12.5 18.7









#40
Hexachlorocyclopentadiene
Concen: 15.077 ng/ul
RT: 12.657 min Scan# 1644
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:237 Resp: 182044
Ion Ratio Lower Upper
237 100
235 62.2 49.8 74.8
272 13.5 9.9 14.9



#41
2,4,6-Trichlorophenol
Concen: 19.038 ng/ul
RT: 12.928 min Scan# 1690
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion:196 Resp: 285083
Ion Ratio Lower Upper
196 100
198 97.0 76.2 114.2
200 31.5 24.2 36.2





#42

2,4,5-Trichlorophenol

Concen: 18.717 ng/ul

RT: 12.999 min Scan# 1702

Delta R.T. -0.006 min


Lab File: BP012127.D

Acq: 14 Oct 2022 08:57

Instrument :

BNA_P

ClientSampleId :






Tgt Ion:196 Resp: 304609

Ion Ratio Lower Upper

196 100

198 96.7 75.4 113.2

200 31.2 24.8 37.2



#43

1,1'-Biphenyl

Concen: 18.483 ng/ul

RT: 13.328 min Scan# 1758

Delta R.T. -0.006 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57




Tgt Ion:154 Resp: 1134714


Ion Ratio Lower Upper

154 100

153 40.2 32.2 48.2

76 12.4 10.3 15.5









#44
2-Chloronaphthalene
Concen: 18.604 ng/ul
RT: 13.369 min Scan# 1765
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:162 Resp: 912086
Ion Ratio Lower Upper
162 100
164 33.0 26.0 39.0
127 37.0 29.4 44.2



#45
2-Nitroaniline
Concen: 19.404 ng/ul
RT: 13.581 min Scan# 1801
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion: 65 Resp: 202933
Ion Ratio Lower Upper
65 100
92 73.4 61.6 92.4
138 115.1 96.5 144.7









#47
Dimethylphthalate
Concen: 18.793 ng/ul
RT: 13.957 min Scan# 1865
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:163 Resp: 1019750
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 9.9 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 19.898 ng/ul
RT: 14.081 min Scan# 1886
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion:165 Resp: 191161
Ion Ratio Lower Upper
165 100
89 54.4 43.4 65.0
121 21.1 18.1 27.1









#50
Acenaphthylene
Concen: 18.708 ng/ul
RT: 14.216 min Scan# 1909
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:152 Resp: 1348138
Ion Ratio Lower Upper
152 100
151 20.0 15.7 23.5
153 13.3 10.4 15.6



#51
3-Nitroaniline
Concen: 18.047 ng/ul
RT: 14.410 min Scan# 1942
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion:138 Resp: 201882
Ion Ratio Lower Upper
138 100
108 10.0 8.5 12.7
92 109.7 85.8 128.8









#52
Acenaphthene
Concen: 18.430 ng/ul
RT: 14.563 min Scan# 1968
Delta R.T. 0.000 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:153 Resp: 945492
Ion Ratio Lower Upper
153 100
152 47.6 38.2 57.2
154 87.0 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 12.292 ng/ul
RT: 14.622 min Scan# 1978
Delta R.T. 0.000 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion:184 Resp: 76291
Ion Ratio Lower Upper
184 100
63 56.0 45.5 68.3
154 55.3 47.0 70.6









#55
4-Nitrophenol
Concen: 16.002 ng/ul
RT: 14.722 min Scan# 1995
Delta R.T. 0.000 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:109 Resp: 111858
Ion Ratio Lower Upper
109 100
139 157.4 124.1 186.1
65 132.2 104.2 156.4



#56
Dibenzofuran
Concen: 18.227 ng/ul
RT: 14.898 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion:168 Resp: 1263705
Ion Ratio Lower Upper
168 100
139 35.7 28.2 42.4









#57
2,4-Dinitrotoluene
Concen: 19.142 ng/ul
RT: 14.869 min Scan# 2020
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:165 Resp: 264297
Ion Ratio Lower Upper
165 100
63 37.9 31.1 46.7
182 2.4 1.8 2.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.533 ng/ul
RT: 15.128 min Scan# 2064
Delta R.T. 0.000 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion:232 Resp: 235085
Ion Ratio Lower Upper
232 100
131 43.7 37.8 56.6
130 2.2 1.8 2.6
166 30.6 26.6 39.8









#59
Diethylphthalate
Concen: 19.179 ng/ul
RT: 15.322 min Scan# 2097
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:149 Resp: 953128
Ion Ratio Lower Upper
149 100
177 22.4 17.5 26.3
150 12.3 9.8 14.8



#61
Fluorene
Concen: 18.537 ng/ul
RT: 15.545 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion:166 Resp: 1008852
Ion Ratio Lower Upper
166 100
165 97.4 79.5 119.3
167 13.6 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 18.671 ng/ul
RT: 15.545 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 491573
Ion Ratio Lower Upper
204 100
206 32.9 25.4 38.2
141 55.5 46.2 69.2



#63
4-Nitroaniline
Concen: 17.946 ng/ul
RT: 15.575 min Scan# 2140
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion:138 Resp: 183314
Ion Ratio Lower Upper
138 100
92 49.8 39.2 58.8
108 66.4 55.4 83.0









#66
4,6-Dinitro-2-methylphenol
Concen: 16.820 ng/ul
RT: 15.634 min Scan# 2150
Delta R.T. 0.000 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:198 Resp: 137717
Ion Ratio Lower Upper
198 100
182 5.7 4.9 7.3
77 22.8 20.0 30.0



#67
N-Nitrosodiphenylamine
Concen: 19.348 ng/ul
RT: 15.763 min Scan# 2172
Delta R.T. 0.000 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion:169 Resp: 821419
Ion Ratio Lower Upper
169 100
168 66.8 53.5 80.3
167 36.8 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 19.401 ng/ul
RT: 16.440 min Scan# 2287
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:248 Resp: 270790
Ion Ratio Lower Upper
248 100
250 97.8 77.9 116.9
141 75.0 59.5 89.3



#69
Hexachlorobenzene
Concen: 18.632 ng/ul
RT: 16.551 min Scan# 2306
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion:284 Resp: 301842
Ion Ratio Lower Upper
284 100
142 39.6 30.5 45.7
249 31.3 25.4 38.0





#70

Atrazine

Concen: 18.536 ng/ul

RT: 16.722 min Scan# 2335

Delta R.T. 0.000 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:200 Resp: 241998

Ion Ratio Lower Upper

200 100

173 26.9 22.1 33.1

215 49.5 38.8 58.2



#71

Pentachlorophenol

Concen: 17.066 ng/ul

RT: 16.904 min Scan# 2366

Delta R.T. -0.006 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57




Tgt Ion:266 Resp: 161980


Ion Ratio Lower Upper

266 100

264 63.4 50.6 75.8

268 66.0 51.8 77.6





#72

Phenanthrene

Concen: 18.621 ng/ul

RT: 17.298 min Scan# 2433

Delta R.T. -0.006 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:178 Resp: 1452504

Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.4

176 19.1 15.8 23.6



#74

Anthracene

Concen: 18.690 ng/ul

RT: 17.392 min Scan# 2449

Delta R.T. 0.000 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57




Tgt Ion:178 Resp: 1442698


Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 18.9 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.624 ng/uL
RT: 13.293 min Scan# 1752
Delta R.T. -0.006 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:216 Resp: 454839
Ion Ratio Lower Upper
216 100
214 78.6 62.6 94.0
218 48.6 38.9 58.3



#76
Pentachlorobenzene
Concen: 19.453 ng/uL
RT: 14.816 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BP012127.D
Acq: 14 Oct 2022 08:57

Tgt Ion:250 Resp: 442802
Ion Ratio Lower Upper
250 100
248 61.5 51.4 77.2
252 63.4 50.7 76.1





#77

Carbazole

Concen: 18.034 ng/ul

RT: 17.663 min Scan# 2495

Delta R.T. -0.006 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57





Instrument :

BNA_P

ClientSampleId :

Tgt Ion:167 Resp: 1241633

Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.0	25.6
139	12.7	10.1	15.1



#78

Di-n-butylphthalate

Concen: 19.667 ng/ul

RT: 18.210 min Scan# 2588




Delta R.T. -0.006 min


Lab File: BP012127.D

Acq: 14 Oct 2022 08:57

Tgt Ion:149 Resp: 1343699

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.2	10.8
104	4.7	3.4	5.2





#80

Fluoranthene

Concen: 17.904 ng/ul

RT: 19.269 min Scan# 2768

Delta R.T. 0.000 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:202 Resp: 1478181

Ion Ratio Lower Upper

202 100

101 11.8 9.0 13.6

100 8.8 7.0 10.6



#82

Pyrene

Concen: 17.778 ng/ul

RT: 19.622 min Scan# 2828

Delta R.T. 0.000 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57




Tgt Ion:202 Resp: 1553313


Ion Ratio Lower Upper

202 100

101 13.9 10.4 15.6

100 11.7 8.5 12.7





#83

Butylbenzylphthalate

Concen: 19.094 ng/ul

RT: 20.492 min Scan# 2976

Delta R.T. -0.006 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:149 Resp: 528848

Ion Ratio Lower Upper

149 100

91 68.7 48.5 72.7

206 19.4 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 17.584 ng/ul

RT: 21.269 min Scan# 3108

Delta R.T. 0.000 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57




Tgt Ion:252 Resp: 509630


Ion Ratio Lower Upper

252 100

254 64.7 51.4 77.2

126 11.1 7.8 11.8





#85

Benzo(a)anthracene

Concen: 18.847 ng/ul

RT: 21.333 min Scan# 3119

Delta R.T. 0.000 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 1560282

Ion Ratio Lower Upper

228 100

229 19.7 15.7 23.5

226 27.1 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 19.615 ng/ul

RT: 21.251 min Scan# 3105

Delta R.T. 0.000 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57




Tgt Ion:149 Resp: 732405


Ion Ratio Lower Upper

149 100

167 26.9 21.9 32.9

279 3.7 3.4 5.2





#87

Chrysene

Concen: 18.428 ng/ul

RT: 21.386 min Scan# 3128

Delta R.T. 0.000 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 1468497

Ion Ratio Lower Upper

228 100

226 30.6 24.0 36.0

229 19.7 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 18.052 ng/ul




RT: 22.180 min Scan# 3263


Delta R.T. 0.000 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57

Tgt Ion:149 Resp: 1283679





#90

Benzo(b)fluoranthene

Concen: 18.774 ng/ul

RT: 23.022 min Scan# 3406

Delta R.T. -0.006 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 1687902

Ion Ratio Lower Upper

252 100

253 22.3 17.6 26.4

125 11.4 7.8 11.8



#91

Benzo(k)fluoranthene

Concen: 17.975 ng/ul

RT: 23.075 min Scan# 3415

Delta R.T. 0.000 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57




Tgt Ion:252 Resp: 1589630


Ion Ratio Lower Upper

252 100

253 21.9 17.6 26.4

125 11.2 8.2 12.2





#93

Benzo(a)pyrene

Concen: 18.369 ng/ul

RT: 23.657 min Scan# 3514

Delta R.T. 0.000 min


Lab File: BP012127.D

Acq: 14 Oct 2022 08:57

Instrument :

BNA_P

ClientSampleId :






Tgt Ion:252 Resp: 1473385

Ion Ratio Lower Upper

252 100

253 21.9 17.6 26.4

125 12.8 8.9 13.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 17.651 ng/ul

RT: 26.280 min Scan# 3960

Delta R.T. 0.000 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57




Tgt Ion:276 Resp: 1790123


Ion Ratio Lower Upper

276 100

138 28.4 19.1 28.7

277 24.8 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 17.786 ng/ul

RT: 26.292 min Scan# 3962

Delta R.T. -0.006 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:278 Resp: 1618295

Ion Ratio Lower Upper

278 100

139 18.7 12.5 18.7

279 24.6 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 17.378 ng/ul

RT: 27.051 min Scan# 4091

Delta R.T. -0.006 min

Lab File: BP012127.D

Acq: 14 Oct 2022 08:57

Tgt Ion:276 Resp: 1589102

Ion Ratio Lower Upper

276 100

138 24.8 16.5 24.7#

277 23.5 19.0 28.4

