

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101422\
 Data File : BPCCC144.D
 Acq On : 14 Oct 2022 19:19
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Oct 14 23:07:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 23:03:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.846	152	355769	20.000	ng/ul	0.00
20) Naphthalene-d8	10.651	136	1594900	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.492	164	927226	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.251	188	1594949	20.000	ng/ul	0.00
79) Chrysene-d12	21.345	240	1277756	20.000	ng/ul	0.00
88) Perylene-d12	23.763	264	1428646	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.269	96	51324	5.688	ng/uL	0.00
4) Pyridine-d5	3.693	84	412449	17.187	ng/ul	0.00
7) Phenol-d5	7.016	99	573499	20.043	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.175	67	331314	20.400	ng/ul	-0.01
11) 2-Chlorophenol-d4	7.375	132	460317	20.267	ng/ul	0.00
15) 4-Methylphenol-d8	8.563	113	452449	20.587	ng/ul	0.00
21) Nitrobenzene-d5	9.010	128	229104	20.339	ng/ul	-0.01
24) 2-Nitrophenol-d4	9.734	143	242450	21.228	ng/ul	-0.01
28) 2,4-Dichlorophenol-d3	10.275	165	439986	19.877	ng/ul	0.00
31) 4-Chloroaniline-d4	10.793	131	615593	18.872	ng/ul	-0.01
46) Dimethylphthalate-d6	13.904	166	1117036	19.022	ng/ul	-0.01
49) Acenaphthylene-d8	14.187	160	1397565	19.246	ng/ul	0.00
54) 4-Nitrophenol-d4	14.704	143	180811	15.684	ng/ul	0.00
60) Fluorene-d10	15.492	176	977948	18.407	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.616	200	139273	16.537	ng/ul	0.00
73) Anthracene-d10	17.351	188	1300409	18.776	ng/ul	0.00
81) Pyrene-d10	19.592	212	1192786	18.222	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.604	264	1288594	18.554	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.305	88	54999	5.792	ng/uL	92
5) Pyridine	3.717	79	400117	17.402	ng/ul	92
6) Benzaldehyde	6.987	77	314224	23.406	ng/ul	97
8) Phenol	7.040	94	601482	20.024	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.269	93	480572	20.277	ng/ul	99
12) 2-Chlorophenol	7.405	128	497276	20.174	ng/ul	98
13) 2-Methylphenol	8.293	108	457450	20.418	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.381	45	512115	21.082	ng/ul	99
16) Acetophenone	8.675	105	710201	20.960	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.657	70	333050	21.581	ng/ul	99
18) 4-Methylphenol	8.622	108	494616	20.483	ng/ul	93
19) Hexachloroethane	8.922	117	195551	20.552	ng/ul	96
22) Nitrobenzene	9.052	77	518184	20.068	ng/ul	99
23) Isophorone	9.587	82	946836	20.673	ng/ul	99
25) 2-Nitrophenol	9.769	139	275551	20.535	ng/ul	97
26) 2,4-Dimethylphenol	9.828	107	515923	19.742	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.063	93	622770	19.815	ng/ul	99
29) 2,4-Dichlorophenol	10.299	162	458106	19.693	ng/ul	99
30) Naphthalene	10.699	128	1582086	18.772	ng/ul	99
32) 4-Chloroaniline	10.816	127	636487	18.926	ng/ul	99
33) Hexachlorobutadiene	10.981	225	269000	18.649	ng/ul	99
34) Caprolactam	11.616	113	119513	18.697	ng/ul	95
35) 4-Chloro-3-methylphenol	11.945	107	448963	20.373	ng/ul	98
36) 2-Methylnaphthalene	12.310	142	1047488	19.301	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101422\
 Data File : BPCCC144.D
 Acq On : 14 Oct 2022 19:19
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 19 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :


Quant Time: Oct 14 23:07:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 12 23:03:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.534	142	1047432	19.272	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.681	216	523095	18.455	ng/ul	98
40) Hexachlorocyclopentadiene	12.657	237	233868	16.987	ng/ul	100
41) 2,4,6-Trichlorophenol	12.928	196	330752	19.372	ng/ul	99
42) 2,4,5-Trichlorophenol	12.998	196	357401	19.261	ng/ul	98
43) 1,1'-Biphenyl	13.322	154	1308817	18.698	ng/ul	100
44) 2-Chloronaphthalene	13.369	162	1042819	18.656	ng/ul	99
45) 2-Nitroaniline	13.581	65	244327	20.490	ng/ul	99
47) Dimethylphthalate	13.951	163	1182606	19.115	ng/ul	99
48) 2,6-Dinitrotoluene	14.081	165	222547	20.317	ng/ul	98
50) Acenaphthylene	14.216	152	1561585	19.006	ng/ul	99
51) 3-Nitroaniline	14.410	138	241500	18.934	ng/ul	100
52) Acenaphthene	14.557	153	1084591	18.543	ng/ul	99
53) 2,4-Dinitrophenol	14.616	184	108221	15.293	ng/ul	96
55) 4-Nitrophenol	14.722	109	127237	15.964	ng/ul	95
56) Dibenzofuran	14.892	168	1451503	18.362	ng/ul	98
57) 2,4-Dinitrotoluene	14.869	165	310900	19.749	ng/ul#	97
58) 2,3,4,6-Tetrachlorophenol	15.122	232	268680	18.577	ng/ul	100
59) Diethylphthalate	15.322	149	1111151	19.610	ng/ul	100
61) Fluorene	15.545	166	1153102	18.583	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.539	204	561972	18.721	ng/ul	99
63) 4-Nitroaniline	15.575	138	212086	18.210	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.634	198	152938	16.798	ng/ul	94
67) N-Nitrosodiphenylamine	15.757	169	942791	19.971	ng/ul	99
68) 4-Bromophenyl-phenylether	16.439	248	308547	19.880	ng/ul	97
69) Hexachlorobenzene	16.551	284	338031	18.765	ng/ul	99
70) Atrazine	16.716	200	271466	18.700	ng/ul	99
71) Pentachlorophenol	16.904	266	178273	16.891	ng/ul	100
72) Phenanthrene	17.298	178	1599088	18.436	ng/ul	99
74) Anthracene	17.386	178	1601450	18.657	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.287	216	519653	20.163	ng/ul	99
76) Pentachlorobenzene	14.810	250	504622	19.936	ng/ul	99
77) Carbazole	17.663	167	1342479	17.536	ng/ul	99
78) Di-n-butylphthalate	18.210	149	1514388	19.934	ng/ul	99
80) Fluoranthene	19.263	202	1506462	18.781	ng/ul	97
82) Pyrene	19.616	202	1560467	18.383	ng/ul	96
83) Butylbenzylphthalate	20.492	149	514456	19.119	ng/ul	92
84) 3,3'-Dichlorobenzidine	21.263	252	481499	17.101	ng/ul#	99
85) Benzo(a)anthracene	21.327	228	1529112	19.012	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.251	149	695949	19.185	ng/ul	99
87) Chrysene	21.380	228	1438679	18.583	ng/ul	100
89) Di-n-octyl phthalate	22.174	149	1213885	16.972	ng/ul	100
90) Benzo(b)fluoranthene	23.021	252	1660773	18.365	ng/ul	98
91) Benzo(k)fluoranthene	23.068	252	1641390	18.453	ng/ul	99
93) Benzo(a)pyrene	23.651	252	1502405	18.622	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.274	276	1875098	18.381	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.286	278	1688233	18.447	ng/ul	97
96) Benzo(g,h,i)perylene	27.051	276	1654641	17.990	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 14 23:07:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101122.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 12 23:03:21 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 5.792 ng/uL
RT: 3.305 min Scan# 54
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 88 Resp: 54999
Ion Ratio Lower Upper
88 100
43 18.5 16.6 25.0
58 58.9 52.6 79.0



#5
Pyridine
Concen: 17.402 ng/uL
RT: 3.717 min Scan# 124
Delta R.T. -0.000 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion: 79 Resp: 400117
Ion Ratio Lower Upper
79 100
52 56.3 51.2 76.8
51 26.3 22.8 34.2









#6
Benzaldehyde
Concen: 23.406 ng/ul
RT: 6.987 min Scan# 680
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 314224
Ion Ratio Lower Upper
77 100
105 103.4 85.2 127.8
106 101.0 83.1 124.7



#8
Phenol
Concen: 20.024 ng/ul
RT: 7.040 min Scan# 689
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion: 94 Resp: 601482
Ion Ratio Lower Upper
94 100
65 25.3 19.4 29.0
66 31.1 24.6 37.0









#10
Bis(2-Chloroethyl)ether
Concen: 20.277 ng/ul
RT: 7.269 min Scan# 71
Delta R.T. -0.012 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 93 Resp: 480572
Ion Ratio Lower Upper
93 100
63 71.5 56.7 85.1
95 32.0 26.2 39.2



#12
2-Chlorophenol
Concen: 20.174 ng/ul
RT: 7.405 min Scan# 751
Delta R.T. -0.012 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:128 Resp: 497276
Ion Ratio Lower Upper
128 100
64 42.3 32.7 49.1
130 32.4 26.3 39.5





#13

2-Methylphenol

Concen: 20.418 ng/ul

RT: 8.293 min Scan# 903

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19

Instrument :

BNA_P





ClientSampleId :

Tgt Ion:108 Resp: 457450

Ion Ratio Lower Upper

108 100

107 88.4 71.9 107.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 21.082 ng/ul

RT: 8.381 min Scan# 917

Delta R.T. -0.000 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19




Tgt Ion: 45 Resp: 512115


Ion Ratio Lower Upper

45 100

77 17.0 14.4 21.6

79 14.1 11.3 16.9









#16
Acetophenone
Concen: 20.960 ng/ul
RT: 8.675 min Scan# 967
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:105 Resp: 710201
Ion Ratio Lower Upper
105 100
77 72.5 57.0 85.4
51 21.3 16.4 24.6



#17
N-Nitroso-di-n-propylamine
Concen: 21.581 ng/ul
RT: 8.657 min Scan# 964
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion: 70 Resp: 333050
Ion Ratio Lower Upper
70 100
42 43.7 34.6 52.0
101 10.2 8.0 12.0
130 22.7 18.3 27.5









#18
4-Methylphenol
Concen: 20.483 ng/ul
RT: 8.622 min Scan# 91
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 494616
Ion Ratio Lower Upper
108 100
107 114.8 86.3 129.5



#19
Hexachloroethane
Concen: 20.552 ng/ul
RT: 8.922 min Scan# 1009
Delta R.T. -0.012 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:117 Resp: 195551
Ion Ratio Lower Upper
117 100
201 97.6 82.3 123.5
199 60.5 50.0 75.0









#22
Nitrobenzene
Concen: 20.068 ng/ul
RT: 9.052 min Scan# 1031
Delta R.T. -0.012 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 518184
Ion Ratio Lower Upper
77 100
123 48.5 39.4 59.0
65 12.2 9.9 14.9



#23
Isophorone
Concen: 20.673 ng/ul
RT: 9.587 min Scan# 1122
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion: 82 Resp: 946836
Ion Ratio Lower Upper
82 100
95 6.5 5.4 8.0
138 19.3 15.2 22.8









#25
2-Nitrophenol
Concen: 20.535 ng/ul
RT: 9.769 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:139 Resp: 275551
Ion Ratio Lower Upper
139 100
65 43.5 36.3 54.5
109 25.0 21.2 31.8



#26
2,4-Dimethylphenol
Concen: 19.742 ng/ul
RT: 9.828 min Scan# 1163
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:107 Resp: 515923
Ion Ratio Lower Upper
107 100
121 52.4 43.0 64.4
122 90.8 71.4 107.2









#27
Bis(2-Chloroethoxy)methane
Concen: 19.815 ng/ul
RT: 10.063 min Scan# 1203
Delta R.T. -0.012 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 93 Resp: 622770
Ion Ratio Lower Upper
93 100
95 32.6 25.7 38.5
123 11.5 9.2 13.8



#29
2,4-Dichlorophenol
Concen: 19.693 ng/ul
RT: 10.299 min Scan# 1243
Delta R.T. -0.012 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:162 Resp: 458106
Ion Ratio Lower Upper
162 100
164 65.0 51.5 77.3
98 34.3 26.3 39.5









#30
Naphthalene
Concen: 18.772 ng/ul
RT: 10.699 min Scan# 1311
Delta R.T. -0.012 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:128 Resp: 1582086
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 13.1 10.2 15.4



#32
4-Chloroaniline
Concen: 18.926 ng/ul
RT: 10.816 min Scan# 1331
Delta R.T. -0.012 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19


Tgt Ion:127 Resp: 636487
Ion Ratio Lower Upper
127 100
129 32.7 25.6 38.4








#33
Hexachlorobutadiene
Concen: 18.649 ng/ul
RT: 10.981 min Scan# 11
Delta R.T. -0.012 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:225 Resp: 269000
Ion Ratio Lower Upper
225 100
223 63.6 50.9 76.3
227 65.5 51.1 76.7



#34
Caprolactam
Concen: 18.697 ng/ul
RT: 11.616 min Scan# 1467
Delta R.T. -0.000 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:113 Resp: 119513
Ion Ratio Lower Upper
113 100
55 152.2 115.2 172.8
56 111.0 85.6 128.4









#35
4-Chloro-3-methylphenol
Concen: 20.373 ng/ul
RT: 11.945 min Scan# 1523
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:107 Resp: 448963
Ion Ratio Lower Upper
107 100
144 27.5 23.0 34.4
142 84.9 69.8 104.8



#36
2-Methylnaphthalene
Concen: 19.301 ng/ul
RT: 12.310 min Scan# 1585
Delta R.T. -0.012 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:142 Resp: 1047488
Ion Ratio Lower Upper
142 100
141 88.2 70.7 106.1





#37

1-Methylnaphthalene

Concen: 19.272 ng/ul

RT: 12.534 min Scan# 1623

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:142 Resp: 1047432

Ion Ratio Lower Upper

142 100

141 91.1 72.9 109.3

116 3.5 2.8 4.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 18.455 ng/ul

RT: 12.681 min Scan# 1648

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19

Tgt Ion:216 Resp: 523095




Ion Ratio Lower Upper


216 100

214 80.1 63.0 94.4

179 20.8 16.3 24.5

108 17.7 12.5 18.7









#40
Hexachlorocyclopentadiene
Concen: 16.987 ng/ul
RT: 12.657 min Scan# 1644
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:237 Resp: 233868
Ion Ratio Lower Upper
237 100
235 62.4 49.8 74.8
272 12.6 9.9 14.9



#41
2,4,6-Trichlorophenol
Concen: 19.372 ng/ul
RT: 12.928 min Scan# 1690
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:196 Resp: 330752
Ion Ratio Lower Upper
196 100
198 95.6 76.2 114.2
200 31.1 24.2 36.2









#42
2,4,5-Trichlorophenol
Concen: 19.261 ng/ul
RT: 12.998 min Scan# 1702
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:196 Resp: 357401
Ion Ratio Lower Upper
196 100
198 96.6 75.4 113.2
200 30.3 24.8 37.2



#43
1,1'-Biphenyl
Concen: 18.698 ng/ul
RT: 13.322 min Scan# 1757
Delta R.T. -0.012 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:154 Resp: 1308817
Ion Ratio Lower Upper
154 100
153 39.9 32.2 48.2
76 13.0 10.3 15.5









#44
2-Chloronaphthalene
Concen: 18.656 ng/ul
RT: 13.369 min Scan# 1765
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:162 Resp: 1042819
Ion Ratio Lower Upper
162 100
164 32.8 26.0 39.0
127 37.3 29.4 44.2



#45
2-Nitroaniline
Concen: 20.490 ng/ul
RT: 13.581 min Scan# 1801
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion: 65 Resp: 244327
Ion Ratio Lower Upper
65 100
92 74.5 61.6 92.4
138 120.8 96.5 144.7









#47
Dimethylphthalate
Concen: 19.115 ng/ul
RT: 13.951 min Scan# 1864
Delta R.T. -0.012 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:163 Resp: 1182606
Ion Ratio Lower Upper
163 100
194 3.8 3.1 4.7
164 10.2 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 20.317 ng/ul
RT: 14.081 min Scan# 1886
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:165 Resp: 222547
Ion Ratio Lower Upper
165 100
89 52.9 43.4 65.0
121 21.0 18.1 27.1









#50
Acenaphthylene
Concen: 19.006 ng/ul
RT: 14.216 min Scan# 1909
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:152 Resp: 1561585
Ion Ratio Lower Upper
152 100
151 20.1 15.7 23.5
153 13.1 10.4 15.6



#51
3-Nitroaniline
Concen: 18.934 ng/ul
RT: 14.410 min Scan# 1942
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:138 Resp: 241500
Ion Ratio Lower Upper
138 100
108 9.6 8.5 12.7
92 107.2 85.8 128.8









#52
Acenaphthene
Concen: 18.543 ng/ul
RT: 14.557 min Scan# 1967
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:153 Resp: 1084591
Ion Ratio Lower Upper
153 100
152 48.6 38.2 57.2
154 88.7 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 15.293 ng/ul
RT: 14.616 min Scan# 1977
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:184 Resp: 108221
Ion Ratio Lower Upper
184 100
63 62.4 45.5 68.3
154 59.0 47.0 70.6









#55
4-Nitrophenol
Concen: 15.964 ng/ul
RT: 14.722 min Scan# 1995
Delta R.T. -0.000 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:109 Resp: 127237
Ion Ratio Lower Upper
109 100
139 161.5 124.1 186.1
65 137.3 104.2 156.4



#56
Dibenzofuran
Concen: 18.362 ng/ul
RT: 14.892 min Scan# 2024
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:168 Resp: 1451503
Ion Ratio Lower Upper
168 100
139 36.3 28.2 42.4









#57
2,4-Dinitrotoluene
Concen: 19.749 ng/ul
RT: 14.869 min Scan# 2020
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:165 Resp: 310900
Ion Ratio Lower Upper
165 100
63 36.8 31.1 46.7
182 2.7 1.8 2.6#



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.577 ng/ul
RT: 15.122 min Scan# 2063
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:232 Resp: 268680
Ion Ratio Lower Upper
232 100
131 47.2 37.8 56.6
130 2.4 1.8 2.6
166 32.7 26.6 39.8





#59

Diethylphthalate

Concen: 19.610 ng/ul

RT: 15.322 min Scan# 2097

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:149 Resp: 1111151

Ion Ratio Lower Upper

149 100

177 21.8 17.5 26.3

150 12.4 9.8 14.8



#61

Fluorene

Concen: 18.583 ng/ul

RT: 15.545 min Scan# 2135

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19




Tgt Ion:166 Resp: 1153102


Ion Ratio Lower Upper

166 100

165 97.4 79.5 119.3

167 13.4 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 18.721 ng/ul
RT: 15.539 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 561972
Ion Ratio Lower Upper
204 100
206 32.4 25.4 38.2
141 56.6 46.2 69.2



#63
4-Nitroaniline
Concen: 18.210 ng/ul
RT: 15.575 min Scan# 2140
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:138 Resp: 212086
Ion Ratio Lower Upper
138 100
92 49.8 39.2 58.8
108 68.9 55.4 83.0









#66
4,6-Dinitro-2-methylphenol
Concen: 16.798 ng/ul
RT: 15.634 min Scan# 2150
Delta R.T. -0.000 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:198 Resp: 152938
Ion Ratio Lower Upper
198 100
182 5.7 4.9 7.3
77 21.6 20.0 30.0



#67
N-Nitrosodiphenylamine
Concen: 19.971 ng/ul
RT: 15.757 min Scan# 2171
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:169 Resp: 942791
Ion Ratio Lower Upper
169 100
168 67.4 53.5 80.3
167 37.2 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 19.880 ng/ul
RT: 16.439 min Scan# 21
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:248 Resp: 308547
Ion Ratio Lower Upper
248 100
250 95.4 77.9 116.9
141 70.5 59.5 89.3



#69
Hexachlorobenzene
Concen: 18.765 ng/ul
RT: 16.551 min Scan# 2306
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:284 Resp: 338031
Ion Ratio Lower Upper
284 100
142 37.0 30.5 45.7
249 32.0 25.4 38.0





#70

Atrazine

Concen: 18.700 ng/ul

RT: 16.716 min Scan# 21

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:200 Resp: 271466

Ion Ratio Lower Upper

200 100

173 27.7 22.1 33.1

215 48.0 38.8 58.2



#71

Pentachlorophenol

Concen: 16.891 ng/ul

RT: 16.904 min Scan# 2366

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19




Tgt Ion:266 Resp: 178273


Ion Ratio Lower Upper

266 100

264 63.2 50.6 75.8

268 64.5 51.8 77.6





#72

Phenanthrene

Concen: 18.436 ng/ul

RT: 17.298 min Scan# 2433

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:178 Resp: 1599088

Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.4

176 19.4 15.8 23.6



#74

Anthracene

Concen: 18.657 ng/ul

RT: 17.386 min Scan# 2448

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19




Tgt Ion:178 Resp: 1601450


Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 18.7 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.163 ng/uL
RT: 13.287 min Scan# 1
Delta R.T. -0.012 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:216 Resp: 519653
Ion Ratio Lower Upper
216 100
214 79.0 62.6 94.0
218 47.5 38.9 58.3



#76
Pentachlorobenzene
Concen: 19.936 ng/uL
RT: 14.810 min Scan# 2010
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:250 Resp: 504622
Ion Ratio Lower Upper
250 100
248 63.4 51.4 77.2
252 62.9 50.7 76.1





#77

Carbazole

Concen: 17.536 ng/ul

RT: 17.663 min Scan# 2495

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:167 Resp: 1342479

Ion Ratio Lower Upper

167 100

166 21.8 17.0 25.6

139 12.8 10.1 15.1



#78

Di-n-butylphthalate

Concen: 19.934 ng/ul

RT: 18.210 min Scan# 2588

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19




Tgt Ion:149 Resp: 1514388


Ion Ratio Lower Upper

149 100

150 9.2 7.2 10.8

104 4.6 3.4 5.2





#80

Fluoranthene

Concen: 18.781 ng/ul

RT: 19.263 min Scan# 21

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:202 Resp: 1506462

Ion Ratio Lower Upper

202 100

101 12.5 9.0 13.6

100 9.4 7.0 10.6



#82

Pyrene

Concen: 18.383 ng/ul

RT: 19.616 min Scan# 2827

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19




Tgt Ion:202 Resp: 1560467


Ion Ratio Lower Upper

202 100

101 14.9 10.4 15.6

100 12.1 8.5 12.7





#83

Butylbenzylphthalate

Concen: 19.119 ng/ul

RT: 20.492 min Scan# 2976

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:149 Resp: 514456

Ion Ratio Lower Upper

149 100

91 67.4 48.5 72.7

206 19.8 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 17.101 ng/ul

RT: 21.263 min Scan# 3107

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19




Tgt Ion:252 Resp: 481499


Ion Ratio Lower Upper

252 100

254 64.3 51.4 77.2

126 13.1 7.8 11.8#









#85
Benzo(a)anthracene
Concen: 19.012 ng/ul
RT: 21.327 min Scan# 3118
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:228 Resp: 1529112
Ion Ratio Lower Upper
228 100
229 20.0 15.7 23.5
226 28.1 21.9 32.9



#86
Bis(2-ethylhexyl)phthalate
Concen: 19.185 ng/ul
RT: 21.251 min Scan# 3105
Delta R.T. -0.000 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:149 Resp: 695949
Ion Ratio Lower Upper
149 100
167 26.6 21.9 32.9
279 3.8 3.4 5.2





#87

Chrysene

Concen: 18.583 ng/ul

RT: 21.380 min Scan# 3128

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 1438679

Ion Ratio Lower Upper

228 100

226 30.1 24.0 36.0

229 19.6 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 16.972 ng/ul




RT: 22.174 min Scan# 3262


Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19

Tgt Ion:149 Resp: 1213885









#90
Benzo(b)fluoranthene
Concen: 18.365 ng/ul
RT: 23.021 min Scan# 3406
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:252 Resp: 1660773
Ion Ratio Lower Upper
252 100
253 21.8 17.6 26.4
125 11.6 7.8 11.8



#91
Benzo(k)fluoranthene
Concen: 18.453 ng/ul
RT: 23.068 min Scan# 3414
Delta R.T. -0.006 min
Lab File: BPCCC144.D
Acq: 14 Oct 2022 19:19

Tgt Ion:252 Resp: 1641390
Ion Ratio Lower Upper
252 100
253 21.9 17.6 26.4
125 11.9 8.2 12.2





#93

Benzo(a)pyrene

Concen: 18.622 ng/ul

RT: 23.651 min Scan# 31

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 1502405

Ion Ratio Lower Upper

252 100

253 22.1 17.6 26.4

125 12.9 8.9 13.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.381 ng/ul

RT: 26.274 min Scan# 3959

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19




Tgt Ion:276 Resp: 1875098


Ion Ratio Lower Upper

276 100

138 29.8 19.1 28.7#

277 25.6 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 18.447 ng/ul

RT: 26.286 min Scan# 3961

Delta R.T. -0.012 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:278 Resp: 1688233

Ion Ratio Lower Upper

278 100

139 18.6 12.5 18.7

279 23.8 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 17.990 ng/ul

RT: 27.051 min Scan# 4091

Delta R.T. -0.006 min

Lab File: BPCCC144.D

Acq: 14 Oct 2022 19:19

Tgt Ion:276 Resp: 1654641

Ion Ratio Lower Upper

276 100

138 24.7 16.5 24.7

277 23.7 19.0 28.4

